

Electrochemical Immunosensor for Ultra-Low Detection of Human Papillomavirus Biomarker for Cervical Cancer

Siwaphiwe Peteni, Okoroike C. Ozoemena, Tobile Khawula, Aderemi B. Haruna, Frankie J. Rawson, Leshweni J. Shai, Oluwafunmilola Ola,* and Kenneth I. Ozoemena*



Cite This: *ACS Sens.* 2023, 8, 2761–2770



Read Online

ACCESS |



Metrics & More




Article Recommendations



Supporting Information

ABSTRACT: Human papillomavirus (HPV) is the causative agent for cervical cancer. Of the various types of HPV, the high-risk HPV-16 type is the most important antigenic high-risk HPV. In this work, the antigenic HPV-16 L1 peptide was immobilized on a glassy carbon electrode and used to detect several concentrations of the anti-HPV-16 L1 antibody, and vice versa. Two electrode platforms were used: onion-like carbon (OLC) and its polyacrylonitrile (OLC-PAN) composites. Both platforms gave a wide linear concentration range (1.95 fg/mL to 6.25 ng/mL), excellent sensitivity ($>5.2 \mu\text{A}/\log$ ([HPV-16 L1, fg/mL]), and extra-ordinarily low limit of detection (LoD) of 1.83 fg/mL (32.7 aM) and 0.61 fg/mL (10.9 aM) for OLC-PAN and OLC-based immunosensors, respectively. OLC-PAN modified with the HPV-16 L1 protein showed low LoD for the HPV-16 L1 antibody (2.54 fg/mL, i.e., 45.36 aM), proving its potential use for screening purposes. The specificity of detection was proven with the anti-ovalbumin antibody (anti-OVA) and native ovalbumin protein (OVA). An immobilized antigenic HPV-16 L1 peptide showed insignificant interaction with anti-OVA in contrast with the excellent interaction with anti-HPV-16 L1 antibody, thus proving high specificity. The application of the immunosensor as a potential point-of-care (PoC) diagnostic device was investigated with screen-printed carbon electrodes, which detected ultra-low (ca. 0.7 fg/mL \approx 12.5 aM) and high (ca. 12 $\mu\text{g}/\text{mL}$ \approx 0.21 μM) concentrations. This study represents the lowest LoD reported for HPV-16 L1. It opens the door for further investigation with other electrode platforms and realization of PoC diagnostic devices for screening and testing of HPV biomarkers for cervical cancer.

KEYWORDS: onion-like carbon, polyacrylonitrile fiber, antigenic HPV-16 L1 peptide, anti-HPV-16 L1 antibody, ovalbumin protein, anti-ovalbumin antibody, ultra-low detection



According to WHO statistics for 2018, cervical cancer alone has killed about 375,000 women and girls out of 500,000 cases, most of these cases are from Asia, Latin America, and sub-Saharan Africa.¹ Although vaccination is one of the preventative ways in which one may follow, there is a problem with the rate of uptake, and it is also advisable to be vaccinated at a younger age for more effectiveness. Human papillomaviruses (HPV) are the causative agents of cervical, anal, and vaginal cancers to name a few. The human papillomavirus L1 protein (HPV-L1) is found in 90% of the HPV capsid and is known to be directly involved in the process of the HPV infection of host cells.² These viruses are grouped as low-risk or high-risk depending on the type of lesion they cause. The low-risk HPVs are known to be benign and, in most cases, can be cleared by the body without any grievous harm, while the high-risk HPVs are associated with malignancy, which is responsible for the cancers mentioned above.³ Of the 14 known HPV types, the high-risk HPV-16 and -18 are mostly associated with cancer of the cervix. HPV-16 and -18 are responsible for the encoding of two oncogenes, E6 and E7,

which are responsible for the progression to cancer.^{4–6} Both high-risk HPV-16 and HPV-18 are thought to be responsible for 93–100% of cervical cancer cases,⁷ with HPV-16 being the most dominant viral strain that affects the cervix, followed by the HPV-18 strain.

The persistence of HPV infection leads to the so-called cervical intraepithelial neoplasia (CIN). If detected early (i.e., when the dysplastic cells are confined within the surface epithelium of the cervix), the CIN can easily be cured. If not detected at the early stage, they can penetrate the basement of the membrane to become invasive cervical cancer and spread into the nearby organs, e.g., the uterus, bladder, rectum, and the pelvic lymph nodes, thereby causing death. The mortality

Received: April 6, 2023
Accepted: June 13, 2023
Published: June 29, 2023






Figure 1. TEM images of (A) OLC and (B) OLC-PAN, SEM images of (C) OLC and (D) OLC-PAN, (E) XRD patterns of powdered OLC and OLC-PAN, and (F) Raman spectra of OLC and OLC-PAN.

rate of cervical cancer in the poorest countries has been estimated as twofold higher than for the wealthy countries.¹ In many low- and middle-income earning countries (LMICs), no organized cervical cancer screening programs exist. LMICs bear the largest burden of human immunodeficiency virus (HIV) infection while persistent high-risk HPV infection is more common among HIV-infected women. Thus, the risk of cervical cancer is increased in women with HIV/AIDS. This underscores the urgent need for the development of low-cost diagnostic devices for HPV.

There are several diagnostic methods for the detection of HPV,⁸ which include Papanicolaou (Pap) smear (aka cervical or vaginal cytology), direct visual inspection of the cervix with acetic acid (VIA) or iodine (VILLI), polymerase chain reaction (PCR), enzyme-linked immunosorbent assay (ELISA), radioimmunoassay (RIA), immunohistochemistry (IHC), and flow cytometry (FC). These methods suffer from several shortcomings. For example, the Pap smear which is the most frequently used diagnostic method has reduced the mortality rate of cervical cancer by up to 50%; however, it is fraught with poor sensitivity (between 50 and 60%).⁹ Pap smear is not sustainable in under-resourced LMIC settings with low socioeconomic and educational level, limited skilled cyto-screener and cytopathology workforce¹⁰ and where, despite a high prevalence of cervical cancer, lack of follow-up and poor adherence to treatment are major impediments for program success. The low-cost VIA and VILLI methods require a trained workforce to deliver and are characterized by poor specificity, which can result in the substantial over-treatment of patients. In addition, it is very difficult to detect small ectocervical and endocervical lesions under visual inspection so, with a once-in-a-lifetime screen or wide screening intervals, pre-cancer can be missed and progress to invasive cancer. PCR, ELISA, RIA, IHC, and FC are efficient but have the disadvantages that they are invasive, expensive, time-consuming, and bulky (i.e., limited to large hospitals).

Considering that the Pap smear method is the most employed technique for the detection of HPV infection in the LMICs, there is a need to design and develop a diagnostic method that complements it, but with advantages of better sensitivity and selectivity (since the level of antibodies in both symptomatic and asymptomatic patients is generally very low), less invasive, and easier to use (preferably self-use) to allow for increased coverage in the communities of the LMICs. Electrochemical detection methods meet the above criteria: they are characterized by their simplicity, high sensitivity, and selectivity and, importantly, electrochemical sensors can be miniaturized (smaller devices) for hand-held, point-of-care (PoC) diagnostic testing and screening devices that require little training for operation in a given community setting.

Today, most of the electrochemical methods of detecting HPV have been focused on nucleic acid (DNA) sensors.^{11–14} To our knowledge, the use of protein biomarkers for the electrochemical detection of HPV is largely unknown. There are only a few reports that attempted the utilization of HPV protein biomarkers.^{13,15–18} For example, Piro et al.¹⁵ reported the use of a very complex conjugated electro-co-polymer to encapsulate the antigenic HPV-16 L1 on a glassy carbon electrode (GCE) and used it to detect just a single concentration of the anti-HPV-16 L1 antibody. Second, Valencia et al.¹⁶ reported the detection of anti-peptide antibodies of HPV-L1 by using a gold electrode modified with a peptide (SPINNTKPHEAR) derived from the HPV-L1, a technique that involved a lot of preparation steps and only tested or suited for the detection of anti-HPV in serum samples (screening). Recently, Wang et al.¹⁷ used quasi-spherical Ag@Au core-shell nanoparticles on graphene oxide (Ag@AuNPs-GO) as the current response amplifier for the detection of HPV-16 L1. To close the knowledge gap, this work used recombinant HPV-16 L1 protein (HPV-16) and anti-HPV16 L1 antibody (anti-HPV) for the fabrication of an electrochemical immunosensor with the potential application in the




Figure 2. (A) Wide scan spectrum of OLC-PAN, and deconvoluted spectra of (B) C1s, (C) O1s, and (D) N1s of OLC-PAN.

screening (i.e., for asymptomatic people) and diagnostic tests (i.e., for symptomatic people) of HPV-16. It is known that the HPV-L1 protein is the dominant epitope (present in 90% of the HPV capsid) directly involved in the process of HPV infection in patients as well as in the prophylactic vaccines, thus the use of the high-risk HPV-16 and anti-HPV-16 is of high importance. We show that polyacrylonitrile (PAN) fiber modified with onion-like carbons (OLC-PAN) nanocomposite represents a viable electrode platform for the encapsulation of the anti-HPV-16 antibodies for the detection of HPV-16 antigen (i.e., testing), and for the encapsulation of the HPV-16 antigen for the detection of the anti-HPV-16 antibody (i.e., screening). OLC is a relatively new carbonaceous material that has been applied as an excellent electrode modifier for electrocatalysis,^{19,20} electrochemical sensing,^{21–23} and electrochemical energy storage^{24,25} due to its attractive properties of high specific surface area, high conductivity, and electrocatalytic properties. In addition, PAN is easy to synthesize and demonstrates good absorptivity for pollutants.²⁶ Although it is a poor conductor, it has shown better electrochemistry when modified with OLC,²³ such that its hybrid has better properties than the individual materials. In addition, from a previous study,²³ it was found that although OLC has high conductivity and high specific surface area, it showed poor electrochemical responses hence a hybrid material was synthesized and tested.

RESULTS AND DISCUSSION

SEM and TEM Characterization. The physicochemical characterizations such as TEM, SEM, XRD, Raman, and surface analysis are discussed in this section. HRTEM images are shown in Figure 1A for OLC and Figure 1B for OLC-PAN. The HRTEM image of OLC is interconnected graphitic layers, which resembles onion-like rings with the outside layers connected to form an oval-like shape, while the OLC-PAN does not show any distinct shape. The SEM images of OLC and OLC-PAN are shown in Figure 1C,D, respectively. The

morphology of OLC is agglomerates of nanoparticles, while the OLC-PAN shows rough and smoothed surfaces, the roughness of the PAN surface confirms the successful incorporation of the OLC into the PAN matrix.

X-ray Diffraction, Raman Spectroscopy, and Specific Surface Area Analysis. Figure 1E compares the powder XRD patterns of the OLC and OLC-PAN. The OLC shows two distinct peaks at 2θ of ~ 26.4 and 44.0° corresponding to the amorphous (i.e., (002) plane) and crystalline graphite (i.e., (101) plane), respectively. The OLC-PAN, on the other hand, shows the same peaks as the OLC, but with an extra peak at $2\theta = 17.1^\circ$ that corresponds to the (100) plane of the polymer. The XRD of the OLC-PAN confirms the presence of the OLC. Figure 1F compares the Raman spectra of the OLC and OLC-PAN. Both OLC and OLC-PAN show the two characteristic distinct D (defect) bands due to the out-of-plane vibrations of sp^3 -bonded carbons, and G (graphitic) band arising from the in-plane vibrations of the sp^3 -bonded carbons. The D and G bands were observed at 337.4 and 1593 cm^{-1} for the OLC, respectively, while those of the OLC-PAN appear at 1347.9 and 1598.2 cm^{-1} , respectively. The degree of graphitization was determined using the conventional I_D/I_G ratio. The I_D/I_G ratio of OLC-PAN was 1.97 compared to the OLC of 1.95, indicating that the OLC-PAN is slightly more defective than the OLC alone.

The BET surface area, pore volume, and pore size were calculated as $279.05\text{ m}^2\text{g}^{-1}$, $1.20\text{ cm}^3\text{g}^{-1}$, and 17.11 nm , respectively for OLC, and as $117.69\text{ m}^2\text{g}^{-1}$, $0.54\text{ cm}^3\text{g}^{-1}$, and 22.89 nm , respectively, for OLC-PAN. The decrease in the surface area of OLC-PAN compared to the OLC should be expected considering the bulky nature of the OLC-PAN as evident from the SEM and TEM images.

X-ray Photoelectron Spectroscopy. Figure 2 shows the XPS wide scan spectrum of the OLC-PAN (Figure 2A) with the expected peaks for C1s, O1s, and N1s at binding energies of 286, 397, and 536 eV, respectively. The deconvoluted




Figure 3. Typical fabrication pathway for HPV-16 L1 immunosensors. The reagents in each path are identified in the [Experimental Section](#).

spectra are shown for C 1s (Figure 2B), O 1s (Figure 2C), and N 1s (Figure 2D). The C1s peak was deconvoluted to two binding energies at 283 eV ($sp^3(C-C)$, defects) and 284.6 eV ($C-C$). It is to be noted that the $sp^3(C-C)$ and $sp^2(C-C)$ defect peaks are characteristics of the nanodiamonds,^{27,28} and the correct peak assignment has been a subject of controversy; most researchers believe that the peak position of sp^2C is lower than that of the sp^3C , few workers believe the opposite.²⁹ Interestingly, however, it has recently been demonstrated (via calculations) that the peak position of sp^2C is greater than that of the sp^3C , which agrees with the experimental results of those few workers.²⁹ According to these workers,²⁹ the reasons for the ambiguous assignments could be related to charging effects and defects (such as pentagons, heptagons, and functional groups) present in diamonds. The values of the $sp^3(C-C)$ range between 283.15 and 283.21 eV,²⁹ and even at 282.8 eV³⁰ which are in close agreement with our value of 283 eV. Importantly, considering that the OLC used in this study was obtained from detonation nanodiamonds (DNDs), we believe that the observed peak at 283 eV ($sp^3(C-C)$) could have arisen from the unconverted DNDs present in the OLC. Similarly, the O1s peak was deconvoluted with two binding energies at 531.2 eV ($C=O$) and 529.8 eV ($C-O$) (Figure 2C). These peak assignments agree with previous reports.^{31,32} In addition, the N 1s was deconvoluted to one peak at 398 eV corresponding to a cyanide ($C=N$) bond arising from the PAN component of the composite structure. Note that the two binding energies of the O 1s peak as well as that of the N 1s are slightly shifted to lower values, but this can be attributed to electron transfer from the PAN to the oxygenated functional groups of the OLC.

Electroanalytical Performance of the Immunosensors Toward HPV-16 L1 Detection. The fabrication of the HPV-16 L1 immunosensors and antigen–antibody interaction mechanisms are described in the [Experimental Section](#) but summarized as shown in Figure 3. The HPV antigen or

antibody is immobilized onto the OLC- or OLC-PAN electrode via the conventional covalent bonding process through the formation of the strong amide bond using the sulfo-NHS ester.^{23,33–35} In practice, route A describes the screening of HPV-16 infection (asymptomatic individuals), while route B is for the testing of the HPV-16 infection (in individual who presents some symptoms).

Cyclic voltammetric evolutions of the bare GCE, the electrochemical immunosensors (EI (1) and EI (2), Figure 3), and their corresponding antigen–antibody interaction stages (AA(I) and AAI (2), Figure 3) were conducted in the redox probe solution (i.e., PBS/AE, pH 7.4, containing 0.1 mM $[Fe(CN)_6]^{4-}/[Fe(CN)_6]^{3-}$) (see Figure S1). It was observed that the OLC-PAN-immunosensor exhibited the strongest current suppression (capacitive behavior) upon interaction with the analytical HPV-16 antigen than the OLC-based counterpart. The difference between the two immunosensor electrodes may be related to the conductivity of the electrode platforms, where OLC shows higher electronic conductivity than the OLC-PAN counterpart.

The developed electrochemical immunosensors were employed to detect HPV-16 L1 antigen at different concentrations (Figure 4). The concentration studies were conducted by exposing/immersing the immunosensor in different concentrations of the HPV-16 L1 antigen (from 1.95×10^{-12} to 5.0×10^{-2} mg/mL) for 5 min to allow for the complexation to occur, after which square wave voltammetry (SWV) was used to monitor the reaction in the presence of the redox probe. The continuous suppression of the redox peaks of the redox probe upon increasing concentration of the HPV-16 L1 antigen is indicative of the excellent complexation between Ab and Ag. The OLC-based immunosensor (Figure 4A) showed a broader peak response compared to its OLC-PAN counterpart (Figure 4C). Aside from this disadvantage of the OLC-based electrode platform, the OLC-PAN-based immunosensor also showed better linearity than the OLC-based




Figure 4. Square wave voltammetric detection. (A) Concentration studies of antigenic HPV-16 L1 protein (Ag) on OLC-based immunosensor and (B) corresponding linear plot of the changes in current vs $\log[\text{HPV-16 L1}]$; (C) concentration studies of antigenic HPV-16 L1 protein (Ag) on OLC-PAN-based platform and (D) corresponding linear plot of the changes in current vs $\log[\text{HPV-16 L1}]$; (E) concentration studies of anti-HPV 16 L1 antibodies on OLC-PAN-based immunosensor and (F) corresponding linear plot of the changes in current vs $\log[\text{anti-HPV 16 L1}]$. Please see the [Experimental Section](#) for details.

immunosensor (i.e., $R^2 = 0.9971$ cf 0.9823). Thus, the OLC-PAN electrode was used to immobilize the HPV-16 L1 antibody for the detection of the HPV-16 L1 antigen. The current response for each concentration was determined by taking the difference from the blank, i.e., ($\Delta i / \mu\text{A}$). For both immunosensors, two linear concentration ranges (Figure 4B,C, for OLC and OLC-PAN-based immunosensors, respectively) were observed, at low (i.e., 1.95×10^{-12} – 6.25×10^{-6} mg/mL) and high (6.25×10^{-6} – 5.0×10^{-2} mg/mL) concentration ranges.

Considering that the OLC-PAN-based electrode platform gives a sharper current response The results are summarized as follows:

OLC-Based Immunosensors for HPV-16 Antigen Detection.

$$\begin{aligned} \Delta i / \mu\text{A} &= 5.406 \pm 0.2479 \log([\text{HPV} - 16 \text{ L1}, \text{ fg/mL}]) \\ &+ 1.175 \pm 1.0366 (R^2) \\ &= 0.9917 \end{aligned} \quad (1)$$

OLC-PAN-Based Immunosensors for HPV-16 Antigen Detection.

$$\begin{aligned} \Delta i / \mu\text{A} &= 5.232 \pm 0.1740 \log([\text{HPV} - 16 \text{ L1}, \text{ fg/mL}]) \\ &+ 16.3757 \pm 0.7278 (R^2) \\ &= 0.9956 \end{aligned} \quad (2)$$

OLC-PAN-Based Immunosensors for HPV-16 Antibody Detection.

Table 1. Comparison of Various Electrochemical and Non-electrochemical Detection Methods for HPV16 L1^a

platform/modifier	target	technique	LCR	LoD	LoQ	ref
GCE-OLC	HPV-16 L1 (antigen)	SWV	1.95 fg/mL–6.25 ng/mL (34.2 aM–0.112 nM)	1.83 fg/mL (32.7 aM)	6.11 fg/mL (0.11 fM)	this work
GCE-OLC-PAN	HPV-16 L1 (antigen)	SWV	1.95 fg/mL–6.25 ng/mL (34.2 aM–0.112 nM)	0.61 fg/mL (10.89 m aM)	2.03 fg/mL (36.3 aM)	this work
GCE-OLC-PAN	HPV-16 L1 (antibody)	SWV	39.1 fg/mL–60 μg/mL (0.70 fM–1.07 μM)	2.54 fg/mL (45.36 aM)	8.46 fg/mL (0.15 fM)	this work
prGO-MoS ₂	HPV-16 L1 (antigen)	DPV	3.5–35.3 pM	1.75 pM		13
GC-polymer	HPV-16 L1 (antigen)	SWV		50 nM		15
Ag@AuNPs-GO/SPA	HPV-16 L1 (antigen)	DPV	0.005–400 ng/mL	2.0 pg./mL		17
Pt/PaN/MWCNTs	HPV-16 L1 (antigen)	CV/SWV	10–80 nM	490 pM		18
melamine-AuNPs	HPV-16 L1 (antigen)	LDI MS	2–80 ng /mL	58.8 pg/mL		40
AuNPs-Aptamer	HPV-16 L1 (antigen)	UV-vis/colorimetry	9.6–201.6 ng /mL	9.6 ng/mL		41

^aKEY: Molar mass of HPV-16 L1 (56 kDa) was used to convert the gravimetric concentrations to their corresponding molar concentrations. LCR (linear concentration range); LoD (limit of detection); LoQ (limit of quantification); PaN (polyaniline); AuNPs (gold nanoparticles); SPA (staphylococcal protein A); LDI MS (laser desorption ionization mass spectrometry); SWV (square wave voltammetry); and DPV (differential pulse voltammetry).




Figure 5. (A) Square wave voltammetric responses of GCE-OLC-PAN-OVAAb-BSA toward OVA-antigen (7.8×10^{-15} fg/mL) and HPV-16 L1 antigen (7.8×10^{-15} fg/mL), (B) Bar chart describing the responses of HPV-16 L1 antibody toward HPV-16 L1 (7.8×10^{-15} fg/mL) before (initial) and after six regeneration steps (reg 1 – reg 6) on day 1 and day 7. At every regeneration step, the detection of HPV-16 L1 antigen was repeated five times.

$$\begin{aligned} \Delta i / \mu\text{A} &= 1.2518 \pm 0.0533 \log([\text{HPV} - 16 \text{ L1}, \text{ fg/mL}]) \\ &+ 6.3744 \pm 0.3519 (R^2 \\ &= 0.9910) \end{aligned} \quad (3)$$

As summarized in Table 1, the immunosensors exhibited wide linear concentration ranges (1.95 fg/mL to 6.0 ng/mL, i.e., 34.82 aM to 0.11 nM), excellent sensitivity at the low concentration range for the detection of antigen (i.e., $5.406 \pm 0.2479 \mu\text{A}/\log([\text{HPV-16 L1}, \text{ fg/mL}])$ for the OLC-based immunosensor, and $5.232 \pm 0.1740 \mu\text{A}/\log([\text{HPV-16 L1}, \text{ fg/mL}])$ for the OLC-PAN-based immunosensor). For the detection of the antibody at OLC-PAN-based immunosensor, the sensitivity is lower than observed for the antigen (i.e., $1.2518 \pm 0.0533 \mu\text{A}/\log([\text{HPV-16 L1}, \text{ fg/mL}])$). The immunosensors showed ultra-low detection limits (i.e., LoD = $3 s/m$, and LoQ = $10 s/m$, where s is the standard deviation of the intercept, while m is the slope/sensitivity of the calibration plot). The values were calculated as the average of six electrodes and are summarized in Table 1, with LoD as ca. 0.61 fg/mL (10.89 aM) and 1.83 fg/mL (32.7 aM) for OLC-PAN and OLC-based immunosensors, respectively. Using the OLC-PAN-based immunosensor, the HPV-16 L1 antibody was

detected at 2.54 fg/mL (45.36 aM). The extraordinary low detection limits shown in this work represent the first of its kind for the detection of HPV-16 L1 using either immunosensor or nucleic acid-based sensors (Table 1).

For possible clinical application, we can compare with the conventional liquid-based pap test (LBPT) (which is still the most easily accessible technique in rural areas of resource-limited countries, or LMICs) and the more advanced PCR-based techniques. In LBPT, the specimen sample is rinsed into a sample container, the liquid is placed onto a microscope slide and then examined under a microscope to check if the cells are normal or not. The detection limit of LBPT is about 2000 copies/mL³⁶ (i.e., 4.04 fg/mL). There is a commercial APTIMA HPV assay that detects the mRNA of high-risk HPV including HPV 16 L1 was reported to be able to detect 106 mRNA copies/reaction (i.e., ca. 0.21 fg). The real-time PCR (RT-PCR) is the most important molecular technique for testing high-risk HPV types in clinical laboratories, especially in high-income settings and advanced countries. In addition, recombinase polymerase amplification (RPA) represents an isothermal alternative to the PCR. Previously, Hesselink et al.³⁷ showed that the HPV-risk assay (i.e., a type of real-time PCR assay) used for testing clinical samples gave LoD of 460

copies/reaction (ca. 0.93 fg) for the HPV-16 L1 genotype). In addition, in 2023, Wongsamart et al.³⁸ reported an advanced form of RPA known as the multiplex recombinase polymerase amplification (mRPA), which gave an LoD of 1000 copies/reaction (i.e., 2.01 fg) for the HPV-16 L1 genotype. The LoD values of our proposed electrochemical immunosensor are lower than the conventional LBPT, and comparable or even lower than the advanced PCR and mRPA techniques. It should be noted, however, that we could not subject our immunosensors to testing real clinical samples at this time due to the inherent difficulty of obtaining ethical clearance for human samples, but efforts are being made to overcome this in the future.

Specificity Studies and Sensor Reusability. To test for specificity, we adopted the use of an anti-ovalbumin antibody (anti-OVA) and native ovalbumin protein (OVA) as reported elsewhere.¹⁵ The anti-OVA (Ab) was immobilized as prescribed during the experimental session and the anti-OVA immunosensor. The constructed immunosensor (GCE-OLC-PAN-OVAAb-BSA) was used to compare the extent to which the OVA-antibody can interact with the corresponding OVA-antigen and HPV-16 L1 antigen of similar concentration (7.8×10^{-15} fg/mL). As shown in Figure 5A, the OVA-antibody exhibited very strong interaction with the OVA-antigen but with weak or no interaction with the HPV-16 L1 antigen. The rejection of the OVA-ab by the HPV-16 L1 antigen confirms the specificity of the HPV-16 L1 antigen with its corresponding antibody.

To test for reusability/reproducibility of the proposed immunosensor, we adopted the well-established procedure of glycine chemistry.³⁹ In a nutshell, the ability to reuse OLC-PAN-based immunosensor after chemically stripping off its active HPV-16 L1 antibody was interrogated using a constant concentration of the HPV-16 L1 antigen (7.8×10^{-15} fg/mL). On the first day, the immunosensor was used to detect the antigen, thereafter the bound antigen was stripped off by dipping the immunosensor into a buffer solution of glycine HCl (pH 2.8) for 5 min and then reused to detect the HPV-16 L1 antigen. The detection–stripping cycle was carried out six different times with each step used to detect the antigen five times. After the repetitive regeneration-detection steps, the immunosensor electrode was stored in the refrigerator at 4 °C, and the process was repeated on the 7th day.

As shown in Figure 5B, the immunosensor was found to be reusable even on the 7th day, albeit with minor loss in the current response compared to the 6th regeneration-detection step on the first day.

Proof of Concept. The possible practical application of the HPV immunosensor was tested using a screen-printed carbon electrode (SPCE) modified with the HPV-16 L1 antibody (Figure 6A,B). Although this was preliminary work, it is evident from Figure 6C that the immunosensor can detect very low (6.98×10^{-13} mg/mL) and high (1.15×10^{-2} mg/mL) concentrations of the antigenic HPV-16 L1. This result promises, upon thorough optimization, to move from lab-based techniques to PoC diagnosis of HPV infections.

CONCLUSIONS

OLC and OLC-PAN were successfully utilized as electrode platforms for the detection of HPV-16 L1 antigen and antibody. Both platforms exhibited a wide linear concentration range (1.95 fg/mL to 6.25 ng/mL) and ultra-low detection of 1.83 fg/mL (32.7 aM) and 0.61 fg/mL (10.9 aM) for OLC-




Figure 6. (A, B) Set-up of a screen-printed carbon electrode (SPCE) modified with HPV-16 L1 antibody immunosensor, (C) typical square wave voltammetric responses toward antigenic HPV-16 L1 protein. Please see the Experimental Section for details.

PAN and OLC-based immunosensors, respectively. The high specificity of detection was proven by experimenting with an anti-ovalbumin antibody (anti-OVA) and native ovalbumin protein (OVA). An immobilized antigenic HPV-16 L1 peptide showed insignificant interaction with anti-OVA in contrast with the excellent interaction with the anti-HPV-16 L1 antibody. The sensor's reusability was also tested, and a couple of runs were done, which showed that even after using the sensor several times, the sensor was still active although there was a slight decrease in the current response. The immunosensor still showed some activity after the 7th day of storage upon surface regeneration. The application of the immunosensor as a potential PoC diagnostic device was investigated with the SPEC, which showed the ability to detect ultra-low (ca. 0.7 fg/mL \approx 12.5 aM) and high (ca. 12 μ g/mL \approx 0.21 μ M) concentrations. This work represents the lowest detection limit ever reported for HPV-16 L1 using electrochemical and non-electrochemical methods. The results open the door of opportunity for studying other electrode platforms and realizing PoC diagnostic devices for screening and testing of HPV biomarkers for cervical cancer.

EXPERIMENTAL SECTION

Materials. *N,N* dimethyl formamide (DMF) was purchased from Merck. Ethanol and Nafion were purchased from Sigma Aldrich. Mono-potassium phosphate (KH_2PO_4), sodium chloride (NaCl), disodium phosphate (Na_2HPO_4), sodium hydroxide (NaOH), ethyl-3-(3-dimethylaminopropyl) carbodiimide (EDC), *N*-hydroxysuccinimide (NHS), ferro(IV) cyanide ($\text{K}_4[\text{Fe}(\text{CN})_6] \cdot 3\text{H}_2\text{O}$), and ferri(III) cyanide ($\text{K}_3[\text{Fe}(\text{CN})_6] \cdot 6\text{H}_2\text{O}$) and were purchased from Sigma Aldrich. Potassium chloride was purchased from Merck. The GCE was purchased from BASi. The carbon screen-printed electrodes were purchased from Metrohm. Recombinant HPV-16 L1 protein (HPV-16, 56 kDa), anti-HPV16 L1 antibody [Cam Vir 1] (anti-HPV-16, 56 kDa), anti-ovalbumin antibody (anti-OVA), and native ovalbumin protein (OVA) were purchased from Celtic molecular diagnostics (Pty) Ltd. Millipore water was obtained from Milli-Q Water Systems (Millipore Corp. Bedford, MA, USA). The saline buffer used was prepared at pH 7.4. OLC-PAN was synthesized according to the literature.²²

Instrumentation. The microstructures were analyzed by scanning electron microscopy (SEM, FEI Nova Nanolab 600). Elemental distribution was determined by energy-dispersive X-ray spectroscopy on an FEI Nova Nanolab 600. The crystal structures of the synthesized catalysts were investigated by powder X-ray diffractom-

eter (D2 Phaser in Bragg–Brentano configuration equipped with a sealed Co $K\alpha$ (\AA) radiation tube and a secondary beam of Fe $K\beta$ filter, Bruker Lynxeye PSD detector, with primary and secondary Soller slits) in the 2θ range of 5 – 90° . Electrochemical experiments were performed with SP300 Potentiostat (BioLogic Science instrument running on EC-Lab software).

Electrochemical Analysis. All electrochemical experiments were performed with bare or modified GCE (diameter = 3.0 mm, BASi) as the working electrode, platinum (Pt) rod as a counter electrode and the potential was measured against the Ag/AgCl electrode (saturated with 3 M KCl) used as the reference electrode. The GCE was first cleaned by polishing with alumina (Al_2O_3) slurry (50 nm powder), washed with ultrapure water, and sonicated in ethanol to remove slurry residues prior to use. Inks were prepared by mixing 1 mg of the electrocatalysts in 1 mL of DMF, and 20 μL of Nafion solution (5 wt %) and were ultrasonicated for 1 h to form a dispersed solution. The pre-cleaned GCE was then coated with 10 μL of the ink suspension and left to dry in an oven for 15 min at 30°C . Every solution used in this work was obtained using ultrapure water (18.2 M Ω cm resistivity). Phosphate buffer solution containing a small amount of sodium azide (as a preservative) and ethylenediaminetetraacetic acid (PBS/AE, pH 7.4) was prepared as previously described^{42,43} and used to prepare the solutions of the HPV-16 L1 antigen and antibody. Redox probe solution of 0.1 mM $\text{K}_4\text{Fe}(\text{CN})_6/\text{K}_3\text{Fe}(\text{CN})_6$ (1:1 mixture) in PBS/AE (pH 7.4) was used for investigating the electrochemical performance of the fabricated immunosensor electrodes.

Synthesis of OLC and OLC-PAN. *Synthesis of OLC.* OLC was synthesized from nanodiamond. Briefly, the nanodiamonds were placed in a sealed cylindrical graphite crucible and were thermally annealed in a water-cooled high-temperature vacuum furnace comprising tungsten heaters. The heating and the cooling rates were $15^\circ\text{C}/\text{min}$, and the chamber pressure ranged between 10 and 100 mPa. To get the final OLC, further annealing at 1300°C for 3 h under argon at 1 L/min was conducted.

Synthesis of OLC-PAN. In two separate conical flasks, 2 g of OLC and PAN were dispersed in 15 mL of DMF, respectively, after which the two solutions were mixed and stirred for 2 h at ambient temperature to form a uniform solution. The solution was further sonicated for 20 min to obtain a homogeneous mixture. The resulting polymer solution was fed into a syringe for electrospinning purposes. For electrospinning, the feed rate was 0.4 mL/h and 15 cm distance from the tip to the collector, while the potential difference between the collector and the grounded plate was 10 kV. The collected fibers were put in water overnight, to remove DMF and were then dried at 60°C for 2 h in an oven.

Fabrication of the Immunosensor. Figure 3 summarizes the fabrication of the sensor. GCE was modified using the drop casting method following a reported method.^{25,33,34} Briefly, 10 μL of electrocatalyst was drop-casted onto GCE and was dried in an oven at 45°C for 30 min, and this was labeled as GCE-OLC-PAN. The GCE-OLC-PAN was functionalized with 3 M NaOH for 2 h, the electrode was washed with Millipore water and PBS, after which the functionalized electrode was dried with N_2 gas, to afford a carboxylic acid functionalized surface (GCE-OLC-PAN-COOH). The COOH functional groups were activated using EDC and NHS pair (1:1% v/v) for 2 h with the above-stated conditions. After the activation of the COOH moiety, the HPV-16 L1 antibody containing NH_2 groups was allowed to interact with the activated GCE-OLC-PAN-COOH overnight to afford a GCE-OLC-PAN-Ab. BSA was then used to block unreacted sites to afford the GCE-OLC-PAN-Ab-BSA sensor. The electrode was stored at 4°C when not used. The screen-printed electrode was modified using the procedure described in Figure 3.

■ ASSOCIATED CONTENT

SI Supporting Information

The Supporting Information is available free of charge at <https://pubs.acs.org/doi/10.1021/acssensors.3c00677>.

SAED, TEM images, TGA, Raman, coulombic efficiency and cycling stability, and the DFT determined structure (PDF)

■ AUTHOR INFORMATION

Corresponding Authors

Oluwafunmilola Ola – *Advanced Materials Group, Faculty of Engineering, The University of Nottingham, Nottingham NG7 2RD, U.K.; Email: Oluwafunmilola.Ola@nottingham.ac.uk*

Kenneth I. Ozoemena – *Molecular Science Institute, School of Chemistry, University of the Witwatersrand, Johannesburg 2050, South Africa; orcid.org/0000-0001-7107-7003; Email: Kenneth.ozoemena@wits.ac.za*

Authors

Siwaphiwe Peteni – *Molecular Science Institute, School of Chemistry, University of the Witwatersrand, Johannesburg 2050, South Africa*

Okoroike C. Ozoemena – *Molecular Science Institute, School of Chemistry, University of the Witwatersrand, Johannesburg 2050, South Africa; Present Address: The author was a visiting student to the School of Chemistry, University of the Witwatersrand during this work. He has relocated to the Department of Chemistry, University of Guelph, Ontario, Canada*

Tobile Khawula – *Molecular Science Institute, School of Chemistry, University of the Witwatersrand, Johannesburg 2050, South Africa*

Aderemi B. Haruna – *Molecular Science Institute, School of Chemistry, University of the Witwatersrand, Johannesburg 2050, South Africa*

Frankie J. Rawson – *School of Pharmacy, Biodiscovery Institute University of Nottingham, Nottingham NG7 2RD, U.K.; orcid.org/0000-0002-4872-8928*

Leshweni J. Shai – *Department of Biomedical Sciences, Tshwane University of Technology, Pretoria 0001, South Africa*

Complete contact information is available at:

<https://pubs.acs.org/10.1021/acssensors.3c00677>

Notes

The authors declare no competing financial interest.

■ ACKNOWLEDGMENTS

This study was supported by the University of the Witwatersrand (Wits) and the South Africa's National Research Foundation (NRF) through the SARChI Chair in Materials Electrochemistry and Energy Technologies (MEET) (UID No. 132739). Siwaphiwe Peteni would like to thank the NRF and the Department of Science and Innovation (DSI) for the PhD Scholarship. The authors are grateful for funding from the University of Nottingham via the Global Challenges Research Fund scheme and the Cancer Research UK (EDDPMA-May21\100042).

■ REFERENCES

- (1) Arbyn, M.; Weiderpass, E.; Bruni, L.; de Sanjosé, S.; Saraiya, M.; Ferlay, J.; Bray, F. Estimates of incidence and mortality of cervical cancer in 2018: a worldwide analysis. *Lancet Glob. Health* **2020**, *8*, e191–e203.
- (2) Insuasty-Cepeda, D. S.; Maldonado, M.; García-Castañeda, J. E.; Rivera-Monroy, Z. J. Obtaining an immunoaffinity monolithic

- material: poly (GMA-co-EDMA) functionalized with an HPV-derived peptide using a thiol–maleimide reaction. *RSC Adv.* **2021**, *11*, 4247–4255.
- (3) Ganguly, N.; Parihar, S. P. Human papillomavirus E6 and E7 oncoproteins as risk factors for tumorigenesis. *J. Biosci.* **2009**, *34*, 113–123.
- (4) Ramesh, T.; Foo, K. L.; Sam, A. J.; Solayappan, M. Gold-Hybridized Zinc oxide nanorods as Real-time Low-cost nano-Biosensors for Detection of virulent DNA signature of HPV-16 in cervical carcinoma. *Sci. Rep.* **2019**, *9*, 17039.
- (5) Campos-Ferreira, D. S.; Souza, E. V.; Nascimento, G. A.; Zanforlin, D. M.; Arruda, M. S.; Beltrao, M. F.; Melo, A. L.; Brunaska, D.; Lima-Filho, J. L. Electrochemical DNA biosensor for the detection of human papillomavirus E6 gene inserted in recombinant plasmid. *Arab. J. Chem.* **2016**, *9*, 443–450.
- (6) Kawamata, Y.; Mitsuhashi, A.; Unno, Y.; Kado, S.; Shino, Y.; Uesugi, K.; Eguchi, O.; Ishii, J.; Seki, K.; Sekiya, S.; Shirasawa, H. HPV 16-E6-mediated degradation of intrinsic p53 is compensated by upregulation of p53 gene expression in normal cervical keratinocytes. *Int. J. Oncol.* **2002**, *21*, 561–567.
- (7) Avelino, K. Y.; Oliveira, L. S.; Lucena-Silva, N.; de Melo, C. P.; Andrade, C. A.; Oliveira, M. D. Metal-polymer hybrid nanomaterial for impedimetric detection of human papillomavirus in cervical specimens. *J. Pharm. Biomed. Anal.* **2020**, *185*, No. 113249.
- (8) Abreu, A. L.; Souza, R. P.; Gimenes, F.; Consolaro, M. E. A review of methods for detect human Papillomavirusinfection. *Virol. J.* **2012**, *9*, 262.
- (9) Waxman, A. G.; Zsemlye, M. M. Preventing cervical cancer: The Pap Test and the HPV vaccine. *Med. Clin. North Am.* **2008**, *92*, 1059–1082.
- (10) Wilson, M. L.; Fleming, K. A.; Kuti, M. A.; Looi, L. M.; Lago, N.; Ru, K. Access to pathology and laboratory medicine services: a crucial gap. *Lancet* **2018**, *391*, 1927–1938.
- (11) Pareek, S.; Jain, U.; Bharadwaj, M.; Chauhan, N. A label free nanosensing platform for the detection of cervical cancer through analysis of ultratrace DNA hybridization. *Sens. Bio-Sens. Res.* **2021**, *33*, No. 100444.
- (12) Espinosa, J. R.; Galván, M.; Quiñones, A. S.; Ayala, J. L.; Ávila, V.; Durón, S. M. Electrochemical resistive DNA biosensor for the detection of HPV type 16. *Molecules* **2021**, *26*, 3436.
- (13) Chekin, F.; Bagga, K.; Subramanian, P.; Jijie, R.; Singh, S. K.; Kurungot, S.; Boukherroub, R.; Szunerits, S. Nucleic aptamer modified porous reduced graphene oxide/MoS₂ based electrodes for viral detection: Application to human papillomavirus (HPV). *Sens. Actuators, B* **2018**, *262*, 991–1000.
- (14) Teengam, P.; Siangproh, W.; Tuantranont, A.; Henry, C. S.; Vilaivan, T.; Chailapakul, O. Electrochemical paper-based peptide nucleic acid biosensor for detecting human papillomavirus. *Anal. Chim. Acta* **2017**, *952*, 32–40.
- (15) Piro, B.; Kapella, A.; Le, V.; Anquetin, G.; Zhang, Q.; Reisberg, S.; Noel, V.; Tran, L.; Duc, H.; Pham, M. Towards the detection of human papillomavirus infection by a reagentless electrochemical peptide biosensor. *Electrochim. Acta* **2011**, *56*, 10688–10693.
- (16) Valencia, D. P.; Dantas, L. M. F.; Lara, A.; García, J.; Rivera, Z.; Rosas, J.; Bertotti, M. Development of a bio-electrochemical immunosensor based on the immobilization of SPINNTKPHEAR peptide derived from HPV-L1 protein on a gold electrode surface. *J. Electroanal. Chem.* **2016**, *770*, 50–55.
- (17) Wang, A.; Zhou, Y.; Chen, Y.; Zhou, J.; You, X.; Liu, H.; Liu, Y.; Ding, P.; Qi, Y.; Liang, C.; Zhu, X.; Zhang, Y.; Liu, E.; Zhang, G. Electrochemical immunosensor for ultrasensitive detection of human papillomavirus type 16 L1 protein based on Ag@AuNPs-GO/SPA. *Anal. Biochem.* **2023**, *660*, No. 114953.
- (18) Tran, L. D.; Nguyen, D. T.; Nguyen, B. H.; Do, Q. P.; Nguyen, H. L. Development of interdigitated arrays coated with functional polyaniline/MWCNT for electrochemical biodetection: application for human papilloma virus. *Talanta* **2011**, *85*, 1560–1565.
- (19) Ejikeme, P. M.; Makgopa, K.; Raju, K.; Ozoemena, K. I. Promotional Effects of Nanodiamond-Derived Onion-Like Carbons on the Electrocatalytic Properties of Pd-MnO₂ for the Oxidation of Glycerol in Alkaline Medium. *ChemElectroChem* **2016**, *3*, 2243–2251.
- (20) Ogada, J. J.; Ipadeola, A. K.; Mwonga, P. V.; Haruna, A. B.; Nichols, F.; Chen, S.; Miller, H. A.; Pagliaro, M. V.; Vizza, F.; Varcoe, J. R.; Meira, D. M.; Wamwangi, D. M.; Ozoemena, K. I. CeO₂ Modulates the Electronic States of a Palladium Onion-Like Carbon Interface into a Highly Active and Durable Electrocatalyst for Hydrogen Oxidation in Anion-Exchange-Membrane Fuel Cells. *ACS Catal.* **2022**, *12*, 7014–7029.
- (21) Ehirim, T. J.; Ozoemena, O. C.; Mwonga, P. V.; Haruna, A. B.; Mofokeng, T. P.; De Wael, K.; Ozoemena, K. I. Onion-like Carbons Provide a Favorable Electrocatalytic Platform for the Sensitive Detection of Tramadol Drug. *ACS Omega* **2022**, *7*, 47892.
- (22) Ozoemena, O. C.; Shai, L. J.; Maphumulo, T.; Ozoemena, K. I. Electrochemical sensing of dopamine using onion-like carbons and their carbon nanofiber composites. *Electrocatalysis* **2019**, *10*, 381–391.
- (23) Ozoemena, O. C.; Mathebula, N. S.; Ehirim, T. J.; Maphumulo, T.; Valikpe, G. M.; Shai, J. L.; Ozoemena, K. I. Onion-like carbon reinforced electrospun polyacrylonitrile fibres for ultrasensitive electrochemical immunosensing of Vibrio cholerae toxin. *Electrochim. Acta* **2020**, *356*, No. 136816.
- (24) Choudhury, S.; Srimuk, P.; Raju, K.; Tolosa, A.; Fleischmann, S.; Zeiger, M.; Ozoemena, K. I.; Borchardt, L.; Presser, V. Carbon onion/sulfur hybrid cathodes via inverse vulcanization for lithium–sulfur batteries. *Sustainable Energy Fuels* **2018**, *2*, 133–146.
- (25) Habib, I.; Ferrer, P.; Ray, S. C.; Ozoemena, K. I. Interrogating the impact of onion-like carbons on the supercapacitive properties of MXene (Ti₂CTx). *J. Appl. Phys.* **2019**, *126*, 134301.
- (26) Roche, R.; Yalcinkaya, F. Electrospun polyacrylonitrile nanofibrous membranes for point-of-use water and air cleaning. *ChemistryOpen* **2019**, *8*, 97–103.
- (27) Lim, D. G.; Kim, K. H.; Kang, E.; Lim, S. H.; Ricci, J.; Sung, S. K.; Kwon, M. T.; Jeong, S. H. Comprehensive evaluation of carboxylated nanodiamond as a topical drug delivery system. *Int. J. Nanomed.* **2016**, *11*, 2381–2395.
- (28) Kirmani, A. R.; Peng, W.; Mahfouz, R.; Amassian, A.; Losovyj, Y.; Idriss, H.; Katsiev, K. On the relation between chemical composition and optical properties of detonation nanodiamonds. *Carbon* **2015**, *94*, 79–84.
- (29) Fujimoto, A.; Yamada, Y.; Koinuma, M.; Sato, S. Origins of sp³C peaks in C1s X-ray Photoelectron Spectra of Carbon Materials. *Anal. Chem.* **2016**, *88*, 6110–6114.
- (30) Du, J.; Zhao, R.; Zhu, Z. A facile approach for synthesis and in situ modification of onion-like carbon with molybdenum carbide. *Phys. Status Solidi A* **2011**, *208*, 878–881.
- (31) Qi, X.; Yang, G.; Jing, M.; Fu, Q.; Chiu, F.-C. Microfibrillated cellulose-reinforced bio-based poly(propylene carbonate) with dual shape memory and self-healing properties. *J. Mater. Chem. A* **2014**, *2*, 20393–20401.
- (32) Xiao, S.; Xu, P.; Peng, Q.; Chen, J.; Huang, J.; Wang, F.; Noor, N. Layer-by-Layer Assembly of Polyelectrolyte Multilayer onto PET Fabric for Highly Tunable Dyeing with Water Soluble Dyestuffs. *Polymers* **2017**, *9*, 735.
- (33) Ozoemena, O. C.; Ehirim, T. J.; Khawula, T.; Makgopa, K.; Shai, L. J.; Ozoemena, K. I. Bovine Serum Albumin-Dependent Charge-Transfer Kinetics Controls the Electrochemical Immunosensitive Detection: Vibrio cholerae as a Model Bioanalyte. *Electrocatalysis* **2021**, *12*, 595–604.
- (34) Ozoemena, O. C.; Maphumulo, T.; Shai, J. L.; Ozoemena, K. I. Electrospun carbon nanofibers as an electrochemical immunosensing platform for vibrio cholerae toxin: Aging effect of the redox probe. *ACS Omega* **2020**, *5*, 5762–5771.
- (35) Agboola, B. O.; Ozoemena, K. I. Self-assembly and heterogeneous electron transfer properties of metallo-octacarboxyphthalocyanine complexes on gold electrode. *Phys. Chem. Chem. Phys.* **2008**, *10*, 2399–2408.

(36) Dockter, J.; Schroder, A.; Eaton, B.; Wang, A.; Sikhamsay, N.; Morales, L.; Giachetti, C. Analytical characterization of the APTIMA HPV Assay. *J. Clin. Virol.* **2009**, *45*, S39–S47.

(37) Hesselink, A. T.; Berkhof, J.; van der Salm, M. L.; van Splunter, A. P.; Geelen, T. H.; van Kemenade, F. J.; Bleeker, M. G.; Heideman, D. A. Clinical validation of the HPV-risk assay, a novel real-time PCR assay for detection of high-risk human papillomavirus DNA by targeting the E7 region. *J. Clin. Microbiol.* **2014**, *52*, 890–896.

(38) Wongsamart, R.; Bhattarakasol, P.; Chaiwongkot, A.; Wongsawaeng, D.; Okada, P. A.; Palaga, T.; Leelahavanichkul, A.; Khovidhunkit, W.; Dean, D.; Somboonna, N. Multiplex recombinase polymerase amplification for high-risk and low-risk type HPV detection, as potential local use in single tube. *Sci. Rep.* **2023**, *13*, 829.

(39) Goode, J.; Rushworth, J.; Millner, P. Biosensor regeneration: a review of common techniques and outcomes. *Langmuir* **2015**, *31*, 6267–6276.

(40) Zhu, L.; Han, J.; Wang, Z.; Yin, L.; Zhang, W.; Peng, Y.; Nie, Z. Competitive adsorption on gold nanoparticles for human papillomavirus 16 L1 protein detection by LDI-MS. *Analyst* **2019**, *144*, 6641–6646.

(41) Zhu, L.; Zhao, Y.; Yao, S.; Xu, M.; Yin, L.; Zhai, X.; Teng, X. A colorimetric aptasensor for the simple and rapid detection of human papillomavirus type 16 L1 proteins. *Analyst* **2021**, *146*, 2712–2717.

(42) Ozoemena, K. I.; Mathebula, N. S.; Pillay, J.; Toschi, G.; Verschoor, J. A. Electron transfer dynamics across self-assembled N-(2-mercaptoethyl) octadecanamide/mycolic acid layers: impedimetric insights into the structural integrity and interaction with anti-mycolic acid antibodies. *Phys. Chem. Chem. Phys.* **2010**, *12*, 345–357.

(43) Mathebula, N. S.; Pillay, J.; Toschi, G.; Verschoor, J. A.; Ozoemena, K. I. Recognition of anti-mycolic acid antibody at self-assembled mycolic acid antigens on a gold electrode: a potential impedimetric immunosensing platform for active tuberculosis. *Chem. Commun.* **2009**, *23*, 3345–3347.