

CORRECTION**Correction: “Commissioning and clinical evaluation of the IDENTIFY™ surface imaging system for frameless stereotactic radiosurgery”**<https://doi.org/10.1002/acm2.14058>**Elizabeth L. Covington^{1,2} | Dennis N. Stanley² | Rodney J. Sullivan² |
Kristen O. Riley² | John B. Fiveash¹ | Richard A. Popple²**¹University of Michigan, Ann Arbor, Michigan, USA²University of Alabama at Birmingham, Birmingham, Alabama, USA

An incorrect table was inadvertently uploaded into the original manuscript. The below table contains the correct values for the median and interquartile range values for the IDENTIFY™ surface imaging system before beam-on at non-zero couch angles.

TABLE 2 Median and interquartile range (IQR) of IDENTIFY™ reported offsets before beam-on at non-zero couch angles.

Couch angle (°)	Vertical (IQR) mm	Longitudinal (IQR) (mm)	Lateral (IQR) (mm)
270	−0.01 (0.47)	−0.03 (0.33)	0.16 (0.32)
225	−0.06 (0.17)	0.21 (0.38)	−0.16 (0.18)
135	−0.08 (0.16)	0.19 (0.38)	−0.41 (0.19)
90	−0.14 (0.20)	0.05 (0.54)	−0.25 (0.24)
0° (end of treatment)	−0.05 (0.18)	−0.03 (0.35)	−0.01 (0.15)

This is an open access article under the terms of the [Creative Commons Attribution](#) License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.

© 2024 The Author(s). *Journal of Applied Clinical Medical Physics* published by Wiley Periodicals, LLC on behalf of The American Association of Physicists in Medicine.