

Rapid detection of SARS-CoV-2 variants by molecular-clamping technology-based RT-qPCR

Shuo Shen,¹ Andrew Y. Fu,¹ Maidar Jamba,¹ Jonathan Li,¹ Zhen Cui,¹ Larry Pastor,¹ Daniel Cataldi,¹ Qing Sun,¹ Joseph A. Pathakamuri,² Daniel Kuebler,² Michael Rohall,² Madison Krohn,² Daniel Kissinger,² Jocelyn Neves,² Isaac Archibeque,² Aiguo Zhang,¹ Chuanyi M. Lu,³ Michael Y. Sha¹

AUTHOR AFFILIATIONS See affiliation list on p. 14.

ABSTRACT Given the challenges that SARS-CoV-2 variants have caused in terms of rapid spread and reduced vaccine efficacy, a rapid and cost-effective assay that can detect new and emerging variants is greatly needed worldwide. We have successfully applied the xenonucleic acid-based molecular-clamping technology to develop a multiplex reverse-transcription quantitative real-time PCR assay for SARS-CoV-2 multivariant detection. The assay was used to test 649 nasopharyngeal swab samples that were collected for clinical diagnosis or surveillance. The assay was able to correctly identify all 36 Delta variant samples as it accurately detected the D614G, T478K, and L452R mutations. In addition, the assay was able to correctly identify all 34 Omicron samples by detecting the K417N, T478K, N501Y, and D614G mutations. This technique reliably detects a variety of variants and has an analytical sensitivity of 100 copies/mL. In conclusion, this novel assay can serve as a rapid and cost-effective tool to facilitate large-scale detection of SARS-CoV-2 variants.

IMPORTANCE We have developed a multiplex reverse-transcription quantitative real-time PCR (RT-qPCR) testing platform for the rapid detection of SARS-CoV-2 variants using the xenonucleic acid (XNA)-based molecular-clamping technology. The XNA-based RT-qPCR assay can achieve high sensitivity with a limit of detection of about 100 copies/mL for variant detection which is much better than the next-generation sequencing (NGS) assay. Its turnaround time is about 4 hours with lower cost and a lot of Clinical Laboratory Improvement Amendments (CLIA) labs own the instrument and meet skillset requirements. This assay provides a rapid, reliable, and cost-effective testing platform for rapid detection and monitoring of known and emerging SARS-CoV-2 variants. This testing platform can be adopted by laboratories that perform routine SARS-CoV-2 PCR testing, providing a rapid and cost-effective method in lieu of NGS-based assays, for detecting, differentiating, and monitoring SARS-CoV-2 variants. This assay is easily scalable to any new variant(s) should it emerge.

KEYWORDS SARS-CoV-2 variant detection, molecular-clamping technology, XNA

More than 3 years after its initial emergence, the SARS-CoV-2 fueled COVID-19 pandemic and continues to spread globally with 755 million infections and 6.8 million deaths to date (13 February 2023) (1). In the United States alone, there have been 102 million documented cases, and the death toll has passed 1.1 millions (13 February 2023) (2). Despite increasing vaccination levels and a rising number of COVID-19-recovered people who have acquired some degree of immunity, the spread has continued. This is largely due to the appearance of more transmissible and partially vaccine-resistant novel variants of the virus. Since the inception of the pandemic, five variants of concern (VOC) and seven variants of interest (VOI) have emerged worldwide

Editor Leiliang Zhang, Shandong First Medical University, Jinan, Shandong, China

Ad Hoc Peer Reviewers Wayne Xianding Deng, University of California San Francisco, San Francisco, California, USA; Junping Peng, Institute of Pathogen Biology, Chinese Academy of Medical Sciences & Peking Union Medical College, Beijing, China

Address correspondence to Michael Y. Sha, miksha168@gmail.com.

Shuo Shen, Andrew Y. Fu, and Maidar Jamba contributed equally to this article. The author order was determined by their contribution to the article

The authors declare no conflict of interest.

This article is dedicated to Dr. Michael J. Powell in recognition of his contributions to XNA in the field of molecular clamping technology during his tenure as CSO at DiaCarta Inc.

Received 21 December 2023

Accepted 30 June 2024

Published 16 October 2024

[This article was published on 16 October 2024 with a typo in the keywords. The first keyword was corrected in the current version, posted on 5 November 2024.]

This is a work of the U.S. Government and is not subject to copyright protection in the United States. Foreign copyrights may apply.

(3–8). These include the Alpha B.1.1.7 (501Y.V1) (3), Beta B.1.351 (501Y.V2) (4), Gamma P.1 (501Y.V3) (5), Epsilon CAL.20C (20 C/S:452R, or B.1.429) (6–8), Delta B.1.617.2, and Omicron (B.1.1.529) (9). As of mid-2023, the infection rates have declined, leading to the removal of certain VOC from the World Health Organization, Centers for Disease Control and Prevention (CDC), and European Center for Disease Prevention and Control websites (10–12). As of June 2024, only VOI and variants under monitoring (VUM) lists are maintained. The prevalent VOIs currently include EG.5, BA.2.86, and JN.1, while the VUMs consist of JN.1.7, KP.2, and KP.3.

The Alpha B.1.1.7 strain, also known as 20I/501Y.V1 and VOC 20DEC-01 (VOC-20DEC-01, previously written as VOC-202012/01), was found initially in the southeast of England in early October 2020 and became prevalent in both Europe and the United States shortly thereafter (13–16). It is estimated to be 40%–80% more transmissible than the original SARS-CoV-2 strain (17, 18).

While the main variant from April to June 2021 in the United States was the Alpha variant, its decline from 70% of cases (8 May to 19 June 2021) to 9.0% of cases (17 July 2021) corresponded with an increase in the prevalence of the Delta variant, which rose from 0.6% (24 April 2021) to 82.8% (17 July 2021) (19–23). The Delta B.1.617.2 variant was first identified in India (24, 25) and is characterized by 13 mutations (26). In addition, a Delta plus variant, which has the Delta mutations plus an additional K417N mutation, has also been detected. Compared to Alpha, these variants are more contagious and are associated with a high degree of mortality. The Delta variants grow more rapidly in the human respiratory tract and have 1,000 times higher viral loads than the original strain (27). In addition, large numbers of breakthrough cases associated with the Delta variants were documented in previously vaccinated individuals. Since November 2021, the emerging Omicron variant (B.1.1.529) was found and reported in Botswana and South Africa. A big family of Omicron variants (BA.2, BA.4, BA.5, Botswana and South Africa, BF.7, BQ.1, etc.) has been reported and studied (28–30).

Each one of these main VOCs has a signature set of spike protein mutations that allow them to be uniquely identified. While each VOC has a unique set of spike protein mutations, the D614G mutation is shared by all five VOC. The D614G mutation involves an amino acid change from aspartic acid to glycine, which is caused by an A-to-G nucleotide transition at position 23403 of the viral genome. This change stabilizes the spike protein and enhances its fitness and infectivity (20) and is associated with increased viral infectivity and transmissibility (21, 22). Another key mutation, N501Y, is found in all the VOCs except for Delta. The N501Y mutation is located within the receptor-binding domain (RBD) of the spike protein and is associated with stronger binding to ACE2 receptor and higher viral infectivity. Of greater concern, though, is the fact that the efficacy of vaccine-mediated immunity against strains containing the N501Y mutation seems to be reduced (23). These two mutations were the two earliest mutations observed and remain the most prominent mutations detected in the SARS-CoV-2 virus.

To date, next-generation sequencing (NGS) has been the standard method used for SARS-CoV-2 variants detection (31). Although the NGS-based assays facilitate variant finding and confirmation, they are expensive and time consuming, require technical expertise, and are not readily available, particularly in low-income countries and regions. These factors limit their utility in large-scale testing and monitoring of the newly spreading and known SARS-CoV-2 variants. Certainly, whole-genome sequence (WGS) can discover the new variants for SARS-CoV-2 and cannot be replaced by qPCR. Thus PCR method and NGS method can well supplement each other in viral variant detection and tracking. Given their importance in fighting the pandemic, the MIT Technology Review magazine listed “COVID variant tracking” as one of the 10 breakthrough technologies in 2022 (32).

Considering the rapid evolution and emergence of various variants, testing platforms capable of rapid detection of specific variants in a cost-effective manner are urgently needed. Toward this end, target-specific RT-PCR screens for specific mutations within the spike protein had been developed (33, 34). In this study, we have developed

a multiplex reverse-transcription quantitative real-time PCR (RT-qPCR) assay that can rapidly and reliably detect known and emerging SARS-CoV-2 variants based on their mutation profiles (Fig. 1). The assay is based on molecular-clamping technology and uses xenonucleic acids (XNA) as molecular clamping probes.

XNAs are artificial genetic polymers that retain the Watson-Crick base-pairing capability and exhibit high chemical and biological stability. In practical applications for disease diagnosis and treatment, XNAs have been used as a source of nuclease-resistant affinity reagents (aptamers) and catalysts (xenzozymes). More notably, they can be employed as molecular clamps in RT-qPCR or as highly specific molecular probes for detecting nucleic acid target sequences due to the stronger hybridizing/binding capability seen in XNA/DNA duplexes than DNA/DNA duplexes (35). Furthermore, even a single base-pair mismatch near the center of the sequence of an XNA/DNA duplex can result in a big drop of 10°C–18°C in melting temperature (T_m) (36). This appealing feature allows for the highly specific clamping of the XNA molecule onto the targeted sequence (usually the wild type, WT) to block WT amplification, thus minimizing the WT background in RT-qPCR and selectively enhancing the signal of the mutant. Robust XNA-based assays have been extensively studied and used for *in vitro* molecular diagnostic assays for early detecting cancer-associated gene mutations (37), fusion gene detection (38) and companion diagnostic (39). XNA can also be used in NGS similarly against




FIG 1 Principle of SARS-CoV-2 variant detection RT-qPCR. (A) SARS-CoV-2 genome structure and its spike gene target sites for our multivariant assay. (B) A high-throughput RT-qPCR workflow for SARS-CoV-2 variant detection, highlighting the mechanism of XNA-based molecular clamping technique capable of distinguishing a targeted mutation from its wild type. The circle indicates the primary structure of the XNA clamping molecule in a typical XNA/DNA duplex.

wild-type background to increase the specificity and sensitivity in mutant detections (40). Broad and diverse applications of XNAs in molecular diagnostics and other related fields including imaging and microscopy have been recently reviewed (35, 41). Based on their molecular properties and prior utility in cancer mutation detection, XNAs represent an ideal tool for developing highly sensitive RT-qPCR assays for detecting SARS-CoV-2 mutations. The XNA we used in this study is a class of chemically modified peptide nucleic acids (PNAs) with improved and more hydrophilic structure.

In this study, we demonstrated for the first time the feasibility, efficiency, reliability, and applicability of using an XNA-based RT-qPCR method to detect various SARS-CoV-2 variants. This assay provides a rapid, reliable, and cost-effective testing platform for the detection and monitoring of known and emerging SARS-CoV-2 variants.

MATERIALS AND METHODS

Study design

Deidentified patient nasopharyngeal swabs (NPS) collected and tested from early 2021 to early January 2022 at San Francisco VA Medical Center clinical laboratory and DiaCarta clinical laboratory for clinical diagnostic or screening purposes were used in this study. The second set of deidentified NPS samples collected and tested from June 2021 to early January 2022 at the COVID-19 screening lab at Franciscan University, Ohio, were also used.

Sample collection and RNA extraction

From early 2021 to early January 2022, patient NPS samples were collected in a virus transportation medium (VTM), tested for SARS-CoV-2 infection, and then stored at -80°C before being used for analysis in this study. An automatic RNA/DNA extraction instrument MGISP-960 (MGI Tech Co., Ltd.) and the MGI Easy Nucleic Acid Extraction Kit (Cat# 1000020261) were used to extract SARS-CoV-2 viral RNA according to the manufacturer's instructions. The starting extraction volume for each sample is 200 μL of NPS sample in VTM. The final RNA extraction volume is 40 μL in RNase-free water, of which 5.5 μL was then used for a single RT-qPCR. The typical turnaround time from sample RNA extraction was about 90 min for 384 samples (Fig. 1) (42).

Multiplex primer and probe design

In order to design primers and probes to detect all the SARS-CoV-2 variants of concern, the conserved regions of ORF1ab gene (34) and areas adjacent to the N501Y, D614G, T478K, L452R, K417T, and K417N mutations in S protein RBD were targeted (Fig. 1A). Unique combinations of these mutants can be used to identify each specific VOC (Table S1).

Gene sequences were retrieved from GenBank and GISAID databases for primer and probe design to ensure coverage of all SARS-CoV-2 variant strains. Multiple alignments of the collected sequences were performed using Qiagen CLC Main Workbench 20.0.4., and conserved regions in each target gene were identified using BioEditor 7.2.5. Primers and probes were designed by using Primer3plus software. All primers were designed with a T_m of approximately 60°C , and the probes were designed with a T_m of about 65°C . The amplicon sizes were kept within the range of 70–150 bp for each primer pair in order to achieve high amplification efficiency and detection sensitivity (Table S2). All designed primers and probes were ordered from Integrated DNA Technologies, Inc. (Coralville, IA) and LGC Biosearch Technologies (Novato, CA), respectively.

XNA design, synthesis, analysis, and optimization

Five groups of XNAs, each specific to one of five SARS-CoV-2 mutations (D614G, N501Y, T478K, L452R, and K417T/K417N), were designed to be exact matches with the WT

sequence in order to facilitate selective blocking of qPCR amplification of WT targets. For each mutation, a number of XNAs of different lengths were synthesized in our chemistry laboratory (Table S3a and b). Each XNA was also designed to partially overlap with and be of the same strand/sense as the corresponding qPCR probe. The best XNA probe sequences were selected based on iterative adjustment of multiple major physicochemical factors: probe sequence length, GC content, purine content and arrangement, self-complementarity, and melting temperature. Five groups of XNAs were synthesized, and each group was tested using qPCR to find the optimal size and concentration of XNA for each group. Fig. S1 showed matrix assisted laser desorption ionization (MALDI)-TOF mass spectra of XNAs.

Real-time reverse-transcription PCR

Two different detection kit assays were developed using this methodology. The QuantiVirus SARS-CoV-2 Variants Detection Kit assay consisted of three multiplex PCR tubes: tube A tested for the presence of D614G and L452R, tube B tested for the presence of K417 T, N501Y, and the human Rp gene as internal control, and tube C tested for the presence of T478K, K417 N, and ORF1ab as wild-type target. The QuantiVirus SARS-CoV-2 Delta Plus Detection Kit assay consisted of only one 4-plex PCR tube that tested for the presence of T478K, K417N, ORF1ab, and Rp.

XNA-based RT-qPCRs with a total volume of 10 μ L were developed using the following reagents: 5.5 μ L of viral RNA, 1.0 μ L of primer and probe mixture (final concentration of 0.2 μ M and 0.1 μ M, respectively), 1.0 μ L 10 \times XNAs mixes, and 2.5 μ L of TaqPath 1-step Multiplex Master Mix (Catalog# A28526, Thermo Fisher, MA). Each test included a positive control (PC), a negative control (NC), and a no template control (NTC) to ensure there were no false positives, false negatives, or contamination in each qPCR run. The PC contained all mutation target genes, the NC included the wild-type gene for each target, and the NTC was water.

The qPCR was performed at 25 $^{\circ}$ C for 2 min with uracil-N-glycosylase (UNG) incubation to remove potential carryover, then 53 $^{\circ}$ C for 10 min for reverse transcription, followed by 95 $^{\circ}$ C for 2 min. Then 45 cycles of 95 $^{\circ}$ C for 3 s and 60 $^{\circ}$ C for 30 s were used for amplification. A Bio-Rad CFX384 (Bio-Rad, CA) was used for RT-qPCR amplification and detection (42–44). Assessment of the results for each individual assay should be based on the C_t values < 40 for each gene (Orf1 ab, D614G, N501Y, T478K, L452R, K417T, and K417N) as positive and C_t > 40 as negative. The results were interpreted according to the Table S1.

Analysis of assay sensitivity and cross-reactivity test

Twist SARS-CoV-2 RNA controls 16 (Beta), 17 (Gamma), 23 (Delta), and 48 (Omicron; Twist Bioscience, CA) were used as references to test the assay sensitivity. We used a twofold dilution series from 800 copies/mL down to 25 copies/mL of the templates in triplicates and identified the lowest concentration that was detectable with 95% confidence to determine the analytical sensitivity limit of detection (LoD).

As control standards, MERS-CoV and SARS-CoV coronavirus samples were ordered from ATCC (Manassas, Virginia), and the ZeptoMetrix NATtrol Respiratory Validation Panel was ordered from ZeptoMetrix (cat# NATRVP-3, Buffalo, NY). RNA/DNA was extracted from high-titer stocks of these potentially cross-reacting microorganisms (estimated 10⁵ units/mL) using the MGI Easy Nucleic Acid Extraction Kit (Cat# 1000020261). The extracted sample RNA/DNA was eluted with sterile RNase-free water to give a 100 μ L solution. A volume of 5.5 μ L of each of the purified RNA/DNA samples was tested in triplicate with QuantiVirus SARS-CoV-2 Variants Detection Kit.

Clinical evaluation of samples

All the clinical samples were evaluated using the QuantiVirus SARS-CoV-2 Variants Detection Test kit. In addition, the presence/absence of SARS-CoV-2 was confirmed for

all samples using the Food and Drug Administration (FDA) Emergency Use Authorization (EUA)-approved QuantiVirus SARS-CoV-2 Multiplex Test Kit at the DiaCarta CLIA-certified clinical laboratory or at the Franciscan University COVID screening lab.

Sanger sequencing and next-generation sequencing verification

All samples that were identified as variant positive by this assay were sent out for Sanger Sequencing to confirm their mutational status (SequeTech, CA), and all sequences were analyzed via the UCSC SARS-CoV-2 Genome Browser (45).

Sanger Sequencing was prepared according to the following procedure. The extracted RNA samples from the SARS-CoV2 virus were amplified in a 10 μ L reaction of RT-PCR using the following reagents: 2.5 μ L of 4 \times 1-step qRT-PCR Master Mix, 2.0 μ L of primer mixture (final concentration of 0.2 μ M), and 5.5 μ L of viral RNA. The RT-PCR procedure consisted of the following steps, and a Bio-Rad CFX384 instrument (Bio-Rad, CA) was utilized: (i) incubation at 25 $^{\circ}$ C for 2 minutes with UNG to eliminate potential carryover; (ii) reverse transcription at 53 $^{\circ}$ C for 10 minutes; (iii) initial denaturation at 95 $^{\circ}$ C for 2 minutes; (iv) amplification through 45 cycles of denaturation at 95 $^{\circ}$ C for 3 seconds and annealing/extension at 60 $^{\circ}$ C for 30 seconds. Subsequently, all the crude PCR products were sent out for Sanger sequencing.

The NGS library was prepared with amplicon methodology using the CleanPlex SARS-CoV-2 FLEX Research and Surveillance Panel (Cat#. 918010, Paragon Genomics, CA) or Illumina COVIDSeq Test (FDA EUA approved, cat# 20049393) and sequenced on the MiSeq instrument (Illumina, CA). All the raw data were analyzed by ARTIC workflow on Galaxy (<https://github.com/galaxyproject/SARS-CoV-2>), using a reference sequence from the NCBI database (NC_045512.2) (46–48). All tools were used with default parameters.

RESULTS

XNA enhances the ability to distinguish viral variants from wild type in RT-qPCR

In order to test whether XNA clamping of the wild-type sequences enhances mutant detection, we compared the RT-qPCR results with and without XNA. An amplification curve of the D614G mutant vs wild type without XNA showed that it is difficult to distinguish the mutant and the wild-type virus (Fig. 2A and C). However, with XNA, the differentiation of the mutant from wild-type SARS-CoV-2 (Fig. 2B and D) was clear (ΔC_t , ~13–20). Figure 2E to L showed L452R, T478K, K417T, and K417N detection with or without XNA. To optimize the assay, we tested different concentrations of XNA. Representatively, the optimization results of D614G and N501Y XNAs are shown in Tables S4 and S5, respectively. The XNA that generated the largest ΔC_t between the wild type and the mutant, within its series, was selected for use in RT-qPCR. For example, 8 μ M D614G XNA001 resulted in the largest ΔC_t 15.9, and 0.25 μ M N510Y XNA003 resulted in the largest ΔC_t 32.7. Through the optimization process, we were able to select D614 XNA001, N501 XNA003, T478 XNA001, L452 XNA003, and K417 XNA001 as the optimal XNAs (their confirmed structural or mass spectra data are shown in Fig. S1 and Table S3a and b).

Analytical sensitivity and cross-reactivity evaluations

In order to determine the analytical sensitivity of the assay, we performed the QuantiVirus SARS-CoV-2 variant detection test on a Bio-Rad CFX384. We diluted SARS-CoV-2 RNA control 16/17/23 which contains the N501Y, D614G, L452R, T478K, K417T, and K417T mutants (Twist Bioscience, CA) from 800 copies/mL down to 25 copies/mL (viral copy number per milliliter of transport media) and tested each level dilution repeatedly in 24 replicates (Table 1; Table S10). The LoD was determined to be 100 copies/mL for each mutant with 95% CI 76.9%–99.8%.

We evaluated the cross-reactivity of the assay to a variety of other pathogens. The results are summarized in Table S6 and indicate that there is no cross-reactivity in our

TABLE 1 LoD of the multiplex RT-qPCR for SARS-CoV-2 variant detection^b

Target	Copies/mL	Avg C _t	SD	CV	Overall tests ^a (N)	Detection rate (95% CI)
N501Y	200	33.76	0.29	0.86%	24/24	100.00% (82.8%–100%)
	100	34.43	0.29	0.83%	23/24	95.83% (76.9%–99.8%)
	50	36.89	0.27	0.73%	22/24	91.67% (71.5%–98.5%)
	25	37.94	0.29	0.77%	21/24	87.50% (66.5%–96.7%)
D614G	200	32.42	0.26	0.79%	24/24	100.00% (82.8%–100%)
	100	34.28	0.34	1.00%	23/24	95.83% (76.9%–99.8%)
	50	37.20	0.25	0.67%	20/24	83.33% (61.8%–94.5%)
	25	38.92	0.27	0.69%	19/24	79.17% (57.3%–92.1%)
L452R	200	34.95	0.27	0.78%	24/24	100.00% (82.8%–100%)
	100	35.78	0.30	0.84%	23/24	95.83% (76.9%–99.8%)
	50	36.84	0.24	0.66%	22/24	91.67% (71.5%–98.5%)
	25	37.90	0.20	0.53%	20/24	83.33% (61.8%–94.5%)
K417T	200	32.32	0.33	1.01%	24/24	100.00% (82.8%–100%)
	100	33.49	0.30	0.90%	24/24	100.00% (82.8%–100%)
	50	34.56	0.28	0.80%	23/24	95.83% (76.9%–99.8%)
	25	36.36	0.32	0.87%	22/24	91.67% (71.5%–98.5%)
K417N	200	32.88	0.29	0.89%	24/24	100.00% (82.8%–100%)
	100	33.71	0.30	0.90%	24/24	100.00% (82.8%–100%)
	50	34.81	0.28	0.82%	22/24	91.67% (71.5%–98.5%)
	25	36.69	0.25	0.68%	22/24	91.67% (71.5%–98.5%)
T478K	200	33.32	0.27	0.82%	24/24	100.00% (82.8%–100%)
	100	34.20	0.27	0.80%	24/24	100.00% (82.8%–100%)
	50	35.49	0.21	0.60%	23/24	95.83% (76.9%–99.8%)
	25	37.32	0.30	0.81%	22/24	91.67% (71.5%–98.5%)
ORF	200	33.32	0.28	0.83%	24/24	100.00% (82.8%–100%)
	100	33.93	0.28	0.82%	24/24	100.00% (82.8%–100%)
	50	34.99	0.23	0.65%	22/24	91.67% (71.5%–98.5%)
	25	36.32	0.28	0.76%	22/24	91.67% (71.5%–98.5%)
RP	200	34.17	0.29	0.85%	24/24	100.00% (82.8%–100%)
	100	34.78	0.28	0.81%	24/24	100.00% (82.8%–100%)
	50	36.84	0.29	0.79%	23/24	95.83% (76.9%–99.8%)
	25	38.04	0.32	0.84%	23/24	95.83% (76.9%–99.8%)

^aEach concentration was tested repeatedly by 24 replicated tests. For details, see Table S9.

^bCV, coefficient of variation; ORF, SARS-CoV-2 open reading frames; RP, human RNase P gene.

XNA-based qPCR assay among the SARS-CoV-2 variants and any of the organisms tested including MERS-CoV.

Use of XNA-based RT-qPCR to confirm rapid surge of Alpha variant in San Francisco Bay Area in early 2021

In order to determine the ability of the QuantiVirus SARS-CoV-2 Variants Detection kit to detect the Alpha variant, 374 confirmed positive samples that were collected between January and March 2021 in the San Francisco Bay Area were analyzed. Among the 139 positive specimens sampled from mid-January 2021, 58 (41.7%) were positive for the D614G mutation but not the N501Y (Tables 2 and 3). None of the samples were positive for both N501Y and D614G mutations, a defining characteristic of the Alpha variant (B.1.1.7). However, of the 139 positive specimens sampled from late February and the 96 positive specimens collected in March 2021, there were 7 (5.04%) and 10 (10.42%) specimens, respectively, that were positive for both mutations. The data suggested an increase in the Alpha variant frequency in Northern California, from 0% in January to above 10% in March 2021. Additionally, the multiplex RT-qPCR test showed high specificity, i.e., amplification in non-N501Y samples and negative controls was




FIG 2 qPCR test with and without XNA. For D614G mutation detection: (A) amplification curve of D614G mutant (MT) vs WT without XNA. Both wild-type and mutant had ~ 30 C_t . It is difficult to distinguish mutant and wild type. (B) Amplification curve of D614G mutant vs wild type with XNA. Wild-type amplification was blocked and had high $C_t \sim 40$, but mutant was enhanced and had $C_t \sim 27$, making the differentiation of WT/MT highly reliable, easy, and accurate. For N501Y mutation detection: (C) N501Y mutant (MT) vs WT without XNA. Both wild-type and mutant had ~ 21 C_t . It is difficult to distinguish the mutant and the wild type. (D) N501Y mutant vs wild type with XNA. Wild-type amplification was blocked and had high $C_t > 40$, but mutant was enhanced and had $C_t \sim 20$, making the distinction apparent and interpretation intuitive. For L452R mutation detection: (E) L452R mutant (MT) vs WT without XNA. The ΔC_t was around 2. (F) L452R MT vs WT with XNA. Wild-type amplification was blocked and had high $C_t > 40$, made ΔC_t to 25. For T478K mutation detection: (G) T478K MT vs WT without XNA (H) T478K MT vs WT with XNA. The ΔC_t between MT and WT increased from 3 to 27. For K417T mutation detection: (I) K417T MT vs WT without XNA (J) K417T MT vs WT with XNA. The ΔC_t between MT and WT increased from 0.5 to 28 and wild-type amplification was fully blocked. For K417N mutation detection: (K) K417N MT vs WT without XNA (L) K417N MT vs WT with XNA. The ΔC_t between MT and WT increased from 1 to 25, and wild-type amplification was fully blocked.

completely blocked by the XNA in the RT-qPCR, while amplification was present in all N501Y positive clinical samples and positive controls (Fig. 3A).

To verify that the D614G and the D614G/N501Y mutations in these mutant samples were detected correctly, we amplified the mutant target by PCR, and the amplicons from these samples were analyzed by Sanger sequencing. All 160 D614G positive samples, 17 D614G/N501Y double positive samples, and 197 wild-type samples, as determined by the Quantivirus SARS-CoV-2 Variants Detection kit, were confirmed by Sanger sequencing indicating that the specificity of the test was 100%. The Sanger sequencing peaks showed the target mutations in the viral cDNA, T < C in the case of D614G (Fig. 3B), and A < T in the case of N501Y (Fig. 3C).

As summarized in Table 2 and Table 3, the XNA-based RT-qPCR has a positive predictive value (PPV) of 100.0% (95% CI: 0.99–1.00) and negative predictive value (NPV) of 100% (95% CI: 0.96–1.00) for detecting the Alpha variant by comparing its Sanger Sequencing data.

Use of XNA-based RT-qPCR to detect breakthrough COVID-19 cases caused by Delta variant and Omicron variant

Despite widespread vaccination, a significant number of breakthrough cases were seen in late 2021 and early 2022 due to emergence of two new SARS-CoV-2 variants, Delta and Omicron. To examine the ability of the Quantivirus SARS-CoV-2 Delta Plus Detection kit

TABLE 2 Summary of SARS-CoV-2 alpha variant detection in January to March 2021^a

Sample collection time	RT-qPCR test	N501Y	D614G	ORF1ab ^b	RNase P gene
January 2021	Positive	0	58	132	139
	Total (N)	139	139	139	139
	Alpha variant detection rate (%)	0	41.7	95.0	100.0
February 2021	Positive	7	78	139	139
	Total	139	139	139	139
	Alpha variant detection rate (%)	5.0	56.1	100.0	100.0
March 2021	Positive	10	41	96	96
	Total	96	96	96	96
	Alpha variant detection rate (%)	10.4	42.7	100.0	100.0

^aThree hundred seventy-four clinical samples collected in middle January, late February, and March 2021 in the California Bay area.
^bORF1ab gene was targeted for wild-type SARS-CoV-2.

to detect these variants, we analyzed 39 NPS that were initially collected from symptomatic COVID-19-positive individuals from Franciscan University in Ohio (Fall 2021) who had been vaccinated (“breakthrough” infection). In addition, six NPS samples collected from healthy donors were analyzed as well. All 36 breakthrough patient samples tested positive for the SARS-CoV-2 Delta variant as they displayed the T478K mutation with the wild-type ORF1ab and three Delta plus variants as K417N and T478K mutant detectable, whereas all six samples collected from healthy donors tested negative for SARS-CoV-2 (Table S7) and an additional 44 negative samples were confirmed negative (Table 4). The data were compared with Sanger sequence and determined that the assay’s PPV for Delta detection was 100% (95% CI: 0.89–1.0), and its NPV was 100% (95% CI: 0.91–1.0).

To assess the ability of QuantiVirus SARS-CoV-2 Variants Detection kit to detect the Omicron variant, we tested an additional 37 NPS samples, 34 of which were collected from vaccinated symptomatic patients in early January 2022. All 34 samples collected from vaccinated symptomatic patients tested positive, while the other three samples tested negative for SARS-CoV-2 by the FDA EUA-approved kit (QuantiVirus SARS-COV-2 Multiplex Test). All 34 positive samples further tested positive for the K417N, T478K, N501Y, and D614G mutations, characteristic of the Omicron variant, using the XNA-based RT-qPCR assay in this kit (Table S8). We also tested 50 negative confirmed samples. The data were compared with Sanger Sequence and determined that the assay’s PPV for Omicron detection was 100% (95% CI: 0.87–1.00), and its NPV was 100% (95% CI: 0.92–1.00; Table 5).

Summary of XNA-based RT-qPCR assay in detecting SARS-CoV-2 variants and comparing with NGS

In total, 649 samples were screened using the kit, including 447 SARS-CoV-2 positive samples and 202 negative samples (Table 6). Among the 447 positive samples, the unique combination sets of detected viral S-gene mutations suggest the following

TABLE 3 Clinical samples evaluation with multiplex qPCR test in early 2021^{a,b}

SARS-CoV-2	Patient samples (N)	Quantivirus SARS-COV-2 variant detection						Sensitivity (95% CI)	Specificity (95% CI)	PPV (%) (95% CI)	NPV (%) (95% CI)
		Variant identification			Variant identification						
		Mutant detection	Positive	Negative	Alpha	Early mutation/wild type	Negative				
Positive	374	N501Y	17	357	17	160/197	102	100% (0.77–1.00)	100% (0.99–1.00)	100% (0.77–1.00)	100% (0.99–1.00)
Negative	102	D614G	177	197				100% (0.97–1.00)	100% (0.98–1.00)	100% (0.97–1.00)	100% (0.98–1.00)
		ORF1ab	374	0				100% (0.99–1.00)	100% (0.96–1.00)	100% (0.99–1.00)	100% (0.96–1.00)
		Rp	476	0							

^aBio-Rad CFX384 qPCR instrument was used for this test; all of 374 positive samples were detected ORF1ab gene (wild type) positive; all variants data were confirmed with Sanger sequencing.
^bPositive predictive value (PPV) and negative predictive value (NPV) were calculated by comparing qPCR data with Sanger sequencing data for each sample.

TABLE 4 Summary of Delta variant detection for breakthrough patient samples from October to November 2021

qPCR kit	Patient sample (N)	Mutant target	Positive	Negative	Variant identification			Comparing with Sanger sequence	
					Delta	Delta+	Negative	PPV (%)	NPV (%)
Quantivirus SARS-CoV-2 delta plus	89	K417N	3	86	36	3	50	100 (95% CI: 0.888–1.0)	100 (95% CI: 0.911–1.0)
		T478K	39	50					
		Orf1ab	39	50					
		RP	89	0					

N501Y/D614G positive, while the remaining one sample was identified as the wild type without any detected mutations. On the other hand, the two SARS-CoV-2 negative samples tested negative using the same kit. All 11 samples were then sequenced by NGS using the CleanPlex SARS-CoV-2 FLEX kit, and all eight positive samples were confirmed as Alpha variants, while the one wild type was confirmed as wild type (Lineage 19A, Nextstrain clad), and the two negative samples were confirmed as negative (Table S9). We additionally sequenced and analyzed 13 positive samples (vaccinated breakthrough cases) and two negative samples collected in early 2022 using both the QuantiVirus SARS-CoV-2 Variants Detection Kit and the Illumina COVIDSeq Test. Again, the results were all congruent (Table S9). The data demonstrated that our RT-qPCR platform gave the same result as the Illumina COVIDSeq Test with 100% concordance.

DISCUSSION

We have described the successful development of a multiplex RT-qPCR testing platform for the rapid detection of SARS-CoV-2 variant strains using XNA-based molecular-clamping technology. In order to develop a sensitive and specific molecular-clamping assay, it is not only necessary to select a suitable set of primers and probe for a given mutant gene target, but it is also imperative to choose an XNA with appropriate sequences and desired performance characteristics. Our XNA selection process included a twofold approach. First, during the sequence design, we excluded those improper sequences with problematic features (e.g., sequences that were too long or too short, had too high or too low T_m values, high purine content, long purine stretch, or unwanted self-complementarity within or between XNA molecules). Second, for each mutation assay, we designed and synthesized multiple XNAs and compared their qPCR clamping robustness and clamping specificity (higher delta C_t difference between wild-type gene and mutation gene amplification). Notably, across the XNA groups, the XNA with a T_m of nearly 80°C stood out in the selection process, which is likely related to the established RT-qPCR temperature-cycling conditions (49).

Using this assay, we were able to track the changes in SARS-CoV-2 variants over time. Due to its higher transmissibility compared to the original SARS-CoV-2 strain, the D614G mutant of SARS-CoV-2 became the dominant variant in the beginning of 2021 in the United States. Its sub-clade, the N501Y mutation, independently emerged in the UK and South Africa and subsequently spread to North America during the first half of 2021. Our data indicated that the spread of this Alpha variant in northern California likely started in February 2021 because patient samples collected in January 2021 were all negative for the N501Y mutation, whereas the N501Y mutation was detectable in up to 5% of the samples collected in late February 2021. Although the U.S. CDC predicted that,

TABLE 5 Summary of Omicron variant detection for breakthrough patients during early 2022

qPCR kit	Patient sample (N)	Mutant target	Positive	Negative	Variant identification		Comparing with Sanger sequence	
					Omicron	Negative	PPV (%)	NPV (%)
Quantivirus SARS-CoV-2 variant detection	87	D614G	34	53	34	53	100 (95% CI: 0.87–1.0)	100 (95% CI: 0.92–1.0)
		N501Y						
		K417N						
		T478K						
		Orf1ab						
		RP	87	0				

TABLE 6 Summary of SARS-CoV-2 multivariant detection in January 2021 to January 2022^{a,b}

Sample total (N)	Status (n) pos/neg	Variant (n)	ORF1ab	D614G	N501Y	T478K	L452R	K417T	K417N	Variant identity		
649	447 (positive)	34	x	x	x	x			x	Omicron		
		36	x	x		x	x			Delta		
		3	x	x			x	x		x	Delta plus	
		5	x	x		x				x	Beta	
		17	x	x		x					Alpha	
		160	x	x							Early mutation	
		192	x								Wild type	
				202								Negative
			202 (negative)									

^aAll data were confirmed by comparing Sanger sequencing data.

^b"X" mean positive detection for each mutation.

according to an epidemiology model (50) at that time, the B.1.1.7 or Alpha variant would quickly dominate soon, the Alpha variant was soon largely replaced around the middle of 2021 by the even more potent and infectious Delta variant. The Delta variant in turn was soon replaced by the more contagious Omicron variants in late 2021 to early 2022. Our RT-qPCR data, confirmed by Sanger or NGS sequencing, were consistent with the sequential surges of COVID-19 cases caused by the D614G mutant and the Alpha, Delta, and Omicron variants in the United States from late 2020 to early 2022.

Availability of rapid and accurate testing platforms is critical for tackling the challenge of emerging SARS-CoV-2 variants, especially to understand breakthrough cases and provide appropriate treatments. Among the many detection methods, a qPCR-based platform could serve as a rapid, cost-effective, and practical testing tool for monitoring SARS-CoV-2 evolution.

A few biotechnology companies and academic institutions have been developing or have reported variant detection methods using qPCR. These methods can be categorized into one of the following: (i) mutant gene-specific or allele-specific primers and probe; (ii) spike gene target failure; (iii) E gene target failure; and (iv) ORF gene deletion (51–54). Some of the methods still require NGS to confirm the results, whereas other platforms require pre-testing by regular SARS-CoV-2 RT-qPCR before carrying out the variant assay. While these assays can be useful in certain circumstances (55), they are limited by the effectiveness and accuracy of the primers and/or probes used. This issue can limit the use of these variant testing methodologies. For example, the allele-specific PCR (AS-PCR) method quickly attracted attention for mutation detection by some test developers (56). However, AS-PCR has two inherent shortcomings: (i) due to the fixed 3' end of the allele-specific primer, it is not always feasible to choose optimal primers for PCR amplification; and (ii) high purity DNA/RNA is essential as low-quality or crude DNA/RNA samples are prone to providing inconclusive results (57).

However, XNA-based qPCR can overcome these aforementioned shortcomings because XNA sequences can be designed with more flexibility compared to restrictive primer sequence design (58), and XNA-based qPCR works better with samples of low-concentration DNA/RNA and samples with high-background mutant gene targets. Noteworthy is that the XNAs have been made with practically appealing cost and high efficiency—the unit cost of XNA is between the oligo primer and the Taqman PCR probe. Compared to AS-PCR, as demonstrated in our studies, the XNA-based RT-qPCR assay can achieve a lower LoD, about 100 copies/mL for variant detection.

Of particular note, the determined LoD 100 copies/mL here refers to the original, collected patient sample, not the extracted concentrated RNA sample. Based on the described procedure of virus RNA extraction, we can calculate the copy number per PCR (10 µL/reaction) to be 2.8–3.7 copies per reaction, this result was well consistent with the theoretical considerations in the MIQE guidelines (59). The molecular-clamping technology used in this study increases the sensitivity and specificity of conventional qPCR as wild-type background amplification is minimized by this method. This XNA clamping-based RT-qPCR assay can also improve the multiplexing of targets, as we were

able to differentiate the wild-type SARS-CoV-2 (ORF1ab gene for wild-type detection) and its variants in a single run. In addition to detecting the existing SARS-CoV-2 mutations D614G and N501Y, the assay described here also detected L452R, T478K, K417N, and K417T. The presence or absence of these major mutations can be used to suggest and identify all five VOCs based on their unique mutation profiles VOCs (60) (Table S1). As a result, this assay covers almost all SARS-CoV-2 variants (both VOC and VOI), including Delta, "Delta Plus," and Omicron, all of which have the potential to breakthrough vaccine-induced protective immunity (60). Of note, it is natural to consider an additional E484A mutation to help detect Omicron, but we finally opted to use the unique combination of N501Y, T478K, and K417N mutations to detect Omicron by using the already-available K417 XNA and K417N primers/probe to maximally speed up the variants assay efforts.

This strategy can provide an effective, rapid, and easily adoptable testing platform for known and emerging SARS-CoV-2 variants in the future, and this technology can be readily adopted by clinical laboratories that perform routine SARS-CoV-2 RT-qPCR testing. Given the established stages of the COVID-19 pandemic, future new variant(s) with increased transmissibility could potentially lead to an exponential increase in cases and/or deaths (13). Therefore, timely and convenient detection of the concerning variants is of high importance.

There are 15 mutations in the RBD of Omicron variant, including Q393R, K417N, T478K, S477N, E484A, G496S, Q498R, Y505H, G446S, N501Y, N440K, S75F, S373P, S371L, and G339D (29). This poses a significant challenge for RT-qPCR-based assays as this high number of mutations can render some primers dysfunctional. Interestingly, the N501Y/D614G primer, which works well for Alpha, Beta, and Delta variant detection, had a lower signal for N501Y/D614G detection in Omicron variants due to the fact that the original N501Y/D614G primer was located in Omicron high mutant region (data in Tables S8 and S9). To avoid this issue in the future, we have re-designed a new set of "degenerate" primers and probes to better fit the mutant area in Omicron, while still covering the original N501Y/D614G region in Alpha, Beta, and Delta variants. Evaluation of this revised N501Y/D614G assay design showed that it is effective in distinguishing these variants effectively. This illustrates how this method can be rapidly adapted to test new and emerging variants while still successfully detecting all known variants.

Despite this issue, our XNA-based RT-qPCR assay kit was able to identify the Omicron variant because of the presence of three mutations N501Y/D614G/T478K/K417N and the absence of L452R mutation, a combination unique to the Omicron variant. Indeed, we were able to detect and differentiate Omicron and Delta variants and successfully detected the Omicron variant in patients who had breakthrough SARS-CoV-2 infections in early January 2022.

Noteworthy, compared to the NGS method in existing SARS-CoV-2 variant detection, the cost (time, reagents, and other resources) of using RT-PCR is considerably cheaper. However, NGS (when WGS is used) is the Gold-Standard method in discovering the emerging unknown variants.

Furthermore, compared to the basic RT-qPCR method, our PCR method using XNA molecular-clamping technology offers a more straightforward and easier interpretation of experimental data, thus saving time and resources; meanwhile, the design and synthesis of XNA molecular claspers were optimized so as to merely add a small extra cost (lower than the synthetic cost of well-known Taqman probe).

The development of mutation-specific RT-qPCR, combined with wild-type gene blockers, has utilized various techniques. For instance, Chaintoutis et al. introduced a method using locked nucleic acids (LNA)-modified oligonucleotides to distinguish target mutations from wild-type genes. This method incorporated LNA modifications into TaqMan probes and non-extendable blocking oligonucleotides to enhance the T_m , thereby improving the test's specificity and sensitivity. However, the same LNA modifications that increased the T_m also posed limitations. LNA-modified probes could potentially hybridize non-specifically, and the non-extendable oligonucleotide blockers,

which had a lower T_m than the LNA probes, might not fully block the wild-type genes (61, 62).

In addition to RT-qPCR and genotyping methods, electrophoresis-based fragment analysis presents an alternative for detecting COVID-19 variants. Clark et al. designed a multiplex fragment analysis technique called CoVarScan, which distinguishes PCR target variants by size and fluorescent color. This method targets eight mutation hotspots in SARS-CoV-2 VOCs, with fragments analyzed via capillary electrophoresis (ABI 3730XL). However, the sensitivity of CoVarScan does not match that of RT-qPCR, and the requirement for capillary electrophoresis constrains its broader application (63).

In contrast to this novel XNA-assisted RT-qPCR method for RNA detection of SARS-CoV-2 variants, there have been no studies reported to our best knowledge to utilize the molecular clampers like PNAs in RT-qPCR SARS-CoV-2 variant detection and identification. One close example was to use PNA clampers in digital PCR for variant detection reported most recently (64). Additionally, other studies using PNA clampers have developed a non-PCR method, such as a biosensor technique (65) and loop-mediated isothermal amplification based on SARS-CoV-2 variants detection (66). These also reflect the wide applications, flexibility, and vast potential of applying molecular clampers like XNAs in clinically relevant molecular diagnosis.

There are a few limitations in this study: (i) relatively limited number of samples ($n < 1,000$); (ii) less attention to newer variants or subvariants due to our main focus on the earlier, most threatening VOCs from Alpha to Delta variants prior to the Omicron variant in the pandemic; (iii) multiple tubes instead of a single tube is included in this kit due to the detection of >4 targets and also consideration of minimizing the assay interference between close targets (e.g., L452 and T478).

In the further development of viral variant detection methods, the following directions may be worthy of attention: PCR probe improvement with even higher specificity; a higher degree of multiplexing; artificial-intelligence-assisted assay design for rapid multiplexing optimization; more integration of multiple techniques/platforms to better supplement each other (isothermal, qPCR, digital PCR, and NGS) for different purposes of variant detection (monitoring, screening, point-of-care, etc.).

In summary, we have developed a multiplex RT-qPCR testing platform for the rapid detection of SARS-CoV-2 variants using the XNA-based molecular-clamping technology. This testing platform can be adopted by laboratories that perform routine SARS-CoV-2 PCR testing, providing a rapid and cost-effective method in lieu of NGS-based assays, for detecting, differentiating, and monitoring SARS-CoV-2 variants. This assay is easily scalable to any new variant(s) should it emerge.

ACKNOWLEDGMENTS

We are grateful to DiaCarta Clinical Laboratory colleagues Yulia Loginova and Eric Abbott for their help in obtaining clinical samples. We also thank the UCSF DeGrado Lab for using MALDI-TOF mass spectrometry instrument.

Data collection, analysis, and interpretation: S.S., M.J., A.Y.F., J.L., L.P., C.Z., M.Y.S., C.M.L., M.R., M.K., D.K., J.N., and I.A. Chemical synthesis, purification, and analysis: A.Y.F. and D.C. Clinical sample processing: J.L., S.S., L.P., and J.A.P. Writing of original draft: S.S., M.J., A.Y.F., and M.Y.S. Revision and editing: A.Y.F., M.Y.S., S.S., D.K., J.A.P., and C.M.L. Conceptualization: M.Y.S., A.Y.F., and Q.S. Project planning and administration: M.S. and A.Z.

S.S., A.Y.F., M.J., J.L., Z.C., L.P., D.C., Q.S., A.Z., and M.Y.S. have been employed by DiaCarta Inc. during the period of this work.

AUTHOR AFFILIATIONS

¹DiaCarta Inc., Pleasanton, California, USA

²Franciscan University of Steubenville, Steubenville, Ohio, USA

³Department of Laboratory Medicine, University of California San Francisco and San Francisco VA Health Care System, San Francisco, California, USA

AUTHOR ORCID*s*

Shuo Shen  <http://orcid.org/0000-0002-3517-8971>

Andrew Y. Fu  <http://orcid.org/0000-0002-1507-0787>

Michael Y. Sha  <http://orcid.org/0000-0002-4552-6079>

AUTHOR CONTRIBUTIONS

Shuo Shen, Data curation, Formal analysis, Investigation, Methodology | Andrew Y. Fu, Conceptualization, Data curation, Investigation, Methodology, Writing – original draft | Maidar Jamba, Data curation, Formal analysis, Methodology | Jonathan Li, Data curation, Formal analysis, Methodology | Zhen Cui, Data curation, Formal analysis | Larry Pastor, Data curation, Formal analysis | Daniel Cataldi, Data curation | Qing Sun, Conceptualization, Methodology, Supervision | Joseph A. Pathakamuri, Investigation, Supervision, Writing – review and editing | Daniel Kuebler, Supervision, Writing – review and editing | Michael Rohall, Data curation | Madison Krohn, Data curation | Daniel Kissinger, Data curation | Jocelyn Neves, Data curation | Isaac Archibeque, Data curation | Aiguo Zhang, project administration, Resources, Supervision | Chuanyi M. Lu, Methodology, Supervision, Writing – review and editing | Michael Y. Sha, Conceptualization, Methodology, Resources, Supervision, Writing – original draft, Writing – review and editing

DATA AVAILABILITY

The data generated and analyzed during this study are available from the corresponding author upon reasonable request.

ETHICS APPROVAL

Relevant portions of the study were approved by the institutional review board (IRB) at University of California at San Francisco (UCSF) (UCSF IRB #11-05207) as a no-subject contact study with waiver of consent and as exempt under category 4 and by the IRB at Franciscan University (FUS IRB #2021-07). All experiments were performed in accordance with relevant guidelines and regulations.

ADDITIONAL FILES

The following material is available [online](#).

Supplemental Material

Supplemental material (Spectrum04248-23-s0001.docx). Tables S1 to S10; Fig. S1.

Open Peer Review

PEER REVIEW HISTORY (review-history.pdf). An accounting of the reviewer comments and feedback.

REFERENCES

1. COVID-19 cases. WHO COVID-19 dashboard. Available from: <https://data.who.int/dashboards/covid19/cases>. Retrieved 13 Feb 2023.
2. CDC. 2020. COVID data tracker. Centers for Disease Control and Prevention. Available from: <https://covid.cdc.gov/covid-data-tracker>. Retrieved 13 Feb 2023.
3. Rambaut A, Loman N, Pybus O, Barclay W, Barrett J, Carabelli A, Connor T, Peacock T, Robertson DL, Volz E, on behalf of COVID-19 Genomics Consortium UK. 2020. Preliminary genomic characterization of an emergent SARS-CoV-2 lineage in the UK defined by a novel set of spike mutations. <https://virological.org/t/preliminary-genomic-characterisation-of-an-emergent-sars-cov-2-lineage-in-the-uk-defined-by-a-novel-set-of-spike-mutations/563>.
4. Tegally H, Wilkinson E, Giovanetti M, Iranzadeh A, Fonseca V, Giandhari J, Doolabh D, Pillay S, San EJ, Msomi N, et al. 2021. Detection of a SARS-CoV-2 variant of concern in South Africa. *Nature New Biol* 592:438–443. <https://doi.org/10.1038/s41586-021-03402-9>
5. Faria NR, Mellan TA, Whittaker C, Claro IM, Candido D da S, Mishra S, Crispim MAE, Sales FCS, Hawryluk I, McCrone JT, et al. 2021. Genomics and epidemiology of the P.1 SARS-CoV-2 lineage in Manaus, Brazil. *Science* 372:815–821. <https://doi.org/10.1126/science.abh2644>
6. Zhang W, Davis BD, Chen SS, Sincuir Martinez JM, Plummer JT, Vail E. 2021. Emergence of a novel SARS-CoV-2 variant in Southern California. *JAMA* 325:1324–1326. <https://doi.org/10.1001/jama.2021.1612>

7. Deng X, Garcia-Knight MA, Khalid MM, Servellita V, Wang C, Morris MK, Sotomayor-González A, Glasner DR, Reyes KR, Gliwa AS, et al. 2021. Transmission, infectivity, and neutralization of a spike L452R SARS-CoV-2 variant. *Cell* 184:3426–3437. <https://doi.org/10.1016/j.cell.2021.04.025>
8. Carroll T, Fox D, van Doremalen N, Ball E, Morris MK, Sotomayor-Gonzalez A, Servellita V, Rustagi A, Yinda CK, Fritts L, Port JR, Ma Z-M, Holbrook MG, Schulz J, Blish CA, Hanson C, Chiu CY, Munster V, Stanley S, Miller CJ. 2022. The B.1.427/1.429 (epsilon) SARS-CoV-2 variants are more virulent than ancestral B.1 (614G) in Syrian hamsters. *PLoS Pathog* 18:e1009914. <https://doi.org/10.1371/journal.ppat.1009914>
9. WHO. 2021. Classification of Omicron (B.1.1.529): SARS-CoV-2 variant of concern. Available from: [https://www.who.int/news/item/26-11-2021-classification-of-omicron-\(b.1.1.529\)-sars-cov-2-variant-of-concern](https://www.who.int/news/item/26-11-2021-classification-of-omicron-(b.1.1.529)-sars-cov-2-variant-of-concern). Retrieved 13 Feb 2023.
10. Tracking SARS-CoV-2 variants. 2024. Available from: <https://www.who.int/activities/tracking-SARS-CoV-2-variants>. Retrieved 10 Jun 2024.
11. CDC. 2020. COVID data tracker. Centers for Disease Control and Prevention. Available from: <https://covid.cdc.gov/covid-data-tracker>. Retrieved 10 Jun 2024.
12. SARS-CoV-2 variants of concern as of 31 May 2024. 2024. Available from: <https://www.ecdc.europa.eu/en/covid-19/variants-concern>. Retrieved 10 Jun 2024.
13. Wise J. 2020. COVID-19: new coronavirus variant is identified in UK. *BMJ*:m4857. <https://doi.org/10.1136/bmj.m4857>
14. ECDC. 2023 Risk of spread of new SARS-CoV-2 variants of concern in the EU/EEA - first update - 21 January 2021. ECDC Stockh
15. CDC. 2020. US COVID-19 cases caused by variants. Centers for Disease Control and Prevention. Available from: <https://www.cdc.gov/coronavirus/2019-ncov/transmission/variant-cases.html>. Retrieved 13 Feb 2023.
16. Helix. The Helix COVID-19 surveillance dashboard. Available from: <https://www.helix.com/pages/helix-covid-19-surveillance-dashboard>. Retrieved 13 Feb 2023.
17. Davies NG, Abbott S, Barnard RC, Jarvis CI, Kucharski AJ, Munday JD, Pearson CAB, Russell TW, Tully DC, Washburne AD, Wenseleers T, Gimma A, Waites W, Wong KLM, van Zandvoort K, Silverman JD, CMMID COVID-19 Working Group, COVID-19 Genomics UK (COG-UK) Consortium, Diaz-Ordaz K, Keogh R, Eggo RM, Funk S, Jit M, Atkins KE, Edmunds WJ. 2021. Estimated transmissibility and impact of SARS-CoV-2 lineage B.1.1.7 in England. *Science* 372:eabg3055. <https://doi.org/10.1126/science.abg3055>
18. Leung K, Shum MH, Leung GM, Lam TT, Wu JT. 2021. Early transmissibility assessment of the N501Y mutant strains of SARS-CoV-2 in the United Kingdom, October to November 2020. *Euro Surveill* 26:2002106. <https://doi.org/10.2807/1560-7917.ES.2020.26.1.2002106>
19. CDC. 2020. COVID data tracker. Centers for Disease Control and Prevention. Available from: <https://covid.cdc.gov/covid-data-tracker/#variant-summary>. Retrieved 13 Feb 2023.
20. Korber B, Fischer WM, Gnanakaran S, Yoon H, Theiler J, Abfalterer W, Hengartner N, Giorgi EE, Bhattacharya T, Foley B, Hastie KM, Parker MD, Partridge DG, Evans CM, Freeman TM, de Silva TI, McDanal C, Perez LG, Tang H, Moon-Walker A, Whelan SP, LaBranche CC, Saphire EO, Montefiori DC, Sheffield COVID-19 Genomics Group. 2020. Tracking changes in SARS-CoV-2 spike: evidence that D614G increases infectivity of the COVID-19 virus. *Cell* 182:812–827. <https://doi.org/10.1016/j.cell.2020.06.043>
21. Hou YJ, Chiba S, Halfmann P, Ehre C, Kuroda M, Dinno KH III, Leist SR, Schäfer A, Nakajima N, Takahashi K, et al. 2020. SARS-CoV-2 D614G variant exhibits efficient replication *ex vivo* and transmission *in vivo*. *Science* 370:1464–1468. <https://doi.org/10.1126/science.abe8499>
22. Volz E, Hill V, McCrone JT, Price A, Jorgensen D, O'Toole Á, Southgate J, Johnson R, Jackson B, Nascimento FF, et al. 2021. Evaluating the effects of SARS-CoV-2 spike mutation D614G on transmissibility and pathogenicity. *Cell* 184:64–75. <https://doi.org/10.1016/j.cell.2020.11.020>
23. Hagen A. SARS-CoV-2 variants vs vaccines. ASM.org. Available from: <https://asm.org/443/Articles/2021/February/SARS-CoV-2-Variants-vs-Vaccines>. Retrieved 25 May 2024.
24. Singh J, Rahman SA, Ehtesham NZ, Hira S, Hasnain SE. 2021. SARS-CoV-2 variants of concern are emerging in India. *Nat Med* 27:1131–1133. <https://doi.org/10.1038/s41591-021-01397-4>
25. Konings F, Perkins MD, Kuhn JH, Pallen MJ, Alm EJ, Archer BN, Barakat A, Bedford T, Bhiman JN, Cally L, et al. 2021. SARS-CoV-2 variants of interest and concern naming scheme conducive for global discourse. *Nat Microbiol* 6:821–823. <https://doi.org/10.1038/s41564-021-00932-w>
26. Expert reaction to cases of variant B.1.617 (the “Indian variant”) being investigated in the UK. Science Media Centre. Available from: <https://www.sciencemediacentre.org/expert-reaction-to-cases-of-variant-b-1-617-the-indian-variant-being-investigated-in-the-uk>. Retrieved 25 May 2024.
27. Li B, Deng A, Li K, Hu Y, Li Z, Shi Y, Xiong Q, Liu Z, Guo Q, Zou L, et al. 2022. Viral infection and transmission in a large, well-traced outbreak caused by the SARS-CoV-2 Delta variant. *Nat Commun* 13:460. <https://doi.org/10.1038/s41467-022-28089-y>
28. Callaway E. 2021. Heavily mutated Omicron variant puts scientists on alert. *Nature New Biol* 600:21–21. <https://doi.org/10.1038/d41586-021-03552-w>
29. European Centre for Disease Prevention and Control. 2021. Implications of the emergence and speed of the SARS-CoV-2 B.1.1.529 variant of concern (Omicron), for the EU/EEA. Stockholm ECDC
30. UK Health Security Agency. 2022. SARS-CoV-2 variants of concern and variants under investigation in England: technical briefing 37. London UK Health Security Agency
31. Yelagandula R, Bykov A, Vogt A, Heinen R, Özkan E, Strobl MM, Baar JC, Uzunova K, Hajdusits B, Kordic D, Suljic E, Kurtovic-Kozaric A, Izetbegovic S, Schaeffer J, Hufnagl P, Zoufaly A, Seitz T, Födinger M, Allerberger F, Stark A, Cochella L, Elling U, VCDI. 2021. Multiplexed detection of SARS-CoV-2 and other respiratory infections in high throughput by SARSseq. *Nat Commun* 12:3132. <https://doi.org/10.1038/s41467-021-22664-5>
32. Breakthrough Technologies. 2022. MIT technology review. Available from: <https://www.technologyreview.com/2022/02/23/1045416/10-breakthrough-technologies-2022>. Retrieved 25 May 2024.
33. Bedotto M, Fournier P-E, Houhamdi L, Levasseur A, Delerac J, Pinault L, Padane A, Chamieh A, Tissot-Dupont H, Brouqui P, Sokhna C, Azar E, Saile R, Mboup S, Bitam I, Colson P, Raoult D. 2021. Implementation of an in-house real-time reverse transcription-PCR assay for the rapid detection of the SARS-CoV-2 Marseille-4 variant. *J Clin Virol* 139:104814. <https://doi.org/10.1016/j.jcv.2021.104814>
34. Annavajhala MK, Mohri H, Wang P, Nair M, Zucker JE, Sheng Z, Gomez-Simmonds A, Kelley AL, Tagliavia M, Huang Y, Bedford T, Ho DD, Uhlemann A-C. 2021. Emergence and expansion of SARS-CoV-2 B.1.526 after identification in New York. *Nature New Biol* 597:703–708. <https://doi.org/10.1038/s41586-021-03908-2>
35. Wang Q, Chen L, Long Y, Tian H, Wu J. 2013. Molecular beacons of xenonucleic acid for detecting nucleic acid. *Theranostics* 3:395–408. <https://doi.org/10.7150/tno.5935>
36. Kyger EM, Krevolin MD, Powell MJ. 1998. Detection of the hereditary hemochromatosis gene mutation by real-time fluorescence polymerase chain reaction and peptide nucleic acid clamping. *Anal Biochem* 260:142–148. <https://doi.org/10.1006/abio.1998.2687>
37. Sun Q, Pastor L, Du J, Powell MJ, Zhang A, Bodmer W, Wu J, Zheng S, Sha MY. 2021. A novel xenonucleic acid-mediated molecular clamping technology for early colorectal cancer screening. *PLoS One* 16:e0244332. <https://doi.org/10.1371/journal.pone.0244332>
38. Lee B, Chern A, Fu AY, Zhang A, Sha MY. 2024. A highly sensitive XNA-based RT-qPCR assay for the identification of ALK, RET, and ROS1 fusions in lung cancer. *Diagn (Basel)* 14:488. <https://doi.org/10.3390/diagnostics14050488>
39. Shen S, Brown R, Kim D, Pastor L, Fu AY, Kaur P, Babu A, Liu W, Zhang A, Tanaka H, Sha MY. 2024. Abstract 7295: a novel XNA technology-based assay to detect JAK2 V617F mutation by real-time PCR. *Cancer Res* 84:7295–7295. <https://doi.org/10.1158/1538-7445.AM2024-7295>
40. Liu W, Brown R, Fu A, Shen S, Sha M, Zhang A. 2023. Abstract 6681: XNA increases assay sensitivity in sanger sequencing, qPCR, NGS and CRISPR mutant screening. *Cancer Res* 83:6681–6681. <https://doi.org/10.1158/1538-7445.AM2023-6681>
41. Anosova I, Kowal EA, Dunn MR, Chaput JC, Van Horn WD, Egli M. 2016. The structural diversity of artificial genetic polymers. *Nucleic Acids Res* 44:1007–1021. <https://doi.org/10.1093/nar/gkv1472>
42. Sun Q, Li J, Ren H, Pastor L, Loginova Y, Madej R, Taylor K, Wong JK, Zhang Z, Zhang A, Lu CM, Sha MY. 2021. Saliva as a testing specimen with or without pooling for SARS-CoV-2 detection by multiplex RT-PCR

- test. *PLoS One* 16:e0243183. <https://doi.org/10.1371/journal.pone.0243183>
43. Pinheiro VB, Holliger P. 2014. Towards XNA nanotechnology: new materials from synthetic genetic polymers. *Trends Biotechnol* 32:321–328. <https://doi.org/10.1016/j.tibtech.2014.03.010>
44. FDA EUA-approved DiaCarta QuantiVirus SARS-CoV-2 test kit. 2020. Available from: <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/in-vitro-diagnostics-euas-molecular-diagnostic-tests-sars-cov-2>. Retrieved 13 Feb 2023.
45. Fernandes JD, Hinrichs AS, Clawson H, Gonzalez JN, Lee BT, Nassar LR, Raney BJ, Rosenbloom KR, Nerli S, Rao AA, Schmelter D, Fyfe A, Maulding N, Zweig AS, Lowe TM, Ares M, Corbet-Detig R, Kent WJ, Haussler D, Haussler M. 2020. The UCSC SARS-CoV-2 genome browser. *Nat Genet* 52:991–998. <https://doi.org/10.1038/s41588-020-0700-8>
46. Afgan E, Baker D, Batut B, van den Beek M, Bouvier D, Cech M, Chilton J, Clements D, Coraor N, Grüning BA, Guerler A, Hillman-Jackson J, Hiltmann S, Jalili V, Rasche H, Soranzo N, Goecks J, Taylor J, Nekrutenko A, Blankenberg D. 2018. The Galaxy platform for accessible, reproducible and collaborative biomedical analyses: 2018 update. *Nucleic Acids Res* 46:W537–W544. <https://doi.org/10.1093/nar/gky379>
47. Blankenberg D, Johnson JE, Team TG, Taylor J, Nekrutenko A. 2014. Wrangling Galaxy's reference data. *Bioinformatics* 30:1917–1919. <https://doi.org/10.1093/bioinformatics/btu119>
48. Grubaugh ND, Gangavarapu K, Quick J, Matteson NL, De Jesus JG, Main BJ, Tan AL, Paul LM, Brackney DE, Grewal S, Gurfield N, Van Rompay KKA, Isern S, Michael SF, Coffey LL, Loman NJ, Andersen KG. 2019. An amplicon-based sequencing framework for accurately measuring intrahost virus diversity using PrimalSeq and iVar. *Genome Biol* 20:8. <https://doi.org/10.1186/s13059-018-1618-7>
49. Giesen U, Kleider W, Berding C, Geiger A, Orum H, Nielsen PE. 1998. A formula for thermal stability (T_m) prediction of PNA/DNA duplexes. *Nucleic Acids Res* 26:5004–5006. <https://doi.org/10.1093/nar/26.21.5004>
50. Galloway SE, Paul P, MacCannell DR, Johansson MA, Brooks JT, MacNeil A, Slayton RB, Tong S, Silk BJ, Armstrong GL, Biggerstaff M, Dugan VG. 2021. Emergence of SARS-CoV-2 B.1.1.7 lineage - United States, December 29, 2020-January 12, 2021. *MMWR* 70:95–99. <https://doi.org/10.15585/mmwr.mm7003e2>
51. Harper H, Burrridge A, Winfield M, Finn A, Davidson A, Matthews D, Hutchings S, Vipond B, Jain N, Edwards K, Barker G, COVID-19 Genomics UK (COG-UK) Consortium. 2021. Detecting SARS-CoV-2 variants with SNP genotyping. *PLoS One* 16:e0243185. <https://doi.org/10.1371/journal.pone.0243185>
52. Washington NL, Gangavarapu K, Zeller M, Bolze A, Cirulli ET, Schiabor Barrett KM, Larsen BB, Anderson C, White S, Cassens T, et al. 2021. Emergence and rapid transmission of SARS-CoV-2 B.1.1.7 in the United States. *Cell* 184:2587–2594. <https://doi.org/10.1016/j.cell.2021.03.052>
53. Artesi M, Bontems S, Göbbels P, Franckh M, Maes P, Boreux R, Meex C, Melin P, Hayette M-P, Bours V, Durkin K. 2020. A recurrent mutation at position 26340 of SARS-CoV-2 is associated with failure of the E gene quantitative reverse transcription-PCR utilized in a commercial dual-target diagnostic assay. *J Clin Microbiol* 58:e01598-20. <https://doi.org/10.1128/JCM.01598-20>
54. COVID-19 test makers are adapting for variants. 2021. *BioWorld*. Available from: <https://www.bioworld.com/articles/503479-covid-19-test-makers-are-adapting-for-variants>. Retrieved 25 May 2024.
55. De Pace V, Bruzzzone B, Orsi A, Ricucci V, Domnich A, Guarona G, Randazzo N, Stefanelli F, Battolla E, Dusi PA, Lillo F, Icardi G. 2022. Comparative analysis of five multiplex RT-PCR assays in the screening of SARS-CoV-2 variants. *Microorganisms* 10:306. <https://doi.org/10.3390/microorganisms10020306>
56. Brito-Mutunayagam S, Maloney D, McAllister G, Dewar R, McHugh M, Templeton K. 2022. Rapid detection of SARS-CoV-2 variants using allele-specific PCR. *J Virol Methods* 303:114497. <https://doi.org/10.1016/j.jviromet.2022.114497>
57. Myakishev MV, Khripin Y, Hu S, Hamer DH. 2001. High-throughput SNP genotyping by allele-specific PCR with universal energy-transfer-labeled primers. *Genome Res* 11:163–169. <https://doi.org/10.1101/gr.157901>
58. Oligo design tools. 2023. Available from: www.thermofisher.com/us/en/home/life-science/oligonucleotides-primers-probes-genes/custom-dna-oligos/oligo-design-tools.html. Retrieved 13 Feb 2023.
59. Bustin SA, Benes V, Garson JA, Hellems J, Huggett J, Kubista M, Mueller R, Nolan T, Pfaffl MW, Shipley GL, Vandesompele J, Wittwer CT. 2009. The MIQE guidelines: minimum information for publication of quantitative real-time PCR experiments. *Clin Chem* 55:611–622. <https://doi.org/10.1373/clinchem.2008.112797>
60. Challen R, Brooks-Pollock E, Read JM, Dyson L, Tsaneva-Atanasova K, Danon L. 2021. Risk of mortality in patients infected with SARS-CoV-2 variant of concern 202012/1: matched cohort study. *BMJ* 372:n579. <https://doi.org/10.1136/bmj.n579>
61. Chaintoutis SC, Chassalevris T, Balaska S, Mouchtaropoulou E, Tsiolas G, Vlatakis I, Tychala A, Koutsoulis D, Argiriou A, Skoura L, Dovas CI. 2021. A novel real-time RT-PCR-based methodology for the preliminary typing of SARS-CoV-2 variants, employing non-extendable LNA oligonucleotides and three signature mutations at the spike protein receptor-binding domain. *Life (Basel)* 11:1015. <https://doi.org/10.3390/life11101015>
62. Owczarzy R, You Y, Groth CL, Tataurov AV. 2011. Stability and mismatch discrimination of locked nucleic acid-DNA duplexes. *Biochemistry* 50:9352–9367. <https://doi.org/10.1021/bi200904e>
63. Clark AE, Wang Z, Ostman E, Zheng H, Yao H, Cantarel B, Kanchwala M, Xing C, Chen L, Irwin P, Xu Y, Oliver D, Lee FM, Gagan JR, Filkins L, Muthukumar A, Park JY, Sarode R, SoRelle JA. 2022. Multiplex fragment analysis for flexible detection of all SARS-CoV-2 variants of concern. *Clin Chem* 68:1042–1052. <https://doi.org/10.1093/clinchem/hvac081>
64. Zhang L, Parvin R, Lin S, Chen M, Zheng R, Fan Q, Ye F. 2024. Peptide nucleic acid clamp-assisted photothermal multiplexed digital pcr for identifying SARS-CoV-2 variants of concern. *Adv Sci (Weinh)* 11:e2306088. <https://doi.org/10.1002/advs.202306088>
65. Li Y, Zhao S, Xu Z, Qiao X, Li M, Li Y, Luo X. 2023. Peptide nucleic acid and antifouling peptide based biosensor for the non-fouling detection of COVID-19 nucleic acid in saliva. *Biosens Bioelectron* 225:115101. <https://doi.org/10.1016/j.bios.2023.115101>
66. Iijima T, Sakai J, Kanamori D, Ando S, Nomura T, Tisi L, Kilgore PE, Percy N, Kohase H, Hayakawa S, Maesaki S, Hoshino T, Seki M. 2023. A new method to detect variants of SARS-CoV-2 using reverse transcription loop-mediated isothermal amplification combined with a bioluminescent assay in real time (RT-LAMP-BART). *Int J Mol Sci* 24:10698. <https://doi.org/10.3390/ijms241310698>