

Published in final edited form as:

Mass Spectrom Rev. 2007 ; 26(1): 51–78.

MASS SPECTROMETRY-BASED METABOLOMICS

Katja Dettmer¹, Pavel A. Aronov¹, and Bruce D. Hammock^{1,2,*}

1 Department of Entomology, University of California at Davis, Davis, California 95616

2 Cancer Research Center, University of California at Davis, Davis, California 95616

Abstract

This review presents an overview of the dynamically developing field of mass spectrometry-based metabolomics. Metabolomics aims at the comprehensive and quantitative analysis of wide arrays of metabolites in biological samples. These numerous analytes have very diverse physico-chemical properties and occur at different abundance levels. Consequently, comprehensive metabolomics investigations are primarily a challenge for analytical chemistry and specifically mass spectrometry has vast potential as a tool for this type of investigation. Metabolomics require special approaches for sample preparation, separation, and mass spectrometric analysis. Current examples of those approaches are described in this review. It primarily focuses on metabolic fingerprinting, a technique that analyzes all detectable analytes in a given sample with subsequent classification of samples and identification of differentially expressed metabolites, which define the sample classes. To perform this complex task, data analysis tools, metabolite libraries, and databases are required. Therefore, recent advances in metabolomics bioinformatics are also discussed.

Keywords

metabolomics; metabolic fingerprinting; metabolic profiling; lipidomics; mass spectrometry

I. INTRODUCTION

Dramatic technological advances in the biological sciences over the past few years have forged a new era of research including the emerging field of systems biology. Although the understanding of living organisms at the molecular system level is still in its infancy, it is evident that comprehensive investigations of the “omics cascade” with genomics, transcriptomics, proteomics, and metabolomics are important building blocks and will play a central role in this new science (see Fig. 1). The integrative analysis of an organism’s response to a perturbation on the transcriptome, proteome, and metabolome levels will lead to a better understanding of the biochemical and biological mechanisms in complex systems. However, whereas genomics, transcriptomics, and proteomics have made significant strides in technology development, the tools for the comprehensive examination of the metabolome are still emerging (Bino et al., 2004). Although metabolomics is the endpoint of the “omics cascade” and is the closest to phenotype, there is no single-instrument platform that currently can analyze all metabolites. Possibly, because there is at least the perception that the other “omic” approaches can be handled by a single platform, metabolomics has lagged behind the

*Correspondence to: Bruce D. Hammock, Department of Entomology, One Shields Avenue, University of California, Davis, CA 95616.
E-mail: bdhammock@ucdavis.edu

Contract grant sponsor: NIEHS; Contract grant numbers: R37 ES002710, R01 ES013933, P30 ES05707; Contract grant sponsor: NIEHS Superfund Basic Research Program; Contract grant number: P42 ES04699; Contract grant sponsor: NIEHS Center for Children’s Environmental Health & Disease Prevention; Contract grant number: P01 ES11269; Contract grant sponsor: Cure Autism Now Foundation Frank del Olmo Memorial Award; Contract grant sponsor: German Research Foundation.

other technologies. This is illustrated in Figure 2, showing the bibliographic search containing the words metabolomics, metabonomics, and proteomics in Chemical Abstracts Plus (SciFinder Scholar). While in 1999 three articles containing the keywords metabolomics or metabonomics were published, the number increased to 147 articles in 2003 and 203 in 2004. Moreover, the journal *Metabolomics* (Springer) was recently launched, which is dedicated to publish research results related to metabolomics technology development, data analysis and storage, integrated studies with other “omics” techniques, and metabolomics applications. The rising number of publications in the field demonstrates that metabolomics is not just a new “omics” word but a valuable emerging tool to study phenotype and changes in phenotype caused by environmental influences, disease, or changes in genotype. The comprehensive investigation of the metabolome is being complicated by its enormous complexity and dynamics. Metabolite distributions are subjected to high temporal and spatial variability; for example, circadian fluctuations in mammalian organisms are well known. In addition, diet-dependent biological variability in mammalian systems can complicate the analysis (Vigneau-Callahan et al., 2001). A careful experimental design is therefore mandatory for the success of these types of investigations. The metabolome represents a vast number of components that belong to a wide variety of compound classes, such as amino acids, lipids, organic acids, nucleotides, etc. These compounds are very diverse in their physical and chemical properties and occur in a wide concentration range. For example, within lipids alone, not only high-abundance compounds, such as fatty acids, triglycerides, or phospholipids, are encountered, but also trace level components with important regulatory effects, such as eicosanoids derived from arachidonic acid. According to Beecher, 2,000 major metabolites seems to be a good estimate for humans (Beecher, 2003). This number can of course be vastly larger as one considers secondary metabolites. Some of these metabolites may be chemical mediators of great biological importance. Up to 200,000 metabolites can be encountered in the plant kingdom (Weckwerth, 2003). Consequently, studying the metabolome is a major challenge to analytical chemistry and a metabolomic analysis in its true sense, namely the quantitative analysis of all metabolites, cannot be achieved with the current analytical instrumentation.

Currently, two complementary approaches are used for metabolomic investigations: metabolic profiling and metabolic fingerprinting (see Fig. 3) (Dettmer & Hammock, 2004). A summary of metabolomics-related definitions is given in Table 1. Metabolic profiling focuses on the analysis of a group of metabolites either related to a specific metabolic pathway or a class of compounds. The quantitative analysis of fatty acids as fatty acid methyl esters by GC-FID (flame ionization detection) or the analysis of amino acids are examples for metabolic profiling. An even more directed approach is target analysis that aims at the measurement of selected analytes, such as biomarkers of disease or toxicant exposure, or substrates and products of enzymatic reactions (Fiehn, 2002). In most cases metabolic profiling is a hypothesis-driven approach rather than a hypothesis-generating one. Based on the questions asked, metabolites are selected for analysis and specific analytical methods are developed for their determination. The tremendous technology advances over the past few years allow a constant expansion of the number of analytes that are quantified simultaneously in a single analysis. Technologically, the analysis of single biomarker is often as complex as profiling all related key metabolites in a given biochemical pathway. However, the latter results will give a more complete and detailed description of metabolic perturbations than a single biomarker can provide. The results of metabolic profiling are quantitative and ideally independent of the technology used for data acquisition. Consequently, the data can be used to build databases that can be integrated with pathway maps or other “omics” data, which will enhance biological understanding. Although quantitative metabolite data from different model organisms are abundant in the literature, their integration in global databases has yet to be accomplished.

The disadvantage of metabolic profiling is that the system is not a global or true “omics” approach. However, numerous quantitative metabolic profiling methods analyzing different

metabolite classes have already been developed and are routinely used. If these methods measuring key metabolites from different biochemical pathways are assembled as building blocks to study the metabolome, a powerful metabolomics approach will evolve.

The second approach towards metabolomics is metabolic fingerprinting. Initially in this approach the intention is not to identify each observed metabolite, but to compare patterns or “fingerprints” of metabolites that change in response to disease, toxin exposure, environmental or genetic alterations. A typical, but simplified workflow for a metabolic fingerprinting analysis is shown in Figure 4. Metabolic fingerprinting has been performed in a wide variety of biological matrices, such as urine, plasma or serum, saliva, and tissues or cells. In addition to metabolic fingerprinting of intracellular metabolites in cell culture systems, the analysis of extracellular metabolites excreted into the culture medium or taken up from the medium by cells can provide valuable information on their phenotype and physiological state. Pattern analysis of metabolites in conditioned cell culture media is called metabolic footprinting (Allen et al., 2003,2004). Since metabolic fingerprinting can be simultaneously applied to a wide range of metabolites, it is a true “omics” approach. The implementation of nuclear magnetic resonance (NMR)-based metabolic fingerprinting has marked the beginning of a metabolomics approach as a tool in biochemistry and has proven to be extremely powerful in screening samples for a variety of signature patterns or clusters. Metabolic fingerprinting can be used as a diagnostic tool, for example, by evaluating a patient’s metabolic fingerprint in comparison to healthy and diseased subjects. In addition, the success of treatment strategies can be monitored by observing if the metabolic phenotype shifts back to the healthy state, or in other words if a sample after treatment falls in the cluster of healthy subjects. However, using metabolomics exclusively for fingerprinting without identifying the metabolites that cause clustering of experimental groups will only deliver a classification tool but not directly contribute to biochemical knowledge and understanding of underlying mechanisms of action. The real power of metabolomics is realized when qualitative and quantitative analyses are performed. The knowledge of metabolite identity and their quantitative perturbation as descriptors of differences in specific phenotypes will provide information that can be interpreted in the light of biochemical pathways. Therefore, metabolites causing group segregation in the fingerprinting approach need to be identified and quantitative methods for the analysis of these metabolites and related compounds should be developed, which will tie metabolic fingerprinting and profiling together. Consequently, the annotation of the metabolome is an important building block for successful metabolomics investigations.

Both metabolic fingerprinting and profiling can be used in the search for new biomarkers. The value of blood glucose and cholesterol tests in medical diagnostics illustrates the value of even simple biomarkers. Metabolomics can yield new biomarkers that can reach the clinic as tools to diagnose health status, disease, or outcome of pharmacological treatment. Metabolomics is not limited to individual biomarkers. It rather represents a new approach to diagnostics where large data sets can be employed in total to develop understanding. For example, evaluating related biochemical pathways in response to drug treatment will give a more complete description of feedback mechanisms or crosstalk than single biomarkers can deliver. Moreover, the concept of individualized health including nutrition but also tailored pharmacological treatment based on metabolic phenotype will rely strongly on metabolomics technology (Watkins & German, 2002). The promise of the technology in clinical medicine to move from milliliter to microliter samples and to move from a few to thousands of analytes is exciting, but the potential of the technology to generate biological understanding is possibly still more significant.

Numerous analytical platforms have been used for metabolomic applications, such as NMR (Nicholson & Wilson, 2003), Fourier transform-infrared spectroscopy (FT-IR) (Harrigan et al., 2004;Johnson et al., 2004) and mass spectrometry (MS) coupled to separation techniques,

or using direct flow injection. The great advantages of NMR are the potential for high-throughput fingerprinting, minimal requirements for sample preparation, and the non-discriminating and non-destructive nature of the technique. However, only medium to high abundance metabolites will be detected with this approach and the identification of individual metabolites based on chemical shift signals that cause sample clustering in multivariate analysis is challenging in complex mixtures. Mass spectrometry-based metabolomics offers quantitative analyses with high selectivity and sensitivity and the potential to identify metabolites. Combination with a separation technique reduces the complexity of the mass spectra due to metabolite separation in a time dimension, provides isobar separation, and delivers additional information on the physico-chemical properties of the metabolites. However, mass spectrometry-based techniques usually require a sample preparation step, which can cause metabolite losses, and based on the sample introduction system and the ionization technique used, specific metabolite classes may be discriminated. Therefore, parallel application of several techniques, for example, GC-MS and LC-MS is desired to study the metabolome comprehensively. Currently, mass spectrometry-based metabolomics is a dynamically emerging field with a number of annual publications exceeding published NMR-based investigations (see Fig. 2).

However, not only the choice of analytical techniques requires careful consideration, but the whole metabolomics experiment should be planned as an integrated unit, because the instrumental data are only as good as the experimental design and sample treatment. In this context Bino et al. (2004) proposed, the minimum information about a metabolomics (MIAMET) experiment, which should be reported with each study in order to facilitate the exchange of information and the establishment of databases. Similar recommendations regarding meta-data have been made for the other “omics” technologies.

In general, for every type of MS-based metabolomics experiment the following steps need to be addressed during method development and validation:

- Sampling,
- Sample preparation,
- Sample analysis including metabolite separation and MS detection,
- Data export,
- Data analysis.

II. SAMPLING AND QUENCHING

Sample acquisition is primarily driven by the experimental design and the experimental type. If possible, a power analysis should be performed to ensure that a sufficient number of samples are acquired and to reduce the influence of biological variability and obtain statistically validated data. In particular, when studying human samples, the influences of diet, gender, age, and genetic factors have to be considered. Therefore, in metabolic fingerprinting analysis a large number of samples are commonly analyzed to detect biologically relevant sample clustering. In addition, representative quality control samples, such as sample replicates, analytical replicates, and blanks including method blanks have to be analyzed.

While processing biological samples, special care must be taken to minimize the formation or degradation of metabolites after sampling due to remaining enzymatic activity or oxidation processes. Several techniques have been used to inhibit metabolism such as freezing in liquid nitrogen, freeze clamping, acid treatment (Theobald et al., 1993) or quenching in salt containing aqueous methanol at low temperatures (Maharjan & Ferenci, 2003). Quenching in cold-buffered methanol kept at -40 to -50°C has been used for the analysis of *Escherichia coli*,

Saccharomyces cerevisiae, and *Aspergillus niger* (Buchholz, Takors, & Wandrey, 2001; Buchholz et al., 2002; Ruijter & Visser, 1996; Gonzalez, Francois, & Renaud, 1997). However, ionic buffer components, such as HEPES (2-[4-(2-hydroxyethyl)-1-piperazinyl]-ethanesulfonic acid) can cause ion suppression in direct infusion-electrospray ionization (ESI) MS (Castrillo et al., 2003). Wittmann et al. (2004) demonstrated that cold shock quenching in buffered methanol could cause metabolite leaching from the cell (*Corynebacterium glutamicum*), which will result in analyte losses if the cells are extracted separately from the quenching solution. Similar results were demonstrated for yeast by Villas-Boas et al. (2005). While cold shock quenching or shock freezing is commonly used when working with cells, plant tissues, or even animal tissues, these techniques are often not feasible in a clinical setting. In contrast, the time between biofluid or tissue sampling and sample extraction or sample storage at low temperatures (-20 or -80°C) is often not controlled and this can introduce variability in the data set.

III. SAMPLE PREPARATION

The quality of sample preparation is an important factor for the success of every analytical procedure. It usually serves several purposes. Mainly, it will extract the analytes from complex biological matrices, such as serum, plasma, whole blood, tissue homogenates, saliva, urine, cell pellets, or cell media and brings them into a format that is compatible with the analytical technique used while removing matrix components that will interfere with the analysis. In the case of low-abundance metabolites, the extraction procedure can include a pre-concentration step to achieve the detection limits of the applied analytical technique.

In terms of metabolic target analysis or quantitative metabolic profiling the sample preparation can be tailored for the target metabolites, because the analytes are known and surrogate compounds or stable isotope-labeled standards can be used to optimize the extraction procedure and matrix removal. However, in a true metabolomics approach, all small molecules are the targets and only salts and macromolecules, such as proteins or larger peptides, can be considered as matrix. Consequently, a sample preparation step should be as simple and universal as possible. Nevertheless, it should be noted that any kind of sample preparation will cause analyte losses. In particular, extraction procedures applied to aqueous samples will result in poor recoveries of very polar compounds. Sample preparation and sample introduction methods for the analysis of biological samples can include, but are not limited to, direct injection, liquid-liquid extraction (LLE), solid-phase extraction (SPE), supercritical fluid extraction, accelerated solvent extraction, microwave-assisted extraction, protein precipitation, and membrane methods, such as dialysis or ultracentrifugation.

A. Volatile Metabolites

Volatile organic compounds (VOCs) are an important component of the metabolome and are found in urine, plasma, or exhaled air. Volatile metabolites can include alcohols, aldehydes, furans, ketones, pyrroles, terpenes, and others (Mills & Walker, 2001). The analysis of volatile compounds is challenging by itself. Conventional sample preparation techniques, such as liquid extraction or SPE, are often not feasible, because the analytes might be incompletely extracted and losses will occur if the extract is concentrated. Moreover, the solvent can interfere with the analytes during gas chromatographic separation due to incomplete separation of the solvent peak from the analyte peaks. Specifically, in GC-MS the analytes will be cut off with the solvent delay. Therefore, solvent-free sample preparation techniques are often the method of choice for the analysis of volatile metabolites. Wahl et al. used headspace sampling with a gastight syringe in combination with cold trapping in a temperature-controlled cold injection system and GC-MS for the analysis of volatile metabolites in urine samples. Thirty-four compounds were identified using MS libraries and reference compounds (Wahl et al., 1999). Another rapid and solvent-less sample preparation technique for the determination of volatile and semi-

volatile compounds is headspace-solid phase micro-extraction (HS-SPME) (Pawliszyn, 1997). SPME uses a silica fiber that is coated with a stationary phase for sampling. The fiber is housed in a syringe-type assembly. A number of different stationary phases comprising non-polar to polar materials are commercially available as fiber coatings. In the case of HS-SPME the sampling takes place in the headspace compartment above the sample, enabling an easy fractionation of volatile analytes and complex matrix. However, sample extraction is not exhaustive, but the amount of analytes enriched on the fiber depends on two equilibrium distribution steps, sample/gaseous phase and gaseous phase/fiber coating. Both distribution coefficients are temperature dependent, and extraction time and temperature need to be optimized to maximize sample enrichment. Moreover, the distribution of the analytes between sample and gaseous phase is influenced by analyte-matrix interactions and has to be considered for quantitative analysis. Desorption of the analytes from the fiber can be achieved by thermal energy in the hot injector of a GC or using solvent for coupling with HPLC. Mills and Walker (2000) reviewed the application of SPME for the GC analysis of volatile compounds in different matrices, such as urine, blood, feces, breast milk, hair, breath, and saliva. The authors also tested different fiber coatings for the profiling of volatile urinary metabolites by HS-SPME-GC-MS and found carboxen-polydimethylsiloxane fibers to be the best choice (Mills & Walker, 2001). In addition, SPME-GC-MS has been used to study VOCs in plants (Verdonk et al., 2003; Bino et al., 2005). Stashenko, Jaramillo, and Martinez (2004a, 2004b) used different extraction methods, such as hydrodistillation, simultaneous distillation-solvent extraction, microwave-assisted hydrodistillation, supercritical fluid extraction (SFE) with CO₂, static headspace, simultaneous purge and trap in CH₂Cl₂, and HS-SPME for sampling volatile plant metabolites.

B. Solid-Phase Extraction

Solid-phase extraction (SPE) has become one of the most important sample preparation techniques to extract analytes from biological fluids. It is widely used in metabolic target and profiling analysis when a sufficient separation of analytes from interfering matrix is needed. SPE describes the non-equilibrium, exhaustive removal of analytes from a flowing liquid sample by retention on a solid sorbent. The retained analytes are subsequently eluted from the sorbent using a solvent or solvent mixtures with sufficient elution strength (Mitra, 2003). Based on the analytes and the sorbent selected, compounds are retained by van der Waals interactions, dipole-dipole interactions, hydrogen bonding, or electrostatic forces. A variety of SPE sorbent materials are commercially available for the broad extraction of metabolites, such as silica, alkylated silica (e.g., C-18), carbon-based sorbents, ion exchange materials, polymer materials, and restricted access materials (RAM) used to remove macromolecules. While C-18 modified silica phases have gained a wide distribution, they are characterized by a number of limitations, including poor retention of polar analytes, instability at high pH, and irreproducible analyte losses when the sorbent runs dry after the equilibration step. In addition to silica-based sorbents, polymer materials have gained more attention over the past years as SPE sorbents. Specifically functionalized polymeric resins, containing polar functional groups embedded in a non-polar polymeric backbone offer improved retention of polar analytes while retaining hydrophobic analytes. Polymeric SPE sorbents also show a better pH-stability and are less prone to analyte losses when running dry. Idborg-Bjorkman et al. (2003) used Oasis HLB (Waters, Inc., Milford, MA), a functionalized copolymer of divinylbenzene and *N*-vinylpyrrolidone, for the extraction of urine for LC-ESI-MS based metabolite screening (Idborg, Edlund, & Jacobsson, 2004). Wang et al. (2003) used a C-18 material to remove salts from serum after separating the serum metabolome from the proteome using a 5-kDa molecular cut-off filter.

A newer development within SPE-sorbent materials is mixed-mode sorbents, which have a high potential for metabolomics investigations. Conventional SPE sorbents rely on a primary mechanism for analyte retention. In contrast, mixed mode materials use multiple retention

mechanisms caused by the incorporation of different ligands in one sorbent. Examples are polymer materials with weak/strong cation-exchange or anion-exchange sites in the same material (Mitra, 2003). For example, using a mixed mode cation exchange material, neutrals, protonated acids, and bases can be extracted from pH-adjusted biofluids using the same solid phase. Neutrals and protonated acids are retained by non-specific hydrophobic interactions with the polymer backbone at low pH, while bases are retained by cation exchange. Consequently, elution with organic solvent will recover a fraction containing neutrals and protonated acids, and a second pH adjusted elution step will recover the basic compounds. Since a wide variety of metabolites can be extracted with these materials, they hold much promise for metabolic fingerprinting analyses. In addition to offline SPE extraction using individual cartridges, disks, or 96-well plates, online sample extraction using a pre-column coupled to LC-MS has been successfully used in our laboratory for metabolomic investigations.

C. Liquid-Liquid Extraction

Liquid-liquid extraction (LLE) for biofluids has a long history as a sample preparation technique. It is often the method of choice for tissue extraction. Upon collection, tissues are typically frozen in liquid nitrogen. In order to obtain an efficient tissue extraction, the tissue has to be homogenized first by grinding the frozen sample to a fine powder using mortar and pestle, mortar grinders (Weckwerth, Wenzel, & Fiehn, 2004), vibration mills (Jonsson et al., 2004), vortexing or mixing with beads made of an inert material (Colebatch et al., 2004), or homogenization of the tissue in a solvent using blenders (Roessner et al., 2000). Metabolite extraction from the tissue is commonly done by shaking or vortexing with organic solvents. Polar metabolites are extracted by isopropanol, ethanol, methanol, acidic methanol, acetonitrile (Aharoni et al., 2002), water, methanol:water (Roessner et al., 2000) or other mixtures of those solvents, while more lipophilic metabolites can be extracted by chloroform or ethyl acetate. The advantage of extracting samples with mixtures of water:-methanol:chloroform or ethyl acetate is the generation of a biphasic sample and the fractionation of the metabolites into polar aqueous and lipophilic organic fractions, which can be analyzed separately (Fiehn et al., 2000; Colebatch et al., 2004). Gullberg et al. (2004) used a “design of experiment approach” to optimize the extraction protocol for plant tissue for subsequent GC-MS analysis. By evaluating the response of 66 metabolites, the authors found extraction in a vibration mill with chloroform (200 μ L) followed by addition of a methanol/water mixture (800 μ L, 75:25) to form a monophasic as the optimum extraction procedure.

Weckwerth, Wenzel, and Fiehn (2004) developed a multiple solvent extraction protocol that sequentially extracts metabolites, proteins, and RNA from tissue samples of *Arabidopsis thaliana*. Different solvent systems were used for the extraction of cells, such as high-temperature extraction with ethanol or methanol (Gonzalez, Francois, & Renaud, 1997; Castrillo et al., 2003; Maharjan & Ferenci, 2003), acetonitrile/water, 0.2% formic acid (50:50) (Vaidyanathan, Kell, & Goodacre, 2002), acid or alkaline treatment (Maharjan & Ferenci, 2003), cold methanol (Maharjan & Ferenci, 2003), chloroform-methanol extraction (Ruijter & Visser, 1996; Maharjan & Ferenci, 2003), or acidic methanol-dichloromethane-ethylacetate (Smedsgaard, 1997). Villas-Boas et al. (2005) compared different extraction protocols, including chloroform:methanol:aqueous buffer, boiling ethanol, perchloric acid, potassium hydroxide, methanol:water, and pure methanol, for the extraction of polar intracellular yeast metabolites and found pure methanol the most and perchloric acid the least suitable for metabolite extraction.

D. Direct Injection

Direct injection of urine samples or diluted urine samples has been reported for LC-ESI-MS-based metabolic fingerprinting (Plumb et al., 2003b; Lenz et al., 2004a,b; Williams et al., 2005). The advantage of direct injection is that ideally no metabolites are lost during the sample

preparation. However, high-salt content in urine can cause ion suppression and adduct formation in the electrospray process and can also cause a rapid deterioration of instrument performance due to contamination by non-volatile residues. The application of nanoelectrospray ionization (nanoESI) generally minimizes ion suppression and decreases source contamination (Scholz et al., 2004). If chromatographic separation is used, the column flow can be diverted in the beginning of the run to reduce contamination of the ESI source (Lafaye et al., 2003). Direct injection of samples with high-protein content can cause poor chromatographic metabolite separation and a rapid increase of column backpressure, resulting in degradation of column performance (Pham-Tuan et al., 2003).

IV. METABOLIC PROFILING—AN ARRAY APPROACH TOWARDS METABOLOMICS

Metabolic profiling or the quantitative analysis of a selected number of metabolites involved in the same biochemical pathway is a widespread tool to study different aspects of metabolism. A thorough coverage of all published methods would greatly exceed the purpose of this review. However, selected contributions in the more recent literature describing the analysis of key metabolite classes will be highlighted below. While each method on its own does not represent a true metabolomic analysis, assembling a whole suite of quantitative methods that analyze key metabolites from different biochemical pathways will turn metabolic profiling into metabolomics. The advantage of this integrative array approach is that quantitative data are generated that will lead to an annotation of the metabolome and the subsequent generation of independent databases containing quantitative information that can be further mined. In addition, using individual assays to probe the metabolome allows the assessment of important low-abundance metabolites with regulatory functions, such as eicosanoids, which would not be detected by a general metabolic fingerprinting approach at the current level of technology development.

A. Mass Spectrometric Profiling of Metabolic Disorders

Mass spectrometric metabolic profiling plays an important role in the diagnosis of inborn errors of metabolism. The analysis of the urinary metabolome for diagnostics of metabolic disorders using GC-MS has recently been reviewed by Kuhara (2005). In addition, electrospray ionization with tandem mass spectrometry (MS/MS) has become an important tool in newborn screening for the assessment of inborn disorders of amino acid, fatty acid, and organic acid biosynthesis. Several reviews have been published recently, describing the methodology and biochemical interpretation of data in detail (Chace, Kalas, & Naylor, 2002,2003;Schulze et al., 2003;Chace & Kalas, 2005). Standard protocols incorporate methanol extraction of disks of dried bloodspots on filter paper, addition of stable isotope-labeled internal standards, solvent evaporation, and derivatization with acidified butanol. Analysis of butyl esters is performed by flow injection analysis (FIA) with positive-mode ESI and MS/MS detection using a neutral loss scan of 102 amu (butyl formate) to selectively detect neutral amino acids, a neutral loss of 119 amu for ornithine and citrulline, and a loss of m/z 161 for arginine. Acylcarnitines are measured by a precursor ion scan of m/z 85 (Zytkovicz et al., 2001).

Piraud et al. (2003) developed a method based on LC-ESI-MS/MS for the analysis of 79 small molecules, mainly amino acids, related to known metabolic disorders. The goal of the study was to analyze underivatized amino acids and related analytes in a single run. Ionization behavior of the metabolites in positive and negative mode and the fragmentation in collision-induced dissociation (CID) were studied to select characteristic and signal-intense transitions for analyte detection. In order to avoid interferences between isobars and in-source fragmentation products, an LC separation was necessary.

B. Lipidomics

The analysis of lipids using mass spectrometry has become a valuable and mature tool for the assessment of changes in lipid metabolism and lipid-mediated signaling processes as a result of disease, toxicant exposure, genetic modifications, or drug therapy (Watkins et al., 2002). The lipidome, comprising lipid classes, subclasses, and lipid signaling molecules, is an important sub-compartment of the metabolome and has been well studied (German, Roberts, & Watkins, 2003). Lipidomics delivers detailed and quantitative information about the constitution of the cellular lipidome and provides insights in biochemical mechanisms of lipid metabolism, lipid–lipid and lipid–protein interactions.

Non-polar lipids have been routinely analyzed using gas chromatographic separation, electron impact ionization (EI), and MS. A derivatization step, such as methylation or silylation, is often needed to further expand the range of lipids accessible with this technique. However, GC-MS analysis is limited to thermally stable compounds with a sufficient vapor pressure for volatilization during injection. Using high-temperature gas chromatography, triglycerides have been analyzed, but the analysis of complex lipids, such as phospholipids, requires a pre-separation of lipid classes followed by hydrolysis, derivatization of free fatty acids and subsequent GC-EI-MS or GC-FID analysis. Although highly quantitative, this approach is very labor intensive and structural information regarding lipid composition is lost.

In addition to conventional hot split/splitless injection for sample introduction, analytical pyrolysis (Py) has been used to analyze the lipid composition in various sample matrices. Analytical pyrolysis is the thermal degradation of a non-volatile sample in an inert gas atmosphere before introduction into the GC. Pyrolysis, in combination with GC-MS, is a very powerful tool for the characterization of complex biological materials. Snyder et al. (1990) used Py-GC-Ion Trap-MS of whole, underivatized microorganism samples for the differentiation of bacterial species based on their lipid profiles. Basile et al. (1998) used *in situ* thermally assisted hydrolysis and methylation (THM) in a pyrolysis system for the differentiation of whole bacterial cells based on the profiles of fatty acid methyl esters formed during THM. Pyrolysis was coupled to GC-MS or used as a direct inlet system for a triple quadrupole mass spectrometer. *In situ* THM is performed using quaternary *N*-alkylammonium salts, such as tetramethylammonium hydroxide (TMAH), which forms tetramethylammonium carboxylates with organic acids and methylation takes place upon heating. Ester-bound fatty acids are saponified and trans-methylated. However, unsaturated fatty acids can undergo isomerization and recoveries are incomplete (Drechsel, Dettmer, & Engewald, 2003).

The introduction of solid-state ion sources, such as fast atom bombardment (FAB) and liquid-phase ion sources, such as ESI or atmospheric pressure chemical ionization (APCI) allowed the direct lipid analysis (Griffiths, 2003). Hermansson et al. (2005) developed a fully automated method for the analysis of the lipidome using LC-ESI-MS that delivers quantitative data for more than 100 polar lipids, originating primarily from phospholipids. The so-called shotgun lipidomics approach uses crude tissue extracts, commonly obtained by Folch or Bligh-Dyer extraction methods. It is based on direct sample injection of the extract, using both positive and negative mode ESI (Han & Gross, 2003,2005). ESI in positive mode allows the detection of acylcarnitines, phosphatidylcholine, lysophosphatidylcholine, phosphatidyl ethanolamine, and sphingomyelins. Free fatty acids, phosphatidic acid, phosphatidyl serine, phosphatidyl inositol, phosphatidyl glycerol, and phosphatidylethanolamine are ionized by negative mode ESI. Neutral lipids, such as triacylglycerols are not readily ionized by ESI, however they can be detected as ammonium, lithium, or sodium adducts with ESI in positive mode (Duffin, Henion, & Shieh, 1991; Hsu & Turk, 1999; Han & Gross, 2005). MS/MS experiments are used to further identify lipid species or to screen for individual lipid classes. For example, phospholipids with specific head groups show characteristic fragmentation patterns in CID spectra that can be used for their identification (Griffiths et al., 2001; Pulfer & Murphy,

2003). A similar approach using characteristic CID fragmentation has been developed for sphingolipid analysis or sphingolipidomics (Merrill et al., 2005). Specifically, neutral-loss scans and precursor-ion scans can be used to identify lipids in samples and to determine the molecular masses of the compounds.

Using ESI in positive mode, precursor-ion scans for m/z 184 ($[\text{C}_5\text{H}_{15}\text{NPO}_4]^+$) or neutral loss scans for 59 amu ($\text{N}(\text{CH}_3)_3$) can be used to identify the $[\text{M} + \text{Na}]^+$ and $[\text{M} + \text{H}]^+$ ions that correspond to phosphatidylcholine and lysophosphatidylcholine (see Fig. 5). Precursor-ion scans for m/z 164, ($[\text{NaHO}_3\text{POC}_2\text{H}_4\text{NH}_3]^+$) can be used to identify the $[\text{M} + \text{Na}]^+$ and $[\text{M} + \text{H}]^+$ ions of phosphatidylethanolamine. Using ESI in negative mode, neutral loss scans for 153 amu ($[\text{HO}_3\text{POCH}_2\text{-C}(\text{OH})\text{CH}_2]$) can be used to identify precursors of phosphatidylglycerol (Pelizzi et al., 2002). In sphingolipidomics the same m/z 184 characteristic ion is used to identify sphingomyelin while m/z 264 and m/z 266 ions can be used for ceramide identification in precursor ion scans (Merrill et al., 2005).

Further MS/MS experiments can be performed for the compositional analysis of lipids. The regio-chemistry of phospholipids is determined based on the intensity ratio of the carboxylate anions originating from *sn*-1 and *sn*-2 position, because the intensity of the fragment ion derived from the *sn*-2 acyl substituent is greater than the product ion corresponding to the *sn*-1 acyl substituent (Fang & Barcelona, 1998). Branched saccharide moieties of higher order sphingolipids (such as gangliosides) requires MS/MS/MS experiments for structure elucidation (Merrill et al., 2005).

Because certain class of lipids may comprise 100 different species, their routine quantification represents a challenge. However, some lipids, such as phospholipids, differentiate only by their head groups and acyl groups (specifically their lengths and number of double bonds). Therefore, computational lipidomics approaches can be developed to automate spectra interpretation and quantification of multiple lipid species to use this similarity as an advantage. For example, automated detection of over 800 phospholipid species in direct injection-ESI-MS/MS analysis of cell extracts has been reported with semi-quantitative data output (lipid array) similar to DNA microarrays (Forrester et al., 2004).

Metabolic profiling of eicosanoids, a class of lipid signaling molecules, was recently termed eicosanomics (Balazy, 2004). Eicosanoids are metabolites of arachidonic acid produced by cyclooxygenases, lipoxygenases, cytochrome P450's and non-enzymatic oxidation and are one of the most important pharmacological targets. Because many eicosanoids are structural isomers, LC or GC separation is required prior to MS detection. The profiling of eicosanoids is performed by LC-ESI-MS/MS in negative mode (Margalit, Duffin, & Isakson, 1996; Newman, Watanabe, & Hammock, 2002; Schmelzer et al., 2005) or LC-Electron Capture-APCI-MS/MS which requires derivatization of carboxylic groups into pentafluorobenzyl esters (Lee et al., 2003).

V. METABOLIC FINGERPRINTING

Metabolic fingerprinting describes the unbiased analysis of the metabolome by examination of metabolite patterns in different experimental groups and subsequent classification of the patterns. Samples can be classified if the metabolite fingerprints change between groups, resulting in sample clustering. In mass spectrometry-based investigation, metabolite fingerprints are described by m/z values and corresponding intensities of detected ions. If a separation step is performed, retention times are also used to index metabolites. Thus, m/z values, retention times, and intensities represent the metabolic fingerprint of the analyzed sample and are exported for sample classification using multivariate data analysis. The chemical structure of the detected metabolites typically remains unknown. In this review, the term metabolic fingerprinting refers not to investigations that aim at the analysis of a pre-

defined set of metabolites but rather to the unbiased analysis of the entire detectable metabolome, even if semi-quantitative data are generated, which could justify using the term metabolomics instead of metabolic fingerprinting.

A. Direct Injection Mass Spectrometry Analysis

Direct injection analysis describes the injection or infusion of a sample into the ionization source of the mass spectrometer without prior chromatographic separation. This technique is commonly used with atmospheric pressure ionization techniques, predominantly ESI. Direct injection is a high-throughput approach, allowing a sample to be processed within a few minutes. The short analysis time increases inter-sample reproducibility and improves the accuracy of subsequent cluster analysis. Direct injection analyses have been performed using ESI and nominal mass resolution mass analyzers. Smedsgaard et al. used ESI in positive mode and a single-stage quadrupole instrument for metabolic fingerprinting in crude fungal extracts (Smedsgaard & Frisvad, 1996,1997). Castrillo et al. (2003) analyzed yeast intracellular metabolites by direct injection ESI in positive mode and triple quadrupole MS.

However, in order to distinguish between isobars (compounds with the same nominal mass) high-resolution mass spectrometers are the instrument of choice (see Fig. 6). High-mass resolution also allows accurate mass measurement of pseudo-molecular ion signals and the calculation of empirical formulas. Direct injection high-resolution mass spectrometry has been performed using time of flight mass spectrometers (TOF-MS) and Fourier transform ion cyclotron mass spectrometers (FT-ICR-MS). Depending on the flight-tube geometry and instrument tuning, TOF-MS instruments provide mass resolution of 6,000–17,000 with mass accuracy in the range of 3–5 mamu. Allen et al. (2003,2004) used TOF-MS with ESI in positive mode for metabolic footprinting analysis in yeast cell media samples.

Hybrid instruments, such as quadrupole-TOF-MS (Q-TOF-MS), have also been used for the rapid screening analysis of cell or plant extracts using the TOF as mass analyzer, which provides accurate mass measurement with high-mass resolution (Vaidyanathan, Kell, & Goodacre, 2002;Scholz et al., 2004). Structural elucidation of biomarkers can be performed by MS/MS experiments without switching to another instrument. Using the quadrupole as a scanning device, precursor ions can be selected for fragmentation in the collision cell, and product ion spectra with high-mass resolution for the fragment ions are produced. Accurate mass measurement of precursor and fragment ions allows clarification of the fragmentation process, thereby significantly facilitating the interpretation of spectra.

FT-ICR-MS offers very high resolution (100,000–1,000,000) and highest currently available mass accuracy (0.1–1 mamu), very low detection limits in the attomole to femtomole range and MSⁿ capabilities (Marshall, Hendrickson, & Jackson, 1998), which makes it an ideal tool for metabolic fingerprinting investigations (Brown, Kruppa, & Dasseux, 2005). However, high-instrument costs have prevented their application from gaining wider distribution in metabolomic research. Aharoni et al. (2002) used FT-ICR-MS to study metabolic changes in ripening strawberry fruit by direct infusion of crude fruit extracts (methanol:water, 0.1% formic acid (50:50), or 100% acetonitrile) and ESI or APCI in positive and negative mode. In the study, a total of 5,844 unique ¹²C single charged masses were obtained by screening both extracts using four different ionization modes, and a single empirical formula could be assigned to over half of them with an error of ±1 ppm. Hirai et al. (2004) used FT-ICR-MS for metabolic fingerprinting of *Arabidopsis thaliana* extracts in an integrated transcriptomics and metabolomics study to understand the global plant response to nutritional stress.

Direct injection experiments for ESI are performed by continuous flow injection or by loop injection of 10–20 µL sample, using a solvent flow rate of 5–50 µL/min. The application of nanoESI has also been described. High-throughput fingerprinting can be achieved using

automated nanoESI systems, such as the Nanomate system (Advion BioSciences, Inc., Ithaca, NY), which comprises a liquid sampler and an ESI chip consisting of an array of nanoelectrospray nozzles (Scholz et al., 2004).

Although direct injection-MS is a high-throughput metabolic fingerprinting technique, it has several disadvantages. For example, chemical isomers cannot be distinguished by this rapid screening technique because they have the same exact mass and therefore would require chromatographic separation (see Fig. 6). In addition, the classification of in-source fragmentation products or adduct formation is complicated. However, in the case of ESI the most important concern is ion suppression. Since all sample components are introduced simultaneously into the ionization source, signal suppression or enhancement can occur. This can impede a meaningful data analysis, especially if the matrix content varies between sample groups, for example, urine samples. An example for ion suppression is given in Figure 7, which shows the direct injection of a rat urine sample spiked with internal standards after online SPE compared to the analysis of internal standards spiked into a saline solution. Although salts are removed by the online-SPE extraction, severe signal suppression is observed for the internal standard signals in the urine sample relative to the saline solution.

B. LC-MS

In order to overcome the drawbacks of directly injecting complex samples, liquid chromatography can be interfaced with the MS detector. Liquid chromatography can reduce ion suppression caused by coeluting compounds, isobaric interferences in the case of low-resolving mass analyzers, and often can separate isomers. In addition, a good analytical separation will result in better detection limits and MS data quality due to reduced background noise. Reversed phase liquid chromatography using C-18 narrow bore columns with particle sizes of 3–5 μm is widely used for metabolomic investigations, but the application of reversed phase capillary column separation is expected to increase in the future. In addition, the application of a monolithic-silica based C-18 capillary column for plant metabolomics has been reported (Tolstikov et al., 2003). However, conventional RP-HPLC separation is often insufficient for the separation of complex biological samples, resulting in poor metabolite resolution (see Fig. 8). One approach to increase chromatographic resolution and peak capacity is to use small particle sizes, such as the 1.7- μm particles used in ultra performance liquid chromatography (UPLC). Sub-2 μm particles provide narrow chromatographic peaks, which results not only in a better resolution but also in lower detection limits. Analysis of mouse urine using a 2.1 \times 100 mm, 1.7 μm particle size C-18 column, and gradient elution produced chromatographic peaks with a peak width in the order of 1.8 sec at the base, resulting in a peak capacity of ca. 250 for a 10 min separation (Wilson et al., 2005a,b). The downside of this approach is the high pressure (10,000–15,000 psi) needed to operate these columns, and thus special UPLC systems are required.

Reversed phase chromatography is a standard tool for the separation of medium polar and non-polar analytes. However, very polar metabolites are not retained on classical reversed phase stationary phases and elute with the void volume. An interesting alternative has been presented by Tolstikov and Fiehn (2002), who used hydrophilic interaction chromatography (HILIC) for the analysis of highly polar compounds in plant extracts. HILIC is orthogonal to reversed phase chromatography and uses either silica or derivatized silica, including amino, diol, amide, polysulfoethyl aspartamide, or polyhydroxyethyl aspartamide groups, with low-aqueous/high-organic solvent systems. Although similar to normal phase chromatography, HILIC requires a substantial amount of water in the mobile phase, which forms a stagnant water layer on the stationary phase surface into which polar analytes can be partitioned and retained. Contrary to normal phase chromatography, HILIC uses water-miscible polar organic solvents, such as acetonitrile or methanol. The application of HILIC for bioanalytical LC-MS/MS has been

reviewed by Naidong, who also discusses the retention mechanisms involved in HILIC separations (Naidong, 2003). LC-MS based metabolic fingerprinting investigations have been performed using different mass analyzers including ion trap instruments (Tolstikov & Fiehn, 2002; Lafaye et al., 2003; Tolstikov et al., 2003; Wang et al., 2003), triple quadrupole instruments (Idborg-Bjorkman et al., 2003), TOF (Wang et al., 2003), and Q-TOF instruments (Plumb et al., 2002, 2003a,b; Lenz et al., 2004a,b; Williams et al., 2005). Triple quadrupole or Q-TOF analyzers provide the ability to perform MS/MS experiments for the structural elucidation of biomarkers. In the case of Q-TOF hybrid instruments, accurate mass measurement can be performed for precursor and product ions. Ion trap instruments have the advantage of MSⁿ capabilities. MS/MS experiments can also be used to probe the metabolome for specific compound classes by screening for characteristic ions and neutral losses, for example, the neutral loss of 44 amu (CO₂) observed in negative mode CID spectra is characteristic for carboxylic acids (Lafaye et al., 2003). In addition, phase II metabolites often show characteristic fragmentation patterns that can be used for rapid screening. Levsen et al. (2005) gives an excellent overview on the CID behavior of major Phase II metabolite classes, including glutathione conjugates, glucuronides, glucosides, malonylglucosides, sulfates, acetates, methyl and glycine conjugates.

The ionization technique of choice for LC-MS based metabolic fingerprinting is ESI. In order to obtain a broad coverage of the metabolome, ionization must be performed in positive and negative mode. This is illustrated in Figure 8, which shows the total ion chromatograms obtained in positive and negative ESI of a human urine sample after online-SPE and HPLC separation. The figure emphasizes that ionization in positive and negative mode leads to the detection of two sets of analytes that can differ significantly. Depending on the mass analyzer used, detection in positive and negative mode can be performed simultaneously in a single run (Idborg-Bjorkman et al., 2003) thus, reducing the time needed for analysis and eliminating possible variability due to injection errors.

The application of LC-MS based metabolic fingerprinting includes the analysis of urine samples in toxicity studies (Idborg-Bjorkman et al., 2003; Lafaye et al., 2003; Lenz et al., 2004a,b; Williams et al., 2005) or genetic studies (Plumb et al., 2003a), plasma samples (Wang et al., 2003), plant extracts (Tolstikov & Fiehn, 2002; Tolstikov et al., 2003) and the analysis of fungal metabolites (Nielsen & Smedsgaard, 2003). Saghatelian et al. (2004) identified novel substrates of fatty acid amide hydrolase (FAAH) by performing metabolic fingerprinting in brain and spinal cord extracts using RP LC-MS.

C. GC-MS

The combination of gas chromatography with electron impact ionization MS (EI-MS) provides high-chromatographic metabolite resolution, analyte-specific detection, and quantification of metabolites, as well as the capability to identify unknowns. However, a major prerequisite for GC-MS analysis is a sufficient vapor pressure and thermal stability of the analytes. The analysis of polar metabolites usually requires derivatization at the functional group to reduce polarity and increase thermal stability and volatility. A summary of common derivatization reagents is given in Table 2. Active hydrogens in functional groups, such as –COOH, –OH, –NH, and –SH can be derivatized by alkylation, acylation or silylation. In silylation, the active hydrogen is replaced by an alkylsilyl group, primarily trimethylsilyl (TMS), with formation of TMS ethers, TMS esters, TMS sulfides or TMS amines. Silyl derivatives show a better thermal stability and higher volatility and can produce more distinct MS spectra than their underivatized precursors. A number of different reagents for trimethylsilylation are available, including BSTFA (*N,O*-bis-(trimethylsilyl)-trifluoroacetamide) and MSTFA (*N*-methyl-trimethylsilyltrifluoroacetamide), which are being predominantly used in metabolomic investigations. BSTFA is highly volatile and produces volatile by-products, for example, the

leaving group is trifluoroacetamide, which causes little interference with early eluting peaks. Comparable in silyl donor strength, MSTFA is the most volatile of the TMS acetamides and is therefore very suitable for GC-MS analysis. Often 1% trimethylchlorosilane (TMCS) is added to the reagents as a catalyst. A disadvantage of silylation is the sensitivity of the derivatives to moisture. Sample extracts must be thoroughly dried before derivatization and care has to be taken when handling the derivatized samples to avoid contact with moisture, resulting in derivative degradation. It has also to be noted that silylation can cause conversion reactions, for example, it has been described that arginine is converted into ornithine by reaction with BSTFA or MSTFA (Halket et al., 2005). Moreover, artifacts can be formed, complicating data interpretation (Little, 1999). In addition to trimethylsilylation, *tert*-butyldimethylsilylation (*tert*-BDMS) is being used for derivatization in metabolic fingerprinting. The EI-MS spectra of *tert*-BDMS derivatives often show characteristic fragmentation with a fragment ion of $[M-57]^+$ caused by σ -bond cleavage at the Si atom and loss of the *tert*-butyl moiety, which can be helpful for the identification of unknown metabolites (Fiehn et al., 2000). Moreover, *tert*-BDMS derivatives have been reported to be stable with 2% residual water in the samples (Birkemeyer, Kolasa, & Kopka, 2003).

In addition to the derivatization of exchangeable protons, carbonyl groups are often transformed into the corresponding oximes with hydroxylamine or alkoxyamines, which primarily stabilizes α -ketoacids and locks sugars in open-ring conformation. Oxime derivatives can be formed as *syn* and *anti* isomers depending on the orientation at the carbon-nitrogen double bond. The isomers can be chromatographically resolved, leading to two GC peaks for each compound. Fiehn et al. (2000) found alkoxyamines superior to hydroxylamines for identification purposes. In order to obtain broad coverage of the metabolome by GC-MS, a two-step derivatization procedure including methoximation followed by silylation is often used for metabolic fingerprinting by GC-MS (Roessner et al., 2000; Colebatch et al., 2004; Jonsson et al., 2004). Gullberg et al. used a "design of experiment approach" to optimize factors influencing the extraction of plant tissue and subsequent derivatization for GC-TOF-MS analysis. The authors used methoximation and studied the influence of oximation temperature and time followed by silylation with MSTFA with 1% trimethylchlorosilane (TMCS) and variation of silylation temperature, amount of reagent, and addition of cosolvent (hexane, acetonitrile) (Gullberg et al., 2004). Using 12 reference compounds from different metabolite classes to study derivatization efficiency, the authors found methoximation with methoxyamine in pyridine for 17 hr at room temperature, followed by silylation with MSTFA for 1 hr at room temperature to be the optimum derivatization procedure.

Sample preparation for the GC-MS analysis of biofluids often includes only a lyophilization step. However, in this case, highly concentrated matrix components can interfere with the analysis. For example, urea in urine samples can cause a significant column and detector overload. Fu et al. (2000) describe a simple urease treatment to remove excessive urea for urine samples prior to silylation and GC-MS analysis.

Metabolomic investigations using GC-EI-MS rely primarily on quadrupole instruments (Roessner et al., 2001; Roessner, Willmitzer, & Fernie, 2001) or TOF mass analyzers. Here, TOF-MS is mainly used for fast analyte detection, rather than for high-mass resolution. For example, the commonly used Pegasus TOF-MS instruments (LECO Instruments) provide high-data acquisition rates with nominal mass resolution. Nevertheless, high-mass resolution TOF-MS will be important for the identification of unknowns. Interestingly, GC-MS with ion trap instruments has not been adapted for general metabolomic investigations.

The combination of GC with TOF-MS provides rapid metabolite detection and has the potential for significant reduction of analysis time by implementing fast GC methods, thereby increasing laboratory throughput. Fast GC methods will reduce conventional GC run times from 40–60

min to less than 10 min by using short GC columns with small inner diameters. Fast GC results in very narrow peaks (peak width 0.5–1 sec) which cannot be sampled by conventional scanning quadrupole mass analyzers.

Although standard capillary GC columns provide a high-peak capacity, in the case of very complex biological samples the separation power of one column is often not enough. As an example, Figure 9 shows the GC chromatogram of a human urine (1 mL) extract after silylation with MSTFA. Deconvolution of the chromatogram using AMDIS resulted in 1,582 components, many of them overlapping.

One way to increase metabolite separation is the use of comprehensive two-dimensional chromatography (GC × GC). In GC × GC, each peak from a first usually non-polar column is transferred to a second (polar) column for further separation. In this way, three-dimensional (3D) chromatograms are obtained, with the retention times on the two columns forming two dimensions and peak intensity representing the third dimension. With MS detection a fourth dimension, the mass spectra, is introduced. The two columns are connected by a so-called modulator. The analytes are transferred from the first to the second column by dividing the eluent from the first column into small segments and focusing them in the modulator, for example, by cooling. In these cryogenic modulators, each individual fraction is cryo-trapped using CO₂ or N₂ cooled jets, focused and then, after rapid heating transferred into the second column as a narrow band. The focusing step reduces the peak width of the elution peaks significantly and therefore decreases the detection limit. Even without a second column dimension, an increased peak capacity can be obtained by the focusing step alone. The second short column separates the individual fractions within a few seconds. The whole process must be carried out very quickly because trapping, release and separation of a fraction must be finished before the next segment reaches the modulator. The resulting peaks are very narrow with baseline peak widths ranging from 80 to 600 msec (Adahchour et al., 2005). The narrow peaks require a fast scanning rate, and for identification purposes a TOF mass spectrometer is commonly used. The introduction of rapid-scanning quadrupoles that have scan speeds of up to 10,000 amu/sec provides a less expensive alternative for detection (Adahchour et al., 2005).

D. CE-MS

Capillary electrophoresis (CE) is a powerful technique for the separation of charged metabolites, offering high-analyte resolution. The combination with mass spectrometry makes CE-MS an ideal tool for the analysis of the metabolome. However, only a few applications have been published to date. Soga et al. (2003) developed a methodology using three CE separations for the analysis of anionic metabolites, cationic metabolites, and nucleotides and CoA compounds, to achieve a comprehensive coverage of the metabolome. Since the authors used a single quadrupole mass analyzer, they choose a scan range of 30 amu to maximize signal intensity for the detection of low-abundance metabolites. In order to cover a range of m/z 70– m/z 1,027, 33 runs were necessary, resulting in a total analysis time of 16 hr for one sample. Using this technique, 1,692 metabolites in a bacterial extract were analyzed. Sato et al. used a similar approach to analyze major metabolites in rice leaves using CE-MS and CE-diode array detection (DAD) (Sato et al., 2004).

Another promising separation technique that has high potential for metabolomics investigations is capillary electro-chromatography (CEC). CEC can be described as a hybrid of liquid chromatography and capillary electrophoresis. It commonly uses capillary columns that are packed with LC-stationary phases or have monolithic stationary phases. The mobile phase flow is caused by the electroosmotic flow, resulting in the plug-like flow profile typical of CE, which in comparison to the parabolic flow profile of LC, leads to narrow peaks, and high resolution. Pressure-assisted CEC has also been performed. The hyphenation of CEC with

MS and the interfaces used has recently been reviewed by Klampfl (2004). The author also discusses different bio-analytical applications of CEC-MS, such as the analysis of proteins, peptides, amino acids, and saccharides.

VI. FLUX ANALYSIS

Metabolism is a dynamic phenomenon. In living organisms there is a continuous flux of matter mediated by fast enzymatic transformation of metabolites. Thus, measurements of intermediates of metabolic pathways represent measurements of steady-state metabolite concentrations or snapshots of metabolite concentrations at the given point in time the samples were taken. In some instances, the steady-state concentration of metabolites may remain constant while flux changes dramatically. Therefore, steady-state concentrations may not reflect the underlying biological processes and do not describe the phenotype completely (Fig. 1). Flux analysis provides a true dynamic picture of the phenotype because it captures the metabolome in its functional interactions with the environment and the genome. Thus, metabolism-wide measurements of fluxes are a basis of phenomics, “phenotypic analysis of genomic information or entire mutant collections with the goal of understanding the relationship between genes and higher levels of organization in the cell” (Sauer, 2004). In addition, flux analysis is important for studying the kinetics and dynamics of xenobiotic biotransformation (e.g., pharmaceuticals or environmental pollutants).

The analysis of metabolite fluxes is performed using isotope-labeled precursors of metabolic pathways. Most of the known biochemical pathways were initially deciphered using radioactive isotope labels. However, detection limits of decay counters are not low enough to perform flux measurements *in vivo* using safe radioisotope doses. Accelerator mass spectrometry (AMS) can precisely quantify atoms in the attomole range and can, therefore, be used to monitor how trace amounts of ^{14}C -labeled compounds and other isotopes below the levels of legally regulated radioactivity move through metabolic pathways (White & Brown, 2004). AMS was applied to study *in vivo* the metabolic flux of human folate (Lin et al., 2004) and beta-carotene (Dueker et al., 2000; Lemke et al., 2003). The metabolic fate of xenobiotics, such as the herbicide atrazine, was studied by HPLC-AMS in humans as well (Buchholz et al., 1999). However, AMS cannot perform structural elucidation of unknowns to confirm identity of metabolites. Therefore, additional techniques are required for identity confirmation (e.g., chromatographic coelution of the labeled compounds with a reference standard). Thus, modern AMS can hardly be used for global flux analysis on the level of the metabolome but it can be an excellent tool to study flux in a specific pathway. The high cost of AMS instruments precludes their widespread application, although the cost is dropping dramatically.

Metabolites labeled with ^{13}C or other stable isotopes can be readily detected with standard NMR and MS instruments. Although their detection limits are significantly higher than the detection limits of AMS, standard instrumentation is commonly available, can confirm metabolite identity and provides structural information about labeled atom positions (positional isotopomers) and the number of labeled atoms in the molecule (mass isotopomers). Lower detection limits and high throughput make MS the preferential technique over NMR for flux analysis (Wittmann, 2002; Sauer, 2004). Because MS detects both natural and artificially added isotopes (e.g., ^{12}C and ^{13}C) and a large number of isotopomers can be derived from an initial precursor label, special data analysis methods are required to interpret the data.

An overview of mass spectrometry application in flux analysis was given recently by Wittmann (2002). Flux analysis is usually performed by GC-MS because EI fragmentation spectra provide information about isotopomer distribution. This technique was used to study metabolic fluxes, for example, in microorganisms *Phaffia rhodozyma* (Cannizzaro et al., 2004),

Escherichia coli (Fischer & Sauer, 2003), in potato (Roessner-Tunali et al., 2004) and in humans (Hellerstein et al., 1997). To analyze non-volatile but ionic metabolites or to avoid derivatization steps, LC-ESI-MS/MS can be used as an alternative technique (Mashego et al., 2004; van Winden et al., 2005). In addition other techniques such as MALDI-TOF (Wittmann & Heinzle, 2001) were applied for flux analysis. Overall, MS-based integrated studies of all metabolite fluxes or MS-based fluxomics is just beginning to emerge (Hellerstein, 2004; Sauer, 2004).

VII. DATA EXPORT AND ANALYSIS

The ultimate goal of data export is to convert metabolic fingerprinting data into a standard and uniform format. This includes data pretreatment to reduce noise and background, alignment of mass spectra or mass chromatograms and automated picking and annotation of mass spectrometric or total-ion/extracted-ion chromatographic peaks of metabolites. Currently, several public and proprietary software tools exist to implement data export and subsequent data analysis. Since metabolomics is essentially the analysis of complex mixtures, any software that facilitates such analysis is referred to as metabolomics software in this review.

A. MS Data Format

With a variety of commercial and custom-built mass spectrometers a number of data file formats exist. This impedes the development of universal metabolomics software that does not depend on an analytical platform, especially the development of General Public License (GNU) software that has undergone tremendous growth in the analysis of genomics and proteomics data. Several solutions to this problem have been implemented. First, instrument manufacturers simply create their own software for the export of metabolomics data (e.g., LECO, Micromass). Second, independent software developers create programs, which are able to access a variety of commercial MS file formats (e.g., Advanced Chemistry Development, Inc.). Finally, open-access file formats such as ASCII text or binary netCDF exist, and instrument manufacturers often provide tools to convert original data files into these formats. Thus, platform-independent software can be created based on such standards.

While ASCII text is a good entry point file format for keeping low resolution or low volume MS data, the size of ASCII data file increases tremendously for high resolution or chromatography-MS data. In addition, there is no widely accepted text markup standard for MS data. Thus, a standard extensible markup language (XML) format for MS metabolomics data is yet to be developed. Currently, a universal mzXML format (Pedrioli et al., 2004) is becoming more widely accepted, however it is not design to contain biological descriptive information. XML format for metabolomics will not only provide standardization of quantitative data output but also a thorough sample annotation, which is required for the biological interpretation of the data.

NetCDF (network common data form) is a binary data format widely used for saving experimental data, including MS data (<http://my.unidata.ucar.edu/content/software/netcdf/index.html>). Major mass spectrometer manufacturers usually provide tools for the conversion of native format data files into a netCDF format (e.g., AIA format in Agilent Chemstation and ANDI in Finnigan Excalibur). Freely distributed libraries to access netCDF files are available for common programming languages such as C, C++, FORTRAN, Matlab, R, Java, and Python. Thus, the development of platform-independent metabolomics software such as COMSPARI (Katz et al., 2004), XCMS (Siuzdak, Scripps Research Institute, San Diego, CA; personal communication) or MZmine (Katajamaa, Miettinen, & Oresic, 2006) should be facilitated. While often it is not possible to convert data from one proprietary data format into another one directly, it may be possible through intermediate netCDF files. Consequently, virtual platform independence can be achieved for data analysis with platform-dependent software.

B. Data Extraction

Raw data are usually stored in sample files as series of mass spectra acquired at a given time point or scans. Each scan is represented by a pair of mass and intensity vectors (see Fig. 10). The vector length varies from scan to scan depending on the number of detected peaks. However, for a metabolomics data analysis it is necessary to extract information about all compounds, including mass and retention time in the case of chromatography or CE, as compound identifiers and intensity as quantitative representation of concentration. Moreover, the data should be extracted from all data files as uniform matrices to allow data comparison and discovery of differentially expressed metabolites. Transformation of continuous mass spectra into centroid spectra can be considered as a first step to reduce spectra complexity before peak extraction.

Software that can perform data extraction from mass spectra and total ion chromatograms (especially extraction of high-resolution data) is still limited. In this review, we only discuss software products that are published in the literature or that we had access to. Proprietary software that is not published and is not openly accessible was not considered for this review.

The straightforward approach for data extraction is to transform mass spectral data into vectors of uniform length (see Fig. 10A). The most trivial way to bring the mass-intensity vectors to uniform length is to bin them in a fashion similar to the approach widely used in NMR metabolomics (Holmes, Nicholson, & Tranter, 2001; Viant, 2003). Binning is highly compatible with nominal mass resolution spectra because no information is lost and the number of bins is usually below 2000 at a bin size of 1 amu. In this case, the binning procedure is simply a rounding of masses to integer values, inserting missing mass values, and assigning them zero intensities. However, for high-resolution mass spectra binning becomes disadvantageous because small drifts in high-resolution mass measurement can occur. Thus, if the accurate mass of a metabolite is close to the margin between two bins, the same centroided peak from different data files may be assigned to different bins by the binning algorithm. Consequently, data become noisy and even false positive results are possible if samples are not randomized and instrument drift occurs over time. The problem of drift can be minimized by using reference standard for accurate mass measurement, which is continuously introduced into the mass spectrometer and allows subsequent alignment of mass spectra. To circumvent the drift problem, the bin size can be increased up to nominal mass resolution with the inevitable sacrifice of mass spectrometric resolution (Jonsson et al., 2005). If the bin size is kept low to represent instrument resolution, then very large data sets are created with only very few bins containing information (sparse matrices). In addition, there is a high probability of signal splitting between adjacent bins, which complicates subsequent data analysis. Another possibility to minimize the mass scale drift problem is to bin original continuous spectra, however, this makes subsequent analysis cumbersome.

Metabolites usually consist of carbon, hydrogen, nitrogen, and oxygen and to a lower extent sulfur and phosphorus atoms. While hydrogen and nitrogen increase the decimal part of the accurate mass, oxygen, sulfur, and phosphorus atoms decrease it. Therefore, the accurate mass of some metabolites rich in oxygen and phosphorus (e.g., fructose 1,6-bisphosphate $\text{C}_6\text{H}_{14}\text{O}_{12}\text{P}_2$) will be slightly less than the integer nominal mass. In addition, since metabolites are considered small molecules, the decimal part of their accurate mass will rarely exceed 0.4 amu for the majority of metabolites with molecular mass below ~500 amu. Therefore, the optimum set of low-resolution bins for majority of primary metabolites except heavy lipids are $(m_{\text{nom}} - 0.1 \text{ amu}; m_{\text{nom}} + 0.4 \text{ amu})$ and $(m_{\text{nom}} + 0.4 \text{ amu}; m_{\text{nom}} + 0.9 \text{ amu})$ where m_{nom} is a nominal mass. An example is given in Figure 11, which shows a histogram of metabolite decimal place distribution as detected by MarkerLynx (Micromass, UK) in metabolic fingerprints of cell culture media (m/z 100–1,200). Consequently, all ions representing

metabolites will be added into the first bin, while the second bin ideally will contain multi-charged species (Smedsgaard & Nielsen, 2005).

An elegant approach to extract masses from high-resolution mass spectra is to build a density function of masses weighted by their intensities by combining all acquired spectra (Scholz et al., 2004). Thus, masses at high-intensity density are determined as the maximum of the density function. Then, each data file is reexamined and mass intensities, which coincide with the determined maxima within the given instrument precision, are extracted.

When the mass intensity vectors are transformed to uniform length, the entire dataset obtained by direct injection analyses (a single scan) can be represented as a 2D matrix ($\mathbf{m} \times \mathbf{f}$), where \mathbf{m} is the length of the uniform mass vector and \mathbf{f} is the number of files. In contrast, the entire dataset obtained with a separation step, which involves multiple scans, can be characterized by a 3D matrix ($\mathbf{m} \times \mathbf{s} \times \mathbf{f}$), where the number of scans \mathbf{s} is added, representing the retention time. Samples characterized by those matrices can be further classified using multivariate data analysis methods, such as principal component analysis (PCA) or parallel factor (PARAFAC) analysis for 3D matrices. Because the statistical analysis of 3D matrices is computationally difficult, it is also possible to sum mass spectra of specific chromatogram segments and convert them into 2D matrices (Plumb et al., 2002) (see also Fig. 10B). Although the information gained by separation is sacrificed, in the case of high-resolution mass spectrometry, only peaks of separated isomers and metabolites with small mass differences in the low-mamu range lose their identity (see Fig. 6).

The approaches described above are suitable for the classification of samples or the discovery of abundant metabolite peaks. However, the detection of low-abundance peaks in 3D data sets, for example, LC-MS or GC-MS runs, may be required for exact sample classification as well as for the discovery of biomarkers which are expressed at low abundances. In this computationally more intense approach, all detected peaks in the 3D data sets can be represented as a 2D matrix ($\mathbf{n} \times \mathbf{f}$), where \mathbf{n} is the number of individual peaks detected in all samples and \mathbf{f} is the number of files in the datasets. Detection of metabolite peaks at nominal resolution can be performed using available peak detection algorithms (see Fig. 10C) (Windig, Phalp, & Payne, 1996; Koradi et al., 1998; Hastings, Norton, & Roy, 2002). For the analysis of high-resolution mass spectrometry data, signal density functions as those described above can be applied (Scholz et al., 2004) (see also Fig. 10D).

Each collected peak is annotated by a retention time and mass. Hard ionization techniques, such as EI, or post-source fragmentation in MS/MS and MSⁿ experiments add another layer of complexity because each metabolite produces several fragment ions. Thus, spectra deconvolution is required, which also reduces the complexity of the data. However, fragmentation also provides additional structural information. Therefore, each peak may be annotated not only by retention time and ion masses but also by a match from the library search (see Section VII.C). Deconvolution can also be used to reduce the complexity of chromatograms obtained with soft ionization techniques by filtering multiple charged species, clusters, and adducts (Jonsson et al., 2005). To match the peaks extracted from \mathbf{f} files, all mass spectra and scans (chromatograms) have to be aligned across the entire dataset and/or matching criteria should be set, usually the mass and retention time windows. For the alignment, internal reference compounds for retention time and mass measurements can be added to the samples, or major peaks present in all chromatograms are used as references. Externally added non-endogenous compounds are preferred as references because they can also be used as retention index references, which further facilitate the identification of unknowns (see Section VII.C).

A few software products are currently available that meet the described requirements for data export. The Micromass Marker-Lynx Applications Manager (Micromass, Manchester, UK)

was used to detect drug metabolites in urine samples analyzed by LC-ESI-Q-TOF-MS (Plumb et al., 2003a). The program detects peaks in LC-MS chromatograms at high-mass resolution, performs automated peak integration and alignment using a single retention time reference standard, matches detected peaks between samples and outputs data as the described ($n \times f$) matrix in a table which can be analyzed further by multivariate data analysis tools. In general, thousands of peaks (markers) can be extracted from a 1-hr gradient run of a complex biological sample. However, the software is not transparent and does not allow assessing the quality of peak extraction, for example, by visual examination of extracted ion chromatogram integration. Thus, a reasonable noise threshold level should be set to identify significant markers among all peaks.

A test alpha version of metAlign software (Plant Research International, Wageningen, Netherlands) was used for the extraction of plant metabolomics data (Tolstikov et al., 2003). The program performs noise prefiltering, baseline correction, and chromatogram alignment. Preprocessed chromatograms are summed into sample groups and summed chromatograms are compared pairwise. The output of the program provides significantly different characteristic peaks, which differentiate the two chromatograms. Integration of detected peaks is performed manually using instrument software. Similar pairwise comparison can be also performed by the COMPARELC function of the MS Processor program (ACD/Labs, Toronto, Canada) (Tolstikov, UC Davis; personal communication). The program applies an advanced component detection (CODA) algorithm to detect peaks. It is transparent and allows reexamination of each extracted-ion chromatogram of a detected peak.

MS Resolver and ReOrder (Pattern Recognition Systems, Bergen, Norway) were used to process data acquired by LC-ESI-MS at nominal resolution (Idborg-Bjorkman et al., 2003; Idborg, Edlund, & Jacobsson, 2004). MS Resolver performs the extraction of peaks with deconvolution, while ReOrder performs peak alignment using retention time windows and similarity of deconvoluted spectra as matching criteria. As a result, 95 component peaks were detected in chromatograms of SPE-extracted rat urine samples.

Data extraction can be also performed using in-house built software. One of the most suitable environments to build such software is MatLab (The MathWorks, Natick, MA), which is specifically designed to process matrices. In addition, multivariate statistics functions are available in MatLab that can be further used for data analysis. Jonsson et al. applied custom written scripts for MatLab for the extraction of GC-EI-TOF-MS (Jonsson et al., 2004) and LC-ESI-Q-TOF-MS (Jonsson et al., 2005) data by a fast deconvolution algorithm. Chromatograms acquired by GC-EI-TOF-MS were aligned and manually divided into 51 retention time windows in chromatogram regions that contained no peaks. Then mass spectra were summed in each window and corresponding windows from all samples were pooled into 51 matrices. Subsequent deconvolution of those matrices by the alternating regression (AR) method produced a final ($n \times f$) matrix with 246 components detected in extracted aspen leaf samples. For the extraction of LC-ESI-Q-TOF-MS data, a similar approach was used except peaks were detected before alignment by peak apex determination using the simple criterion of a first derivative sign change. Therefore, peak shapes could be omitted and retention time windows were automatically set equal to two scans. As a result, 347 components were detected in a diluted urine sample.

While we were preparing this review, a beta version of XCMS Analyte Profiling software was released (Siuzdak, Scripps Research Institute, San Diego, CA; personal communication). The software is available through Bioconductor, an open source software project (<http://www.bioconductor.org/>) or METLIN database (<http://metlin.scripps.edu/download/>). The program is capable of peak picking, non-linear retention time alignment, and relative quantitation and works with universal netCDF file format. The program is integrated into the

freeware statistical package R (<http://www.r-project.org/>), which provides an excellent environment for subsequent statistical data analysis. MZmine (Katajamaa, Miettinen, & Oresic, 2006) is another promising data analysis program, which was recently introduced.

C. Data Normalization, Transformation, and Analysis

A detailed description of statistical methods for data analysis is beyond the scope of this review. A thorough review of currently available statistical methods was published for related NMR metabolomics (metabolomics) applications (Lindon, Holmes, & Nicholson, 2001). The statistical tools for metabolomics data analysis should be selected according to the aim of the study. If the aim is sample classification and prior information about the sample identity is unknown (e.g., in identification of silent mutation phenotype in progeny), unsupervised methods such as hierarchical clustering analysis (HCA), principal component analysis (PCA), or independent component analysis (ICA) are used. On the other hand, sample identity is often known and the aim of the study is to discover characteristic biomarkers (e.g., search for biomarkers of a disease comparing samples from healthy and diseased subjects). In this case, supervised methods such as principal least square (PLS) or soft-independent method of class analogy (SIMCA) can be used as well. The supervised methods use prior information about sample class and perform better in biomarker discovery (Jonsson et al., 2005). The supervised methods are also the preferred option for diagnostics studies (especially SIMCA), when classes of some samples are known and identities of others need to be determined (e.g., cancer diagnosis by comparing a patient's metabolic fingerprint with the fingerprints of healthy and diseased subjects). The most popular data analysis tool in metabolomics is PCA which is available in many statistical software packages. However, if the aim of the study is a mere biomarker discovery in samples of known classes, common statistical methods such as ANOVA or the recently described ASCA (ANOVA-simultaneous component analysis) can be used (Smilde et al., 2005).

When significantly different markers are discovered, PCA can be used for data visualization. Strictly speaking, PCA is rather a data visualization tool that projects most variable components of a multidimensional data on a 2D or 3D plot, which can be comprehended by human mind in contrast to the multidimensional space. Thus, since variability of the data usually increases with absolute values, data transformation should be performed along with PCA. Otherwise, PCA will discover insignificant but variable components with high intensities and neglect significant but low-abundance biomarkers. The impact of variable components is actually one of the major reasons why high-intersample reproducibility is desired in metabolomics experiments. To minimize the impact of variability of high-intensity peaks, data normalization, and data transformation are used. Each sample is usually normalized to the sum of its intensities or to Euclidean norm. Transformation is intended to reduce high-intensity values but keep low-intensity values. Commonly, a logarithm function is used for the transformation. Because $\log 0 \rightarrow \infty$, an infinitesimal constant (Idborg-Bjorkman et al., 2003) or unity are added to all intensity values. An improvement of log transformation is a general logarithm (glog) transformation that was successfully applied to NMR metabolomics data (Purohit et al., 2004). The glog transformation takes advantage of the fact that the variance does not depend on mean values at low-signal intensities. Thus, more precise cluster analysis can be performed for signals close to the limit of detection.

Another perspective method to apply in metabolomics data analysis is ICA. ICA is similar to PCA but is less strict to the choice of components. It was shown that ICA is able to detect intersample variability caused by instrument drift as a separate component of the data while PCA failed to separate it (Scholz et al., 2004). As a result, improved separation between the samples was achieved because intersample variability was filtered out.

VIII. IDENTIFICATION OF METABOLITES

A. *De Novo* Identification of Metabolites

An important advantage of mass spectrometry tools in metabolomics investigation is achieving low-detection limits, which enables the detection of low-abundance metabolites. However, identification of those compounds can become a problem if other spectroscopic techniques are needed as alternative tools for structure elucidation, as these tools often lack the necessary sensitivity. In this case, multiple fraction collection steps and sample pre-concentration are needed, which is not only labor-intensive but also bears the risk of metabolite decomposition during sample workup. For high-abundance metabolites, proton or other types of NMR is the technique of choice for structure elucidation. Often multiple metabolites can be produced by a single enzyme. Hence, changes in enzyme expression may change the expression of multiple but related metabolites that can be structurally similar due to the substrate specificity of the enzyme. Thus, structure determination of the most abundant differentially expressed metabolite by NMR may provide leads for the structure identification of less-abundant differentially expressed metabolites, which subsequently can be confirmed by mass spectrometry.

Experimental approaches for the structure elucidation by mass spectrometry are extensively described in the literature. Valuable information can be obtained by FT-ICR-MS because of its ultra-high resolution and mass accuracy, which allows determination of elemental composition and its capability for high-resolution MSⁿ experiments, which provides detailed information about molecular structure units. A less expensive alternative to FT-ICR-MS for MSⁿ experiments is a 3D ion trap mass analyzer. Finally, high-resolution Q-TOF instruments can be used for MS/MS experiments providing high-resolution data on fragment ions.

Often a combination of different mass spectrometric techniques is required for the structural elucidation of unknowns. For example, the structure of *N*-acyl taurines, representing a novel class of metabolites, was identified by FT-ICR-MS accurate mass measurements and triple quadrupole MS fragmentation studies (Saghatelian et al., 2004). From accurate mass measurements, the elemental composition was determined, while MS/MS spectra provided evidence of taurine residue in the molecules.

Moreover, derivatization can be used to identify functional groups and number of functional groups based on the observed mass shift. A hydrogen/deuterium exchange assay can be used to determine the number of exchangeable protons as a tool to screen for functional groups such as alcohols, amines, carboxylic acids, etc.

Metabolites can be also identified by library search. However, library search results should be confirmed by comparison with a reference spectrum of the standard and desirably by chromatographic retention of the standard because spectral differences for structural isomers can be minimal or not present at all. Thus, the use of retention time indices provides more confidence in assigning metabolite structure. Retention indices, or more precisely the Kovats indices, are routinely used in GC with the homologous series of *n*-alkanes used as references compounds. However, a retention index system for LC is yet to be established.

B. GC-MS Libraries

EI-MS spectra have been studied for several decades and, therefore, many EI-MS spectra libraries exist containing thousands of spectra. The creation of EI-MS libraries was facilitated by high-interinstrument reproducibility and the standardized electron energy of 70 eV commonly used for ionization. The largest spectra libraries are available from Wiley (~400,000 spectra) and the National Institute of Standards and Technology (NIST) (~200,000 spectra) as well as a combination of both with some additional spectra from Palisade (~600,000 spectra).

Large commercial libraries contain many spectra of synthetic compounds, which are unlikely to appear in biological samples. Although their spectra may provide leads for the structure identification of unknown metabolites, searching through thousands of spectra can be time consuming if hundreds of unknown metabolites need to be identified. In addition, less common metabolites are often not represented in large commercial libraries. However, there are metabolite libraries available, which contain mass spectra of metabolites found in specific biological samples. For example, a set of publicly available libraries of quadrupole and TOF spectra of plant metabolites was recently created as a part of the Golm Metabolome Database (GMD) (Schauer et al., 2005). These MS libraries also contain retention indices (MSRI libraries), which tremendously improve the identification of spectrally similar compounds. One of the tools performing spectra deconvolution and application of retention indices is the Automated Mass Spectral Deconvolution and Identification System (AMDIS) freely available from NIST (<http://chemdata.-nist.gov/mass-spc/amdis/>) (Stein, 1999). The program was used for profiling of carboxylic acids in human urine (Halket et al., 1999) and for the identification of metabolites in phloems of *Curucbita maxima* (Fiehn, 2003).

C. LC-MS Libraries

LC systems are usually interfaced with APCI or ESI ion sources that produce even-electron pseudo-molecular ions ($[M + H]^+$ or $[M - H]^-$) usually without fragmentation. Therefore, additional fragmentation techniques are required such as in-source CID or conventional CID in triple quadrupole instruments or ion traps, which are more selective than in-source techniques and will be discussed below. Fragmentation spectra vary between different types of mass analyzers (triple quadrupole instruments and ion traps are usually compared) and even between instruments of the same mass analyzer type but different brands (Bristow et al., 2004; Jansen, Lachatre, & Marquet, 2005). Thus, the creation of universal libraries for LC-MS is impeded.

To minimize instrument-dependent variability, calibration ions can be used to tune fragmentation conditions according to conditions used to acquire the library spectra (Lemire & Busch, 1996). Ion-trap spectra were shown to be less dependent on collision energy parameters than triple quadrupole spectra, as observed for fragmentation spectra of steroids (Josephs & Sanders, 2004). To reduce the dependence of triple quadrupole instrument libraries on collision energy parameters, the library spectra can be generated by summing several spectra acquired at different collision energies (Josephs & Sanders, 2004). If the criteria for spectrum matching are lowered, for example, as a percentage fit of $\geq 60\%$, satisfactory inter-instrumental identification through a library search can be achieved (Bristow et al., 2004). However, it is unclear if this criterion will be specific enough for the analysis of complex mixtures of metabolites. For metabolites that readily ionize in both ionization modes, similarity of both positive and negative ionization spectra can be a strong-fit criterion, even if a low-fit criterion is used for individual spectra.

Despite the described difficulties, MS/MS spectral libraries are being created. Recently NIST has published a new update of their library (NIST 2005) containing 5,191 MS/MS spectra of 1,943 different ions (1,671 positive and 341 negative ions). The spectra were acquired with different triple quadrupole and ion trap instruments. In addition, several user libraries exist, which are usually built using the same instrument model. To date 79 MS/MS spectra of common metabolites and drugs acquired with a Q-TOF instrument are available at the METLIN database (<http://metlin.scripps.edu/>) (Smith et al., 2005). To increase the specificity of library search, retention time is also included in the library. This approach was used for the identification of drug metabolites in urine samples (Pelander et al., 2003). A more detailed overview on currently available LC-MS and GC-MS libraries was recently published (Halket et al., 2005).

IX. DATABASES

Although a number of genomics and proteomics databases are available now, metabolomics databases are still limited. Biochemical databases can be used to identify unknown metabolites, for example, to identify structure from known elemental composition, or to determine the biological function of the identified metabolite. Information about biochemical pathways and metabolites involved in those pathways are available in the KEGG (<http://www.genome.jp/kegg/>) and BioCyc (<http://biocyc.org/>) databases. More specific lipidomics databases exist, such as LipidMaps (<http://www.lipidmaps.org/data/index.html>), SphinGOMAP (<http://sphingomap.org/>), and Lipid Bank (<http://lipidbank.jp/index00.shtml>), which contain structural and nomenclatural information as well as standard analytical protocols. General information about physico-chemical properties of metabolites can be obtained by searching general chemical databases such as PubChem or CAS. A general database search can also be useful for identifying the chemical structure of unknowns by starting with the elemental composition, which can be obtained from accurate mass measurements. For example, a search in the Merck Index was used to identify unknown water pollutants (Ibanez et al., 2005). The development of specific metabolomics mass spectrometry databases is also in progress. The Golm metabolome database (GMD) (<http://csbdb.mpimp-golm.mpg.de/csbdb/gmd/gmd.html>) contains publicly available mass spectra and retention index metabolite (MSRI) libraries and GC-MS metabolic profiles of plant samples (Kopka et al., 2005). The METLIN database (<http://metlin.scripps.edu/>) catalogues metabolites, MS/MS spectra, and LC-MS profiles of human plasma and urine samples (Smith et al., 2005). To date, existing metabolomics databases aim primarily at the structural identification of metabolites in various biological samples. However, once a better annotation of the metabolome in various organisms is achieved, the generation of databases containing quantitative metabolite data can be expected. An example for this type of database is the human metabolite database that contains more than 1,400 metabolites found in the human body. Each metabolite is described by a MetaboCard designed to contain chemical data, clinical data and molecular biology/biochemistry data (<http://www.hmdb.ca/>).

X. CONCLUSIONS

Technological advances in NMR and mass spectrometry have opened a new chapter in biochemistry by introducing metabolomics as an approach to study metabolism and its regulation in relation to disease, genetic, and environmental factors. Since the metabolome is the most predictive of phenotype, metabolomics holds the promise to extensively contribute to the understanding of phenotypic changes as an organism's answer to disease, genetic changes, and nutritional, toxicological, environmental, and pharmacological influences. With regards to human health alone, multiple benefits of metabolomics investigations can be envisioned. It can deliver new tools to diagnose disease or monitor the success of nutritional and pharmacological interventions, for example, by using an unbiased metabolic fingerprinting approach. Based on research studies, metabolomics can provide new biomarkers to assess human health, and over time a powerful list of diagnostic markers will evolve, which can be measured using high-throughput assays. However, in order to proceed from the single biomarker concept to the global metabolome evaluation outside a research environment, the technology has yet to be developed to provide the clinician with the tools to assess entire wide classes of metabolites in biofluids and automatically process the data to evaluate the biochemical status of an individual.

The high selectivity of mass spectrometers in combination with low-detection limits, as well as their compatibility with separation techniques and their ability to deliver quantitative data makes mass spectrometry an ideal tool for metabolomic applications. So far, several excellent scientific contributions have been published demonstrating the enormous power of mass spectrometry applied to metabolomics. However, the field is rapidly developing and exciting

new techniques and applications are anticipated in the near future, as we have seen in the rapidly growing field of proteomics.

Nevertheless, many technical and methodological issues have to be addressed to create analytical platforms that readily answer biological questions efficiently.

A major obstacle in this type of investigation is high diversity and variability encountered in the metabolome. The physico-chemical properties of metabolites are so diverse that none of the currently available mass spectrometric techniques can analyze all metabolites simultaneously. In general, an efficient sample introduction and ionization of all metabolite classes is of high importance. While EI and ESI ion sources are very efficient for certain types of analytes they have their own limitations. Among liquid MS interfaces, a promising ionization technique is the combination of ESI/APCI and atmospheric pressure photo ionization (APPI) or laser-based ionization that enhances ionization efficiency for non-polar species. In addition, a more efficient separation of metabolites using chromatographic or electrophoretic techniques will decrease interferences and can provide lower detection limits. Multidimensional chromatographic techniques, such as GC \times GC, are expected to gain wider distribution. A major goal is of course to further decrease the detection limits of the techniques used, not only to expand metabolomics studies towards the detection of low-abundance metabolites, but also to move to small sample volumes. This is not only important in the area of medicine but also opens up the exciting field of single-cell metabolomics. The trend towards miniaturization is already in progress with the evolving lab-on-the-chip technologies. Moreover, automated high-throughput sample preparation tools that are interfaced with the analytical platform are a vital technological component for metabolomics. A major obstacle in global metabolomics investigations using mass spectrometry is the quantification of the metabolites encountered, because signal intensity is not only function of concentration or mass but also depends on the chemical structure of the analytes and can be influenced by matrix interferences as observed in ESI. Stable isotope-labeled internal standards are therefore commonly used in metabolic profiling techniques. However, it seems unrealistic at this point that isotopic-labeled standards will become available, for example, for all plasma metabolites. Nevertheless, a centralized effort to design standard sets of isotopic-labeled reference compounds covering a wide range of metabolites would tremendously accelerate the work in metabolomics and would aid in the inter-laboratory comparison of semi-quantitative metabolic fingerprinting data.

Besides the technological aspects, data standardization, export, and analysis are essential components of metabolomics (Mendes, 2002). While high-throughput and semi-quantitative proteomics, with automated identification of proteins by computer algorithms, is available now (e.g., SEQUEST algorithm search), data analysis and visualization tools, libraries, and databases for metabolomics have yet to be developed.

Acknowledgements

This work was supported by NIEHS Grants R37 ES002710, R01 ES013933, P30 ES05707, NIEHS Superfund Basic Research Program Grant P42 ES04699, and NIEHS Center for Children's Environmental Health & Disease Prevention Grant P01 ES11269, the Cure Autism Now Foundation Frank del Olmo Memorial Award and in part by the German Research Foundation. We thank D. Hanna for the discussion of the manuscript.

References

- Adahchour M, Brandt M, Baier HU, Vreuls RJ, Batenburg AM, Brinkman UA. Comprehensive two-dimensional gas chromatography coupled to a rapid-scanning quadrupole mass spectrometer: Principles and applications. *J Chromatogr A* 2005;1067:245–254. [PubMed: 15844530]
- Aharoni A, Ric de Vos CH, Verhoeven HA, Maliepaard CA, Kruppa G, Bino R, Goodenow DB. Nontargeted metabolome analysis by use of Fourier transform ion cyclotron mass spectrometry. *Omics* 2002;6:217–234. [PubMed: 12427274]

- Allen J, Davey HM, Broadhurst D, Heald JK, Rowland JJ, Oliver SG, Kell DB. High-throughput classification of yeast mutants for functional genomics using metabolic footprinting. *Nat Biotechnol* 2003;21:692–696. [PubMed: 12740584]
- Allen J, Davey HM, Broadhurst D, Rowland JJ, Oliver SG, Kell DB. Discrimination of modes of action of antifungal substances by use of metabolic footprinting. *Appl Environ Microbiol* 2004;70:6157–6165. [PubMed: 15466562]
- Balazy M. Eicosanomics: Targeted lipidomics of eicosanoids in biological systems. *Prostaglandins Other Lipid Mediat* 2004;73:173–180. [PubMed: 15287150]
- Basile F, Beverly MB, Abbas-Hawks C, Mowry CD, Voorhees KJ, Hadfield TL. Direct mass spectrometric analysis of in situ thermally hydrolyzed and methylated lipids from whole bacterial cells. *Anal Chem* 1998;70:1555–1562. [PubMed: 9569765]
- Beecher, CWW. The human metabolome. In: George, G.; Harrigan, RG.; George, G.; Harrigan, RG., editors. *Metabolic profiling: Its role in biomarker discovery and gene function analysis*. New York, NY: Springer; 2003.
- Bino RJ, de Vos CH, Lieberman M, Hall RD, Bovy A, Jonker HH, Tikunov Y, Lommen A, Moco S, Levin I. The light-hyperresponsive high pigment-2 mutation of tomato: Alterations in the fruit metabolome. *New Phytol* 2005;166:427–438. [PubMed: 15819907]
- Bino RJ, Hall RD, Fiehn O, Kopka J, Saito K, Draper J, Nikolau BJ, Mendes P, Roessner-Tunali U, Beale MH, Trethewey RN, Lange BM, Wurtele ES, Sumner LW. Potential of metabolomics as a functional genomics tool. *Trends Plant Sci* 2004;9:418–425. [PubMed: 15337491]
- Birkemeyer C, Kolasa A, Kopka J. Comprehensive chemical derivatization for gas chromatography-mass spectrometry-based multi-targeted profiling of the major phytohormones. *J Chromatogr A* 2003;993:89–102. [PubMed: 12735441]
- Bristow AW, Webb KS, Lubben AT, Halket J. Reproducible production tandem mass spectra on various liquid chromatography/mass spectrometry instruments for the development of spectral libraries. *Rapid Commun Mass Spectrom* 2004;18:1447–1454. [PubMed: 15216504]
- Brown SC, Kruppa G, Dasseux JL. Metabolomics applications of FT-ICR mass spectrometry. *Mass Spectrom Rev* 2005;24:223–231. [PubMed: 15389859]
- Buchholz A, Hurlebaus J, Wandrey C, Takors R. Metabolomics: Quantification of intracellular metabolite dynamics. *Biomol Eng* 2002;19:5–15. [PubMed: 12103361]
- Buchholz A, Takors R, Wandrey C. Quantification of intracellular metabolites in *Escherichia coli* K12 using liquid chromatographic-electrospray ionization tandem mass spectrometric techniques. *Anal Biochem* 2001;295:129–137. [PubMed: 11488613]
- Buchholz BA, Fultz E, Haack KW, Vogel JS, Gilman SD, Gee SJ, Hammock BD, Hui X, Wester RC, Maibach HI. HPLC-accelerator MS measurement of atrazine metabolites in human urine after dermal exposure. *Anal Chem* 1999;71:3519–3525. [PubMed: 10464479]
- Cannizzaro C, Christensen B, Nielsen J, von Stockar U. Metabolic network analysis on *Phaffia rhodozyma* yeast using ¹³C-labeled glucose and gas chromatography-mass spectrometry. *Metab Eng* 2004;6:340–351. [PubMed: 15491863]
- Castrillo JI, Hayes A, Mohammed S, Gaskell SJ, Oliver SG. An optimized protocol for metabolome analysis in yeast using direct infusion electrospray mass spectrometry. *Phytochemistry* 2003;62:929–937. [PubMed: 12590120]
- Chace DH, Kalas TA. A biochemical perspective on the use of tandem mass spectrometry for newborn screening and clinical testing. *Clin Biochem* 2005;38:296–309. [PubMed: 15766731]
- Chace DH, Kalas TA, Naylor EW. The application of tandem mass spectrometry to neonatal screening for inherited disorders of intermediary metabolism. *Annu Rev Genomics Hum Genet* 2002;3:17–45. [PubMed: 12142359]
- Chace DH, Kalas TA, Naylor EW. Use of tandem mass spectrometry for multianalyte screening of dried blood specimens from newborns. *Clin Chem* 2003;49:1797–1817. [PubMed: 14578311]
- Colebatch G, Desbrosses G, Ott T, Krusell L, Montanari O, Kloska S, Kopka J, Udvardi MK. Global changes in transcription orchestrate metabolic differentiation during symbiotic nitrogen fixation in *Lotus japonicus*. *Plant J* 2004;39:487–512. [PubMed: 15272870]
- Dettmer K, Hammock BD. Metabolomics—A new exciting field within the “omics” sciences. *Environ Health Perspect* 2004;112:A396–397. [PubMed: 15159211]

- Drechsel D, Dettmer K, Engewald W. Studies of thermally assisted hydrolysis and methylation—GC-MS of fatty acids and triglycerides using different reagents and injection systems. *Chromatographia* 2003;57:S283–S289.
- Dueker SR, Lin Y, Buchholz BA, Schneider PD, Lame MW, Segall HJ, Vogel JS, Clifford AJ. Long-term kinetic study of beta-carotene, using accelerator mass spectrometry in an adult volunteer. *J Lipid Res* 2000;41:1790–1800. [PubMed: 11060348]
- Duffin KL, Henion JD, Shieh JJ. Electrospray and tandem mass spectrometric characterization of acylglycerol mixtures that are dissolved in nonpolar solvents. *Anal Chem* 1991;63:1781–1788. [PubMed: 1789441]
- Fang J, Barcelona MJ. Structural determination and quantitative analysis of bacterial phospholipids using liquid chromatography/electrospray ionization/mass spectrometry. *J Microbiol Methods* 1998;33:23–35.
- Fiehn O. Metabolomics—The link between genotypes and phenotypes. *Plant Mol Biol* 2002;48:155–171. [PubMed: 11860207]
- Fiehn O. Metabolic networks of *Cucurbita maxima* phloem. *Phytochemistry* 2003;62:875–886. [PubMed: 12590115]
- Fiehn O, Kopka J, Trethewey RN, Willmitzer L. Identification of uncommon plant metabolites based on calculation of elemental compositions using gas chromatography and quadrupole mass spectrometry. *Anal Chem* 2000;72:3573–3580. [PubMed: 10952545]
- Fischer E, Sauer U. Metabolic flux profiling of *Escherichia coli* mutants in central carbon metabolism using GC-MS. *Eur J Biochem* 2003;270:880–891. [PubMed: 12603321]
- Forrester JS, Milne SB, Ivanova PT, Brown HA. Computational lipidomics: A multiplexed analysis of dynamic changes in membrane lipid composition during signal transduction 10.1124/mol.65.4.813. *Mol Pharmacol* 2004;65:813–821. [PubMed: 15044609]
- Fu X, Iga M, Kimura M, Yamaguchi S. Simplified screening for organic acidemia using GC/MS and dried urine filter paper: A study on neonatal mass screening. *Early Hum Dev* 2000;58:41–55. [PubMed: 10785335]
- German JB, Roberts MA, Watkins SM. Genomics and metabolomics as markers for the interaction of diet and health: Lessons from lipids. *J Nutr* 2003;133:2078S–2083S. [PubMed: 12771369]
- Gonzalez B, Francois J, Renaud M. A rapid and reliable method for metabolite extraction in yeast using boiling buffered ethanol. *Yeast* 1997;13:1347–1355. [PubMed: 9392079]
- Griffiths WJ. Tandem mass spectrometry in the study of fatty acids, bile acids, and steroids. *Mass Spectrom Rev* 2003;22:81–152. [PubMed: 12820273]
- Griffiths WJ, Jonsson AP, Liu S, Rai DK, Wang Y. Electrospray and tandem mass spectrometry in biochemistry. *Biochem J* 2001;355:545–561. [PubMed: 11311115]
- Gullberg J, Jonsson P, Nordstrom A, Sjoström M, Moritz T. Design of experiments: An efficient strategy to identify factors influencing extraction and derivatization of *Arabidopsis thaliana* samples in metabolomic studies with gas chromatography/mass spectrometry. *Anal Biochem* 2004;331:283–295. [PubMed: 15265734]
- Halket JM, Przyborowska A, Stein SE, Mallard WG, Down S, Chalmers RA. Deconvolution gas chromatography/mass spectrometry of urinary organic acids—potential for pattern recognition and automated identification of metabolic disorders. *Rapid Commun Mass Spectrom* 1999;13:279–284. [PubMed: 10097403]
- Halket JM, Waterman D, Przyborowska AM, Patel RK, Fraser PD, Bramley PM. Chemical derivatization and mass spectral libraries in metabolic profiling by GC/MS and LC/MS/MS. *J Exp Bot* 2005;56:219–243. [PubMed: 15618298]
- Han X, Gross RW. Global analyses of cellular lipidomes directly from crude extracts of biological samples by ESI mass spectrometry: A bridge to lipidomics. *J Lipid Res* 2003;44:1071–1079. [PubMed: 12671038]
- Han X, Gross RW. Shotgun lipidomics: Electrospray ionization mass spectrometric analysis and quantitation of cellular lipidomes directly from crude extracts of biological samples. *Mass Spectrom Rev* 2005;24:367–412. [PubMed: 15389848]
- Harrigan GG, LaPlante RH, Cosma GN, Cockerell G, Goodacre R, Maddox JF, Luyendyk JP, Ganey PE, Roth RA. Application of high-throughput Fourier-transform infrared spectroscopy in toxicology

- studies: Contribution to a study on the development of an animal model for idiosyncratic toxicity. *Toxicol Lett* 2004;146:197–205. [PubMed: 14687757]
- Hastings CA, Norton SM, Roy S. New algorithms for processing and peak detection in liquid chromatography/mass spectrometry data. *Rapid Commun Mass Spectrom* 2002;16:462–467. [PubMed: 11857732]
- Hellerstein MK. New stable isotope-mass spectrometric techniques for measuring fluxes through intact metabolic pathways in mammalian systems: Introduction of moving pictures into functional genomics and biochemical phenotyping. *Metab Eng* 2004;6:85–100. [PubMed: 14734258]
- Hellerstein MK, Neese RA, Linfoot P, Christiansen M, Turner S, Letscher A. Hepatic gluconeogenic fluxes and glycogen turnover during fasting in humans. A stable isotope study. *J Clin Invest* 1997;100:1305–1319. [PubMed: 9276749]
- Hermansson M, Uphoff A, Kakela R, Somerharju P. Automated quantitative analysis of complex lipidomes by liquid chromatography/mass spectrometry. *Anal Chem* 2005;77:2166–2175. [PubMed: 15801751]
- Hirai MY, Yano M, Goodenowe DB, Kanaya S, Kimura T, Awazuhara M, Arita M, Fujiwara T, Saito K. Integration of transcriptomics and metabolomics for understanding of global responses to nutritional stresses in *Arabidopsis thaliana*. *Proc Natl Acad Sci USA* 2004;101:10205–10210. [PubMed: 15199185]
- Holmes E, Nicholson JK, Tranter G. Metabonomic characterization of genetic variations in toxicological and metabolic responses using probabilistic neural networks. *Chem Res Toxicol* 2001;14:182–191. [PubMed: 11258967]
- Hsu FF, Turk J. Structural characterization of triacylglycerols as lithiated adducts by electrospray ionization mass spectrometry using low-energy collisionally activated dissociation on a triple stage quadrupole instrument. *J Am Soc Mass Spectrom* 1999;10:587–599. [PubMed: 10384723]
- Ibanez M, Sancho JV, Pozo OJ, Niessen W, Hernandez F. Use of quadrupole time-of-flight mass spectrometry in the elucidation of unknown compounds present in environmental water. *Rapid Commun Mass Spectrom* 2005;19:169–178. [PubMed: 15593265]
- Idborg H, Edlund PO, Jacobsson SP. Multivariate approaches for efficient detection of potential metabolites from liquid chromatography/mass spectrometry data. *Rapid Commun Mass Spectrom* 2004;18:944–954. [PubMed: 15116421]
- Idborg-Bjorkman H, Edlund PO, Kvalheim OM, Schuppe-Koistinen I, Jacobsson SP. Screening of biomarkers in rat urine using LC/electrospray ionization-MS and two-way data analysis. *Anal Chem* 2003;75:4784–4792. [PubMed: 14674455]
- Jansen R, Lachatre G, Marquet P. LC-MS/MS systematic toxicological analysis: Comparison of MS/MS spectra obtained with different instruments and settings. *Clin Biochem* 2005;38:362–372. [PubMed: 15766737]
- Johnson HE, Broadhurst D, Kell DB, Theodorou MK, Merry RJ, Griffith GW. High-throughput metabolic fingerprinting of legume silage fermentations via Fourier transform infrared spectroscopy and chemometrics. *Appl Environ Microbiol* 2004;70:1583–1592. [PubMed: 15006782]
- Jonsson P, Bruce SJ, Moritz T, Trygg J, Sjostrom M, Plumb R, Granger J, Maibaum E, Nicholson JK, Holmes E, Antti H. Extraction, interpretation and validation of information for comparing samples in metabolic LC/MS data sets. *Analyst* 2005;130:701–707. [PubMed: 15852140]
- Jonsson P, Gullberg J, Nordstrom A, Kusano M, Kowalczyk M, Sjostrom M, Moritz T. A strategy for identifying differences in large series of metabolomic samples analyzed by GC/MS. *Anal Chem* 2004;76:1738–1745. [PubMed: 15018577]
- Josephs JL, Sanders M. Creation and comparison of MS/MS spectral libraries using quadrupole ion trap and triple-quadrupole mass spectrometers. *Rapid Commun Mass Spectrom* 2004;18:743–759. [PubMed: 15052556]
- Katajamaa M, Miettinen J, Oresic M. MZmine: Toolbox for processing and visualization of mass spectrometry based molecular profile data. *Bioinformatics* 2006;22:1023–1030. doi:10.1093/bioinformatics/btk039
- Katz JE, Dumlao DS, Clarke S, Hau J. A new technique (COMSPARI) to facilitate the identification of minor compounds in complex mixtures by GC/MS and LC/MS: Tools for the visualization of matched datasets. *J Am Soc Mass Spectrom* 2004;15:580–584. [PubMed: 15047062]

- Klampfl CW. Review coupling of capillary electrochromatography to mass spectrometry. *J Chromatogr A* 2004;1044:131–144. [PubMed: 15354433]
- Kopka J, Schauer N, Krueger S, Birkemeyer C, Usadel B, Bergmuller E, Dormann P, Weckwerth W, Gibon Y, Stitt M, Willmitzer L, Fernie AR, Steinhauser D. GMD@CSB.DB: The Golm Metabolome Database. *Bioinformatics* 2005;21:1635–1638. [PubMed: 15613389]
- Koradi R, Billeter M, Engeli M, Guntert P, Wuthrich K. Automated peak picking and peak integration in macromolecular NMR spectra using AUTOPSY. *J Magn Reson* 1998;135:288–297. [PubMed: 9878459]
- Kuhara T. Gas chromatographic-mass spectrometric urinary metabolome analysis to study mutations of inborn errors of metabolism. *Mass Spectrom Rev* 2005;24:814–827. [PubMed: 15376278]
- Lafaye A, Junot C, Ramounet-Le Gall B, Fritsch P, Tabet JC, Ezan E. Metabolite profiling in rat urine by liquid chromatography/electrospray ion trap mass spectrometry. Application to the study of heavy metal toxicity. *Rapid Commun Mass Spectrom* 2003;17:2541–2549. [PubMed: 14608626]
- Lee SH, Williams MV, DuBois RN, Blair IA. Targeted lipidomics using electron capture atmospheric pressure chemical ionization mass spectrometry. *Rapid Commun Mass Spectrom* 2003;17:2168–2176. [PubMed: 14515314]
- Lemire SW, Busch KL. Calibration point for liquid secondary ion mass spectrometry tandem mass spectra measured with an EBqQ hybrid mass spectrometer. *J Mass Spectrom* 1996;31:280–288.
- Lemke SL, Dueker SR, Follett JR, Lin Y, Carkeet C, Buchholz BA, Vogel JS, Clifford AJ. Absorption and retinol equivalence of beta-carotene in humans is influenced by dietary vitamin A intake. *J Lipid Res* 2003;44:1591–1600. [PubMed: 12777466]
- Lenz EM, Bright J, Knight R, Wilson ID, Major H. A metabonomic investigation of the biochemical effects of mercuric chloride in the rat using ¹H NMR and HPLC-TOF/MS: Time dependent changes in the urinary profile of endogenous metabolites as a result of nephrotoxicity. *Analyst* 2004a; 129:535–541. [PubMed: 15152332]
- Lenz EM, Bright J, Knight R, Wilson ID, Major H. Cyclosporin A-induced changes in endogenous metabolites in rat urine: A metabonomic investigation using high field ¹H NMR spectroscopy, HPLC-TOF/MS and chemometrics. *J Pharm Biomed Anal* 2004b;35:599–608. [PubMed: 15137985]
- Levsen K, Schiebel HM, Behnke B, Dotzer R, Dreher W, Elend M, Thiele H. Structure elucidation of phase II metabolites by tandem mass spectrometry: An overview. *J Chromatogr A* 2005;1067:55–72. [PubMed: 15844510]
- Lin Y, Dueker SR, Follett JR, Fadel JG, Arjomand A, Schneider PD, Miller JW, Green R, Buchholz BA, Vogel JS, Phair RD, Clifford AJ. Quantitation of in vivo human folate metabolism. *Am J Clin Nutr* 2004;80:680–691. [PubMed: 15321809]
- Lindon JC, Holmes E, Nicholson JK. Pattern recognition methods and applications in biomedical magnetic resonance. *Prog Nucl Magn Reson Spectrosc* 2001;39:1–40.
- Little JL. Artifacts in trimethylsilyl derivatization reactions and ways to avoid them. *J Chromatogr A* 1999;844:1–22. [PubMed: 10399322]
- Maharjan RP, Ferenci T. Global metabolite analysis: The influence of extraction methodology on metabolome profiles of *Escherichia coli*. *Anal Biochem* 2003;313:145–154. [PubMed: 12576070]
- Margalit A, Duffin KL, Isakson PC. Rapid quantitation of a large scope of eicosanoids in two models of inflammation: Development of an electrospray and tandem mass spectrometry method and application to biological studies. *Anal Biochem* 1996;235:73–81. [PubMed: 8850549]
- Marshall AG, Hendrickson CL, Jackson GS. Fourier transform ion cyclotron resonance mass spectrometry: A primer. *Mass Spectrom Rev* 1998;17:1–35. [PubMed: 9768511]
- Mashego MR, Wu L, Van Dam JC, Ras C, Vinke JL, Van Winden WA, Van Gulik WM, Heijnen JJ. MIRACLE: Mass isotopomer ratio analysis of U-¹³C-labeled extracts. A new method for accurate quantification of changes in concentrations of intracellular metabolites. *Biotechnol Bioeng* 2004;85:620–628. [PubMed: 14966803]
- Mendes P. Emerging bioinformatics for the metabolome. *Brief Bioinform* 2002;3:134–145. [PubMed: 12139433]
- Merrill AH Jr, Sullards MC, Allegood JC, Kelly S, Wang E. Sphingolipidomics: High-throughput, structure-specific, and quantitative analysis of sphingolipids by liquid chromatography tandem mass spectrometry. *Methods* 2005;36:207–224. [PubMed: 15894491]

- Mills GA, Walker V. Headspace solid-phase microextraction procedures for gas chromatographic analysis of biological fluids and materials. *J Chromatogr A* 2000;902:267–287. [PubMed: 11192159]
- Mills GA, Walker V. Headspace solid-phase microextraction profiling of volatile compounds in urine: Application to metabolic investigations. *J Chromatogr B Biomed Sci Appl* 2001;753:259–268. [PubMed: 11334339]
- Mitra, S. *Sample preparation techniques in analytical chemistry*. Hoboken; Wiley: 2003.
- Naidong W. Bioanalytical liquid chromatography tandem mass spectrometry methods on underivatized silica columns with aqueous/organic mobile phases. *J Chromatogr B Analyt Technol Biomed Life Sci* 2003;796:209–224.
- Newman JW, Watanabe T, Hammock BD. The simultaneous quantification of cytochrome P450 dependent linoleate and arachidonate metabolites in urine by HPLC-MS/MS. *J Lipid Res* 2002;43:1563–1578. [PubMed: 12235189]
- Nicholson JK, Wilson ID. Opinion: Understanding ‘global’ systems biology: Metabonomics and the continuum of metabolism. *Nat Rev Drug Discov* 2003;2:668–676. [PubMed: 12904817]
- Nielsen KF, Smedsgaard J. Fungal metabolite screening: Database of 474 mycotoxins and fungal metabolites for dereplication by standardised liquid chromatography-UV-mass spectrometry methodology. *J Chromatogr A* 2003;1002:111–136. [PubMed: 12885084]
- Pawliszyn, J. *Solid phase microextraction theory and practice*. New York: Wiley-VCH, Inc; 1997.
- Pedrioli PGA, Eng JK, Hubley R, Vogelzang M, Deutsch EW, Raught B, Pratt B, Nilsson E, Angeletti RH, Apweiler R, Cheung K, Costello CE, Hermjakob H, Huang S, Julian RK, Kapp E, McComb ME, Oliver SG, Omenn G, Paton NW, Simpson R, Smith R, Taylor CF, Zhu W, Aebersold R. A common open representation of mass spectrometry data and its application to proteomics research. *Nat Biotech* 2004;22:1459–1466.
- Pelander A, Ojanpera I, Laks S, Rasanen I, Vuori E. Toxicological screening with formula-based metabolite identification by liquid chromatography/time-of-flight mass spectrometry. *Anal Chem* 2003;75:5710–5718. [PubMed: 14588010]
- Pelizzi N, Catinella S, Barbosa S, Zanol M. Different electrospray tandem mass spectrometric approaches for rapid characterization of phospholipid classes of Curosurf, a natural pulmonary surfactant. *Rapid Commun Mass Spectrom* 2002;16:2215–2220. [PubMed: 12478563]
- Pham-Tuan H, Kaskavelis L, Daykin CA, Janssen HG. Method development in high-performance liquid chromatography for high-throughput profiling and metabonomic studies of biofluid samples. *J Chromatogr B Analyt Technol Biomed Life Sci* 2003;789:283–301.
- Piraud M, Vianey-Saban C, Petritis K, Elfakir C, Steghens JP, Morla A, Bouchu D. ESI-MS/MS analysis of underivatized amino acids: A new tool for the diagnosis of inherited disorders of amino acid metabolism. Fragmentation study of 79 molecules of biological interest in positive and negative ionisation mode. *Rapid Commun Mass Spectrom* 2003;17:1297–1311. [PubMed: 12811753]
- Plumb R, Granger J, Stumpf C, Wilson ID, Evans JA, Lenz EM. Metabonomic analysis of mouse urine by liquid-chromatography-time of flight mass spectrometry (LC-TOFMS): Detection of strain, diurnal and gender differences. *Analyst* 2003a;128:819–823. [PubMed: 12894816]
- Plumb RS, Stumpf CL, Gorenstein MV, Castro-Perez JM, Dear GJ, Anthony M, Sweatman BC, Connor SC, Haselden JN. Metabonomics: The use of electrospray mass spectrometry coupled to reversed-phase liquid chromatography shows potential for the screening of rat urine in drug development. *Rapid Commun Mass Spectrom* 2002;16:1991–1996. [PubMed: 12362392]
- Plumb RS, Stumpf CL, Granger JH, Castro-Perez J, Haselden JN, Dear GJ. Use of liquid chromatography/time-of-flight mass spectrometry and multivariate statistical analysis shows promise for the detection of drug metabolites in biological fluids. *Rapid Commun Mass Spectrom* 2003b;17:2632–2638. [PubMed: 14648901]
- Pulfer M, Murphy RC. Electrospray mass spectrometry of phospholipids. *Mass Spectrom Rev* 2003;22:332–364. [PubMed: 12949918]
- Purohit PV, Rocke DM, Viant MR, Woodruff DL. Discrimination models using variance-stabilizing transformation of metabolomic NMR data. *Omics* 2004;8:118–130. [PubMed: 15268771]
- Roessner U, Luedemann A, Brust D, Fiehn O, Linke T, Willmitzer L, Fernie A. Metabolic profiling allows comprehensive phenotyping of genetically or environmentally modified plant systems. *Plant Cell* 2001;13:11–29. [PubMed: 11158526]

- Roessner U, Wagner C, Kopka J, Trethewey RN, Willmitzer L. Technical advance: Simultaneous analysis of metabolites in potato tuber by gas chromatography-mass spectrometry. *Plant J* 2000;23:131–142. [PubMed: 10929108]
- Roessner U, Willmitzer L, Fernie AR. High-resolution metabolic phenotyping of genetically and environmentally diverse potato tuber systems. Identification of phenocopies. *Plant Physiol* 2001;127:749–764. [PubMed: 11706160]
- Roessner-Tunali U, Liu J, Leisse A, Balbo I, Perez-Melis A, Willmitzer L, Fernie AR. Kinetics of labelling of organic and amino acids in potato tubers by gas chromatography-mass spectrometry following incubation in (13)C labelled isotopes. *Plant J* 2004;39:668–679. [PubMed: 15272882]
- Ruijter GJG, Visser J. Determination of intermediary metabolites in *Aspergillus niger*. *J Microbiol Methods* 1996;25:295–302.
- Saghatelian A, Trauger SA, Want EJ, Hawkins EG, Siuzdak G, Cravatt BF. Assignment of endogenous substrates to enzymes by global metabolite profiling. *Biochemistry* 2004;43:14332–14339. [PubMed: 15533037]
- Sato S, Soga T, Nishioka T, Tomita M. Simultaneous determination of the main metabolites in rice leaves using capillary electrophoresis mass spectrometry and capillary electrophoresis diode array detection. *Plant J* 2004;40:151–163. [PubMed: 15361149]
- Sauer U. High-throughput phenomics: Experimental methods for mapping fluxomes. *Curr Opin Biotechnol* 2004;15:58–63. [PubMed: 15102468]
- Schauer N, Steinhäuser D, Strelkov S, Schomburg D, Allison G, Moritz T, Lundgren K, Roessner-Tunali U, Forbes MG, Willmitzer L, Fernie AR, Kopka J. GC-MS libraries for the rapid identification of metabolites in complex biological samples. *FEBS Lett* 2005;579:1332–1337. [PubMed: 15733837]
- Schmelzer KR, Kubala L, Newman JW, Kim IH, Eiserich JP, Hammock BD. Soluble epoxide hydrolase is a therapeutic target for acute inflammation. *Proc Natl Acad Sci USA* 2005;102:9772–9777. [PubMed: 15994227]
- Scholz M, Gatzek S, Sterling A, Fiehn O, Selbig J. Metabolite fingerprinting: Detecting biological features by independent component analysis. *Bioinformatics* 2004;20:2447–2454. [PubMed: 15087312]
- Schulze A, Lindner M, Kohlmüller D, Olgemöller K, Mayatepek E, Hoffmann GF. Expanded newborn screening for inborn errors of metabolism by electrospray ionization-tandem mass spectrometry: Results, outcome, and implications. *Pediatrics* 2003;111:1399–1406. [PubMed: 12777559]
- Smedsgaard J. Micro-scale extraction procedure for standardized screening of fungal metabolite production in cultures. *J Chromatogr A* 1997;760:264–270. [PubMed: 9062989]
- Smedsgaard J, Frisvad JC. Using direct electrospray mass spectrometry in taxonomy and secondary metabolite profiling of crude fungal extracts. *J Microbiol Meth* 1996;25:5–17.
- Smedsgaard J, Frisvad JC. Terverticillate penicillia studied by direct electrospray mass spectrometric profiling of crude extracts. I. Chemo-systematics. *Biochem Syst Ecol* 1997;25:51–64.
- Smedsgaard J, Nielsen J. Metabolite profiling of fungi and yeast: From phenotype to metabolome by MS and informatics. *J Exp Bot* 2005;56:273–286. [PubMed: 15618299]
- Smilde AK, Jansen JJ, Hoefsloot HC, Lamers RJ, van der Greef J, Timmerman ME. ANOVA-simultaneous component analysis (ASCA): A new tool for analyzing designed metabolomics data. *Bioinformatics* 2005;21:3043–3048. [PubMed: 15890747]
- Smith CA, O'Maille G, Want EJ, Qin C, Trauger SA, Brandon TR, Custodio DE, Abagyan R, Siuzdak G. METLIN: A metabolite mass spectral database. *Ther Drug Monit* 2005;27:747–751. [PubMed: 16404815]
- Snyder AP, McClennen WH, Dworzanski JP, Meuzelaar HL. Characterization of underivatized lipid biomarkers from microorganisms with pyrolysis short-column gas chromatography/ion trap mass spectrometry. *Anal Chem* 1990;62:2565–2573. [PubMed: 2288413]
- Soga T, Ohashi Y, Ueno Y, Naraoka H, Tomita M, Nishioka T. Quantitative metabolome analysis using capillary electrophoresis mass spectrometry. *J Proteome Res* 2003;2:488–494. [PubMed: 14582645]
- Stashenko EE, Jaramillo BE, Martínez JR. Analysis of volatile secondary metabolites from Colombian *Xylopiia aromatica* (Lamarck) by different extraction and headspace methods and gas chromatography. *J Chromatogr A* 2004a;1025:105–113. [PubMed: 14753677]

- Stashenko EE, Jaramillo BE, Martinez JR. Comparison of different extraction methods for the analysis of volatile secondary metabolites of *Lippia alba* (Mill.) N.E. Brown, grown in Colombia, and evaluation of its in vitro antioxidant activity. *J Chromatogr A* 2004b;1025:93–103. [PubMed: 14753676]
- Stein SE. An integrated method for spectrum extraction and compound identification from gas chromatography/mass spectrometry data. *J Am Soc Mass Spectrom* 1999;10:770–781.
- Theobald U, Mailinger W, Reuss M, Rizzi M. In vivo analysis of glucose-induced fast changes in yeast adenine nucleotide pool applying a rapid sampling technique. *Anal Biochem* 1993;214:31–37. [PubMed: 8250239]
- Tolstikov VV, Fiehn O. Analysis of highly polar compounds of plant origin: Combination of hydrophilic interaction chromatography and electrospray ion trap mass spectrometry. *Anal Biochem* 2002;301:298–307. [PubMed: 11814300]
- Tolstikov VV, Lommen A, Nakanishi K, Tanaka N, Fiehn O. Monolithic silica-based capillary reversed-phase liquid chromatography/electrospray mass spectrometry for plant metabolomics. *Anal Chem* 2003;75:6737–6740. [PubMed: 14640754]
- Vaidyanathan S, Kell DB, Goodacre R. Flow-injection electrospray ionization mass spectrometry of crude cell extracts for high-throughput bacterial identification. *J Am Soc Mass Spectrom* 2002;13:118–128. [PubMed: 11838015]
- van Winden WA, van Dam JC, Ras C, Kleijn RJ, Vinke JL, van Gulik WM, Heijnen JJ. Metabolic-flux analysis of *Saccharomyces cerevisiae* CEN.PK113-7D based on mass isotopomer measurements of ¹³C-labeled primary metabolites. *FEMS Yeast Res* 2005;5:559–568. [PubMed: 15780655]
- Verdonk JC, Ric de Vos CH, Verhoeven HA, Haring MA, van Tunen AJ, Schuurink RC. Regulation of floral scent production in petunia revealed by targeted metabolomics. *Phytochemistry* 2003;62:997–1008. [PubMed: 12590126]
- Viant MR. Improved methods for the acquisition and interpretation of NMR metabolomic data. *Biochem Biophys Res Commun* 2003;310:943–948. [PubMed: 14550295]
- Vigneau-Callahan KE, Shestopalov AI, Milbury PE, Matson WR, Kristal BS. Characterization of diet-dependent metabolic serotypes: Analytical and biological variability issues in rats. *J Nutr* 2001;131:924S–932S. [PubMed: 11238787]
- Villas-Boas SG, Hojer-Pedersen J, Akesson M, Smedsgaard J, Nielsen J. Global metabolite analysis of yeast: Evaluation of sample preparation methods. *Yeast* 2005;22:1155–1169. [PubMed: 16240456]
- Wahl HG, Hoffmann A, Luft D, Liebich HM. Analysis of volatile organic compounds in human urine by headspace gas chromatography-mass spectrometry with a multipurpose sampler. *J Chromatogr A* 1999;847:117–125. [PubMed: 10431355]
- Wang W, Zhou H, Lin H, Roy S, Shaler TA, Hill LR, Norton S, Kumar P, Anderle M, Becker CH. Quantification of proteins and metabolites by mass spectrometry without isotopic labeling or spiked standards. *Anal Chem* 2003;75:4818–4826. [PubMed: 14674459]
- Watkins SM, German JB. Toward the implementation of metabolomic assessments of human health and nutrition. *Curr Opin Biotechnol* 2002;13:512–516. [PubMed: 12459346]
- Watkins SM, Reifsnnyder PR, Pan HJ, German JB, Leiter EH. Lipid metabolome-wide effects of the PPARgamma agonist rosiglitazone. *J Lipid Res* 2002;43:1809–1817. [PubMed: 12401879]
- Weckwerth W. Metabolomics in systems biology. *Annu Rev Plant Biol* 2003;54:669–689. [PubMed: 14503007]
- Weckwerth W, Wenzel K, Fiehn O. Process for the integrated extraction, identification and quantification of metabolites, proteins and RNA to reveal their co-regulation in biochemical networks. *Proteomics* 2004;4:78–83. [PubMed: 14730673]
- White IN, Brown K. Techniques: The application of accelerator mass spectrometry to pharmacology and toxicology. *Trends Pharmacol Sci* 2004;25:442–447. [PubMed: 15276714]
- Williams RE, Major H, Lock EA, Lenz EM, Wilson ID. D-Serine-induced nephrotoxicity: A HPLC-TOF/MS-based metabolomics approach. *Toxicology* 2005;207:179–190. [PubMed: 15596249]
- Wilson ID, Nicholson JK, Castro-Perez J, Granger JH, Johnson KA, Smith BW, Plumb RS. High resolution “ultra performance” liquid chromatography coupled to oa-TOF mass spectrometry as a tool for differential metabolic pathway profiling in functional genomic studies. *J Proteome Res* 2005a;4:591–598. [PubMed: 15822939]

- Wilson ID, Plumb R, Granger J, Major H, Williams R, Lenz EM. HPLC-MS-based methods for the study of metabolomics. *J Chromatogr B Analyt Technol Biomed Life Sci* 2005b;817:67–76.
- Windig W, Phalp JM, Payne AW. A noise and background reduction method for component detection in liquid chromatography mass spectrometry. *Anal Chem* 1996;68:3602–3606.
- Wittmann C. Metabolic flux analysis using mass spectrometry. *Adv Biochem Eng Biotechnol* 2002;74:39–64. [PubMed: 11991183]
- Wittmann C, Heinzle E. Application of MALDI-TOF MS to lysine-producing *Corynebacterium glutamicum*: A novel approach for metabolic flux analysis. *Eur J Biochem* 2001;268:2441–2455. [PubMed: 11298764]
- Wittmann C, Kromer JO, Kiefer P, Binz T, Heinzle E. Impact of the cold shock phenomenon on quantification of intracellular metabolites in bacteria. *Anal Biochem* 2004;327:135–139. [PubMed: 15033521]
- Zytkovicz TH, Fitzgerald EF, Marsden D, Larson CA, Shih VE, Johnson DM, Strauss AW, Comeau AM, Eaton RB, Grady GF. Tandem mass spectrometric analysis for amino, organic, and fatty acid disorders in newborn dried blood spots: A two-year summary from the New England Newborn Screening Program. *Clin Chem* 2001;47:1945–1955. [PubMed: 11673361]

Biographies

Katja Dettmer, Ph.D., studied Chemistry at the University of Leipzig and received her Ph.D. in 2001 from the Institute of Analytical Chemistry, University of Leipzig. She spent 4 years as a postdoctoral fellow and later as an assistant research scientist at the University of California, Davis, USA, where she started to work in the field of metabolomics. Currently, she is a research scientist at the Institute of Functional Genomics at the University of Regensburg, Germany. She is conducting research in the field of metabolomics, focusing on the development of mass spectrometry-based tools for metabolic fingerprinting in biofluids as well as metabolic-profiling methods.

Pavel A. Aronov studied Environmental Chemistry at the Novosibirsk State University and graduated with honors in 2001. His interest in Analytical and Biological Chemistry brought him to the laboratory of Bruce Hammock at the University of California, Davis, where he started to work in the field of metabolomics. Currently he works on his Ph.D. focused on metabolic fingerprinting, metabolic profiling of biologically active lipids, and chemical derivatization for electrospray ionization.

Bruce D. Hammock received his Ph.D. in Entomology and Toxicology at U.C. Berkeley in 1973 and has subsequently taught at Northwestern University, UC Riverside and since 1980 at UC Davis where he is a distinguished professor in Entomology and the Cancer Research Center. He has been PI of a NIEHS Superfund Basic Research Program for over 20 years, is the director of a NIH Biotechnology Training Grant and is a member of the US National Academy of Sciences. The UCD-NIEHS analytical laboratory that he directs has had a metabolomic emphasis for 10 years. His laboratory works in aspects of analytical chemistry, biosensor development, mammalian pharmacology, and agricultural-medical pest control.




FIGURE 1.

The “Omic” cascade comprises complex datasets that as an entity comprehensively describe the response of biological systems to disease, genetic, and environmental perturbations. The most powerful database will integrate data from all omic levels. However, of these databases the metabolome is the most predictive of phenotype.




FIGURE 2.

Bibliographic search in Chemical Abstracts Plus containing the keywords metabolomics and metabonomics using SciFinder Scholar (as of June 10th, 2005). A total of 696 journal articles were found. The dataset was further mined using the search parameter mass spectrometry or NMR. The diagram shows the frequency of total metabolomics publications (black squares), publications that mention mass spectrometry (black bars) and NMR (gray bars) from 1999 to 2004.




FIGURE 3.
Strategies for metabolomic investigations.




FIGURE 4. Simplified workflow for a metabolic fingerprinting analysis.





FIGURE 5. Product ion spectra of dipentadecyl-phosphatidylcholine using ESI in positive $[M + H]^+$ (upper spectrum) and negative mode $[M + HCOO]^-$ (lower spectrum).

**FIGURE 6.**

Combined MS spectra from 2.5 to 35 min obtained after online-SPE extraction of 10 μ L human urine, reversed phase HPLC separation, and TOF-MS detection (LCT, Micromass, UK) with ESI in positive mode. Expanded section of the spectrum illustrates the need for high-mass resolution to distinguish isobaric peaks, which originate from different metabolites as shown by the extracted-ion chromatogram. The third extracted-ion chromatogram shows that m/z 202.129 originates from two different metabolites that can be chromatographically separated. Direct infusion would not readily distinguish these two metabolites.




FIGURE 7.

Online SPE extraction of (A) 10 µL saline solution and (B) 10 µL rat urine spiked with internal standards coupled to direct infusion analysis using ESI in positive mode and TOF-MS detection (LCT, Micromass). Internal standards used: 1. Indole-3-acetic acid-d₂, 2. 11-Aminoundecanoic acid, 3. Melatonin-d₄. Rat urine spectrum (B) shows significant ion suppression for the internal standards in the range of 60%.




FIGURE 8.

Simultaneously acquired total-ion chromatogram with ESI in positive (upper chromatogram) and negative (lower chromatogram) mode obtained by online SPE extraction of 10 μ L human urine, RP-chromatography using gradient elution and TOF-MS detection (LCT, Micromass).




FIGURE 9. Total ion chromatogram of a human urine (1 mL) extract after silylation with MSTFA and GC-MS analysis. Peak deconvolution using AMDIS resulted in 1,582 components.


**FIGURE 10.**

Diagram showing different approaches for metabolomics data analysis. **A:** Mass-intensity vectors in each scan of the raw LC-MS or GC-MS data are binned to form mass-aligned vectors of uniform length. **B:** Scans across chromatogram area of interest are combined. **C:** Components are detected in binned data by peak detection algorithms in all scans belonging to one bin (selected ion chromatograms) or several bins (peak deconvolution). **D:** Peaks are detected in raw data using peak picking algorithms.




FIGURE 11.

A histogram of decimal places of metabolite accurate masses which demonstrates the optimum bin margins. Accurate masses of 1,008 metabolites were used which were discovered by MarkerLynx (MicroMass, UK) in a typical LC-MS fingerprint of cell culture media samples.

TABLE 1

Metabolomics-related definitions

Metabolite	Small molecules that participate in general metabolic reactions and that are required for the maintenance, growth and normal function of a cell [*] .
Metabolome	The complete set of metabolites in an organism.
Metabolomics	Identification and quantification of all metabolites in a biological system.
Metabolic profiling	Quantitative analysis of set of metabolites in a selected biochemical pathway or a specific class of compounds. This includes target analysis, the analysis of a very limited number of metabolites, e.g. single analytes as precursors or products of biochemical reactions.
Metabolic fingerprinting	Unbiased, global screening approach to classify samples based on metabolite patterns or "fingerprints" that change in response to disease, environmental or genetic perturbations with the ultimate goal to identify discriminating metabolites.
Metabolic footprinting	Fingerprinting analysis of extra-cellular metabolites in cell culture medium as a reflection of metabolite excretion or uptake by cells.

^{*} This is in contrast to xenobiotic or foreign compound metabolites which may overshadow natural metabolites in an analytical procedure. However, they can be very valuable in evaluating the physiological status of the organism.

TABLE 2

Common derivatization reactions used for GC-MS based metabolic fingerprinting investigations

Derivatization reagent	Product	Reagent leaving group
Silylation	Example: Organic acid	
	$\begin{array}{c} \text{O} \\ \parallel \\ \text{R}-\text{C}-\text{OH} \end{array}$	
MSTFA N-Methyl-N-trimethylsilyl-trifluoroacetamide	Trimethylsilylester	N-Methyl trifluoroacetamide
$\begin{array}{c} \text{O} \\ \parallel \\ \text{F}_3\text{C}-\text{C}-\text{N}-\text{Si}(\text{CH}_3)_3 \\ \\ \text{CH}_3 \end{array}$	$\begin{array}{c} \text{O} \\ \parallel \\ \text{R}-\text{C}-\text{O}-\text{Si}(\text{CH}_3)_3 \end{array}$	$\begin{array}{c} \text{O} \\ \parallel \\ \text{F}_3\text{C}-\text{C}-\text{NH} \\ \\ \text{CH}_3 \end{array}$
MTBSTFA N-Methyl-N- <i>tert</i> -butyl-dimethylsilyl-trifluoroacetamide	<i>tert</i> -Butyldimethylsilylester	N-Methyl trifluoroacetamide
$\begin{array}{c} \text{O} \\ \parallel \\ \text{F}_3\text{C}-\text{C}-\text{N}-\text{Si}(\text{CH}_3)_2-\text{C}(\text{CH}_3)_3 \\ \quad \\ \text{CH}_3 \quad \text{CH}_3 \end{array}$	$\begin{array}{c} \text{CH}_3 \\ \\ \text{R}-\text{O}-\text{Si}-\text{C}(\text{CH}_3)_3 \\ \\ \text{CH}_3 \end{array}$	$\begin{array}{c} \text{O} \\ \parallel \\ \text{F}_3\text{C}-\text{C}-\text{NH} \\ \\ \text{CH}_3 \end{array}$
BSTFA N,O-Bis(trimethylsilyl)-trifluoroacetamide	Trimethylsilylester	N-trimethylsilyltrifluoro-acetamide
$\begin{array}{c} \text{O}-\text{Si}(\text{CH}_3)_3 \\ \\ \text{F}_3\text{C}-\text{C}=\text{N}-\text{Si}(\text{CH}_3)_3 \end{array}$	$\begin{array}{c} \text{O} \\ \parallel \\ \text{R}-\text{C}-\text{O}-\text{Si}(\text{CH}_3)_3 \end{array}$	$\begin{array}{c} \text{O} \\ \parallel \\ \text{F}_3\text{C}-\text{C}-\text{N}-\text{Si}(\text{CH}_3)_3 \\ \\ \text{H} \\ \parallel \\ \text{O} \\ \parallel \\ \text{F}_3\text{C}-\text{C}-\text{NH}_2 \\ \text{Trifluoroacetamide} \\ \parallel \\ \text{O} \\ \parallel \\ \text{F}_3\text{C}-\text{C}-\text{NH}_2 \end{array}$
Oximation	Example: Aldehyde	
	$\begin{array}{c} \text{O} \\ \parallel \\ \text{R}_1-\text{C}-\text{R}_2 \end{array}$	
O-Methylhydroxylamine hydrochloride H ₃ C—O—NH ₂ *HCl	Methoxime	HCL, H ₂ O
	$\begin{array}{c} \text{O}-\text{CH}_3 \\ \\ \text{N} \\ \parallel \\ \text{R}_1-\text{C}-\text{R}_2 \end{array}$	