

## Erratum: “Technical note: Heterogeneity dose calculation accuracy in IMRT: Study of five commercial treatment planning systems using an anthropomorphic thorax phantom” [Med. Phys. 35, 5434–5439 (2008)]

Scott E. Davidson<sup>a)</sup>

*Department of Radiation Physics, The University of Texas M. D. Anderson Cancer Center,  
Houston, Texas 77030*

Richard A. Popple

*Department of Radiation Oncology, The University of Alabama-Birmingham,  
Birmingham, Alabama 35233*

Geoffrey S. Ibbott and David S. Followill

*Department of Radiation Physics, The University of Texas M. D. Anderson Cancer Center,  
Houston, Texas 77030*

(Received 31 March 2009; revised 8 April 2009; accepted for publication 9 April 2009;  
published 28 May 2009)

[DOI: [10.1118/1.3129160](https://doi.org/10.1118/1.3129160)]

The authors report an error in the published paper.<sup>1</sup> The variation in dose between the dose calculations and measurement for the low-dose critical structures was not reported correctly. The correct range for the low-dose critical structures is an underestimation of 14.9% to an overestimation of 11.2%. While this is reflected correctly in Table I of the aforementioned publication, it was not stated correctly in

Sec. III on p. 5437 and in Sec. IV on p. 5438.

<sup>a)</sup>Author to whom correspondence should be addressed. Electronic mail: [sedavids@mdanderson.org](mailto:sedavids@mdanderson.org)

<sup>1</sup>S. E. Davidson, R. A. Popple, G. S. Ibbott, and D. S. Followill, “Technical note: Heterogeneity dose calculation accuracy in IMRT: Study of five commercial treatment planning systems using an anthropomorphic thorax phantom,” *Med. Phys.* **35**(12), 5434–5439 (2008).