

Ultrasonics in endodontic surgery: a review of the literature

Gianfranco De Paolis, DDS, PhD

Valentina Vincenti, DDS

Matteo Prencipe, DDS

Valerio Milana, DDS

Gianluca Plotino, DDS, PhD

Dentists, private practice in Rome, Italy

Corresponding author:

Dott. Gianluca Plotino

Via Eleonora, Duse 22

00197 Rome, Italy

Tel.: 3396910098

E-mail: endo@gianlucaplotino.com

Summary

Ultrasonics in endodontic surgery: a review of the literature.

Currently, although ultrasonics (US) is used in dentistry for therapeutic and diagnostic applications as well as for cleaning of instruments before sterilization, its main use is for scaling and root planing of teeth and in root canal therapy, both for orthograde and retrograde therapy. Both in conventional and surgical treatments, US in endodontics has enhanced quality of clinical procedures and represents an important adjunct in the treatment of difficult cases. More precisely it has become increasingly more useful in applications such as gaining access to canal openings, cleaning and shaping, obturation of root canals, removal of intracanal materials and obstructions, and endodontic surgery. This review of the literature aims at presenting the numerous advantages of US in surgical endodontics and emphasizes its application in a modern-day endodontic practice.

Key words: endodontics, surgery, ultrasonics.

Introduction

In order to understand the basic concepts of the use of ultrasonics (US) in dentistry, it must be underlined that ultrasound is sound energy with a frequency above the range of human hearing, which is 20 kHz. In dentistry the range of frequencies employed in the first ultrasonic units was between 25 and 40 kHz (1). Later, low-frequency ultrasonic handpieces operating from 1 to 8 kHz were developed (2,3). This low-frequency devices were found to produce lower shear stresses (4), thus causing less alteration to the tooth surface (5). Currently, there are two basic methods of producing ultrasound (6,7,8). The first is magnetostric-

tion, which converts electromagnetic energy into mechanical energy. A stack of magnetostrictive metal strips in a handpiece is subjected to a standing and alternating magnetic field, as a result of which vibrations are produced. The second method is based on the piezoelectric principle, in which a crystal is used that changes dimension when an electrical charge is applied. Deformation of this crystal is converted into mechanical oscillation without producing heat (1).

In the last decade piezoelectric units have become the most common ultrasonic devices in dentistry. They have some advantages compared with earlier magnetostrictive units because they offer more cycles per second, 40 versus 24 kHz. The tips of these units work in a linear, back-and-forth, "piston-like" motion, which is ideal for endodontics. Lea et al. (9) demonstrated that the position of nodes and antinodes of an unconstrained and unloaded endosonic file activated by a 30-kHz piezon generator was along the file length. As a result the file vibration displacement amplitude does not increase linearly with increasing generator power. This applies in particular when "troughing" for hidden canals or when removing posts and separated instruments. In addition, this motion is ideal in surgical endodontics when creating a preparation for a retrograde filling. A magnetostrictive unit, on the other hand, creates more of a figure eight (elliptical) motion, which is not ideal for either surgical or nonsurgical endodontic use. In endodontic surgery, for example, this characteristic does not allow a precise cut of a cavity. The magnetostrictive units also have the disadvantage that the stack generates heat, thus requiring adequate cooling (1). Once again, this overheating is not desirable in surgical endodontics.

In dentistry ultrasonics (US) or ultrasonic instrumentation was first introduced to for cavity preparations (10,11), using an abrasive slurry. Although the technique received favorable comments (12,13), it never became popular, because it had to compete with the much more effective and convenient instruments, i.e. the burs mounted on high-speed handpieces (14). However, a different application was introduced in 1955, when Zinner (15) reported on the use of an ultrasonic instrument to remove deposits from the tooth surface. This was improved upon by Johnson and Wilson (16), and the ultrasonic scaler became an established tool in the removal of dental calculus and plaque. Currently, although US is used in dentistry for therapeutic and diagnostic applications as well as for cleaning of instruments before sterilization, its main use is for scaling and root planing of teeth and in root canal therapy, both for surgical and non-surgical approach. (1,17,18). More recently, the concept of minimally invasive dentistry (19,20) and the desire for preparations with small dimensions has stimulated new approaches in cavity design and tooth-cutting concepts, including ultrasound for cavity preparation (21).

The concept of using US in endodontics was first introduced

by Richman (22) in 1957. However, it was not until Martin et al. (23,24,25) demonstrated the ability of ultrasonically activated K-type files to cut dentin that this application found common use in the preparation of root canals before filling and also obturation. The term *endosonics* was coined by Martin and Cunningham (26,27) and was defined as the ultrasonic and synergistic system of root canal instrumentation and disinfection. The most frequent applications of US in endodontics are the following. (i) Access refinement, finding orifices and calcified canals, and removal of attached pulp stones. (ii) Removal of intracanal obstructions (separated instruments, root canal posts, silver points, and fractured metallic posts). (iii) Root canal preparation using ultrasonically activated k-files. (iv) Root canal irrigation with an increased action of irrigating solutions, due to cavitation and microstreaming action. (v) Ultrasonic condensation of gutta-percha. (vi) Placement of calcium hydroxide and mineral trioxide aggregate (MTA). (vii) Surgical endodontics: root-end cavity preparation and refinement and placement of root-end obturation material

The aim of the present literature review is to analyze and discuss the use of US in surgical endodontics, focusing attention mainly of the advantages of US-activated instruments for cavity preparation and refinement. Moreover the use of US for the placement of root-end filling materials will also be addressed.

The use of ultrasonics in Surgical Endodontics

In the past decades root-end cavities have traditionally been prepared by means of small round or inverted cone burs in a micro-handpiece. In the mid-1980s, standardized instruments and aluminum oxide ceramic pins were introduced for retrograde filling (28), but that system could not be used in cases with limited working space or in teeth with large oval canals. Soon after sonically or ultrasonically driven microsurgical retrotips became commercially available in the early 1990s (29,30,31,32), this new technique of retrograde root canal instrumentation has become rapidly popular and been established as an essential adjunct in periradicular surgery (33,34). However, the cutting properties of the retrotips at that time were limited and seemed to be dependent on loading, power setting, and orientation of the tip to the long axis of the handpiece (35,36). Moreover in some retrotips, cooling of the working tip was insufficient, and dentin and bone were at risk of being overheated. However, the technique was promising and led to significant improvement in the instruments, which have significantly enhanced the treatment outcome in apicoectomy with retrofilling (37). As the prognosis of endodontic surgery is highly dependent on good obturation and sealing of the root canal, an optimal cavity preparation is an essential prerequisite for an adequate root-end filling after apicoectomy (38,39). The first root-end preparation using modified ultrasonic inserts following an apicoectomy is attributed to Bertrand et al. (40). Others followed, but it was not until 1987 that Flath and Hicks (41) further reported on the use of ultrasonics and sonics for root-end cavity preparation.

One of the reasons for the success of US retrotips is due to the fact that conventional root-end cavity preparation using rotary burs in a micro-handpiece is faced with several problems (42,43), such as a cavity preparation not be-

ing parallel to the canal, difficult access to the root end, and risk of lingual perforation of the root. Furthermore, the inability to prepare to a sufficient depth, thus compromising retention of the root-end filling material, means that the root-end resection procedure requires a longer cutting bevel, thus exposing more dentinal tubules and isthmus tissue, of which the latter is difficult to remove. The development of ultrasonic and sonic retrotips has revolutionized root-end therapy, improving the surgical procedure with better access to the root end, resulting in better canal preparation (44,45). Ultrasonic retrotips come in a variety of shapes and angles, thus improving some steps during the surgical procedures (46,47).

Probably, the most relevant clinical advantages are the enhanced access to root-ends in a limited working space. This leads to a smaller osteotomy for surgical access because of the advantage of using various angulations and the small size of the retrotips (48). However, a number of studies compared root-end preparations made with microsurgical tips to those made with burs. They demonstrated additional advantages of this technique, such as deeper and more conservative cavities that follow the original path of the root canal more closely (49,50,51,52,53,54). A better-centered root-end preparation also lessens the risk of lateral perforation. Furthermore, the geometry of the retrotip design does not require a beveled root-end resection for surgical access, thus decreasing the number of exposed dentinal tubules (55,56,57) and minimizing apical leakage (58,59,60,61). They also enable the removal of isthmus tissue present between two canals within the same root (62,63,64). It is considered a time-saving technique that seems to have a lower failure rate.

More recently, the cleaning effect and the cutting ability of ultrasonic retrotips have been described as satisfactory by many authors (65,66). Furthermore, US produced less smear layer in a retro-end cavity compared to a slow-speed handpiece (67). Moreover, the refinement of cavity margins that were obtained with the ultrasonic tips may positively affect the delivery of materials into the cavities and enhance their seal (68,69,70), even if cavities prepared with erbium:YAG lasers have been shown to produce significantly lower microleakage than ultrasonic preparations (71). Amongst the possible iatrogenic errors, in a study by Walsley et al. (72) the breakage of ultrasonic root-end preparation tips was investigated and attributed to the design of the tip. Increased angulation of retrotips increases the transverse oscillation and decreases the longitudinal oscillation, putting the greatest strain at the bend of the instrument. The authors suggested reducing the angulation and increasing the dimensions of the tip to resist breakage. This may be true, but a straighter design will restrict access and a thicker instrument prevents instrumentation of isthmuses. A controversial issue with sonic or ultrasonic root-end preparation is the formation of cracks or microfractures and its implications for healing success (73,74). Some studies indicated that this was a possible drawback (75,76,77,78,79). Other studies, however, disputed these findings and did not report a higher prevalence of microfractures (80,81,82,83,84,85,86,87,88.). Khabbaz et al. (53) found that cracks did not correlate directly with the surface area of the root-end surfaces but rather with the type of retrotip used. Preparation with smooth stainless steel ultrasonic tips produced fewer intradentin cracks than diamond-coated stainless steel ultrasonic tips and sonic diamond-coated tips. It is recommended that the ultrason-

ic unit be set at medium power and the cavities be prepared to a depth of 2.5-3 mm (89). This depth allows for a minimum thickness of material that can still provide an effective apical seal (90). The cavity walls should be parallel and follow the anatomic outline of the pulpal space (91). It has also been suggested that root-end cavities should be initiated with a diamond-coated retrotip, using its better cutting ability to provide the main cavity. This aids in the removal of root canal obturation materials and should be followed by a smooth retrotip to smooth and clean cavity walls. It must be underlined the fact that the influence of root-end microfractures on the periradicular healing process and apical leakage should be clarified. Apical resorption after healing (92) may eliminate the surface defects and contribute to the overall success of treatment. Also, such defects can be removed by finishing resected and retrofilled root-end surfaces (93). Several *in vivo* studies reported excellent success rates when the root-end preparation was performed using ultrasonic retrotips (94,95), thus demonstrating that modern surgical endodontic treatment using an operating microscope and ultrasonic tips significantly improves the outcome compared to the traditional techniques (96). In order to improve retrograde filling techniques, it has been suggested the use of condenser tip ultrasonically activated for placement of retrograde filling materials, as the ultrasonic vibration is meant to improve the flow, settling and compaction of these materials to root-end dentinal walls. This should improve the delivery of materials into the cavity thus enhancing their seal. Ultrasonic tips can also be used to polish root end material and apical surfaces. Utilizing specific ultrasonic tips for refinement of the external radicular surface may be beneficial in the elimination of extraradicular bacteria, which may be responsible for infection (97).

More recently, Witherspoon and Ham (98) described the use of US to aid in the placement of MTA. The inherent irregularities and divergent nature of some open apices may predispose the material to marginal gaps at the dentin interface. It was demonstrated that, with the adjunct of US, a significantly better seal with MTA was achieved. Placement of MTA with ultrasonic vibration and an endodontic condenser improved the flow, settling, and compaction of MTA. Furthermore, the ultrasonically condensed MTA appeared denser radiographically, with fewer voids. These results contradicted those of Aminoshariae et al. (99), who concluded from an *in vitro* study that hand condensation was superior. The recommended placement method consists of selecting a condenser tip, then picking up and placing the MTA with the ultrasonic tip, followed by activating the tip and slowly moving the MTA material down using a 1- to 2-mm vertical packing motion. Direct ultrasonic energy will vibrate and generate a wavelike motion, which facilitates moving and adapting the cement to the canal walls. In a case of repairing a defect apical to the canal curvature, Ruddle (100) recommends incrementally placing MTA deep into a canal, then shepherding it around the curvature with a flexible trimmed gutta-percha cone utilized as a plugger. A precurved 15 or 20 stainless steel file is then inserted into the material and placed to within 1 or 2 mm of the working length. This is followed by indirect ultrasound, which involves placing the working end of an ultrasonic instrument on the shaft of the file. This vibratory energy encourages MTA to move and conform to the configurations of the canal laterally as well as controlling its movement. This technique was recommended initially for placing MTA

in open and diverging apices, but it can also be used to put the material in root-end cavities, in perforations, and especially in perforations of the floor of the pulp chamber.

References

1. Stock CJR. Current status of the use of ultrasound in endodontics. *Int Dent J.* 1991;41:175-182.
2. Ahmad M, Roy RA, Kamarudin AG, Safar M. The vibratory pattern of ultrasonic files driven piezoelectrically. *Int Endod J.* 1993;26:120-124.
3. Lumley PJ, Walmsley AD, Marquis PM. Effect of air inlet ring opening on sonic handpiece performance. *J Dent.* 1994;22:376-379.
4. Ahmad M, Pitt Ford TR, Crum LA. Ultrasonic debridement of root canals: an insight into the mechanisms involved. *J Endod.* 1987;13:93-101.
5. Layton CA, Marshall JG, Morgan LA, Baumgartner JC. Evaluation of cracks associated with ultrasonic root-end preparation. *J Endod.* 1996;22:157-160.
6. Walmsley AD. Ultrasound and root canal treatment: the need for scientific evaluation. *Int Endod J.* 1987;20:105-111.
7. Lumley PJ, Walmsley AD, Laird WRE. An investigation into the occurrence of cavitation activity during endosonic instrumentation. *J Dent.* 1988;16:120-122.
8. Laird WRE, Walmsley AD. Ultrasound in dentistry: Part 1 (Biophysical interactions). *J Dent.* 1991;19:14-17.
9. Lea SC, Walmsley AD, Lumley PJ, Landini G. A new insight into the oscillation characteristics of endosonic files used in dentistry. *Phys Med Biol.* 2004;49:2095-2102.
10. Catuna MC. Ultrasonic energy: a possible dental application (Preliminary report of an ultrasonic cutting method). *Ann Dent.* 1953;12:256-260.
11. Postle HH. Ultrasonic cavity preparation. *J Prosthet Dent.* 1958;8:153-160.
12. Oman CR, Applebaum E. Ultrasonic cavity preparation II (Progress report). *J Am Dent Assoc.* 1954;50:414-417.
13. Nielsen AG, Richards JR, Wolcott RB. Ultrasonic dental cutting instrument: I. *J Am Dent Assoc.* 1955;50:392-399.
14. Street EV. Critical evaluation of ultrasonics in dentistry. *J Prosthet Dent.* 1959;9:32-41.
15. Zinner DD. Recent ultrasonic dental studies including periodontia, without the use of an abrasive. *J Dent Res.* 1955;34:748-749.
16. Johnson WN, Wilson JR. Application of the ultrasonic dental unit to scaling procedures. *J Periodontol.* 1957;28:264-271.
17. Walmsley AD. Applications of ultrasound in dentistry. *Ultrasound Med Biol.* 1988;14:7-14.
18. Walmsley AD, Laird WRE, Lumley PJ. Ultrasound in dentistry: Part 2 (Periodontology and endodontics). *J Dent.* 1992;20:11-17.
19. Peters MC, McLean ME. Minimally invasive operative care (I. Minimal intervention and concepts for minimally invasive cavity preparations). *J Adhes Dent.* 2001;3:7-16.
20. Peters MC, McLean ME. Minimally invasive operative care (II. Contemporary techniques and materials: an overview). *J Adhes Dent.* 2001;3:17-31.
21. Sheets CG, Paquette JM. Ultrasonic tips for conservative restorative dentistry. *Dent Today.* 2002;21:102-104.
22. Richman RJ. The use of ultrasonics in root canal therapy and root resection. *Med Dent J.* 1957;12:12-18.
23. Martin H. Ultrasonic disinfection of the root canal. *Oral Surg Oral Med Oral Pathol.* 1976;42:92-99.
24. Martin H, Cunningham WT, Norris JP, Cotton WR. Ultrasonic versus hand filing of dentin: a quantitative study. *Oral Surg Oral Med Oral Pathol.* 1980;49:79-81.
25. Martin H, Cunningham WT, Norris JP. A quantitative com-

- parison of the ability of diamond and K-type files to remove dentin. *Oral Surg Oral Med Oral Pathol.* 1980;50:566-568.
- 26. Martin H, Cunningham W. Endosonic endodontics: the ultrasonic synergistic system. *Int Dent J.* 1984;34:198-203.
 - 27. Martin H, Cunningham W. Endosonics: the ultrasonic synergistic system of endodontics. *Endod Dent Traumatol.* 1985;1:201-206.
 - 28. Keller U. Aluminium oxide ceramic pins for retrograde root filling: experiences with a new system. *Oral Surg Oral Med Oral Pathol.* 1990;69:737-742.
 - 29. Pannkuk TF. Endodontic surgery: principles, objectives and treatment of posterior teeth: Part I. *Endod Rep.* 1991;6:8-14.
 - 30. Carr G. Advanced techniques and visual enhancement for endodontic surgery. *Endod Rep.* 1992;7:6-9.
 - 31. Pannkuk TF. Endodontic surgery: the treatment phase and wound healing: Part II. *Endod Rep.* 1992;7:14-19.
 - 32. Fong CD. A sonic instrument for retrograde preparation. *J Endod.* 1993;19:374-375.
 - 33. Gutmann JL, Pitt Ford TR. Management of the resected root-end: a clinical review. *Int Endod J.* 1993;26:273-283.
 - 34. Wuchenich G, Meadows D, Torabinejad M. A comparison between two root end preparation techniques in human cadavers. *J Endod.* 1994;20:279-282.
 - 35. Waplington M, Lumley PJ, Walmsley AD, Blunt L. Cutting ability of an ultrasonic retrograde cavity preparation instrument. *Endod Dent Traumatol.* 1995;11:177-180.
 - 36. Devall R, Lumley P, Waplington M, Blunt L. Cutting characteristics of a sonic root-end preparation instrument. *Endod Dent Traumatol.* 1996;12:96-99.
 - 37. Sumi Y, Hattori H, Hayashi K, Ueda M. Ultrasonic root-end preparation: clinical and radiographic evaluation of results. *J Oral Maxillofac Surg.* 1996;54:590-593.
 - 38. Carr G. Surgical endodontics. In: Cohen S, Burns R editor. *Pathways of the pulp.* 4th ed.. St. Louis: Mosby; 1994;p. 546-552.
 - 39. Sutimuntanakul S, Worayoskowit W, Mangkornkarn C. Retrograde seal in ultrasonically prepared canals. *J Endod.* 2000;26:444-446.
 - 40. Bertrand G, Festal F, Barailly R. Use of ultrasound in apicoectomy. *Quintessence Int.* 1976;7:9-12.
 - 41. Flath RK, Hicks ML. Retrograde instrumentation and obturation with new devices. *J Endod.* 1987;13:546-549.
 - 42. von Arx T, Walker WA. Microsurgical instruments for root-end cavity preparation following apicoectomy: a literature review. *Endod Dent Traumatol.* 2000;16:47-62.
 - 43. Kim S, Kratchman S. Modern endodontic surgery concepts and practice: a review. *J Endod.* 2006;32:601-623.
 - 44. Gutmann JL, Harrison JW. Posterior endodontic surgery: anatomical considerations and clinical techniques. *Int Endod J.* 1985;18:8-34.
 - 45. Kim S. Principles of endodontic microsurgery. *Dent Clin North Am.* 1997;41:481-497.
 - 46. Kellert M, Solomon C, Chalfin H. A modern approach to surgical endodontics: ultrasonic apical preparation. *N Y State Dent J.* 1994;60:25-28.
 - 47. Rubinstein R, Torabinejad M. Contemporary endodontic surgery. *J Calif Dent Assoc.* 2004;32:485-492.
 - 48. Mehlhoff DS, Marshall JG, Baumgartner JC. Comparison of ultrasonic and high-speed-bur root-end preparations using bilaterally matched teeth. *J Endod.* 1997;23:448-452.
 - 49. Gutmann JL, Saunders WP, Nguyen L, Guo IY, Saunders EM. Ultrasonic root-end preparation: Part 1 (SEM analysis). *Int Endod J.* 1994;27:318-324.
 - 50. Engel TK, Steiman HR. Preliminary investigation of ultrasonic root end preparation. *J Endod.* 1995;21:443-445.
 - 51. Gorman MC, Steiman HR, Gartner AH. Scanning electron microscopic evaluation of root-end preparations. *J Endod.* 1995;21:113-117.
 - 52. Lin CP, Chou HG, Kuo JC, Lan WH. The quality of ultrasonic root-end preparation: a quantitative study. *J Endod.* 1998;24:666-670.
 - 53. Khabbaz MG, Kerezoudis NP, Aroni E, Tsatsas V. Evaluation of different methods for the root-end cavity preparation. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2004;98:237-242.
 - 54. Abedi HR, Van Mierlo BL, Wilder-Smith P, Torabinejad M. Effects of ultrasonic root-end cavity preparation on the root apex. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 1995;80:207-213.
 - 55. Rud J, Andreassen JO, Moller-Jensen JE. A follow-up study of 1000 cases treated by endodontic surgery. *Int J Oral Surg.* 1972;1:15-28.
 - 56. Amagasa T, Nagase M, Sato T, Shioda S. Apicoectomy with retrograde gutta-percha root filling. *Oral Surg Oral Med Oral Pathol.* 1989;68:339-342.
 - 57. Tidmarsh BG, Arrowsmith MG. Dentinal tubules at the root ends of apicected teeth: a scanning electron microscopic study. *Int Endod J.* 1989;22:184-189.
 - 58. Shani J, Friedman S, Stabholz A, Abed JA. Radionuclide model for evaluating sealability of retrograde filling materials. *Int J Nucl Med Biol.* 1984;11:46-51.
 - 59. Vertucci F, Beatty RG. Apical leakage associated with retrofilling techniques: a dye study. *J Endod.* 1986;12:331-336.
 - 60. Gilheany PA, Figidor D, Tyas MJ. Apical dentin permeability and microleakage associated with root end resection and retrograde filling. *J Endod.* 1994;20:22-26.
 - 61. Gagliani M, Taschieri S, Molinari R. Ultrasonic root-end preparation: influence of cutting angle on the apical seal. *J Endod.* 1998;24:726-730.
 - 62. Weller RN, Niemczyk SP, Kim S. Incidence and position of the canal isthmus. *J Endod.* 1995;21:380-383.
 - 63. Hsu Y-Y, Kim S. The resected root surface: the issue of canal isthmuses. *Dent Clin North Am.* 1997;41:529-540.
 - 64. Zuolo ML, Perin FR, Ferreira MOF, Faria FP. Ultrasonic root-end preparation with smooth and diamond-coated tips. *Endod Dent Traumatol.* 1999;15:265-268.
 - 65. O'Connor RP, Hutter JW, Roahen JO. Leakage of amalgam and Super-EBA root-end fillings using two preparation techniques and surgical microscopy. *J Endod.* 1995;21:74-78.
 - 66. Sultan M, Pitt-Ford TR. Ultrasonic preparation and obturation of root-end cavities. *Int Endod J.* 1995;28:231-238.
 - 67. Pashley DH. Smear layer: physiological considerations. *Oper Dent.* 1984;3(Suppl):13-29.
 - 68. Saunders WP, Saunders EM, Gutman JL. Ultrasonic root-end preparation: Part 2 (Microleakage of EBA root-end fillings). *Int Endod J.* 1994;27:325-329.
 - 69. Lloyd A, Gutmann J, Dummer P, Newcombe R. Microleakage of Diaket and amalgam in root-end cavities prepared using MicroMega sonic retro-prep tips. *Int Endod J.* 1997;30:196-204.
 - 70. Chailertvanitkul P, Saunders WP, Saunders EM, MacKenzie D. Polymicrobial coronal leakage of super EBA root-end fillings following two methods of root-end preparation. *Int Endod J.* 1998;31:348-353.
 - 71. Karlovic Z, Pezelj-Ribaric S, Miletic I, Jukic S, Grgurevic J, Anic I. Erbium:YAG laser versus ultrasonic in preparation of root-end cavities. *J Endod.* 2005;31:821-823.
 - 72. Walmsley AD, Lumley PJ, Johnson WT, Walton RE. Breakage of ultrasonic root-end preparation tips. *J Endod.* 1996;22:287-289.
 - 73. Taschieri S, Testori T, Francetti L, Del Fabbro M. Effects of ultrasonic root end preparation on resected root surfaces: SEM evaluation. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2004;98:611-618.
 - 74. Gondim E, Figueiredo Almeida de Gomes BP, Ferraz CC, Tei-

- xeira FB, de Souza-Filho FJ. Effect of sonic and ultrasonic retrograde cavity preparation on the integrity of root apices of freshly extracted human teeth: scanning electron microscopy analysis. *J Endod*. 2002;28:646-650.
75. Onnick PA, Davis RD, Wayman BE. An in vitro comparison of incomplete root fractures associated with three obturation techniques. *J Endod*. 1994;20:32-37.
76. Waplington M, Lumley PJ, Walmsley AD. Incidence of root face alteration after ultrasonic retrograde cavity preparation. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod*. 1997;83:387-392.
77. Brent PD, Morgan LA, Marshall JG, Baumgartner JC. Evaluation of diamond-coated ultrasonic instruments for root-end preparation. *J Endod*. 1999;25:672-675.
78. Lin CP, Chou HG, Chen RS, Lan WH, Hsieh CC. Root deformation during root-end preparation. *J Endod*. 1999;25:668-671.
79. Frank RJ, Antrim DD, Bakland LK. Effect of retrograde cavity preparations on root apices. *Endod Dent Traumatol*. 1996;12:100-103.
80. Beling KL, Marshall JG, Morgan LA, Baumgartner JC. Evaluation for cracks associated with ultrasonic root-end preparation of gutta-percha filled canals. *J Endod*. 1997;23:323-326.
81. Min MM, Brown CE, Legan JJ, Kafrawy AH. In vitro evaluation of effects of ultrasonic root-end preparation on resected root surfaces. *J Endod*. 1997;23:624-628.
82. Calzonetti KJ, Iwanowski T, Komorowski R, Friedman S. Ultrasonic root-end cavity preparation assessed by an in situ impression technique. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod*. 1998;85:210-215.
83. Morgan LA, Marshall JG. A scanning electron microscopic study of in vivo ultrasonic root-end preparations. *J Endod*. 1999;25:567-570.
84. Gray GJ, Hatton JF, Holtzmann DJ, Jenkins DB, Nielsen CJ. Quality of root-end preparations using ultrasonic and rotary instrumentation in cadavers. *J Endod*. 2000;26:281-283.
85. Rainwater A, Jeansonne BG, Sarkar N. Effects of ultrasonic root-end preparation on microcrack formation and leakage. *J Endod*. 2000;26:72-75.
86. Peters Cl, Peters OA, Barbakow F. An in vitro study comparing root-end cavities prepared by diamond-coated and stainless steel ultrasonic retrotips. *Int Endod J*. 2001;34:142-148.
87. Navarre SW, Steiman HR. Root-end fracture during retro-preparation: a comparison between zirconium nitride-coated and stainless steel microsurgical ultrasonic instruments. *J Endod*. 2002;28:330-332.
88. Ishikawa H, Sawada N, Kobayashi C, Suda H. Evaluation of root-end cavity preparation using ultrasonic retrotips. *Int Endod J*. 2003;36:586-590.
89. Tanzilli JP, Donald R, Moodnik RM. A comparison of the marginal adaptation of retrograde techniques: a scanning electron microscopic study. *Oral Surg Oral Med Oral Pathol*. 1980;50:74-80.
90. Mattison GD, Von Fraunhofer A, Delivanis PD, Anderson AN. Microleakage of retrograde amalgams. *J Endod*. 1985;11:340-345.
91. Carr G. Ultrasonic root-end preparation. *Dent Clin North Am*. 1997;41:541-554.
92. Holland R, Otoboni Filho JA, Bernabè PF, de Souza V, Nery MJ, Dezan Junior E. Effect of root canal filling material and level of surgical injury on periodontal healing in dogs. *Endod Dent Traumatol*. 1998;14:199-205.
93. Gondim E, Zaia AA, Figueiredo Almeida de Gomes BP, Ferreira CC, Teixeira FB, de Souza-Filho FJ. Investigation of the marginal adaptation of root-end filling materials in root-end cavities prepared with ultrasonic tips. *Int Endod J*. 2003;36:491-499.
94. Maddalone M, Gagliani M. Periapical endodontic surgery: a 3-year follow-up study. *Int Endod J*. 2003;36:193-198.
95. Taschieri S, Del Fabbro M, Testori T, Francetti L, Weinstein R. Endodontic surgery with ultrasonic retrotips: one-year follow-up. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod*. 2005;100:380-387.
96. Tsesis I, Rosen E, Schwartz-Arad D, Fuss Z. Retrospective evaluation of surgical endodontic treatment: traditional versus modern technique. *J Endod*. 2006;32:412-416.
97. Siqueira JF. Endodontic infections: concepts, paradigms, and perspectives. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod*. 2002;94:281-293.
98. Witherspoon D, Ham K. One-visit apexification: technique for inducing root-end barrier formation in apical closures. *Pract Proced Aesthet Dent*. 2001;13:455-460.
99. Aminoshariae A, Hartwell GR, Moon PC. Placement of mineral trioxide aggregate using two different techniques. *J Endod*. 2003;29:679-682.
100. Ruddle CJ. Nonsurgical endodontic retreatment. *J Calif Dent Assoc*. 2004;32:474-484.