

RESEARCH ARTICLE

Open Access

Metal ion levels in large-diameter total hip and resurfacing hip arthroplasty-Preliminary results of a prospective five year study after two years of follow-up

W. Maurer-Ertl, J. Friesenbichler*, P. Sadoghi, M. Pechmann, M. Trennheuser and A. Leithner

Abstract

Background: Metal-on-metal hip resurfacing is an alternative to metal-on-metal total hip arthroplasty, especially for young and physically active patients. However, wear which might be detected by increased serum ion levels is a matter of concern.

Methods: The aims of this preliminary study were to determine the raise of metal ion levels at 2-years follow-up in a prospective setting and to evaluate differences between patients with either resurfacing or total hip arthroplasty. Furthermore we investigated if the inclination of the acetabular component and the arc of cover would influence these findings. Therefore, 36 patients were followed prospectively.

Results: The results showed increments for Co and Cr in both implant groups. Patients treated with large-diameter total hip arthroplasty showed fourfold and threefold, respectively, higher levels for Co and Cr compared to the resurfacing group (Co: $p < 0,001$ and Cr: $p = 0,005$). Nevertheless, we observed no significant correlation between serum ion levels, inclination and arc of cover.

Discussion: In order to clarify the biologic effects of ion dissemination and to identify risks concerning long-term toxicity of metals, the exposure should be monitored carefully. Therefore, long-term studies have to be done to determine adverse effects of Co and Cr following metal-on-metal hip replacement.

Keywords: Hip resurfacing, Total hip arthroplasty, Prosthetic wear, Cobalt, Chromium

Background

Metal-on-metal hip resurfacing or total hip arthroplasty (THA) became an accepted and widespread procedure for joint replacement due to the favourable wear pattern, especially in young and physically active patients [1-5]. Therefore, several manufacturers introduced different hip resurfacing systems and large-diameter hip arthroplasty devices.

Nevertheless, the number of revisions for failed hip arthroplasties using metal-on-metal articulation increased, especially following hip resurfacing. The main drawback of metal articulation is the production of metal wear debris leading to elevated concentrations of metal

ions (Cobalt (Co), Chromium (Cr) and Molybdenum (Mo)) within the hip joint, the periarticular soft tissues and systemically [3,6-20]. Furthermore, there are several other factors leading to a variability in the serum levels of metal ions such as different implant size, incorrect adjustment of the components (inclination, anteversion, arc of cover), and differences in manufacture and metallurgy as well as metal corrosion [4-8,16,17,19,21,22].

Up to now there have been many studies investigating the characteristics of metal ions in the human body after hip arthroplasty with metal-on-metal bearings [1-18,21-28]. Increased concentrations have been found to be associated with hypersensitivity, local soft tissue reactions (pseudotumors), metallosis, and several other adverse effects [1-3,5-7,11,18-20,23,25]. Nevertheless, it remains unclear whether adverse reactions are dose-

* Correspondence: joerg.friesenbichler@medunigraz.at
Department of Orthopaedic Surgery, Medical University of Graz, Graz, Austria

dependent and whether they are mediated by an immune response or if they are a toxicological effect [6].

Herein, we present the preliminary results of serial serum metal ion determination after 2 years in a prospective series over a five year period in patients who had undergone metal-on-metal hip resurfacing or large-diameter total hip arthroplasty. The primary aim of this preliminary study was to determine the raise of metal ion levels at 2-years follow-up in a prospective setting. The secondary aim was to evaluate differences between patients with either resurfacing or total hip arthroplasty and if the inclination of the acetabular component and the arc of cover would influence these findings.

The first study hypothesis was, that metal ion levels would increase after 2-years of follow-up in both groups. The second hypothesis was that inclination of the acetabular component and the arc of cover would positively influence elevated metal ion levels.

Methods

A series of 36 patients who underwent metal-on-metal THA or hip resurfacing were followed prospectively over a 5-year period. For resurfacing hip arthroplasty a porous coated anatomical system was used (Articular Surface Replacement System-ASRTM, DePuy, Warsaw, IN). For metal-on-metal THA a porous coated anatomical system with a large diameter femoral head (both ASRTM XL Head, DePuy) was combined with a standard femoral shaft (Corail[®] or Future[®], DePuy). All prostheses were manufactured from Co-Cr-Mo alloy according to ISO 5832-4.

The surgeries were performed by six senior orthopaedic surgeons, using a modified antero-lateral (THA) or a posterior approach (resurfacing arthroplasty). Press fit fixation of the acetabular components and femoral stems was performed in all patients and the femoral resurfacing component was cemented.

Five men and 3 women were included in the resurfacing group (n = 8 resurfacings). These patients were chosen for this procedure due to their physical activity and their young age at time of operation (mean: 47-years, range, 33-57 years). The THA group included 28 patients (15 men and 13 women) with a mean age at operation of 52-years (range, 40-61 years). During the study period 4 patients of the THA group required replacement of the contralateral hip (n = 32 THAs) using the same metal-on-metal device. Patient demographics are reported in Table 1.

Determination of metal ion concentrations was performed preoperatively, 6 weeks, 3 and 6 months postoperatively, and annually thereafter.

Metal ion sampling

Blood was taken from 36 patients treated with metal-on-metal hip resurfacing or metal-on-metal total hip

Table 1 Patient demographics, gender distribution and component details

Demographics	ASR Resurfacing	ASR XL Head
Number of patients	8	28
Number of hips	8	32
M:F (% female)	5:3 (37,5)	15:13 (46,4)
Mean age in yrs (range)	47 (33 to 57)	52 (40 to 61)
Mean body mass index in kg/m ²	24,5 (21,3 to 27,7)	27,8 (21,6 to 37,6)
Mean acetabular size in mm	55 (50 to 58)	52 (46 to 58)
Mean femoral size in mm	49 (45 to 51)	46 (41 to 51)

All values are mean and shown with their range.

arthroplasty using stainless-steel needles attached to 9 mL no additive plastic vacuum tubes (VACUETTE[®]). All needles and tubes were used from the same batch. None of the patients had a history of renal impairment.

All specimens were centrifuged for 10 minutes at 4300 rpm within 2 hours and stored at 4°C until analysis. The concentrations of Co, Cr and Mo were determined using electrothermal graphite furnace atomic absorption spectrometry (ET ASS) in the same laboratory (Medical and Chemical Laboratory Diagnostic Lorenz & Petek GmbH). This method of analysis has been chosen due to high sensitivity and reduced matrix effects (Zeeman Effect). A second tube was stored at the department at -20°C due to the availability in case of further studies. Furthermore, the metal ion measurements can be repeated at any time if necessary.

For analysis, 300 µl of each serum sample were diluted with 50 µl Modifier und 550 µl Aqua dest. (Rotipuran). Control samples as well as a standard samples were diluted in the same way. Fifty µl Modifier and 850 µl Aqua dest. (Rotipuran) were used as blank values of the reagents. The ET ASS has been repeated twice for each sample and the levels of metal ions in the serum were recorded in concentrations expressed as µg/dl.

The sample to be examined was evaporated in an atomization apparatus and transformed into atomic condition. With graphite furnace atomic absorption spectrometry, solved samples were charged in a graphite tube with a micro pipette and liberated from the solvent and other concomitant agents by stepwise heating before being atomized. This produced a signal with an area proportional to the element of interest. The concentration of the dilution could be calculated by using the dosed volume of the sample.

Radiological analysis

Weight-bearing, anteroposterior digital radiographs from the pelvis and the operated hip were made one week, 3 months, 1 year and 2 years postoperatively. The inclination angle of the acetabular component was measured from the intersection of two lines on the anteroposterior

radiograph of the pelvis, one line across the edges of the component and the other joining both tear drop signs (Figure 1a). According to the study of De Haan et al. [22] the inclination was considered steep if the angle was higher than 55°. Furthermore, the implants' arc of cover was calculated (Figure 1b) [22]. If the arc of cover was less than 10 mm, elevated metal ion levels in serum were expected because the smaller the arc of cover the greater the tendency to rim-loading [6]. All measurements were performed using the Synedra view program, personal version 1.0.12.3 (Synedra information technologies GmbH, Innsbruck, Austria). We correlated metal ion levels with inclination and the arc of cover.

Statistical analysis

Spearman's rank correlation coefficient was calculated separated for each implant group to show if there was a

correlation between the concentrations of Co, Cr and Mo in the serum, the inclination of the acetabular component and the arc of cover. A *Mann-Whitney-U test* was performed to determine differences in median serum metal ion concentrations at the follow-up checks. This non-parametric test was chosen due to the asymmetric distribution of metal ion levels. A *p-value* of <0,05 was considered to be statistical significant. For statistic analysis the PASW Statistics 16.0 program (SPSS Inc., Chicago, IL) was used.

The study was approved by the Ethics Committee of the Medical University of Graz (EK17-265 ex 05/06) and written informed consent was obtained from all patients.

Results

The median results for cobalt and chromium in the serum of patients with the ASR™ resurfacing device




Figure 1 a&b. Plain radiographs of the pelvis of a 46-year-old female patient with an ASR™ XL Head device two years following implantation. a) Measuring cup inclination. b) Measurements to determine the arc of cover according to the method described by De Haan et al. in 2008.




Figure 2 a-c. Box plot of metal ion concentrations devided by resurfacing and THA group, a) 6-months, b) one year and c) two yeras following implantation.

were 0,09 µg/dl (range, 0,03-0,54 µg/dl) and 0,141 µg/dl (range, 0,092-0,990 µg/dl) 2 years following implantation (Figure 2a-c; Table 2). In comparison to that, patients with the ASR™ XL Head device revealed median concentrations of 0,36 µg/dl (range, 0,03-7,82 µg/dl) and 0,458 µg/dl (range, 0,071-5,198 µg/dl) for Co and Cr (Figure 2a-c; Table 2). The values for Co and Cr were fourfold and threefold, respectively, higher in the THA group compared to the resurfacing group. These differences for Co ($p < 0,001$) and Cr ($p = 0,005$) were statistically significant (Table 2).

Determining the correlation between the concentration of Co and Cr in the resurfacing group revealed a statistically not significant result (Spearman: 0,695; $p = 0,056$), while this correlation was significant in the THA group (Spearman: 0,805; $p < 0,001$; Table 3).

Two years following implantation, the mean inclination of the acetabular component was 41° (range, 35°-61°) in the resurfacing group and 45,5° (range, 28°-68°) in the THA group (Table 2). Nevertheless, differences in inclination were statistically not significant ($p = 0,351$). Consequently, there were not significant differences in the arc of cover ($p = 0,281$; Table 2). Calculations revealed no significant correlations between the concentrations of metal ions in the serum, the inclination of the acetabular component, as well as the arc of cover in both groups (Table 3). Probably, the current study failed to show a correlation between metal ion levels, inclination and arc of cover due to the small number of patients included.

Discussion

The primary aim of this preliminary study was to determine the raise of metal ion levels at 2-years follow-up in a prospective setting. The secondary aim was to evaluate differences between patients with either resurfacing or total hip arthroplasty and if the inclination of the acetabular component and the arc of cover would influence these findings.

The first study hypothesis was, that metal ion levels would increase after 2-years of follow-up in both groups. The second hypothesis was that inclination of the acetabular component and the arc of cover would positively influence elevated metal ion levels.

Our findings showed an increment of Co and Cr concentrations in the serum of patients following metal-on-metal hip resurfacing as well as THA within the first 2-years following implantation with significant differences in Co and Cr concentrations between the resurfacing and large-diameter THA group (Co: $p < 0,011$, Cr:

Table 2 Preoperative and postoperative median Co and Cr concentrations (range) following metal-on-metal hip arthroplasty

Serum metal ion levels	ASR Resurfacing	ASR XL Head	Significance
<i>Median Co(µg/dl)</i>			
preoperative	0,02 (0 to 0,03)	0,00 (0 to 0,54)	0,504
6 weeks	0,06 (0,01 to 0,19)	0,07 (0 to 0,30)	0,780
3 months	0,04 (0 to 0,33)	0,10 (0 to 0,66)	0,143
6 months	0,07 (0,01 to 0,38)	0,14 (0,00 to 2,81)	0,134
1 year	0,14 (0 to 0,39)	0,27 (0 to 5,85)	0,068
2 years	0,09 (0,03 to 0,54)	0,36 (0,03 to 7,82)	0,011
<i>Median Cr (µg/dl)</i>			
preoperative	0,020 (0,010 to 0,030)	0,026 (0,007 to 0,753)	0,293
6 weeks	0,100 (0,060 to 0,140)	0,101 (0,050 to 0,443)	0,802
3 months	0,137 (0,097 to 0,398)	0,176 (0,059 to 0,724)	0,404
6 months	0,170 (0,070 to 0,475)	0,253 (0,064 to 1,240)	0,099
1 year	0,207 (0,081 to 0,968)	0,324 (0,111 to 2,656)	0,051
2 years	0,141 (0,092 to 0,990)	0,458 (0,071 to 5,198)	0,005
<i>Inclination in °</i>			
1 week	43,8 (35,1 to 60,0)	43,6 (30,4 to 60,3)	0,892
3 months	40,9 (35,4 to 61,2)	44,1 (30,7 to 61,3)	0,465
1 year	40,9 (34,9 to 59,3)	44,5 (34,4 to 60,4)	0,465
2 years	41,0 (35,0 to 61,0)	45,5 (28,0 to 68,0)	0,351
<i>Arc of cover in mm</i>			
1 week	14,6 (7,7 to 19,2)	13,4 (5,4 to 19,9)	0,372
3 months	15,5 (7,0 to 19,0)	13,9 (5,2 to 19,9)	0,244
1 year	15,2 (8,2 to 19,1)	13,1 (5,7 to 17,9)	0,223
2 years	16,0 (7,6 to 19,5)	13,8 (6,0 to 20,2)	0,281

Mean values with their range are given for the inclination and the arc of cover.

$p < 0,005$) 2-years after implantation (Figure 2a-c) and therefore, we confirm the first hypothesis. On the other hand, we refute the second study hypothesis as there was no significant correlation between the inclination of the acetabular component and the arc of cover with metal ion concentration neither in the resurfacing nor in the THA group (Table 3).

Several authors demonstrated that the concentrations of Co and Cr ions increase within the first 3 months to 2 years following metal-on-metal hip arthroplasty (so called run-in period) [3,8,10-14,24,25]. Imanishi et al. [8] related a not significant additional increasement of serum metal ion levels at one year follow-up, whilest Back et al.[10] and Daniel et al. [11,15] observed a decreasing trend within the first 6-years. Sauvè et al. [9] reported threefold and fivefold higher serum concentrations of Co and Cr following metal-on-metal hip arthroplasty compared to a control group during a 30-year follow-up. In 2010, deSouza et al. [2] related their 10-year results following hip resurfacing with two key findings. They found a second increment of metal ion

concentrations, five years following hip resurfacing and second they found no significant differences between women and men, which is contradict to other data in the literature. In the present study, we can not provide further information whether the metal ion levels will show a further increasement, whether they will reach a plateau phase or whether they will decrease. It seems that the patients are still in the prosthesis' run-in period and further follow-up has to be done to observe the development of metal ion concentrations.

Matthies et al. [1] recently showed that metal-on-metal hip resurfacing and modular total hip replacements have same wear rates and both types of devices are associated with increased blood levels of metal ions. Therefore, this group suggested that the problems related to high wear due to metal-on-metal articulation are likely to be similar in hip resurfacing as well as large-diameter hip arthroplasty [1]. Moroni et al. [4] also found no differences in mean Co, Cr and Mo levels between patients with metal-on-metal hip resurfacing and metal-on-metal THA at 2- and 5-year follow-up. Furthermore, there were no

Table 3 Results of correlation analysis

ASR Resurfacing				
Follow-up: 1 year	Cr	Co	Inclination	Arc of cover
Co	Spearman: ,855; p=0,007		Spearman: -,229; p=0,586	Spearman: ,229; p=0,586
Cr	Spearman: ,855; p=0,007		Spearman: -,048; p=0,911	Spearman: ,048; p=0,911
Follow-up: 2 years	Cr	Co	Inclination	Arc of cover
Co	Spearman: ,695; p=0,056		Spearman: -,530; p=0,177	Spearman: ,539; p=0,168
Cr	Spearman: ,695; p=0,056		Spearman: -,778; p=0,023	Spearman: ,476; p=0,233
ASR XL Head				
Follow-up: 1 year	Cr	Co	Inclination	Arc of cover
Co	Spearman: ,819; p<0,001		Spearman: ,192; p=0,357	Spearman: -,176; p=0,389
Cr	Spearman: ,819; p<0,001		Spearman: ,103; p=0,623	Spearman: -,107; p=0,601
Follow-up: 2 years	Cr	Co	Inclination	Arc of cover
Co	Spearman: ,805; p<0,001		Spearman: ,240; p=0,201	Spearman: -,326; p=0,073
Cr	Spearman: ,805; p<0,001		Spearman: ,234; p=0,213	Spearman: -,258; p=0,161

Spearman's correlation coefficient was calculated to determine correlations between metal ion concentration, inclination and arc of cover at, one and two years postoperatively.

significant differences within the implant groups at each time of determination. Only the ion concentrations were greater in patients with metal-on-metal devices compared to a control group ($p < 0,001$) [4].

Otherwise, the current study and even the studies of Garbuz et al. [3] and Vendittoli et al. [28] showed that patients with metal-on-metal total hip arthroplasties had significantly higher concentrations of blood metal ions compared to patients undergoing hip resurfacing (Table 2).

The variability of the serum levels of metal ions is influenced by several factors such as different implant size, incorrect adjustment of the components, differences in manufacture and metallurgy as well as metal corrosion [4-8,16,17,21,22]. Shimmin et al. [21] reported the results of the Australian National Joint Registry showing an inverse relationship between the size of the femoral component and the risk of revision (component size <44 mm have a fivefold higher risk of revision) following hip resurfacing. In the series of Langton et al. [7], 17 patients with the ASRTM device were revised due to adverse reaction to metal debris (ARMD). This group had significantly smaller components, higher acetabular component anteversion and higher whole blood concentrations of Co and Cr in contrast to asymptomatic patients (all $p < 0,001$).

In several other studies increased components' wear, edge loading and raised blood and/or serum ion levels were also correlated positively with the inclination of the acetabular component [1,5-8,16,17,21,22]. A comparative study of Langton et al. [17] showed increased metal ion levels in patients with resurfacing devices and an inclination greater than 45°, while Witzleb et al. [14] showed inverse results. Furthermore, a significant inverse relationship could be found between femoral size and Co

and Cr concentrations [17]. Therefore, the authors stated that larger resurfacing implants seem to be more resistant to suboptimal position in terms of metal ion generation. This might be due to the thicker fluid film as well as the greater arc of cover [17]. On the other hand, the recent study showed no correlation between the metal ion levels in the serum, the inclination of the acetabular component, and the arc of cover (Table 3), possibly due to the small number of patients included in the study.

This study has some limitations: First, we only observed metal ion levels after a minimum follow-up of 2 years and therefore lack information of possible long-term effects and outcome. Second, the radiological evaluation of inclination and arc of cover lacks inter- and intraobserver coefficients. Third we observed a relatively small number of patients and the patients enrolled are split into two subpopulations. Fourth we did not perform an a priori power analysis but differences of outcome between the THA group compared to the resurfacing group were large enough yielding statistical significance and therefore adequate post hoc power.

Conclusion

Determining the serum concentrations of metal ions in patients following metal-on-metal hip resurfacing as well as metal-on-metal THA showed increments for Co and Cr. Furthermore, patients treated with large-diameter metal-on-metal total hip prostheses showed threefold and fourfold, respectively, higher levels for Co and Cr compared to patients with metal-on-metal resurfacing devices, 2-years following implantation. The effects of systemic metal ion exposure in patients with implants made of Co and Cr continue to be a matter of concern.

In order to clarify the biologic effects of ion dissemination and to identify risks concerning long-term toxicity of metals, the exposure should be monitored carefully. Therefore, long-term studies have to be done to determine adverse effects of Co and Cr following metal-on-metal hip replacement.

Competing interests

The authors declare that they have no competing interests.

Authors' contributions

WME, JF, MP and MT participated in the design and the coordination of the study. WME, JF, PS and AL performed acquisition of the data. WME, JF, MP and PS drafted the manuscript. AL reviewed the manuscript critically. All authors read and approved the final manuscript.

Acknowledgements

We would like to thank Prof. Lorenz and his laboratory team for performing metal ion measurements in their laboratory.

Received: 3 August 2011 Accepted: 27 March 2012

Published: 11 April 2012

References

1. Matthies A, Underwood R, Cann P, Ilo K, Nawaz Z, Skinner J, Hart AJ: Retrieval analysis of 240 metal-on-metal hip components, comparing modular total hip replacement with hip resurfacing. *J Bone Joint Surg Br* 2010, **93**:307–314.
2. deSouza RM, Parsons NR, Oni T, Dalton P, Costa M, Krikler S: Metal ion levels following resurfacing arthroplasty of the hip: serial results over a ten-year period. *J Bone Joint Surg Br* 2010, **92**:1642–1647.
3. Garbuz DS, Tanzer M, Greidanus NV, Masri BA, Duncan CP: The John Charnley Award: Metal-on-metal hip resurfacing versus large-diameter head metal-on-metal total hip arthroplasty: a randomized clinical trial. *Clin Orthop Relat Res* 2010, **468**:318–325.
4. Moroni A, Savarino L, Hoque M, Cadossi M, Baldini N: Do ion levels in hip resurfacing differ from metal-on-metal THA at midterm? *Clin Orthop Relat Res* 2010, **469**:180–187.
5. De Haan R, Campbell PA, Su EP, De Smet KA: Revision of metal-on-metal resurfacing arthroplasty of the hip: the influence of malpositioning of the components. *J Bone Joint Surg Br* 2008, **90**:1158–1163.
6. Langton DJ, Joyce TJ, Jameson SS, Lord J, Van Orsouw M, Holland JP, Nargol AV, De Smet KA: Adverse reaction to metal debris following hip resurfacing: the influence of component type, orientation and volumetric wear. *J Bone Joint Surg Br* 2011, **93**:164–171.
7. Langton DJ, Jameson SS, Joyce TJ, Hallab NJ, Natsu S, Nargol AV: Early failure of metal-on-metal bearings in hip resurfacing and large-diameter total hip replacement: A consequence of excess wear. *J Bone Joint Surg Br* 2010, **92**:38–46.
8. Imanishi T, Hasegawa M, Sudo A: Serum metal ion levels after second-generation metal-on-metal total hip arthroplasty. *Arch Orthop Trauma Surg* 2010, **130**:1447–1450.
9. Sauve P, Mountney J, Khan T, De Beer J, Higgins B, Grover M: Metal ion levels after metal-on-metal Ring total hip replacement: a 30-year follow-up study. *J Bone Joint Surg Br* 2007, **89**:586–590.
10. Back DL, Young DA, Shimmin AJ: How do serum cobalt and chromium levels change after metal-on-metal hip resurfacing? *Clin Orthop Relat Res* 2005, **438**:177–181.
11. Daniel J, Ziaee H, Pradhan C, Pynsent PB, McMinn DJ: Blood and urine metal ion levels in young and active patients after Birmingham hip resurfacing arthroplasty: four-year results of a prospective longitudinal study. *J Bone Joint Surg Br* 2007, **89**:169–173.
12. Savarino L, Granchi D, Ciapetti G, Cenni E, Nardi Pantoli A, Rotini R, Veronesi CA, Baldini N, Giunti A: Ion release in patients with metal-on-metal hip bearings in total joint replacement: a comparison with metal-on-polyethylene bearings. *J Biomed Mater Res* 2002, **63**:467–474.
13. Engh CA Jr, MacDonald SJ, Sritulanondha S, Thompson A, Naudie D, Engh CA: 2008 John Charnley award: metal ion levels after metal-on-metal total hip arthroplasty: a randomized trial. *Clin Orthop Relat Res* 2009, **467**:101–111.
14. Witzleb WC, Ziegler J, Krummenauer F, Neumeister V, Guenther KP: Exposure to chromium, cobalt and molybdenum from metal-on-metal total hip replacement and hip resurfacing arthroplasty. *Acta Orthop* 2006, **77**:697–705.
15. Daniel J, Ziaee H, Pradhan C, McMinn DJ: Six-year results of a prospective study of metal ion levels in young patients with metal-on-metal hip resurfacings. *J Bone Joint Surg Br* 2009, **91**:176–179.
16. Langton DJ, Jameson SS, Joyce TJ, Webb J, Nargol AV: The effect of component size and orientation on the concentrations of metal ions after resurfacing arthroplasty of the hip. *J Bone Joint Surg Br* 2008, **90**:1143–1151.
17. Langton DJ, Sprowson AP, Joyce TJ, Reed M, Carluke I, Partington P, Nargol AV: Blood metal ion concentrations after hip resurfacing arthroplasty: a comparative study of articular surface replacement and Birmingham Hip Resurfacing arthroplasties. *J Bone Joint Surg Br* 2009, **91**:1287–1295.
18. Pazzaglia UE, Apostoli P, Congiu T, Catalani S, Marchese M, Zarattini G: Cobalt, chromium and molybdenum ions kinetics in the human body: data gained from a total hip replacement with massive third body wear of the head and neuropathy by cobalt intoxication. *Arch Orthop Trauma Surg* 2011, **131**(9):1299–308.
19. Steens W, Loehr JF, von Foerster G, Katzer A: Chronic cobalt poisoning in endoprosthetic replacement. *Orthopade* 2006, **35**:860–864.
20. Oldenburg M, Wegner R, Baur X: Severe cobalt intoxication due to prosthesis wear in repeated total hip arthroplasty. *J Arthroplasty* 2009, **24**:825 e815–820.
21. Shimmin AJ, Walter WL, Esposito C: The influence of the size of the component on the outcome of resurfacing arthroplasty of the hip: a review of the literature. *J Bone Joint Surg Br* 2010, **92**:469–476.
22. De Haan R, Pattyn C, Gill HS, Murray DW, Campbell PA, De Smet K: Correlation between inclination of the acetabular component and metal ion levels in metal-on-metal hip resurfacing replacement. *J Bone Joint Surg Br* 2008, **90**:1291–1297.
23. Hart AJ, Hester T, Sinclair K, Powell JJ, Goodship AE, Pele L, Fersht NL, Skinner J: The association between metal ions from hip resurfacing and reduced T-cell counts. *J Bone Joint Surg Br* 2006, **88**:449–454.
24. Antoniou J, Zukor DJ, Mwale F, Minarik W, Petit A, Huk OL: Metal ion levels in the blood of patients after hip resurfacing: a comparison between twenty-eight and thirty-six-millimeter-head metal-on-metal prostheses. *J Bone Joint Surg Am* 2008, **90**(Suppl 3):142–148.
25. Savarino L, Granchi D, Ciapetti G, Cenni E, Greco M, Rotini R, Veronesi CA, Baldini N, Giunti A: Ion release in stable hip arthroplasties using metal-on-metal articulating surfaces: a comparison between short- and medium-term results. *J Biomed Mater Res A* 2003, **66**:450–456.
26. Grubl A, Marker M, Brodner W, Giurea A, Heinze G, Meisinger V, Zehetgruber H, Kotz R: Long-term follow-up of metal-on-metal total hip replacement. *J Orthop Res* 2007, **25**:841–848.
27. De Smet K, De Haan R, Calistri A, Campbell PA, Ebramzadeh E, Pattyn C, Gill HS: Metal ion measurement as a diagnostic tool to identify problems with metal-on-metal hip resurfacing. *J Bone Joint Surg Am* 2008, **90**(Suppl 4):202–208.
28. Vendittoli PA, Amzica T, Roy AG, Lusignan D, Girard J, Lavigne M: Metal Ion release with large-diameter metal-on-metal hip arthroplasty. *J Arthroplasty* 2010, **26**:282–288.

doi:10.1186/1471-2474-13-56

Cite this article as: Maurer-Ertl et al.: Metal ion levels in large-diameter total hip and resurfacing hip arthroplasty-Preliminary results of a prospective five year study after two years of follow-up. *BMC Musculoskeletal Disorders* 2012 **13**:56.