

Published in final edited form as:

Lab Chip. 2013 August 7; 13(15): 2990–2998. doi:10.1039/c3lc50424g.

A microfluidic platform for generating large-scale nearly identical human microphysiological system arrays

Yu-Hsiang Hsu^{a,+}, Monica L. Moya^{a,+}, Christopher C.W. Hughes^{a,b}, Steven C. George^{c,d,¶}, and Abraham P. Lee^{a,e,¶,*}

^aDepartment of Biomedical Engineering, University of California, Irvine, CA 92697 USA

^bMolecular Biology and Biochemistry, University of California, Irvine, CA 92697 USA

^cChemical Engineering and Materials Scienc, University of California, Irvine, CA 92697 USA

^dMedicine, University of California, Irvine, CA 92697 USA

^eMechanical and Aerospace Engineering, University of California, Irvine, CA 92697 USA

Abstract

This paper reports a polydimethylsiloxane microfluidic model system that can develop an array of nearly identical human microtissues with interconnected vascular networks. The microfluidic system design is based on an analogy with an electric circuit, applying resistive circuit concepts to design pressure dividers in serially-connected microtissue chambers. A long microchannel (550, 620 and 775 μm) creates a resistive circuit with a large hydraulic resistance. Two media reservoirs with a large cross-sectional area and of different heights are connected to the entrance and exit of the long microchannel to serve as a pressure source, and create a near constant pressure drop along the long microchannel. Microtissue chambers (0.12 μl) serve as a two-terminal resistive component with an input impedance > 50-fold larger than the long microchannel. Connecting each microtissue chamber to two different positions along the long microchannel creates a series of pressure dividers. Each microtissue chamber enables a controlled pressure drop of a segment of the microchannel without altering the hydrodynamic behaviour of the microchannel. The result is a controlled and predictable microphysiological environment within the microchamber. Interstitial flow, a mechanical cue for stimulating vasculogenesis, was verified by finite element simulation and experiments. The simplicity of this design enabled the development of multiple microtissue arrays (5, 12, and 30 microtissues) by co-culturing endothelial cells, stromal cells, and fibrin within the microchambers over two and three week periods. This methodology enables the culturing of a large array of microtissues with interconnected vascular networks for biological studies and applications such as drug development.

© The Royal Society of Chemistry [year]

* Tel: +1-(949) 824-9691; Fax: +1-(949) 824-1727; aplee@uci.edu, 3120 Natural Sciences II, Irvine, CA 92697-2715, USA.

⁺These authors contributed equally to this paper

[¶]These authors contributed equally as senior authors

Introduction

Recent advances in microfluidic technology have enabled the *in vitro* study of cellular behaviours in 3-D microenvironments.¹⁻² By designing the microfluidic system to control both the flow and concentration profiles of the constituents, the mechanical and chemical environment in a small volume can be controlled temporally and spatially. Based on these unique capabilities, microsystems that simulate *in vivo* physiological microenvironments have been demonstrated for studying endothelial cell migration,³ angiogenesis,^{4,5} flow in single capillaries,⁶ and flow in a network of endothelial-lined channels.⁷⁻⁸ However, in order to perform large-scale parallel assays with 3-D tissues, it is critical to develop a microfluidic array platform with a user-friendly interface. Current microfluidic methodologies and systems (e.g. on-chip passive pumps) pose major limitations to grow 3-D microtissues in a high-throughput fashion. In these systems, considerable pressure loss takes place in a short amount of time (ranging from minutes to a day).⁹⁻¹⁰ Constant adjustment of these passive pumps is needed to maintain the 3-D tissue microenvironments, both mechanically and chemically. On the other hand, active pumping systems are bulky and typically do not provide a user-friendly interface. Furthermore, the set up time and complex fluid manifolds make it challenging for large-scale applications.¹¹⁻¹² For example, it is not possible to rapidly change medium through the microfluidic system without influencing or interrupting the created 3-D microenvironment. Thus, a microfluidic model system that can control 3-D microenvironments in large-scale arrays for long culturing times (weeks and above) and with a simple user-interface (as simple as changing the medium of a culture dish) would be a major stepping stone for in depth tissue engineering studies.

In this paper, we present an *in vitro* model system that can generate multiple human microtissues with a self-assembled and connected microvascular network that are nearly identical. The user interface is simple and can be used in regular tissue culture incubators. Instead of constructing an artificial substrate⁶⁻⁸ or controlling concentration gradients of chemical stimuli,⁴⁻⁵ our system provides proper microphysiological conditions to allow endothelial cells to self-assemble (vasculogenesis) into a vascular network by controlling interstitial flow. Interstitial flow ranging from 0.5 to 10 $\mu\text{m/s}$ stimulates the vasculogenic process mechanically,¹³⁻¹⁶ and also manipulates the spatial concentration of growth factors and morphogens.

The physiological microenvironment can be controlled using the concept of a pressure divider and a resistive circuit, applying principles used in the design of electric circuits.¹⁷ This analogy has been used by others to design proportional networks¹⁸ and serial networks¹⁹ for serial dilution, as well as pyramidal networks for creating concentration gradients.²⁰ A microfluidic resistance network has also been applied to culture cells in a large array,²¹ and to control the number of trapped cells for controlling the size of embryoid bodies.²² In the present work, multiple human microtissues with a continuous and connected (within and between chambers) microvascular network were developed over two to three weeks, and the physiological microenvironment was verified by finite element simulation. Using this microfluidic platform, metabolically active microtissues with perfused microvascular networks were demonstrated. The detailed experimental findings of the perfused vascular network were presented in a separate paper.²³ Herein, we present the

design, modelling, and simulation of a microfluidic platform for producing arrays of nearly identical microtissues. The methodology of the presented platform forms the basis of new high-throughput tissue culture systems for studying tumour metastasis, drug discovery, vascular disease, and chemical toxicity.

Materials and methods

A Design of microfluidic network

The basic design is a long, square (100 μm by 100 μm) cross-sectional microchannel with culture media flowing at low Reynolds number ($\ll 1$) driven by a pressure difference between the source (P_S) and the end of the channel (P_E). A steady-state incompressible Navier-Stokes equation with homogeneous viscosity can be applied to derive the velocity and pressure distribution within the channels and the microtissue chambers. Since a long (relative to the size of the microchambers; 1 mm by 2 mm by 0.1 mm in diamond shape) microchannel was used in this study (550, 620 and 775 μm), a fully developed laminar flow could be assumed. Thus, the Hagen–Poiseuille law can be used to describe the fully developed laminar flow throughout the long microchannel, and is generalized to $P = R_T Q$,¹⁷ where $P = P_S - P_E$, and R_T and Q are the total hydraulic resistance and volumetric flow rate. The long microchannel can then be considered as a series of microchannel segments connected to each other, and each segment can be lumped into an equivalent hydraulic resistor R_i . This arrangement is analogous to Ohm's law of linear electric circuits. Furthermore, if the pressure gradient along the long microchannel is constant, pressure dividers in the microfluidic channel become analogous to voltage dividers in electric circuit design.

Using this configuration, the specific pressure drop across a microtissue chamber can be achieved by applying the design concept of a voltmeter.²⁴ To accurately measure the voltage drop across a component of an electric circuit, the voltmeter connects a large resistor in series with a Galvanometer in parallel to the electric component to measure the voltage across it. Thus, a very low current is drawn from the electric circuit without altering the voltages or currents in the circuit. Following the same design concept, the long microfluidic channel and the microtissue chamber can be considered as the electric circuit and the voltmeter, respectively. However, different from the purpose of a voltmeter, the “measured pressure” is used as the “driving pressure” for creating a physiological environment in the microtissue chamber.

To realize this design, the microtissue chamber was designed to have high input impedance with two communication pores, which served as a two-terminal resistive loading component. By connecting the communication pores at different positions along the long microchannel, the pressure drop across the microtissue chamber can be controlled using the concept of the voltage divider. Furthermore, since only a very small amount of media flows into the microtissue chamber, a constant pressure drop along the long microchannel is maintained. This concept has been applied previously to capture the full physiological range of mass transport in 3-D tissue cultures to study vasculogenesis.¹⁶

Figure 1(A) shows the basic design concept, where the constant pressure drop (P_{L1}) across the first two-terminal component (microtissue chamber R_{L1}) can be written as:

$$\Delta P_{L1} = \frac{[(R_2+R_3) // R_L] \Delta P}{R_1 + [(R_2+R_3) // R_L] + R_4 + R_5}, \quad \text{where (1)}$$

$$\Delta P_{L1} \approx \frac{R_2+R_3}{R_1+R_2+R_3+R_4+R_5} \Delta P, \quad \text{when } [(R_2+R_3)/R_L] \ll 1. \quad (2)$$

Similarly, the constant pressure drop of the second two-terminal component (microtissue chamber R_{L2}) can be written as:

$$\Delta P_{L2} = \frac{[(R_3+R_4) // R_L] \Delta P}{R_1+R_2 + [(R_3+R_4) // R_L] + R_5}, \quad \text{where (3)}$$

$$\Delta P_{L2} \approx \frac{R_3+R_4}{R_1+R_2+R_3+R_4+R_5} \Delta P, \quad \text{when } [(R_3+R_4)/R_L] \ll 1. \quad (4)$$

where // is a symbol to represent the equivalent resistance of resistors connected in parallel. If the microchannel segments are equal in length, then the resistance of R_2 , R_3 , and R_4 are equal in magnitude, R . The constant pressure drop applied to both microtissue chambers becomes identical and can be written as:

$$\Delta P_{Li} \approx \frac{2R}{R_1+3R+R_5} \Delta P, \quad \text{when } R \ll R_L, \quad (5)$$

where subscript $i = 1$ or 2 . By using this design strategy, multiple microtissue chambers with high input impedance (R_L) can be connected evenly along a long microfluidic channel at equal distances from each other. As a result, the pressure drop across each microtissue chamber is identical. Figure 1(B) shows an equivalent circuit of n microtissue chambers connected to a long microchannel separated into $2n+1$ hydraulic resistors, where R_1 and R_{2n+1} are the hydraulic resistors of the entrance and the exit portions of the long microchannel, and $R_2 = R_3 = \dots = R_{2n} = R$. The pressure drop P_L across each microchamber (R_{Li}) can be written as

$$\Delta P_L = \left[\left(\sum_{i=1}^n R_{i+1} \right) / \left(\sum_{j=1}^{2n+1} R_j \right) \right] \Delta P = \frac{nR}{R_T} \Delta P, \quad (6)$$

where R_T is the total hydraulic resistance of the long microchannel, and $P = P_s - P_E$. Note also that the same concept can be applied to create different pressure drops by connecting the two terminals of microchambers to a long microchannel with different separation distances. This enables a spectrum of physiological conditions to be realized on one microfluidic platform. An example verifying this alternate design and simulation is discussed in the supplementary document.

With the design described here, only a limited number of microchambers can fit on the microfluidic platform due to the need to route the long microchannel around the microchambers. In order to bypass the need to maintain each segment of the long microchannel to be the same ($R_2 = R_3 = \dots = R_{2n} = R$), the hydraulic resistor R_{n+1} between the two terminals of every microchamber can be increased to dominate the resistive circuit. For example, if R_3 is dominant [Fig. 1(A)], the difference between R_2 and R_4 can be ignored, and Eq (5) can be modified to describe pressure drops in both resistors.

$$\Delta P_{Li} = \frac{R_3 + R + e}{R_1 + R_3 + R + e + R_5} \Delta P \approx \frac{R_3 + R}{R_1 + R_3 + R + R_5} \Delta P, \quad \text{when } R_L \gg R_3 \gg R \gg e, \quad (7)$$

where e is the small variation of resistance generated from slight length variations between R_2 and R_4 . With this design principle, the variation of hydraulic resistors of each segment of a large resistance circuit can be neglected. The pressure drop P_{Li} of the i^{th} microchamber (R_{Li}) can be simplified to

$$\Delta P_{Li} = \frac{\sum_{c=i}^{n-1} R_{c+1} + R_{n+1} + \sum_{d=n+2}^{n+i} R_d}{\sum_{a=1}^n R_a + R_{n+1} + \sum_{b=n+2}^{2n+1} R_b} \Delta P = \frac{R_{n+1} + R_e + e_i}{R_T} \Delta P \approx \frac{R_{eq}}{R_T} \Delta P, \quad (8)$$

where $R_e + e_i$ is a lumped resistor from other microchannel segments between two terminals [Fig. 1(C)], e_i is the small variation of resistance generated from slight length variations for each i^{th} microchamber, $R_{n+1} \gg R_e \approx (n-1)R > e_i$, and $R_{eq} = R_{n+1} + R_e$. By using this method, the pressure drop across all microchambers is nearly identical and a large array of nearly identical microchambers can be realized.

B High input impedance microtissue chamber design

Figure 2(A) shows the basic design of the microtissue chamber, which is filled with a fibrin gel seeded with stromal cells (normal human lung fibroblasts, NHLFs) and endothelial cells (ECs) (see below) to form a 3-D tissue construct. Stromal cells are required for the development of stable microvessels.²⁵

Based on our design methodology, two key requirements must be considered for the microchamber: 1) a high input impedance, and 2) a proper physiological microenvironment for initiating and maintaining vasculogenesis. The high input impedance was created by combining the hydraulic permeability of the porous matrix with small cross-section communication pores [30 μm by 100 μm , Fig. 2(A)]. The hydraulic impedance (R_{Li}) between the two communication pores was more than 50-fold higher than the total hydraulic impedance of the long microchannel R_T . This allowed the microtissue chamber to be connected to any two arbitrary positions along the long microchannel without influencing the resistive circuit. Since the pressure drop along the long microchannel served as a constant pressure source to the microtissue chamber, the pressure drop inside the microchamber was determined solely by the distance between the two communication pores along the microchannel. The constant pressure was achieved by using the large hydraulic resistance of the microchannel itself and two media reservoirs with a large cross-sectional

area (12.5mm in diameter) connected to the entrance and exit of the long microchannel [Fig. 2(B)]. The maximum rate of pressure drop was only 20 $\mu\text{m/hr}$.

The physiological environment to stimulate vasculogenic process was created by manipulating the mass transport in the microtissue chamber. The mechanical stimulus, interstitial flow, was introduced and controlled by the pressure drop across the microtissue chamber. The nutrient supply was controlled by the diamond shaped and millimeter-sized microchamber [Fig. 2(A), height 100 μm]. In order to control the oxygen supply, the thickness of the top PDMS was larger than 8 mm to provide a long diffusion length for oxygen,²⁶ and a 1 mm thick glass slide was bonded to the bottom PDMS sheet to block oxygen diffusion from the bottom. Thus, the majority of oxygen supplied to the 3-D cell construct was from the small communication pore positioned at the high pressure side, and delivered throughout the microchamber by the resulting interstitial flow. Moreover, to mimic physiological oxygen levels in adult tissue, the microplatform was cultured under 5% oxygen, thus maintaining the oxygen level inside the long microchannel and media reservoirs was at approximately 5%. Under this microphysiological environment, the 3-D cell constructs initiated vasculogenesis and the formation of a continuous vascular network. Furthermore, the diamond-shaped microchamber provided a symmetrical chamber to balance the contractile force generated by the stromal cells in the microtissue. The necking channels shown in Fig. 2(A) were used to load the cell-matrix solution, and they also served to connect multiple microchambers in series for large scale loading. In addition, these necking channels also minimized the influence of vasculogenic stimuli from adjacent microchambers. Finally, the concept of a capillary burst valve was used for the design of communication pores to inhibit leaking of the fibrin solution into the microfluidic channel during loading as previously described.^{16, 23,27}

C Cell culture and cell-matrix preparation

Endothelial colony forming cell-derived endothelial cells (ECFC-ECs) were isolated from cordblood²⁸ and expanded on gelatin-coated flasks in endothelial growth medium-2 (EGM-2, Lonza). Normal lung human fibroblasts (NHLFs, Lonza) were cultured in fibroblast growth media (Lonza). ECFC-ECs (ECs, used at passages 4-7) and NHLFs (used at passage 3-5) were both grown in a 37 °C / 5% CO₂ / 20% O₂ incubator in 100% humidified air. The cell-matrix solution was prepared by dissolving bovine fibrinogen (Sigma-Aldrich) in DPBS (10 mg/ml), and ECs (5 $\times 10^6$ cells/mL) and NHLFs (2.5 $\times 10^6$ cells/mL) were suspended in the fibrinogen solution at a ratio of 2:1. Thrombin (50 U/ml) was then added to the cell-matrix solution for a final concentration of 3 U/ml. This allowed the gel to be fully polymerized within 30 minutes after being pipetted into the microtissue chambers. Fully supplemented EGM-2 media and 20% oxygen tension was supplied through the long microchannel for the first 12 hours. Then, the fully supplemented media was replaced by EGM-2 without VEGF and bFGF, and the levels of the media in the two media reservoirs were levelled to the constant pressure drop P of the resistive network. The microfluidic platform was then placed in a 5% O₂ incubator for 2 to 3 weeks to allow cell constructs to develop into microtissues with a continuous vascular network. The media levels of the two media reservoirs were adjusted every other day to maintain a constant P .

The cell nuclei and developed microvessels were then fixed and stained with DAPI (nuclear) and anti-CD31 (ECs) antibody for imaging.

D Finite Element Simulation

As previously described, we used finite element simulation to estimate the hydraulic resistances of the designed microfluidic network.¹⁶ Briefly, COMSOL Multiphysics® 3.5a was used to simulate the mass transport induced in each microchamber. The incompressible Navier-Stokes equation was used to perform the steady-state analysis of the 3-D microfluidic models. No-slip boundary condition was applied to all surfaces except the entrance and exit of the long microchannel. Simulated velocity field was saved, and used to solve the time-dependent convection and diffusion problem for mass transfer. Dextran at $1 \text{ mol}\cdot\text{m}^{-3}$ concentration (diffusion coefficient of $7\times 10^{-11} \text{ m}^2/\text{s}$) was simulated flowing from the entrance into the long microchannel. The dynamic viscosity and density of water were $0.748 \text{ mPa}\cdot\text{s}$ and $1 \text{ kg}/\text{m}^3$. The porosity and permeability of fibrin gel were estimated to be 0.99 and $1.5\times 10^{-13} \text{ m}^2$ by using the Brinkman equation.¹⁵⁻¹⁶ These parameters were retrieved from the best fitted parameters used in our previous published paper.¹⁶

E Microfabrication of the PDMS microfluidic platform

A $100 \mu\text{m}$ thick SU-8 100 (Micro Chem) was first spun on a cleaned silicon wafer [RCA-1 cleaned and 2% HF (Hydrofluoric acid) dipped]. A single mask photolithography process was then employed to create the master mold for the long microchannel, communication pores, and microchambers. The SU-8 molds were silanized with trichlorosilane ($\text{C}_8\text{H}_4\text{Cl}_3\text{F}_3\text{Si}$), and a 10-to-1 mixture of PDMS prepolymer and curing agent (Dow Corning) was cast on the SU-8 mold. A cured PDMS microfluidic network was then demolded, and plasma bonded to a 1 mm thick PDMS sheet that was itself plasma bonded to a 1 mm thick glass slide. Two glass vials with the bottom cleaved were then glued to the entrance and exit of the long microchannel using a PDMS mixture and then cured in an $85 \text{ }^\circ\text{C}$ oven for 4 hours. The finished microfluidic platform was then sterilized at $121 \text{ }^\circ\text{C}$ for 20 minutes.

Results and Discussion

A Multi-microtissue platform based on pressure divider

Figures 3(A) and 3(B) shows the design and finite element simulated pressure field of a microtissue platform based on the equivalent circuit presented in Fig. 1(B) and Eq. (6). A total of 12 microtissue chambers and a 550 mm long microchannel were used for this microplatform. The total pressure drop P across the long microchannel was maintained at $10 \text{ mmH}_2\text{O}$ (98.1 Pa). The simulated hydraulic resistances were $R_T = 122.5\times 10^{12} \text{ Pa}\cdot\text{s}/\text{m}^3$, $R_i = 4.42\times 10^{12} \text{ Pa}\cdot\text{s}/\text{m}^3$ with a standard deviation of $0.29\times 10^{12} \text{ Pa}\cdot\text{s}/\text{m}^3$. The average hydraulic resistance R_{Li} of each microchamber was $5094\times 10^{12} \text{ Pa}\cdot\text{s}/\text{m}^3$, which was large enough to have negligible influence on the resistive circuit of the long microchannel (i.e., $R_{Li} \gg R_T$). Notice that the microchannel segment [R_{n+1} specified between two arrows; Fig. 3 (A)] between all of the two communication pores of each microtissue chamber could not have a similar length to other microchannel segments (R_i) between adjacent microchambers, its hydraulic resistance was 3.56 times higher ($15.75\times 10^{12} \text{ Pa}\cdot\text{s}/\text{m}^3$). A microfluidic jumper

made by a short silicone tube was added in its path to minimize the length of this segment [see the fabricated PDMS device shown in Fig. 3(C)]. In addition, this jumper allowed breaking up the long microchannel in to two, and had one segment on each side of the microtissue chambers. This design offered the capability to minimize the amount of shear stress exerted on exposed fibrin gel at the pores while media entered the long microchannel during loading. The jumper can also be removed for additional investigation of the system following the development of the human microtissues which required the pressure on either side of the microtissue to be decoupled.

The calculated hydraulic resistance of the microchannel segment connected in parallel with each microchamber was $0.51R_T$ with standard deviation of 0.014. It is more than 80 times smaller than the input impedance of the microtissue chamber (R_{Li}). Thus, a very small amount of media flows into the microtissue chamber. This demonstrates the voltmeter analogy, as the pressure drop across each microtissue chamber was nearly identical. Figure 3(B) shows the simulated velocity field and pattern of interstitial flow (streamlines) of two consecutive microtissue chambers which have essentially identical velocity fields. Figures 3(D) and 3(E) are the simulated pressure field and velocity field of each microtissue chamber starting from the communication high pressure pore at (0 mm) and ending at the low pressure pore (1 mm). These data lines are numbered C1 to C12 sequentially to represent microtissue chambers from the left to right order shown in Fig 3(A). It is evident from Fig 3(D) that the magnitude of hydraulic pressure applied to each communication pore decreases with a near constant value while having a similar pressure gradient within the microtissue. This design results in a nearly identical velocity field in all the microtissue chambers [Fig. 3(E)]. Further, the range of interstitial flow was controlled in the range that was identified for stimulating vasculogenic process.¹³⁻¹⁶ In Fig. 3(E), the velocity of interstitial flow is between $1.18 \mu\text{m/s}$ to $7.10 \mu\text{m/s}$ over the majority of the microchamber (90%), but increases to $65 \mu\text{m/s}$ near the communication pores due the narrow opening of the pore and rapidly expanding field for flow. Since this central line has the highest velocity profile in the microchamber, the range of created interstitial flow in the majority of the chamber is controlled below $7.1 \mu\text{m/s}$.

The effect of the necking channel was also verified by observing the velocity profiles of adjacent microtissue chambers [Fig 3(B)]. The velocity of mass transport in the necking channel was approximately $10^{-8} \mu\text{m/s}$; thus, there is essentially zero communication between adjacent microchambers. This effect was verified experimentally by flowing fluorescein isothiocyanate (FITC) dextran (Sigma Aldrich, 70 kDa MW) in the long microchannel [Fig 3(F)]. The FITC dextran is transported from bottom to top direction, but not to the adjacent microtissue chamber. Note also that the microvessel network is visible as dark lines demonstrating that the 70 kDa MW FITC-dextran does not penetrate the microvessel through the basal surface. This phenomenon was also observed with 150 kDa dextran (data not shown).²³ Finally, all 12 microtissue chambers developed an interconnected of microvessels after 2-week and 3-week culture. Figure 3(G) shows the developed vascular network in two of the 12 microtissues. The microvascular network and nuclei were stained with CD-31 anti-body (green) and DAPI (blue). Figure 3(H) shows 3 independent experiments characterized by measuring the total vessel length of each

microtissue chamber after a 3-week culture. The average total vessel length for the experiments are 14.85 mm (Exp1; standard deviation = 2.47 mm), 14.95 mm (Exp2; standard deviation = 2.3 mm), and 15.42 mm (Exp3; standard deviation = 2.68 mm), respectively.

B Multi-microtissue platform with a high hydraulic resistance

Figure 4 shows an alternate design, and finite element simulated pressure field, of a microtissue platform based on a large resistor R_{n+1} [Eq. (8)]. Five microtissue chambers are connected along a 620 mm long microchannel. The length of the microchannel segment between the two communication pores of the first and last microtissue chambers was increased to 560 mm long such that a very large resistance separated them [R_{n+1} of 92.6×10^{12} Pa·s/m³, Fig. 4(A); specified between two arrows]. This was 62.4 times larger than each microchannel segment R_i , where $R_e = 3.64 \times 10^{12}$ Pa·s/m³ and $0.9 \times 10^{12} < e_i < 1.1 \times 10^{12}$ Pa·s/m³ [Fig. 1(C)]. Since R_{n+1} was nearly 93 times larger than e_i , the variations in hydraulic resistance between two terminals of each microchamber could be neglected. The approximation presented in Eq. (8) could then be applied. The calculated hydraulic resistance R_{L_i} of each microchamber was 6273×10^{12} Pa·s/m³ and thus was also significantly higher than the long microchannel. The calculated hydraulic resistance of the microchannel segment in parallel with each microchamber was $0.86R_T$ with a standard deviation of 0.0076. The total pressure drop P of the long microchannel was maintained at 15 mmH₂O (147.1 Pa). The finite element simulated velocity field and the pattern of interstitial flow are shown in Fig. 4(B), and the microfabricated microplatform is shown in Fig. 4(C). The simulated pressure field and velocity field of each microtissue chamber are shown in Fig. 4(D) and 4(E), respectively. The simulated pressure field clearly demonstrates that by introducing a long segment of microchannel between the two communication pores of every microtissue chamber, a high hydraulic resistance R_{n+1} is introduced and dominates the pressure divider [Eq. (8)]. The magnitude of pressure applied to the communication pores was thus nearly the same to every microtissue chamber, and a nearly identical velocity profile was created [Fig. 4(E)]. Note also that induced interstitial flow is within the identified physiological range for stimulating vasculogenesis.¹³⁻¹⁶ Similar to the first microtissue platform (Fig. 3), the velocity of interstitial flow was controlled between 2.88 μm/s and 10.00 μm/s in 80% of the central area. The remaining 20% area was increased to 95 μm/s due to the narrow pore and rapidly enlarging cross-sectional area. The interstitial flow is controlled below 10 μm/s in the majority of the microchamber.

The simulation results verified nearly equal pressure drop P_L across each microchamber. To visualize the distribution of nutrient supply from the communication pore at high pressure, fluorescent dextran (70 kDa) was introduced into the long microchannel under identical P . The time for dextran to fill the whole microchamber was approximately 10 hours [Fig. 4(F); left panel experimental results, right panel simulated result for each time point]. This result demonstrates that similar microphysiological environments can be created within microtissue chambers with negligible influence from adjacent chambers. Note that the U-shaped necking channel between microtissue chambers can significantly suppress the interaction from adjacent chambers. Evidence of initiation of vasculogenesis was verified by vacuole formation within 12-hours of culture [Fig. 4(G)].²⁹

Fully developed (i.e., continuous) microvascular networks were found in all microtissue chambers after two weeks in culture. Figure 4(H) shows 3 independent experiments characterized by measuring the total vessel length of each microtissue chamber after a 2-week culture. The average total vessel length for the experiments are 16.38 mm (Exp1; standard deviation = 0.55 mm), 16.41 mm (Exp2; standard deviation = 0.68 mm), and 16.13 mm (Exp3; standard deviation = 0.60 mm), respectively. Figure 4(I) shows a fluorescent image of the five microtissues developed in this microplatform, where microvessels and nuclei were stained with anti-CD31 (green) and DAPI (blue), respectively. Furthermore, the necking channels connected to the loading pipette tips (first and last microchamber) minimized the influence of the cell suspension remaining in the tips (white arrows). Thus, each microtissue chamber can be considered as an independent environment, and both cell and tissue level (i.e., microvessel network) phenotype can be investigated in parallel.

C A large-scale multi-microtissue platform

Utilizing the design criteria shown in Fig. 4 and Eq. (8), a microtissue platform with a larger array of 30 microtissue chambers was designed. The total length of the long microchannel was 775 mm. Fig. 5(A) shows a fluorescent image of a 2-week cultured microtissue array taken by a digital camera. The developed microvascular network was labelled with CD31 anti-body (white). Fig 5(B) shows the microfabricated microfluidic platform (filled with red dye for demonstration) and a zoomed-in image of 30 microtissue chambers. The 30 microtissues were created by 3 separate loadings of cell-matrix suspensions from the 3 bottom loading holes. With this microplatform, 30 microtissues with a continuous vascular network were formed after two weeks in culture [Fig. 5(A)]. Figures 5(C) and 5(D) represent two of the microchambers at higher magnification (white rectangles in Fig 5A labeled “1” and “2”). Figure 5(E) shows the measured total vessel length of each microtissue chamber shown in Fig. 5(A). The average total length is 12.46 mm and the standard deviation is 1.87mm. This experimental result demonstrated that a larger array of microtissue chambers with nearly identical physiological environments can be created simultaneously to create nearly identical microvessel networks in parallel.

Conclusions

We present a novel microfluidic model system that can develop nearly identical multiple human microtissues in parallel on a PDMS microplatform by controlling the mechanical and chemical microphysiological environment. For our application, we used the device to create self-assembled and continuous microvascular networks over two and three weeks in culture in three different designs that consisted of 5, 12, and 30 microtissues. The experimental and theoretical results demonstrate that the platform design can control the major determinants of vasculogenesis [interstitial flow in the physiological range (0.5 to 10 $\mu\text{m/s}$)] in every microtissue chamber.¹³⁻¹⁶

The microfluidic system is flexible and can be generalized to create a host of physiological environments for tissue engineering applications. First, the nearly identical mechanical and chemical environments can be manipulated and controlled to create parallel investigation of microtissue development. Second, the composition of the cell-matrix suspension can be

selectively seeded to investigate the impact of cell type or matrix composition. Third, as suggested in the supplementary document, the magnitude of the pressure gradient in microchambers need not be identical in all chambers, but can be manipulated to generate different physiological conditions. This can be realized by utilizing a long microchannel with different separation lengths. Using this method, 3-D cellular behaviours in response to different physiological conditions can be investigated simultaneously on one microfluidic platform. This method is simple, and can be easily adapted to essentially all current microfluidic systems for creating a specific microphysiological environment for 3-D tissue engineering studies.

The platform design is based on an analogy between an electrical circuit and a microfluidic network. Two microfluidic components were used to create the microfluidic model system. The first is a microchannel that spans several hundred millimeters and serves as a constant pressure source for each microtissue chamber. The second is the millimeter-sized microtissue chamber with high input impedance. The microchamber draws a very small flow while using the measured pressure drop to control the internal microenvironment. The method introduced here provides a simple and straightforward means to create nearly identical microphysiological environments in an array of microtissue chambers. This microplatform and its application can potentially impact such fields as tumour metastasis, drug discovery, vascular disease, and chemical toxicity.

Supplementary Material

Refer to Web version on PubMed Central for supplementary material.

Acknowledgments

This research was supported by National Institutes of Health RC1 ES018361 and UH2 TR000481.

References

1. Griffith LG, Swartz MA. Capturing complex 3D tissue physiology in vitro. *Nat Rev Mol Cell Biol.* 2006; 7:211–224. [PubMed: 16496023]
2. Chung S, Sudo R, Vickerman V, Zervantonakis IK, Kamm RD. Microfluidic Platforms for Studies of Angiogenesis, Cell Migration, and Cell–Cell Interactions. *Ann Biomed Eng.* 2010; 38:1164–1177. [PubMed: 20336839]
3. Chung S, Sudo R, Mack PJ, Wan C-R, Vickerman V, Kamm RD. Cell migration into scaffolds under co-culture conditions in a microfluidic platform. *Lab Chip.* 2009; 9:269–275. [PubMed: 19107284]
4. Shin Y, Jeon JS, Han S, Jung G-S, Shin S, Lee S-H, Sudo R, Kamm RD, Chung S. In vitro 3D collective sprouting angiogenesis under orchestrated ANG-1 and VEGF gradients. *Lab Chip.* 2011; 11:2175–2181. [PubMed: 21617793]
5. Chung S, Sudo R, Zervantonakis IK, Rimchala T, Kamm RD. Surface - treatment-induced three-dimensional capillary morphogenesis in a microfluidic platform. *Adv Mater.* 2009; 21:4863–4867. [PubMed: 21049511]
6. Chrobak KM, Potter DR, Tien J. Formation of perfused, functional microvascular tubes in vitro. *Microvasc Res.* 2006; 71:185–196. [PubMed: 16600313]
7. Borenstein JT, Weinberg EJ, Orrick BK, Sundback C, Kaazempur-mofrad MR, Vacanti JP. Microfabrication of three-dimensional engineered scaffolds. *Tissue Eng.* 2007; 13:1837–1844. [PubMed: 17590149]

8. Shin M, King K, Matsuda K, Ishii O, Terai H, Weinberg E, Kaazempur-Mofrad M, Borenstein J, Detmar M, Vacanti J. Endothelialized networks with a vascular geometry in microfabricated poly(dimethyl siloxane). *Biomed Microdevices*. 2004; 6:269–278. [PubMed: 15548874]
9. Berthiera E, Beebe DJ. Flow rate analysis of a surface tension driven passive micropump. *Lab Chip*. 2007; 7:1475–1478. [PubMed: 17960274]
10. Sudo R, Chung S, Zervantonakis IK, Vickerman V, Toshimitsu Y, Griffith LG, Kamm RD. Transport-mediated angiogenesis in epithelialcoculture. *The FASEB Journal*. 2009; 23:2155–2164.
11. Jaasma MJ, Plunkett NA, O'Brien FJ. Design and validation of a dynamic flow perfusion bioreactor for use with compliant tissue engineering scaffolds. *J Biotechnol*. 2008; 133:490–496. [PubMed: 18221813]
12. Hegab HM, ElMekawy A, Stakenborg T. Review of microfluidic microbioreactor technology for high-throughput submerged microbiological cultivation. *Biomicrofluidics*. 2013; 7:021502.
13. Bonvin C B, Overney J, Shieh AC, Dixon JB, Swartz MA. A multichamber fluidic device for 3D cultures under interstitial flow with live imaging: development, characterization, and applications. *Biotechnol Bioeng*. 2010; 105:982–991. [PubMed: 19953672]
14. Nga CP, Helma C-LE, Swartz MA. Interstitial flow differentially stimulates blood and lymphatic endothelial cell morphogenesis in vitro. *Microvasc Res*. 2004; 68:258–264. [PubMed: 15501245]
15. Helm C-LE, Fleury ME, Zisch AH, Boschetti F, Swartz MA. Synergy between interstitial flow and VEGF directs capillary morphogenesis in vitro through a gradient amplification mechanism. *PNAS*. 2005; 102:15779–15784. [PubMed: 16249343]
16. Hsu Y-H, Moya ML, Abiri P, Hughes CCW, George SC, Lee AP. Full range physiological mass transport control in 3D tissue cultures. *LabChip*. 2013; 13:81–89.
17. Oh KW, Lee K, Ahn B, Furlani EP. Design of pressure-driven microfluidic networks using electric circuit analogy. *Lab Chip*. 2012; 12:515–545. [PubMed: 22179505]
18. Walker GM, Monteiro-Riviere N, Rousea J, O'Neill AT. A linear dilution microfluidic device for cytotoxicity assays. *Lab Chip*. 2007; 7:226–232. [PubMed: 17268625]
19. Lee K, Kim C, Ahn B, Panchapakesan R, Full AR, Nordee L, Kang JY, Oh KW. Generalized serial dilution module for monotonic and arbitrary microfluidic gradient generators. *Lab Chip*. 2009; 9:709–717. [PubMed: 19224022]
20. Jeon NL, Dertinger SKW, Chiu DT, Choi IS, Stroock AD, Whitesides GM. Generation of solution and surface gradients using microfluidic systems. *Langmuir*. 2000; 16:8311–8316.
21. Hung PJ, Lee PJ, Sabounchi P, Lin R, Lee LP. Continuous perfusion microfluidic cell culture array for high-throughput cell-based assays. *Biotechnol Bioeng*. 2005; 89:1–8. [PubMed: 15580587]
22. Kim C, Lee KS, Bang JH, Kim YE, Kim MC, Oh KW, Lee SH, Kang JY. 3-Dimensional cell culture for on-chip differentiation of stem cells in embryoid body. *Lab Chip*. 2011; 11:874–882. [PubMed: 21249238]
23. Moya ML, Hsu Y-H, Lee AP, Hughes CCW, George SC. In vitro perfused human capillary networks. *Tissue Eng PT C-method*. 2013 doi:10.1089/ten.TEC.2012.0430.
24. Horowitz, H.; Hill, W. *The Art of Electronics*. 2nd ed.. Cambridge University Press; New York, USA: 1995.
25. Griffith CK, Miller C, Sainson RCA, Calvert JW, Jeon NL, Hughes CCW, George SC. Diffusion limits of an in-vitro thick prevascularized tissue. *Tissue Eng*. 2005; 11:257–266. [PubMed: 15738680]
26. Cox ME, Dunn B. Oxygen diffusion in poly(dimethyl siloxane) using fluorescence quenching. I. measurement technique and analysis. *J Polym Sci A1*. 1986; 24:621–36.
27. Cho H, Kim H-Y, Kang JY, Kim TS. How the capillary burst microvalve works. *J Colloid Interf Sci*. 2007; 306:379–85.
28. Chen X, Aledia AS, Popson SA, Him L, Hughes CCW, George SC. Rapid anastomosis of endothelial progenitor cell-derived vessels with host vasculature is promoted by a high density of cotransplanted fibroblasts. *Tissue Eng Part A* 2010. 16:585–594.
29. Kamei M, Saunders WB, Bayless KJ, Dye L, Davis GE, Weinstein BM. Endothelial tubes assemble from intracellular vacuoles in vivo. *Nature*. 2006; 442:453–456. [PubMed: 16799567]




Fig. 1. Equivalent circuits of (A) a microchannel connected by two external loadings (R_{L1} & R_{L2}), (B) a long microchannel connected by n external loadings with equal pressure drop, and (C) the lumped resistors (R_e & e_i) between the i^{th} external loading.




Fig. 2. (A) Conceptual illustrations of the high impedance 3-D microtissue chamber, where cell construct is formed by fibrin gel seeded with ECs and NHFLs, (B) a picture of the microfluidic platform and media reservoirs.




Fig. 3. (A) Microfluidic configuration and the simulated pressure field of the multi-microtissue platform, (B) simulated velocity field and pattern of interstitial flow (white lines), (C) the microtissue platform, (D) & (E) the pressure field and velocity field of each microtissue chamber, (F) the distribution of dextran flowing into a microchamber, (G) one of the 12 developed human microtissues, where microvascular network and nuclei were labeled with CD31 (green) and DAPI (blue). (H) total vessel lengths of each microtissue chamber from three independent experiments.




Fig. 4. (A) Microfluidic configuration and the simulated pressure field of the multi-microtissue platform with a dominated hydraulic resistor R_{n+1} , (B) simulated velocity field and pattern of interstitial flow (white lines), (C) the microtissue platform, (D) & (E) the pressure field and velocity field of each microtissue chamber, (F) time-lapse distributions of dextran flowing into two microchamber and corresponding simulated results at 1, 100, and 240 min, (G) vacuoles found after 12-hour culture, and (H) total vessel lengths of each microtissue chamber from three independent experiments, (I) a fluorescent image of developed 5 human

microtissues, where microvascular network and nuclei were labeled with CD31 (green) and DAPI (blue).




Fig. 5. (A) A Fluorescent image of the microtissues developed by the high throughput microtissue platform, where the microvessels were stained with CD-31. (B) The high throughput microtissue platform, (C) and (D) are fluorescent images of two microtissues (4X magnification), where microvascular network and nuclei were labeled with CD31 (green) and DAPI (blue), (E) total vessel lengths of each microtissue chamber.