

Comparative study of MALDI-TOF MS and VITEK 2 in bacteria identification

Ling Guo, Liyan Ye, Qiang Zhao, Yanning Ma, Jiyong Yang, Yanping Luo

Department of Microbiology, General Hospital of PLA, Beijing 100853, China

Correspondence to: Yanping Luo. Department of Microbiology, PLA General Hospital, 28 Fuxing Road, Beijing 100853, China. Email: ypluo301@aliyun.com.

Background: Matrix-assisted laser desorption ionization-time of flight mass spectrometry (MALDI-TOF MS) has recently been introduced in diagnostic microbiology laboratories for the identification of bacterial and yeast strains isolated from clinical samples. This study aimed to evaluate the accuracy of MALDI-TOF MS in clinical microbiology diagnosis by comparing it with commonly-used VITEK 2 or gene sequencing.

Methods: The performances of MALDI-TOF MS and VITEK 2 were compared retrospectively for identifying routine isolates. Discrepancies were analyzed by gene sequencing analysis of the *16S* genes.

Results: For 1,025 isolates, classified as 55 species of 25 genera, 1,021 (99.60%) isolates were accurately identified at the genus level, and 957 (93.37%) isolates at the species level by using MALDI-TOF MS. A total of 949 (92.59%) isolates were completely matched by both methods. Both methods found 76 unmatched isolates among which one strain had no definite identification by MALDI-TOF MS and VITEK 2 respectively. However, MALDI-TOF MS made no errors at the genus level while VITEK 2 made 6 (0.58%) errors at the genus level. At the species level, the identification error rates for MALDI-TOF MS and VITEK 2 were 5.56% and 6.24%, respectively.

Conclusions: With a lower identification error rate, MALDI-TOF MS has better performance than VITEK 2 in identifying bacteria found routinely in the clinical laboratory. It is a quick and cost-effective technique, and has the potential to replace conventional phenotype methods in identifying common bacterial isolates in clinical microbiology laboratories.

Keywords: Matrix-assisted laser desorption ionization-time of flight mass spectrometry (MALDI-TOF MS); VITEK2; bacteria identification

Submitted Jan 18, 2014. Accepted for publication Feb 26, 2014.

doi: 10.3978/j.issn.2072-1439.2014.02.18

View this article at: <http://www.jthoracdis.com/article/view/2351/2939>

Introduction

Traditionally, bacterial and fungal identifications in clinical microbiology laboratories are mainly carried out according to phenotype characteristics, including identifications of culture media, colony morphology, gram stain and various biochemical reactions (1). Although all of these methods can achieve high accuracies, it usually takes minimum one day or longer to complete the whole identification process. Molecular methods, such as real-time PCR, gene sequencing and microarray analysis, are quick methods for bacterial and fungal identification, but they come at a very high cost and require highly-trained technicians. Therefore,

molecular methods are not routinely used for bacterial identification. A faster and easier technique for microbial identification would greatly enhance the conventional laboratory in providing more timely feedback to clinic. This is especially true in cases when patients are critically ill suffering from infectious diseases and where therapeutic intervention is urgently needed.

Matrix-assisted laser desorption ionization-time of flight mass spectrometry (MALDI-TOF MS) can be used to obtain protein fingerprinting from whole bacterial cells (2). Through comparing these fingerprints to a reference database by the use of various algorithms, bacteria can be rapidly identified. The earliest application of this technique

for bacterial identification dates back to 1975 (3), while the first related article was published in 1996 (Holland, *et al.*) (4). Studies in this field have been progressively advancing for the last decade (5). Until now, domestic comparative studies on mass spectrometry and other methods are relatively few. In order to compare the performance differences between MALDI-TOF MS and VITEK 2, the latter being the most-used automated identification technique in current microbiology laboratories, we employed the two systems in parallel to identify and analyze 1,025 isolates routinely isolated in 2012 at a microbiology laboratory of PLA General Hospital.

Methods

Bacteria isolates

A total of 1,025 isolates, composed of 1,020 bacterial strains and 5 fungal strains representing 25 different genera were selected for analysis. These bacteria were routinely isolated from clinical patients, such as *Pseudomonas spp.*, *Acinetobacter spp.* that cause lung infections, *Enterobacteriaceae* that cause blood and urinary tract infections. In addition, there are also some bacteria which grow slowly but have clinical significance such as *Eikenella corrodens*, *Listeria monocytogenes*, *Haemophilus influenzae*, etc.

VITEK 2 identification

GP, GN, YST, ANC and NH were selected to run identification analysis according to the different strains to be tested. The identification rate was expected to be above 93% by VITEK 2 (bioMérieux) (5), with quality control by stages via ATCC25922, ATCC27853 and ATCC25923.

MALDI-TOF MS identification

Strains were identified by MALDI-TOF MS using the VITEK MS system (bioMérieux). Loops were used to select and smear the subject isolates onto the sample spots on the target slide. Then 1 μ L VITEK MS-CHCA matrix was applied over the sample and air dried until the matrix and sample co-crystallized. The target slide with all prepared samples then was loaded into the VITEK MS system to acquire the mass spectra of whole bacterial cell protein, mainly composed of ribosomal protein, for each sample. In the case of yeasts, 0.5 μ L VITEK MS-FA was added to each sample on the target slide and allowed to air dry it

before adding 1 μ L VITEK MS-CHCA matrix. Finally, the mass spectra acquired for each sample were compared to the known mass spectra contained in the database for, and given a confidence score according to how close the acquired spectra matched those contained in the database. ATCC8739 was used as the quality control strain.

Gene sequencing

Discrepancies between VITEK 2 and MALDI-TOF MS were resolved by *16S rRNA* gene sequencing. The results would be considered valid if the homologous rate was above 99%.

Standards for identification

In cases where the isolate was correctly identified at the species level or at the generic level by both VITEK MS and VITEK 2, the results were considered to be in accordance with one another. In instances that the two methods produced different results, or if the results from one method were not definite, *16S rRNA* gene sequencing was used to resolve identification discrepancies. For isolates misidentified at the genus level, the results were considered a major mistake in the identification analysis. For isolates misidentified at species level but correctly identified at the genus level, the results were considered a minor mistake. Results not able to distinguish two different bacteria of the same genus also were considered as minor mistakes (6,7).

Results

Identification results by MALDI-TOF MS

Among the 1,025 total isolates, 1,021 (99.60%) isolates were accurately identified at the genus level and 957 (93.37%) isolates at the species level by MALDI-TOF MS.

Matched results by the two identification methods

Identification results for 949 (92.59%) isolates, belonging to 23 genera and 48 species, were in agreement (see *Table 1*) using both methods.

Unmatched results by the two identification methods

Among the 1,025 isolates, 76 (7.4%) isolates produced discordant results between the two identification methods. One strain had no definitive identification for each

Table 1 Matched results by both MALDI-TOF MS and VITEK 2

Bacterial genera	Number of strains
<i>Acinetobacter baumannii</i> complex	30
<i>Corynebacterium jeikeium</i>	1
<i>Citrobacter braakii</i>	2
<i>Citrobacter freundii</i>	6
<i>Citrobacter koseri</i>	1
<i>Escherichia coli</i>	297
<i>Enterobacter aerogenes</i>	13
<i>Enterococcus casseliflavus</i>	2
<i>Enterococcus faecalis</i>	10
<i>Enterococcus faecium</i>	24
<i>Enterococcus gallinarum</i>	1
<i>Enterococcus hirae</i>	3
<i>Klebsiella oxytoca</i>	10
<i>Klebsiella pneumoniae</i>	66
<i>Micrococcus ssp</i>	3
<i>Morganella morganii</i>	7
<i>Pseudomonas aeruginosa</i>	126
<i>Pseudomonas putida</i>	1
<i>Stenotrophomonas maltophilia</i>	33
<i>Proteus mirabilis</i>	75
<i>Proteus vul.gr/proteus pen.</i>	9
<i>Providencia rettgeri</i>	2
<i>Staphylococcus aureus</i>	76
<i>Staphylococcus epidermidis</i>	36
<i>Staphylococcus hominis</i>	37
<i>Staphylococcus saprophyticus</i>	1
<i>Staphylococcus haemolyticus</i>	14
<i>Staphylococcus cohnii</i>	1
<i>Staphylococcus capitis</i>	5
<i>Serratia marcescens</i>	23
<i>Streptococcus parasanguinis</i>	1
<i>Streptococcus pneumoniae</i>	4
<i>Streptococcus mitis/Streptococcus oralis</i>	3
<i>Streptococcus gallolyticus</i>	1
<i>Streptococcus gordonii</i>	1
<i>Streptococcus salivarius</i>	1
<i>Streptococcus agalatae</i>	5
<i>Streptococcus sanguinis</i>	1
<i>Shewanella algea</i>	3
<i>Vibrio parahaemolyticus</i>	1
<i>Candida parapsilosis</i>	2
<i>Aeromonas</i> sp.	3
<i>Chryseobacterium indologenes</i>	1
<i>Eikenella corrodens</i>	1
<i>Salmonella</i> group	2
<i>Listeria monocytogenes</i>	1
<i>Haemophilus influenzae</i>	1
<i>Burkholderia cepacia</i>	1
Total	949

MALDI-TOF MS and VITEK 2. However, MALDI-TOF MS made no errors at the genus level while VITEK 2 made 6 (0.58%) errors at the genus level. At the species level, the identification error rates of the two methods were 5.56% and 6.24% for MALDI-TOF MS and VITEK 2 respectively (Table 2).

Discussion and Conclusions

In our laboratory, MALDI-TOF MS is an efficient, quick and relatively inexpensive per isolate method for identifying pathogenic microorganisms including bacteria and fungi. MALDI-TOF MS identification also is largely compatible with a large range of culture media and culture conditions, and is the fastest means to detect microbes in positive blood culture (8). Furthermore, MALDI-TOF MS has broader application prospects in the field of testing for drug resistance (9,10).

Shortening the turnaround time to pathogen identification and offering quicker results to the clinic remain the most important issues needed to be solved in the microbiology laboratory. Based on our data, experienced clinicians can treat critically infected patients appropriate antibiotic treatment given effective and timely identification of pathogenic bacteria. Shortening the turnaround time permits the clinician to treat patients faster, ultimately reducing the fatality rates, the length of patient stay in the clinic, and healthcare costs associated with patient care. MALDI-TOF MS offers one promising solution to shorten the turnaround time and relieve these pressures in the clinic.

In the present study, we chose 1,025 pathogenic bacteria routinely found in our laboratory for identification, and compared the performance of MALDI-TOF MS to the VITEK 2 system, the latter which is based on conventional biochemical identification that is presently found in most microbiology laboratories. Our results showed that MALDI-TOF MS offered higher identification accuracy and lower error rates at the species level when compared to VITEK 2. Additionally, MALDI-TOF MS dramatically shortened identification time from 6-8 hours to just a few minutes.

MALDI-TOF MS identified 99.60% of isolates to the genus level and 93.37% of isolates to the species level, which are slightly higher rates than those previously reported (5,7,11-14). Van Veen *et al.*, for example, achieved accuracy rates of 97.1% and 92% to the genus and species levels, respectively. The difference in accuracy is most likely due to the different choices of strains. The strains in van Veen's study included 61 yeast strains, among which 96.7%

Table 2 Unmatched results by both MALDI-TOF MS and VITEK 2

Bacterial species	Number of strains	VITEK MS		VITEK 2	
		Generic errors	Species errors	Generic errors	Species errors
<i>Acinetobacter baumannii</i>	2			2	
<i>Acinetobacter jonnosonii</i>	2			2	
<i>Citrobacter freundii</i>	1		1		
<i>Enterobacter cloacae</i>	46		46		46
<i>Enterococcus faecium</i>	1		1		
<i>Pseudomonas putida</i>	1				1
<i>Pseudomonas aeruginosa</i>	1		1		
<i>Staphylococcus hominis</i>	4				4
<i>Staphylococcus aureus</i>	4			1	3
<i>Staphylococcus epidermidis</i>	1				1
<i>Streptococcus agalactiae</i>	2				2
<i>Streptococcus mitis</i>	3		3		3
<i>Candida guilliermondii</i>	1				1
<i>Rhodotorula mucilaginosa</i>	1		1		1
<i>Achromobacter xylosoxidans</i>	3		3		
<i>Chryseindologenes</i>	1		1		1
<i>Listeria monocytogenes</i>	1				1
<i>Cryptococcus neoformans</i>	1			1	
Total	76	0 (0)	57 (5.56%)	6 (0.58%)	6 (6.24%)

were identified to the genus level and 82.5% to the species level. *Streptococcus viridans*, *streptococcus pneumoniae* and anaerobic bacteria, all difficult pathogens to identify notably were included in van Veen's study and not in our study. Most of the strains analyzed in our study were largely more commonly found pathogens, and the construction of the MALDI-TOF MS database may offer higher identification accuracies for these pathogens.

In conclusion, MALDI-TOF MS is a rapid, easy, relatively inexpensive and high-throughput method for identifying clinically relevant bacteria and fungi. MALDI-TOF MS offers very high accuracy, often higher than conventional methods, in identifying common microorganisms. Due to these key advantages and with continued development of MALDI-TOF MS technology and its clinical knowledgebase, MALDI-TOF MS will become a rapid, routine method for identifying pathogenic bacteria found in the conventional clinical microbiology laboratory.

Acknowledgements

Disclosure: The authors declare no conflict of interest.

References

1. Carroll KC, Weinstein MP. Manual and automated systems for detection and identification of microorganisms. In: Murray PR, Baron EJ, Jorgensen JH, et al. eds. Manual of clinical microbiology. 9th ed. Washington, DC: American Society for Microbiology Press, 2007:192-217.
2. Fenselau C, Demirev PA. Characterization of intact microorganisms by MALDI mass spectrometry. *Mass Spectrom Rev* 2001;20:157-71.
3. Anhalt JP, Fenselau C. Identification of bacteria using mass spectrometry. *Anal Chem* 1975;47:219-25.
4. Holland RD, Wilkes JG, Rafii F, et al. Rapid identification of intact whole bacteria based on spectral patterns using matrix-assisted laser desorption/ionization with time-of-flight mass spectrometry. *Rapid Commun Mass Spectrom* 1996;10:1227-32.
5. Seng P, Drancourt M, Gouriet F, et al. Ongoing revolution in bacteriology: routine identification of bacteria by matrix-assisted laser desorption ionization time-of-flight mass spectrometry. *Clin Infect Dis* 2009;49:543-51.
6. Bizzini A, Durussel C, Bille J, et al. Performance of

- matrix-assisted laser desorption ionization-time of flight mass spectrometry for identification of bacterial strains routinely isolated in a clinical microbiology laboratory. *J Clin Microbiol* 2010;48:1549-54.
7. van Veen SQ, Claas EC, Kuijper EJ. High-throughput identification of bacteria and yeast by matrix-assisted laser desorption ionization-time of flight mass spectrometry in conventional medical microbiology laboratories. *J Clin Microbiol* 2010;48:900-7.
 8. Drancourt M. Detection of microorganisms in blood specimens using matrix-assisted laser desorption ionization time-of-flight mass spectrometry: a review. *Clin Microbiol Infect* 2010;16:1620-5.
 9. Burckhardt I, Zimmermann S. Using matrix-assisted laser desorption ionization-time of flight mass spectrometry to detect carbapenem resistance within 1 to 2.5 hours. *J Clin Microbiol* 2011;49:3321-4.
 10. Hrabák J, Walková R, Studentová V, et al. Carbapenemase activity detection by matrix-assisted laser desorption ionization-time of flight mass spectrometry. *J Clin Microbiol* 2011;49:3222-7.
 11. Blondiaux N, Gaillot O, Courcol RJ. MALDI-TOF mass spectrometry to identify clinical bacterial isolates: evaluation in a teaching hospital in Lille. *Pathol Biol (Paris)* 2010;58:55-7.
 12. Dupont C, Sivadon-Tardy V, Bille E, et al. Identification of clinical coagulase-negative staphylococci, isolated in microbiology laboratories, by matrix-assisted laser desorption/ionization-time of flight mass spectrometry and two automated systems. *Clin Microbiol Infect* 2010;16:998-1004.
 13. Dubois D, Leyssene D, Chacornac JP, et al. Identification of a variety of Staphylococcus species by matrix-assisted laser desorption ionization-time of flight mass spectrometry. *J Clin Microbiol* 2010;48:941-5.
 14. Spanu T, De Carolis E, Fiori B, et al. Evaluation of matrix-assisted laser desorption ionization-time-of-flight mass spectrometry in comparison to rpoB gene sequencing for species identification of bloodstream infection staphylococcal isolates. *Clin Microbiol Infect* 2011;17:44-9.

Cite this article as: Guo L, Ye L, Zhao Q, Ma Y, Yang J, Luo Y. Comparative study of MALDI-TOF MS and VITEK 2 in bacteria identification. *J Thorac Dis* 2014;6(5):534-538. doi: 10.3978/j.issn.2072-1439.2014.02.18