



Published in final edited form as:

Acta Biomater. 2014 July ; 10(7): 2935–2944. doi:10.1016/j.actbio.2014.03.005.

Near-Infrared Fluorescence Imaging as an Alternative to Bioluminescent Bacteria to Monitor Biomaterial-Associated Infections

Nina Dinjaski¹, Suri Shalu², Jaione Valle³, Susan M. Lehman², Iñigo Lasa³, María Auxiliadora Prieto¹, and Andrés J. García^{2,*}

¹Centro de Investigaciones Biológicas, CSIC, Madrid 28040, Spain

²Woodruff School of Mechanical Engineering and Petit Institute for Bioengineering and Bioscience, Georgia Institute of Technology, Atlanta 30332-0363, USA

³Instituto de Agrobiotecnología, UPNA-CSIC-Gobierno de Navarra, 31006 Pamplona, Spain

Abstract

Biomaterial-associated infection is one of the most common complications related with the implantation of any biomedical device. Several *in vivo* imaging platforms have emerged as powerful diagnostic tools to longitudinally monitor biomaterial-associated infections in small animal models. In this study, we directly compared two imaging approaches: bacteria engineered to produce luciferase to generate bioluminescence and reactive oxygen species (ROS) imaging of the inflammatory response associated with the infected implant. We performed longitudinal imaging of bioluminescence associated with bacteria strains expressing plasmid-integrated luciferase driven by different promoters or a strain with the luciferase gene integrated into the chromosome. These luminescent strains provided adequate signal for acute (0–4 days) monitoring of the infection, but the bioluminescence signal decreased over time and leveled off by 7 days post-implantation. This loss in bioluminescence signal was attributed to changes in the metabolic activity of the bacteria. In contrast, near-infrared fluorescence imaging of ROS associated with inflammation to the implant provided sensitive and dose-dependent signals of biomaterial-associated bacteria. ROS imaging exhibited higher sensitivity than the bioluminescence imaging and was independent of the bacteria strain. Near-infrared fluorescence imaging of inflammatory responses represents a powerful alternative to bioluminescence imaging for monitoring biomaterial-associated bacterial infections.

© 2014 Acta Materialia Inc. Published by Elsevier Ltd. All rights reserved.

*Corresponding author: Woodruff School of Mechanical Engineering and Petit Institute for Bioengineering and Bioscience, Georgia Institute of Technology, 2310 IBB Building, 315 Ferst Drive, Atlanta 30332-0363, USA. Fax: 404-385-1397. andres.garcia@me.gatech.edu (A. J. García).

Publisher's Disclaimer: This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final citable form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

Keywords

Biomaterial-associated infection; Bioluminescence; Near infrared Fluorescence; Noninvasive monitoring; *Staphylococcus aureus*

1. Introduction

Device-related bacterial infections are a growing healthcare problem [1–3], accounting for more than 50% of the 2,000,000 annual hospital-acquired infections associated with indwelling devices and implants in the United States [2]. *Staphylococcus aureus* is one of the most common pathogens associated with these cases. Bacterial colonization and biofilm development can lead to both malfunction of the device and systemic infection, since biofilms are complex cooperative communities, and such biofilm bacteria are nearly impervious to antimicrobial therapy or host defense mechanisms [4,5]. In most cases, the affected devices must be removed to eliminate the infection, given the fact that currently there are no drugs that specifically target bacteria in biofilms [6–8].

A requirement to efficiently treat implant-associated infections are *in vivo* monitoring approaches that allow better understanding and control of biofilm formation, together with novel methods for targeting efficient drug candidates [9]. Optical imaging of bacterial infections *in vivo* using engineered bioluminescent bacterial strains is a widely used approach for spatial and temporal assessment of the infection [10]. This method is based on bioluminescent bacteria expressing a luciferase-based reporter system that emits light that can be monitored longitudinally and nondestructively in the same animal.

In view of the fact that biomaterial-associated infections modulate the inflammatory response to the biomaterial, changes in inflammatory markers may be used to improve monitoring of an ongoing infection [11]. In particular, reactive oxygen species (ROS) form part of the oxygen-dependent bactericidal mechanisms that phagocytic cells employ [12]. Near-infrared fluorescence (NIRF) imaging probes, such as hydrocyanines, allow real-time fluorescence imaging and ROS detection in the vicinity of an implant [13]. Moreover, NIRF imaging is an excellent noninvasive method for whole-body scanning that can determine the extent of the infectious disease throughout the body, especially in clinically challenging cases involving trauma, infection, and compromised tissue beds.

Herein, we directly compared two imaging approaches of implant-associated infection: bacteria engineered to produce luciferase to generate bioluminescence and ROS imaging of the inflammatory response associated with the infected implant. These approaches were correlated to bacterial counts before and after 7 days of implantation (Fig. 1).

2. Materials and methods

2.1. Disk fabrication

Poly(3-hydroxyoctanoate-*co*-hydroxyhexanoate), PHOHHx, was kindly provided by Bioplolis S.L. The monomer composition of PHOHHx was determined by gas chromatography-mass spectrometry (GC-MS) as previously described [14] and consisted in

8.5% 3-hydroxyhexanoate (OH-C6) and 91.5% 3-hydroxyoctanoate (OH-C8). An optimized downstream processing was applied to eliminate endotoxins as previously described [15]. Briefly, 1 g of PHOHHx was dissolved in 100 mL of chloroform at 40 °C under vigorous stirring, the suspension was pressure filtered and the polymer was precipitated by addition of non-solvent methanol. Finally, the polymer was dried under vacuum at 40 °C for 48 h. This procedure was repeated two times to obtain PHOHHx with endotoxin units (EU) <20 EU g⁻¹, in compliance with the endotoxin requirements for biomedical applications (FDA) [16]. The endotoxin content was measured using a Limulus amoebocyte lysate (LAL)-test (Pyrogen Plus Single Test Kit, Lonza) and the endotoxin content was determined to be <15 EU g⁻¹.

Poly(ethylene terephthalate) (PET) disks (6 mm diameter) were coated with PHOHHx by solvent-casting. PHOHHx dissolved in chloroform (2% w/v) was applied over sterile, endotoxin-free PET disks (kindly supplied by ACCIONA, Barcelona, Spain), in a dust-free atmosphere. The coatings were allowed to dry for 72 h at room temperature and the resulting PHOHHx-coated disks (referred to hereafter as PHOHHx disks) were sterilized with ethylene oxide at 40 °C.

2.2. Bacterial strains, media and growth conditions

The bacterial strains used throughout this study were *Staphylococcus aureus* subsp. *aureus* ATCC 12600, its two derivative luminescent strains *S. aureus* (pAmiBlaz) and *S. aureus* (pAmiSPA), and the *S. aureus* Xen29 strain containing *lux* operon stably inserted into the chromosome (Caliper Life Sciences, PerkinElmer Company). All bacterial strains were pre-cultured in trypticase soy agar (TSA) plates and incubated at 37 °C for 24 h. The appropriate selection antibiotics, chloramphenicol (10 µg mL⁻¹) or kanamycin (200 µg mL⁻¹), were added when indicated. Trypticase soy broth (TSB, Difco) was used as the growth medium to culture all bacterial pathogens.

2.3. Construction of *S. aureus* luminescent strains

Bioluminescent *S. aureus* strains were generated by transforming ATCC 12600 strain with a modified *Photobacterium luminescens luxCDABE (lux)* gene cluster using the pAmiBlaz or pAmiSPA plasmids. To construct the vectors, *blaZ* (β-lactamase) and *spa2* (protein A) promoters were inserted into promoterless-*lux* cloning vector pAmilux [17] to yield pAmiBlaz or pAmiSPA plasmid, respectively. Vectors were introduced into the cells by electroporation as previously described [18]. Transformants were selected on TSA plates containing chloramphenicol (10 µg mL⁻¹). Successful transformation was confirmed by bioluminescent colonies screening using an IVIS Lumina bioimaging system (Xenogen).

Expression of *lux* operon in pAmiBlaz vector was driven by the *BlaZ* promoter, whereas in pAmiSPA the *lux* operon was controlled by the protein A promoter. These two promoters were used with the aim to generate different expression patterns. The *BlaZ* promoter was used for constitutive expression of luciferase. In contrast, the *S. aureus* protein A is involved in the development of biofilm-associated infections [19]; therefore the *spa* promoter drives the expression of luciferase during biofilm development. *S. aureus* Xen29 (Caliper,

PerkinElmer) is a commonly used, commercially available strain containing luciferase construct stably integrated into the chromosome.

2.4. Preparation of hydro-indocyanine green

Hydro-indocyanine green (H-ICG) was synthesized from indocyanine green dye (ICG) (Acros Organics) by reduction with sodium borohydride as described [13]. Briefly, 2 mg of ICG was dissolved in 4 mL of methanol and reduced with 2–3 mg of sodium borohydride (Aldrich). Solvent was removed by stirring reaction mix for 5 min under reduced pressure. The dye was nitrogen capped and stored overnight at $-20\text{ }^{\circ}\text{C}$.

2.5. Sample preparation, implantation and bioimaging

To test the biofilm formation of the *S. aureus* parental strain and its bioluminescent derivatives *in vivo*, bacterial strains were first cultivated overnight in TSB with aeration at $37\text{ }^{\circ}\text{C}$. The cultures were diluted with fresh TSB to reach absorbance $A_{595} = 0.1$ and incubated for 3–4 h to $A_{595} = 0.7$. Afterwards, bacterial test inocula were prepared in 2 mL of phosphate-buffered saline, pH 7.2 (PBS), to serve as a cell suspension. Bacterial suspensions (3.5×10^2 to 3.5×10^8 CFU/mm²) were placed on sterile endotoxin-free PHOHHx disks. Prior to implantation, disks pre-colonized with bacteria were placed in sterile containers and incubated 30 min under static conditions at $37\text{ }^{\circ}\text{C}$.

National Institutes of Health (NIH) guidelines for the care and use of laboratory animals were followed [20]. All surgical procedures were approved by the Institutional Animal Care and Use Committee at the Georgia Institute of Technology. Sterile, endotoxin-free disks as well as bacteria pre-colonized disks were implanted subcutaneously in the back of 6–8 weeks old male BALB/c mice (Jackson Laboratories) anesthetized by isoflurane. A single 1 cm incision was made on the dorsum proximal to the spine, and a subcutaneous pocket laterally spanning the dorsum was created. Sterile disks (two per mouse, one on either side of the spine) were implanted, and the incision was closed using sterile wound clips. For analysis of each experimental group, 3 or more mice were imaged.

For bioimaging of luminescent bacteria, mice were anesthetized with isoflurane and imaged with a CCD camera (IVIS Lumina[®] bioimaging system, Xenogen) directly following implantation and 1, 4, 7 days post implantation. Bioluminescence was integrated using Living Image[®] software Version 3.1 (Xenogen). Total counts from *S. aureus* were collected during a 2 min exposure using the IVIS Imaging System and Living Image software (Xenogen Corporation). Bioluminescence images were displayed using a pseudo-color scale (blue representing the least-intense light and red representing the most-intense light) that was overlaid on a gray-scale image to generate a two-dimensional image of the distribution of bioluminescent bacteria in the animal. To account for the background luminescence, one uninfected mouse was imaged along with the infected animals. The total counts from a region were quantified using the Living Image software package (Xenogen Corporation), and the data are presented as total counts contained within each region.

For bioimaging of ROS, 30 μL of H-ICG at a concentration of 1 mg mL^{-1} in sterile water was injected near the vicinity of the implant as described [13]. The same animals used to

image bioluminescence were used to evaluate NIRF signals. Thirty minutes after dye injection, the whole body of the animal was scanned in an IVIS Lumina[®] bioimaging system (Xenogen). Biofluorescence was integrated using Living Image[®] software Version 3.1 (Xenogen). ROS bioimaging was performed 7 days post-implantation.

2.6. Explant analysis

After euthanasia, disks were carefully explanted with the intact surrounding tissue to avoid disrupting the cell-material interface. For immunostaining, explants were embedded in optimal cutting temperature compound (Tissue-Tek) and cryosectioned at 10 μm . Sections were fixed in 4% paraformaldehyde and stained with monoclonal antibodies (Abcam) against macrophage (CD68) or neutrophil (NIMP-R14) markers. Alternatively, samples were incubated with monoclonal antibodies against *S. aureus* (Abcam). AlexaFluor 488-conjugated goat antibody (Invitrogen) was used as a secondary antibody. The sections were mounted with antifade mounting media containing 4',6-diamidino-2-phenylindole (DAPI, Vector Labs) and imaged under a Nikon C1 confocal microscope system. Five-six fields per sample were acquired and ImageJ software was used to count the fluorescently labeled cells.

For CFU bacteria counting, each explant was placed in a glass tube containing 1 mL of PBS and sonicated for 10 min in an ultrasonic bath to remove adhered bacteria. Afterwards, two more sonicating cycles were applied (5 min and 30 s) interspersed with 30 s of vortexing. Serial dilutions were plated on TSA plates supplemented with kanamycin (50 $\mu\text{g mL}^{-1}$) to determine the number of viable *S. aureus* Xen29 or chloramphenicol (10 $\mu\text{g mL}^{-1}$) to determine the number of *S. aureus* carrying pAmiBlaZ or pAmiSPA. CFU were determined after 24 h of incubation at 37 $^{\circ}\text{C}$.

2.7. Statistical analysis

Statistical analysis was performed by two-way ANOVA using Tukey post-hoc test with P 0.05 considered significant. Pair-wise comparisons were performed using Bonferroni post-hoc test with P 0.05 considered significant.

3. Results

3.1. Bioluminescent *S. aureus* permits short-term monitoring of Biomaterial-Associated Infections

To monitor infection profile in an *in vivo* murine model, PHOHHx polymer disks were loaded with bioluminescent *S. aureus* strains carrying different genetic configurations of the *lux* operon. *S. aureus* (pAmiBlaz) and *S. aureus* (pAmiSPA) contained the luciferase construct in an antibiotic-selective plasmid, whereas *S. aureus* Xen29 contained the luciferase gene integrated into the bacterial chromosome.

PHOHHx disks pre-colonized with different inocula of *S. aureus* luminescent strains were implanted subcutaneously in BALB/c mice and bioluminescent signal intensity was monitored over a 7-day period. For infected implants, the bioluminescence signal was significantly higher than readings for sterile PHOHHx implants (Fig. 2a,b). Bioluminescent signal for *S. aureus* (pAmiBlaz) and *S. aureus* (pAmiSPA) strains was detected only in a

local site with high bacterial densities ($3.5 \times 10^6 - 3.5 \times 10^8$ CFU/mm²), while lower bacterial densities did not emit detectable signal at any time point (Fig. 2a). The higher inocula, 3.5×10^7 and 3.5×10^8 CFU/mm², of both strains had concentration-dependent increases in bioluminescent signals and showed maximal value 1 h post-implantation (Fig. 2b). Differences in bioluminescence levels due to differences in promoter activity were evident. Over the next 4 days post-implantation, bioluminescent signal progressively decreased for both *S. aureus* (pAmiBlaZ) and *S. aureus* (pAmiSPA) strains, and only the highly expressed *spa* promoter provided detectable signal. At day 7 post-implantation, the bioluminescence signal for all strains was equivalent, although the signal was higher than sterile controls (Fig. 2b).

Because bacterial luciferase is an energy-requiring oxygenase and as such a reporter of cell metabolic activity, it is likely that during the stationary phase of bacterial growth the intensity of luminescent signal is related to the cell metabolic activity rather than to the promoter activity. Additionally, the drop-off in signal could possibly be due to loss of the plasmid and/or reductions in bacteria numbers. The correlation between signal intensity and the number of bacteria containing plasmid was investigated by growing *S. aureus* (pAmiBlaz) and *S. aureus* (pAmiSPA) cells on the plates containing selective marker. CFU of bacteria carrying pAmiBlaz or pAmiSPA prior to implantation was compared with CFU of bacteria harboring plasmid up on disk retrieval on day 7. A decreased number of bacteria carrying plasmid was clearly observed (Fig. 3a) when grown on selective plates (see Materials and methods section for details), indicating a possible cause for the loss of bioluminescence signal.

To examine whether the stable integration of the *lux* operon into the bacterial chromosome overcomes this limitation of the *S. aureus* (pAmiBlaz) and *S. aureus* (pAmiSPA) strains carrying the luciferase construct in a plasmid, we tested the *S. aureus* Xen29 strain. Using *S. aureus* Xen29, we were able to monitor the infection in the experimental model applying lower bacterial inoculums (3.5×10^3 and 3.5×10^4 CFU/mm²) (Fig. 2a,b). Following implantation of pre-colonized disks, the bioluminescence measurements increased exponentially over 24 h and peaked on day 1 post-implantation (approximately 5.5 log (luminescence-total counts/disk) for bacterial inoculum 3.5×10^4 CFU/mm² and 4.8 log (luminescence-total counts/disk) for bacterial inoculum 3.5×10^3 CFU/mm²). The high signal for Xen29 compared to the other strains is attributed to differences in the promoter and construct configuration. However, similar to our observations for the strains carrying the luciferase plasmid, the bioluminescence signal decreased at day 4 post-implantation and reached equivalent levels as the other bioluminescent strains by day 7 (Fig. 2b). Bacteria counts indicated significant loss in the number of viable bacteria at explant (Fig. 3b); this loss in viable bacteria is mostly likely due to an inflammatory response and accounts for the loss in bioluminescence signal. Taken together, these results demonstrate that bacteria strains expressing luciferase can be used to image infection short-term but none of the strains tested was suitable for monitoring chronic infections above a bacteria threshold ($3.5 \log$ CFU/mm²) in this model of biomaterial-associated infection.

3.2. NIRF imaging of infection-associated inflammation

We next applied NIRF imaging as a tool to examine biomaterial-associated infections using hydrocyanine dyes as an alternative to bioluminescent bacteria. Because inflammation patterns are a key correlate of the presence of an infection [21,22], the NIRF ROS sensor H-ICG was used to measure the levels of local inflammation associated with the infected biomaterial. Based on previous studies where biomaterial associated inflammation was longitudinally monitored using H-ICG [13,23], we observed the clearest evidence of NIRF signal at 7 days. Therefore, at 7 days post-implantation, H-ICG was injected in the vicinity of infected PHOHHx disks with a low bacteria inoculum (3.5×10^3 or 3.5×10^4 CFU/mm²) and animals were imaged in order to directly compare the two imaging modalities, NIRF and bioluminescence (Fig. 4). Increases in bacteria dose were clearly detected by imaging ROS via H-ICG signal, and these levels were significantly higher than the ROS signal for inflammation associated with sterile implants (Fig. 4b). Importantly, the H-ICG signal was independent of the bacteria strain used. In contrast, these lower bacteria densities were undetectable by imaging of bioluminescence (Fig. 5). NIRF imaging of infection-related inflammation using H-ICG provided readings that correlated with bacterial concentration independently of the *S. aureus* strain used (Fig. 4b; Fig. 5).

3.4. ROS signal correlate with inflammatory cell recruitment of infected biomaterials

We analyzed macrophage (CD68) and neutrophil (NIMP-R14) recruitment to the implant at day 7 post-implantation by immunostaining (Fig. 6a–f). Both macrophages and neutrophils were recruited to the implant. Quantification of the number of cells staining for these markers revealed an increasing trend in neutrophil and macrophage recruitment with increased bacteria dose (Fig. 6g).

Sections were also stained for *S. aureus* using a commercial antibody. Grape-like clusters, characteristic for *Staphylococcus* species, stained positive for *S. aureus* were widely present in all samples that were pre-colonized by bacteria prior to implantation (Fig. 7). Bacterial cells were observed only in the vicinity on contaminated implants and not surrounding tissues (SFig. 1), demonstrating the pattern of localized infection. Additionally, initial signs of tissue damage around infected implants became evident with increasing number of bacteria used to pre-colonize disks. Tissue necrosis was present for inoculums of 3.5×10^8 CFU/mm² or higher. DAPI-stained nuclei lost sharp borders and were completely destroyed (Fig. 7d,h). This phenomenon was observed only in the near vicinity of contaminated implant, whereas the cells of tissue located farther from the implant did not show signs of cell death (SFig. 1).

4. Discussion

Implantation of biomedical devices facilitates infection, since the biomaterial provides a surface for bacterial colonization and biofilm formation. Upon implantation, proteins and other biomacromolecules immediately coat the device and promote bacteria adhesion [24,25]. *S. aureus* harbors numerous cell wall-bound surface proteins that contain binding domains for mammalian proteins [26]. Early detection of these infections prior to formation of a recalcitrant biofilm is of great clinical importance. Hence, there is compelling need for

the development of new sensitive diagnostic techniques for biomaterial-associated infections.

Conventional methodologies for monitoring pathogens *in vivo* are cumbersome and include biological assessment such as biopsies, biochemical and genetic testing. Therefore, we focused on developing a minimally-invasive imaging strategy for *in vivo* monitoring of bacterial infection on polymeric implants. In this study, we directly compared two imaging approaches: bacteria engineered to produce luciferase to generate bioluminescence and ROS imaging of the inflammatory response associated with the infected implant. Bioluminescent bacteria strains are widely used in the field to study the progression of a bacterial infection. We performed longitudinal imaging of bioluminescence associated with bacteria strains expressing luciferase plasmids driven by different promoters or a commercially available strain with the luciferase gene integrated into the chromosome. These luminescent strains provided excellent signal for acute (0–4 days) monitoring of the infection, but the bioluminescence signal decreased over time and leveled off by 7 days post-implantation (Fig. 2). We attribute this loss in bioluminescence signal primarily to changes in the metabolic activity of the bacteria. Because biofilm-associated bacteria mostly exist in a stationary phase-like state where transcription and translation are markedly reduced [4], we expect for luciferase expression to decrease as bacteria form biofilms. Notably, in one of the strains evaluated, luciferase was driven by the promoter of the *spa2* gene responsible for expression of a *S. aureus* surface protein synthesized during biofilm formation [20]. However, this construct did not increase the intensity of bioluminescence signal. Additionally, the decrease of number of bacteria carrying plasmid contributed to the loss of bioluminescence signal. The loss in bioluminescence signal was even observed when using a *S. aureus* strain carrying the *lux* operon stably integrated in the chromosome. In addition to the signal loss attributed to changes in metabolic activity in the bacteria, the intrinsic blue-green spectral output of *lux* ($\lambda_{\text{max}} = 425 \text{ nm}$) limited tissue penetration and the detection limit of this approach. Importantly, the magnitude and kinetics of luciferase expression were dependent on the specifics of the promoter and gene construct. Bioluminescence monitoring has been used as a tool to validate the efficiency of antibacterial treatment that implies short term screening [27,28]. Importantly, it has been reported that bioluminescence imaging could monitor chronic infection, but is strongly related to the infection model and bacterial strain used [29,30].

As an alternative to bioluminescence imaging of luciferase-expressing bacteria, we evaluated NIRF imaging of ROS generated by the inflammation associated with bacterial infection. NIRF imaging offers excellent characteristics for optical imaging enabling deeper tissue penetration and sensitivity [31]. Indeed, several NIRF probes for *in vivo* imaging of bacterial infections have been reported [32–38]. The sensing mechanism for these probes is based on metabolic conversion of probe inside bacteria or molecules with high affinity for bacterial membrane proteins. As discussed previously, the metabolic activity in many bacterial species vary among planktonic and biofilm states, therefore limiting the wide applicability of these probes. In addition, many of these probes have not been validated for imaging biofilm associated with a biomaterial. In contrast, we show that the ROS sensor H-ICG could provide sensitive and dose-dependent signals of biomaterial-associated bacteria.

In addition, ROS imaging exhibited higher sensitivity than the bioluminescence imaging. Importantly, the ROS signal was independent of the bacteria strain (Fig. 4); this is a major advantage over bioluminescence imaging because it does not require the use of a luciferase-expressing bacteria strain. Additional characterization with other bacterial species and biomaterials is necessary to fully establish NIRF imaging of inflammatory responses as an effective strategy to monitor biomaterial-associated infections. Although the ROS signal for bacteria-colonized implants was significantly higher than the signal for sterile implants, the ROS probe would require calibration to discriminate between biofilm-containing and sterile implants, and this calibration may vary significantly due to variability among patient, device, implant location, and biofilm characteristics. Nevertheless, NIRF imaging of inflammatory responses represents a powerful alternative to bioluminescence imaging for monitoring biomaterial-associated bacterial infections in animal models.

Supplementary Material

Refer to Web version on PubMed Central for supplementary material.

Acknowledgments

We thank Biópolis S.L. for the supplied polymer. This work was supported by the Ministerio of Economía y Competitividad (BIO2010-21049, 201120E092), the U.S.A. National Institutes of Health grant R21 AI094624 (A.J.G.), the Georgia Tech/Emory Center for the Engineering of Living Tissues, the Atlanta Clinical and Translational Science Institute under PHS Grant UL RR025008 from the Clinical and Translational Science Award Program.

References

- Zimmerli W, Trampuz A, Ochsner PE. Prosthetic-joint infections. *N Engl J Med*. 2004; 351:1645–1654. [PubMed: 15483283]
- Darouiche RO. Treatment of infections associated with surgical implants. *N Engl J Med*. 2004; 350:1422–1429. [PubMed: 15070792]
- Bryers JD. Medical biofilms. *Biotechnol Bioeng*. 2008; 100:1–18. [PubMed: 18366134]
- Kiedrowski MR, Horswill AR. New approaches for treating staphylococcal biofilm infections. *Ann N Y Acad Sci*. 2011; 1241:104–121. [PubMed: 22191529]
- Lasa I. Towards the identification of the common features of bacterial biofilm development. *Int Microbiol*. 2006; 9:21–28. [PubMed: 16636986]
- Costerton JW, Stewart PS, Greenberg EP. Bacterial biofilms: a common cause of persistent infections. *Science*. 1999; 284:1318–1322. [PubMed: 10334980]
- Mack, D.; Davies, AP.; Harris, LG.; Jeeves, R.; Pascoe, B.; Knobloch, JK-M.; Rohde, H.; Wilkinson, TS. *Staphylococcus epidermidis* in Biomaterial-Associated Infection. In: Moriarty, TF.; Zaat, SAJ.; Busscher, HJ., editors. *Biomaterials Associated Infection*. New York: Springer; 2013. p. 25-57.
- Bjarnsholt T, Ciofu O, Molin S, Givskov M, Høiby N. Applying insights from biofilm biology to drug development - can a new approach be developed? *Nat Rev Drug Discov*. 2013; 12:791–808. [PubMed: 24080700]
- Auer JA, Goodship A, Arnoczky S, Pearce S, Price J, Claes L, von Rechenberg B, Hofmann-Amttenbrinck M, Schneider E, Müller-Terpitz R, Thiele F, Rippe KP, Grainger DW. Refining animal models in fracture research: seeking consensus in optimising both animal welfare and scientific validity for appropriate biomedical use. *BMC Musculoskelet Disord*. 2007; 8:72. [PubMed: 17678534]
- Doyle TC, Burns SM, Contag CH. In vivo bioluminescence imaging for integrated studies of infection. *Cell Microbiol*. 2004; 6:303–317. [PubMed: 15009023]

11. Zaat, SAJ. Tissue Colonization in Biomaterial-Associated Infection. In: Moriarty, TF.; Zaat, SAJ.; Busscher, HJ., editors. Biomaterials Associated Infection. New York: Springer; 2013. p. 175-209.
12. Anderson, JM.; Patel, JD. Biomaterial-Dependent Characteristics of the Foreign Body Response and S. epidermidis Biofilm Interactions. In: Moriarty, TF.; Zaat, SAJ.; Busscher, HJ., editors. Biomaterials Associated Infection. New York: Springer; 2013. p. 119-151.
13. Selvam S, Kundu K, Templeman KL, Murthy N, García AJ. Minimally invasive, longitudinal monitoring of biomaterial-associated inflammation by fluorescence imaging. Biomaterials. 2011; 32:7785–7792. [PubMed: 21813173]
14. Escapa IF, Morales V, Martino VP, Pollet E, Avérous L, García JL, Prieto MA. Disruption of β -oxidation pathway in *Pseudomonas putida* KT2442 to produce new functionalized PHAs with thioester groups. Appl Microbiol Biotechnol. 2011; 89:1583–1598. [PubMed: 21267558]
15. Furrer P, Panke S, Zinn M. Efficient recovery of low endotoxin medium-chain-length poly([R]-3-hydroxyalkanoate) from bacterial biomass. J Microbiol Methods. 2007; 69:206–213. [PubMed: 17316850]
16. FDA. U.S. Department of Health and Human Services FaDA. Guideline on validation of the Limulus amoebocyte lysate test as an end-product endotoxin test for human animal parenteral drugs, biological products, and medical devices. Rockville, MD: 1987. <http://www.gmpua.com/Validation/Method/LAL/FDAGuidelineForTheValidationA.pdf>
17. Mesak LR, Yim G, Davies J. Improved *lux* reporters for use in *Staphylococcus aureus*. Plasmid. 2009; 61:182–187. [PubMed: 19399993]
18. Cucarella C, Solano C, Valle J, Amorena B, Lasa I, Penadés JR. Bap, a *Staphylococcus aureus* surface protein involved in biofilm formation. J Bacteriol. 2001; 183:2888–2896. [PubMed: 11292810]
19. Merino N, Toledo-Arana A, Vergara-Irigaray M, Valle J, Solano C, Calvo E, Lopez JA, Foster TJ, Penadés JR, Lasa I. Protein A-mediated multicellular behavior in *Staphylococcus aureus*. J Bacteriol. 2009; 191:832–43. [PubMed: 19047354]
20. National Research Council. United States Department of Health and Human Services PHS, National Institutes of Health. Committee on Care and Use of Laboratory Animals of the Institute of Laboratory Animal Resources, Commission on Life Sciences. Bethesda, MD: National Research Council; 1985. Guide for the care and use of laboratory animals. NIH Publication No. 85-23
21. Hannan TJ, Mysorekar IU, Hung CS, Isaacson-Schmid ML, Hultgren SJ. Early severe inflammatory responses to uropathogenic *E. coli* predispose to chronic and recurrent urinary tract infection. PLoS Pathog. 2010; 6(8):e1001042.10.1371/journal.ppat.1001042. [PubMed: 20811584]
22. Nathan C. Points of control in inflammation. Nature. 2002; 420:846–852. [PubMed: 12490957]
23. Dinjaski N, Fernández-Gutiérrez M, Selvam S, Parra-Ruiz FJ, Lehman SM, San Román J, García E, García JL, García AJ, Prieto MA. PHACOS, a functionalized bacterial polyester with bactericidal activity against methicillin-resistant *Staphylococcus aureus*. Biomaterials. 2014; 35:14–24. [PubMed: 24094939]
24. Francois P, Vaudaux P, Lew PD. Role of plasma and extracellular matrix proteins in the physiopathology of foreign body infections. Ann Vasc Surg. 1998; 12:34–40. [PubMed: 9451994]
25. Francois P, Schrenzel J, Stoerman-Chopard C, Favre H, Herrmann M, Foster TJ, Lew DP, Vaudaux P. Identification of plasma proteins adsorbed on hemodialysis tubing that promote *Staphylococcus aureus* adhesion. J Lab Clin Med. 2000; 135:32–42. [PubMed: 10638692]
26. Foster TJ, Hook M. Surface protein adhesins of *Staphylococcus aureus*. Trends Microbiol. 1998; 6:484–488. [PubMed: 10036727]
27. Kadurugamuwa JL, Sin LV, Yu J, Francis KP, Kimura R, Purchio T, Contag PR. Rapid direct method for monitoring antibiotics in a mouse model of bacterial biofilm infection. Antimicrob Agents Chemother. 2003; 47:3130–3137. [PubMed: 14506020]
28. Keyaerts M, Cavelliers V, Lahoutte T. Bioluminescence imaging: looking beyond the light. Trends Mol Med. 2012; 18:164–172. [PubMed: 22321645]
29. Dai T, Tegos GP, Lu Z, Huang L, Zhiyentayev T, Franklin MJ, Baer DG, Hamblin MR. Photodynamic therapy for *Acinetobacter baumannii* burn infections in mice. Antimicrob Agents Chemother. 2009; 53:3929–3934. [PubMed: 19564369]

30. Waidmann MS, Bleichrodt FS, Laslo T, Riedel CU. Bacterial luciferase reporters: the Swiss army knife of molecular biology. *Bioeng Bugs*. 2011; 2:8–16. [PubMed: 21636983]
31. Hilderbrand SA, Weissleder R. Near-infrared fluorescence: application to *in vivo* molecular imaging. *Curr Opin Chem Biol*. 2010; 14:71–79. [PubMed: 19879798]
32. Leevy WM, Gammon ST, Jiang H, Johnson JR, Maxwell DJ, Jackson EN, Marquez M, Piwnica-Worms D, Smith BD. Optical imaging of bacterial infection in living mice using a fluorescent near-infrared molecular probe. *J Am Chem Soc*. 2006; 128:16476–16577. [PubMed: 17177377]
33. Bettegowda C, Foss CA, Cheong I, Wang Y, Diaz L, Agrawal N, Fox J, Dick J, Dang LH, Zhou S, Kinzler KW, Vogelstein B, Pomper MG. Imaging bacterial infections with radiolabeled 1-(2'-deoxy-2'-fluoro-beta-D-arabinofuranosyl)-5-iodouracil. *Proc Natl Acad Sci U S A*. 2005; 102:1145–1150. [PubMed: 15653773]
34. White AG, Gray BD, Pak KY, Smith BD. Deep-red fluorescent imaging probe for bacteria. *Bioorg Med Chem Lett*. 2012; 22:2833–2836. [PubMed: 22424976]
35. Kong Y, Yao H, Ren H, Subbian S, Cirillo SL, Sacchetti JC, Rao J, Cirillo JD. Imaging tuberculosis with endogenous beta-lactamase reporter enzyme fluorescence in live mice. *Proc Natl Acad Sci U S A*. 2010; 107:12239–12244. [PubMed: 20566877]
36. Leevy WM, Gammon ST, Johnson JR, Lampkins AJ, Jiang H, Marquez M, Piwnica-Worms D, Suckow MA, Smith BD. Noninvasive optical imaging of *Staphylococcus aureus* bacterial infection in living mice using a Bis-dipicolylamine-Zinc(II) affinity group conjugated to a near-infrared fluorophore. *Bioconjug Chem*. 2008; 19:686–692. [PubMed: 18260609]
37. Piatkevich KD, Subach FV, Verkhusha VV. Far-red light photoactivatable near-infrared fluorescent proteins engineered from a bacterial phytochrome. *Nat Commun*. 2013; 4:2153. [PubMed: 23842578]
38. Ning X, Lee S, Wang Z, Kim D, Stubblefield B, Gilbert E, Murthy N. Maltodextrin-based imaging probes detect bacteria *in vivo* with high sensitivity and specificity. *Nat Mater*. 2011; 10:602–707. [PubMed: 21765397]





Figure 1.

Experimental timeline for comparison of bioimaging approaches of biomaterial-associated infection. PHOHHx disks were pre-colonized with engineered bioluminescent bacteria. Following counting bacteria and bioluminescence imaging, disks were implanted subcutaneously. Bioluminescence was measured at 0, 1, 4, 7 days. After ROS imaging at day 7, disks were retrieved and analyzed for bacterial counts and immunostaining.

**Figure 2.**

Temporal progression of PHOHHx implant associated infection using different *S. aureus* luminescent strains. (a) Bioimaging data of animals scanned in an IVIS[®] imaging system for monitoring the intensity of *S. aureus* (pAmiBlaz), *S. aureus* (pAmiSPA) and *S. aureus* Xen29 luminescent signal. (b) Quantification of luminescence data from animals receiving subcutaneous PHOHHx implant pre-colonized with different *S. aureus* (pAmiBlaz), *S. aureus* (pAmiSPA) and *S. aureus* Xen29 CFU (n = 3 mice/time point).




Figure 3. Quantification of plasmid-containing bacteria (a) and number of viable cells (b). (a) *S. aureus* cells containing pAmiBLAZ or pAmiSPA previous to implantation and 7 days-post implantation was screened on TSA plates supplemented with chloramphenicol ($10 \mu\text{g mL}^{-1}$). (b) The number of viable *S. aureus* Xen29 cells previous to implantation and 7 days-post implantation determined on TSA kanamycin ($50 \mu\text{g mL}^{-1}$) plates.




Figure 4. NIRF imaging approach to detect PHOHHx implant-associated infection with different *S. aureus* strains at 7 days post-implantation. (a) Bioimaging data of animals scanned in an IVIS® imaging system for *in vivo* ROS imaging of inflammation associated with implant infection using H-ICG sensor. (b) Quantification of ROS fluorescence data from mice with PHOHHx implants incubated with 3.5×10^3 CFU/mm² (open bars), 3.5×10^4 CFU/mm² (closed bars) and sterile PHOHHx implant (open black bar) (n = 3 mice/time point).




Figure 5. Comparative representation of near IR fluorescence total efficiency (left axis) and bioluminescence total count (right axis) of *S. aureus* pre-colonized disks. NIRF imaging (bars) and bioluminescence imaging (line) of same *S. aureus* pre-colonized disks was compared in the same animals at 7 days post-implantation.




Figure 6. Immunohistochemical staining for macrophages (CD68+) and neutrophils (NIMP-R14) in infected implant-associated inflammation. (a–c) Representative co-localization images of CD68+ and DAPI stained nuclei in (a) sterile PHOHHx implant, (b) PHOHHx implant pre-colonized with 3.5×10^3 CFU/mm², and (c) PHOHHx implant pre-colonized with 3.5×10^5 CFU/mm². (d–f) Representative co-localization images of NIMP-R14+ and DAPI stained nuclei in (d) sterile PHOHHx implant, (e) PHOHHx implant pre-colonized with 3.5×10^3 CFU/mm², and (f) PHOHHx implant pre-colonized with 3.5×10^5 CFU/mm². (g) Quantification of CD68+, NIMP-R14+ and total number of cells stained positive for CD68+ and NIMP-R14+ (n = 4 disks/bacterial concentration, *P < 0.05). Scale bar 10 μm.




Figure 7. Immunofluorescence staining for (a–d) *S. aureus* (pAmiBLAZ) and (f–g) *S. aureus* (pAmiSPA) in PHOHHx implant associated infection. (a,e) Sterile PHOHHx implant, (b,f) 3.5×10^6 CFU/mm², (c,g) 3.5×10^7 CFU/mm², and (d,h) 3.5×10^8 CFU/mm². *S. aureus* marked with AlexaFluor 488-conjugated antibody shown in green, nucleus of cells surrounding implant DAPI stained and represented in blue. White arrows indicate zones of tissue necrosis. Scale bar 10 μ m.