



Published in final edited form as:

*Am J Clin Oncol*. 2013 June ; 36(3): 310–315. doi:10.1097/COC.0b013e31826e053d.

## American College of Radiology (ACR) and American Society for Radiation Oncology (ASTRO) Practice Guideline for the Performance of Stereotactic Radiosurgery (SRS)

Steven K. Seung, MD, PhD<sup>\*</sup>, David A. Larson, MD, PhD<sup>†</sup>, James M. Galvin, DSc<sup>‡</sup>, Minesh P. Mehta, MD<sup>§</sup>, Louis Potters, MD<sup>||</sup>, Christopher J. Schultz, MD<sup>¶</sup>, Santosh V. Yajnik, MD<sup>#</sup>, Alan C. Hartford, MD, PhD<sup>\*\*</sup>, and Seth A. Rosenthal, MD<sup>††</sup>

<sup>\*</sup>Department of Radiation Oncology, The Oregon Clinic, Portland, OR

<sup>†</sup>Department of Radiation Oncology, University of California-San Francisco, San Francisco, CA

<sup>‡</sup>Department of Radiation Oncology, Thomas Jefferson University, Philadelphia, PA

<sup>§</sup>Department of Radiation Oncology, Northwestern University Feinberg School of Medicine, Chicago, IL

<sup>||</sup>Department of Radiation Oncology, North Shore-LIJ Health System, New Hyde Park, NY

<sup>¶</sup>Department of Radiation Oncology, Medical College of Wisconsin, Milwaukee, WI

<sup>#</sup>Radiation Oncology Consultants, Park Ridge, IL

<sup>\*\*</sup>Section of Radiation Oncology, Department of Medicine, Dartmouth Medical School, Hanover, NH

<sup>††</sup>Radiation Oncology Centers, Sacramento, CA

### Abstract

American College of Radiology and American Society for Radiation Oncology Practice Guideline for the Performance of Stereotactic Radiosurgery (SRS). SRS is a safe and efficacious treatment option of a variety of benign and malignant disorders involving intracranial structures and selected extracranial lesions. SRS involves a high dose of ionizing radiation with a high degree of precision and spatial accuracy. A quality SRS program requires a multidisciplinary team involved in the patient management. Organization, appropriate staffing, and careful adherence to detail and to established SRS standards is important to ensure operational efficiency and to improve the likelihood of procedural success. A collaborative effort of the American College of Radiology and American Society for Therapeutic Radiation Oncology has produced a practice guideline for SRS.

Copyright © 2013 American College of Radiology (ACR) and American Society for Radiation Oncology (ASTRO)

Reprints: Steven K. Seung, MD, PhD, Department of Radiation Oncology, The Oregon Clinic, 4805 NE Glisan, Level G, Portland, OR 97213-2991. steven.seung@providence.org.

The other authors declare no conflicts of interest.

This guideline was revised according to the process described under the heading *The Process for Developing ACR Practice Guidelines and Technical Standards* on the ACR web site (<http://www.acr.org/guidelines>) by the Guidelines and Standards Committee of the Commission on Radiation Oncology in collaboration with the ASTRO.

This report was previously published as ACR-ASTRO Practice Guideline for the Performance of Stereotactic Radiosurgery, ACR Practice Guidelines and Technical Standards CD, 2011.

The guideline defines the qualifications and responsibilities of all the involved personnel, including the radiation oncologist, neurosurgeon, and qualified medical physicist. Quality assurance is essential for safe and accurate delivery of treatment with SRS. Quality assurance issues for the treatment unit, stereotactic accessories, medical imaging, and treatment-planning system are presented and discussed. Adherence to these practice guidelines can be part of ensuring quality and patient safety in a successful SRS program.

### Keywords

stereotactic radiation therapy; stereotactic radiosurgery; quality assurance; safety; practice guidelines; radiation oncology

---

### PREAMBLE

These guidelines are an educational tool designed to assist practitioners in providing appropriate radiation oncology care for patients. They are not inflexible rules or requirements of practice and are not intended, nor should they be used, to establish a legal standard of care. For these reasons and those set forth below, the American College of Radiology cautions against the use of these guidelines in litigation in which the clinical decisions of a practitioner are called into question.

The ultimate judgment regarding the propriety of any specific procedure or course of action must be made by the physician or medical physicist in light of all the circumstances presented. Thus, an approach that differs from the guidelines, standing alone, does not necessarily imply that the approach was below the standard of care. To the contrary, a conscientious practitioner may responsibly adopt a course of action different from that set forth in the guidelines when, in the reasonable judgment of the practitioner, such course of action is indicated by the condition of the patient, limitations of available resources, or advances in knowledge or technology subsequent to publication of the guidelines. However, a practitioner who uses an approach substantially different from these guidelines is advised to document in the patient record information sufficient to explain the approach taken.

The practice of medicine involves not only the science, but also the art of dealing with the prevention, diagnosis, alleviation, and treatment of disease. The variety and complexity of human conditions make it impossible to always reach the most appropriate diagnosis or to predict with certainty a particular response to treatment.

Therefore, it should be recognized that adherence to these guidelines will not assure an accurate diagnosis or a successful outcome. All that should be expected is that the practitioner will follow a reasonable course of action based on current knowledge, available resources, and the needs of the patient to deliver effective and safe medical care. The sole purpose of these guidelines is to assist practitioners in achieving this objective.

### INTRODUCTION

This guideline was revised by the American College of Radiology (ACR) and the American Society for Radiation Oncology (ASTRO).

Stereotactic radiosurgery (SRS) historically referred to targeting intracranial lesions. As various technologies have advanced, SRS has come to refer also to targeting extracranial lesions. For the purpose of this document SRS is strictly defined as radiation therapy delivered via stereotactic guidance with approximately 1 mm targeting accuracy to intracranial targets in 1 to 5 fractions. For information regarding extracranial target treatments, refer to the ACR-ASTRO Practice Guideline for the Performance of Stereotactic Body Radiation Therapy.<sup>1</sup>

SRS has been applied to a number of benign and malignant intracranial conditions.<sup>2–11</sup> The delivery of a high dose of ionizing radiation that conforms to the shape of the lesion mandates an overall accuracy of approximately 1 mm.<sup>12</sup>  $\gamma$ -ray photons, x-ray photons, protons, helium ions, and neutrons have been used for SRS. SRS can be delivered using a medical linear accelerator, a  $\gamma$ -ray treatment device, or a particle beam accelerator. Despite the variety of stereotactic radiosurgical techniques, many commonalities exist.<sup>13–16</sup>

The shape of the beam aperture used with linear accelerator-based systems is usually defined by secondary collimation positioned near the patient to reduce the beam penumbra. A large number of such beams sequentially irradiate the target, typically using a dynamic delivery,  $\gamma$ -ray treatment devices also position the collimation near the patient's skin surface to control the penumbra. In this case, numerous  $\gamma$ -ray beams, depending on the model, simultaneously irradiate a single point (called the isocenter) within the patient. In the early implementation of either the x-ray or  $\gamma$ -ray treatment approach, all beams were focused to converge to this single point in space so that a dose distribution devoid of surface concavities was produced. More recently, multiple points in space (isocenters) are irradiated to shape the dose distribution to allow for critical structures that invaginate the target.

Robotic, nonisocentric, frameless SRS is a type of SRS treatment consisting of dozens of nonisocentric beams with distinctive quality assurance (QA) procedures and continuous target tracking that result in comparable dose conformality and reduction in intrafraction systematic error.

Intensity modulated radiation therapy (IMRT) is also used for SRS. In this case a single isocenter can be used with off-axis beams created by a multileaf collimator so that the equivalent effect obtained with multiple isocenters is achieved. The multileaf collimator is often placed as a tertiary device nearer to the patient and with narrow leaves to improve penumbra. A related approach that is used for robotic dose delivery does not have a mechanical isocenter as a reference in space for the treatment beams. Like the IMRT, this technology also uses a large number of beams that crisscross through the target(s) from different directions that are not necessarily oriented toward any single point in space. As is the case for a number of the other SRS approaches, the robotic delivery approach usually irradiates a subregion of the target with any 1 beam to paint the dose for complex irregularly shaped targets. Stereotactic localization of the lesion uses an appropriate imaging modality to identify a reference point for positioning the individual treatment beams. Traditionally, a rigid frame that included a fiducial system for precisely locating coordinate positions within the frame was attached to the patient's head. More recently, "frameless" approaches have been introduced. These approaches are referred to as image-guided radiation therapy (IGRT)

techniques. While being irradiated, the patient may be immobilized when appropriate, and patient and target positioning are verified to ensure accurate treatment delivery.

Imaging, planning, and treatment typically are performed in close temporal proximity. Treatment delivery should be accurate to within approximately 1 mm. This leaves little room for error in the overall process. Strict protocols for QA must be followed and multiple-checking, preferably repeated by different individuals, is required at critical junctures. SRS requires the participation of a multidisciplinary team as outlined below.

The guidelines outlined in this document describe a minimal set of criteria for an SRS QA program. The reader is also referred to other publications regarding quality control for SRS and its related procedures.<sup>17–33</sup> In some cases, SRS may be an IGRT procedure. The recommendations are described in both the ACR-ASTRO Practice Guideline for IGRT<sup>34,35</sup> and the ACR Technical Standard for Medical Physics Performance Monitoring of Image-Guided External Beam Radiation Therapy.<sup>36</sup>

## QUALIFICATIONS AND RESPONSIBILITIES OF PERSONNEL

See the ACR Practice Guideline for Radiation Oncology where qualifications, credentialing, professional relationships, and development are outlined.<sup>37</sup>

The following are minimal recommendations for staffing levels and staff responsibilities while participating in an SRS procedure. Specific duties may be reassigned where appropriate.

### Radiation Oncologist

1. Certification in Radiology by the American Board of Radiology (ABR) of a physician who confines his/her professional practice to radiation oncology, or certification in Radiation Oncology or Therapeutic Radiology by the ABR, the American Osteopathic Board of Radiology, the Royal College of Physicians and Surgeons of Canada, or the Collège des Médecins du Québec may be considered proof of adequate physician qualifications.  
  
and/or
2. Satisfactory completion of a residency program in radiation oncology approved by the Accreditation Council for Graduate Medical Education, the Royal College of Physicians and Surgeons of Canada, the Collège des Médecins du Québec, or the American Osteopathic Association.

If the radiation oncology residency training did not include SRS training and direct clinical experience, then specific training or mentoring in SRS should be obtained before performing any radiosurgical procedures. In addition there may be vendor-specific delivery systems that require additional training.

For SRS treatment devices that use sealed isotope sources, the radiation oncologist is the “authorized user” as defined by Nuclear Regulatory Commission regulations. The

responsibilities of the radiation oncologist must be clearly defined and, irrespective of the treatment device, his or her duties should include the following:

1. Participating in initial treatment decision making and obtaining informed consent.
2. Overseeing radiation therapy management of the patient.
3. In concert with the neurosurgeon, neuroradiologist, or other physicians, specifying the target volume and relevant critical normal tissues.
4. Participating in the iterative process of plan development and approving the final treatment plan and dose.
5. Ensuring that patient positioning on the treatment unit is appropriate.
6. Attending and directing the radiosurgical treatment delivery, according to Nuclear Regulatory Commission regulations where appropriate.
7. Following the patient and participating in the monitoring of disease control and complications

### **Neurosurgeon**

Satisfactory completion of an Accreditation Council for Graduate Medical Education-approved neurosurgical residency program.

If the neurosurgical residency training did not include SRS training and direct clinical experience, then specific training or mentoring in SRS should be obtained before performing any radiosurgical procedures. In addition there may be vendor-specific delivery systems that require additional training.

An appropriately trained neurosurgeon is an integral member of the multidisciplinary SRS team and his or her services may include:

1. Participating in initial treatment decision making.
2. Placement of stereotactic head frame, where necessary.
3. Locating and specifying the target volume and relevant critical normal tissues in concert with the radiation oncologist and neuroradiologist or other physicians.
4. Participating in the iterative process of plan development and approving the final treatment plan and dose.
5. Ensuring that patient positioning on the treatment unit is appropriate.
6. Following the patient and participating in the monitoring of disease control and management of treatment complications.

### **Continuing Medical Education**

The radiation oncologist's continuing medical education should be in accordance with the ACR Practice Guideline for Continuing Medical Education (CME).<sup>38</sup> The physician should also meet the CME requirements as is the standard at the physician's institution.

### Qualified Medical Physicist

A qualified medical physicist is an individual who is competent to practice independently one or more of the sub-fields in medical physics. The ACR considers that certification and continuing education and experience in the appropriate subfield(s) to demonstrate that an individual is competent to practice one or more of the subfields in medical physics, and to be a qualified medical physicist. The ACR recommends that the individual be certified in the appropriate subfield(s) by the ABR, the Canadian College of Physics in Medicine, or for magnetic resonance imaging (MRI), by the American Board of Medical Physics in MRI physics.

The appropriate subfields of medical physics for this guideline are therapeutic radiological physics and radiological physics.

A qualified medical physicist should meet the ACR Practice Guideline for CME (ACR Resolution 17, 1996—revised in 2008, Resolution 7).

If the above training did not include SRS, then specific training or mentoring in SRS should be obtained before performing any radiosurgical procedures. There may be vendor-specific delivery systems that require additional training.

The medical physicist is responsible for many technical aspects of radiosurgery and must be available for consultation throughout the entire procedure: imaging, treatment planning, and dose delivery. Those responsibilities must be clearly defined and should include the following:

1. Acceptance testing and commissioning of the radiosurgery system to assure its initial geometric and dosimetric precision and accuracy.<sup>12,39</sup> This includes:
  - Localization devices used for accurate determination of target coordinates.
  - The treatment-planning system.<sup>40</sup>
  - The radiosurgery external beam delivery unit.
  - The precision of the imaging device, such as the MRI scanner, used for target and critical structure identification.
2. Implementing and managing a QA program for the radiosurgery system to monitor and assure its proper functioning.<sup>41–43</sup>
  1. The radiosurgery external beam delivery unit.
  2. The treatment-planning system.
  3. The precision of the imaging device, such as the MRI scanner, used for target and critical structure identification.
3. Initiation and maintenance of a comprehensive QA checklist that acts as a detailed guide to the entire treatment process.

4. Directly planning, supervising, or overseeing the treatment-planning process, including verification of dosimetric calculations using monitor unit double-check software.
5. Consulting with the radiation oncologist and/or medical dosimetrist to determine the optimal patient plan.
6. Using the plan approved by the radiation oncologist and an appropriate patient-specific measurement technique and checks the appropriate beam-delivery parameters.
7. Supervising the technical aspects of the beam-delivery process on the treatment unit to assure accurate fulfillment of the prescription of the radiation oncologist.

### **Radiation Therapist (When Applicable)**

A radiation therapist must fulfill state licensing requirements and should have American Registry of Radiologic Technologists certification in radiation therapy.

The responsibilities of the radiation therapist must be clearly defined and may include the following:

1. Preparing the treatment room for the SRS procedure.
2. Assisting the treatment team with patient positioning/immobilization.
3. Operating the treatment unit after the clinical and technical aspects of beam delivery are approved.

If the radiation therapy training did not include SRS training and direct clinical experience, then specific training or mentoring in SRS should be obtained before performing any radiosurgical procedures. In addition there may be vendor-specific delivery systems that require additional training.

### **Medical Dosimetrist (When Applicable)**

The responsibilities of the medical dosimetrist or other designated treatment planner must be clearly defined and may include the following:

1. Contour clearly discernible critical normal structures.
2. Ensure proper orientation of volumetric patient image data on the radiation therapy treatment-planning system (from computed tomography and other fused image data sets).
3. Design and generate the treatment plan under the direction of the radiation oncologist and medical physicist as required.
4. Generate all technical documentation required to implement the treatment plan.
5. Be available for the first treatment and assist with verification for subsequent treatments as necessary.

If the radiation therapy training did not include SRS training and direct clinical experience, then specific training or mentoring in SRS should be obtained before performing any radiosurgical procedures. In addition there may be vendor-specific delivery systems that require additional training.

## QA OF THE TREATMENT UNIT

The mechanical precision and electronic complexity of the treatment-delivery unit require the implementation of and adherence to an ongoing QA program. This program assures that the SRS treatment unit is in compliance with recommendations of the treatment unit manufacturer, with the specified clinical tolerances recommended by the ACR, AAPM, and ASTRO and with applicable regulatory requirements. It is recognized that various test procedures, with equal validity, may be used to ascertain that the treatment-delivery unit is functioning properly and safely. However, it is the responsibility of the medical physicist to determine that the testing procedure used is equivalent to the recommendations listed above. The test results should be documented, signed by the person doing the testing, and archived.

Important elements of the treatment-delivery unit QA program are as follows:

1. Radiation-beam alignment testing to assure the beam can be correctly aimed at the targeted tissues (see the Quality control of stereotactic accessories section for a complete list of the references describing this test).<sup>39</sup>
2. Calculation of radiation dose per unit time (or per monitor unit) based on physical measurements for the treatment field size at the location of the target.

## QUALITY CONTROL OF STEREOTACTIC ACCESSORIES

In some cases, SRS may be an IGRT procedure. As such, all of the recommendation previously stated in the ACR-ASTRO Practice Guideline for IGRT and the ACR Technical Standard for Medical Physics Performance Monitoring of Image-Guided Radiation External Beam Therapy<sup>34,36</sup> apply to this treatment modality. In addition, the AAPM Task Group 142 report<sup>44</sup> was written to extend the information in the previous AAPM Task Group 40 report<sup>45</sup> to specifically include guidelines for SRS. This document calls for daily verification of the correspondence of the treatment and imaging reference coordinate systems. Tolerance limits for this test are also stated in the AAPM Task Group 142 report.<sup>44</sup> For the use of frameless localization systems, a precise description of the required test is given in the practice guideline and technical standard referred to above. For frame-based systems, the classic Winston/Lutz test is recommended periodically on a regular basis when applicable.<sup>22</sup>

## QUALITY CONTROL OF IMAGES

SRS is image-based treatment. All salient anatomic features of the SRS patient, both normal and abnormal, are defined with computed tomography (CT), MRI, angiography, and/or other applicable imaging modalities. Both high 3D spatial accuracy and tissue-contrast definition are very important imaging features if one is to utilize SRS to its fullest positional accuracy. When the imager is located in the radiology department and not under direct control of the



radiation oncology department, considerable cooperation is required for good quality control specific to the needs of SRS.

The medical images used in SRS are critical to the entire process. They are used for focalizing target boundaries and generating target coordinates at which the treatment beams are to be aimed (see the Quality control of stereotactic accessories section). They are used for creating an anatomic patient model (virtual patient) for treatment planning, and they contain the morphology required for the treatment plan evaluation and dose calculation. The accuracy and precision required by SRS are to be assured. This assurance issue is addressed in the QA for SRS treatment-planning systems section below. However, general consideration should be given to the following issues.

Imaging, whether by CT, MRI, or other applicable modalities, should assure creation of a spatially accurate anatomic patient model for use in the treatment-planning process. The chosen image sets should also allow optimal definition of target(s) and critical structure(s). The chosen imaging modality must be thoroughly investigated before use in the SRS treatment-planning process. Some imaging considerations are the following: partial volume averaging, pixel size, slice thickness, distance between slices, image reformatting for the treatment-planning system, spatial distortion and image warping, motion artifacts, magnetic susceptibility artifacts, and others.

## QA FOR SRS TREATMENT-PLANNING SYSTEMS

SRS treatment-planning systems are very complex. Data from medical imaging devices are used in conjunction with a mathematical description of the external radiation beams to produce an anatomically detailed patient model illustrating the dose distribution with a high degree of precision. The level of complexity is also related to the treatment-planning techniques used for SRS. When IMRT techniques are used, inverse treatment-planning methodologies are necessary. However, these same inverse planning approaches are used for some of the multi-isocenter and nonisocentric treatment approaches. Inverse treatment planning provides computer-selected weights for a very large number of independent treatment beams. As such, it significantly complicates the treatment-planning process and requires QA steps that are different than the information provided in some earlier reports on treatment-planning QA (AAPM TG-53 report).<sup>40</sup> Because of the system's complexity, the medical physicist may elect to release the system in stages, and the required validation and verification testing will only reflect the features of the system that are in current clinical use at the facility. Documentation must exist indicating that the medical physicist has authorized the system for clinical use and has established a QA program to monitor the system's performance as it relates to the treatment-planning process.

The QA program for SRS involves elements that may be considered to be both dosimetric and nondosimetric. In addition to the IGRT recommendations given in the Quality control of stereotactic accessories section, it is further recommended that the ACR-ASTRO Practice Guideline for IMRT<sup>46</sup> be followed when IMRT is used and when other techniques that use inverse planning are used. Furthermore, it is recognized that various testing methods may be used, with equal validity, to assure that a system feature or component is performing

correctly. However, it is the responsibility of the medical physicist to determine that the alternative testing methods are equivalent to the testing procedures presented in the ACR-ASTRO Practice Guideline for IGRT, the ACR Technical Standard for Medical Physics Performance Monitoring of Image-Guided External Beam Radiation Therapy, and the ACR-ASTRO Practice Guideline on IMRT.<sup>34,36,47</sup> Although the AAPM document on QA for treatment planning does not fully include recommendations on IMRT, it should be used as a reference for general QA of treatment-planning systems (AAPM TG-53 report).<sup>40</sup> It is also noted that the commercial manufacturer may recommend specific QA tests to be performed on its planning systems. Although a manufacturer's testing procedure can be very helpful, it is the medical physicist's responsibility to guarantee that the total QA is complete in that it addresses all modes of possible failure. The references given above should be consulted to make this determination.

### **System Log**

Maintain an ongoing system log indicating system component failures, error messages, corrective actions, and system hardware or software changes.

### **System Data Input Devices**

Check the input devices of image-based planning systems for functionality and accuracy. Devices include digitizer tablet, input interface for medical imaging data (CT, MRI, angiography, etc.) and video digitizers. Assure correct anatomic registration: left, right, anterior, posterior, cephalad, and caudad from all the appropriate input devices.

### **System Output Devices**

Assure the functionality and accuracy of all printers, plotters, and graphical display units that produce, using digitally reconstructed radiographs or the like, a beam's-eye view rendering of anatomic structures near the treatment beam isocenter. Assure correct information transfer and appropriate dimensional scaling.

### **System Software**

Assure the continued integrity of the radiation therapy treatment-planning system information files used for modeling the external radiation beams. Confirm agreement of the beam modeling to currently accept clinical data derived from physical measurements. Similarly, assure the integrity of the system to render the anatomic modeling correctly.

## **VALIDATION OF THE TECHNIQUE AS IMPLEMENTED**

Once the individual components of the SRS planning and treatment technique are commissioned, it is recommended that the QA program include an "operational test" of the SRS system before clinical treatment begins, or whenever a plan modification is implemented for a fractionated treatment schedule. This testing should mimic the patient treatment and should use all of the same equipment used for treating the patient. The testing is given the name "patient specific end-to-end testing" and is described in the ACR-ASTRO Practice Guideline for IMRT.<sup>46</sup> An added benefit to this approach is that it provides training for each team member who will participate in the procedure.

## FOLLOW-UP

There should be follow-up of all patients treated and maintenance of appropriate records. The data should be collected in a manner that complies with statutory and regulatory guidelines to protect confidentiality.

## DOCUMENTATION

Procedure documentation should be in accordance with the ACR Practice Guideline for Communication: Radiation Oncology.<sup>48</sup>

## SUMMARY

The quality of a SRS program is defined by the strength of the multidisciplinary team involved in the management of the patient, and the attention to detail for this highly complex and demanding procedure. Radiosurgery is an involved procedure requiring participants from many disciplines. High spatial accuracy is expected, and there may be time constraints. Numerous systems to achieve optimal accuracy have been developed, and specific training in their use is required. All of the above demands a highly organized and efficient SRS team. Checklists are required to ensure that all aspects of the procedure are completed properly by each team member. The procedure must be appropriately staffed.

## Acknowledgments

M.P.M.: Relevant financial activities outside the submitted work; Board membership: Pharmaceutics (money paid to him). Consultancy: Adnexus, Bayer, Bristol Meyers, Merck, Quark, Tomotherapy (money paid to him). Stock/ Stock Options: Pharmaceutics, Tomotherapy, Procetus, Colby, Stemina. Payment for lectures: Merck, MCM, Robert-Michael Institute, Grace Foundation (money paid to him). Patents: Procetus (money paid to him). Royalties: Warf. Development of educational presentations: MCM (money paid to him). Other: Apogenix, DSMB (money paid to him). SVY: Shareholder in Illinois Gamma Knife Center, Elk Grove, IL. Shareholder in CDH Proton Center, Warrenville, IL. Shareholder in a Cyberknife center in Ridge Park, IL at Advocate Lutheran General Hospital.

## REFERENCES

1. Potters L, Kavanagh B, Galvin JM, et al. American society for therapeutic radiology and oncology (Astro) and American college of radiology (ACR) practice guideline for the performance of stereotactic body radiation therapy. *Int J Radiat Oncol Biol Phys.* 2010; 76:326–332. [PubMed: 20117285]
2. Andrade-Souza YM, Zadeh G, Scora D, et al. Radiosurgery for basal ganglia, internal capsule, and thalamus arteriovenous malformation: clinical outcome. *Neurosurgery.* 2005; 56:56–63. discussion 63–54. [PubMed: 15617586]
3. Chopra R, Kondziolka D, Niranjan A, et al. Long-term follow-up of acoustic schwannoma radiosurgery with marginal tumor doses of 12 to 13Gy. *Int J Radiat Oncol Biol Phys.* 2007; 68:845–851. [PubMed: 17379451]
4. Davey P, Schwartz ML, Scora D, et al. Fractionated (split dose) radiosurgery in patients with recurrent brain metastases: implications for survival. *Br J Neurosurg.* 2007; 2:491–495. [PubMed: 17852114]
5. Linskey ME, Andrews DW, Asher AL, et al. The role of stereotactic radiosurgery in the management of patients with newly diagnosed brain metastases: a systematic review and evidence-based clinical practice guideline. *J Neurooncol.* 2010; 96:45–68. [PubMed: 19960227]

6. Lopez BC, Hamlyn PJ, Zakrzewska JM. Stereotactic radiosurgery for primary trigeminal neuralgia: state of the evidence and recommendations for future reports. *J Neurol Neurosurg Psychiatry*. 2004; 75:1019–1024. [PubMed: 15201363]
7. Mehta MP, Tsao MN, Whelan TJ, et al. The American Society for Therapeutic Radiology and Oncology (ASTRO) evidence-based review of the role of radiosurgery for brain metastases. *Int J Radiat Oncol Biol Phys*. 2005; 63:37–46. [PubMed: 16111570]
8. Romanelli P, Conti A, Pontoriero A, et al. Role of stereotactic radiosurgery and fractionated stereotactic radiotherapy for the treatment of recurrent glioblastoma multiforme. *Neurosurg Focus*. 2009; 27:E8. [PubMed: 19951061]
9. Schwer AL, Damek DM, Kavanagh BD, et al. A phase I dose-escalation study of fractionated stereotactic radiosurgery in combination with gefitinib in patients with recurrent malignant gliomas. *Int J Radiat Oncol Biol Phys*. 2008; 70:993–1001. [PubMed: 17967517]
10. Tsao MN, Mehta MP, Whelan TJ, et al. The American Society for Therapeutic Radiology and Oncology (ASTRO) evidence-based review of the role of radiosurgery for malignant glioma. *Int J Radiat Oncol Biol Phys*. 2005; 63:47–55. [PubMed: 16111571]
11. Vesper J, Bolke B, Wille C, et al. Current concepts in stereotactic radiosurgery—a neurosurgical and radiooncological point of view. *Eur J Med Res*. 2009; 14:93–101. [PubMed: 19380278]
12. Schell, MC.; Bova, FJ.; Larson, DA., et al. Stereotactic Radiosurgery. AAPM Report 54; Task Group 42. College Park, MD: American Association of Physicists in Medicine; 1995.
13. Branch CL Jr, Coric D, Olds W, et al. Stereotactic radiosurgery. A review of “gamma knife” and “linac knife” technology and the unit at the Wake Forest University Medical Center. *N C Med J*. 1992; 53:395–399. [PubMed: 1513382]
14. Sahgal A, Ma L, Chang E, et al. Advances in technology for intracranial stereotactic radiosurgery. *Technol Cancer Res Treat*. 2009; 8:271–280.
15. Theodorou K, Stathakis S, Lind B, et al. Dosimetric and radiobiological evaluation of dose distribution perturbation due to head heterogeneities for Linac and Gamma Knife stereotactic radiotherapy. *Acta Oncol*. 2008; 47:917–927. [PubMed: 17957500]
16. Yu C, Shepard D. Treatment planning for stereotactic radiosurgery with photon beams. *Technol Cancer Res Treat*. 2003; 2:93–104. [PubMed: 12680789]
17. Curran, BH.; Starkschall, G. A program for quality assurance of dose planning computer. In: Starkschall, G.; Horton, J., editors. *Quality Assurance in Radiotherapy Physics*. Madison, WI: Medical Physics Publishing; 1991. p. 207-228.
18. Drzymala RE, Klein EE, Simpson JR, et al. Assurance of high quality linac-based stereotactic radiosurgery. *Int J Radiat Oncol Biol Phys*. 1994; 30:459–472. [PubMed: 7928474]
19. Grebe G, Pfaender M, Roll M, et al. Dynamic arc radiosurgery and radiotherapy: commissioning and verification of dose distributions. *Int J Radiat Oncol Biol Phys*. 2001; 49:1451–1460. [PubMed: 11286853]
20. Kooy HM, van Herk M, Barnes PD, et al. Image fusion for stereotactic radiotherapy and radiosurgery treatment planning. *Int J Radiat Oncol Biol Phys*. 1994; 28:1229–1234. [PubMed: 8175410]
21. Larson DA, Bova F, Eisert D, et al. Current radiosurgery practice: results of an ASTRO survey. Task Force on Stereotactic Radiosurgery, American Society for Therapeutic Radiology and Oncology. *Int J Radiat Oncol Biol Phys*. 1994; 28:523–526. [PubMed: 8276670]
22. Lutz W, Winston KR, Maleki N. A system for stereotactic radiosurgery with a linear accelerator. *Int J Radiat Oncol Biol Phys*. 1988; 14:373–381. [PubMed: 3276655]
23. Mack A, Czempiel H, Kreiner HJ, et al. Quality assurance in stereotactic space. A system test for verifying the accuracy of aim in radiosurgery. *Med Phys*. 2002; 29:561–568.
24. Maitz AH, Wu A, Lunsford LD, et al. Quality assurance for gamma knife stereotactic radiosurgery. *Int J Radiat Oncol Biol Phys*. 1995; 32:1465–1471. [PubMed: 7635790]
25. Ramaseshan R, Heydarian M. Comprehensive quality assurance for stereotactic radiosurgery treatments. *Phys Med Biol*. 2003; 48:N199–N205. [PubMed: 12894984]
26. Rice RK, Hansen JL, Svensson GK, et al. Measurements of dose distributions in small beams of 6MV x-rays. *Phys Med Biol*. 1987; 32:1087–1099. [PubMed: 3671496]

27. Scheib SG, Gianolini S, Lomax NJ, et al. High precision radiosurgery and technical standards. *Acta Neurochir Suppl.* 2004; 91:9–23. [PubMed: 15707022]
28. Shaw E, Kline R, Gillin M, et al. Radiation therapy oncology group: radiosurgery quality assurance guidelines. *Int J Radiat Oncol Biol Phys.* 1993; 27:1231–1239. [PubMed: 8262852]
29. Soanes T, Hampshire A, Vaughan P, et al. The commissioning and quality assurance of the Automatic Positioning System on the Leksell gamma knife. *J Neurosurg.* 2002; 97:574–578. [PubMed: 12507099]
30. Tsai JS, Buck BA, Svensson GK, et al. Quality assurance in stereotactic radiosurgery using a standard linear accelerator. *Int J Radiat Oncol Biol Phys.* 1991; 21:737–748. [PubMed: 1869467]
31. Yeung D, Palta J, Fontanesi J, et al. Systematic analysis of errors in target localization and treatment delivery in stereotactic radiosurgery (SRS). *Int J Radiat Oncol Biol Phys.* 1994; 28:493–498. [PubMed: 8276666]
32. Consensus statement on stereotactic radiosurgery quality improvement. The American Society for Therapeutic Radiology and Oncology, Task Force on Stereotactic Radiosurgery and the American Association of Neurological Surgeons, Task Force on Stereotactic Radiosurgery. *Int J Radiat Oncol Biol Phys.* 1994; 28:527–530. [PubMed: 8276671]
33. Consensus statement on stereotactic radiosurgery: quality improvement. *Neurosurgery.* 1994; 34:193–195. [PubMed: 8121561]
34. American College of Radiology. ACR-ASTRO Practice Guideline for Image-Guided Radiation Therapy (IGRT). Available at: [http://www.acr.org/SecondaryMainMenuCategories/quality\\_safety/guidelines/ro/IGRT.aspx](http://www.acr.org/SecondaryMainMenuCategories/quality_safety/guidelines/ro/IGRT.aspx).
35. Potters L, Gaspar LE, Kavanagh B, et al. American Society for Therapeutic Radiology and Oncology (ASTRO) and American College of Radiology (ACR) practice guidelines for image-guided radiation therapy (IGRT). *Int J Radiat Oncol Biol Phys.* 2010; 76:319–325. [PubMed: 20117284]
36. American College of Radiology. ACR Technical Standard for Medical Physics Performance Monitoring of Image-guided External Beam Radiation Therapy (IGRT). Available at: [http://www.acr.org/SecondaryMainMenuCategories/quality\\_safety/guidelines/med\\_phys/monitor\\_IGRT.aspx](http://www.acr.org/SecondaryMainMenuCategories/quality_safety/guidelines/med_phys/monitor_IGRT.aspx).
37. American College of Radiology. ACR Practice Guideline for Radiation Oncology. Available at: [http://www.acr.org/SecondaryMainMenuCategories/quality\\_safety/guidelines/ro/radiation\\_oncology.aspx](http://www.acr.org/SecondaryMainMenuCategories/quality_safety/guidelines/ro/radiation_oncology.aspx).
38. American College of Radiology. ACR practice guideline for continuing medical education (CME). Available at: [http://www.acr.org/SecondaryMainMenuCategories/quality\\_safety/guidelines/cme/cme.aspx](http://www.acr.org/SecondaryMainMenuCategories/quality_safety/guidelines/cme/cme.aspx).
39. Hartman, GH. Quality Assurance Program on Stereotactic Radiosurgery: Report from a Quality Assurance Task Group. Berlin, Germany: Springer-Verlag; 1995.
40. Fraass B, Doppke K, Hunt M, et al. American Association of Physicists in Medicine Radiation Therapy Committee Task Group 53: quality assurance for clinical radiotherapy treatment planning. *Med Phys.* 1998; 25:1773–1829. [PubMed: 9800687]
41. Dieterich S, Pawlicki T. Cyberknife image-guided delivery and quality assurance. *Int J Radiat Oncol Biol Phys.* 2008; 71:S126–S130. [PubMed: 18406910]
42. Goetsch SJ. Linear accelerator and gamma knife-based stereotactic cranial radiosurgery: challenges and successes of existing quality assurance guidelines and paradigms. *Int J Radiat Oncol Biol Phys.* 2008; 71:S118–S121. [PubMed: 18406908]
43. Solberg TD, Medin PM, Mullins J, et al. Quality assurance of immobilization and target localization systems for frameless stereotactic cranial and extracranial hypofractionated radiotherapy. *Int J Radiat Oncol Biol Phys.* 2008; 71:S131–S135. [PubMed: 18406912]
44. Klein EE, Hanley J, Bayouth J, et al. Task Group 142 report: quality assurance of medical accelerators. *Med Phys.* 2009; 36:4197–4212. [PubMed: 19810494]
45. Kutcher GJ, Coia L, Gillin M, et al. Comprehensive QA for radiation oncology: report of AAPM Radiation Therapy Committee Task Group 40. *Med Phys.* 1994; 21:581–618. [PubMed: 8058027]
46. Hartford AC, Palisca MG, Eichler TJ, et al. American Society for Therapeutic Radiology and Oncology (ASTRO) and American College of Radiology (ACR) Practice Guidelines for Intensity-

- Modulated Radiation Therapy (IMRT). *Int J Radiat Oncol Biol Phys.* 2009; 73:9–14. [PubMed: 19100920]
47. American College of Radiology. ACR-ASTRO Practice Guideline for Intensity-Modulated Radiation Therapy (IMRT). Available at: [http://www.acr.org/SecondaryMainMenuCategories/quality\\_safety/guidelines/ro/imrt.aspx](http://www.acr.org/SecondaryMainMenuCategories/quality_safety/guidelines/ro/imrt.aspx).
  48. American College of Radiology. ACR Practice Guideline for Communication: Radiation Oncology. Available at: [http://www.acr.org/SecondaryMainMenuCategories/quality\\_safety/guidelines/ro/comm\\_radiation\\_oncology.aspx](http://www.acr.org/SecondaryMainMenuCategories/quality_safety/guidelines/ro/comm_radiation_oncology.aspx).