

Comparison of FilmArray and Quantitative Real-Time Reverse Transcriptase PCR for Detection of Zaire Ebolavirus from Contrived and Clinical Specimens

Timothy R. Southern,^{a,b} Lori D. Racsa,^c César G. Albariño,^d Paul D. Fey,^a Steven H. Hinrichs,^a Caitlin N. Murphy,^{a,b} Vicki L. Herrera,^b Anthony R. Sambol,^b Charles E. Hill,^c Emily L. Ryan,^c Colleen S. Kraft,^{c,e} Shelley Campbell,^d Tara K. Sealy,^d Amy Schuh,^d James C. Ritchie,^c G. Marshall Lyon III,^e Aneesh K. Mehta,^e Jay B. Varkey,^e Bruce S. Ribner,^e Kent P. Brantly,^f Ute Ströher,^d Peter C. Iwen,^{a,b} Eileen M. Burd^{c,e}

Department of Pathology and Microbiology, University of Nebraska Medical Center, Omaha, Nebraska, USA^a; Nebraska Public Health Laboratory, Omaha, Nebraska, USA^b; Department of Pathology and Laboratory Medicine, Emory University, Atlanta, Georgia, USA^c; Centers for Disease Control and Prevention, Atlanta, Georgia, USA^d; Department of Medicine, Division of Infectious Diseases, Emory University School of Medicine, Atlanta, Georgia, USA^e; Samaritan's Purse, Boone, North Carolina, USA^f

Rapid, reliable, and easy-to-use diagnostic assays for detection of Zaire ebolavirus (ZEBOV) are urgently needed. The goal of this study was to examine the agreement among emergency use authorization (EUA) tests for the detection of ZEBOV nucleic acids, including the BioFire FilmArray BioThreat (BT) panel, the FilmArray BT-E panel, and the NP2 and VP40 quantitative real-time reverse transcriptase (qRT) PCR assays from the Centers for Disease Control and Prevention (CDC). Specimens used in this study included whole blood spiked with inactivated ZEBOV at known titers and whole-blood, plasma, and urine clinical specimens collected from persons diagnosed with Ebola virus disease (EVD). The agreement for FilmArray and qRT-PCR results using contrived whole-blood specimens was 100% (6/6 specimens) for each ZEBOV dilution from 4×10^7 to 4×10^2 50% tissue culture infective dose (TCID₅₀)/ml, as well as the no-virus negative-control sample. The limit of detection for FilmArray and qRT-PCR assays with inactivated ZEBOV, based on duplicate positive results, was determined to be 4×10^2 TCID₅₀/ml. Rates of agreement between FilmArray and qRT-PCR results for clinical specimens from patients with EVD were 85% (23/27 specimens) for whole-blood specimens, 90% (18/20 specimens) for whole-blood specimens tested by FilmArray testing and matched plasma specimens tested by qRT-PCR testing, and 85% (11/13 specimens) for urine specimens. Among 60 specimens, eight discordant results were noted, with ZEBOV nucleic acids being detected only by FilmArray testing in four specimens and only by qRT-PCR testing in the remaining four specimens. These findings demonstrate that the rapid and easy-to-use FilmArray panels are effective tests for evaluating patients with EVD.

Ebolavirus is an enveloped, single-stranded RNA virus that is the cause of Ebola virus disease (EVD). EVD is characterized by fever, emesis, diarrhea, and a hemorrhagic disorder that can include maculopapular rash, petechiae, ecchymoses, and mucosal hemorrhage (1). Zaire ebolavirus (ZEBOV) was the cause of the 2014–2015 outbreak of EVD in West Africa, which to date has resulted in 27,352 total cases, 15,052 laboratory-confirmed cases, and 11,178 deaths (2, 3).

Early detection of ZEBOV is critical for the management of cases of EVD and for outbreak control. A significant challenge in areas without Ebola, such as the United States, is the rapid assessment of individuals with a history of travel to West Africa who present with symptoms of EVD. Currently, the standard protocol for EVD testing involves collection of whole blood or plasma, followed by testing using quantitative real-time reverse transcriptase (qRT) PCR assays. Although they are highly sensitive and specific, qRT-PCR assays for ZEBOV are complex, which limits their use to state public health laboratories and the Centers for Disease Control and Prevention (CDC) (4, 5). Depending on the location of the patient being tested and the laboratory performing the testing, the turnaround time for qRT-PCR results could be measured in days, whereas initial testing performed at or near the point of care using a rapid test could be completed within hours. Rapid, reliable, and easy-to-use tests for the detection of ZEBOV are needed not only for testing in areas in which the disease is

endemic but also for screening of health care workers, international travelers, and other potentially exposed individuals.

The first U.S. nationals infected in Africa were returned to our facilities for treatment in the Serious Communicable Diseases Unit (SCDU) at Emory University (EU) and the Nebraska Biocontainment Unit at Nebraska Medicine, the academic medical hospital affiliated with the University of Nebraska Medical Center (UNMC), between 2 August and 28 October 2014 (6–8). At that

Received 19 May 2015 Returned for modification 26 May 2015

Accepted 30 June 2015

Accepted manuscript posted online 8 July 2015

Citation Southern TR, Racsa LD, Albariño CG, Fey PD, Hinrichs SH, Murphy CN, Herrera VL, Sambol AR, Hill CE, Ryan EL, Kraft CS, Campbell S, Sealy TK, Schuh A, Ritchie JC, Lyon GM, III, Mehta AK, Varkey JB, Ribner BS, Brantly KP, Ströher U, Iwen PC, Burd EM. 2015. Comparison of FilmArray and quantitative real-time reverse transcriptase PCR for detection of Zaire ebolavirus from contrived and clinical specimens. *J Clin Microbiol* 53:2956–2960. doi:10.1128/JCM.01317-15.

Editor: A. J. McAdam

Address correspondence to Eileen M. Burd, eburd@emory.edu.

T.R.S. and L.D.R. contributed equally to this article.

Supplemental material for this article may be found at <http://dx.doi.org/10.1128/JCM.01317-15>.

Copyright © 2015, American Society for Microbiology. All Rights Reserved.

doi:10.1128/JCM.01317-15

time, the assays available for the detection of ZEBOV were the research-use-only FilmArray BioThreat (BT) panel (BioFire Defense, Salt Lake City, UT) and the ZEBOV nucleoprotein 2 (NP2) and matrix protein (VP40) gene assays from the Centers for Disease Control and Prevention (Atlanta, GA). The NP2 and VP40 qRT-PCR assays and the FilmArray BT-E assay (a modified version of the BT panel) were granted emergency use authorization (EUA) in October 2014, during the course of the treatment of our patients. The FilmArray BT assay detects a panel of biothreat agents, but the ZEBOV primers detecting the L-gene are identical in the FilmArray BT test and the FilmArray BT-E (EUA) test. The FilmArray BT-E assay also includes a freeze-dried protease to add with the blood and loading buffer and has an additional primer in the second-stage PCR that perfectly matches the current circulating strain. This study provides a comparative analysis of the FilmArray BT-E panel and the NP2 and VP40 qRT-PCR assays for the detection of ZEBOV using contrived whole-blood specimens. This study also provides a prospective analysis of the FilmArray BT panel and qRT-PCR assays using whole-blood, plasma, and urine specimens from 6 persons with EVD who were treated in our facilities.

MATERIALS AND METHODS

FilmArray assay. The BioFire FilmArray system (BioFire Diagnostics LLC, Salt Lake City, UT) is an automated nested PCR system that allows the extraction and detection of nucleic acid targets in a closed system. The BT and BT-E panels (BioFire Defense, Salt Lake City, UT) are used with the FilmArray system to detect ZEBOV in whole-blood or urine specimens (9). Briefly, each panel was rehydrated with the provided rehydration solution, and whole blood (100 μ l) or urine (200 μ l) was mixed with the sample buffer provided. The resulting solution was injected into the panel pouch, and the panel pouch was loaded into the FilmArray instrument. One sample was tested at a time, with a result of detected or not detected being provided in approximately 1 h. FilmArray testing was performed in a satellite laboratory within the SCU at EU, at the Nebraska Public Health Laboratory (NPHL) biosafety level 3 (BSL3) facility located on the UNMC campus, and at the CDC in the Viral Special Pathogens Branch (VSPB) laboratory. Specimens were processed for FilmArray testing in a class II biological safety cabinet, by individuals wearing enhanced personal protective equipment, as defined by the biosafety committees at each institution.

qRT-PCR assays. Testing was also performed by the VSPB at the CDC using the qRT-PCR assays for the detection of NP2 and VP40 genes (10, 11). While only whole blood was evaluated with the FilmArray system, plasma and whole-blood samples were tested with qRT-PCR assays at various times. Results were generated in about 3 h and up to 96 specimens could be analyzed simultaneously. Total RNA was extracted from whole-blood, plasma, or urine specimens by using the BeadRetriever system (Invitrogen, Grand Island, NY) and the MagMax Pathogen RNA/DNA isolation kit (Applied Biosystems, Grand Island, NY). The NP2 qRT-PCR assay was then performed with the extracted nucleic acids, and a cycle threshold (C_T) value was reported back to each institution for use in patient care. Cycle threshold values of ≤ 40 were interpreted as positive. Specimens with no specific amplification or with amplification curves that did not cross the baseline threshold were interpreted as negative. Specimens that yielded positive results with C_T values of > 38 in the NP2 assay were confirmed using the VP40 assay and/or an additional serological assay, at the CDC VSPB laboratory (the EUA provides for equivocal interpretation for specimens with C_T values of 38 to 40 and suggests that additional analysis may be required).

Specimens. Contrived specimens were prepared at the VSPB using inactivated ZEBOV at known titers, to examine concordance between the FilmArray BT-E panel, the NP2 qRT-PCR assay, and the VP40 qRT-PCR

assay for detection of ZEBOV nucleic acids. ZEBOV (specimen number 812592) at a titer of 4×10^7 50% tissue culture infective dose (TCID₅₀)/ml was inactivated by gamma irradiation. Ten-fold serial dilutions of the inactivated virus, from 4×10^7 to 4×10^1 TCID₅₀/ml, and a no-virus negative-control sample were prepared in whole blood. Contrived specimens at each dilution of virus, and the negative-control sample, were tested in duplicate using the FilmArray BT-E panel and the NP2 and VP40 qRT-PCR assays.

Clinical specimens from individuals with EVD who were treated at EU or UNMC were evaluated using the FilmArray BT panel and qRT-PCR assays. Whole-blood ($n = 27$), plasma ($n = 20$), and urine ($n = 13$) specimens were collected and tested at various times throughout the course of patient management. Twenty-seven matched whole-blood specimens and 13 matched urine specimens were tested using the FilmArray BT panel and qRT-PCR testing. An additional 20 whole-blood specimens were tested using the FilmArray BT panel with qRT-PCR testing on matched plasma specimens, for a total of 60 pairwise tests on clinical specimens from patients being treated for EVD. FilmArray BT panel testing was performed at the clinical sites, using whole-blood and urine specimens. Matched clinical specimens were tested by qRT-PCR at the VSPB, once per sample. Only the qRT-PCR results were used for patient management at that time. The interval between on-site FilmArray testing and testing at the VSPB ranged from 24 h before to 5 days after qRT-PCR results were received.

Calculations. Agreement was calculated for FilmArray and qRT-PCR assay results.

Ethics statement. Approval was obtained from the institutional review boards at Emory University and the University of Nebraska.

RESULTS

Ten-fold serial dilutions of inactivated ZEBOV in whole blood, from 4×10^7 to 4×10^1 TCID₅₀/ml, were tested in duplicate using the FilmArray BT-E panel, the NP2 qRT-PCR assay, and the VP40 qRT-PCR assay, at the VSPB (Table 1). ZEBOV nucleic acids were detected in contrived whole-blood specimens at virus concentrations of 4×10^7 to 4×10^2 TCID₅₀/ml, resulting in 100% agreement among the 3 ZEBOV tests (6/6 specimens for each titer and 36/36 specimens overall). The VP40 qRT-PCR assay detected ZEBOV nucleic acids in 1 of 2 tests at the 4×10^1 virus dilution, while ZEBOV nucleic acids were not detected by either the FilmArray panel or the NP2 qRT-PCR assay, which resulted in 83.3% agreement (5/6 specimens) for that dilution. ZEBOV nucleic acids were not detected in whole-blood samples that did not contain inactivated virus, resulting in 100% agreement (6/6 specimens) among the three assays. The limit of detection (LOD), i.e., the lowest virus dilution at which each assay detected ZEBOV nucleic acids in duplicate tests, was determined to be 4×10^2 TCID₅₀/ml for the FilmArray panel and both qRT-PCR assays.

Sixty specimens (47 whole-blood specimens and 13 urine specimens) from individuals with EVD were tested using the FilmArray system (see Table S1 in the supplemental material). Testing of matched whole-blood, plasma, and urine specimens was performed by qRT-PCR testing at the CDC. Twenty-seven whole-blood specimens were tested using both the FilmArray and qRT-PCR assays. An additional 20 whole-blood specimens were tested using the FilmArray system and the corresponding plasma specimens were tested using the qRT-PCR assay. Thirteen urine specimens were also tested using both assays.

Of the 60 specimens tested, ZEBOV nucleic acids were detected in 40 specimens using both FilmArray and qRT-PCR assays, while 12 specimens were negative by both assays (Table 2). Eight discrepant results were noted, with ZEBOV nucleic acid results being

TABLE 1 Results of FilmArray BioThreat E panel and qRT-PCR testing using serial dilutions of inactivated Zaire ebolavirus at known titers in whole blood^a

Virus titer (TCID ₅₀ /ml)	NP2 results		VP40 results		FilmArray results	
	Test 1	Test 2	Test 1	Test 2	Test 1	Test 2
40,000,000	Positive (<i>C_T</i> = 20)	Positive (<i>C_T</i> = 20)	Positive (<i>C_T</i> = 18)	Positive (<i>C_T</i> = 18)	Detected	Detected
4,000,000	Positive (<i>C_T</i> = 23)	Positive (<i>C_T</i> = 23)	Positive (<i>C_T</i> = 22)	Positive (<i>C_T</i> = 22)	Detected	Detected
400,000	Positive (<i>C_T</i> = 26)	Positive (<i>C_T</i> = 26)	Positive (<i>C_T</i> = 25)	Positive (<i>C_T</i> = 25)	Detected	Detected
40,000	Positive (<i>C_T</i> = 30)	Positive (<i>C_T</i> = 30)	Positive (<i>C_T</i> = 28)	Positive (<i>C_T</i> = 28)	Detected	Detected
4,000	Positive (<i>C_T</i> = 33)	Positive (<i>C_T</i> = 33)	Positive (<i>C_T</i> = 32)	Positive (<i>C_T</i> = 32)	Detected	Detected
400	Positive (<i>C_T</i> = 37)	Positive (<i>C_T</i> = 35)	Positive (<i>C_T</i> = 34)	Positive (<i>C_T</i> = 34)	Detected	Detected
40	Negative	Negative	Negative	Positive (<i>C_T</i> = 37)	Not detected	Not detected
0	Negative	Negative	Negative	Negative	Not detected	Not detected

^a Zaire ebolavirus (strain 812592) stock at a known titer of 4×10^8 TCID₅₀/ml was inactivated by gamma irradiation prior to testing. EUA-approved NP2 and VP40 qRT-PCR assays and FilmArray BT-E panel testing were performed in duplicate, using 10-fold serial dilutions of inactivated virus in whole blood, at the CDC.

positive by the FilmArray test but negative by the qRT-PCR assay in four cases and positive by the qRT-PCR assay but negative by the FilmArray test in four cases. The specimens that yielded discrepant results included four whole-blood specimens, two plasma specimens (paired with whole-blood specimens), and two urine specimens (see Table S1 in the supplemental material). The overall agreement between FilmArray and qRT-PCR results was 87% (52/60 specimens). The agreements were 85% (23/27 specimens) for whole-blood specimens tested by both the FilmArray and qRT-PCR assays, 90% (18/20 specimens) for whole-blood specimens tested by the FilmArray assay and matched plasma specimens tested by the qRT-PCR assay, and 85% (11/13 specimens) for urine specimens (Table 3).

Two plasma specimens tested ZEBOV negative by qRT-PCR testing but the corresponding whole-blood specimens tested positive by FilmArray testing. To explore this further, a series of six specimens from one of the patients at EU were collected every 24 h for 6 days and tested on the day of collection, using whole blood and plasma obtained from the same collection tube. ZEBOV was detected in whole blood by both the FilmArray and qRT-PCR assays for all six specimens; however, detectable virus was found in only two of the corresponding plasma specimens by FilmArray testing, suggesting a trend toward the virus being cleared from plasma before being cleared from whole blood (Table 4).

DISCUSSION

Rapid, reliable, and easy-to-use assays for the detection of ZEBOV in clinical specimens are needed in response to the unprecedented outbreak in West Africa and the emergence of infected individuals beyond outbreak zones. This study evaluated the FilmArray assay and two qRT-PCR assays for detection of ZEBOV in whole-blood

specimens spiked with inactivated virus at known titers and whole-blood, plasma, and urine clinical specimens from 6 individuals with EVD who were treated in the United States.

Parallel testing of contrived whole-blood specimens at virus dilutions of 4×10^7 to 4×10^2 TCID₅₀/ml revealed 100% agreement between the three diagnostic assays evaluated in this study. The only disagreement noted was a single positive result in VP40 qRT-PCR testing for the dilution of 4×10^1 TCID₅₀/ml, which resulted in 83.3% agreement (5/6 specimens). Detection of the VP40 qRT-PCR target in one sample at a 10-fold lower dilution, compared to FilmArray and NP2 qRT-PCR results, may suggest somewhat greater sensitivity for the VP40 qRT-PCR assay.

Parallel testing of contrived whole-blood specimens with inactivated virus also provided insight into the limits of detection, which, given the concentrations tested, were determined to be 4×10^2 TCID₅₀/ml for all three assays. Inactivation is known to affect nucleic acid integrity and to shift the detection limit, usually about 10- to 100-fold higher, compared to that obtained with viable virus (10, 11). Studies performed by BioFire Defense with the BT-E panel indicated an LOD of 6×10^5 PFU per ml of whole blood with inactivated ZEBOV (9). BioFire Defense also reported successful BT-E panel detection of a synthetic ZEBOV L-gene RNA template in Tris-EDTA buffer at 1×10^0 to 1×10^5 genome equivalents (8). Studies at the CDC documented limits of detection for both the NP and VP40 qRT-PCR assays of 30 TCID₅₀/reaction ($\sim 5,400$ TCID₅₀/ml), using inactivated ZEBOV in both whole blood and urine (10, 11). The apparent differences likely have more to do with variations in the harshness of the protocols used to inactivate the virus in the analyte preparations than the ability of the test systems to detect ZEBOV. Our studies did not evaluate the specificity of the assays, but the EUA documentation includes results of extensive studies that were performed with the help of the U.S. Army Medical Research Institute of Infectious Diseases (USAMRIID) and the U.S. Department of Defense, which confirmed the high analytical specificities of all three tests (9–11). Further, clinical specificity was documented in a study in Sierra Leone in which the FilmArray assay was used for research purposes to test patients and health care workers who had been referred for diagnostic testing (12). In that study, 83 individuals were tested. Six of the individuals tested positive, with 5 of the 6 being confirmed as positive by CDC mobile laboratories using the NP2 and VP 40 tests in Sierra Leone. The remaining patient had a throat swab that tested positive in FilmArray testing, but the patient died with typical EVD symptoms before a specimen for con-

TABLE 2 Results of FilmArray BioThreat panel and qRT-PCR testing of clinical specimens throughout the course of disease for patients diagnosed with EVD^a

FilmArray result	No. with NP2 qRT-PCR result of:	
	Positive (<i>C_T</i> of ≤ 40)	Negative
Detected	40	4 ^b
Not detected	4 ^c	12

^a Testing was performed using both the FilmArray BioThreat panel and the CDC qRT-PCR assay, with whole-blood, plasma, and urine specimens.

^b Includes two samples that involved whole blood paired with plasma.

^c The *C_T* values for these four samples were 38, 39, 39, and 39.

TABLE 3 Comparison of FilmArray BioThreat panel and CDC NP2 qRT-PCR assay results for detection of Zaire ebolavirus in whole-blood, plasma, and urine specimens^a

Specimen type		No. of specimens evaluated	No. with result of:				Agreement (%)
RT	FA		RT+/FA+	RT-/FA-	RT+/FA-	RT-/FA+	
Whole blood	Whole blood	27	19	4	3	1	85
Plasma	Whole blood	20	13	5	0	2	90
Urine	Urine	13	8	3	1	1	85

^a Testing was performed using both the FilmArray (FA) BioThreat panel and the CDC NP2 qRT-PCR (RT) assay.

firmatory testing could be obtained. Notably, one asymptomatic health care worker tested positive, became symptomatic the following day, and also tested positive by the CDC assay with a blood specimen collected 4 days later. There were 19 asymptomatic individuals who had been exposed to patients with confirmed EVD but tested negative in FilmArray testing, did not meet the suspected EVD case definition, and so were not tested further. Whole-blood specimens from 57 symptomatic patients tested negative in FilmArray testing and were confirmed as negative by the CDC mobile laboratory. During follow-up monitoring, none of those patients developed EVD. One symptomatic patient with a urine specimen that tested negative in FilmArray testing was diagnosed as having EVD on the basis of a blood specimen that was tested 3 days later by the CDC mobile laboratory. This study showed high specificity and perfect correlation between the FilmArray and CDC assays when they were used as diagnostic tests with whole blood as the specimen type.

Testing of clinical specimens in the current study using FilmArray and qRT-PCR assays resulted in overall agreement of 87% (52/60 specimens), with agreements of 85% (23/27 specimens) for whole-blood specimens and 85% (11/13 specimens) for urine specimens. Eight discrepancies between FilmArray and qRT-PCR results were noted, with four from whole-blood specimens, two from plasma specimens, and two from urine specimens, which occurred when EVD was resolving and viral loads were waning, as indicated by patient improvement and qRT-PCR results with high or negative C_T values. Agreement between the FilmArray and NP2 qRT-PCR results, regardless of specimen type, was 100% (30/30 specimens) for testing of clinical specimens obtained initially, during the early stages of disease, and through early recovery in our patients. These specimens all had C_T values of 37 or lower. In the later stages of recovery, there was some variability between the assays. Of the 14 specimens with qRT-PCR C_T values of 38 or higher, 4 were not detected by FilmArray testing. Conversely, of

the 16 specimens that tested negative by qRT-PCR assay, 4 had detectable virus by FilmArray testing.

Specimen type, sample volume, handling, and storage are important to consider when interpreting the results of FilmArray and qRT-PCR testing for ZEBOV. Plasma and whole blood were tested with qRT-PCR testing, while only whole blood was evaluated with FilmArray testing. In this study, two discrepancies were noted when plasma specimens tested by qRT-PCR testing were negative for ZEBOV while corresponding whole-blood specimens were positive using FilmArray testing. Evaluation of six additional specimens corroborated the results and showed that this discordance occurred when viral loads were waning, as evidenced by C_T values of 36 or higher. This observation gives some insight into viral kinetics and might suggest that whole blood is a more-appropriate specimen type for ZEBOV detection, since corresponding plasma specimens may not contain virus at detectable levels. This is perhaps not unexpected, since monocytes are infected early in EVD and the plasma compartment clearing first has been shown for other RNA viruses (13, 14). Sample volume is also an important consideration for molecular detection of ZEBOV. We initially performed FilmArray testing with 100 μ l of whole blood, but the EUA FilmArray BT-E instructions for use indicate that 200 μ l of whole blood should be used (9). We tested two whole-blood specimens using the FilmArray assay to compare input volumes of 100 μ l and 200 μ l. Positive FilmArray results were obtained with both input volumes. The corresponding C_T values from the NP assay were 33 and 37. Specimens with higher C_T values were not tested. Four whole-blood specimens that tested negative using FilmArray testing were also negative in qRT-PCR testing. Finally, specimen handling and storage are important considerations for molecular detection of ZEBOV. Whole-blood specimens should be collected in plastic collection tubes, with EDTA, sodium polyanethol sulfonate, or citrate as preservative, and tested promptly following collection. If necessary, specimens may be stored for a short time at 4°C or frozen to prevent degradation of the specimens and viral nucleic acids. If shipping is required, then the specimens should be packaged and shipped at 2 to 8°C with cold packs, according to the Department of Transportation recommendations for a category A infectious substance, using an approved courier (15).

The unprecedented outbreak of ZEBOV in West Africa highlights the need for rapid, reliable, and easy-to-use tests for the detection of ZEBOV in clinical specimens. Data presented here suggest that the FilmArray BT and BT-E panels perform comparably to the CDC qRT-PCR assays for rapid detection of ZEBOV in blood and urine specimens from individuals suspected of having EVD. Reduced manipulation of clinical specimens, ease of use, and rapid turnaround time make this an appropriate screening test for health care institutions and public health laboratories that

TABLE 4 Detection of low-level ZEBOV viremia using whole blood versus plasma, in serial samples from a single patient^a

Sample no.	FilmArray results		NP2 qRT-PCR result for whole blood
	Whole blood	Plasma	
1	Detected	Detected	Positive ($C_T = 36$)
2	Detected	Not detected	Positive ($C_T = 36$)
3	Detected	Detected	Positive ($C_T = 37$)
4	Detected	Detected	Positive ($C_T = 37$)
5	Detected	Not detected	Positive ($C_T = 37$)
6	Detected	Not detected	Positive ($C_T = 40$)

^a Testing was performed using both the FilmArray BioThreat panel and the CDC NP2 qRT-PCR assay.

lack qRT-PCR capabilities but need the ability to provide presumptive identification of ZEBOV for individuals suspected of having EVD. According to the EUA, all FilmArray BT-E results should be confirmed by state public health laboratories or the CDC using EUA methods.

ACKNOWLEDGMENTS

None of the authors received external grant support for this study.

P.D.F. previously received grant support from BioFire to support clinical validation studies of the FilmArray respiratory and gastrointestinal panels.

The findings and conclusions in this report are those of the authors and do not necessarily represent the official position of the Centers for Disease Control and Prevention.

REFERENCES

- Feldmann H, Geisbert TW. 2011. Ebola haemorrhagic fever. *Lancet* 377:849–862. [http://dx.doi.org/10.1016/S0140-6736\(10\)60667-8](http://dx.doi.org/10.1016/S0140-6736(10)60667-8).
- Baize S, Pannetier D, Oestereich L, Rieger T, Koivogui L, Magassouba N, Soropogui B, Sow MS, Keita S, De Clerck H, Tiffany A, Dominguez G, Loua M, Traoré A, Kolié M, Malano ER, Heleze E, Bocquin A, Mély S, Raoul H, Caro V, Cadar D, Gabriel M, Pahlmann M, Tappe D, Schmidt-Chanasit J, Impouma B, Diallo AK, Formenty P, Van Herp M, Günther S. 2014. Emergence of Zaire Ebola virus disease in Guinea: preliminary report. *N Engl J Med* 371:1418–1425. <http://dx.doi.org/10.1056/NEJMoa1404505>.
- Centers for Disease Control and Prevention. 2015. 2014 Ebola outbreak in West Africa. <http://www.cdc.gov/vhf/ebola/outbreaks/2014-west-africa/index.html>. Accessed 20 June, 2015.
- Towner JS, Rollin PE, Bausch DG, Sanchez A, Crary SM, Vincent M, Lee WF, Spiropoulou CF, Ksiazek TG, Lukwiya M, Kaducu F, Downing R, Nichol ST. 2004. Rapid diagnosis of Ebola hemorrhagic fever by reverse transcription-PCR in an outbreak setting and assessment of patient viral load as a predictor of outcome. *J Virol* 78:4330–4341. <http://dx.doi.org/10.1128/JVI.78.8.4330-4341.2004>.
- He J, Kraft AJ, Fan J, Van Dyke M, Wang L, Bose ME, Khanna M, Metallo JA, Henrickson KJ. 2009. Simultaneous detection of CDC category “A” DNA and RNA bioterrorism agents by use of multiplex PCR & RT-PCR enzyme hybridization assays. *Viruses* 1:441–459. <http://dx.doi.org/10.3390/v1030441>.
- Lyon GM, Mehta AK, Varkey JB, Brantly K, Plyler L, McElroy AK, Kraft CS, Towner JS, Spiropoulou C, Ströher U, Uyeki TM, Ribner BS. 2014. Clinical care of two patients with Ebola virus disease in the United States. *N Engl J Med* 371:2402–2409. <http://dx.doi.org/10.1056/NEJMoa1409838>.
- Kraft CS, Hewlett AL, Koepsell S, Winkler AM, Kratochvil CJ, Larson L, Varkey JB, Mehta AK, Lyon GM, Friedman-Moraco RJ, Marconi VC, Hill CE, Sullivan JN, Johnson DW, Lisco SJ, Mulligan MJ, Uyeki TM, McElroy AK, Sealy T, Campbell S, Spiropoulou D, Ströher U, Crozier I, Sacra R, Connor MJ, Jr, Sueblivong V, Franch HA, Smith PW, Ribner BS. 22 April 2015. The use of TKM-100802 and convalescent plasma in 2 patients with Ebola virus disease in the United States. *Clin Infect Dis* <http://dx.doi.org/10.1093/cid/civ334>.
- Liddell AM, Davey RT, Jr, Mehta AK, Varkey JB, Kraft CS, Tseggay GK, Badidi O, Faust AC, Brown KV, Suffredini AF, Barrett K, Wolcott MJ, Marconi VC, Lyon GM, III, Weinstein GL, Weinmeister K, Sutton S, Hazbun M, Albariño CG, Reed Z, Cannon D, Ströher U, Feldman M, Ribner BS, Lane HC, Fauci AS, Uyeki TM. 12 May 2015. Characteristics and clinical management of a cluster of 3 patients with Ebola virus disease, including the first domestically acquired cases in the United States. *Ann Intern Med* <http://dx.doi.org/10.7326/M15-0530>.
- BioFire Defense. 2014. FilmArray™ BioThreat-E instructions for use. BioFire Defense, Salt Lake City, UT. <http://www.biofiredefense.com/media/RFIT-PRT-0302-FA-BioThreat-E-Instructions-for-Use.pdf>.
- Centers for Disease Control and Prevention. 2014. Ebola virus NP real-time RT-PCR assay: instructions for use. Centers for Disease Control and Prevention, Atlanta, GA. <http://www.fda.gov/downloads/MedicalDevices/Safety/EmergencySituations/UCM418810.pdf>.
- Centers for Disease Control and Prevention. 2014. Ebola virus VP40 real-time RT-PCR assay: instructions for use. Centers for Disease Control and Prevention, Atlanta, GA. <http://www.fda.gov/downloads/MedicalDevices/Safety/EmergencySituations/UCM418815.pdf>.
- Leski TA, Ansumana R, Taitt CR, Lamin JM, Bangura U, Lahai J, Mbayo G, Kanneh MB, Bawo B, Bockarie AS, Scullion M, Phillips CL, Horner CP, Jacobsen KH, Stenger DA. 13 May 2015. Use of the FilmArray system for detection of Zaire ebolavirus in a small hospital in Bo, Sierra Leone. *J Clin Microbiol* <http://dx.doi.org/10.1128/JCM.00527-15>.
- Ansari AA. 2014. Clinical features and pathobiology of Ebolavirus infection. *J Autoimmun* 55:1–9. <http://dx.doi.org/10.1016/j.jaut.2014.09.001>.
- Lanteri M, Lee T, Wen L, Kaidarova Z, Bravo MD, Kiely NE, Kamel HT, Tobler LH, Norris PJ, Busch MP. 2014. West Nile virus nucleic acid persistence in whole blood months after clearance in plasma: implication for transfusion and transplantation safety. *Transfusion* 54:3232–3241. <http://dx.doi.org/10.1111/trf.12764>.
- Centers for Disease Control and Prevention. 30 January 2015. Guidance for collection, transport and submission of specimens for Ebola virus testing. <http://www.cdc.gov/vhf/ebola/healthcare-us/laboratories/specimens.html>. Accessed 5 April, 2015.