



Published in final edited form as:

Bio Protoc. 2015 ; 5(18): .

## Measurement of Resting Energy Metabolism in Mice Using Oxymax Open Circuit Indirect Calorimeter

Yaohui Nie<sup>1,2</sup>, Timothy P. Gavin<sup>2</sup>, and Shihuan Kuang<sup>1,3,\*</sup>

<sup>1</sup>Department of Animal Sciences, Purdue University, West Lafayette, USA

<sup>2</sup>Department of Health and Kinesiology and Max E. Wastl Human Performance Laboratory, Purdue University, West Lafayette, USA

<sup>3</sup>Center for Cancer Research, Purdue University, West Lafayette, USA

### Abstract

Indirect calorimeter is a powerful tool to monitor resting energy metabolism through the measurement of oxygen (O<sub>2</sub>) consumption and carbon dioxide (CO<sub>2</sub>) production. From the measurement of VO<sub>2</sub> and VCO<sub>2</sub>, the respiratory exchange ratio (RER) can be calculated to assess energy fuel utilization and energy expenditure (Evan *et al.*, 2012). Previously, indirect calorimeter has been widely used in metabolic disease research in mice to reveal the potential roles of specific genes or treatments in regulating energy metabolism (for example: Bi *et al.*, 2014; Feng *et al.*, 2014). Here, we described a protocol to evaluate the resting energy metabolism of C57BL/6 mice during dark and light cycles using the Oxymax Open Circuit indirect calorimeter.

### Materials and Reagents

1. Adult mice (C57BL/6 male mice at 3-month old were used for data acquisition in this protocol, but male or female mice of other genetic backgrounds or strains, at different ages can be used)
2. Food (normal chow diet or high fat diet) and water (ad lib)
3. Compressed gas mixture with the components of 4,929 PPM CO<sub>2</sub>, 20.47% O<sub>2</sub> and Balance N<sub>2</sub>

### Equipment

1. Oxymax Open Circuit Indirect Calorimeter (Columbus Instruments, model: Open Circuit Indirect Calorimeter) (Figure 1)
2. Computer with software provided by the manufacture (Columbus Instruments, model: Oxymax v4.91)

\*For correspondence: skuang@purdue.edu.

## Procedure

1. Turn on the Oxymax instrument and computer, allowing the system to warm up for 2 h;
2. Start the Oxymax v4.91 program;
3. Perform CO<sub>2</sub> calibration;
  - a. Select calibration in the program, turn on the gas tank and set it to 5–10 psi;
  - b. Press the CO<sub>2</sub> button in “calibration” window (Figure 2a). Gain should be close to 1 after calibration (Figure 2b);
4. Perform O<sub>2</sub> calibration;
  - a. Press the O<sub>2</sub> button in “calibration” window (Figure 2a);
  - b. Desired O<sub>2</sub> level is listed in the first prompt window (Figure 2d). Use the fine and coarse knobs (Figure 2c) to adjust the O<sub>2</sub> level to desired level;
  - c. Turn off the gas tank after finishing the O<sub>2</sub> calibration.
5. Setup a new experiment, choose the chambers to be used, input the Identification number and weight of mice to be measured (Figure 3), only one mouse is allowed per chamber;
6. Setup the measurement schedule including number of Intervals, time of cage settle, cage measure, reference settle, reference measure, and reference method (Figure 4).
7. Place the mice in the corresponding calorimetric chambers, standard cages are used (Figure 5);
8. Fill the cage with sufficient food and water for a period of 24 h. Ensure that food and water are available *ad libitum* (Figure 5);
9. Check if there is any leakage in the system with software as follows:  
 Open *Tools* → Select *Sample Pump* from *Oxymax utility* → Ensure *Test in Valve* open and *N2* and *Ref Air/Cal Valve* closed → Turn *Sample Pump* ON → Select the chamber to be tested in *Expansion Interface* Disconnect drier input tubing from tested chamber → put the finger over it, If there is no system leaks, the ball on the front of the system sample pump will drop to 0; If not, check all air fittings to assure an air-tight connection and test it again (Figure 6).
10. Run the experiment for 24 h and export the data in Excel file format, which include the data listed as below:
  - a. O<sub>2</sub> consumption = (VO<sub>2</sub> input) – (VO<sub>2</sub> output), ml/kg/h;
  - b. CO<sub>2</sub> production = (VCO<sub>2</sub> output) – (VCO<sub>2</sub>) input, ml/kg/h;
  - c. Respiratory exchange ratio (RER) = VCO<sub>2</sub>/VO<sub>2</sub> Ratio;
  - d. Energy Expenditure (Heat production) = calorific value (Cv) x VO<sub>2</sub> =(3.815 + 1.232 x RER) x VO<sub>2</sub>, cal/h;


11. Close the experiment and return the mice to their home cages;
12. Turn off the system and clean the calorimetry with water and appropriate disinfectant.

## Notes


1. Begin data collection of mice after 1-day of acclimation in the metabolic chambers;
2. Install the Oxymax system under a constant environmental temperature (22 °C) and 12 h light (6 am–6 pm), 12 h dark cycle (6 pm–6 am);
3.  $\text{VO}_2$  and  $\text{VCO}_2$  were increased by approximately 15% and 28% in dark cycle, respectively;
4. RER was increased from 0.90 (light cycle) to 0.99 (dark cycle), suggesting a shift in macronutrient source from a mix of fat + carbohydrates to predominant carbohydrates in the dark cycle (the  $\text{VCO}_2/\text{VO}_2$  ratio of fatty acid oxidation is 0.7 and carbohydrates oxidation is 1.0);
5. Energy expenditure was 20% greater in dark than light cycle, with the most active phase of mice being between 7 pm–12 pm;
6. Indirect calorimetry is a versatile system to investigate alternations of metabolic rate under different conditions. For example:
  - a. To compare metabolic homeostasis and energy expenditure in wild type and mutant mice fed with normal chow diet or high-fat-diet;
  - b. To investigate changes in metabolic rate with aging, a potential indicator of improved health status.
  - c. Oxymax open circuit indirect calorimeter can also be incorporated with other chamber systems, such as activity, body mass, feeding, drinking, food access control, running wheel, urine collection, sleep detection, body core temperature and heart rate to fulfill different experimental designs.

## References


1. Bi P, Shan T, Liu W, Yue F, Yang X, Liang XR, Wang J, Li J, Carlesso N, Liu X, Kuang S. Inhibition of Notch signaling promotes browning of white adipose tissue and ameliorates obesity. *Nat Med.* 2014; 20(8):911–918. [PubMed: 25038826]
2. Even PC, Nadkarni NA. Indirect calorimetry in laboratory mice and rats: principles, practical considerations, interpretation and perspectives. *Am J Physiol Regul Integr Comp Physiol.* 2012; 303(5):R459–476. [PubMed: 22718809]
3. Feng B, Jiao P, Helou Y, Li Y, He Q, Walters MS, Salomon A, Xu H. Mitogen-activated protein kinase phosphatase 3 (MKP-3)-deficient mice are resistant to diet-induced obesity. *Diabetes.* 2014; 63(9):2924–2934. [PubMed: 24722245]




**Figure 1.**  
Open circuit indirect calorimeter components



**Figure 2.**  
CO<sub>2</sub> and O<sub>2</sub> calibration




**Figure 3.**  
Input of mouse information





**Figure 4. Setup schedule for measurement**

“0” for intervals means the experiment will run indefinitely until stopped by the user. “8” for reference methods means the experiment will measure the reference air after every 8 subjects.




**Figure 5.**  
Test chamber for mice (standard cage)





**Figure 6.**  
Gas leakage check



**Figure 7. O<sub>2</sub> consumption, CO<sub>2</sub> production, RER and Energy Expenditure of 3-month old C57BL6 mice during light and dark cycle**

The daily rhythms of metabolic parameters were recorded under a 12 h-light (open bar) and 12 h-dark cycle (black bar) (Left). Data were presented as Means ± SE ( $n = 5$ ) during light and dark cycle (right). \*,  $p < 0.05$ ; \*\*,  $p < 0.01$  analyzed by the Student's  $t$ -test (comparison of mean values between the light and dark cycles).