



HHS Public Access

Author manuscript

Abdom Radiol (NY). Author manuscript; available in PMC 2017 March 01.

Published in final edited form as:

Abdom Radiol (NY). 2016 March ; 41(3): 444–451. doi:10.1007/s00261-016-0680-4.

Magnetic Resonance Safety

Steffen Sammet, M.D., Ph.D., DABR, DABMRS, FAMP

Abstract

Magnetic Resonance Imaging (MRI) has a superior soft-tissue contrast compared to other radiological imaging modalities and its physiological and functional applications have led to a significant increase in MRI scans worldwide. A comprehensive MRI safety training to protect patients and other healthcare workers from potential bio-effects and risks of the magnetic fields in an MRI suite is therefore essential. The knowledge of the purpose of safety zones in an MRI suite as well as MRI appropriateness criteria is important for all healthcare professionals who will work in the MRI environment or refer patients for MRI scans.

The purpose of this article is to give an overview of current magnetic resonance safety guidelines and discuss the safety risks of magnetic fields in an MRI suite including forces and torque of ferromagnetic objects, tissue heating, peripheral nerve stimulation and hearing damages. MRI safety and compatibility of implanted devices, MRI scans during pregnancy and the potential risks of MRI contrast agents will also be discussed and a comprehensive MRI safety training to avoid fatal accidents in an MRI suite will be presented.

Keywords

Magnetic Resonance Imaging (MRI); Safety; Magnetic Fields; Implants; Pregnancy; Contrast Agents

Introduction

The increasing clinical demand for Magnetic Resonance Imaging (MRI) with its superior soft-tissue contrast compared to other radiological imaging modalities and potential physiological and functional applications has contributed to the installation of almost 30,000 MRI scanners worldwide. Therefore, more and more healthcare professionals need to be trained in MRI safety to protect patients and other healthcare workers from the potential risks of MRI [1,2]. It is also important that radiologists, referring physicians and MR technologists are able to evaluate MRI safety and compatibility of medical devices and implants because they are often the first health care professionals who will talk to a patient about an MRI exam, potential risks, and MRI safety [3].

Radiologists are well trained about MRI appropriateness criteria but they require support from referring physicians to estimate the risks and benefits of MR imaging procedures [4].

Corresponding Author: Steffen Sammet, M.D, Ph.D., DABR, DABRMS, FAMP, Associate Professor of Radiology and Medical Physics, Director of Clinical MR Physics, University of Chicago Medicine, Department of Radiology, 5841 South Maryland Avenue, MC2026, Chicago, Illinois 60637, USA, Tel.: +1 (773) 702-3162, Fax: +1 (773) 702-1161, ssammet@uchicago.edu.

Especially referring physicians who know details of a patient's medical history can improve the MRI safety screening process when they are aware of the risks of an MRI scan by pre-screening their patients before an MRI exam [5]. This is particularly important in high-risk patients and in patients with new implants that have not yet been tested for MRI compatibility [6,7].

An expert panel has developed the American College of Radiology (ACR) Guidance Document for Safe MR practices [8]. The following sections will review potential bio-effects and risks of the magnetic fields that interact with patients and health care professionals in an MRI suite [9].

Magnetic Fields in an MRI suite

There are three major magnetic fields in an MRI suite that that have potential safety risks [10–14]:

1. The static magnetic field B_0 of clinical MRI scanners that ranges from 0.2T to 3T [15–17]. B_0 is orders of magnitudes larger than the magnetic field of the earth and can torque, attract and accelerate ferromagnetic objects in direction of the opening of the bore of an MRI scanner. B_0 can also interfere with implanted devices such as pumps and pace makers [18–26,1,27,28,2,29,30,15,31,12,16,32–47,7,48–62,11,63–73].
2. The radiofrequency (RF) field B_1 that is in the order of μT [74,30] and is produced by RF-coils. It can potentially cause tissue heating, especially when implants are present [19,75–89,16,90–100,4,101–103,10,104–116,69,117].
3. The magnetic field gradients with amplitudes in the order of 100 mT/m and with slew rates up to 200 mT/m/ms [63,118]. The fast switching gradient fields are applied for spatial encoding of the MRI signal and can cause peripheral nerve stimulation and implant heating. They are also responsible for the noise in the MRI scanner room, which can reach levels of 100 dB or more and potentially lead to hearing damages [74,119–121,118,122–126,30,31,127,128,16,90,129,37,130–133,57,58,134,135,69].

MRI Zones

In the ACR guidance document on MR safe practices from 2013 four different zones are suggested around the MRI scanner [136,137]. The access to these zones is restricted in MRI facilities and hospitals and the boundary of each zone in this four-zone safety system is defined by its purpose and distance from the MRI scanner [30]. Some zones may extend into other areas or floors of the facility due to the three dimensional extend of the magnetic field [138].

Zone I includes all areas freely accessible to the general public where the magnet field poses no hazards, such as the entrance to the MR facility.

Zone II is located between Zone I and the more restrictive Zone III. In Zone II patients are under general supervision of MR personnel. Zone II often includes the reception area, dressing rooms and MRI screening rooms.

Zone III is access-restricted by physical barriers such as doors with coded access. Inside Zone III, only approved MR personnel and patients that have undergone MRI screening are allowed. The MR control room is in Zone III.

Zone IV is the room where the magnet is located. Access to Zone IV should only be possible by passing through Zone III. Zone IV is designed so that the walls of magnet room contain the five 0.5 mT line (or 5 Gauss) line of the fringe field of the magnet.

The 5 Gauss line of an MRI suite defines a border to an area in which the magnetic field could affect implanted devices such as pacemakers [7]. Special warning signs about the strong magnetic field and its associated hazards need to be set up in the MRI facility. An MR safety program should be established to train employees about the dangers of the magnetic fields in the MRI suite [139] and to warn about potential interferences of the magnetic fringe field with implanted devices [140].

Attractive forces and torque on ferromagnetic objects by the static magnetic field B_0

The static magnetic field B_0 of an MRI machine attracts ferromagnetic objects and accelerates them toward the center of the bore of the MRI scanner. Ferromagnetic objects such as coins, hairpins, steel oxygen tanks or scissors can be accelerated or torqued by B_0 [4,10] and become dangerous projectiles [51]. The MRI safety program of the facility needs to warn about the misconception that larger objects will resist attraction to the field and need to emphasize the relationship between object size, material components, and projectile risk. Insufficient MRI safety training of ancillary medical personnel has led to fatal accidents when medical and other equipment was accelerated into the bore of the magnet [50,51].

Thermal effects induced by the radiofrequency field B_1

Physicians and MRI technologists need to be aware of the bio-effects of radiofrequency fields [99] because RF-fields can cause tissue heating in the human body [100,105]. The specific absorption rate (SAR) is a measure of the RF-power absorbed per mass of tissue and has the units of watts per kilogram (W/kg) [75]. The absorbed RF-energy is transformed into heat in the human body [141,110] and especially infants, children and patients with thermo-regulatory disorders might experience an increase in the body core temperature due to RF-induced heating during an MRI exam [76,141,114,142].

A proper preparation of each patient before an MRI exam is necessary to avoid burns even for patients without implants. The Guidelines to Prevent Excessive Heating and Burns Associated with Magnetic Resonance Procedures [143] recommend to

1. remove metallic objects contacting the patient's skin (e.g., jewelry, drug delivery patches with metallic components),

2. use insulation material of 1 cm or thicker to prevent skin-to-skin contact and the formation of closed-loops from touching body parts,
3. allow only devices, equipment, accessories (e.g., ECG leads, electrodes), and materials that have been thoroughly tested and determined to be MRI safe [6,144,7].

Peripheral nerve stimulation and hearing damage caused by the gradient system

The time-varying fast-switching gradient magnetic fields may stimulate nerves or muscles in patients by inducing electrical fields [121]. There are multiple factors influencing the interactions of gradient fields with biological tissues and they depend on the frequency of the gradient field, the maximum and average flux densities, the presence of harmonic frequencies, the waveform characteristics of the signal, the polarity of the signal, the current distribution in the body, the electrical properties, and the sensitivity of the cell membrane [119–121,118,126,31,128,37,131,133,63,134]. The acoustic noise during an MRI exam is also caused by the gradient system [121,118]. Head phones and earplugs are essential hearing protections for patients during and anybody present during an MRI examination [37].

MRI screening procedures

MRI screening before any MRI exam is indispensable and evaluates the magnetic and geometric properties of implants or foreign bodies, and their potential interactions with the magnetic fields in an MRI system [16,145,51,4,10,146,147]. Before anybody is allowed to enter the MRI suite, it is essential to remove all objects that could potentially interact with the magnetic fields [50,51,58,11]. It is recommended that patients will wear gowns in the MRI environment to avoid metallic fasteners, loose metallic components, or metallic threads on clothing in the MRI suite [148].

Implanted Devices

It is essential for any referring physician, radiologist or MR technologist to know where to find details about the MRI safety and compatibility of medical implants and devices [7,6,144]. Frank Shellock, Ph.D. and his team work on MRI safety and have published several books and a searchable on-line catalog that lists MRI-safe devices and MRI-safe implants with their allowed magnetic field strengths and gradient limitations (http://www.mrisafety.com/TheList_search.asp) [100,4,101–103,58,10,104,59,105,60,106,107]. When a patient with implants is scheduled for an MRI exam it is vital to understand that the injury risk increases with the proximity of the implants to vital vascular, neural or soft tissue structures [58]. Implants, medical materials, and devices from the US and Europe that were manufactured in last three decades are made from non-ferromagnetic materials and are usually labeled *MR safe* or *MR conditional* [100]. Other implants or devices have to be considered *MR unsafe* and are contraindicated for an MRI scan. Interactions of these objects with the magnetic fields can cause severe artifacts and heating [7]. Even non-ferromagnetic

implants can cause heating due to eddy currents that propagate in metals exposed to oscillating magnetic fields [102,103]. Especially orthopedic implants such as external fixation systems can cause heating in an MRI scanner [6,144,7,145,149,4,103,10,107] and some of the new implants or devices have not been tested or labeled for MRI compatibility [150]. It is therefore essential to establish a close collaboration between the radiology team, the ordering physician, and the physician who placed the implant [13].

MRI during Pregnancy

The practice guidelines collaboratively developed by the American College of Radiology (ACR) and the Society for Pediatric Radiology (SPR) recommends no special consideration for any trimester in pregnancy [151] since there is no evidence in the current literature for deleterious effects of MR imaging at 1.5 T on the developing fetus [152–154]. The risk to benefit ratio to perform an MR scan on a pregnant patient needs to be evaluated by a level 2 MR personnel-designated attending radiologist [139,151]. The radiologist should discuss and agree with the referring physician to perform the magnetic resonance imaging exam. Patients can then undergo an MRI scan in any stage of the pregnancy and the following statements should be added to the radiology report or to the patients' medical record [151]:

1. The information requested from the MRI study cannot be acquired by ultrasonography.
2. The data are needed to potentially affect the care of the patient or fetus during the pregnancy.
3. The referring physician does not feel it is prudent to wait until the patient is no longer pregnant to obtain these data.

The ACR Manual on Contrast Media recommends that each case should be reviewed carefully by members of the clinical and radiology team and gadolinium based contrast agents should be administered only when there is a potential significant benefit to the patient or fetus that outweighs the possible but unknown risk of fetal exposure to free gadolinium ions [136]. Gadolinium based contrast agents are consequently classified as pregnancy class C drugs because potential benefits may warrant the use in pregnant women despite potential risks [8,155].

MR Contrast Agents

FDA approved MRI contrast agents are gadolinium chelates with different stability, viscosity, and osmolality [156].

Radiologists and MRI technologists should be aware of adverse effects of gadolinium based contrast agents. Gadolinium chelates are in general well tolerated and acute adverse reactions are observed with a lower frequency than after the administration of iodinated contrast media [157]. The reported frequencies of all acute adverse events after an injection of 0.1 or 0.2 mmol/kg of gadolinium chelate range from 0.07% to 2.4% [156,158]. Coldness at the injection site, nausea with or without vomiting, headache, warmth or pain at the injection site, paresthesias, dizziness, and itching are potential adverse effects after the

administration of gadolinium chelates [159,160,157,161–167,158,156,168–170]. Severe allergic reactions including rash hives, urticaria and bronchospasm range from 0.004% to 0.7% [160,161]. Severe, life-threatening anaphylactoid or nonallergic anaphylactic reactions are exceedingly rare with frequencies between 0.001% and 0.01% [160,162,164,168]. In an accumulated series of 687,000 doses there were only 5 severe reactions and fatal reactions to gadolinium chelate are extremely rare [156,160,157,161].

Nevertheless, it is important to know that Gadolinium chelates administered to patients with acute renal failure or severe chronic kidney disease can result in a syndrome of nephrogenic systemic fibrosis (NSF) [159]. There are no reports of NSF in patients with normal kidney function, therefore, the U.S. Food and Drug Administration (FDA) requires that manufacturers include a boxed warning on the product labeling of all gadolinium-based contrast agents that patients with severe kidney insufficiency who receive gadolinium-based agents are at risk for developing NSF [157,160].

Claustrophobia during MRI exams

Patients can become claustrophobic during an MRI scan and refuse to complete the exam especially when they are not well informed about the procedure [171,172]. Larger bore sizes and Open MRI systems can help to reduce claustrophobia [173,174]. Referring physicians can help to reduce claustrophobia by discussing the details of the MRI procedure with the patients before their exam [175,176].

Sedation and Anesthesia for MRI exams

Non-compliant patients such as children or claustrophobic adults may require sedation or anesthesia during an MRI exam because MRI scans often require long scan times and are sensitive to motion [177]. Patients with pain can benefit from pain management to help them to remain motionless during the MRI exam. An MRI safety program should emphasize specific risks and benefits of sedation, anesthesia, and pain management in these patients [178].

MRI Safety Training and Emergency Procedures in an MRI Suite

All personnel working in the MR environment need to be trained with a comprehensive MRI safety course [13]. For new employees who will work in the MR environment this course should be included in the employee orientation program and be repeated annually. The MRI safety training should include the presentation of technical and medical background of MRI safety. Hands-on demonstrations of missile effects of ferromagnetic objects can help to better understand and experience the dangers in an MRI suite [9,13]. Detailed screening procedures of patients with a questionnaire for ferromagnetic objects, implants, devices, body piercing, allergies to MRI contrast agents, kidney disease, pregnancy, breast feeding and also the screening of patients that have a history injuries by a metallic foreign body such as bullets, shrapnel, or other type of metallic fragments help to avoid severe accidents in an MRI suite. An important topic to discuss in an MRI safety course are severe burn wounds that were experienced by patients when limbs or other body parts of the patients were in

direct contact with transmit RF coils of the MR systems or when skin-to-skin contact points were responsible for these injuries. The safety course needs to warn about high acoustic noise levels of the gradient system during an MRI scan and the potential noise reduction with earplugs and headphones to avoid potential hearing damage. Videos from quenching magnets can help to understand how powerful a sudden loss of the superconductivity of the magnet might be and emergency procedures during a quench should be discussed. It is important that medical personal entering the MRI scanner room to evaluate the patient, administer medications or interventions need to be trained in emergency procedures in an MRI suite [140]. Healthcare professionals need to know which objects can be brought into the different MRI zones in order to prevent fatal injuries and medical equipment failure and how to remove a patient from the MRI magnet room to resuscitate or treat the patient in emergency cases [145].

Acknowledgments

Steffen Sammet, M.D., Ph.D., DABR, DABMRS, FAMP is funded by the National Institute of Neurological Disorders and Stroke (R25NS080949) National Cancer Institute Education and Career Development program R25 Cancer Nanotechnology in Imaging and Radiotherapy (R25CA132822), the Cancer Research Foundation, the University of Chicago Comprehensive Cancer Center and Philips Healthcare.

References

1. Chakeres DW, de Vocht F. Static magnetic field effects on human subjects related to magnetic resonance imaging systems. *Progress in biophysics and molecular biology*. 2005; 87(2–3):255–265.10.1016/j.pbiomolbio.2004.08.012 [PubMed: 15556664]
2. Feychting M. Health effects of static magnetic fields--a review of the epidemiological evidence. *Progress in biophysics and molecular biology*. 2005; 87(2–3):241–246.10.1016/j.pbiomolbio.2004.08.007 [PubMed: 15556662]
3. Hipp, E.; Sammet, S.; Straus, C. MR Safety Standards for Medical Students Nationwide. *Proceedings of the 19th Annual Meeting of ISMRM; Melbourne, Australia*. 2012. p (abstract 2731)
4. Shellock, FG. *Magnetic resonance procedures : health effects and safety*. CRC Press; Boca Raton: 2001.
5. Karpowicz J, Gryz K. Health risk assessment of occupational exposure to a magnetic field from magnetic resonance imaging devices. *International journal of occupational safety and ergonomics : JOSE*. 2006; 12(2):155–167. [PubMed: 16790173]
6. ASTM F2052-00 Standard Test Method for Measurement of Magnetically Induced Displacement Force on Passive Implants in the Magnetic Resonance Environment.
7. Sammet CL, Yang X, Wassenaar PA, Bourekas EC, Yuh BA, Shellock F, Sammet S, Knopp MV. RF-related heating assessment of extracranial neurosurgical implants at 7T. *Magnetic resonance imaging*. 2013; 31(6):1029–1034.10.1016/j.mri.2012.10.025 [PubMed: 23643158]
8. Kanal E, Barkovich AJ, Bell C, Borgstede JP, Bradley WG Jr, Froelich JW, Gimbel JR, Gosbee JW, Kuhni-Kaminski E, Larson PA, Lester JW Jr, Nyenhuis J, Schaefer DJ, Sebek EA, Weinreb J, Wilkoff BL, Woods TO, Lucey L, Hernandez D. Expert Panel on MRS. ACR guidance document on MR safe practices: 2013. *Journal of magnetic resonance imaging : JMRI*. 2013; 37(3):501–530.10.1002/jmri.24011 [PubMed: 23345200]
9. Sammet, S. Implementation of a Comprehensive MR Safety Course for Medical Students. *Proceedings of the 20th Annual Meeting of ISMRM; Salt Lake City*. 2013. p (abstract 4071)
10. Shellock, FG.; Kanal, E. *Magnetic resonance : bioeffects, safety, and patient management*. 2. Lippincott-Raven; Philadelphia: 1996.
11. Simmons A, Hakansson K. Magnetic resonance safety. *Methods in molecular biology*. 2011; 711:17–28.10.1007/978-1-61737-992-5_2 [PubMed: 21279596]

12. Hansson Mild K, Hand J, Hietanen M, Gowland P, Karpowicz J, Keevil S, Lagroye I, van Rongen E, Scarfi MR, Wilen J. Exposure classification of MRI workers in epidemiological studies. *Bioelectromagnetics*. 2013; 34(1):81–84.10.1002/bem.21728 [PubMed: 22532229]
13. Sammet S, Sammet CL. Implementation of a comprehensive MR safety course for medical students. *Journal of magnetic resonance imaging : JMRI*. 201510.1002/jmri.24993
14. Tsai LL, Grant AK, Morteale KJ, Kung JW, Smith MP. A Practical Guide to MR Imaging Safety: What Radiologists Need to Know. *Radiographics*. 2015; 35(6):1722–1737.10.1148/rg.2015150108 [PubMed: 26466181]
15. Ghodbane S, Lahbib A, Sakly M, Abdelmelek H. Bioeffects of static magnetic fields: oxidative stress, genotoxic effects, and cancer studies. *BioMed research international*. 2013; 2013:602987.10.1155/2013/602987 [PubMed: 24027759]
16. Hartwig V, Giovannetti G, Vanello N, Lombardi M, Landini L, Simi S. Biological effects and safety in magnetic resonance imaging: a review. *International journal of environmental research and public health*. 2009; 6(6):1778–1798.10.3390/ijerph6061778 [PubMed: 19578460]
17. Glover PM, Cavin I, Qian W, Bowtell R, Gowland PA. Magnetic-field-induced vertigo: a theoretical and experimental investigation. *Bioelectromagnetics*. 2007; 28(5):349–361.10.1002/bem.20316 [PubMed: 17427890]
18. Heinrich A, Szostek A, Nees F, Meyer P, Semmler W, Flor H. Effects of static magnetic fields on cognition, vital signs, and sensory perception: a meta-analysis. *Journal of magnetic resonance imaging : JMRI*. 2011; 34(4):758–763.10.1002/jmri.22720 [PubMed: 21751291]
19. International Electrotechnical Commission (IEC), Medical Electrical Equipment. International Standard IEC 6060, 1-2-33, 2002. Nov 8. 2015 Particular requirements for the safety of magnetic resonance equipment for medical diagnosis.
20. Atkinson IC, Renteria L, Burd H, Pliskin NH, Thulborn KR. Safety of human MRI at static fields above the FDA 8 T guideline: sodium imaging at 9.4 T does not affect vital signs or cognitive ability. *Journal of magnetic resonance imaging : JMRI*. 2007; 26(5):1222–1227.10.1002/jmri.21150 [PubMed: 17969172]
21. Atkinson IC, Sonstegaard R, Pliskin NH, Thulborn KR. Vital signs and cognitive function are not affected by 23-sodium and 17-oxygen magnetic resonance imaging of the human brain at 9.4 T. *Journal of magnetic resonance imaging : JMRI*. 2010; 32(1):82–87.10.1002/jmri.22221 [PubMed: 20578014]
22. Besson JA, Foreman EI, Eastwood LM, Smith FW, Ashcroft GW. Cognitive evaluation following NMR imaging of the brain. *Journal of neurology, neurosurgery, and psychiatry*. 1984; 47(3):314–316.
23. Brockway JP, Bream PR Jr. Does memory loss occur after MR imaging? *Journal of magnetic resonance imaging : JMRI*. 1992; 2(6):721–728. [PubMed: 1446117]
24. Brody AS, Embury SH, Mentzer WC, Winkler ML, Gooding CA. Preservation of sickle cell blood-flow patterns during MR imaging: an in vivo study. *AJR American journal of roentgenology*. 1988; 151(1):139–141.10.2214/ajr.151.1.139 [PubMed: 3259797]
25. Brody AS, Sorette MP, Gooding CA, Listerud J, Clark MR, Mentzer WC, Brasch RC, James TL. AUR memorial Award. Induced alignment of flowing sickle erythrocytes in a magnetic field. A preliminary report. *Investigative radiology*. 1985; 20(6):560–566. [PubMed: 2933361]
26. Cavin ID, Glover PM, Bowtell RW, Gowland PA. Thresholds for perceiving metallic taste at high magnetic field. *Journal of magnetic resonance imaging : JMRI*. 2007; 26(5):1357–1361.10.1002/jmri.21153 [PubMed: 17969179]
27. de Vocht F, Stevens T, Glover P, Sunderland A, Gowland P, Kromhout H. Cognitive effects of head-movements in stray fields generated by a 7 Tesla whole-body MRI magnet. *Bioelectromagnetics*. 2007; 28(4):247–255.10.1002/bem.20311 [PubMed: 17290435]
28. de Vocht F, Stevens T, van Wendel-de-Joode B, Engels H, Kromhout H. Acute neurobehavioral effects of exposure to static magnetic fields: analyses of exposure-response relations. *Journal of magnetic resonance imaging : JMRI*. 2006; 23(3):291–297.10.1002/jmri.20510 [PubMed: 16463303]

29. Franco G, Mora E, Perduri R. Focusing ethical dilemmas of evidence-based practice in SMF-exposed MRI-workers: a qualitative analysis. *International archives of occupational and environmental health*. 2010; 83(4):417–421.10.1007/s00420-009-0476-8 [PubMed: 19888595]
30. Fuentes MA, Trakic A, Wilson SJ, Crozier S. Analysis and measurements of magnetic field exposures for healthcare workers in selected MR environments. *IEEE transactions on bio-medical engineering*. 2008; 55(4):1355–1364.10.1109/TBME.2007.913410 [PubMed: 18390326]
31. Glover PM. Interaction of MRI field gradients with the human body. *Physics in medicine and biology*. 2009; 54(21):R99–R115.10.1088/0031-9155/54/21/R01 [PubMed: 19826206]
32. Heinrich A, Szostek A, Meyer P, Nees F, Rauschenberg J, Grobner J, Gilles M, Paslakis G, Deuschle M, Semmler W, Flor H. Cognition and sensation in very high static magnetic fields: a randomized case-crossover study with different field strengths. *Radiology*. 2013; 266(1):236–245.10.1148/radiol.12112172 [PubMed: 23091174]
33. Hong CZ, Shellock FG. Short-term exposure to a 1.5 tesla static magnetic field does not affect somato-sensory-evoked potentials in man. *Magnetic resonance imaging*. 1990; 8(1):65–69. [PubMed: 2325519]
34. Hsieh CH, Lee MC, Tsai-Wu JJ, Chen MH, Lee HS, Chiang H, Herbert Wu CH, Jiang CC. Deleterious effects of MRI on chondrocytes. *Osteoarthritis and cartilage/OARS, Osteoarthritis Research Society*. 2008; 16(3):343–351.10.1016/j.joca.2007.07.001
35. Innis NK, Ossenkopp KP, Prato FS, Sestini E. Behavioral effects of exposure to nuclear magnetic resonance imaging: II. Spatial memory tests. *Magnetic resonance imaging*. 1986; 4(4):281–284. [PubMed: 3669943]
36. Kangarlu A, Burgess RE, Zhu H, Nakayama T, Hamlin RL, Abduljalil AM, Robitaille PM. Cognitive, cardiac, and physiological safety studies in ultra high field magnetic resonance imaging. *Magnetic resonance imaging*. 1999; 17(10):1407–1416. [PubMed: 10609989]
37. Kannala S, Toivo T, Alanko T, Jokela K. Occupational exposure measurements of static and pulsed gradient magnetic fields in the vicinity of MRI scanners. *Physics in medicine and biology*. 2009; 54(7):2243–2257.10.1088/0031-9155/54/7/026 [PubMed: 19293469]
38. Karpowicz J, Gryz K, Politanski P, Zmyslony M. Exposure to static magnetic field and health hazards during the operation of magnetic resonance scanners. *Medycyna pracy*. 2011; 62(3):309–321. [PubMed: 21870421]
39. Karpowicz J, Zradzinski P, Gryz K. Measures of occupational exposure to time-varying low frequency magnetic fields of non-uniform spatial distribution in the light of international guidelines and electrodynamic exposure effects in the human body. *Medycyna pracy*. 2012; 63(3): 317–328. [PubMed: 22880453]
40. Kay HH, Herfkens RJ, Kay BK. Effect of magnetic resonance imaging on *Xenopus laevis* embryogenesis. *Magnetic resonance imaging*. 1988; 6(5):501–506. [PubMed: 2465471]
41. Laszlo J, Gyires K. 3 T homogeneous static magnetic field of a clinical MR significantly inhibits pain in mice. *Life sciences*. 2009; 84(1–2):12–17.10.1016/j.lfs.2008.10.009 [PubMed: 19000698]
42. McRobbie DW. Occupational exposure in MRI. *The British journal of radiology*. 2012; 85(1012): 293–312.10.1259/bjr/30146162 [PubMed: 22457400]
43. Muller S, Hotz M. Human brainstem auditory evoked potentials (BAEP) before and after MR examinations. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 1990; 16(3):476–480.
44. Ossenkopp KP, Innis NK, Prato FS, Sestini E. Behavioral effects of exposure to nuclear magnetic resonance imaging: I. Open-field behavior and passive avoidance learning in rats. *Magnetic resonance imaging*. 1986; 4(4):275–280. [PubMed: 3669942]
45. Patel M, Williamsom RA, Dorevitch S, Buchanan S. Pilot study investigating the effect of the static magnetic field from a 9.4-T MRI on the vestibular system. *Journal of occupational and environmental medicine/American College of Occupational and Environmental Medicine*. 2008; 50(5):576–583.10.1097/JOM.0b013e318162f5d6 [PubMed: 18469627]
46. Prasad N, Wright DA, Ford JJ, Thornby JI. Safety of 4-T MR imaging: study of effects on developing frog embryos. *Radiology*. 1990; 174(1):251–253.10.1148/radiology.174.1.2294557 [PubMed: 2294557]

47. Sakurai T, Terashima S, Miyakoshi J. Effects of strong static magnetic fields used in magnetic resonance imaging on insulin-secreting cells. *Bioelectromagnetics*. 2009; 30(1):1–8.10.1002/bem.20433 [PubMed: 18521845]
48. Schaap K, Christopher-De Vries Y, Slottje P, Kromhout H. Inventory of MRI applications and workers exposed to MRI-related electromagnetic fields in the Netherlands. *European journal of radiology*. 2013; 82(12):2279–2285.10.1016/j.ejrad.2013.07.023 [PubMed: 24055183]
49. Schenck JF. Health and physiological effects of human exposure to whole-body four-tesla magnetic fields during MRI. *Annals of the New York Academy of Sciences*. 1992; 649:285–301. [PubMed: 1580500]
50. Schenck JF. MR safety at high magnetic fields. *Magnetic resonance imaging clinics of North America*. 1998; 6(4):715–730. [PubMed: 9799852]
51. Schenck JF. Safety of strong, static magnetic fields. *Journal of magnetic resonance imaging : JMRI*. 2000; 12(1):2–19. [PubMed: 10931560]
52. Schenck JF, Dumoulin CL, Redington RW, Kressel HY, Elliott RT, McDougall IL. Human exposure to 4.0-Tesla magnetic fields in a whole-body scanner. *Medical physics*. 1992; 19(4): 1089–1098. [PubMed: 1518472]
53. Schlamann M, Voigt MA, Maderwald S, Bitz AK, Kraff O, Ladd SC, Ladd ME, Forsting M, Wilhelm H. Exposure to high-field MRI does not affect cognitive function. *Journal of magnetic resonance imaging : JMRI*. 2010; 31(5):1061–1066.10.1002/jmri.22065 [PubMed: 20432339]
54. Schlamann M, Yoon MS, Maderwald S, Pietrzyk T, Bitz AK, Gerwig M, Forsting M, Ladd SC, Ladd ME, Kastrup O. Short term effects of magnetic resonance imaging on excitability of the motor cortex at 1.5T and 7T. *Academic radiology*. 2010; 17(3):277–281.10.1016/j.acra.2009.10.004 [PubMed: 20036585]
55. Schwartz JL, Crooks LE. NMR imaging produces no observable mutations or cytotoxicity in mammalian cells. *AJR American journal of roentgenology*. 1982; 139(3):583–585.10.2214/ajr.139.3.583 [PubMed: 6981326]
56. Schwenzer NF, Bantleon R, Maurer B, Kehlbach R, Herberts T, Claussen CD, Rodegerdts E. In vitro evaluation of magnetic resonance imaging at 3.0 tesla on clonogenic ability, proliferation, and cell cycle in human embryonic lung fibroblasts. *Investigative radiology*. 2007; 42(4):212–217.10.1097/01.rli.0000255831.40115.83 [PubMed: 17351426]
57. Schwenzer NF, Bantleon R, Maurer B, Kehlbach R, Schraml C, Claussen CD, Rodegerdts E. Do static or time-varying magnetic fields in magnetic resonance imaging (3.0 T) alter protein-gene expression?-A study on human embryonic lung fibroblasts. *Journal of magnetic resonance imaging : JMRI*. 2007; 26(5):1210–1215.10.1002/jmri.21145 [PubMed: 17969170]
58. Shellock FG, Crues JV. MR procedures: biologic effects, safety, and patient care. *Radiology*. 2004; 232(3):635–652.10.1148/radiol.2323030830 [PubMed: 15284433]
59. Shellock FG, Schaefer DJ, Crues JV. Exposure to a 1.5-T static magnetic field does not alter body and skin temperatures in man. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 1989; 11(3):371–375.
60. Shellock FG, Schaefer DJ, Gordon CJ. Effect of a 1.5 T static magnetic field on body temperature of man. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 1986; 3(4):644–647.
61. Short WO, Goodwill L, Taylor CW, Job C, Arthur ME, Cress AE. Alteration of human tumor cell adhesion by high-strength static magnetic fields. *Investigative radiology*. 1992; 27(10):836–840. [PubMed: 1399440]
62. Silva AK, Silva EL, Egito ES, Carrico AS. Safety concerns related to magnetic field exposure. *Radiation and environmental biophysics*. 2006; 45(4):245–252.10.1007/s00411-006-0065-0 [PubMed: 17021785]
63. Tomasi DG, Wang R. Induced magnetic field gradients and forces in the human head in MRI. *Journal of magnetic resonance imaging : JMRI*. 2007; 26(5):1340–1345.10.1002/jmri.21143 [PubMed: 17969175]

64. Toyomaki A, Yamamoto T. Observation of changes in neural activity due to the static magnetic field of an MRI scanner. *Journal of magnetic resonance imaging : JMRI*. 2007; 26(5):1216–1221.10.1002/jmri.21151 [PubMed: 17969171]
65. Valiron O, Peris L, Rikken G, Schweitzer A, Saoudi Y, Remy C, Job D. Cellular disorders induced by high magnetic fields. *Journal of magnetic resonance imaging : JMRI*. 2005; 22(3):334–340.10.1002/jmri.20398 [PubMed: 16106367]
66. van Nierop LE, Slottje P, van Zandvoort MJ, de Vocht F, Kromhout H. Effects of magnetic stray fields from a 7 tesla MRI scanner on neurocognition: a double-blind randomised crossover study. *Occupational and environmental medicine*. 2012; 69(10):759–766.10.1136/oemed-2011-100468 [PubMed: 22930737]
67. Vaughan T, DelaBarre L, Snyder C, Tian J, Akgun C, Shrivastava D, Liu W, Olson C, Adriany G, Strupp J, Andersen P, Gopinath A, van de Moortele PF, Garwood M, Ugurbil K. 9.4T human MRI: preliminary results. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 2006; 56(6):1274–1282.10.1002/mrm.21073
68. Vogl TJ, Paulus W, Fuchs A, Krafczyk S, Lissner J. Influence of magnetic resonance imaging on evoked potentials and nerve conduction velocities in humans. *Investigative radiology*. 1991; 26(5):432–437. [PubMed: 2055741]
69. Weintraub MI, Houry A, Cole SP. Biologic effects of 3 Tesla (T) MR imaging comparing traditional 1.5 T and 0.6 T in 1023 consecutive outpatients. *Journal of neuroimaging : official journal of the American Society of Neuroimaging*. 2007; 17(3):241–245.10.1111/j.1552-6569.2007.00118.x [PubMed: 17608910]
70. Weiss J, Herrick RC, Taber KH, Contant C, Plishker GA. Bio-effects of high magnetic fields: a study using a simple animal model. *Magnetic resonance imaging*. 1992; 10(4):689–694. [PubMed: 1501539]
71. Yamaguchi-Sekino S, Nakai T, Imai S, Izawa S, Okuno T. Occupational exposure levels of static magnetic field during routine MRI examination in 3 T MR system. *Bioelectromagnetics*. 2013.10.1002/bem.21817
72. Yamaguchi-Sekino S, Sekino M, Ueno S. Biological effects of electromagnetic fields and recently updated safety guidelines for strong static magnetic fields. *Magnetic resonance in medical sciences : MRMS : an official journal of Japan Society of Magnetic Resonance in Medicine*. 2011; 10(1):1–10.
73. Yuh WT, Fisher DJ, Shields RK, Ehrhardt JC, Shellock FG. Phantom limb pain induced in amputee by strong magnetic fields. *Journal of magnetic resonance imaging : JMRI*. 1992; 2(2):221–223. [PubMed: 1562774]
74. Abart J, Eberhardt K, Fischer H, Huk W, Richter E, Schmitt F, Storch T, Tomandl B, Zeitler E. Peripheral nerve stimulation by time-varying magnetic fields. *Journal of computer assisted tomography*. 1997; 21(4):532–538. [PubMed: 9216757]
75. Alon L, Deniz CM, Brown R, Sodickson DK, Zhu Y. Method for in situ characterization of radiofrequency heating in parallel transmit MRI. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 2013; 69(5):1457–1465.10.1002/mrm.24374
76. Atalar E. Radiofrequency safety for interventional MRI procedures. *Academic radiology*. 2005; 12(9):1149–1157.10.1016/j.acra.2005.06.007 [PubMed: 16112515]
77. Boss A, Graf H, Berger A, Lauer UA, Wojtczyk H, Claussen CD, Schick F. Tissue warming and regulatory responses induced by radio frequency energy deposition on a whole-body 3-Tesla magnetic resonance imager. *Journal of magnetic resonance imaging : JMRI*. 2007; 26(5):1334–1339.10.1002/jmri.21156 [PubMed: 17969173]
78. Bottomley PA. Turning up the heat on MRI. *Journal of the American College of Radiology : JACR*. 2008; 5(7):853–855.10.1016/j.jacr.2008.04.003 [PubMed: 18585665]
79. Bottomley PA, Edelstein WA. Power deposition in whole-body NMR imaging. *Medical physics*. 1981; 8(4):510–512. [PubMed: 7322070]
80. Budinger TF. Nuclear magnetic resonance (NMR) in vivo studies: known thresholds for health effects. *Journal of computer assisted tomography*. 1981; 5(6):800–811. [PubMed: 7033311]

81. Collins CM. Numerical field calculations considering the human subject for engineering and safety assurance in MRI. *NMR in biomedicine*. 2009; 22(9):919–926.10.1002/nbm.1251 [PubMed: 18384179]
82. de Greef M, Ipek O, Raaijmakers AJ, Crezee J, van den Berg CA. Specific absorption rate intersubject variability in 7T parallel transmit MRI of the head. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 2013; 69(5):1476–1485.10.1002/mrm.24378
83. Drinkwater BL, Horvath SM. Heat tolerance and aging. *Medicine and science in sports*. 1979; 11(1):49–55. [PubMed: 481156]
84. El-Sharkawy AM, Qian D, Bottomley PA, Edelstein WA. A multichannel, real-time MRI RF power monitor for independent SAR determination. *Medical physics*. 2012; 39(5):2334–2341.10.1118/1.3700169 [PubMed: 22559603]
85. Gorny KR, Bernstein MA, Felmlee JP, Ward HA, McGee KP, Lanners DM, Lee KH. Calorimetric calibration of head coil SAR estimates displayed on a clinical MR scanner. *Physics in medicine and biology*. 2008; 53(10):2565–2576.10.1088/0031-9155/53/10/008 [PubMed: 18441413]
86. Gowland PA, De Wilde J. Temperature increase in the fetus due to radio frequency exposure during magnetic resonance scanning. *Physics in medicine and biology*. 2008; 53(21):L15–18.10.1088/0031-9155/53/21/L01 [PubMed: 18843171]
87. Graesslin I, Homann H, Biederer S, Bornert P, Nehrke K, Vernickel P, Mens G, Harvey P, Katscher U. A specific absorption rate prediction concept for parallel transmission MR. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 2012; 68(5):1664–1674.10.1002/mrm.24138
88. Hand JW, Li Y, Hajnal JV. Numerical study of RF exposure and the resulting temperature rise in the foetus during a magnetic resonance procedure. *Physics in medicine and biology*. 2010; 55(4):913–930.10.1088/0031-9155/55/4/001 [PubMed: 20090188]
89. Hand JW, Li Y, Thomas EL, Rutherford MA, Hajnal JV. Prediction of specific absorption rate in mother and fetus associated with MRI examinations during pregnancy. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 2006; 55(4):883–893.10.1002/mrm.20824
90. International Commission on Non-Ionizing Radiation P. Medical magnetic resonance (MR) procedures: protection of patients. *Health physics*. 2004; 87(2):197–216. [PubMed: 15257220]
91. Israel M, Zaryabova V, Ivanova M. Electromagnetic field occupational exposure: non-thermal vs. thermal effects. *Electromagnetic biology and medicine*. 2013; 32(2):145–154.10.3109/15368378.2013.776349 [PubMed: 23675617]
92. Jauchem JR. Effects of drugs on thermal responses to microwaves. *General pharmacology*. 1985; 16(4):307–310. [PubMed: 2864303]
93. Kangarlu A, Shellock FG, Chakeres DW. 8.0-Tesla human MR system: temperature changes associated with radiofrequency-induced heating of a head phantom. *Journal of magnetic resonance imaging : JMRI*. 2003; 17(2):220–226.10.1002/jmri.10236 [PubMed: 12541230]
94. Kikuchi S, Saito K, Takahashi M, Ito K. Temperature elevation in the fetus from electromagnetic exposure during magnetic resonance imaging. *Physics in medicine and biology*. 2010; 55(8):2411–2426.10.1088/0031-9155/55/8/018 [PubMed: 20360633]
95. Murbach M, Neufeld E, Kainz W, Pruessmann KP, Kuster N. Whole-body and local RF absorption in human models as a function of anatomy and position within 1.5T MR body coil. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 2013; 10.1002/mrm.24690
96. Neufeld E, Gosselin MC, Murbach M, Christ A, Cabot E, Kuster N. Analysis of the local worst-case SAR exposure caused by an MRI multi-transmit body coil in anatomical models of the human body. *Physics in medicine and biology*. 2011; 56(15):4649–4659.10.1088/0031-9155/56/15/002 [PubMed: 21734334]
97. Oh S, Webb AG, Neuberger T, Park B, Collins CM. Experimental and numerical assessment of MRI-induced temperature change and SAR distributions in phantoms and in vivo. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 2010; 63(1):218–223.10.1002/mrm.22174

98. Rowell LB. Cardiovascular aspects of human thermoregulation. *Circulation research*. 1983; 52(4): 367–379. [PubMed: 6339107]
99. Schaefer DJ. Safety aspects of radiofrequency power deposition in magnetic resonance. *Magnetic resonance imaging clinics of North America*. 1998; 6(4):775–789. [PubMed: 9799855]
100. Shellock FG. Radiofrequency energy-induced heating during MR procedures: a review. *Journal of magnetic resonance imaging : JMRI*. 2000; 12(1):30–36. [PubMed: 10931562]
101. Shellock FG, Crues JV. Temperature, heart rate, and blood pressure changes associated with clinical MR imaging at 1.5 T. *Radiology*. 1987; 163(1):259–262.10.1148/radiology.163.1.3823445 [PubMed: 3823445]
102. Shellock FG, Crues JV. Corneal temperature changes induced by high-field-strength MR imaging with a head coil. *Radiology*. 1988; 167(3):809–811.10.1148/radiology.167.3.3363146 [PubMed: 3363146]
103. Shellock FG, Crues JV. Temperature changes caused by MR imaging of the brain with a head coil. *AJNR American journal of neuroradiology*. 1988; 9(2):287–291. [PubMed: 3128077]
104. Shellock FG, Rothman B, Sarti D. Heating of the scrotum by high-field-strength MR imaging. *AJR American journal of roentgenology*. 1990; 154(6):1229–1232.10.2214/ajr.154.6.2110733 [PubMed: 2110733]
105. Shellock FG, Schaefer DJ, Crues JV. Alterations in body and skin temperatures caused by magnetic resonance imaging: is the recommended exposure for radiofrequency radiation too conservative? *The British journal of radiology*. 1989; 62(742):904–909. [PubMed: 2819358]
106. Shellock FG, Schaefer DJ, Kanal E. Physiologic responses to an MR imaging procedure performed at a specific absorption rate of 6.0 W/kg. *Radiology*. 1994; 192(3):865–868.10.1148/radiology.192.3.8058962 [PubMed: 8058962]
107. Shellock FG, Schatz CJ. Increased corneal temperature caused by MR imaging of the eye with a dedicated local coil. *Radiology*. 1992; 185(3):697–699.10.1148/radiology.185.3.1438747 [PubMed: 1438747]
108. Shrivastava D, Hanson T, Kulesa J, DelaBarre L, Iaizzo P, Vaughan JT. Radio frequency heating at 9.4T (400.2 MHz): in vivo thermoregulatory temperature response in swine. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/ Society of Magnetic Resonance in Medicine*. 2009; 62(4):888–895.10.1002/mrm.22072
109. Shrivastava D, Hanson T, Kulesa J, Tian J, Adriany G, Vaughan JT. Radiofrequency heating in porcine models with a “large” 32 cm internal diameter, 7 T (296 MHz) head coil. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/ Society of Magnetic Resonance in Medicine*. 2011; 66(1):255–263.10.1002/mrm.22790
110. Shrivastava D, Hanson T, Schlentz R, Gallagher W, Snyder C, Delabarre L, Prakash S, Iaizzo P, Vaughan JT. Radiofrequency heating at 9.4T: in vivo temperature measurement results in swine. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 2008; 59(1):73–78.10.1002/mrm.21425
111. Shuman WP, Haynor DR, Guy AW, Wesbey GE, Schaefer DJ, Moss AA. Superficial- and deep-tissue temperature increases in anesthetized dogs during exposure to high specific absorption rates in a 1.5-T MR imager. *Radiology*. 1988; 167(2):551–554.10.1148/radiology.167.2.3357971 [PubMed: 3357971]
112. Stuchly MA, Abrishamkar H, Strydom ML. Numerical evaluation of radio frequency power deposition in human models during MRI. *Conference proceedings : Annual International Conference of the IEEE Engineering in Medicine and Biology Society IEEE Engineering in Medicine and Biology Society Annual Conference*. 2006; 1:272–275.10.1109/IEMBS.2006.259880
113. van Lier AL, Kotte AN, Raaymakers BW, Lagendijk JJ, van den Berg CA. Radiofrequency heating induced by 7T head MRI: thermal assessment using discrete vasculature or Pennes’ bioheat equation. *Journal of magnetic resonance imaging : JMRI*. 2012; 35(4):795–803.10.1002/jmri.22878 [PubMed: 22068916]
114. van Rhooen GC, Samaras T, Yarmolenko PS, Dewhirst MW, Neufeld E, Kuster N. CEM43 degrees C thermal dose thresholds: a potential guide for magnetic resonance radiofrequency exposure

- levels? *European radiology*. 2013; 23(8):2215–2227.10.1007/s00330-013-2825-y [PubMed: 23553588]
115. Voigt T, Homann H, Katscher U, Doessel O. Patient-individual local SAR determination: in vivo measurements and numerical validation. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 2012; 68(4):1117–1126.10.1002/mrm.23322
 116. Wang Z, Lin JC, Vaughan JT, Collins CM. Consideration of physiological response in numerical models of temperature during MRI of the human head. *Journal of magnetic resonance imaging : JMRI*. 2008; 28(5):1303–1308.10.1002/jmri.21556 [PubMed: 18972342]
 117. Wolf S, Diehl D, Gebhardt M, Mallow J, Speck O. SAR simulations for high-field MRI: how much detail, effort, and accuracy is needed? *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 2013; 69(4):1157–1168.10.1002/mrm.24329
 118. Budinger TF, Fischer H, Hentschel D, Reinfelder HE, Schmitt F. Physiological effects of fast oscillating magnetic field gradients. *Journal of computer assisted tomography*. 1991; 15(6):909–914. [PubMed: 1939767]
 119. Andreuccetti D, Contessa GM, Falsaperla R, Lodato R, Pinto R, Zoppetti N, Rossi P. Weighted-peak assessment of occupational exposure due to MRI gradient fields and movements in a nonhomogeneous static magnetic field. *Medical physics*. 2013; 40(1):011910.10.1118/1.4771933 [PubMed: 23298101]
 120. Bencsik M, Bowtell R, Bowley R. Electric fields induced in the human body by time-varying magnetic field gradients in MRI: numerical calculations and correlation analysis. *Physics in medicine and biology*. 2007; 52(9):2337–2353.10.1088/0031-9155/52/9/001 [PubMed: 17440238]
 121. Bourland JD, Nyenhuis JA, Schaefer DJ. Physiologic effects of intense MR imaging gradient fields. *Neuroimaging clinics of North America*. 1999; 9(2):363–377. [PubMed: 10318720]
 122. Cohen MS, Weisskoff RM, Rzedzian RR, Kantor HL. Sensory stimulation by time-varying magnetic fields. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 1990; 14(2):409–414.
 123. de Vocht F, Liket L, De Vocht A, Mistry T, Glover P, Gowland P, Kromhout H. Exposure to alternating electromagnetic fields and effects on the visual and visuomotor systems. *The British journal of radiology*. 2007; 80(958):822–828.10.1259/bjr/22263979 [PubMed: 17875601]
 124. Doherty JU, Whitman GJ, Robinson MD, Harken AH, Simson MB, Spear JF, Josephson ME. Changes in cardiac excitability and vulnerability in NMR fields. *Investigative radiology*. 1985; 20(2):129–135. [PubMed: 3988463]
 125. Ehrhardt JC, Lin CS, Magnotta VA, Fisher DJ, Yuh WT. Peripheral nerve stimulation in a whole-body echo-planar imaging system. *Journal of magnetic resonance imaging : JMRI*. 1997; 7(2): 405–409. [PubMed: 9090599]
 126. Feldman RE, Hardy CJ, Aksel B, Schenck J, Chronik BA. Experimental determination of human peripheral nerve stimulation thresholds in a 3-axis planar gradient system. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 2009; 62(3):763–770.10.1002/mrm.22050
 127. Glover PM, Eldeghaidy S, Mistry TR, Gowland PA. Measurement of visual evoked potential during and after periods of pulsed magnetic field exposure. *Journal of magnetic resonance imaging : JMRI*. 2007; 26(5):1353–1356.10.1002/jmri.21155 [PubMed: 17969178]
 128. Ham CL, Engels JM, van de Wiel GT, Machielsen A. Peripheral nerve stimulation during MRI: effects of high gradient amplitudes and switching rates. *Journal of magnetic resonance imaging : JMRI*. 1997; 7(5):933–937. [PubMed: 9307922]
 129. Kangarlu A, Tang L, Ibrahim TS. Electric field measurements and computational modeling at ultrahigh-field MRI. *Magnetic resonance imaging*. 2007; 25(8):1222–1226.10.1016/j.mri.2007.01.115 [PubMed: 17368794]
 130. King KF, Schaefer DJ. Spiral scan peripheral nerve stimulation. *Journal of magnetic resonance imaging : JMRI*. 2000; 12(1):164–170. [PubMed: 10931576]

131. Li Y, Hand JW, Wills T, Hajnal JV. Numerically-simulated induced electric field and current density within a human model located close to a z-gradient coil. *Journal of magnetic resonance imaging : JMRI*. 2007; 26(5):1286–1295.10.1002/jmri.21137 [PubMed: 17969144]
132. Mansfield P, Harvey PR. Limits to neural stimulation in echo-planar imaging. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 1993; 29(6):746–758.
133. Schaefer DJ, Bourland JD, Nyenhuis JA. Review of patient safety in time-varying gradient fields. *Journal of magnetic resonance imaging : JMRI*. 2000; 12(1):20–29. [PubMed: 10931561]
134. Vogt FM, Ladd ME, Hunold P, Mateiescu S, Hebrank FX, Zhang A, Debatin JF, Gohde SC. Increased time rate of change of gradient fields: effect on peripheral nerve stimulation at clinical MR imaging. *Radiology*. 2004; 233(2):548–554.10.1148/radiol.2332030428 [PubMed: 15358851]
135. Weinberg IN, Stepanov PY, Fricke ST, Probst R, Urdaneta M, Warnow D, Sanders H, Glidden SC, McMillan A, Starewicz PM, Reilly JP. Increasing the oscillation frequency of strong magnetic fields above 101 kHz significantly raises peripheral nerve excitation thresholds. *Medical physics*. 2012; 39(5):2578–2583.10.1118/1.3702775 [PubMed: 22559628]
136. American College of Radiology (ACR). [Accessed November 8, 2015] Manual on Contrast Media, Version 9. 2013. <http://www.acr.org/quality-safety/resources/~media/37D84428BF1D4E1B9A3A2918DA9E27A3.pdf>
137. U.S. Department of Health and Human Services, Food and Drug Administration, Center for Devices and Radiological Health, Guidance for Industry and FDA Staff. Criteria for Significant Risk Investigations of Magnetic Resonance Diagnostic Devices. Jul 14. 2003 [Accessed November 8, 2015]
138. U.S. Department of Health and Human Services, FDA, CDRH, Guidance for Industry and FDA Staff. Criteria for Significant Risk Investigations of Magnetic Resonance Diagnostic Devices. Jul 14. 2003 [Accessed November 8, 2015]
139. [Accessed November 8, 2015] ACR MR Safety. Available at <http://www.acr.org/quality-safety/radiology-safety/mr-safety>
140. Sammet S, Koch RM, Aguila F, Knopp MV. Residual magnetism in an MRI suite after field-rampdown: what are the issues and experiences? *Journal of magnetic resonance imaging : JMRI*. 2010; 31(5):1272–1276.10.1002/jmri.22141 [PubMed: 20432367]
141. Collins CM, Wang Z. Calculation of radiofrequency electromagnetic fields and their effects in MRI of human subjects. *Magnetic resonance in medicine : official journal of the Society of Magnetic Resonance in Medicine/Society of Magnetic Resonance in Medicine*. 2011; 65(5):1470–1482.10.1002/mrm.22845
142. Machata AM, Willschke H, Kabon B, Prayer D, Marhofer P. Effect of brain magnetic resonance imaging on body core temperature in sedated infants and children. *British journal of anaesthesia*. 2009; 102(3):385–389.10.1093/bja/aen388 [PubMed: 19174372]
143. Woods TO. Standards for medical devices in MRI: present and future. *Journal of magnetic resonance imaging : JMRI*. 2007; 26(5):1186–1189.10.1002/jmri.21140 [PubMed: 17969160]
144. ASTM F2119-01 Standard Test Method for Evaluation of MR Image Artifacts From Passive Implants.
145. Kanal E, Shellock FG, Talagala L. Safety considerations in MR imaging. *Radiology*. 1990; 176(3):593–606.10.1148/radiology.176.3.2202008 [PubMed: 2202008]
146. IEC 601-2-33 - Medical Electrical Equipment - Part 2: Particular requirements for the safety of magnetic resonance equipment for medical diagnosis.
147. International Electrotechnical Commission (IEC). Medical Electrical Equipment, Particular requirements for the safety of magnetic resonance equipment for medical diagnosis, International Standard IEC 6060, 1-2-33, 2002.
148. [Accessed November 8, 2015] Guidelines to Prevent Excessive Heating and Burns Associated with Magnetic Resonance Procedures developed by the Institute for Magnetic Resonance Safety, Education, and Research (IMRSER). <http://www.mrisafety.com/SafetyInfov.asp?SafetyInfoID=166>

149. Kangarlu A, Ibrahim TS, Shellock FG. Effects of coil dimensions and field polarization on RF heating inside a head phantom. *Magnetic resonance imaging*. 2005; 23(1):53–60.10.1016/j.mri.2004.11.007 [PubMed: 15733788]
150. Hassepass F, Stabenau V, Arndt S, Beck R, Bulla S, Grauvogel T, Aschendorff A. Magnet dislocation: an increasing and serious complication following MRI in patients with cochlear implants. *RoFo : Fortschritte auf dem Gebiete der Rontgenstrahlen und der Nuklearmedizin*. 2014; 186(7):680–685.10.1055/s-0033-1356238 [PubMed: 24497090]
151. [Accessed November 8, 2015] ACR_SPR Practice parameter for the safe and optimal performance of fetal magnetic resonance imaging (MRI), Amended 2014 (Resolution 39)*. <http://www.acr.org/~media/CB384A65345F402083639E6756CE513F.pdf>
152. Baker PN, Johnson IR, Harvey PR, Gowland PA, Mansfield P. A three-year follow-up of children imaged in utero with echo-planar magnetic resonance. *American journal of obstetrics and gynecology*. 1994; 170(1 Pt 1):32–33. [PubMed: 8296840]
153. Chew S, Ahmadi A, Goh PS, Foong LC. The effects of 1.5T magnetic resonance imaging on early murine in-vitro embryo development. *Journal of magnetic resonance imaging : JMRI*. 2001; 13(3):417–420. [PubMed: 11241816]
154. Clements H, Duncan KR, Fielding K, Gowland PA, Johnson IR, Baker PN. Infants exposed to MRI in utero have a normal paediatric assessment at 9 months of age. *The British journal of radiology*. 2000; 73(866):190–194.10.1259/bjr.73.866.10884733 [PubMed: 10884733]
155. Gilk T, Kanal E. Interrelating sentinel event alert #38 with the ACR guidance document on MR safe practices: 2013. An MRI accreditation safety review tool. *Journal of magnetic resonance imaging : JMRI*. 2013; 37(3):531–543.10.1002/jmri.24027 [PubMed: 23345250]
156. Thomsen, HS. *Medical radiology*. Springer; Berlin: 2006. Contrast media : safety issues and esur guidelines.
157. Behra-Miellet J, Gressier B, Brunet C, Dine T, Luyckx M, Cazin M, Cazin JC. Free gadolinium and gadodiamide, a gadolinium chelate used in magnetic resonance imaging: evaluation of their in vitro effects on human neutrophil viability. *Methods and findings in experimental and clinical pharmacology*. 1996; 18(7):437–442. [PubMed: 8900215]
158. Mendoza FA, Artlett CM, Sandorfi N, Latinis K, Piera-Velazquez S, Jimenez SA. Description of 12 cases of nephrogenic fibrosing dermopathy and review of the literature. *Seminars in arthritis and rheumatism*. 2006; 35(4):238–249.10.1016/j.semarthrit.2005.08.002 [PubMed: 16461069]
159. [Accessed November 8, 2015] The International Center for Nephrogenic Systemic Fibrosis Research (ICNSFR) by Cowper SE. Available at <http://www.icnfr.org>
160. Food and Drug Administration (FDA). [Accessed November 8, 2015] Information on Gadolinium-Based Contrast Agents. 2014. <http://www.fda.gov/Drugs/DrugSafety/PostmarketDrugSafetyInformationforPatientsandProviders/ucm142882.htm>
161. Brown JJ, Hynes MR, Wible JH Jr. Measurement of serum calcium concentration after administration of four gadolinium-based contrast agents to human volunteers. *AJR American journal of roentgenology*. 2007; 189(6):1539–1544.10.2214/AJR.07.2464 [PubMed: 18029898]
162. Dillman JR, Ellis JH, Cohan RH, Caoili EM, Hussain HK, Campbell AD, Strouse PJ. Safety of gadolinium-based contrast material in sickle cell disease. *Journal of magnetic resonance imaging : JMRI*. 2011; 34(4):917–920.10.1002/jmri.22666 [PubMed: 21769966]
163. Evenepoel P, Zeegers M, Segaeert S, Claes K, Kuypers D, Maes B, Flamen P, Fransis S, Vanreenterghem Y. Nephrogenic fibrosing dermopathy: a novel, disabling disorder in patients with renal failure. *Nephrology, dialysis, transplantation : official publication of the European Dialysis and Transplant Association - European Renal Association*. 2004; 19(2):469–473.
164. Grobner T. Gadolinium--a specific trigger for the development of nephrogenic fibrosing dermopathy and nephrogenic systemic fibrosis? *Nephrology, dialysis, transplantation : official publication of the European Dialysis and Transplant Association - European Renal Association*. 2006; 21(4):1104–1108.10.1093/ndt/gfk062
165. Kanal E. An overview of electromagnetic safety considerations associated with magnetic resonance imaging. *Annals of the New York Academy of Sciences*. 1992; 649:204–224. [PubMed: 1580494]

166. Marckmann P, Skov L, Rossen K, Dupont A, Damholt MB, Heaf JG, Thomsen HS. Nephrogenic systemic fibrosis: suspected causative role of gadodiamide used for contrast-enhanced magnetic resonance imaging. *Journal of the American Society of Nephrology : JASN*. 2006; 17(9):2359–2362.10.1681/ASN.2006060601 [PubMed: 16885403]
167. Marckmann P, Skov L, Rossen K, Heaf JG, Thomsen HS. Case-control study of gadodiamide-related nephrogenic systemic fibrosis. *Nephrology, dialysis, transplantation : official publication of the European Dialysis and Transplant Association - European Renal Association*. 2007; 22(11):3174–3178.10.1093/ndt/gfm261
168. Umans H, Haramati N, Flusser G. The diagnostic role of gadolinium enhanced MRI in distinguishing between acute medullary bone infarct and osteomyelitis. *Magnetic resonance imaging*. 2000; 18(3):255–262. [PubMed: 10745133]
169. Westwood MA, Shah F, Anderson LJ, Strange JW, Tanner MA, Maceira AM, Howard J, Porter JB, Walker JM, Wonke B, Pennell DJ. Myocardial tissue characterization and the role of chronic anemia in sickle cell cardiomyopathy. *Journal of magnetic resonance imaging : JMRI*. 2007; 26(3):564–568.10.1002/jmri.21018 [PubMed: 17729345]
170. Zimmerman RA. MRI/MRA evaluation of sickle cell disease of the brain. *Pediatric radiology*. 2005; 35(3):249–257.10.1007/s00247-005-1420-z [PubMed: 15703900]
171. Thorpe S, Salkovskis PM, Dittner A. Claustrophobia in MRI: the role of cognitions. *Magnetic resonance imaging*. 2008; 26(8):1081–1088.10.1016/j.mri.2008.01.022 [PubMed: 18524527]
172. Conquering claustrophobia during your MRI. *The Johns Hopkins medical letter health after 50*. 2002; 14(9):3.
173. Dantendorfer K, Wimberger D, Katschnig H, Imhoff H. Claustrophobia in MRI scanners. *Lancet*. 1991; 338(8769):761–762. [PubMed: 1679897]
174. Murphy KJ, Brunberg JA. Adult claustrophobia, anxiety and sedation in MRI. *Magnetic resonance imaging*. 1997; 15(1):51–54. [PubMed: 9084025]
175. Phelps LA. MRI and claustrophobia. *American family physician*. 1990; 42(4):930. [PubMed: 2220520]
176. Wahba H. Minimizing claustrophobia in an MRI scanner. *Nursing*. 1995; 25(7):32C–32D.
177. Serafini G, Ongaro L, Mori A, Rossi C, Cavalloro F, Tagliaferri C, Mencherini S, Braschi A. Anesthesia for MRI in the paediatric patient. *Minerva anesthesiologica*. 2005; 71(6):361–366. [PubMed: 15886602]
178. Pilling D, Abernethy L, Wright N, Carty H. Sedation, safety and MRI. *The British journal of radiology*. 2001; 74(885):875–876.10.1259/bjr.74.885.740875b [PubMed: 11560840]