

CORRECTION

Correction: Correlation between Serum Levels of 3,3',5'-Triiodothyronine and Thyroid Hormones Measured by Liquid Chromatography-Tandem Mass Spectrometry and Immunoassay

Hiroyuki Sakai, Hidenori Nagao, Mamoru Sakurai, Takako Okumura, Yoshiyuki Nagai, Junpei Shikuma, Rokuro Ito, Tetsuya Imazu, Takashi Miwa, Masato Odawara

There are numerical errors in the rT3 columns of Table 2 of the published article. Please see the correct [Table 2](#) here.



OPEN ACCESS

Citation: Sakai H, Nagao H, Sakurai M, Okumura T, Nagai Y, Shikuma J, et al. (2016) Correction: Correlation between Serum Levels of 3,3',5'-Triiodothyronine and Thyroid Hormones Measured by Liquid Chromatography-Tandem Mass Spectrometry and Immunoassay. PLoS ONE 11(7): e0159169. doi:10.1371/journal.pone.0159169

Published: July 7, 2016

Copyright: © 2016 Sakai et al. This is an open access article distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Table 2. Precision and Accuracy of intra-day assay.

Concentration (ng/mL)	Accuracy (%)		
	T4	T3	rT3
0.05	104.0 ± 8.0 (7.7)	110.0 ± 4.0 (3.6)	108.0 ± 6.0 (5.6)
0.1	105.0 ± 6.0 (5.7)	104.0 ± 7.0 (6.7)	102.0 ± 11.0 (10.8)
2.5	108.1 ± 3.1 (2.9)	100.3 ± 5.5 (5.5)	101.8 ± 4.5 (4.4)
80	102.5 ± 1.3 (1.3)	96.2 ± 6.6 (6.9)	98.5 ± 6.4 (6.5)

Data are expressed as the mean ± S.D. (n = 5).

Values in parentheses are coefficients of variance (%).

doi:10.1371/journal.pone.0159169.t001

Reference

1. Sakai H, Nagao H, Sakurai M, Okumura T, Nagai Y, Shikuma J, et al. (2015) Correlation between Serum Levels of 3,3',5'-Triiodothyronine and Thyroid Hormones Measured by Liquid Chromatography-Tandem Mass Spectrometry and Immunoassay. PLoS ONE 10(10): e0138864. doi: [10.1371/journal.pone.0138864](https://doi.org/10.1371/journal.pone.0138864) PMID: 26426328