



Published in final edited form as:

Cornea. 2017 May ; 36(5): 631–635. doi:10.1097/ICO.0000000000001157.

Use of Cross-Polarizers to Enhance Images of the Eyelids

Ryan O'Sullivan, BA^{1,*}, Lisa M. Tom, MD^{1,*}, Vatinee Y. Bunya, MD^{1,**}, William C. Nyberg^{1,+},
Mina Massaro-Giordano, MD¹, Ebenezer Daniels, MPH, PhD², Eli Smith, BA², David H.
Brainard, PhD³, James Gee, PhD⁴, Maureen G. Maguire, PhD², and Richard A. Stone, MD¹

¹Department of Ophthalmology, Scheie Eye Institute, University of Pennsylvania, Philadelphia, PA

²Center for Preventive Ophthalmology and Biostatistics, University of Pennsylvania, Philadelphia, PA

³Department of Psychology, University of Pennsylvania, Philadelphia, PA

⁴Department of Radiology, University of Pennsylvania, Philadelphia, PA

Abstract

Purpose—To describe imaging the external eye with crossed polarizers to enhance clinically important features in digital photographs of the eyelids.

Methods—External photographs with and without crossed polarizing filters were taken of patients with blepharitis and controls with no clinical eye pathology.

Results—Photographing eyelid skin through crossed polarizers decreased reflections on the skin surface and improved visualization of eyelid telangiectasias and blood vessels in patients with a broad range of skin pigmentation and ethnicities.

Conclusions—The use of crossed polarizers in imaging the external eye reduces reflections and glare from the eyelid skin and margins, thereby allowing for a more detailed evaluation of underlying structures and analysis of images. These findings suggest that including crossed polarizers in clinical photography has informative applications for assessing eyelid disease.

Keywords

conjunctiva; cornea; eyelids; imaging; cross-polarizers

****Corresponding Author:** Vatinee Y. Bunya, MD, Scheie Eye Institute, University of Pennsylvania, 51 N. 39th Street, Philadelphia, PA 19104, vatinee.bunya@uphs.upenn.edu, 215-662-9791.

*Co-first authors (listed alphabetically)

+Deceased

Conflict of Interest Statement:

The Eyelid Photography Protocol is copyright-protected by the University of Pennsylvania. Use of the protocol in investigational studies sponsored in whole or part by for-profit entities or for commercial purposes by any entities is prohibited without the express written consent of the University of Pennsylvania. The authors retain intellectual rights to the photography protocol for the possibility of future licensing (VYB, DHB, MMG, WCN, MGM, RAS).

Disclosures not relevant to submitted work: Dr. Massaro-Giordano owns stock in Physician Recommended Nutraceuticals. Dr. Stone received an honorarium from Santen, Inc., in 2016 for a seminar and travel expenses; and is co-inventor on 3 patents licensed to EyeIC but does not receive any royalties.

INTRODUCTION

The use of polarized light for diagnostic imaging has been studied in dermatology¹, pathology², and ophthalmology^{3, 4}, where it can reveal the reflectance, polarizing, depolarizing, and/or birefringence properties of tissues under study. In dermatology, polarization methods have been applied to a variety of conditions including pigmented and non-pigmented skin lesions⁵, skin malignancies⁶, photoaging⁷, and acne.^{8, 9} Photography using a polarizing filter parallel to the polarization of the light source (*parallel polarizers*) enhances superficial structures of the skin, while using a filter at cross-polarization to the light source (*crossed polarizers*) improves visualization of subsurface tissue structures.¹ By reducing specular reflections from superficial epidermal tissue and skin surface lipids, imaging with crossed polarizers specifically enhances the assessment of blood vessels and chromophores in skin^{10, 11}, thus improving assessments of pigmented lesions and vascular lesions^{10, 11}, and reducing the impact of epidermal melanin.¹²

In ophthalmology, polarization manipulations have been applied to image various parts of the human eye, including the conjunctiva¹³, cornea^{13, 14}, lens¹⁵, sclera¹³, and retinal nerve fiber layer.¹⁶ Because approximately 90% of the cornea is composed of parallel sheets of collagen fibers, the cornea behaves as a uniaxial birefringent or anisotropic medium.¹⁷ When imaged using crossed polarizers, a characteristic interference pattern develops, known as an “isogyre”, which was first described in 1941 by Cogan.¹⁸ The mechanism for the appearance of the corneal isogyre is due to a combination of the optical properties created by the curvature of the cornea and distribution of collagen fibrils.¹⁹ Briefly, light that is parallel or perpendicular to the corneal surface is not altered, whereas light that is oblique to the surface of the cornea is rotated and retarded, thereby appearing as a dark cross shape.¹⁹ Birefringence of the cornea is influenced by intraocular pressure (IOP) and corneal birefringence was once explored as a non-invasive measure of IOP but found to have limited utility.²⁰

While the lens similarly produces an isogyre interference pattern¹⁷, its birefringence is much weaker than that of the cornea.¹⁵ Scanning laser polarimetry utilizes the birefringence of the retinal nerve fiber layer to estimate its thickness.²¹ Polarization sensitive ocular coherence tomography (PS-OCT) can measure tissue birefringence and has been studied for several applications, including measuring scleral elasticity and retinal nerve fiber layer damage.²²

In the present study, we used crossed polarizers to image the eyelids. We hypothesized that photography through cross-polarizers would reduce specular reflectance and glare and thus improve visualization of sub-epidermal structures and the vasculature of the eyelids.

MATERIALS AND METHODS

This study was approved by the University of Pennsylvania Institutional Review Board and complied with the Health Insurance Portability and Accountability Act. Informed consent was obtained from all subjects prior to participation, and the patients were photographed using the protocol described below (a modification of the Eyelid Photography Protocol previously published by our group).²³ The use of polarized filters with an external camera is

not registered as a device with the United States Food and Drug Administration for the diagnosis of external eye disease, and its use here should be considered investigational.

A Canon EOS Rebel T2i camera and external flash (Canon 600 EX-RT) were attached to a custom-made frame that screwed onto a monopod (Manofrotto 608B monopod with ball joint attachment) for stability. This arrangement enabled the external electronic flash to be aligned at a 45 degree angle downward towards the subject's eye. The external flash was set to ETTL (exposure through the lens) mode, with 1/60s shutter speed. A Canon 100mm f/2.8L macro lens, attached to the camera, was set to manual focus at 0.34 meters and F22 aperture. The camera was connected via a USB cable to a laptop computer with Canon EOS Utility Software that permitted real-time viewing and management of the images.

Photographs were taken in examination rooms with fluorescent overhead lighting of approximately 320 lux. Photos of the eyelids were taken with the patient looking straight ahead, up, or down with gentle traction of the lids to allow full visualization of the eyelid margins. For each photo, the camera was positioned so the lens focused on the region of interest on the eyelid, with the camera re-positioned for each image.

For the polarized photographs, a circular polarizing filter (Heliopan ES 67mm SH PMC high transmission circular polarizer) was attached over the lens of the camera and an orthogonally oriented piece of TECHSPEC linear polarizing laminated film (Edmund Optics, Barrington, NJ, custom-cut to fit a filter holder) was placed over the external flash (Figure 1).

RESULTS

Polarized and non-polarized photos were taken of patients with blepharitis and controls with no clinical eyelid pathology. The novel observations here relate to improved visualization of eyelid blood vessels and telangiectasias. These observations are summarized in Table 1.

Eyelids

Eyelid photographs with crossed polarizers decreased reflections from the skin surface and revealed blood vessels and telangiectasias that were either poorly visualized or not visualized in the non-polarized photographs (Figure 2). Crossed polarizers improved the visualization of eyelid telangiectasias and vasculature for patients with a broad range of skin pigmentation and ethnicities.

Cornea and conjunctiva

Use of crossed polarizers revealed the typical isogyre in the cornea (Figures 2B, 2D). As previously described^{13, 14}, photographs through crossed polarizers decreased reflections from the corneal and conjunctival surfaces (Figure 2).

DISCUSSION

The use of crossed polarizers in imaging the external eye reduces reflections and glare from the eyelid skin and margins, as well as from conjunctival and corneal surfaces.¹³ Polarized digital images using commercially available photography equipment can be easily acquired

in many clinical settings. The isogyre in cornea can confirm the correct orthogonal orientation of the two polarizing filters¹³, when the cornea is fully visible in the image.

For the eyelids, photography through crossed polarizers improves visualization of telangiectasias beneath the skin and lid margins that were not readily visible in non-polarized images. Importantly, this method improves vascular visualization in subjects with both lightly and heavily pigmented skin, suggesting potentially useful applications.

For example, polarized photography could be useful for diagnosis and research in eyelid diseases such as blepharitis and meibomian gland dysfunction (MGD). While the enhancement of vessels in rosacea of the skin¹ and other disorders⁶ has been described in the dermatology literature, we are not aware of prior reports describing the utility of cross-polarizing filters for imaging the eyelids. Blepharitis and MGD are under-diagnosed in ophthalmology, particularly in patients with heavily pigmented skin²⁴; and patients often present with non-specific signs or symptoms.²⁵ Earlier clinical diagnosis of blepharitis and MGD, along with improved research end-points, could result from the application of cross-polarization imaging.

While reduction of corneal and conjunctival surface reflections from photography with crossed polarizers has been described previously¹³, this imaging technique has not been extensively applied to investigating conjunctival or corneal disease. Besides potential use in diagnosing eyelid pathology, removing reflections could improve the accuracy of objective assessment algorithms now being developed to quantify redness or other biomarkers from digital conjunctival and corneal photographs.

Acknowledgments

We would like to thank Dominique Caggiano and Trenton Rivera for their assistance in obtaining the photograph used for Figure 1.

Funding:

This work was supported by National Eye Institute grants K12 EY015398 (VYB) and core grant P30 EY001583, the Paul and Evanina Bell Mackall Foundation Trust (RAS), and an unrestricted grant from Research to Prevent Blindness (VYB, MMG, RAS).

References

1. Anderson RR. Polarized light examination and photography of the skin. *Arch Dermatol.* 1991; 127:1000–5. [PubMed: 2064396]
2. Yaroslavsky AN, Neel V, Anderson RR. Demarcation of nonmelanoma skin cancer margins in thick excisions using multispectral polarized light imaging. *J Invest Dermatol.* 2003; 121:259–66. [PubMed: 12880416]
3. Gotzinger E, Pircher M, Dejaco-Ruhswurm I, et al. Imaging of birefringent properties of keratoconus corneas by polarization-sensitive optical coherence tomography. *Invest Ophthalmol Vis Sci.* 2007; 48:3551–8. [PubMed: 17652723]
4. Weinreb RN, Bowd C, Zangwill LM. Glaucoma detection using scanning laser polarimetry with variable corneal polarization compensation. *Arch Ophthalmol.* 2003; 121:218–24. [PubMed: 12583788]

5. Benvenuto-Andrade C, Dusza SW, Agero AL, et al. Differences between polarized light dermoscopy and immersion contact dermoscopy for the evaluation of skin lesions. *Arch Dermatol.* 2007; 143:329–38. [PubMed: 17372097]
6. Jacques SL, Ramella-Roman JC, Lee K. Imaging skin pathology with polarized light. *J Biomed Opt.* 2002; 7:329–40. [PubMed: 12175282]
7. Muccini JA, Kollias N, Phillips SB, et al. Polarized light photography in the evaluation of photoaging. *J Am Acad Dermatol.* 1995; 33:765–9. [PubMed: 7593775]
8. Rizova E, Kligman A. New photographic techniques for clinical evaluation of acne. *J Eur Acad Dermatol Venereol.* 2001; 15(Suppl 3):13–8. [PubMed: 11843228]
9. Phillips SB, Kollias N, Gillies R, et al. Polarized light photography enhances visualization of inflammatory lesions of acne vulgaris. *J Am Acad Dermatol.* 1997; 37:948–52. [PubMed: 9418762]
10. Jung B, Choi B, Durkin AJ, et al. Characterization of port wine stain skin erythema and melanin content using cross-polarized diffuse reflectance imaging. *Lasers Surg Med.* 2004; 34:174–81. [PubMed: 15004831]
11. Kang H, Jung B, Nelson JS. Polarization color imaging system for on-line quantitative evaluation of facial skin lesions. *Dermatol Surg.* 2007; 33:1350–6. [PubMed: 17958588]
12. Jung B, Kim CS, Choi B, et al. Use of erythema index imaging for systematic analysis of port wine stain skin response to laser therapy. *Lasers Surg Med.* 2005; 37:186–91. [PubMed: 16175634]
13. Fariza E, O'Day T, Jalkh AE, Medina A. Use of cross-polarized light in anterior segment photography. *Arch Ophthalmol.* 1989; 107:608–10. [PubMed: 2705935]
14. Stanworth A, Naylor EJ. The polarization optics of the isolated cornea. *Br J Ophthalmol.* 1950; 34:201–11. [PubMed: 15411483]
15. Bueno JM, Campbell MC. Polarization properties of the in vitro old human crystalline lens. *Ophthalmic Physiol Opt.* 2003; 23:109–18. [PubMed: 12641698]
16. Weinreb RN, Shakiba S, Zangwill L. Scanning laser polarimetry to measure the nerve fiber layer of normal and glaucomatous eyes. *Am J Ophthalmol.* 1995; 119:627–36. [PubMed: 7733188]
17. Pierscionek BK. Explanation of isogyre formation by the eye lens. *Ophthalmic Physiol Opt.* 1993; 13:91–4. [PubMed: 8510954]
18. Cogan D. Some ocular phenomena produced with polarized light. *Arch Ophthalmol.* 1941; 25:391–400.
19. Cope WT, Wolbarsht ML, Yamanashi BS. The corneal polarization cross. *J Opt Soc Am.* 1978; 68:1139–41. [PubMed: 712454]
20. Stanworth A. Polarized light studies of the cornea II: the effect of intra-ocular pressure. *J Exp Biol.* 1953; 30:164–9.
21. Greenfield DS, Knighton RW, Huang XR. Effect of corneal polarization axis on assessment of retinal nerve fiber layer thickness by scanning laser polarimetry. *Am J Ophthalmol.* 2000; 129:715–22. [PubMed: 10926978]
22. Pircher M, Hitzenberger CK, Schmidt-Erfurth U. Polarization sensitive optical coherence tomography in the human eye. *Prog Retin Eye Res.* 2011; 30:431–51. [PubMed: 21729763]
23. Bunya VY, Brainard DH, Daniel E, et al. Assessment of signs of anterior blepharitis using standardized color photographs. *Cornea.* 2013; 32:1475–82. [PubMed: 24055901]
24. Browning DJ, Rosenwasser G, Lugo M. Ocular rosacea in blacks. *Am J Ophthalmol.* 1986; 101:441–4. [PubMed: 2938475]
25. Nichols KK, Foulks GN, Bron AJ, et al. The international workshop on meibomian gland dysfunction: executive summary. *Invest Ophthalmol Vis Sci.* 2011; 52:1922–9. [PubMed: 21450913]




Figure 1. Camera set-up. The wooden frame was custom-made to hold the flash above the body of the camera, so that the flash could be angled down at an appropriate 45 degree angle. A linear polarizer was inserted over the flash and a circular polarizer over the camera lens for polarized images, and were removed for non-polarized images.




Figure 2. Non-polarized and cross-polarized eyelid photographs. Conventional, non-polarized images of the eyelids of patients with blepharitis (panels A, C, E, and G) and the corresponding images of the same eyes through crossed-polarizers (panels B, D, F, and H). The method is illustrated in patients with blepharitis of different races (A, B: Caucasian; C, D: South Asian; E, F: lightly pigmented African-American; G, H: more heavily pigmented African-American). The control subject (I and J) is Asian. Black arrows indicate a representative blood vessel in each patient that is made more visible with cross-polarizers. One patient

(E,F) has a loose eyelash visible on the upper eyelid (not highlighted). White arrows in B, D indicate the typical isogyre when the cornea is visualized through crossed-polarizers. No adjustments were made to any of the images other than the insertion of arrows and labeling.

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript

Table 1

Summary of effects of crossed polarized imaging on eyelids and anterior segment

	Effect of crossed polarizers on image
Eyelids	Enhanced visualization of blood vessels and telangiectasias of eyelid. Decreased reflections from skin surface.
Conjunctiva	Decreased reflections from conjunctival surface.
Cornea	Appearance of corneal isogyre. Decreased reflections from corneal surface.

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript