

ORIGINAL ARTICLE

Open Access



# Controlling *Streptococcus mutans* and *Staphylococcus aureus* biofilms with direct current and chlorhexidine

Hao Wang<sup>1,2</sup> and Dacheng Ren<sup>1,2,3,4\*</sup> 

## Abstract

Microbial biofilms formed on biomaterials are major causes of chronic infections. Among them, Gram-positive bacteria *Streptococcus mutans* and *Staphylococcus aureus* are important pathogens causing infections associated with dental caries (tooth-decay) and other medical implants. Unfortunately, current antimicrobial approaches are ineffective in disrupting established biofilms and new methods are needed to improve the efficacy. In this study, we report that the biofilm cells of *S. mutans* and *S. aureus* can be effectively killed by low-level direct current (DC) and through synergy in concurrent treatment with DC and chlorhexidine (CHX) at low concentrations. For example, after treatment with 28  $\mu\text{A}/\text{cm}^2$  DC and 50  $\mu\text{g}/\text{mL}$  CHX for 1 h, the viability of biofilm cells was reduced by approximately 4 and 5 logs for *S. mutans* and *S. aureus*, respectively. These results are useful for developing more effective approaches to control pathogenic biofilms.

**Keywords:** Biofilm, Electrochemical control, Chlorhexidine, Synergistic effects

## Introduction

Biofilms are formed by microbial cells embedded in a matrix comprised of extracellular polymeric substance (EPS) containing polysaccharide, proteins, and DNA. The presence of this extracellular matrix provides protection to microbial pathogens from antimicrobials to host immune cells/factors (Liu et al. 2016; Hall and Mah 2017). Biofilms can form on both biotic and abiotic surfaces and are common causes of chronic infections including dental plaques (Smith et al. 2011; Song et al. 2015). The protection of EPS plus the dormancy of biofilm cells render these multicellular structures extremely difficult to eradicate (Kouidhi et al. 2015; Smith et al. 2011; Song et al. 2015).

*Streptococcus mutans* is a Gram-positive bacterium commonly found in human dental biofilms. It is a dominant species with higher biomass in dental biofilms than other *Streptococcus* species, including *S. sanguinis*, *S.*

*mitis*, and *S. salivarius*, due to its acid tolerance and thus the capability to live in low pH environment of oral cavities (Bender et al. 1986; Harper and Loesche 1984; Kreth et al. 2005). *S. mutans* expresses multiple exoenzymes (glucosyltransferases) that make it the primary EPS producer in oral cavity (Falsetta et al. 2014), while it is also highly acidogenic and aciduric. *S. mutans* can rapidly colonize tooth surface and establish cariogenic biofilms with extracellular polysaccharides (EPS). This acidifies the local microenvironment and promotes the growth of an acidogenic microbiota, facilitating the development of dental caries (Falsetta et al. 2012, 2014).

*Staphylococcus aureus* is also an abundant Gram-positive bacterium, which usually harbors in the nasal passages and ears of patients (Smith et al. 2011). Previous studies have shown that *S. aureus* is not only a significant cause of many localized and systemic infections such as osteomyelitis (Lew and Waldvogel 2004), chronic wound infection (Hansson et al. 1995), and chronic rhinosinusitis (Stephenson et al. 2010), but also has a strong connection to dental implant infections (Salvi et al. 2008; Harris et al. 2004). The established biofilms of *S. aureus*,

\*Correspondence: dren@syr.edu

<sup>1</sup> Department of Biomedical and Chemical Engineering, Syracuse University, Syracuse, NY 13244, USA

Full list of author information is available at the end of the article

especially the methicillin-resistant *S. aureus* (MRSA), are highly tolerant to common antimicrobial treatments (Jones et al. 2001; O'Donnell et al. 2015; Lewis et al. 2015).

Few approaches are currently available for controlling cariogenic biofilms (Liu et al. 2016). Chlorhexidine (CHX) is considered the “gold standard” for oral antimicrobial therapy (Jones 1997). However, use of high dose CHX has adverse side effects such as tooth staining and calculus formation. Also, CHX is not recommended for long term daily therapeutic use (Flotra et al. 1971). In 1994, Costerton et al. (1994) reported bacterial killing by synergistic effects between low-level electric currents and antibiotics, a phenomenon named “bioelectric effects”. Since 1990s, direct currents (DCs) ranging from  $\mu\text{A}$  to mA have been reported for their bactericidal effects after a relatively long period (from several hours to days) of treatment time (Costerton et al. 1994; del Pozo et al. 2009; Schmidt-Malan et al. 2015; Spadaro et al. 1974) either by DC alone or with antibiotics together (Wattanakaroon et al. 2000; Niepa et al. 2012, 2016a). Recent studies reported that mA level DC could enhance the killing effect of 0.2% (200  $\mu\text{g}/\text{mL}$ ) chlorhexidine on biofilms of Gram-negative *Porphyromonas gingivalis*, although there was no bactericidal effect by DC alone (Lasserre et al. 2015). To explore the potential of lower levels of DC and CHX in killing dental biofilms of Gram-positive bacteria, we conducted this study using *S. mutans* and *S. aureus* as model species. We demonstrate that stainless steel electrode derived DC and CHX have strong synergy in killing *S. mutans* and *S. aureus* biofilms; and the levels of DC and CHX appear to be lower than other reported systems.

## Materials and methods

### Bacteria strains and growth media

*Staphylococcus mutans* Clarke strain (ATCC 25175) was cultured in brain heart infusion (BHI) broth (BD Biosciences, San Jose, CA, U.S.) (Murchison et al. 1982). The *S. aureus* ALC2085 (strain RN6390 containing pALC2084) was obtained from Dr. Karin Sauer at Binghamton University (Sauer et al. 2009) and cultured in Lysogeny broth (LB) (Sambrook and Russell 2001) containing 10 g/L tryptone, 5 g/L yeast extract, and 10 g/L NaCl, supplemented with 10  $\mu\text{g}/\text{mL}$  chloramphenicol (Sigma-Aldrich, St. Louis, MO, U.S.). Both strains were routinely cultured overnight at 37 °C with shaking at 200 rpm.

### Biofilm formation

Biofilms were formed on acrylic coupons (3.5 cm  $\times$  0.5 cm  $\times$  0.1 cm; McMaster-Carr, Aurora, OH, U.S.). Briefly, 25  $\mu\text{L}$  of an overnight culture of *S. mutans*

was used to inoculate a petri dish containing 25 mL of BHI medium and acrylic coupons. The culture was incubated at 37 °C for 48 h without shaking. Then the coupons with biofilms were removed from petri dish and washed three times with 0.85% NaCl solution to remove all planktonic cells and only retained the firmly attached cells for DC and CHX treatments. The *S. aureus* biofilm samples were prepared in the same way except that the medium was LB plus 10  $\mu\text{g}/\text{mL}$  chloramphenicol and the incubation time was reduced to 24 h due to a higher growth rate of *S. aureus*.

### Electrochemical treatment

The experimental system for DC treatment is the same as we described previously (Niepa et al. 2012, 2016b). Briefly, an electrochemical cell was constructed with two electrodes on the opposite sides of a plastic cuvette (Thermo Fisher Scientific, Pittsburg, PA, U.S.). DC was generated using a potentiostat (Potentiostat WaveNow, Pine Research Instrumentation, Raleigh, NC, U.S.) in the three electrode system with a silver wire (0.015” diameter, A-M Systems, Sequim, WA, U.S.) placed in bleach for 30 min to create an Ag/AgCl reference electrode. The DC level and voltage across the electric field were monitored and recorded using the AfterMath software (Potentiostat WaveNow, Pine Research Instrumentation, Raleigh, NC, U.S.) in the galvanostatic mode during the treatment.

### DC treatment of biofilms

Each DC treatment was carried out in 3 mL 0.85% NaCl solution. First, a sterile SS304 electrode (3.5 cm  $\times$  0.95 cm  $\times$  0.05 cm) was inserted into a cuvette, followed by an acrylic coupon with *S. mutans* or *S. aureus* biofilm attached. Another sterile SS304 electrode was then inserted on the opposite side. The biofilm was treated galvanostatically with direct current (DC) for 1 h in the absence or presence of CHX (MP Biomedicals, Solon, OH, U.S.). Samples treated with DC or CHX alone and untreated samples were used as controls. After treatment, each acrylic coupon was transferred to a 10 mL tube containing 5 mL 0.85% NaCl solution. The biofilm cells were removed from the surface by gentle sonication for 1 min. The number of viable cells detached from acrylic coupons was quantified by counting colony forming units (CFUs) in the solution.

To further evaluate the effects in an environment similar to that of oral cavity, the test medium was replaced with artificial saliva medium or a mixture of 0.85% NaCl and artificial saliva medium (2:1). The recipe of artificial saliva from Pratten et al. (1998) was followed. It contains 2 g/L yeast extract, 5 g/L peptone, 2.5 g/L type III hog gastric mucin, 0.2 g/L NaCl, 0.2 g/L KCl, and 0.3 g/L  $\text{CaCl}_2$ , supplemental with 1.25 mL of sterile 40% urea.

The CHX was tested at 50  $\mu\text{g}/\text{mL}$  to 500  $\mu\text{g}/\text{mL}$ . The treatment process was the same as described above for 0.85% NaCl solution.

#### Live/dead staining

To corroborate the CFU results, another set of acrylic coupons with biofilms treated with DC and CHX in the same way were stained with Live/Dead staining kit (Life Technologies Inc., Carlsbad, CA, U.S.) for 10 min. Then the biofilm samples were imaged using a fluorescence microscope (Axio Imager M1, Carl Zeiss Inc., Berlin, Germany).

#### Statistical analysis


All data are presented as mean  $\pm$  standard deviation. Statistical significance was assessed with two-way ANOVA

followed by Tukey test. Statistical significance was set as  $p < 0.05$ . All analyses were performed using SAS 9.4 software (SAS Institute, Cary, NC, USA).

## Results

### Effects of DC and CHX on *S. mutans* and *S. aureus* biofilms in 0.85% NaCl solution

As shown in Fig. 1, treatment with either CHX (at 5, 10, 20, 50, 100 and 200  $\mu\text{g}/\text{mL}$ , Fig. 1a) or DC (at 7, 14 and 28  $\mu\text{A}/\text{cm}^2$ , Fig. 1b) showed moderate but statistically significant killing ( $p < 0.05$ , two-way ANOVA followed by Tukey test). For example, up to 1.2 log and 0.7 log of killing was obtained with 28  $\mu\text{A}/\text{cm}^2$  DC and 50  $\mu\text{g}/\text{mL}$  CHX, respectively. Furthermore, synergy was observed between DC and CHX in killing *S. mutans* biofilms dose dependently. Among the tested conditions, the maximum



**Fig. 1** Viability of *S. mutans* biofilm cells after 1 h treatment. **a** Treatment with CHX alone. **b** Treatment with DC alone. **c** Concurrent treatment with CHX and DC. All treatments were tested in 0.85% NaCl solution

killing effect (4 logs) was observed under the condition of  $28 \mu\text{A}/\text{cm}^2$  DC and  $50 \mu\text{g}/\text{mL}$  CHX ( $p = 0.02$ , two-way ANOVA followed by Tukey test; Figs. 1c, 2a). The higher concentration of CHX (100 and  $200 \mu\text{g}/\text{mL}$ ) did not show significant increase in killing of *S. mutans* biofilm cells (compared to  $50 \mu\text{g}/\text{mL}$ ) both in the absence and presence of DC ( $p > 0.6$ , two-way ANOVA followed by Tukey test; Fig. 1a, c).

Similar synergistic effects were also observed for *S. aureus* biofilms under the same treatment conditions. The number of viable *S. aureus* biofilm cells was reduced by more than 5 logs ( $p < 0.0001$ , two-way ANOVA with Tukey test; Fig. 3a) after treatment with  $28 \mu\text{A}/\text{cm}^2$  DC and  $50 \mu\text{g}/\text{mL}$  CHX for 1 h in 0.85% NaCl solution. In comparison, treatment with the same level of DC or CHX alone only reduced the number of viable biofilm cells by  $60.0 \pm 7.9$  and  $74.3 \pm 2.5\%$  (less than 1 log for both conditions), respectively.


The CFU results were corroborated with fluorescence microscopy. According to the images from Live/Dead staining of *S. mutans* and *S. aureus* biofilms, the number of live cells (green) decreased when treated with DC and CHX even at low doses ( $7 \mu\text{A}/\text{cm}^2$  DC and  $5 \mu\text{g}/\text{mL}$  CHX for *S. mutans*,  $28 \mu\text{A}/\text{cm}^2$  DC and  $20 \mu\text{g}/\text{mL}$  CHX for *S. aureus*); and almost no live cells (green fluorescence only) were found on the surface of acrylic coupons after concurrent treatment with DC and CHX together (Figs. 4, 5). Compared with sample treated with DC alone, samples treated with both CHX and DC concurrently only have



patches of cell debris in red, suggesting that substantial cell lysis might have occurred.


During treatment, we also observed that some small particles and tiny gas bubbles were released from anode to cathode, respectively. The particles are metal oxides as we reported in our previous study on *P. aeruginosa* under the same experimental setup (Niepa et al. 2012, 2016a). Based on the half reaction potential of related species (Lide 2006), we speculate that hydrogen peroxide was generated by the reduction reaction of the oxygen in the solution. It will be interesting to further characterize these reactions and link the electrochemical reactions products to the killing effects. This is part of our ongoing work.

#### Effects in the presence of artificial saliva

Since the surface of dental implants is commonly covered with saliva, we also tested the effects of DC and CHX in the presence of artificial saliva. When artificial saliva was added to 0.85% NaCl solution as treatment medium, the killing effects were reduced but still significant ( $p < 0.05$ , two-way ANOVA followed by Tukey test; Additional file 1: Figures S1 and S2). For example, the reduction of biofilm cell viability was  $98.0 \pm 0.4\%$  ( $\sim 1.7$  log) when *S. aureus* biofilm was treated with  $50 \mu\text{g}/\text{mL}$  CHX and  $28 \mu\text{A}/\text{cm}^2$  DC in a mixture of artificial saliva and 0.85% NaCl solution (1:2 v/v) (Additional file 1: Figure S1). No killing effect was observed when biofilms were treated in pure artificial saliva medium under the same dosage of CHX or DC level.







**Fig. 5** Live/dead staining of *S. aureus* biofilms after treatment with CHX, DC, or concurrent treatment. **a** Without treatment. **b** Treatment with 20 µg/mL CHX. **c** Treatment with 28 µA/cm<sup>2</sup> DC. **d** Concurrent treatment with 20 µg/mL CHX and 28 µA/cm<sup>2</sup> DC. Bar = 20 µm

However, when the concentration of CHX increased to 500 µg/mL (0.05 w/v%; the dosage used in commercial oral rinsing products is 0.12 w/v%) while keeping the DC level at 28 µA/cm<sup>2</sup>, the number of viable *S. aureus* cells in biofilm was reduced by 2.5 logs compared to untreated control ( $p = 0.005$ , two-way ANOVA followed by Tukey test; Fig. 3b). The viability of biofilm cells treated with CHX alone was reduced by approximately 1 log and no significant killing effect was observed for 28 µA/cm<sup>2</sup> DC treatment alone (Fig. 3b). Similar results were observed for *S. mutans* biofilms (Fig. 2b and Additional file 1: Figure S1), although the killing of *S. mutans* biofilm cells in artificial saliva medium was lower than *S. aureus*. The number of viable cells was reduced by 0.54 log, 0.17 log, and 1.63 log when treated with CHX alone, DC alone, or concurrent treatment with CHX and DC, respectively ( $p = 0.02$ , two-way ANOVA followed by Tukey test; Fig. 2b).

## Discussion

Direct currents and alternative currents (AC) are known to kill biofilm cells in the presence or absence of antibiotics, and treatment time tested to date varies from hours to days (del Pozo et al. 2009; Schmidt-Malan et al. 2015; Spadaro et al. 1974). Our group recently found synergistic effect between low level DC and the antibiotic tobramycin in killing *Pseudomonas aeruginosa* biofilm and persister cells (Niepa et al. 2012; 2016b). However, most of previous studies focus on biofilms formed directly on the surface of electrodes.

To mimic real applications, it is important to test biofilms that are not in direct contact with electrodes. In this study, we set a sandwich structure with biofilms formed on acrylic coupons in the middle of the electric field and about 1.5 mm from each electrode. Our results show that the viability of *S. mutans* and *S. aureus*

biofilm cells (placed between two electrodes) on the surface of denture material can be reduced by low level DC and CHX through concurrent treatment in 1 h; and the effect was approximately 1–3 logs stronger than that obtained with the same level of DC or CHX alone indicating synergistic effects between DC and CHX in killing biofilm cells of these two dental bacteria. The effect was more profound in 0.85% NaCl solution than in the artificial saliva medium. The images of Live/Dead staining also confirmed that there was profound killing by concurrent treatment. We speculated that mucin and other proteins in artificial saliva medium might repress the killing effects of CHX and DC, since DC kills cells partially through the generation and movement of reactive species from electrochemical reactions (Niepa et al. 2012).

We speculated that this synergy was primarily resulted from the interaction between the products of DC treatment and CHX. In the recent studies, it showed that hydrogen peroxide was generated from electrode surface during electrical treatment of bacteria biofilms (Istanbulu et al. 2012; Sultana et al. 2015), which had been reported for its synergetic antibacterial effect with CHX against *streptococcus* and *staphylococcus* species (Steinberg et al. 1999). Furthermore, some metal ions ( $Zn^{2+}$ ,  $Cu^{2+}$ ) were shown for their capabilities to enhance the effect of CHX on different oral pathogens (Cronan et al. 2006; Drake et al. 1993). The stainless-steel electrodes used in this study have a larger surface area and can release multiple types of metal ions including  $Fe^{2+}$ ,  $Fe^{3+}$ ,  $Cr^{2+}$ ,  $Cr^{3+}$  and  $Cr^{6+}$  during DC treatment (Niepa et al. 2012).  $Fe^{2+}$  and  $Fe^{3+}$  ions were found to kill *P. aeruginosa* persister cells in the presence of antibiotics in an electric field (Niepa et al. 2016a). Our lab also found that  $Cr^{3+}$  and  $Cr^{6+}$  can form complex with certain antibiotic compounds, and thus increase the affinity between antibiotics and intracellular targets (Niepa et al. 2016b). It is possible that some released ions interact with CHX molecules and result in the observed synergy in killing *S. mutans* and *S. aureus*. This is part of our ongoing study.

Recently, Lasserre et al. (2015) reported that the viability of *P. gingivalis* biofilm could be reduced by 81.1 and 98.9% in 10 min when treated with 2000  $\mu\text{g/mL}$  (0.2 w/v%) CHX alone and concurrent treatment with same dosage of CHX and 5882  $\mu\text{A/cm}^2$  DC, respectively; while the treatment with DC itself did not kill *P. gingivalis* cells. The biofilms were cultured on the discs of a Modified Robbins Device (MRD), which were placed between two electrodes of platinum wires in the MRD's chamber. This is an exciting discovery, but the DC level appears high and may not be suitable for in vivo therapy, especially for the implants close to nervous systems that do not tolerate more than a maximum current density of 30  $\mu\text{A/cm}^2$  (McCreery et al. 1990; Shannon 1992; Clark

2003). Hence, it is necessary to reduce DC to  $\mu\text{A}$  level for future in vivo applications. In this study, we treated *S. aureus* and *S. mutans* biofilm without direct contact to electrodes by placing an acrylic coupon in the middle of a low-level electric field and parallel to the electrode surfaces. By using stainless steel as electrode material, the level of DC and CHX in our study are much lower (28  $\mu\text{A/cm}^2$  DC and 50  $\mu\text{g/mL}$  CHX), and strong killing effects (3–4 logs) were obtained.

CHX is bacteriostatic at low concentrations by affecting the integrity of bacterial cell wall and bactericidal at high concentrations by disrupting the cell (McDonnell and Russell 1999). *S. mutans* and *S. aureus* appear to be quite susceptible to CHX according to MIC data (< 8  $\mu\text{g/mL}$ ) (Chung et al. 2006). However, the maximum killing of preformed biofilms by CHX alone in our experimental system was only less than 1.5 logs even with a dosage up to 500  $\mu\text{g/mL}$ .

Through synergy with DC, CHX was found to be more effective in killing biofilm cells. The effective doses of CHX we used were only 50  $\mu\text{g/mL}$  (0.005 w/v%) in 0.85% NaCl solution and 500  $\mu\text{g/mL}$  (0.05 w/v%) in artificial saliva medium. This CHX level is expected to be safe because the commercial products for oral wash have approximately 1200  $\mu\text{g/mL}$  (0.12 w/v%)–2000  $\mu\text{g/mL}$  (0.2 w/v%) of CHX. The exact mechanism for such synergistic killing is unknown and is part of our ongoing research.

In summary, we demonstrated that the biofilm cells of two Gram-positive pathogenic bacteria, *S. mutans* and *S. aureus*, could be efficiently killed by concurrent treatment with low level DC and CHX in 1 h. This electrochemical control is effective against the biofilms formed on the acrylic materials. The synergistic effect between DC and CHX can help design new devices and strategies for controlling pathogenic biofilms. The interaction between electrochemical products and CHX may play a significant role in the observed synergy in biofilm killing, which deserves further study.

## Additional file

**Additional file 1: Figure S1.** Viability of *S. aureus* biofilm cells after 1 h treatment with 50  $\mu\text{g/mL}$  CHX alone, 28  $\mu\text{A/cm}^2$  DC alone or concurrent treatment with CHX and DC. The treatments were tested in a mixture of 0.85% NaCl and artificial saliva medium (2:1, v/v). **Figure S2.** Viability of *S. mutans* biofilm cells after 1 h treatment with 50  $\mu\text{g/mL}$  CHX alone, 28  $\mu\text{A/cm}^2$  DC alone or concurrent treatment with CHX and DC. The treatments were tested in a mixture of 0.85% NaCl and artificial saliva medium (2:1, v/v).

## Abbreviations

AC: alternative currents; ANOVA: analysis of variance; BHI: brain heart infusion; CHX: chlorhexidine; DC: direct current; EPS: extracellular polysaccharides; MRD: modified Robbins device; LB: lysogeny broth; *S. aureus*: *Staphylococcus aureus*; *S. mutans*: *Streptococcus mutans*.

**Authors' contributions**

HW and DR designed the research. DR supervised the research. HW performed all experiments and analyzed data with DR. HW and DR wrote the manuscript. Both authors read and approved the final manuscript.

**Author details**

<sup>1</sup> Department of Biomedical and Chemical Engineering, Syracuse University, Syracuse, NY 13244, USA. <sup>2</sup> Syracuse Biomaterials Institute, Syracuse University, Syracuse, NY 13244, USA. <sup>3</sup> Department of Civil and Environmental Engineering, Syracuse University, Syracuse, NY 13244, USA. <sup>4</sup> Department of Biology, Syracuse University, Syracuse, NY 13244, USA.

**Acknowledgements**

We thank Dr. Hyun Koo at University of Pennsylvania for helpful comments on the manuscript.

**Competing interests**

The authors declare that they have no competing interests.

**Availability of data and materials**

The datasets supporting the conclusions of this article are included within the article and its supplement materials.

**Consent for publication**

Not applicable.

**Ethics approval and consent to participate**

Not applicable.

**Funding**

The authors are grateful for the support of the Milton and Ann Stevenson Endowment.

**Publisher's Note**

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Received: 24 September 2017 Accepted: 7 November 2017

Published online: 15 November 2017

**References**

- Bender GR, Sutton SV, Marquis RE (1986) Acid tolerance, proton permeabilities, and membrane ATPases of oral *streptococci*. *Infect Immun* 53:331–338
- Chung JY, Choo JH, Lee MH, Hwang JK (2006) Anticariogenic activity of macelignan isolated from *Myristica fragrans* (nutmeg) against *Streptococcus mutans*. *Phytomedicine* 13:261–266
- Clark G (2003) *Cochlear Implants: Fundamentals and applications*. Springer Verlag, New York, p 171
- Costerton JW, Ellis B, Lam K, Johnson F, Khoury AE (1994) Mechanism of electrical enhancement of efficacy of antibiotics in killing biofilm bacteria. *Antimicrob Agents Chemother* 38:2803–2809
- Cronan CA, Potempa J, Travis J, Mayo JA (2006) Inhibition of *Porphyromonas gingivalis* proteinases (gingipains) by chlorhexidine: synergistic effect of Zn(II). *Oral Microbiol Immunol* 21:212–217
- del Pozo JL, Rouse MS, Mandrekar JN, Steckelberg JM, Patel R (2009) The electricidal effect: reduction of *Staphylococcus* and *Pseudomonas* biofilms by prolonged exposure to low-intensity electrical current. *Antimicrob Agents Chemother* 53:41–45. <https://doi.org/10.1128/AAC.00680-08>
- Drake DR, Grigsby W, Cardenzana A, Dunkerson D (1993) Synergistic, growth-inhibitory effects of chlorhexidine and copper combinations on *Streptococcus mutans*, *Actinomyces viscosus*, and *Actinomyces naeslundii*. *J Dent Res* 72:524–528
- Falsetta ML, Klein MI, Lemos JA, Silva BB, Agidi S, Scott-Anne KK, Koo H (2012) Novel antibiofilm chemotherapy targets exopolysaccharide synthesis and stress tolerance in *Streptococcus mutans* to modulate virulence expression in vivo. *Antimicrob Agents Chemother* 56:6201–6211. <https://doi.org/10.1128/AAC.01381-12>
- Falsetta ML, Klein MI, Colonne PM, Scott-Anne K, Gregoire S, Pai CH, Gonzalez-Begne M, Watson G, Krysan DJ, Bowen WH, Koo H (2014) Symbiotic relationship between *Streptococcus mutans* and *Candida albicans* synergizes virulence of plaque biofilms in vivo. *Infect Immun* 82:1968–1981. <https://doi.org/10.1128/IAI.00087-14>
- Flotra L, Gjermo P, Rolla G, Waerhaug J (1971) Side effects of chlorhexidine mouth washes. *Scand J Dent Res* 79:119–125
- Hall CW, Mah TF (2017) Molecular mechanisms of biofilm-based antibiotic resistance and tolerance in pathogenic bacteria. *FEMS Microbiol Rev* 41:276–301. <https://doi.org/10.1093/femsre/fux010>
- Hansson C, Hoborn J, Moller A, Swanbeck G (1995) The microbial flora in venous leg ulcers without clinical signs of infection. Repeated culture using a validated standardised microbiological technique. *Acta Derm Venereol* 75:24–30
- Harper DS, Loesche WJ (1984) Growth and acid tolerance of human dental plaque bacteria. *Arch Oral Biol* 29:843–848
- Harris LG, Richards PG (2004) *Staphylococcus aureus* adhesion to different treated titanium surfaces. *J Mat Sci Mat Med* 15:311–314
- Istanbul O, Babauta J, Duc Nguyen H, Beyenal H (2012) Electrochemical biofilm control: mechanism of action. *Biofouling* 28(8):769–778. <https://doi.org/10.1080/08927014.2012.707651>
- Jones CG (1997) Chlorhexidine: is it still the gold standard? *Periodontol* 2000 15:55–62
- Jones SM, Morgan M, Humphrey TJ, Lappin-Scott H (2001) Effect of vancomycin and rifampicin on methicillin-resistant *Staphylococcus aureus* biofilms. *Lancet* 357:40–41
- Kouidhi B, Al Qurashi YM, Chaieb K (2015) Drug resistance of bacterial dental biofilm and the potential use of natural compounds as alternative for prevention and treatment. *Microb Pathog* 80:39–49. <https://doi.org/10.1016/j.micpath.2015.02.007>
- Kreth J, Merritt J, Shi W, Qi F (2005) Competition and coexistence between *Streptococcus mutans* and *Streptococcus sanguinis* in the dental biofilm. *J Bacteriol* 187:7193–7203
- Lasserre JF, Leprince JG, Toma S, Brex MC (2015) Electrical enhancement of chlorhexidine efficacy against the periodontal pathogen *Porphyromonas gingivalis* within a biofilm. *New Microbiol* 38:511–519
- Lew DP, Waldvogel FA (2004) Osteomyelitis. *Lancet* 364:369–379
- Lewis N, Parmar N, Hussain Z, Baker G, Green I, Howlett J, Kearns A, Cookson B, McDonald A, Wilson M, Ready D (2015) Colonization of dentures by *Staphylococcus aureus* and MRSA in out-patient and in-patient populations. *Eur J Clin Microbiol Infect Dis* 34:1823–1826. <https://doi.org/10.1007/s10096-015-2418-6>
- Lide DR (2006) *CRC handbook of chemistry and physics*, 87th edn. CRC Press, Boca Raton
- Liu Y, Kamesh AC, Xiao Y, Sun V, Hayes M, Daniell H, Koo H (2016) Topical delivery of low-cost protein drug candidates made in chloroplasts for biofilm disruption and uptake by oral epithelial cells. *Biomaterials* 105:156–166. <https://doi.org/10.1016/j.biomaterials.2016.07.042>
- McCreery DB, Agnew WF, Yuen TG, Bullara L (1990) Charge density and charge per phase as cofactors in neural injury induced by electrical stimulation. *IEEE Trans Biomed Eng* 7:996–1001
- McDonnell G, Russell AD (1999) Antiseptics and disinfectants: activity, action, and resistance. *Clin Microbiol Rev* 12:147–179
- Murchison H, Larrimore S, Hull S, Curtiss R 3rd (1982) Isolation and characterization of *Streptococcus mutans* mutants with altered cellular morphology or chain length. *Infect Immun* 38(1):282–291
- Niepa TH, Gilbert JL, Ren D (2012) Controlling *Pseudomonas aeruginosa* persister cells by weak electrochemical currents and synergistic effects with tobramycin. *Biomaterials* 33:7356–7365. <https://doi.org/10.1016/j.biomaterials.2012.06.092>
- Niepa TH, Snepenger LM, Wang H, Sivan S, Gilbert JL, Jones MB, Ren D (2016a) Sensitizing *Pseudomonas aeruginosa* to antibiotics by electrochemical disruption of membrane functions. *Biomaterials* 74:267–279. <https://doi.org/10.1016/j.biomaterials.2015.10.007>
- Niepa TH, Wang H, Gilbert JL, Dabrowiak JC, Ren D (2016b) Synergy between tobramycin and trivalent chromium ion in electrochemical control of *Pseudomonas aeruginosa*. *Acta Biomater* 36:286–295. <https://doi.org/10.1016/j.actbio.2016.03.028>
- O'Donnell LE, Smith K, Williams C, Nile CJ, Lappin DF, Bradshaw D, Lambert M, Robertson DP, Bagg J, Hannah V, Ramage G (2015) Dentures are a



- reservoir for respiratory pathogens. *J Prosthodont* 25(2):99–104. <https://doi.org/10.1111/jopr.12342>
- Pratten J, Wills K, Barnett P, Wilson M (1998) In vitro studies of the effect of antiseptic-containing mouthwashes on the formation and viability of *Streptococcus sanguis* biofilms. *J Appl Microbiol* 84:1149–1155
- Salvi GE, Furst MM, Lang NP, Persson GR (2008) One-year bacterial colonization patterns of *Staphylococcus aureus* and other bacteria at implants and adjacent teeth. *Clin Oral Implants Res* 19:242–248
- Sambrook, J, Russell, DW (2001) *Molecular cloning: a laboratory manual*. Cold spring harbor laboratory press
- Sauer K, Steczko J, Ash SR (2009) Effect of a solution containing citrate/methylene blue/parabens on *Staphylococcus aureus* bacteria and biofilm, and comparison with various heparin solutions. *J Antimicrob Chemother* 63:937–945
- Schmidt-Malan SM, Karau MJ, Cede J, Greenwood-Quaintance KE, Brinkman CL, Mandrekar JN, Patel R (2015) Antibiofilm activity of low-amperage continuous and intermittent direct electrical current. *Antimicrob Agents Chemother* 59:4610–4615. <https://doi.org/10.1128/AAC.00483-15>
- Shannon RV (1992) A model of safe levels for electrical stimulation. *IEEE Trans Biomed Eng* 39:424–426
- Smith A, Buchinsky FJ, Post JC (2011) Eradicating chronic ear, nose, and throat infections: a systematically conducted literature review of advances in biofilm treatment. *Otolaryngol Head Neck Surg* 144:338–347. <https://doi.org/10.1177/0194599810391620>
- Song F, Koo H, Ren D (2015) Effects of material properties on bacterial adhesion and biofilm formation. *J Dent Res* 94:1027–1034. <https://doi.org/10.1177/0022034515587690>
- Spadaro JA, Berger TJ, Barranco SD, Chapin SE, Becker RO (1974) Antibacterial effects of silver electrodes with weak direct current. *Antimicrob Agents Chemother* 6:637–642
- Steinberg D, Heling I, Daniel I, Ginsburg I (1999) Antibacterial synergistic effect of chlorhexidine and hydrogen peroxide against *Streptococcus sobrinus*, *Streptococcus faecalis* and *Staphylococcus aureus*. *J Oral Rehabil* 26:151–156
- Stephenson MF, Mfuna L, Dowd SE, Wolcott RD, Barbeau J, Poisson M, James G, Desrosiers M (2010) Molecular characterization of the polymicrobial flora in chronic rhinosinusitis. *J Otolaryngol Head Neck Surg* 39:182–187
- Sultana ST, Atci E, Babauta JT, Falghoush AM, Snekvik KR, Call DR, Beyenal H (2015) Electrochemical scaffold generates localized, low concentration of hydrogen peroxide that inhibits bacterial pathogens and biofilms. *Sci Rep* 5:14908. <https://doi.org/10.1038/srep14908>
- Wattanakaroon W, Stewart PS (2000) Electrical enhancement of *Streptococcus gordonii* biofilm killing by gentamicin. *Arch Oral Biol* 45:167–171

Submit your manuscript to a SpringerOpen<sup>®</sup> journal and benefit from:

- Convenient online submission
- Rigorous peer review
- Open access: articles freely available online
- High visibility within the field
- Retaining the copyright to your article

---

Submit your next manuscript at ► [springeropen.com](http://springeropen.com)

---