



Evaluation of a New Immunochromatography Technology Test (LDBio Diagnostics) To Detect *Toxoplasma* IgG and IgM: Comparison with the Routine Architect Technique

Caroline Mahinc,^a Pierre Flori,^a Edouard Delaunay,^b Cécile Guillerme,^a Sana Charaoui,^a Hélène Raberin,^a Jamal Hafid,^{a,c} Coralie L'Ollivier^b

Pôle de Biologie-Pathologie, Parasitology and Mycology Laboratory, University Hospital of Saint Etienne, Saint Etienne, France^a; Laboratoire de Parasitologie-Mycologie, Hôpital de la Timone, Marseille, France^b; Unité d'Immunologie et de Physiologie, Faculté des Sciences et Techniques, Marrakech, Morocco^c

ABSTRACT A study comparing the ICT (immunochromatography technology) *Toxoplasma* IgG and IgM rapid diagnostic test (LDBio Diagnostics, France) with a fully automated system, Architect, was performed on samples from university hospitals of Marseille and Saint-Etienne. A total of 767 prospective sera and 235 selected sera were collected. The panels were selected to test various IgG and IgM parameters. The reference technique, *Toxoplasma* IgGII Western blot analysis (LDBio Diagnostics), was used to confirm the IgG results, and commercial kits Platelia Toxo IgM (Bio-Rad) and Toxo-ISAgA (bioMérieux) were used in Saint-Etienne and Marseille, respectively, as the IgM reference techniques. Sensitivity and specificity of the ICT and the Architect IgG assays were compared using a prospective panel. Sensitivity was 100% for the ICT test and 92.1% for Architect (cutoff at 1.6 IU/ml). The low-IgG-titer serum results confirmed that ICT sensitivity was superior to that of Architect. Specificity was 98.7% (ICT) and 99.8% (Architect IgG). The ICT test is also useful for detecting IgM without IgG and is both sensitive (100%) and specific (100%), as it can distinguish nonspecific IgM from specific *Toxoplasma* IgM. In comparison, IgM sensitivity and specificity on Architect are 96.1% and 99.6%, respectively (cutoff at 0.5 arbitrary units [AU]/ml). To conclude, this new test overcomes the limitations of automated screening techniques, which are not sensitive enough for IgG and lack specificity for IgM (rare IgM false-positive cases).

KEYWORDS toxoplasmosis, *Toxoplasma gondii*, immunoglobulin G, immunoglobulin M, serology, immunochromatography test, Architect

Toxoplasmosis, which is caused by *Toxoplasma gondii*, is usually asymptomatic and benign in immunocompetent humans. In pregnant women, maternal transmission may result in congenital toxoplasmosis, which may cause severe disease or sequelae (1). Medical follow-up of obstetrical toxoplasmosis is essential for seronegative women. In severe cellular immunodeficiency, reactivation of the infection causes acute neurological damage and can be lethal if not successfully treated. For these patients, detection of *Toxoplasma*-specific antibodies showing serological reactivation or primary infection is essential to properly diagnose and prevent severe toxoplasmosis.

The follow-up of obstetrical toxoplasmosis primarily involves the detection of anti-*Toxoplasma*-specific immunoglobulin M (IgM) and immunoglobulin G (IgG) antibodies (2–4). Detection of *Toxoplasma*-specific IgM with the first blood test is usually cause for concern. The presence of *Toxoplasma*-specific IgG without IgM confirms the immunization of the patient, thus avoiding unnecessary and expensive follow-up.

Received 17 July 2017 Returned for modification 7 August 2017 Accepted 8 September 2017

Accepted manuscript posted online 27 September 2017

Citation Mahinc C, Flori P, Delaunay E, Guillerme C, Charaoui S, Raberin H, Hafid J, L'Ollivier C. 2017. Evaluation of a new immunochromatography technology test (LDBio Diagnostics) to detect *Toxoplasma* IgG and IgM: comparison with the routine Architect technique. *J Clin Microbiol* 55:3395–3404. <https://doi.org/10.1128/JCM.01106-17>.

Editor Michael J. Loeffelholz, University of Texas Medical Branch

Copyright © 2017 American Society for Microbiology. All Rights Reserved.

Address correspondence to Caroline Mahinc, caroline.mahinc@chu-st-etienne.fr.




FIG 1 Distribution of nonselected IgG serum titers with Architect. The dark columns correspond to sera that are positive by ICT IgG-IgM (LDBIO Diagnostics). The gray columns correspond to sera that are negative by ICT IgG-IgM (LDBIO Diagnostics).

As the prevalence of the disease has declined in Europe and France in particular (5), the number of women to be controlled during pregnancy has increased (46% to 63% between 1995 and 2010 in France), which influences the cost-benefit aspect of a mass systematic screening program.

In this study, we assessed the performance of the immunochromatography technology (ICT) *Toxoplasma* IgG and IgM rapid diagnostic test (LDBio Diagnostics, France) and compared the results with those of the Architect system. This automated technique is reliable for first-line serodiagnosis (6) and was chosen as the screening technique in Saint-Etienne and Marseille. This assessment is critical to define a good serological strategy based on the specificity and sensitivity of the two techniques. The aim of this study was to determine whether the ICT test can overcome the limits of the screening technique and ultimately be used as a second-line test.

RESULTS

The evaluation of the ICT *Toxoplasma* IgG and IgM test performance was performed as described below.

Panel 1: 767 nonselected prospective sera. (i) IgG analysis. The IgG value distribution provided by the Architect system and ICT test are shown in Fig. 1. The IgG concordance analysis between the two techniques is also detailed in Table 1. Of the 767

TABLE 1 Analysis of IgG concordant and discrepant sera between ICT and Toxo IgG Architect test by testing of 767 nonselected samples

IgG analysis	n	Architect value ^a (UI/ml)	ICT ^b	Reference Western blot ^c (LDBio Toxo II IgG)	Conclusion
Concordant	582	0–1.3 (Neg)	Neg	NR	Neg IgG
Concordant	143	≥3	Pos	NR	Pos IgG
Minor discrepant	20	1.6–2.9 (gray zone)	Pos	Pos	Pos IgG
Discrepant	14	0.6–1.5 (Neg)	Pos	Pos	Pos IgG (false-negative Architect)
Discrepant	1	2.5 (gray zone)	Pos	Neg	Neg IgG (false-positive ICT)
Discrepant	7	0.1–0.3 (Neg)	Pos	Neg	Neg IgG (false-positive ICT)

^aIgG concentrations on Architect: negative, ≤1.5 IU/ml; gray zone, 1.6 to 2.9 IU/ml; positive, ≥3 IU/ml. Neg, negative.

^bPos, positive.

^cNR, not realized.

TABLE 2 Analysis of IgM concordant and discrepant sera between ICT IgG-IgM and Toxo IgM Architect by testing of 767 nonselected samples

IgM analysis	No. of samples ^a (total <i>n</i> = 767, total M = 356, total SE = 411)	Architect IgM value (AU/ml)	ICT (IgM-IgG)	Reference IgM (M = ISAgA, SE = Platelia)	IgG analysis	Conclusion (follow-up)
Concordant with IgG analysis	737 (M = 340, SE = 397)	0–0.49 (Neg)	Neg (571), Pos (166)	NR	Neg IgG (571), Pos IgG (166)	Neg IgM
Concordant	15 (M = 8, SE = 7)	>0.60 (Pos)	Pos	Pos	Pos IgG	Pos IgM (residual with IgG)
Concordant	6 (M = 2, SE = 4)	>0.60 (Pos)	Pos	Pos	Neg or equivocal IgG	Beginning primary infection: IgM without IgG
Minor discrepant	4 (M = 3, SE = 1)	0.50–0.59 (gray zone)	Pos	Pos	Pos IgG	Pos IgM (residual with IgG)
Minor discrepant	1 (M = 1)	0.50–0.59 (gray zone)	Pos	Pos	Neg IgG	Beginning primary infection: IgM without IgG
Minor discrepant	1 (M = 1)	0.50–0.59 (gray zone)	Neg	Neg	Neg IgG	Neg IgM
Discrepant	3 (M = 2, SE = 1)	>0.60 (Pos)	Neg	1 Neg (M), 2 Pos ^b (1 M, 1 SE)	Neg IgG	Neg IgM

^aM, sera from Marseille tested with reference technique IgM immunosorbent agglutination assay (ISAgA; bioMérieux, Marcy l'Etoile, France). SE, sera from Saint-Etienne tested with reference technique Platelia Toxo IgM (Platelia; Bio-Rad, Marne la Coquette, France).

^bThese two patients were positive by Architect IgM and immunocapture, but the serological IgG follow-up at 2 weeks and 1 month did not reveal the appearance of IgG; nonspecific IgM was found.

serum samples tested, 582 sera were negative according to both the Architect and ICT assays. All of the 143 sera with a positive result according to the Architect (≥ 3.0 IU/ml) system were also positive with the ICT. Regarding 20 sera that were in the gray zone using the Architect system (1.6 to 2.9 IU/ml), these samples were positive according to the ICT test, which was confirmed by the IgGII Western blot. Additionally, 14 negative samples that fell between 0.6 and 1.5 IU/ml on the Architect system were positive with the ICT test and confirmed by the IgGII Western blot. We also found eight false-positive results using the ICT test (7 negative sera with negative IgM and IgG between 0.1 and 0.3 IU/ml on Architect, one negative IgM with IgG at 2.5 IU/ml on Architect).

(ii) IgM analysis. Among 737 IgM-negative sera on Architect (0 to 0.49 arbitrary units [AU]/ml), 571 were also IgG negative and had a negative ICT. For the other 166 negative IgM sera, they were positive for IgG and had a positive ICT (IgG and IgM) (Table 2).

Concerning the positive results on the Architect system (>0.6 AU/ml), 15 were concordant with the ICT test and confirmed by immunocapture, while 6 were positive, but without IgG, according to both the ICT test and immunocapture (indicating early primary infection). The other 9 sera are major or minor discrepant and are detailed in Table 2.

(iii) Sensitivity, specificity, NPV, and PPV. The concordance between the Architect and ICT systems concerning sensitivity, specificity, positive predictive value (PPV), and negative predictive value (NPV) are shown in Table 3. Considering equivocal samples with Architect as positive (6) (see results described below for low IgG titers), the correlation between the two techniques was 97.2% for IgG and 99.6% for IgM.

Taking reference techniques and clinical data together, the ICT assay shows a sensitivity of 100% and a specificity of 98.7% for IgG, 100% sensitivity and specificity for IgM, and an NVP and PPV for IgG of 100% and 95.8%, respectively.

Analysis of the 235 selected sera in panel 2. (i) Panel 2.1: low IgG titers (*n* = 92). The results of 92 selected sera with low IgG titers (0.6 to 3.4 IU/ml) are shown in Table 4. IgG-positive results on the Architect system were set to 3 IU/ml. We found that for values between 2.1 and 3.4 IU/ml, both the ICT and IgGII Western blot results were positive for 100% of the 38 sera tested. For 47 cases in the large range from 0.8 to 2.0 IU/ml (31 IgG negative and 16 IgG Architect gray zone), the concordance between the ICT and the IgGII Western blot was perfect. Most of the cases were positive (41/47), whereas some were IgG negative with Architect (31/47). For the 7 sera with IgG values ranging from 0.6 to 0.7 IU/ml (negative on Architect), 3 ICT results were positive while 4 sera subjected to IgGII Western blot were positive.

TABLE 3 Concordance, Se, Sp, PPV, NPV, Youden index, and Yule Q by testing 767 nonselected samples for anti-*Toxoplasma* IgG and/or IgM^d

Test	Concordance with ICT (%)	Se ^a (%)	NPV (%)	Sp ^b (%)	PPV (%)	Youden index (Se + Sp - 1)	Yule Q (%)
ICT for IgG	NA ^c	100 (98.5–100)	100 (98.5–100)	98.7 (97.8–99.6)	95.8 (94.2–97.4)	0.99	100
Architect IgG (cutoff, 1.6 IU/ml)	97.2 (96.0–98.4)	92.1 (88.1–96.1)	97.7 (95.5–99.8)	99.8 (99.4–100)	99.4 (98.8–100)	0.92	100
Architect IgG (cutoff, 3.0 IU/ml)	94.5 (92.9–96.1)	80.7 (74.9–86.5)	94.6 (91.3–97.9)	100 (99.6–100)	100 (99.6–100)	0.81	100
ICT for IgM	NA	100 (92.6–100)	NA	100 (98.3–100)	NA	NA	NA
Architect IgM (cutoff, 0.5 AU/ml)	99.6 (99.1–100)	96.1 (88.7–100)	NA	99.6 (98.3–100)	NA	NA	NA
Architect IgM (cutoff, 0.6 AU/ml)	99.1 (98.4–99.8)	80.8 (65.7–95.9)	NA	99.7 (98.3–100)	NA	NA	NA

^aSe, sensitivity defined as true positive/(true positive + false negative).

^bSp, specificity defined as true negative/(true negative + false positive).

^cNA, not applicable.

^dValues in parentheses are 95% confidence intervals.

(ii) Panel 2.2: high IgG titers (n = 20). Regarding the sera with high IgG titers (range, 125 to 2,000 IU/ml), all ICT were positive (weakly or strongly positive depending on strip intensity), but there was no correlation between the titer and the intensity of the test band.

(iii) Panel 2.3: sera followed by seroconversion (n = 50). For 33 sera, IgG and IgM were associated. Regarding 14 sera with IgM of >0.5 AU/ml without IgG (IgG of <1.5 IU/ml), both the immunocapture test and the ICT were positive. For the last 3 sera corresponding to the first sample of a proved seroconversion, 2 were positive with ICT and immunosorbent agglutination assay (ISAgA), whereas only one was negative with all techniques.

(iv) Panel 2.4: sera without seroconversion (n = 33) and with nonspecific IgM on Architect (n = 23). The 10 negative or equivocal sera (0.35 < IgM < 0.59 AU/ml) were negative according to the ICT and immunocapture assays. Concerning the 23 other sera (IgM positive by Architect), we conclude that all of these corresponded to false-positive results, since the IgG follow-up did not show the appearance of IgG and increasing IgM. Ten out of 23 sera tested positive on the Architect system and negative by immunocapture and ICT test, which indicates a false-positive result on Architect. Eleven out of 23 samples yielded positive immunocapture and Architect results but negative ICT results and corresponded to false-positive results on both the Architect system and immunocapture assay. We found that 2/23 sera tested positive for IgM with the three techniques.

(v) Panel 2.5: false-positive IgG (n = 11). The 11 sera with IgG between 3.2 and 21.6 IU/ml on the Architect system had a negative ICT. The Western blot (Toxo II IgG LDBio Diagnostics) results were negative with different types of profiles (Fig. 2). Profile 1 displayed IgG at 3.2 IU/ml on Architect with 2 bands (negative) (P35 to P40). Profiles

TABLE 4 Analysis of IgG concordant and discrepant sera between ICT and Toxo IgG Architect for ninety-two selected sera with low IgG titer

Architect IgG value ^a (IU/ml)	No. of sera	Positive ICT [no. positive/total no. (%)]	Conclusion [no. with positive IgGII Western blot/total no. (%)]
0.6–0.7	7	3/7 (43)	4/7 (57)
0.8–0.9	7	5/7 (71)	5/7 (71)
1–1.2	12	11/12 (92)	11/12 (92)
1.3–1.5	12	10/12 (83)	10/12 (83)
1.6–2	16	15/16 (94)	15/16 (94)
2.1–2.9	27	27/27 (100)	27/27 (100)
3–3.4	11	11/11 (100)	11/11 (100)

^aIgG concentrations on Architect: negative, ≤1.5 IU/ml; gray zone, 1.6 to 2.9 IU/ml; positive, ≥3 IU/ml.




FIG 2 *Toxoplasma* IgGII Western blot (LDBIO Diagnostics) profiles. These profiles were obtained from 11 sera of the panel corresponding to nonspecific IgG on Architect with a negative ICT. C, control; C1, positive control with IgG of 5 UI/ml on Architect; C2, positive control with IgG of 1.2 UI/ml on Architect; C3, positive control of a confirmed seroconversion with IgG of 0.6 UI/ml on Architect; P, panel. Eleven sera were used: P1, IgG of 3.2 UI/ml on Architect, negative ICT, 2 bands (i.e., negative) with Toxo II IgG Western blot analysis; P2 to P4, IgG of 3.2 to 21.6 UI/ml on Architect, negative ICT, 1 band (i.e., negative) with Toxo II IgG Western blot analysis; P5 to P11, IgG of 3.8 to 6.7 UI/ml on Architect, negative ICT, 0 band (i.e., negative) with Toxo II IgG Western blot analysis; P6 to P10, not presented.

2 to 4 displayed IgG of 3.2 to 21.6 IU/ml on the Architect system with a single band (negative) (P40). Profiles 5 to 11 displayed no bands.

(vi) Panel 2.6: analysis of selected sera of toxoplasmosis-seronegative patients and with potentially cross-reacting sera for CMV infection, EBV infection, HIV infection, and presence of RF. The panel of 6 sera from patients with anti-CMV (cytomegalovirus) antibodies included 4 sera with IgG- and IgM-specific anti-CMV antibodies, which probably corresponded to acute infection. Among 13 sera with positive Epstein-Barr virus (EBV) serology, 8 had both IgM and VCA-IgG; these results corresponded to a probable primary EBV infection. We selected 6 HIV-positive sera that corresponded to a recent primary infection (positive P24 and antigen viral load between 3.5 and 5 log copies/ml). All sera were seronegative for *Toxoplasma*, but 4 ICT were false-positive tests, indicating probable HIV serology, EBV infection, and rheumatoid factor (RF). To summarize, among these 29 sera, we found the following nonspecific reactions: one positive for IgG on the Architect system, three equivocal or positive for IgM on the Architect system, and four equivocal or positive for IgG-IgM according to the ICT test (Table 5).

DISCUSSION

Serological diagnosis of toxoplasmosis requires quantitative tests, most of which are automated techniques. The automated techniques have several advantages: they are

TABLE 5 Twenty-nine selected sera from toxoplasmosis-seronegative patients with CMV, EBV, HIV, and the presence of RF

Toxoplasmosis-seronegative sera	No. of sera	No. positive by:		
		IgG Architect	IgM Architect	ICT
CMV infection	6	0/6	0/6	0/6
EBV infection	13	0/13	3/13 (gray zone or positive)	1/13 (weakly positive)
HIV infection	5	0/5	0/5	2/5
RF	5	1/5 (gray zone)	0/5	1/5
Total	29	1/29	3/29	4/29

adapted to the routine analysis of large quantities of samples and produce quantitative and reproducible IgG and IgM results, particularly for the Architect technique, with excellent specificity for IgG (near 100%) and excellent sensitivity for IgM (6–8).

However, automated techniques have drawbacks: concerning the IgG parameter, most of these techniques lack sensitivity (Vidas, Access, Centaur, Immulite, AxSym, Platelia, and Enzygnost) (6, 7) found with the Architect system. Concerning IgM, some automated techniques can display an imperfect sensitivity, except the Platelia system, which is well correlated with the ISAgA system (6).

This study was carried out to evaluate the ICT *Toxoplasma* IgG and IgM rapid diagnostic test. We aimed to determine whether the ICT test is able to overcome the sensitivity limits of automated techniques, particularly those of the Architect system, which is used as a first-line diagnostic approach. A similar recent study conducted in Lyon by Chapey et al. also evaluated the ICT test (8) and found sensitivity of 97% and specificity of 96%, which are similar to results of our study. However, in this study, the reliability of the ICT test has been compared to that of the Architect system and not to the reference technique. Therefore, the expertise of discrepant cases is uncertain and does not highlight the interests and the limits of this new test. Our study cohort was larger than that of the Lyon study, as it included 1,002 sera from two centers (Marseille and Saint-Etienne). Furthermore, we selected various categories of sera to evaluate the relevance of the test in critical situations (e.g., low IgG titers, seroconversion, nonspecific IgM titers, etc.).

Another recent study has evaluated this new technique (9). The methodological approach is different and complementary (in relation to our work). This study uses 180 sera: 115 chronically infected persons with known serotype (48 serotype II, 14 serotype I-III, serotype I-IIIa, and 28 serotype atypical, haplogroup 12), 51 seronegative samples, and 13 samples from recently infected persons. In this study, the ICT test had 100% sensitivity and specificity. The authors also developed the theme of economic considerations, and we agree that the ICT test offers new options for improved prenatal care in low- and middle-income countries; it facilitates early identification and diagnosis, with similar or better sensitivity and specificity than automated techniques.

The ICT is a qualitative test that simultaneously detects specific *Toxoplasma* IgG and IgM. A positive result can be caused by the presence of IgG and/or IgM anti-*Toxoplasma*. Regarding nonselected sera, our study revealed that sensitivity for IgG was 100% with the ICT test (Table 3), which correlated with the Western blot results. This is not surprising, as the ICT test is calibrated with the *Toxoplasma* IgGII Western blot assay marketed by the same company (LDBio Diagnostics). All samples with IgG values in the gray zone and 14 negative samples according to the Architect system were positive with the ICT test (Table 1).

To better explore the sensitivities of the ICT and Architect IgG system, we selected a panel of 92 sera with IgG titers close to the threshold (Table 4). The analysis of these results clearly shows a higher sensitivity of the ICT than the Architect IgG technique. This observation confirms that the IgG cutoffs chosen by Abbott are too high and stringent. Figure 1 highlights the gain in sensitivity of the ICT test, which deciphers the equivocal results of the Architect system. These results are in line with those of Chapey

TABLE 6 Comparison of results between Toxo IgM Architect and ICT from 50 sera (corresponding to 24 women) with seroconversion and 33 sera with nonspecific IgM^a

No. of sera tested	Architect IgM (AU/ml; range)	ICT (% positive)	IgM reference technique [Platelia or ISAgA (range)]	Architect IgG (IU/ml)	Conclusion
Proved seroconversion (<i>n</i> = 50)		98			
1 (1 M)	0.34	Neg	Neg; 1 ISAgA (5)	0	Beginning of IgM increasing
2 (2 M)	0.39–0.39	Pos	Pos; 2 ISAgA (9–9)	<0.2	Beginning of IgM increasing
14 (2 M, 12 SE)	0.78–9.04	Pos	Pos; 2 ISAgA (10–12); 12 Platelia (2.84–9.87)	<1.5	IgM positive (all techniques)
33 (19 M, 14SE)	0.55–20.61	Pos	Pos; 19 ISAgA (9–12); 14 Platelia (1.18–9.38)	≥1.5	Proved seroconversion
No IgG seroconversion (<i>n</i> = 33), no specific IgM		6			
8 (8 SE)	0.30–0.49	Neg	Neg; 8 Platelia (0.10–0.78)	<0.2	Absence of IgM
2 (2 SE)	0.5–0.59	Neg	Neg; 2 Platelia (0.45–0.65)	<0.2	Absence of IgM
10 (4 M, 6 SE)	0.63–1.08	Neg	Neg; 4 ISAgA (0–2); 6 Platelia (0.23–0.78)	<0.2	False-positive IgM Architect
11 (7 M, 4 SE)	0.76–1.80	Neg	Pos; 7 ISAgA (9–11); 4 Platelia (1.30–2.17)	<0.2	False-positive IgM Architect and reference technique
2 (2 SE)	0.69–3.96	Pos	Pos; 2 Platelia (1.23–2.58)	1, 0.2	False-positive Architect, ICT, and reference technique

^aThe following cutoffs were used: Architect IgM (Abbott Diagnostics, Wiesbaden, Germany), 0.5 to 0.6 AU/ml; Platelia Toxo IgM (Bio-Rad, Marne la Coquette, France), 0.8 to 1.0; ISAgA IgM (bioMérieux, Marcy l'Etoile, France), 6 to 9.

et al. (8) and Villard et al. (6), who have worked on a similar but smaller panel (21 and 35 low-IgG-titer sera, respectively). However, our results show the limits of Chapey's study, which resulted in 8/400 false-positive ICT tests. For these specific cases, only a sensitive enough confirmatory technique (Western blot or dye test) would lead to a reliable conclusion.

Overall, with this sensitive rapid test (IgG), it is possible to avoid performing large numbers of unnecessary serological follow-up tests, thus providing significant economic savings. Concerning test specificity, false-positive IgG results on the Architect system in routine practice are rare (10), which is due to the use of *T. gondii* recombinant P30 (SAG-1) and P35 (GRA-8) antigens in the Architect *Toxoplasma* IgG assay. False-positive results often are associated with nonspecific anti-p35 antibodies. Systematic confirmation of positive IgG results is critical in parasitology laboratories. Using the ICT test as a second-line diagnostic approach enables the identification of sera that yield a false-positive result on the Architect system. Indeed, all 11 false-positive IgG selected sera were negative according to the ICT test and confirmed via *Toxoplasma* IgGII Western blot analysis, which showed different types of profiles (0, 1, or 2 bands) but did not correspond to serological profiles of toxoplasmosis.

False-positive IgG results with the ICT test occurred at low frequency (specificity of 98.7%), and discrepancies with the Architect system corresponded to negative IgG values (7/767) or a gray-zone IgG value (1/767) and need complementary expertise with reference techniques. Therefore, in rare cases of discrepant results between the Architect and ICT tests, we systematically confirmed the results using the *Toxoplasma* IgGII Western blot assay.

Concerning IgM detection, the ICT test displayed outstanding sensitivity and specificity. This technique was similar to the Architect system in cases of seroconversion. Furthermore, for two cases the ICT test exhibited better sensitivity and specificity for IgM than the Architect technique (Table 6). The ICT test is very useful for detecting IgM without IgG, and as has been mentioned, it also can distinguish nonspecific IgM (i.e., natural IgM) from specific *Toxoplasma* IgM.

To conclude, the *Toxoplasma* ICT IgG-IgM rapid test appears quite reliable. ICT sensitivity for IgG is equivalent to that of the reference technique (*Toxoplasma* IgGII Western blot analysis). ICT is also less expensive than the Western blot approach. The cost of ICT and automated techniques analyses (IgG and IgM) is on the order of \$10, whereas it is about \$30 for the IgGII Western blot. The ICT test can detect Architect

TABLE 7 Distribution of 235 selected sera (panel 2) based on the results obtained using the Architect assay and corresponding to different critical serological status

Panel	Serological status	Result with Architect and confirmatory techniques	No. of sera
2.1	Equivocal IgG titers (0.6–3.4 IU/ml)	No IgM	92
2.2	Serological reactivation	High IgG titers of >200 IU/ml	20
2.3	Seroconversion, acute toxoplasmosis	IgG of ≤ 1.5 IU/ml	17
2.4	False-positive IgM	IgG of >1.5 IU/ml	33
		IgM between 0.35 and 0.59 AU/ml	10
		IgM between 0.60 and 3.96 AU/ml	23
2.5	False-positive IgG	Negative IgGII Western blot	11
2.6	Crossing reactions with CMV, EBV, HIV, RF	Negative serology	29

false-positive IgG results and is capable of differentiating between nonspecific IgM and specific IgM in the majority of cases. In contrast, the reference immunocapture techniques (ISAgA and Platelia) often yielded false-positive results. In cases of positive IgM results, the serological follow-up 2 weeks apart is always performed to detect the possible appearance of IgG (11). Combined with the automated technique, the ICT test enables reliable orientation of the diagnostic results.

Lastly, the ICT IgG-IgM test could be promoted as a first-line technique in developing countries and could be particularly interesting for the early follow-up of pregnant women. This indication was also mentioned by Begeman et al. (9). Indeed, the simple determination of the serological status in early pregnancy compared to the serological status at birth seems simple and feasible in developing countries. This would allow to follow-up of the discrepant cases by targeting the cases requiring expertise, by limiting the cost of follow-up, and by increasing accessibility to screening through the simplicity of this technique (i.e., point of care). Obviously, in cases of serological status transition from negative to positive, these patients should be secondarily monitored in a reference laboratory.

MATERIALS AND METHODS

Serological samples. The study was carried out with 1,002 serum samples from patients tested in the laboratories of the University Hospital of Saint-Etienne (Saint-Etienne, France) and the La Timone University Hospital (Marseille, France). Two panels were performed.

Panel 1 included 767 nonselected prospective sera, 356 from Marseille and 411 from Saint-Etienne. Panel 2 (Table 7) was comprised of 235 sera selected based on results obtained using the Architect system; the serum characteristics were the following. For panel 2.1, 92 sera were derived from patients with low IgG titers (0.6 to 3.4 IU/ml) without IgM. For panel 2.2, 20 sera with high IgG titers, which correspond to serological reactivation, were used. For panel 2.3, 50 samples from 24 patients with acute toxoplasmosis and seroconversion, which then showed IgG appearance, were used. For all of these patients, a previous serum test was negative for IgG and IgM. Seventeen sera had an IgG titer of ≤ 1.5 IU/ml, and 33 sera had an IgG titer of >1.5 IU/ml (Table 7). For panel 2.4, we used 33 sera from 33 patients with IgM values between 0.35 and 3.96 AU/ml without IgG and not succeeded by seroconversion. Systematic IgG follow-up was done 2 weeks and 1 month later and did not show IgG appearance or increasing of IgM (data not shown). These samples corresponded to 23 false-positive results (IgM of >0.6 AU/ml), 2 equivocal IgM results ($0.5 < \text{IgM} < 0.6$ AU/ml), and 8 negative results close to the threshold ($0.3 < \text{IgM} < 0.5$ AU/ml) using the Architect system. For panel 2.5, 11 false-positive IgG sera from 11 patients not confirmed by the reference technique IgGII Western blot were used. These cases are rare (10) and have been identified thanks to IgG complementary techniques (negative hemagglutination or negative indirect immunofluorescence) and Western blot analysis. The estimated frequency of these cases is 1/10,000. For panel 2.6, we used 29 sera with the absence of IgG and IgM anti-*Toxoplasma gondii* (reference techniques) and with positive serology (IgG and/or IgM) for cytomegalovirus (CMV), Epstein-Barr virus (EBV), and HIV and the presence of rheumatoid factor (RF), which could interfere with the specificity of the ICT *Toxoplasma* IgG and IgM tests in the assessment of patients with unrelated antibodies.

Methods and serological diagnostics panel. In the serological diagnostics panel, diagnosis of *T. gondii* infection was previously established by testing the samples using routine techniques for the detection of anti-*T. gondii* IgG and IgM antibodies. In the two centers, IgG and IgM antibodies were determined using Architect Toxo IgG and IgM assays (Abbott Diagnostics, Wiesbaden, Germany) and, if necessary, an IgM immunosorbent agglutination assay (Toxo ISAgA IgM [bioMérieux, Mercy l'Etoile,

France] or Platelia Toxo IgM [Bio-Rad, Marne la Coquette, France]) and LDBio Toxo II IgG immunoblotting (LDBio, Lyon, France).

Architect *Toxoplasma* IgG and IgM (Abbott) is a screening method for the serological diagnosis of toxoplasmosis used at the hospitals in Saint-Etienne and Marseille. The routinely used assay is based on chemiluminescent microparticle immunoassay (CMIA) technology (10). Specimens with IgG concentrations of ≥ 3.0 IU/ml are considered reactive for IgG antibodies to *T. gondii*, concentrations ranging from 1.6 to 2.9 IU/ml are considered the gray zone, and concentrations of < 1.6 IU/ml are considered nonreactive. For IgM, reactive results were defined as index values of ≥ 0.60 AU/ml, gray-zone values ranged from 0.50 to 0.60 AU/ml, and nonreactive results were defined as index values of < 0.50 AU/ml.

Reference techniques for IgM detection used in each center. To confirm specificity for IgM, the following routine reference tests were used: the Platelia Toxo IgM (Bio-Rad, Marne la Coquette, France) in Saint-Etienne and the IgM immunosorbent agglutination assay (ISAgA; bioMérieux, Marcy l'Etoile, France) in Marseille. Both techniques, which are based on the immunocapture principle, can be considered reference tests due to their reliability and high sensitivity (12, 13). Platelia Toxo IgM has 97.9% sensitivity and 92.6% specificity and shows 97.6% concordance with ISAgA (6). In cases where IgM was detected without IgG, the secondary detection of IgG was monitored for serum classification (panel 2.3 or 2.4) (reference technique). Serological follow-up revealed the nature of IgM: specific IgM associated with IgG seroconversion, nonspecific IgM, or natural IgM.

Platelia Toxo IgM (Bio-Rad) is a qualitative test used for the detection of IgM antibodies against *T. gondii* via capture of IgM in the solid phase (the microplate wells are coated with anti-human μ chains). A mixture of antigens and the monoclonal anti-*T. gondii* antibody labeled with peroxidase is used as the conjugate. In the present study, values ranging from > 0.8 to < 1 were considered equivocal, while values of ≥ 1 were considered positive according to the manufacturer's instructions.

ISAgA-IgM (bioMérieux) is based on the agglutination of *Toxoplasma* antigens by specific IgM antibodies in the patient serum. The technique is based on a combination of two methods: direct agglutination and enzyme-linked immunosorbent assay (ELISA). The commercial kit Toxo-ISAgA (bioMérieux, Marcy l'Etoile, France) uses the IgM monoclonal antibody linked to the solid phase. In the absence of specific antibodies, *Toxoplasma* precipitates into wells. Agglutination occurs for positive reactions.

Assays were conducted as recommended by the manufacturers, and the cutoffs for the interpretations of the serologic values for adults are the following: positive, ≥ 9 ; negative, ≤ 6 ; equivocal, between 6 and 9. Those for infants are the following: positive, ≥ 3 ; negative, < 3 .

Reference technique for IgG. The LDBio *Toxoplasma* IgGII Western blot test is an immunoenzymatic test that involves immunoblotting to nitrocellulose strips. After standardization, incubation with sera, and fixation of specific IgG onto the band, the anti-*Toxoplasma* IgG bound to the strip then is detected using an alkaline phosphatase-conjugated antibody and specific substrate. The resulting bands on the patient strip correspond to 30, 31, 33, 40, and 45 kDa (Fig. 2). A positive result is defined by the presence of at least three matching bands on the patient strip, including the specific band at 30 kDa.

The sensitivity and specificity of the Architect Toxo IgG and ICT tests were evaluated by comparing the results with those of the LDBio *Toxoplasma* IgGII Western blot assay, which is used as a reference confirmatory test for low titers of anti-*Toxoplasma* IgG (14).

New technique: *Toxoplasma* ICT IgG-IgM (LDBio Diagnostics). Sera were tested by a rapid test in a second step. The test is comprised of a cassette with a nitrocellulose strip with a test band (*T. gondii* antigens) and a control band (rabbit gamma globulins), as well as a fiberglass support (conjugate pad) impregnated with red latex particles coupled to the toxoplasmic antigen (latex test, or T) and blue latex particles coupled with an anti-rabbit IgG goat antiserum (latex control, or C). The test consists of successively depositing a sample of serum (30 μ l) and an eluting solution (3 drops of eluent) in the well provided for this purpose. The reading must be done between 20 and 30 min. The test was considered positive when 2 lines, T and C, appear in the corresponding areas.

ACKNOWLEDGMENTS

We are grateful to Sandra Moore for English language editing and the LDBio Diagnostics team for technical assistance.

This study was sponsored by LDBio Diagnostics. LDBio Diagnostics participated in the study design but did not interfere with the analysis or conclusions reported.

Toxoplasma gondii antigen as well as diverse parasite and fungal antigens sold to LDBio Diagnostics are produced at the institution where C. Mahinc, C. Guillerme, H. Raberin, and P. Flori are currently employed.

REFERENCES

1. Montoya J, Liesenfeld O. 2004. Toxoplasmosis. *Lancet* 363:1965–1976. [https://doi.org/10.1016/S0140-6736\(04\)16412-X](https://doi.org/10.1016/S0140-6736(04)16412-X).
2. Roberts A, Hedman K, Luyasu V, Zufferey J, Bessières MH, Blatz RM, Candolfi E, Decoster A, Enders G, Gross U, Guy E, Hayde M, Ho-Yen D, Johnson J, Lécolier B, Naessens A, Pelloux H, Thulliez P, Petersen E. 2001. Multicenter evaluation of strategies for serodiagnosis of primary infection with *Toxoplasma gondii*. *Eur J Clin Microbiol Infect Dis Off Publ Eur Soc Clin Microbiol* 20:467–474. <https://doi.org/10.1007/PL00011289>.
3. Remington JS, Thulliez P, Montoya JG. 2004. Recent developments for diagnosis of toxoplasmosis. *J Clin Microbiol* 42:941–945. <https://doi.org/10.1128/JCM.42.3.941-945.2004>.
4. Sensini A. 2006. *Toxoplasma gondii* infection in pregnancy: opportuni-

- ties and pitfalls of serological diagnosis. *Clin Microbiol Infect Off Publ Eur Soc Clin Microbiol Infect Dis* 12:504–512.
5. Nogareda F, Le Strat Y, Villena I, De Valk H, Goulet V. 2014. Incidence and prevalence of *Toxoplasma gondii* infection in women in France, 1980–2020: model-based estimation. *Epidemiol Infect* 142:1661–1670. <https://doi.org/10.1017/S0950268813002756>.
 6. Villard O, Cimon B, L'Ollivier C, Fricker-Hidalgo H, Godineau N, Houze S, Paris L, Pelloux H, Villena I, Candolfi E. 2016. Help in the choice of automated or semiautomated immunoassays for serological diagnosis of toxoplasmosis: evaluation of nine immunoassays by the French National Reference Center for Toxoplasmosis. *J Clin Microbiol* 54:3034–3042. <https://doi.org/10.1128/JCM.01193-16>.
 7. Maudry A, Chene G, Chatelain R, Patural H, Belleste B, Tisseur B, Hafid J, Raberin H, Beretta S, Tran Manh Sung R, Belot G, Flori P. 2009. Bicentric evaluation of six anti-toxoplasma immunoglobulin G (IgG) automated immunoassays and comparison to the Toxo II IgG Western blot. *Clin Vaccine Immunol* 16:1322–1326. <https://doi.org/10.1128/CVI.00128-09>.
 8. Chapey E, Wallon M, Peyron F. 2017. Evaluation of the LDBIO point of care test for the combined detection of toxoplasmic IgG and IgM. *Clin Chim Acta Int J Clin Chem* 464:200–201. <https://doi.org/10.1016/j.cca.2016.10.023>.
 9. Begeman IJ, Lykins J, Zhou Y, Lai BS, Levigne P, El Bissati K, Boyer K, Withers S, Clouser F, Noble AG, Rabiah P, Swisher CN, Heydemann PT, Contopoulos-Ioannidis DG, Montoya JG, Maldonado Y, Ramirez R, Press C, Stillwaggon E, Peyron F, McLeod R. 2017. Point-of-care testing for *Toxoplasma gondii* IgG/IgM using Toxoplasma ICT IgG-IgM test with sera from the United States and implications for developing countries. *PLoS Negl Trop Dis* 11:e0005670. <https://doi.org/10.1371/journal.pntd.0005670>.
 10. Gay-Andrieu F, Fricker-Hidalgo H, Sickinger E, Espern A, Brenier-Pinchart M-P, Braun H-B, Pelloux H. 2009. Comparative evaluation of the ARCHITECT Toxo IgG, IgM, and IgG Avidity assays for anti-Toxoplasma antibodies detection in pregnant women sera. *Diagn Microbiol Infect Dis* 65:279–287. <https://doi.org/10.1016/j.diagmicrobio.2009.07.013>.
 11. Villard O, Cimon B, L'Ollivier C, Fricker-Hidalgo H, Godineau N, Houze S, Paris L, Pelloux H, Villena I, Candolfi E. 2016. Serological diagnosis of *Toxoplasma gondii* infection: recommendations from the French National Reference Center for Toxoplasmosis. *Diagn Microbiol Infect Dis* 84:22–33. <https://doi.org/10.1016/j.diagmicrobio.2015.09.009>.
 12. Meek B, van Gool T, Gilis H, Peek R. 2001. Dissecting the IgM antibody response during the acute and latent phase of toxoplasmosis. *Diagn Microbiol Infect Dis* 41:131–137. [https://doi.org/10.1016/S0732-8893\(01\)00291-7](https://doi.org/10.1016/S0732-8893(01)00291-7).
 13. Liesenfeld O, Press C, Montoya JG, Gill R, Isaac-Renton JL, Hedman K, Remington JS. 1997. False-positive results in immunoglobulin M (IgM) toxoplasma antibody tests and importance of confirmatory testing: the Platelia Toxo IgM test. *J Clin Microbiol* 35:174–178.
 14. Franck J, Garin YJ-F, Dumon H. 2008. LDBio-Toxo II immunoglobulin G Western blot confirmatory test for anti-*Toxoplasma* antibody detection. *J Clin Microbiol* 46:2334–2338. <https://doi.org/10.1128/JCM.00182-08>.