

ERRATA

In the article below, there are several corrections.

Magn Reson Med Sci 2017; 16; 209–216

Comparison of Silent and Conventional MR Imaging for the Evaluation of Myelination in Children

Chisato Matsuo-Hagiya, Yoshiyuki Watanabe, Hisashi Tanaka, Hiroto Takahashi, Atsuko Arisawa, Eri Yoshioka, Shin Nabatame, Sayaka Nakano, and Noriyuki Tomiyama

In the column “NEX” in Table 1, the data in 1st, 2nd, 5th, and 6th lines are wrong.

Wrong

Table 1. Sequence parameters

Corrected ages (weeks)	Sequence	Imaging plane	ST (mm)	TR (ms)	TE (ms)	FOV (cm)	NEX	Matrix	Time (min:s)
>44	Silent T ₁ W	sagittal	1	880	1.600E ⁻²	24	1	240*240	5:10
	Silent T ₂ W	axial	5	5700	105–109	22	1	320*320	2:13
	SE T ₁ W	axial	4–5	500	10	22	1	256*224	3:16
	FSE T ₂ W	coronal	3	6500	100–102	22	1	512*320	3:09
<44	Silent T ₁ W	sagittal	0.8	911	1.600E ⁻²	20	1	250*250	5:45
	Silent T ₂ W	axial	4–5	7600–8000	109–142	18–22	1	320*320	2:32
	SE T ₁ W	axial	4	500	10	18–20	1	256*256	3:42
	FSE T ₂ W	coronal	3–4	5800–7000	102–128	18–20	1	512*320	2:55

ST, slice thickness; TR, repetition time; TE, echo time; FOV, field of view; NEX, number of excitations; T₁W, T₁-weighted; T₂W, T₂-weighted.

Correct

Table 1. Sequence parameters

Corrected ages (weeks)	Sequence	Imaging plane	ST (mm)	TR (ms)	TE (ms)	FOV (cm)	NEX	Matrix	Time (min:s)
>44	Silent T ₁ W	sagittal	1	880	1.600E ⁻²	24	1.5	240*240	5:10
	Silent T ₂ W	axial	5	5700	105–109	22	1.8	320*320	2:13
	SE T ₁ W	axial	4–5	500	10	22	1	256*224	3:16
	FSE T ₂ W	coronal	3	6500	100–102	22	1	512*320	3:09
<44	Silent T ₁ W	sagittal	0.8	911	1.600E ⁻²	20	1.5	250*250	5:45
	Silent T ₂ W	axial	4–5	7600–8000	109–142	18–22	2	320*320	2:32
	SE T ₁ W	axial	4	500	10	18–20	1	256*256	3:42
	FSE T ₂ W	coronal	3–4	5800–7000	102–128	18–20	1	512*320	2:55

ST, slice thickness; TR, repetition time; TE, echo time; FOV, field of view; NEX, number of excitations; T₁W, T₁-weighted; T₂W, T₂-weighted.