



HHS Public Access

Author manuscript

Proc Int Soc Magn Reson Med Sci Meet Exhib Int Soc Magn Reson Med Sci Meet Exhib.

Author manuscript; available in PMC 2018 April 09.

Published in final edited form as:

Proc Int Soc Magn Reson Med Sci Meet Exhib Int Soc Magn Reson Med Sci Meet Exhib. 2017 April ; 25: 3944–.

Accurate High-speed 3D-Registration of EPI vNavs for Head Motion Correction

Yingzhuo Zhang¹, Iman Aganj^{2,3}, André J.W. van der Kouwe^{2,3}, and Matthew Dylan Tisdall⁴

¹John A. Paulson School of Engineering and Applied Sciences, Harvard University, Cambridge, MA, United States

²Athinoula A. Martinos Center for Biomedical Imaging, Radiology Department, Massachusetts General Hospital, Charlestown, MA, United States

³Department of Radiology, Harvard Medical School, Boston, MA, United States

⁴Department of Radiology, Perelman School of Medicine, University of Pennsylvania, Philadelphia, PA, United States

Synopsis

Low-resolution, whole-head volumes can be acquired rapidly with EPI-based volumetric navigators (vNavs). vNavs interspersed in a longer scan are widely used for prospective motion correction in a variety of sequences. To further improve the accuracy and flexibility of vNavs, we present a novel registration algorithm, tailored specifically for the vNavs application. Accuracy of the algorithm is tested on navigator volumes acquired with human volunteers at three isotropic resolutions, 6.4mm, 8mm, and 10mm, using a series of field of view (FOV) rotations and translations to provide ground truth rigid “motion”.

Purpose

Volumetric navigators (vNavs)^{1,2}, based on 3D-encoded gradient echo EPI, are widely used prospective motion correction systems for neuroimaging, particularly in pediatric imaging. Previous applications of vNavs used the PACE³ registration algorithm, designed for higher-resolution fMRI volumes, constraining vNav design choices. In this work, we present a novel registration algorithm, achieving both the accuracy and efficiency requirements of prospective motion correction with lower-resolution, and thus faster, vNavs than PACE allowed.

Method

Imaging was performed on a 3 T TIM Trio (Siemens Healthcare, Erlangen, Germany), using the body coil to ensure spatial uniformity. Three human volunteers, having given informed consent, were scanned at three isotropic resolutions, 6.4 mm, 8 mm, and 10 mm (see acquisition parameters in Table 1).

Motion was simulated by a programmed series of FOV rotations and translations. “Small rotations” data contained a volume at iso-center and on-axis (the “reference”), followed by volumes with rotations of 0.5°–5° at 0.5° increments and translations of 1–5 mm at 1 mm

increments. “Large rotations” data was acquired on two of the volunteers with rotations of 5° – 25° at 5° increments. Rotations were repeated around the x-, y-, and z-axes paired with z-direction translations, and around axes oriented along xy, xz, and yz paired with xy-direction translations. Scans were acquired in <30 s sets, during which the volunteers were instructed to hold their breath and remain still. 432 volumes per resolution per subject were acquired, from which 420 pairs (comparing each set’s “reference” with subsequent volumes) were extracted.

The volumes were masked before registration with a spherical mask (radius = image width/2 – 1 pixels) in the Fourier domain. Corners of the mask were Hann windowed to reduce ringing artifacts.

Our algorithm registers a *new* volume $V_N(\mathbf{x})$ to a given *reference* volume $V_R(\mathbf{x})$ by minimizing the residual

$$r(\mathbf{p}) = \| \{ W(T(\mathbf{p}, \mathbf{x})) [V_N(\mathbf{x}) - V_R(T(\mathbf{p}, \mathbf{x}))] \mid \mathbf{x} \in X \} \|_2,$$

where $T(\mathbf{p}, \mathbf{x})$ is a rigid transformation parameterized by $\mathbf{p} \in \mathbf{R}^6$. $W(\mathbf{x})$ is the same mask function initially applied in the Fourier domain, and X is the set of points inside our image FOV. We transform the reference to the new volume, interpolating only the reference volume. We can then precompute the coefficients of the cubic B-spline interpolator, accelerating each interpolation step^{4,5}.

We performed Gauss-Newton minimization, giving the update equation

$$\frac{dr(\mathbf{p})}{d\mathbf{p}}^T \frac{dr(\mathbf{p})}{d\mathbf{p}} \delta\mathbf{p} = \frac{dr(\mathbf{p})}{d\mathbf{p}}^T r(\mathbf{p})$$

We update our estimated \mathbf{p} by composing its transformation with that given by $\delta\mathbf{p}$. A line search scales $\delta\mathbf{p}$, ensuring the residual decreases at each step. Minimization stopped when updates move all points in a 100 mm radius of the image center less than 0.05 mm.

We evaluate the gradient with the approximation

$$\frac{dr(\mathbf{p})}{d\mathbf{p}} \approx -W(T(\mathbf{p}, \mathbf{x})) \frac{dV_N(\mathbf{x})}{d\mathbf{x}} \mathbf{M}(\mathbf{x}),$$

where

$$M(x) = \begin{bmatrix} 1 & 0 & 0 & 0 & x_z & -x_y \\ 0 & 1 & 0 & -x_z & 0 & -x_x \\ 0 & 0 & 1 & x_y & -x_x & 0 \end{bmatrix}$$

and

$$x = \{x_x, x_y, x_z\}.$$

We computed the “residual error transformation” between the estimated transformation and the known true transformation of each reference/new pair. We computed the RMS displacement that the “error transformation” would produce in a 100 mm sphere centered at isocenter⁶.

Results and Discussion

In all conditions RMS error was within 1 mm, with the exception of large rotations for one subject at 8mm resolution (Detailed results in Figures 2 and 3). We compared vNav resolutions using the Wilcoxon signed-rank test on the RMS errors (Results in Table 2). While 6.4 mm resolution produced the highest accuracy, assuming our motion tracking goal is ~1 mm, our new algorithm achieved functionally equivalent accuracy with 10 mm vNavs, requiring less than half of the acquisition time of 6.4 mm vNavs.

The algorithm is implemented in C++, using the Eigen⁷ library and is available on Github⁸. Median registration times for small rotations, running single-threaded on a 2.30 GHz Intel Xeon, were 12.97 ms, 25.10 ms, and 54.53, ms for 10 mm, 8 mm, and 6.4 mm vNavs respectively. Large rotations took 18.30 ms, 36.46 ms, and 88.47 ms, respectively.

To ensure that involuntary motion did not dominate registration error, we performed the same tests using a pineapple. Median RMS error in human subjects were 0.18, 0.14, 0.12, and in pineapple were 0.23, 0.21, 0.19, at 10 mm, 8mm and 6.4 mm resolutions, respectively.

Conclusions

We have developed and validated novel registration algorithm for vNavs that achieves fast and highly accurate motion tracking on a wider range of vNav protocols than was previously possible. Our algorithm supports motion correction with 10 mm vNavs, reducing navigator durations compared to previous designs, and increasing the number of sequences in which vNavs can be inserted.

Acknowledgments

This work was funded by NIH awards: R00HD074649, R21AG046657, R01HD071664, R21EB008547, P41RR014075, R01HD085813, K01DK101631, and the BrightFocus Foundation (A2016172S).

References

1. Tisdall, M Dylan, Hess, Aaron T., Reuter, Martin, Meintjes, Ernesta M., Fischl, Bruce, van der Kouwe, André JW. Volumetric navigators for prospective motion correction and selective reacquisition in neuroanatomical MRI. *Magnetic resonance in medicine*. 2012; 68(2):389–399. [PubMed: 22213578]
2. Hess, Aaron T., Dylan Tisdall, M., Andronesi, Ovidiu C., Meintjes, Ernesta M., van der Kouwe, André JW. Real-time motion and B0 corrected single voxel spectroscopy using volumetric navigators. *Magnetic resonance in medicine*. 2011; 66(2):314–323. [PubMed: 21381101]
3. Thesen, Stefan, Heid, Oliver, Mueller, Edgar, Schad, Lothar R. Prospective acquisition correction for head motion with image-based tracking for real-time fMRI. *Magnetic Resonance in Medicine*. 2000; 44(3):457–465. [PubMed: 10975899]
4. Lekien, Francois, Marsden, J. Tricubic interpolation in three dimensions. *International Journal for Numerical Methods in Engineering*. 2005; 63(3):455–471.
5. Zhang, Yingzhuo, Aganj, Iman, van der Kouwe, Andre J., Dylan Tisdall, M. Effects of Resolution and Registration Algorithm on the Accuracy of EPI vNavs for Real Time Head Motion Correction in MRI. *Proceedings of the IEEE Conference on Computer Vision and Pattern Recognition Workshops*. 2016:143–151.
6. Jenkinson, Mark. Studholme, C., Hill, DLG, Hawkes, DJ. 1999. Measuring transformation error by RMS deviation.
7. Guennebaud, G., Jacob, B., et al. Eigen v3. 2010. <http://eigen.tuxfamily.org>
8. Zhang, Yingzhuo. GitHub repository. <https://github.com/zyzdiana/MotionCorrection>




Figure 1.
Sagittal slices from vNavs at different resolutions.




Figure 2.

Distribution of RMS error displacement on three different subjects with small rotations at three vNav resolutions. Each violin represents results from 420 pairs of registration with rotations ranging from 0 to 5 degrees and translation from 1 to 5 mm. Median value is shown as a white point in each violin, and on the left side of each violin is the scatter plot of the values of RMS error.




Figure 3. Distribution of RMS error displacement on two different subjects with large rotations at three vNav resolutions. Each violin represents results from 420 pairs of registration with rotations ranging from 5 to 25 degrees and translation from 1 to 5 mm. Median value is shown as a white point in each violin, and on the left side of each violin is the scatter plot of the values of RMS error.

Table 1

Scanner Parameters for vNavs.

	6.4 mm	8 mm	10 mm
TR	15 ms	11 ms	10 ms
TE	6.7 ms	5.0 ms	4.1 ms
FA	3°	3°	3°
BW	4310 Hz/Px	4596 Hz/Px	4578 Hz/Px
FOV	256 mm	256 mm	260 mm
Total Scan Time	600 ms	352 ms	260 ms

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript

Table 2

Z-score computed from the Wilcoxon signed-rank test statistic for comparing registration accuracy across resolutions. The test was performed on each pair of vNav resolutions on small and large rotations. The test statistic is computed on the pairs of registration on all subjects combined. A negative score indicates that there is an improvement in the pair from the first resolution to the second, and a positive score indicates the opposite. All values in the table are z-scores that are significant for a one-sided difference at $p < 0.01$.

	8 mm to 6.4 mm	10 mm to 6.4 mm	10 mm to 8 mm
Small Rotations	-3.49	-10.38	-7.33
Large Rotations	-9.45	-6.85	3.77