

Magnetic resonance imaging of patients with airway stents

Yang Xia[#], Rui Jin[#], Wen Li, Huahao Shen

Key Laboratory of Respiratory Disease of Zhejiang Province, Department of Respiratory and Critical Care Medicine, Second Affiliated Hospital, Zhejiang University School of Medicine, Hangzhou 310009, China

Contributions: (I) Conception and design: Y Xia, H Shen; (II) Administrative support: H Shen; (III) Provision of study materials or patients: W Li; (IV) Collection and assembly of data: R Jin, W Li; (V) Data analysis and interpretation: R Jin, Y Xia; (VI) Manuscript writing: All authors; (VII) Final approval of manuscript: All authors.

[#]These authors contributed equally to this work.

Correspondence to: Huahao Shen, MD, FCCP; Wen Li, MD. Department of Respiratory and Critical Care Medicine, Second Affiliated Hospital, Zhejiang University School of Medicine, Hangzhou 310052, China. Email: huahaoshen@zju.edu.cn; liwen@zju.edu.cn.

Abstract: Airway stent implantation is a highly effective treatment for airway stenosis. However, it is presently unclear whether patients with airway stents can safely undergo magnetic resonance imaging (MRI). Such stents may be metallic or non-metallic, and MRI may induce stent dislodgment or heating and may be associated with stent-induced artifacts. We thoroughly reviewed the literature, experimental data, and manufacturer information on non-metallic, stainless steel (SS) and nickel-titanium alloy stents. Non-metallic stents are made of non-ferromagnetic materials associated with no MRI concerns. SS stents may shift in a magnetic field, generating significant artifacts. Nickel-titanium alloy stents are not at risk of dislodgement or heating, but may create some artifacts affecting image quality. Both non-metallic and nickel-titanium alloy stents are safe for patients who must undergo MRI. However, the safety of SS stents depends on the type of steel used.

Keywords: Magnetic resonance imaging (MRI); airway stents; safety; stainless steel stents (SS stents); nickel-titanium alloy stents

Submitted May 09, 2018. Accepted for publication Sep 20, 2018.

doi: 10.21037/jtd.2018.09.132

View this article at: <http://dx.doi.org/10.21037/jtd.2018.09.132>

Introduction

Airway stenosis is a severe life-threatening condition in both adults and children, caused by a variety of conditions, and is divided into benign and malignant strictures. Common causes of benign airway stenosis include iatrogenic factors, inflammation, infection, dynamic collapse, and other miscellaneous causes. Malignant causes include lung cancer, a salivary gland tumor, thyroid cancer, and metastatic diseases presenting as extrinsic airway compression by a tumor, an intrinsic airway tumor, or a combination of these (1,2).

Airway stent implantation is a safe and highly effective treatment for both malignant and benign tracheobronchial stenosis (3-8), promptly palliating dyspnea, reducing the risk for asphyxia (5,9) and enhancing survival (10-12). Airway

stents are generally divided into metallic and non-metallic stents. As an example of the latter, the Dumon stent first introduced by Dr. Jean-François Dumon in the 1990s (13) remains the most widely used (7,14). Non-metallic stents are made of non-ferromagnetic materials and thus do not affect MRI. However, safety issues associated with MRI of patients with metallic stents remain of concern. It is unclear whether such patients can safely undergo MRI. Hence, in this review, we focus on the safety of metallic and hybrid stents (covered metallic stents). We performed a structured PubMed search using the key phrase “stents and magnetic resonance imaging.” We screened the literature on MRI safety issues associated with metallic tracheobronchial stents. We also included manufacturer data.

Classification of airway stents

Stent materials are of prime concern in terms of their safety when undergoing MRI. Metallic stents can be divided into balloon-expandable and self-expandable stents; the latter include uncovered and partially and fully membrane-covered stents (of which the last two are hybrid stents). Metallic stents consist of stainless steel (SS) and nickel-titanium memory alloy stents. The former includes the Palmaz mesh, Gianturco-Z, Sigma, and Dynamic (Y-type) stents, and the latter include the Wallstent mesh, Ultraflex knitting, Silmet, Leufen, and Micro Tech stents. As SS is ferromagnetic, SS stents may interact with a magnetic field.

Apart from traditional stents, drug-eluting and three-dimensional (3D)-printed stents afford more options for patients with airway stenosis. Drug-eluting stents feature a bare metallic stent, drugs, and a carrier, and can inhibit restenosis (15). In contrast, 3D-printed airway stents, which were first introduced in 2014 by a team from the University of Girona, Spain, are made of silicone (16). Currently, most materials used to make 3D-printed stents are non-magnetic, and include silicone and polymers (17). However, more sophisticated printing technologies using alternative materials are under investigation. Notably, it remains unknown whether these emerging stents compromise the safety of MRI.

Effects of MRI on metallic airway stents

During MRI, metallic implants may interact with the magnetic field, harming both the patient and the imaging device. When encountering a magnetic field, metallic objects become magnetized and thus are affected by a displacement force and torque, triggering dislodgment (18,19). In addition, the field gradient and radiofrequency (RF) pulsing raise the temperatures of metallic materials and their environments (19-21). The RF wavelength, type of RF transmission coil used, and specific absorption rate (SAR) of stent material affect the temperature increase. MRI-associated heating varies markedly by RF (22). Furthermore, it should be noted that when describing the SAR during MRI, the whole-body-averaged SAR for each scan sequence is required (23). Finally, magnetic artifacts are principally attributable to the destruction of static and gradient magnetic fields. Severe artifacts hamper image quality (19). When labeling artifacts in the context of MRI safety, artifact extents should be described (18).

MR safety ratings of medical devices

In 1997, the Food and Drug Administration (FDA) Center for Devices and Radiological Health first proposed that the terms “MR safe” and “MR compatible” should be used to label medical devices (23). Given the variety of MR systems and MR conditions in clinical use today, the standards for MR device safety were revised by the American Society for Testing Materials (ASTM) International in 2005, and are documented in the ASTM International F2503 guidelines. The US FDA requires all implants and medical devices to be labeled as “MR safe,” “MR unsafe,” or “MR conditional” (23,24). However, this revised terminology has not been retrospectively applied to the many implants and devices that previously received FDA approval using the terms “MR safe” or “MR compatible” (23). The website www.mrisafety.com provides exhaustive MR safety data for most medical devices. The information is regularly updated and is very accessible. Each device is categorized as “Safe,” “Conditional,” or “Unsafe.” The details are summarized in *Table 1*.

MR safety of airway stents

SS stents

SS stents were the first clinical stents on the market. The vast majority are made of 316L SS, an austenitic containing 16–18% chromium, 10–14% nickel, 2.0–3.0% molybdenum, and no more than 0.03% carbon, 2% manganese, 0.75% silicon, 0.045% phosphorus, 0.03% sulfur, and 0.1% nitrogen by weight; the remainder is iron (25). 316L SS was recognized in the 1960s by the ASTM Medical and Surgical Materials and Devices Committee as the standard material for fabrication of surgical implants and has found many applications in clinical settings (26). SS implants meeting the International Organization for Standardization (ISO) 5832-1 standard exhibit stable austenitic structures and are completely nonmagnetic (27). These implants will not move or become heated during MRI, although artifacts compromising MRI clarity may arise (27,28). However, austenitic SSs may become martensitic after machining and thus exhibit strong magnetism. Hence, stents made of martensitic SSs will be affected by MRI.

It is accepted that patients with ferromagnetic implants should not undergo MRI unless both the deflection angle and the extent of dislodgment are extremely small (29). Taal (30) used an *ex vivo* Perspex device to show that the Gianturco esophageal stent (an SS stent) was attracted

Table 1 MR safety ratings of medical devices. Information is summarized from <http://www.mrisafety.com> (reproduced from Dr. Frank G. Shellock with permission)

Status	Definition
Safe	The device poses no known hazards in all MR imaging environments. MR safe items are nonconducting, nonmetallic, and nonmagnetic items (23)
Conditional	The device has been demonstrated to pose no known hazards in a specified MRI environment with specified conditions of use. Field conditions include static magnetic field strength, spatial gradient, dB/dt (time varying magnetic fields), radio frequency (RF) fields, and specific absorption rate (SAR). Specified conditions require attention, that is, a medical device that is MR Conditional in a 1.5T scanner may not be safe to scan in an MR system with a higher or lower field strength (18)
Conditional 1	The device is acceptable for the patient in the MRI environment, even though it showed positive findings for magnetic field interactions during testing. The device is considered to be “weakly” ferromagnetic
Conditional 2	The devices (these particular “weakly” ferromagnetic coils, filters, stents, clips, cardiac occluders, or other implants) are usually impossible to be moved or displaced during MRI testing because they typically become firmly incorporated into the tissue six weeks following placement. Furthermore, there has been no report of an injury to a patient or individual in association with an MRI procedure for these devices
Conditional 3	Certain transdermal patches with metallic foil or other metallic components, although not attracted to an MRI system, have been reported to heat excessively during MRI procedures. It is recommended that the patch be removed prior to the MRI procedure and a new patch be applied immediately after the examination
Conditional 4	Halo vest or cervical fixation device may have ferromagnetic component parts, MRI-related heating may exist, however, the magnetic field interactions have not been determined. Nevertheless, there has been no report of patient injury. As such, guidelines provided in the Product Instructions should be carefully followed
Conditional 5	This device is acceptable for a patient undergoing an MRI procedure only if specific guidelines or recommendations are followed
Conditional 6	A patient with this device can be scanned safely immediately after placement in a static magnetic field of 3-Tesla or less, spatial gradient magnetic field of 720-Gauss/cm or less and with a maxima whole-body-averaged specific absorption rate (SAR) of 2-W/kg for 15 minutes of scanning
Conditional 7	This device is not intended for use during an MRI procedure. That is, this device should not be inside of the bore of the MRI system and exposed to the time-varying and RF fields activated during an MRI procedure
Conditional 8	This information only pertains to a device that has MRI labeling at 1.5- and 3-Tesla
Unsafe	The device is known to pose hazards in all MRI environments
Unsafe 1	The device is considered to pose a potential or realistic risk or hazard to an individual in the MRI environment primarily due to movement or displacement. This object is considered to be a contraindication for an MRI procedure
Unsafe 2	This object displays minor magnetic field interactions and is unlikely to pose a hazard or risk in association with movement or displacement. Nevertheless, potential risks of currents, excessive heating, or other potentially hazardous conditions may exist. This object is considered to be a contraindication for an MRI procedure

to, and exhibited torque in, a magnetic field; however, direct evidence of dislodgment was lacking. Although no study has yet addressed the Gianturco tracheobronchial stent in terms of safety during MRI, the stent is made of the same material. Holton (31) used a suture to suspend an SS stent horizontally in a 4.1 T magnetic field. The angles of deflection from the vertical axis after alignment of the longitudinal axis were measured using a protractor. The 316L SS stent experienced a force of 5.2 μN at 4.1 T;

theoretically, this would not injure tissues. In another study, the magnetic forces on coronary stents fabricated of surgical SS were determined by measuring stent magnetic dipole moments employing the on-axis magnetic field profile of an MRI magnet (32). The maximum force was 0.18 μN , thus even smaller than the gravitational force on the stent; no physiological impact would be expected. In general, SS stents are unlikely to undergo displacement. However, extreme caution should be exercised during MRI.

In addition, many studies have found that heating of SS stents in a magnetic field does not endanger patients. Wang (33) measured heating of metal implants in a 1.5 T MRI field; SS stents were heated to greater extents than titanium alloy stents. However, the temperatures of adjacent tissues increased by less than 1.0 °C; this was acceptable. Ho (34) used a computer model to estimate RF-mediated heating of tissues surrounding metallic implants during MRI. A metallic cylinder with 16 wires was implanted in the heart region of a realistic human and the maximum SAR was calculated. The temperature did not increase appreciably. Notably, the wire material differed from that of metallic stents. Hence, the results are not immediately applicable to all metallic implants.

Many studies have confirmed that SS stents induce significant artifacts during MRI. In one study, stents were embedded in 1% (w/v) gadolinium-doped agar, 3D phase images were obtained in a 1.5 T field, and susceptibility and shielding factors were determined; it was apparent that magnetization of a Palmaz stent would completely destroy the signal (35). In addition, Holton (31) showed that a 316 low-carbon SS stent caused dramatic *ex vivo* signal loss in a 4.1 T field. In another study, 18 stents made of different materials were placed in an environment simulating brain blood vessels and were studied in terms of 3 T MRI artifacts; SS stents were associated with severe artifacts (36). Adams (37) placed two SS stents in cadaveric femoral arteries subjected to 1.5 T MRI before and after stent deployment; the stent images were seriously distorted. Heinrich (38) suspended stents between two wooden sticks and immersed them in copper sulfate (CuSO₄) solution within 1.5 and 3 T magnetic fields. The maximum artifact width in terms of stent diameter was measured using MR-Susceptibility Artifact Measurement (SAM) software. The maximum SS stent signal losses beyond the diameter were 6 mm [turbo spin echo (TSE) sequence] and 10 mm [gradient echo (GRE) sequence]. Burg (39) evaluated the *in vitro* lumen visibility of 22 different peripheral arterial stents (iliac, renal, and carotid) via magnetic resonance angiography (MRA). Stents were imaged at 1.5 T and lumen visibility was classified on a three-point scale (good, intermediate, and poor). Two SS stents were graded intermediate and five were poor. Another MRA assessment of coronary stents reached a comparable conclusion: SS stents were associated with the most pronounced intraluminal signal voids (40).

Nickel-titanium alloy stents

Nickel-titanium alloy, also termed nitinol, is a shape

memory alloy with outstanding plasticity and excellent biocompatibility, wear-resistance, and corrosion-resistance, much superior to those of SS stents (41). Nitinol is non-ferromagnetic, exhibiting no dislodgment and only slight heating during MRI. Shellock (19) studied a nitinol airway stent [Elgiloy; 40% cobalt, 20% chromium, 15% nickel, 16% iron, 7% molybdenum, and 2% manganese (all w/w)] at 0° deflection under no torque. The temperature increased by 0.1 °C in a 1.5 T magnetic field, and no dislodgement, rotation, or tissue damage were noted. Holton (31) suspended a 54/46 nickel-titanium stent *ex vivo* in a 4.1 T magnetic field and observed no deflection.

Nitinol-associated artifacts during MRI are minimal. Wang (35) calculated the susceptibility and shielding factors of a nitinol stent in an *in vitro* magnetic field; signal loss was negligible. Similarly, Holton (31) observed only a slight increase in the signal void of an *ex vivo* nickel-titanium stent in a 4.1 T magnetic field as the echo time increased. In an *ex vivo* MRI assessment of 27 stents, the maximum signal loss beyond the actual diameter was less than 1 mm (TSE) and was 4 mm (GRE) for nitinol stents, thus much lower than for SS stents (38). MRA assessments have come to similar conclusions (39,40).

MR safety information for common metallic stents

Clinical stents are produced by a variety of manufacturers using different materials, and the MRI safety ratings differ. We integrated data from www.mrisafety.com and manufacturers' instructions to introduce the safety recommendation (Table 2).

Conclusions

In general, MRI safety varies by stent material. Non-metallic stents are non-ferromagnetic and thus completely safe. SS stents that meet the ISO 5832-1 standard are completely nonmagnetic and thus safe, although they produce artifacts. Magnetic SS stents may shift in the magnetic field and generate significant artifacts, affecting image quality. Thus, patients with such SS stents should not undergo MRI unless the risks are weighed and the manufacturer's instructions are carefully followed. Nitinol stents cannot be dislodged, any heating effect is minimal, and no severe artifacts affecting image quality are created. Routine MRI is usually safe for patients with such stents.

Table 2 MR safety information of commonly used metallic stents. Information is summarized from <http://www.mrisafety.com> (reproduced from Dr. Frank G. Shellock with permission) and the official websites of stent companies

Manufacturer	Object	Material	MRI Status	Field strength
Johnson & Johnson Interventional (USA)	Palmaz stent	Stainless steel	Undetermined	–
Cook Medical (USA)	Cook-Z stent, Gianturco-Rosch, Tracheobronchial Design	Stainless steel	Conditional 5	1.5
Huaian Sigma Medical (China)	Sigma stent	Silicone, stainless steel	Undetermined	–
Boston Scientific (USA)	Dynamic (Y) stent, Bifurcated Tracheobronchial stent	Stainless steel	Unsafe	–
Merit Medical (USA)	AERO Tracheobronchial stent	Nitinol	Conditional 6	3
Schneider (USA)	Tracheobronchial Wallstent, Endoprosthesis, 14×80 Tracheobronchial Wallstent, Endoprosthesis, 24×70	Elgiloy	Safe	1.5
Boston Scientific (USA)	Ultraflex Tracheobronchial stent	Nitinol	Conditional 5	1.5, 3
Novatech (France)	SILMET stent	Nitinol	Undetermined	–
Leufen Medical (Germany)	Leufen stent	Nitinol	Undetermined	–
MICRO-TECH (China)	MICRO-TECH Tracheal stents/Bronchial stents MICRO-TECH Y-Tracheal stents MICRO-TECH J-Tracheal stents MICRO-TECH Bronchial Stump Fistula-occluding stents	Nitinol	Conditional 8	1.5, 3

Acknowledgements

We thank Dr. Frank G. Shellock for the kind advice to the paper.

Funding: This work was supported by the National Natural Science Foundation of China (81500012, 81870022), the Zhejiang Provincial Natural Science Foundation of China (LQ16H010001), the Medical and Health Technology Program of Zhejiang Province (2015111464, 2017204226) and the Program of Zhejiang Province Health High-level Personnel [2017] to Yang Xia.

Footnote

Conflicts of Interest: The authors have no conflicts of interest to declare.

References

1. Puchalski J, Musani AI. Tracheobronchial stenosis: causes and advances in management. *Clin Chest Med* 2013;34:557-67.
2. Li YQ, Li Q, Bai C, et al. Causes of benign central airway stenoses and the efficacy of interventional treatments through flexible bronchoscopy. *Zhonghua Jie He He Hu Xi Za Zhi* 2008;31:364-8.
3. Murgu SD, Egressy K, Laxmanan B, et al. Central Airway Obstruction: Benign Strictures, Tracheobronchomalacia, and Malignancy-related Obstruction. *Chest* 2016;150:426-41.
4. Schmidt B, Olze H, Borges AC, et al. Endotracheal balloon dilatation and stent implantation in benign stenoses. *Ann Thorac Surg* 2001;71:1630-4.
5. Ernst A, Feller-Kopman D, Becker HD, et al. Central airway obstruction. *Am J Respir Crit Care Med* 2004;169:1278-97.
6. Sonett JR, Keenan RJ, Ferson PF, et al. Endobronchial management of benign, malignant, and lung transplantation airway stenoses. *Ann Thorac Surg* 1995;59:1417-22.
7. Wood DE, Liu YH, Vallieres E, et al. Airway stenting for malignant and benign tracheobronchial stenosis. *Ann*

- Thorac Surg 2003;76:167-72; discussion 173-4.
8. Husain SA, Finch D, Ahmed M, et al. Long-term follow-up of ultraflex metallic stents in benign and malignant central airway obstruction. *Ann Thorac Surg* 2007;83:1251-6.
9. Tanigawa N, Sawada S, Okuda Y, et al. Symptomatic improvement in dyspnea following tracheobronchial metallic stenting for malignant airway obstruction. *Acta Radiol* 2000;41:425-8.
10. Yerushalmi R, Fenig E, Shitrit D, et al. Endobronchial stent for malignant airway obstructions. *Isr Med Assoc J* 2006;8:615-7.
11. McGrath EE, Warriner D, Anderson P. The insertion of self expanding metal stents with flexible bronchoscopy under sedation for malignant tracheobronchial stenosis: a single-center retrospective analysis. *Arch Bronconeumol* 2012;48:43-8.
12. Okiror L, Jiang L, Oswald N, et al. Bronchoscopic management of patients with symptomatic airway stenosis and prognostic factors for survival. *Ann Thorac Surg* 2015;99:1725-30.
13. Dumon JF. A dedicated tracheobronchial stent. *Chest* 1990;97:328-32.
14. Dutau H, Breen D, Bugalho A, et al. Current Practice of Airway Stenting in the Adult Population in Europe: A Survey of the European Association of Bronchology and Interventional Pulmonology (EABIP). *Respiration* 2018;95:44-54.
15. Dutau H, Musani AI, Laroumagne S, et al. Biodegradable Airway Stents - Bench to Bedside: A Comprehensive Review. *Respiration* 2015;90:512-21.
16. Vallicrosa G, Serenó L, Melgoza EL, et al. Rapid tooling using 3D printing system for manufacturing of customized tracheal stent. *Rapid Prototyping J* 2014;20:2-12.
17. Freitag L, Gordes M, Zarogoulidis P, et al. Towards Individualized Tracheobronchial Stents: Technical, Practical and Legal Considerations. *Respiration* 2017;94:442-56.
18. Administration USFAD. Establishing Safety and Compatibility of Passive Implants in the Magnetic Resonance (MR) Environment. 2014.
19. Shellock FG, Shellock VJ. Metallic stents: evaluation of MR imaging safety. *AJR Am J Roentgenol* 1999;173:543-7.
20. ASTM International (ASTM). ASTM F2182-11a, Standard Test Method for Measurement of Radio Frequency Induced Heating On or Near Passive Implants During Magnetic Resonance Imaging. West Conshohocken, PA, 2011.
21. Shellock FG. Magnetic resonance safety update 2002: implants and devices. *J Magn Reson Imaging* 2002;16:485-96.
22. Shellock FG, Crues JV. MRI bioeffects, safety, and patient management. Biomedical Research Publishing Group, 2014.
23. Shellock FG, Woods TO, Crues JV 3rd. MR labeling information for implants and devices: explanation of terminology. *Radiology* 2009;253:26-30.
24. Tsai LL, Grant AK, Mortelet KJ, et al. A Practical Guide to MR Imaging Safety: What Radiologists Need to Know. *Radiographics* 2015;35:1722-37.
25. Disegi JA, Zardiackas LD. Metallurgical and Mechanical Evaluation of 316L Stainless Steel Orthopaedic Cable. *Stainless Steels for Medical and Surgical Applications*. ASTM International, 2003.
26. Wu T, Chen X, Fan D, et al. Development and application of metal materials in terms of vascular stents. *Biomed Mater Eng* 2015;25:435-41.
27. Disegi JA, Eschbach L. Stainless steel in bone surgery. *Injury* 2000;31 Suppl 4:2-6.
28. Disegi JA. Magnetic resonance imaging of AO/ASIF stainless steel and titanium implants. *Injury* 1992;23 Suppl 2:S1-4.
29. Shellock FG. Reference manual for magnetic resonance safety, implants, and devices: 2006 edition. Los Angeles: Biomedical Research Publishing Group, 2006.
30. Taal BG, Muller SH, Boot H, et al. Potential risks and artifacts of magnetic resonance imaging of self-expandable esophageal stents. *Gastrointest Endosc* 1997;46:424-9.
31. Holton A, Walsh E, Anayiotos A, et al. Comparative MRI compatibility of 316 L stainless steel alloy and nickel-titanium alloy stents. *J Cardiovasc Magn Reson* 2002;4:423-30.
32. Lopič N, Jelen A, Vrtnik S, et al. Quantitative determination of magnetic force on a coronary stent in MRI. *J Magn Reson Imaging* 2013;37:391-7.
33. Wang CB, Zhang L, Li DP. Influence of the Security of Thermal Effects in 1.5T MR Imaging of Metal Implants. *China Medical Devices* 2015:27-9.
34. Ho HS. Safety of metallic implants in magnetic resonance imaging. *J Magn Reson Imaging* 2001;14:472-7.
35. Wang Y, Truong TN, Yen C, et al. Quantitative evaluation of susceptibility and shielding effects of nitinol, platinum, cobalt-alloy, and stainless steel stents. *Magn Reson Med* 2003;49:972-6.
36. Reinhardt J, Nguyen-Trong TH, Hahnel S, et al. Magnetic resonance imaging of stents: quantitative in vitro

- examination at 3 Tesla. *Z Med Phys* 2009;19:278-87.
37. Adams GJ, Baltazar U, Karmonik C, et al. Comparison of 15 different stents in superficial femoral arteries by high resolution MRI ex vivo and in vivo. *J Magn Reson Imaging* 2005;22:125-35.
38. Heinrich A, Guttler FV, Schlesies F, et al. In vitro stent assessment by MRI: visibility of lumen and artifacts for 27 modern stents. *Biomed Tech (Berl)* 2017;62:565-73.
39. Burg MC, Bunck AC, Seifarth H, et al. MR Angiography of Peripheral Arterial Stents: In Vitro Evaluation of 22 Different Stent Types. *Radiol Res Pract* 2011;2011:478175.
40. Maintz D, Botnar RM, Fischbach R, et al. Coronary magnetic resonance angiography for assessment of the stent lumen: a phantom study. *J Cardiovasc Magn Reson* 2002;4:359-67.
41. Miyazawa T, Yamakido M, Ikeda S, et al. Implantation of ultraflex nitinol stents in malignant tracheobronchial stenoses. *Chest* 2000;118:959-65.

Cite this article as: Xia Y, Jin R, Li W, Shen H. Magnetic resonance imaging of patients with airway stents. *J Thorac Dis* 2018;10(10):5939-5945. doi: 10.21037/jtd.2018.09.132