



Single-Port Trans-Perineal Approach to Cystoprostatectomy with Intracorporeal Ileal Conduit Urinary Diversion and Lymph-Nodes Dissection using a Purpose-Built Robotic System: Surgical Steps in a Preclinical Model

Juan Garisto¹, Riccardo Bertolo¹, Eddie Chan², Jihad Kaouk¹

¹ Glickman Urological & Kidney Institute, Cleveland Clinic, Cleveland OH, USA; ² Chinese University of Hong Kong and Division of Urology at Prince of Wales Hospital, Shatin, Hong Kong

ABSTRACT

Aim: To report the technique for single-port trans-perineal cystoprostatectomy with intracorporeal ileal conduit urinary diversion and lymph nodes dissection using a purpose-built robotic platform (da Vinci SP 1098, Intuitive Surgical, Sunny-vale, CA, USA).

Materials and Methods: In a male cadaver the SP1098 robotic system was used to perform cystoprostatectomy with intracorporeal ileal conduit urinary diversion and lymph nodes dissection by single-port trans-perineal approach. The surgery was completed through a 2.5-cm perineal incision through which a GelPOINT Mini advanced access platform (Applied Medical, Rancho Santa Margarita, CA, USA) and a dedicated 25-mm multichannel port accommodating a 12 x 10-mm oval articulating robotic camera, three 6-mm double-jointed articulating robotic instruments and a 6-mm accessory laparoscopic instrument were positioned. At the planned level of the stoma for the ileal conduit, a 12-mm port was placed through which the EndoGIA® stapler was used to mature the urinary diversion

Results: The total operative time was 185 min. The procedure was successfully completed without the need for additional ports placement. The benefits of the trans-perineal approach, particularly in longer procedures as radical cystectomy with intracorporeal urinary diversion, might include the avoided need of Trendelenburg position, with undoubtful advantages for the patient and the anesthesiologist in terms of respiratory mechanics and hemodynamics.

Conclusions: The feasibility of single-port trans-perineal cystoprostatectomy with intracorporeal ileal conduit urinary diversion and lymph nodes dissection using the SP1098 purpose-built robotic platform is demonstrated. The duplication of the described surgical steps in the clinical model is awaited when the platform will be available on the market.

CONFLICT OF INTEREST

None declared.

ARTICLE INFO

D Riccardo Bertolo

http://orcid.org/0000-0003-0260-4601

Available at: http://www.intbrazjurol.com.br/video-section/20180524_Garisto_et_al Int Braz J Urol. 2019; 45 (Video #15): 854-5

Submitted for publication: July 30, 2018

Accepted after revision: November 17, 2018

Published as Ahead of Print: November 17, 2018 Correspondence address: Jihad H. Kaouk, MD Glickman Urology and Kidney Institute, Cleveland Clinic 9500 Euclid Ave, Q10 Cleveland, OH 44195, USA Telephone: +1 216 444-2976 E-mail: kaoukj@ccf.org