

Published in final edited form as:

Processes (Basel). ; 8(5): . doi:10.3390/pr8050580.

Ultrasensitive $Ti_3C_2T_x$ MXene/Chitosan Nanocomposite-Based Amperometric Biosensor for Detection of Potential Prostate Cancer Marker in Urine Samples

Stefania Hroncekova¹, Tomas Bertok¹, Michal Hires¹, Eduard Jane¹, Lenka Lorencova¹, Alica Vikartovska¹, Aisha Tanvir², Peter Kasak², Jan Tkac^{1,*}

Stefania Hroncekova: stefania.hroncekova@savba.sk; Tomas Bertok: tomas.bertok@savba.sk; Michal Hires: michal.hires@savba.sk; Eduard Jane: eduard.jane@savba.sk; Lenka Lorencova: lenka.lorencova@savba.sk; Alica Vikartovska: alica.vikartovska@savba.sk; Aisha Tanvir: atanvir@qu.edu.qa; Peter Kasak: peter.kasak@qu.edu.qa

¹Institute of Chemistry, Slovak Academy of Sciences, Dubravská cesta 9, Bratislava 845 38, Slovak Republic

²Center for Advanced Materials, Qatar University, P. O. BOX 2713, Doha, Qatar

Abstract

Two-dimensional layered nanomaterial $Ti_3C_2T_x$ (a member of the MXene family) was used to immobilise enzyme sarcosine oxidase to fabricate a nanostructured biosensor. The device was applied for detection of sarcosine, a potential prostate cancer biomarker, in urine for the first time. The morphology and structures of MXene have been characterised by atomic force microscopy (AFM) and scanning electron microscopy (SEM). Electrochemical measurements, SEM and AFM analysis revealed that MXene interfaced with chitosan is an excellent support for enzyme immobilisation to fabricate a sensitive biosensor exhibiting a low detection limit of 18 nM and a linear range up to 7.8 μ M. The proposed biosensing method also provides a short response time of 2 s and high recovery index of 102.6% for detection of sarcosine spiked into urine sample in a clinically relevant range.

Keywords

MXene; nanocomposite; biosensor; sarcosine; sarcosine oxidase; prostate cancer

1 Introduction

Early stage diagnostics of prostate cancer (PCa) and therapy monitoring of PCa patients is done by analysis of the level of prostate-specific antigen (PSA) in blood. PSA is a

Submitted for possible open access publication under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

*Correspondence: jan.tkac@savba.sk.

Author Contributions: Conceptualization, T.B. and J.T.; Methodology, S.H., T.B., M.H., L.L., A.T. and P.K.; Validation, T.B., P.K. and J.T.; Formal Analysis, T.B. and P.K.; Investigation, S.H., M.H., E.J., L.L. and A.T.; Resources, A.V., P.K. and J.T.; Data Curation, S.H., T.B. and E.J.; Writing-Original Draft Preparation, S.H., T.B. and J.T.; Writing-Review & Editing, S.H., T.B., M.H., E.J., L.L., A.V., A.T., P.K. and J.T.; Visualization, E.J.; Supervision, T.B. and J.T.; Project Administration, A.V.; Funding Acquisition, P.K. and J.T.

Conflicts of Interest: The authors declare no conflict of interest.

glycoprotein released into the blood stream by the prostate and the protein is responsible for seminal fluid liquefaction. Since PSA is organ-specific rather than cancer-specific, the protein is released into the blood stream not only as a result of PCa development and progression, but under other conditions such as prostate inflammation. Hence, PSA cannot be used as a reliable early stage diagnostic PCa biomarker because of the high false-positive/negative results provided [1, 2], with an AUC (Area Under Curve in Receiver Operating Curve) of 0.68 [1]. PCa is one of the most frequently diagnosed cancers, with the second highest mortality rate among men globally [3]. Thus, there is an urgent need to find new sensitive and specific biomarkers for PCa to reduce the number of unnecessary biopsies and to enhance PCa patient survival. One of the alternative PCa biomarkers is sarcosine, which is able to provide significant discrimination between PCa patients and non-cancer individuals, with an AUC of 0.833 [4]. A wide range of (bio)analytical approaches are available for determination of sarcosine, such as a colorimetric [5], enzymatic [6], electrochemical and/or spectrometric [7] and liquid chromatography with tandem mass spectrometry-based approaches [8]. All of these methods have drawbacks such as low sensitivity and/or specificity, time-consuming sample preparation and a need for having skilled persons to operate such instrumentation [9].

Two-dimensional (2D) nanomaterials, namely graphene [10], find numerous applications because of their high surface area and unique electronic/optical properties that differ significantly from their bulk counterparts [11]. Although graphene has attracted more attention than any other 2D nanomaterial so far, a simple interfacial chemistry limits some applications. 2D nanomaterials containing more than one element offer novel possibilities as they provide larger numbers of compositional variables, which can be finely tuned to achieve specific properties [12]. Since their appearance in 2011, 2D transition metal carbides/nitrides, known as MXenes, due to their metallic conductivity, large electrochemically active surface, excellent mechanical properties and water solubility, are extensively used in many applications such as batteries/super-capacitors, sensors and biosensors [13–22].

So far, only a limited number of MXenes-based biosensors have been developed [17]. A pristine or modified Ti_3C_2 MXene and haemoglobin-based device was developed for determination of H_2O_2 [23] or NaNO_2 [24]. MXene, with a hydrophilic surface and excellent metallic conductivity, is a prospective interface for effective docking of a wide range of biorecognition molecules for preparation of biosensors. Enzymatic MXene-based electrochemical biosensors have been recently prepared by immobilisation of glucose oxidase [25], acetylcholinesterase [26] and tyrosinase [27] for ultrasensitive and rapid detection of glucose or toxic compounds.

In this work, $\text{Ti}_3\text{C}_2\text{T}_x$ (where T represents terminating functional groups such as =O, -OH or -F) MXene was used as an enzyme immobilisation support to construct a biosensor for sensitive detection of a small molecule—sarcosine (*N*-methylglycine, a by-product of glycine synthesis and degradation, Figure 1), a potential PCa biomarker [28, 29]. Sarcosine is also present in urine, which is a less complex matrix compared to human serum. Quantification of sarcosine was achieved indirectly via amperometric detection of H_2O_2 produced during the enzymatic reaction. A sensitive nanostructured MXene-based biosensor

selectively detected sarcosine in urine in the concentration range of 0.02–5 μM , which is a clinically relevant range [30].

For the experimental section, see the Supplementary Materials.

2 Material and Methods

2.1 Materials

Sarcosine oxidase (SOx, enzyme commission number 1.5.3.1) from *Bacillus sp.* (lyophilised powder, 25–50 U mg^{-1}), sarcosine (98%), glutaraldehyde solution (50 wt. % in H_2O), chitosan (medium molecular weight), acetic acid, phosphate buffer (PB) components (KH_2PO_4 and K_2HPO_4) and SurineTM (a negative urine control for toxicology) were purchased from Sigma-Aldrich (St. Louis, MO, USA). Micropolish alumina powder (1.0 and 0.3 μm) and polishing pads for disc electrodes were purchased from Buehler (lake Bluff, IL, USA).

2.2 Preparation of MXene

MXene was prepared by in situ-formed hydrofluoric acid (HF) (reaction of lithium fluoride (LiF) with hydrochloric acid (HCl)) using a previously described protocol [32]. This MXene was previously shown to exhibit a good sensitivity towards H_2O_2 and some other small molecules as well.

2.3 Electrochemical Procedures

All electrochemical procedures were carried out on a laboratory potentiostat/galvanostat Autolab PGSTAT 302N with an impedimetric module (Ecochemie, Utrecht, Netherlands) with a GCE ($d = 3$ mm, Bioanalytical systems, West Lafayette, IN, USA) used as a working electrode. An Ag/AgCl/3 M KCl reference electrode and a counter platinum electrode (Bioanalytical systems, West Lafayette, IN, USA) were applied in a three-electrode cell system. Chronoamperometric detection of H_2O_2 (at different rpm) was performed on a rotating disc electrode employed as a working electrode (Metrohm Autolab, Utrecht, Netherlands). Cyclic voltammetry and chronoamperometry were performed in 0.1 M phosphate buffer (PB) pH 7.4 purged with nitrogen for 15 min prior to use to eliminate the interference from an oxygen reduction reaction. Cyclic voltammetry measurements were run in the potential window from 0.1 V (volt) to -1.0 V at a scan rate of 100 mV s^{-1} . Chronoamperometry measurements were run at a fixed potential value of -0.7 V. All electrochemical measurements were run under Nova Software 1.10 (Metrohm Autolab, Utrecht, Netherlands), and data acquired were evaluated using OriginPro 9.1 (OriginLab, Northampton, MA, USA).

2.4 Construction of MXene-Based Sarcosine Biosensor

Prior to modification, a GCE with 3 mm diameter was polished on a polishing cloth with 1.0 and 0.3 μm alumina powder, then washed with distilled water and sonicated in ethanol and water successively. The MXene-based biosensor for determination of sarcosine was prepared by a simple drop-casting method. Firstly, 200 μL chitosan solution (Chi, 0.1% in 0.3% acetic acid) and 40 μL MXene dispersion (3 mg mL^{-1} in distilled water) were mixed and shaken

overnight at a temperature of 20 °C and at a rotation speed of 1500 rpm. After preparation, 20 µL of the above-prepared mixture was cast onto the surface of polished GCEs. After drying, 10 µL of 5% or 1% glutaraldehyde solution was applied to the surface for 10 min. Construction and behaviour of the biosensor without using glutaraldehyde solution was also studied. Finally, 20 µL of the enzyme solution was cast onto the surface and allowed to dry at room temperature in a laminar box. Before electrochemical measurements, all the as-prepared electrodes were immersed in PB pH 7.4 to remove residual components. For the purpose of biosensor preparation and its performance, different configurations and nanocomposite preparation protocols were used (Supplementary Figure S1), i.e., (i) MXene+SOx (stock) mixed prior to drop-casting, (ii) MXene+SOx (10x diluted) mixed prior to drop-casting, (iii) layer-by-layer (LBL) drop-casting of MXene, SOx (10x diluted) and chitosan solution, (iv) MXene+SOx (10x diluted)+chitosan solution mixed prior to drop-casting, (v) all components mixed prior to drop-casting with SOx stock solution being desalted using Zeba Spin desalting columns (7k MWCO; Thermo, Waltham, MA, USA), alternatively MXene+chitosan mixed prior to drop-casting and subsequently activated using glutaraldehyde (GA) solutions (5%, 1% or no glutaraldehyde - GA) and incubated with desalted and diluted SOx. Among all these configurations, from the stability point of view (meaning reproducible performance, high sensitivity and no obvious dissolving of the prepared nanocomposite), only the configuration with mixed chitosan+MXene (overnight, 20 °C, 1500 rpm) prior to drop-casting, subsequently modified by LBL using desalted and diluted SOx with no GA activation, was chosen.

2.5 Contact Angle Measurements

Contact angle measurements were run on a portable instrument System E (Advex Instruments, Brno, Czech Republic) to obtain wetting angle and free surface energy for MXene-based interfaces. The droplet volume was 1 µL and the testing liquid was distilled water and diiodomethane. Free surface energy was determined using the two-liquid Owens–Wendt method, where the total surface energy γ consists of disperse γ^d and polar γ^p components. For each sample, the water and diiodomethane wetting angle was obtained as an average value of assays performed using 3 droplets.

2.6 Characterisation of MXene and MXene-Chitosan Composite

The morphology and structure of MXene and the as-prepared MXene-chitosan mixture were characterised by scanning electron microscopy (SEM) and atomic force microscopy (AFM). SEM images of MXene and the MXene-chitosan composite were recorded using a Nova NanoSEM 450 (FEI, Hillsboro, OR, USA) microscope, applying accelerating voltage. A peak force tapping mode atomic force microscopy (AFM, Scan Asyst, Bruker, Billerica, MA, USA) in air was carried out on a Bioscope Catalyst instrument and Olympus IX71 microscope in conjunction with NanoScope 8.15 software (Bruker, Billerica, MA, USA).

3 Results and Discussion

Five different configurations were tested during biosensor preparation (Supplementary Figure S1). Among all these configurations, from the stability point of view (meaning reproducible performance, high sensitivity and no obvious dissolving of the prepared

nanocomposite), only the configuration with mixed chitosan+MXene (overnight, 20 °C, 1500 rpm) prior to drop-casting, subsequently modified by LBL using desalted and diluted SOx with no glutaraldehyde activation, was chosen. The main reason behind that decision was the fact that such biosensor offered the highest response, as shown below.

3.1 Characterisation of MXene

Initial scanning electron microscopy-based characterisation of MXene showed the presence of MXene flakes with the size from a few μm to tens of μm (Supplementary Figure S2a–d). Atomic force microscopy showed the presence of MXene flakes with thickness of 20.2 ± 3.1 nm (Supplementary Figure S3c), which is in agreement with the literature [24, 27]. The thickness of the MXene-chitosan flake is shown in Supplementary Figure S3b. When MXene flakes were stored in water, after about a week, the MXene thickness decreased to 1 nm (Figure 2).

Wetting angle and surface energy calculation using the Owens–Wendt method for two liquids (water and diiodomethane) were applied to further investigate the surface modifications during individual steps of biosensor preparation. The hydrophilicity of bare rotating glassy carbon electrode (GCE) (90.0 ± 2.3)° increased upon modification of the electrode by MXene (71.9 ± 5.0)° and further by the MXene-chitosan composite (69.5 ± 3.6)°. After SOx adsorption, the biosensor surface becomes quite hydrophobic again (87.4 ± 4.9)° (Supplementary Figure S4).

3.2 Electrochemical Measurements

Since MXene undergoes irreversible anodic oxidation at +0.43 V [15, 16], cyclic voltammetry (CV) experiments were performed in a potential window from 0.1 V to -1.0 V in N_2 -purged PB pH 7.4. At this pH value, SOx has a maximum stability [33]. A direct comparison of the electrochemical activity towards 1.5 mM H_2O_2 on unmodified GCE and MXene-modified GCE is shown in Supplementary Figure S5, confirming significant redox activity of MXene-modified GCE towards H_2O_2 reduction. The effect of concentration of SOx applied for patterning of MXene-modified GCE was optimised in the range from 0.19 mg/mL to 0.77 mg/mL, with an optimal SOx concentration of 0.38 mg/mL identified.

To evaluate the bioelectrochemical activity of the biosensor and to construct the corresponding calibration curves, the final sarcosine concentration was in the range of 2.5–50 μM (Figure 3, left). Electrochemical reduction of H_2O_2 produced by catalytic decomposition of sarcosine by SOx present on the surface of the electrode started at -0.4 V, with the peak maximum present at -0.7 V. The cathodic peak for H_2O_2 reduction increased with increasing biomarker concentration. Calibration curves were constructed by plotting current response versus sarcosine concentration in the first CV scan with subtraction of a background CV.

Immobilisation of SOx on the modified electrode was optimised using various glutaraldehyde concentrations (Figure 3, left inset). Fitting of calibration curves using a Hill model and Michaelis-Menten kinetics renders the following maximal current outputs (I_{max}): (2.61 ± 0.15) μA (0% GA), (1.96 ± 0.26) μA (1% GA), and (1.42 ± 0.10) μA (5% GA). Thus, glutaraldehyde has a negative effect on the SOx activity, which can be explained by

the formation of a multi-point attachment of SO_x with the carrier, leading to the enzyme inactivation [34, 35]. Wang et al. [36] reported similar findings on deactivation of the enzyme by high concentrations of glutaraldehyde. One of the most important performance parameters of the biosensor is the limit of detection (LOD). The biosensor prepared without glutaraldehyde exhibits a limit of detection (LOD) (based on signal to noise ratio of 3) of 91.4 nM using CV. The biosensor devices constructed with the aid of glutaraldehyde exhibit a LOD of 79.9 nM (1% GA) or 110 nM (5% GA).

Chronoamperometry is a more sensitive electrochemical method compared to CV for indirect sarcosine detection. In order to obtain low LOD values, chronoamperometry on a rotating disc GCE at -0.7 V and 900 rpm was performed. Figure 3, right, shows a calibration plot of the SO_x/MXene-Chi/GCE biosensor with a linear range up to 7.83 μM ($R^2=0.998$, $n=20$) and LOD=18.0 nM ($S/N=3$). Moreover, the biosensor exhibits a very short response time of 2 s (Figure 3, right inset).

Our biosensor device belongs to the most sensitive MXene devices published so far, when considering LOD (Table 1). Additionally, our MXene-based sarcosine biosensor belongs to the most sensitive electrochemical biosensor devices for analysis of sarcosine, when comparing LOD (Table 2). For example the electrochemical SO_x-based biosensor offered the following LOD: 430 nM for SO_x immobilised on hollow magnetic Pt-Fe₃O₄@C nanospheres [37], 50 nM for organic transistor with grafted SO_x on platinum-plated anodized aluminium oxide gate electrodes [30] and 10 nM for immobilisation of SO_x nanoparticles onto gold electrode [38]. An optical-detection platform was also applied for preparation of SO_x-biosensors with LOD down to a low μM range achieved for fused protein of SO_x and fluorescent protein co-immobilised on silica particles with horseradish peroxidase [28, 39]. A hybrid type of the biosensor based on SO_x immobilised on a non-structured gold electrode modified with CdSe/ZnS quantum dots offered a LOD of 0.11 mM, while detecting a photocurrent [40]. It is important to remember that the above-mentioned optical biosensor cannot be reused and requires sample pre-treatment. The performance of other SO_x-based biosensors is summarised in a recent review paper [41]. Moreover, the MXene-biosensor developed here offers a really quick response towards the analyte, with a response time of 2 s, value equal to the best sarcosine biosensors [41].

3.3 Clinical Application of SO_x/MXene-chi/GCE Biosensor

In order to evaluate potential application of the biosensor for analysis of real samples, the SO_x/MXene-Chi/GCE device was applied for determination of sarcosine (0.1 to 0.5 μM) spiked into 10x diluted artificial urine (Surine™ negative urine control, Merck). The value of the obtained recovery index was 102.6%. This means that under such conditions, there were no interferences affecting the biosensor's performance. The results prove that the as-fabricated biosensor is reliable tool for sarcosine detection in urine samples.

4 Conclusions

In this paper, a MXene-based biosensing platform was successfully fabricated for indirect sensitive detection of sarcosine. The sarcosine biosensor displayed low detection limit of 18 nM and a linear range up to 7.8 μM for sarcosine, with correlation coefficient of 0.998 and a

response time of 2 s. The SO_x/MXene-Chi/GCE device was further used for sarcosine determination in artificial urine with a recovery index (defined as the concentration of the analyte measured by the assay divided by concentration of sarcosine spiked into artificial urine and expressed in %) of 102.6%. These excellent operational parameters make MXene a potential electrochemical biosensing platform for protein-based biosensors with application in biomedical detection and diagnostics. In future work, we would like to test the device for analysis of human urine samples from healthy individuals and patients with prostate cancer.

Supplementary Material

Refer to Web version on PubMed Central for supplementary material.

Funding

The financial support received from, from the Slovak Research and Development Agency APVV 17-0300 is acknowledged. We would like to acknowledge the support from the ERC Proof of Concept grant (No. 825586). This work was made possible by NPRP grant # 9 - 219-2-105 from the Qatar National Research Fund (A Member of The Qatar Foundation). The findings achieved herein are solely the responsibility of the authors. We would like to acknowledge the support received from the Ministry of Health of the Slovak Republic under the project registration number 2019/68-CHÚSAV-1.

References

1. Tkac J, Gajdosova V, Hroncekova S, Bertok T, Hires M, Jane E, Lorencova L, Kasak P. Prostate-specific antigen glycoprofiling as diagnostic and prognostic biomarker of prostate cancer. *Interface Focus*. 2019; 9(2)
2. Bertok T, Lorencova L, Hroncekova S, Gajdosova V, Jane E, Hires M, Kasak P, Kaman O, Sokol R, Bella V, Eckstein AA, et al. Advanced impedimetric biosensor configuration and assay protocol for glycoprofiling of a prostate oncomarker using Au nanoshells with a magnetic core. *Biosens Bioelectron*. 2019; 131:24–29. [PubMed: 30798249]
3. Damborský P, Damborská D, Belický Š, Tkáč J, Katrlík J. Sweet Strategies in Prostate Cancer Biomarker Research: Focus on a Prostate Specific Antigen. *BioNanoScience*. 2017; 8(2):690–700.
4. Markin PA, Brito A, Moskaleva N, Fodor M, Lartsova EV, Shpot YV, Lerner YV, Mikhajlov VY, Potoldykova NV, Enikeev DV. Plasma Sarcosine Measured by Gas Chromatography-Mass Spectrometry Distinguishes Prostatic Intraepithelial Neoplasia and Prostate Cancer from Benign Prostate Hyperplasia. *Lab Med*. 2020; doi: 10.1093/labmed/lmaa008
5. Mazzu-Nascimento T, Gomes Carneiro Leão PA, Catai JR, Morbioli GG, Carrilho E. Towards low-cost bioanalytical tools for sarcosine assays for cancer diagnostics. *Anal Methods*. 2016; 8(40):7312–7318.
6. Burton C, Gamagedara S, Ma Y. A novel enzymatic technique for determination of sarcosine in urine samples. *Anal Methods*. 2012; 4(1):141–146.
7. Cernei N, Zitka O, Ryvolova M, Adam V, Masarik M, Hubalek J, Kizek R. Spectrometric and Electrochemical Analysis of Sarcosine as a Potential Prostate Carcinoma Marker. *Int J Electrochem Sci*. 2012; 7:4286–4301.
8. Jiang Y, Cheng X, Wang C, Ma Y. Quantitative Determination of Sarcosine and Related Compounds in Urinary Samples by Liquid Chromatography with Tandem Mass Spectrometry. *Anal Chem*. 2010; 82:9022–9027. [PubMed: 20939533]
9. Narwal V, Kumar P, Joon P, Pundir CS. Fabrication of an amperometric sarcosine biosensor based on sarcosine oxidase/chitosan/CuNPs/c-MWCNT/Au electrode for detection of prostate cancer. *Enzyme Microb Technol*. 2018; 113:44–51. [PubMed: 29602386]
10. Naguib M, Mashtalir O, Carle J, Presser V, Lu J, Hultman L, Gogotsi Y, Barsoum MW. Two-Dimensional Transition Metal Carbides. *ACS Nano*. 2012; 6:1322–1331. [PubMed: 22279971]

11. Naguib M, Gogotsi Y. Synthesis of two-dimensional materials by selective extraction. *Acc Chem Res.* 2015; 48(1):128–35. [PubMed: 25489991]
12. Naguib M, Kurtoglu M, Presser V, Lu J, Niu J, Heon M, Hultman L, Gogotsi Y, Barsoum MW. Two-dimensional nanocrystals produced by exfoliation of Ti₃AlC₂. *Adv Mater.* 2011; 23(37):4248–53. [PubMed: 21861270]
13. Barsoum MW. The MN+1AXN Phases: A New Class of Solids; Thermodynamically Stable Nanolaminates. *Prog Solid State Chem.* 2000; 28:201–281.
14. Naguib M, Mochalin VN, Barsoum MW, Gogotsi Y. 25th anniversary article: MXenes: a new family of two-dimensional materials. *Adv Mater.* 2014; 26(7):992–1005. [PubMed: 24357390]
15. Lorencova L, Bertok T, Dosekova E, Holazova A, Paprckova D, Vikartovska A, Sasinkova V, Filip J, Kasak P, Jerigova M, Velic D, et al. Electrochemical performance of Ti₃C₂T_x MXene in aqueous media: towards ultrasensitive H₂O₂ sensing. *Electrochim Acta.* 2017; 235:471–479. [PubMed: 29109588]
16. Lorencova L, Bertok T, Filip J, Jerigova M, Velic D, Kasak P, Mahmoud KA, Tkac J. Highly stable Ti₃C₂T_x (MXene)/Pt nanoparticles-modified glassy carbon electrode for H₂O₂ and small molecules sensing applications. *Sens Actuat B: Chem.* 2018; 263:360–368.
17. Lorencova L, Gajdosova V, Hroncekova S, Bertok T, Blahutova J, Vikartovska A, Parrakova L, Gemeiner P, Kasak P, Tkac J. 2D MXenes as Perspective Immobilization Platforms for Design of Electrochemical Nanobiosensors. *Electroanalysis.* 2019; 31(10):1833–1844.
18. Dong Y, Wu ZS, Zheng S, Wang X, Qin J, Wang S, Shi X, Bao X. Ti₃C₂ MXene-Derived Sodium/Potassium Titanate Nanoribbons for High-Performance Sodium/Potassium Ion Batteries with Enhanced Capacities. *ACS Nano.* 2017; 11(5):4792–4800. [PubMed: 28460161]
19. Lukatskaya MR, Mashtalir O, Ren CE, Dall'Agnese Y, Rozier P, Taberna PL, Naguib M, Simon P, Barsoum MW, Gogotsi Y. Cation intercalation and high volumetric capacitance of two-dimensional titanium carbide. *Science.* 2013; 341(6153):1502–5. [PubMed: 24072919]
20. Xie X, Chen S, Ding W, Nie Y, Wei Z. An extraordinarily stable catalyst: Pt NPs supported on two-dimensional Ti₃C₂X₂ (X = OH, F) nanosheets for oxygen reduction reaction. *Chem Commun.* 2013; 49(86):10112–4.
21. Mashtalir O, Naguib M, Mochalin VN, Dall'Agnese Y, Heon M, Barsoum MW, Gogotsi Y. Intercalation and delamination of layered carbides and carbonitrides. *Nat Commun.* 2013; 4
22. Wang F, Yang C, Duan M, Tang Y, Zhu J. TiO₂ nanoparticle modified organ-like Ti₃C₂ MXene nanocomposite encapsulating hemoglobin for a mediator-free biosensor with excellent performances. *Biosens Bioelectron.* 2015; 74:1022–8. [PubMed: 26264270]
23. Wang F, Yang C, Duan C, Xiao D, Tang Y, Zhu J. An Organ-Like Titanium Carbide Material (MXene) with Multilayer Structure Encapsulating Hemoglobin for a Mediator-Free Biosensor. *J Electrochem Soc.* 2015; 162(1):B16–B21.
24. Liu H, Duan C, Yang C, Shen W, Wang F, Zhu Z. A novel nitrite biosensor based on the direct electrochemistry of hemoglobin immobilized on MXene-Ti₃C₂. *Sens Actuat B: Chem.* 2015; 218:60–66.
25. Rakhi RB, Nayak P, Xia C, Alshareef HN. Novel amperometric glucose biosensor based on MXene nanocomposite. *Sci Rep.* 2016; 6
26. Zhou L, Zhang X, Ma L, Gao J, Jiang Y. Acetylcholinesterase/chitosan-transition metal carbides nanocomposites-based biosensor for the organophosphate pesticides detection. *Biochem Eng J.* 2017; 128:243–249.
27. Wu L, Lu X, Dhanjai, Wu ZS, Dong Y, Wang X, Zheng S, Chen J. 2D transition metal carbide MXene as a robust biosensing platform for enzyme immobilization and ultrasensitive detection of phenol. *Biosens Bioelectron.* 2018; 107:69–75. [PubMed: 29448223]
28. Jornet-Martínez N, Henderson CJ, Campíns-Falcó P, Daly R, Hall EAH. Towards sarcosine determination in urine for prostatic carcinoma detection. *Sens Actuat B: Chem.* 2019; 287:380–389.
29. Cernei N, Heger Z, Gumulec J, Zitka O, Masarik M, Babula P, Eckschlager T, Stiborova M, Kizek R, Adam V. Sarcosine as a potential prostate cancer biomarker--a review. *Int J Mol Sci.* 2013; 14(7):13893–908. [PubMed: 23880848]

30. Hu J, Wei W, Ke S, Zeng X, Lin P. A novel and sensitive sarcosine biosensor based on organic electrochemical transistor. *Electrochim Acta*. 2019; 307:100–106.
31. Wagner MA, Trickey P, Chen Z-w, Mathews FS, Jorns MS. Monomeric Sarcosine Oxidase: 1. Flavin Reactivity and Active Site Binding Determinants. *Biochemistry*. 2000; 39(30):8813–8824. [PubMed: 10913292]
32. Alhabeb M, Maleski K, Anasori B, Lelyukh P, Clark L, Sin S, Gogotsi Y. Guidelines for Synthesis and Processing of Two-Dimensional Titanium Carbide (Ti₃C₂T_x MXene). *Chem Mater*. 2017; 29(18):7633–7644.
33. Matsuda Y, Hoshika H, Inouye Y, Ikuta S, Matsuura K, Nakamura S. Purification and Characterization of Sarcosine oxidase of *Bacillus* Origin. *Chem Pharm Bull*. 1987; 35:711–717.
34. Chen H, Zhang Q, Dang Y, Shu G. The Effect of Glutaraldehyde Cross-Linking on the Enzyme Activity of Immobilized β -Galactosidase on Chitosan Bead. *Advance J Food Sci Technol*. 2013; 5(7):932–935.
35. Ang LF, Por LY, Yam MF. Development of an amperometric-based glucose biosensor to measure the glucose content of fruit. *PLoS One*. 2015; 10(3):e0111859. [PubMed: 25789757]
36. Wang H-S, Pan Q-X, Wang G-X. A Biosensor Based on Immobilization of Horseradish Peroxidase in Chitosan Matrix Cross-linked with Glyoxal for Amperometric Determination of Hydrogen Peroxide. *Sensors*. 2005; 5:266–276.
37. Yang QG, Li N, Li Q, Chen SQ, Wang HL, Yang HP. Amperometric sarcosine biosensor based on hollow magnetic Pt-Fe₃O₄@C nanospheres. *Anal Chim Acta*. 2019; 1078:161–167. [PubMed: 31358215]
38. Kumar P, Narwal V, Jaiwal R, Pundir CS. Construction and application of amperometric sarcosine biosensor based on SOxNPs/AuE for determination of prostate cancer. *Biosens Bioelectron*. 2018; 122:140–146. [PubMed: 30248641]
39. Henderson CJ, Pumford E, Seevaratnam DJ, Daly R, Hall EAH. Gene to diagnostic: Self immobilizing protein for silica microparticle biosensor, modelled with sarcosine oxidase. *Biomaterials*. 2019; 193:58–70. [PubMed: 30562636]
40. Zhao S, Volkner J, Riedel M, Witte G, Yue Z, Lisdat F, Parak WJ. Multiplexed Readout of Enzymatic Reactions by Means of Laterally Resolved Illumination of Quantum Dot Electrodes. *ACS Appl Mater Interf*. 2019; 11(24):21830–21839.
41. Pundir CS, Deswal R, Kumar P. Quantitative analysis of sarcosine with special emphasis on biosensors: a review. *Biomarkers*. 2019; 24(5):415–422. [PubMed: 31050554]
42. Rebelo TS, Pereira CM, Sales MG, Noronha JP, Costa-Rodrigues J, Silva F, Fernandes MH. Sarcosine oxidase composite screen-printed electrode for sarcosine determination in biological samples. *Anal Chim Acta*. 2014; 850:26–32. [PubMed: 25441156]
43. Wang Q, Zhao Y, Yang Q, Du D, Yang H, Lin Y. Amperometric sarcosine biosensor with strong anti-interference capabilities based on mesoporous organic-inorganic hybrid materials. *Biosens Bioelectron*. 2019; 141
44. Liu T, Fu B, Chen J, Li K. An electrochemical sarcosine sensor based on biomimetic recognition. *Microchim Acta*. 2019; 186(3):136.
45. Yang H, Wang J, Yang C, Zhao X, Xie S, Ge Z. Nano Pt@ZIF8 Modified Electrode and Its Application to Detect Sarcosine. *J Electrochem Soc*. 2018; 165(5):H247–H250.




Figure 1.

A graphical presentation of a glassy carbon electrode (GCE) modified using a MXene/chitosan nanocomposite as a support for sarcosine oxidase (SOx) immobilisation and indirect sarcosine detection in urine, based on hydrogen peroxide electrochemical reduction. SOx structure is adapted from the Protein Data Bank (code 1EL5) with permission from Reference [31]. Copyright (2000) American Chemical Society.




Figure 2. Atomic force microscopy (AFM) images showing individual MXene sheets (1 nm) observed in MXene solution after a week of storage in an aqueous solution, accompanied with a slight change in colour, most likely due to progressive oxidation and delamination. To prevent this from affecting the measurements, freshly prepared MXene solutions were used for every experiment.




Figure 3.

Left image: cyclic voltammograms CVs of the prepared SO_x/MXene-Chi/GCE biosensor in 0.1 M PB pH 7.4 purged by N₂ containing 0 (a), 2.5 (b), 5 (c), 10 (d), 20 (e), 30 (f), 40 (g), 50 (h) µM sarcosine at a scan rate of 100 mV s⁻¹ in a potential window from -1.0 V to +0.1 V. Inset: Calibration plots obtained when using no glutaraldehyde (a, blue line), 1% glutaraldehyde (b, green line) and 5% glutaraldehyde (c, red line) during fabrication of the biosensor. Right image: Plot of current responses after subtraction of blank (current response obtained in the absence of sarcosine) of the SO_x/MXene-Chi/GCE biosensor measured on a rotating disc glassy carbon electrode at -0.7 V and 900 rpm versus sarcosine concentration. Inset: a current versus time response of the SO_x/MXene-Chi/GCE biosensor at -0.7 V to successive addition of stock sarcosine solution in stirred nitrogen-purged 0.1 M phosphate buffer, pH 7.4.

Table 1
Electrochemical biosensors based on MXene with integration of redox proteins/enzymes.

Enzyme/protein	MXene patterning	Immobilization	Analyte	LOD (nM)	Reference
haemoglobin (Hb)	MXene	Hb glued <i>via</i> Nf	H ₂ O ₂	20	[23]
haemoglobin (Hb)	MXene	Hb glued <i>via</i> Nf	NO ₂ ⁻	120	[24]
haemoglobin (Hb)	TiO ₂ on MXene	Hb glued <i>via</i> Nf	H ₂ O ₂	14	[22]
tyrosinase (Tyr)	MXene	Tyr glued <i>via</i> Chi	phenol	12	[27]
glucose oxidase	AuNPs on MXene	GOx adsorbed on Nf-AuNP/MXene	glucose	5900	[25]
SOx	MXene	SOx glued <i>via</i> Chi	sarcosine	18	This work

Abbreviations: LOD—limit of detection; Nf—nafion; Chi—chitosan; NPs—nanoparticles.

Table 2
Analytical performance of SO_x-based electrochemical biosensors.

Interface	LOD (nM)	Working range (nM)	Reference
platinum-plated anodised aluminium oxide electrode	50	50–100,000	[30]
Hybrid: Pt nanoparticles with hollow Fe ₃ O ₄ nanospheres	430	500–60,000	[37]
Modified SPCE	16	10–100	[42]
SO _x NPs/AuE	10	100–100,000	[38]
SO _x /Pt–Fe ₃ O ₄ @C nanocomposite/GCE	430	500–60,000	[37]
Pt-supported organic/inorganic hybrid mesoporous NPs	130	1000–70,000	[43]
Riboflavin/AuPt-PPy/graphene-chitosan-modified GCE	680	2500–600,000	[44]
nanoPt@porous zeolitic imidazolate framework-8	1060	5000–30,000	[45]
SO _x /MXene-Chi/GCE	18	36–7800	This work

Abbreviations: SPCE – screen printed carbon electrode; PPy – polypyrrole; AuE – gold electrode; Chi - chitosan