

CORRESPONDENCE

Open Access

SARS-CoV-2 detection with CRISPR diagnostics

Lu Guo^{1,2,3}, Xuehan Sun^{1,2,3}, Xinge Wang^{1,2,3}, Chen Liang^{1,2,3}, Haiping Jiang^{1,2,3}, Qingqin Gao^{1,2,3}, Moyu Dai^{1,2,3}, Bin Qu^{1,2,3}, Sen Fang^{1,2,3}, Yihuan Mao^{1,2,3}, Yangcan Chen^{1,2,3}, Guihai Feng^{1,2}, Qi Gu^{1,2}, Ruiqi Rachel Wang^{1,2}, Qi Zhou^{1,2,3} and Wei Li^{1,2,3}

Dear Editor,

The novel coronavirus (CoV) disease termed COVID-19 (coronavirus disease-19) caused by SARS-CoV-2 (severe acute respiratory syndrome coronavirus-2)¹ is causing a massive pandemic worldwide, threatening public health systems across the globe. During this ongoing COVID-19 outbreak, nucleic-acid detection has played an important role in early diagnosis². To date, four protocols based on CRISPR for detecting SARS-CoV-2 have been published^{3–6}. Using lateral flow protocols, RNA samples harboring more than 1×10^4 – 1×10^5 copies/mL (SHERLOCK) or 1×10^4 copies/mL (DETECTR) can be detected within 1 hour. In addition to these reported efforts, we have also established a SARS-CoV-2 detection protocol based on our previously reported platform—CDetection (Cas12b-mediated DNA detection)⁷. By combining sample treatment protocols and nucleic-acid amplification methods with CDetection, we have established an integrated viral nucleic-acid detection platform—CASdetec (CRISPR-assisted detection). The detection limit of CASdetec for SARS-CoV-2 pseudovirus is 1×10^4 copies/mL, with no cross-reactivity observed. Here, we present our assay design and optimization process, which could provide guidance for future CRISPR-based nucleic-acid detection assay development and optimization.

To optimize the output of fluorescence signal, we designed and synthesized poly-T fluorescence-quenchers of varying nucleotide lengths, including 4 nt, 5 nt, 7 nt, 12 nt, 17 nt, 22 nt, and 27 nt. Of all the lengths tried, the 7 nt

poly-T reporter provided the highest signals in the shortest time (Supplementary Fig. S1a, b). Based on this observation, we applied the 7 nt poly-T reporter in later experiments.

According to the published SARS-CoV-2 whole-genome sequence⁸, we designed seven sgRNAs around the RdRp locus, as recommended by World Health Organization (WHO)² (Supplementary Fig. S2a and Table S1). Owing to its high similarity to SARS-CoV, we ran initial experiments on both SARS-CoV-2 and SARS-CoV plasmids. According to fluorescence kinetics studies, sgRNA-3 stood out in not only being able to distinguish between the two similar coronaviruses but also being able to produce the most distinct fluorescence signal (Fig. 1a and Supplementary Fig. S2b–g).

As previously demonstrated, CRISPR is unable to detect any target DNA when there is <1 – 10 nM of amplification product within the reaction mix⁷. Hence, increasing the molecular collisions between CRISPR and target would be essential to improve sensitivity. We found out that increasing the sgRNA concentration by threefolds not only enhances the fluorescence signal and the signal-to-background ratio, but also increases the rate of reaction (Fig. 1b and Supplementary Fig. S3a, b).

Given that the average viral load in the plasma of SARS patients ranged from <1 to ~ 1000 copies per microliter⁹, or 1×10^3 – 1×10^6 copies/mL, nucleic-acid amplification techniques are needed to produce sufficient DNA for CRISPR-based DNA detection methods. Recombinase-aided amplification (RAA) can amplify substrates 10^{10} times at most (from 1 aM to 10 nM) within 10–30 minutes at constant temperature between 37 °C and 42 °C, complementing the needs of CRISPR-based detection. Thus, we designed and screened RAA primers that matched our previously optimized sgRNA-3 (Supplementary Fig. S4a and Table S1). Based on our screenings, we found that by using the best primer pairs together with sgRNA-3, we

Correspondence: Ruiqi Rachel Wang (wangrr@iozac.cn) or Qi Zhou (zhouqi@iozac.cn) or Wei Li (liweili@iozac.cn)

¹State Key Laboratory of Stem Cell and Reproductive Biology, Institute of Zoology, Chinese Academy of Sciences, Beijing 100101, China

²Institute for Stem Cell and Regeneration, Chinese Academy of Sciences, Beijing 100101, China

Full list of author information is available at the end of the article

These authors contributed equally: Xuehan Sun, Xinge Wang, Chen Liang, Haiping Jiang, Qingqin Gao, Moyu Dai

© The Author(s) 2020



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.




Fig. 1 (See legend on next page.)

(see figure on previous page)

Fig. 1 CASdetec used for SARS-CoV-2 detection. **a** Fluorescence kinetics of sgRNA-3 for RdRp detection. *E. coli* cells bearing Blunt-SARS-CoV-RdRp or Blunt-SARS-CoV-2-RdRp were pre-incubated at 95 °C for 10 min and used as templates for RAA and CDetection. PAM sequences are colored in green, protospacers are colored in blue, base pair mismatches are colored in red. Error bars indicate standard errors of the mean (s.e.m.), $n = 3$. RFU, relative fluorescence units. **b** Fluorescence kinetics of RdRp detection using 108 nM sgRNA-3. Plasmid bearing SARS-CoV-2-RdRp was serially diluted as shown in the legend. $n = 2$. ΔR_n , Δ Fluorescence, which refers to the R_n value of an experimental reaction minus the R_n value of the baseline signal generated by ABI 7500. **c** Fluorescence kinetics of F1- and R1-based RdRp detection. SARS-CoV-2-RdRp RNA was serially diluted as shown in the legend. Error bars indicate (s.e.m.), $n = 3$. **d** Evaluation of cross-reactivity. Plasmids containing target RdRp region from six human epidemic coronaviruses were serially diluted as the shown in the legend. $n = 2$. **e** Detection of SARS-CoV-2 pseudovirus. Virus genome was extracted using the virus RNA extraction kit (spin column). SARS-CoV were diluted to 5×10^5 copies/mL. $n = 2$. **f** Detection of SARS-CoV-2 pseudovirus. Virus was treated by direct lysis. SARS-CoV was diluted to 5×10^5 copies/mL. $n = 2$. **g** CASdetec results could be directly observed under blue LED. 3 replicates of products from Fig. 1c were imaged upon blue LED illumination. **h** Schematics showing the workflow of CASdetec. Virus genome was extracted by kit or direct lysis. Target sequences were pre-amplified by isothermal amplification, followed was CDetection. Fluorescence signals were obtained either from fluorescence reader or direct observation under blue light.

can detect SARS-CoV-2 RNA samples containing 5×10^3 copies/mL (Fig. 1c and Supplementary Fig. S4b, c).

In addition, we analyzed all SARS-CoV-2 sequences that have been uploaded to GISAID up till March 26th 2020. Out of 1792 sequences on GISAID, 1673 of them contained sequences matching our chosen primers and sgRNA. Only three of them have one mismatch to the forward primer and only two of them have one mismatch to the reverse primer (Supplementary Fig. S5), suggesting that our selected sgRNA and primers can be used for nearly all of the reported SARS-CoV-2 genomes. Meanwhile, we aligned the selected primers and sgRNAs to twelve typical human coronaviruses to evaluate their specificity, and found that none of the whole set of primers and sgRNA showed high similarity (Supplementary Fig. S6a–c).

The experiments above were conducted by executing RT-RAA nucleic-acid amplification and CDetection separately. However, it would be best to conduct both reactions within a single tube for convenience and, more importantly, to prevent aerosol contamination, which happens when the reaction mixture has to be exposed to the environment midway through the protocol. Hence, we tried to execute both the RT-RAA and CDetection concurrently within a single tube. However, the combination resulted in a drastic decrease in sensitivity (Fig. 1c and Supplementary Fig. S7). Therefore, in order to keep both the RT-RAA and CRISPR reactions within a single tube, we executed the RT-RAA reaction within the tube while keeping the CDetection reagents within the cap of the tube for 30 minutes, following which, the CDetection reagents were spun down into the tube for nucleic-acid detection, and the resultant reaction mixture was imaged for fluorescence.

To validate the specificity of our method for SARS-CoV-2 nucleic-acid detection, we tested our protocols against six coronaviruses known to cause respiratory diseases (SARS-CoV, MERS-CoV, CoV-HKU1, CoV-229E, CoV-OC43, and CoV-NL63). Consistent with alignment analysis (Supplementary Fig. S6a–c), no cross-reactivity with other endemic human coronavirus were detected (Fig. 1d). Our results

suggested our set of sgRNA and primers showed high sensitivity and specificity.

However, viral genomes are packaged inside capsid protein and need to be released. Thus, to investigate the virus handling process, we produced pseudoviruses by packaging the target sequences of SARS-CoV-2, SARS-CoV, MERS-CoV into actual lentivirus particles (Supplementary Fig. S8). These pseudoviruses were diluted serially and treated with either virus genome extraction kits (spin column) or lysis buffer, respectively. Our results demonstrated that the spin column treatment gave a lower detection limit of 1×10^4 copies/mL (Fig. 1e). On the other hand, the lysis buffer offered higher usability, but raised the detection limit to 5×10^4 copies/mL (Fig. 1f). Owing to the difference in detection limit between spin column and lysis buffer, we suggest using the spin columns in hospitals, and using the lysis buffer for point-of-care testing (POCT).

To make CASdetec more amenable for POCT, we have also constructed a portable dark box containing a blue LED and demonstrated that the positive fluorescence signal generated from the protocol can be visualized upon illumination by a blue LED (Fig. 1g). The detection limit was decreased under blue LED compared with quantitative PCR instrument. Optimizing the extensity of excitation light and the sensitivity of optical detector may increase the detection limit of optical visualization. What's more, a more impersonal readout method, such as taking photos followed by mobile software-mediated analyses, can be developed to analyze the optical results.

In conclusion, we have established a CASdetec platform, which consists of procedures including virus handling, nucleic-acid amplification, and CRISPR-based detection (Fig. 1h). CASdetec can detect pseudovirus samples with $>1 \times 10^4$ copies/mL, with no cross-reactivity to other endemic human coronaviruses. In addition, we optimized the workflow to run both reactions within one single tube without lid opening. This will thus prevent aerosol contamination and reduce the false positive rate.

To guarantee the accuracy of CASdetec, efforts should be taken to avoid nucleic-acid aerosol contamination, including (1) conducting reagent preparation, sample treatment, amplification, and product analysis in separate rooms and (2) being careful on handling samples and reagents to avoid contamination from touch.

Acknowledgements

We thank professor Ng Shyh-Chang from Institute of Zoology, CAS for his critical support with this study. We thank Hanxing Zhang from Institute of Microbiology, CAS for her kind help on equipment. This work was supported by the National Key Research and Development Program (2020YFA070009 to R.R.W.), the Key Research Projects of the Frontier Science of the Chinese Academy of Sciences (QYZDY-SSW-SMC002 to Q.Z. and QYZDB-SSW-SMC022 to W.L.), the Strategic Priority Research Program of the Chinese Academy of Sciences (XDA16030400 to W.L.).

Author details

¹State Key Laboratory of Stem Cell and Reproductive Biology, Institute of Zoology, Chinese Academy of Sciences, Beijing 100101, China. ²Institute for Stem Cell and Regeneration, Chinese Academy of Sciences, Beijing 100101, China. ³University of Chinese Academy of Sciences, Beijing 100049, China

Author contributions

L.G., R.R.W., Q.Z. and W.L. conceived and designed the experiments. L.G., X.S., X.W., C.L., H.J., Q.G., M.D., B.Q., S.F., Y.M. and Y.C. participated in multiple experiments; L.G., X.W., X.S., C.L. and G.F. analyzed the data. L.G. wrote the manuscript. W.L., R.R.W., Q.Z. and L.W. provided the final approval of the manuscript.

Conflict of interest

The authors declare that they have no conflict of interest.

Publisher's note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Supplementary Information accompanies the paper at (<https://doi.org/10.1038/s41421-020-0174-y>).

Received: 10 April 2020 Accepted: 26 April 2020

Published online: 19 May 2020

References

1. Coronaviridae Study Group of the International Committee on Taxonomy of Viruses. The species severe acute respiratory syndrome-related coronavirus: classifying 2019-nCoV and naming it SARS-CoV-2. *Nat. Microbiol.* **5**, 536–544 (2020).
2. Corman, V. M. et al. Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. *Eur. Surveill.* **25**, 2000045 (2020).
3. Zhang, F., Abudayyeh, O. O. & Gootenberg, J. S. A protocol for detection of COVID-19 using CRISPR diagnostics. [https://www.broadinstitute.org/files/publications/special/COVID-19%20detection%20\(updated\).pdf](https://www.broadinstitute.org/files/publications/special/COVID-19%20detection%20(updated).pdf) (2020).
4. Metsky, H. C. et al. CRISPR-based COVID-19 surveillance using a genomically-comprehensive machine learning approach. bioRxiv. <https://doi.org/10.1101/2020.02.26.967026> (2020).
5. Lucia, C., Federico, P. B. & Alejandra, G. C. An ultrasensitive, rapid, and portable coronavirus SARS-CoV-2 sequence detection method based on CRISPR-Cas12. bioRxiv. <https://doi.org/10.1101/2020.02.29.971127> (2020).
6. Broughton, J. P. et al. CRISPR-Cas12-based detection of SARS-CoV-2. *Nat. Biotechnol.* <https://doi.org/10.1038/s41587-020-0513-4> (2020).
7. Teng, F. et al. CDetection: CRISPR-Cas12b-based DNA detection with sub-attomolar sensitivity and single-base specificity. *Genome Biol.* **20**, 132 (2019).
8. Zhang, Y. Novel 2019 coronavirus genome. <http://virological.org/t/novel-2019-coronavirus-genome/319> (2020).
9. Grant, P. R. et al. Detection of SARS coronavirus in plasma by real-time RT-PCR. *N. Engl. J. Med.* **349**, 2468–2469 (2003).