




Since January 2020 Elsevier has created a COVID-19 resource centre with free information in English and Mandarin on the novel coronavirus COVID-19. The COVID-19 resource centre is hosted on Elsevier Connect, the company's public news and information website.

Elsevier hereby grants permission to make all its COVID-19-related research that is available on the COVID-19 resource centre - including this research content - immediately available in PubMed Central and other publicly funded repositories, such as the WHO COVID database with rights for unrestricted research re-use and analyses in any form or by any means with acknowledgement of the original source. These permissions are granted for free by Elsevier for as long as the COVID-19 resource centre remains active.



## Letter to the Editor

## Evaluating the newly developed BioFire COVID-19 test for SARS-CoV-2 molecular detection

Flora Marzia Liotti<sup>1,2,†</sup>, Giulia Menchinelli<sup>1,2,†</sup>, Simona Marchetti<sup>2</sup>,  
Grazia Angela Morandotti<sup>2</sup>, Maurizio Sanguinetti<sup>1,2,\*</sup>, Brunella Posteraro<sup>1,3,§</sup>,  
Paola Cattani<sup>1,2,§</sup>

<sup>1</sup> Dipartimento di Scienze Biotecnologiche di Base, Cliniche Intensivologiche e Perioperatorie, Università Cattolica del Sacro Cuore, Rome, Italy

<sup>2</sup> Dipartimento di Scienze di Laboratorio e Infettivologiche, Rome, Italy

<sup>3</sup> Dipartimento di Scienze Gastroenterologiche, Endocrino-Metaboliche e Nefro-Urologiche, Fondazione Policlinico Universitario A. Gemelli IRCCS, Rome, Italy

## ARTICLE INFO

## Article history:

Received 19 June 2020

Received in revised form

16 July 2020

Accepted 18 July 2020

Available online 28 July 2020

Editor: L. Leibovici

It was not until 24 March 2020 that the newly developed BioFire coronavirus disease 2019 (COVID-19) test (BioFire Defense, Salt Lake City, UT, USA) for PCR-based detection of RNA from severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) in clinical samples received a US Food and Drug Administration emergency-use authorization (<https://www.fda.gov/media/136356/download>). It is thus not surprising that no published studies to date have evaluated the BioFire COVID-19 test in clinical microbiology practice.

We compared the performance of BioFire COVID-19 test with that of Quanty COVID-19 assay (Clonit, Milan, Italy), which also provides quantitative results, for detection of SARS-CoV-2 RNA in nasal/oropharyngeal (N/OP) patient samples. Both molecular tests detect SARS-CoV-2 specifically, with the first targeting two viral open reading frame (ORF) sequences (ORF1ab and ORF8)—in three independent PCR assays—and the second targeting three viral

nucleocapsid (N) sequences (N1, N2 and N3). Such comparison is essential for defining the cause of potential false-negative results [1], which may undermine the clinical utility of various molecular diagnostic tests available currently [2,5].

We analysed the results of 120 N/OP samples tested with both the BioFire COVID-19 test and the Quanty COVID-19 assay. Samples that had been kept frozen at  $-70^{\circ}\text{C}$  until testing to ensure RNA integrity were randomly selected from among those that were SARS-CoV-2 positive ( $n = 86$ ) and negative ( $n = 34$ ), as tested with the Allplex 2019-nCoV assay (Arrow Diagnostics, Genova, Italy) [3], and then confirmed (as positive or negative respectively) by a real-time PCR assay (here used as the reference method) based on the Corman et al. method [4]. The agreement between the BioFire COVID-19 test and Quanty COVID-19 assay was 95.0% (114/120) for overall results and 100% (34/34) for negative results. Eighty (93.0%) of 86 positive samples yielded results with the BioFire COVID-19 test that matched those with the Quanty COVID-19 assay. For six remaining positive samples, BioFire COVID-19 test results did not match those with the Quanty COVID-19 assay. As shown in [Supplementary Table S1](#), two of six samples—falsely negative by BioFire COVID-19 test—had no detections in all three assays (hence interpreted as ‘not detected’), whereas four samples initially yielded detection in only one assay (hence interpreted as ‘equivocal’) but resulted as ‘not detected’ at retesting. Interestingly, virus loads (expressed as RNA copies/mL) of the six samples were  $2.20 \times 10^1$  to  $1.60 \times 10^2$ . However, these loads were below the limit of detection of  $3.30 \times 10^2$  RNA copies/mL estimated for the BioFire COVID-19 test (<https://www.biofiredefense.com/covid-19test/>). In 80 samples with results agreeing between the assays, the median (interquartile range) virus load was  $7.89 \times 10^3$  ( $2.48 \times 10^3$ – $2.75 \times 10^5$ ) RNA copies/mL, which was consistent with an average (range) value of  $1.24 \times 10^8$  ( $3.82 \times 10^2$ – $7.83 \times 10^9$ ) RNA copies/mL. Compared to the reference method, the BioFire COVID-19 test sensitivity, specificity, positive predictive value and negative predictive value (with their 95% confidence intervals) were 93.0 (85.4–97.4), 100.0

\* Corresponding author: Maurizio Sanguinetti, Dipartimento di Scienze di Laboratorio e Infettivologiche, Fondazione Policlinico Universitario A. Gemelli IRCCS, Largo A. Gemelli 8, 00168 Rome, Italy.

E-mail address: [maurizio.sanguinetti@unicatt.it](mailto:maurizio.sanguinetti@unicatt.it) (M. Sanguinetti).

† The first two authors contributed equally to this letter, and both should be considered first author.

§ The last two authors contributed equally to this letter, and both should be considered senior author.

(89.7–100.0), 100.0 (95.5–100.0) and 85.0 (70.2–94.3), respectively.

These findings suggest that the lower analytical sensitivity of the BioFire COVID-19 test might have caused false-negative results in our study. Consequently, compared to molecular tests such as the Quanta COVID-19 assay, the analytical sensitivity shown by BioFire COVID-19 test would result in a slight reduction in its clinical sensitivity in COVID-19 diagnosis. Additionally, 'equivocal' results that at repeated testing with the BioFire COVID-19 test are claimed as 'not detected' may actually be truly positive, but this requires further investigation. Relying on fully automated FilmArray platforms, BioFire COVID-19 test provides results in approximately 45 minutes from N/OP sample collection. Thus, the possibility of shortening the time to results merits consideration when deciding which SARS-CoV-2 molecular test to implement in the clinical microbiology laboratory.

### Transparency declaration

Reale Group and Fondazione Valentino Garavani & Giancarlo Giammetti provided financial support for COVID-19 research. bio-Mérieux (Marcy l'Étoile, France) provided the reagents for this

study. All authors report no conflicts of interest relevant to this letter.

### Appendix A. Supplementary data

Supplementary data to this article can be found online at <https://doi.org/10.1016/j.cmi.2020.07.026>.

### References

- [1] Winichakoon P, Chaiwarith R, Liwsrisakun C, Salee P, Goonna A, Limsukon A, et al. Negative nasopharyngeal and oropharyngeal swabs do not rule out COVID-19. *J Clin Microbiol* 2020;58:e00297-20.
- [2] Cheng MP, Papenburg J, Desjardins M, Kanjilal S, Quach C, Libman M, et al. Diagnostic testing for severe acute respiratory syndrome-related coronavirus-2: a narrative review. *Ann Intern Med* 2020;172:726–34.
- [3] World Health Organization. Laboratory testing for coronavirus disease (COVID-19) in suspected human cases. Interim guidance. Geneva: World Health Organization; 2020. Available at: <https://apps.who.int/iris/bitstream/handle/10665/331501/WHOCOVID-19-laboratory-2020.5-eng.pdf>.
- [4] Corman VM, Landt O, Kaiser M, Molenkamp R, Meijer A, Chu DK, et al. Detection of 2019 novel coronavirus (2019-nCoV) by realtime RT-PCR. *Euro Surveill* 2020;25:2000045.
- [5] Patel R, Babady E, Theel ES, Storch GA, Pinsky BA, St George K. Report from the American Society for Microbiology COVID-19 international summit, 23 March 2020: value of diagnostic testing for SARS-CoV-2/COVID-19. *mBio* 2020;11:e00722-20.