



Published in final edited form as:

Nano Lett. 2019 October 09; 19(10): 6862–6868. doi:10.1021/acs.nanolett.9b02070.

Plasmonic Microneedle Arrays for in Situ Sensing with Surface-Enhanced Raman Spectroscopy (SERS)

Ji Eun Park[†], Nihan Yonet-Tanyeri[‡], Emma Vander Ende[†], Anne-Isabelle Henry[†], Bethany E. Perez White[§], Milan Mrksich^{*,†,‡}, Richard P. Van Duyne^{*,||,†,‡}

[†]Department of Chemistry, Northwestern University, 2145 Sheridan Road, Evanston, Illinois 60208, United States

[‡]Department of Biomedical Engineering, Northwestern University, 2145 Sheridan Road, Evanston, Illinois 60208, United States

[§]Skin Tissue Engineering Core and Department of Dermatology, Feinberg School of Medicine, Northwestern University, Chicago, Illinois 60611 United States

Abstract

Surface-enhanced Raman spectroscopy (SERS) is a sensitive, chemically specific, and short-time response probing method with significant potential in biomedical sensing. This paper reports the integration of SERS with microneedle arrays as a minimally invasive platform for chemical sensing, with a particular view toward sensing in interstitial fluid (ISF). Microneedle arrays were fabricated from a commercial polymeric adhesive and coated with plasmonically active gold nanorods that were functionalized with the pH-sensitive molecule 4-mercaptobenzoic acid. This sensor can quantitate pH over a range of 5 to 9 and can detect pH levels in an agar gel skin phantom and in human skin in situ. The sensor array is stable and mechanically robust in that it exhibits no loss in SERS activity after multiple punches through an agar gel skin phantom and human skin or after a month-long incubation in phosphate-buffered saline. This work is the first to integrate SERS-active nanoparticles with polymeric microneedle arrays and to demonstrate in situ sensing with this platform.

Graphical Abstract

*Corresponding Authors: milan.mrksich@northwestern.edu. vanduyne@northwestern.edu.

Author Contributions

J.P. and N.Y.-T. contributed equally with designing and conducting the experiments, characterizing the sensor, and analyzing the data. E.V.E. and A.-I.H. mentored and supported with many fruitful discussions to develop this work. A.-I.H. contributed to the general direction of the project and helped put the project together. B.E.P.W. helped facilitate experiments on human skin through experimental design. A.-I.H., M.M. and R.P.V.D. supervised the experimental aspect and the direction of the project.


^{||}(R.P.V.D.) Died July 28, 2019.

The authors declare no competing financial interest.

Supporting Information

The Supporting Information is available free of charge on the ACS Publications website at DOI: 10.1021/acs.nanolett.9b02070.

Experimental section with chemicals and materials, instrumentation, detailed information regarding the plasmonic microneedle arrays and Au film-over-nano-spheres fabrication and characterization, gold nanorod synthesis and characterization, experimental procedures, data acquisition and analysis, and additional experimental results (Figures S1–S11) (PDF)



Keywords

Plasmonic microneedle arrays; SERS; pH in situ sensing; agar gel skin phantom; human skin

Surface-enhanced Raman spectroscopy (SERS) is a well-established analytical technique that can quantitate analytes with high sensitivity, even down to the single molecule level in special cases.^{1,2} In addition to its sensitivity, SERS is rapid and provides structural information on analytes, which is a significant benefit in reducing false positive measurements. Advances in fabricating highly enhanced SERS substrates have enabled Raman scattering enhancement factors of up to 10^8 relative to normal Raman,^{3,4} leading to its application in both fundamental studies and in a variety of applied studies, including catalysis,⁵ electrochemistry,⁶ art conservation,⁷ and in vivo biosensing.⁸ SERS for biomedical applications is an area of interest, particularly for continuous glucose monitoring due to its short-time response, but one impediment to such an application has been the difficulty of integrating plasmonically enhancing substrates with living tissue, in part because implanted sensors undergo biofouling and lose activity,^{9,10} and because the implantation is invasive. Important prior work to detect analytes of interest with SERS in a minimally biofouling or invasive way has utilized a polymeric film,¹¹ microcapsules,¹² and metalized^{13,14} and paper-based plasmonic microneedles.¹⁵ In the present work, we describe an approach to minimize invasive SERS devices for biomedical applications by integrating plasmonically enhanced nanoparticles (gold nanorods) for SERS with a polymeric microneedle array. We demonstrate the effectiveness of the plasmonic microneedle array by modifying the gold nanorods with the pH-sensitive molecule 4-mercaptobenzoic acid (4-MBA) and measuring the pH of solutions, the pH in an agar gel skin phantom, and the pH in human skin.

Microneedle arrays supported on a flexible substrate have been developed for minimally invasive biomedical applications,^{16,17} including transdermal drug and vaccine delivery^{16,18–20} and extraction of biofluids such as interstitial fluid (ISF) or blood through the skin.^{21–24} Microneedle sensors have also been used to measure biologically relevant analytes in the ISF, including glucose,^{25,26} pH,²⁶ lactate,^{26,27} glutamate,²⁸ nerve agents,²⁹ and alcohol³⁰ using enzyme-based electrochemical methods. ISF biomarker monitoring using microneedle arrays and SERS would circumvent potential problems to the standard electrochemical techniques (e.g., enzyme degradation³¹ or instability³²). In addition to biological analytes in the ISF, the pH of ISF is also important. ISF pH is critical in sustaining homeostasis and normal bodily functions.³³ Compared to the blood pH that is strictly controlled in the normal range (pH 7.35–7.45), the ISF pH can deviate from the normal range (pH 6.60–7.60) due to fewer pH-buffering molecules than in blood.^{33,34} Deviation from the normal pH range in ISF has been suggested as a biomarker for skin

cancers, diabetes mellitus, and other diseases.^{33,35–38} Thus, in this work we describe an in situ pH sensor, both as a demonstration of the SERS activity of the microneedle array platform and because pH is important in diagnostic applications.

A schematic representation of the plasmonic microneedle array and pH sensing with 4-MBA is shown in Figure 1. The microneedle arrays are prepared from Norland Optical Adhesive (NOA) 65 (Figure 2). This polymer consists of mercaptoester functional group, which we hypothesize that it allows us to functionalize AuNRs on the polymer surface and is also transparent at 785 nm, therefore enabling the collection of the SERS signal through the microneedle array and permitting in situ sensing. In addition, biocompatibility of various NOAs has been demonstrated; NOA patterns have been used as a cell culture scaffold for endothelial cells,³⁹ fibroblasts,⁴⁰ HeLa cells and neurons,⁴¹ and human embryonic stem cells⁴² without any adverse effects. The polymer was cured inside a PDMS mold with 10 min of 365 nm exposure to produce a microneedle array having a size of approximately 1 cm² and with pyramidal microneedles each 300 μm \times 200 μm (height \times width) with a tip-to-tip spacing of 500 μm (Figure 3). The Supporting Information describes the fabrication in detail.

Gold nanorods (AuNRs) were synthesized using the silver-assisted seed-mediated method and stabilized with cetyltrimethylammonium bromide (CTAB).^{43–45} We selected AuNRs with transverse and longitudinal dimensions of 15 and 55 nm, respectively, because particles of this size have a plasmon resonance at 785 nm (based on the UV–vis extinction spectrum shown in Figure S1) and the Raman scattering enhancement is highest when the excitation wavelength overlaps with the extinction maximum of the plasmonic material.

We first treated the microneedle array with ozone, which both sterilizes the surface and promotes attachment of the AuNRs to the surface of the microneedles.^{46,47} The microneedle array was then incubated for 24 h in a solution composed of 40 μL of aqueous AuNRs (2×10^{13} particles/mL) and 1 mL of 10 mM 4-MBA (in 1:1 ethanol/water) that was premixed immediately prior to incubation. As the 4-MBA binds to the AuNRs through Au–thiolate bond formations, the AuNRs aggregate and deposit onto the surface of the microneedle array (Figure 2). 4-MBA functions as a Raman reporter molecule for pH because its carboxylate vibrational mode shifts between 1400 and 1425 cm^{-1} depending on the protonation state, which can be detected by SERS and the integrated intensity of that mode provides a measure (when calibrated) for the pH of the sensing environment.⁴⁸ We used this plasmonic microneedle array to measure the pH of solutions and an agar gel skin phantom, as well as to demonstrate suitable thermal and mechanical stability of the platform. The attachment of the AuNRs onto the surface of the microneedles was confirmed by optical microscopy, scanning electron microscopy (SEM), and UV–vis extinction spectroscopy (Figure 3). The extinction spectrum of the microneedle array with AuNRs shows a broad peak centered around 780 nm that represents the AuNR aggregates on the microneedle surface (Figure 3d).

We first calibrated the sensors for pH measurements by collecting SER spectra (Figure 4a,b) from three replicate 4-MBA plasmonic microneedle array sensors in Britton-Robinson buffer solutions (pH 2–12) and averaging the spectra to produce a measure of precision (Figure 4c). Britton-Robinson buffer was chosen to generate a calibration curve in the range between pH

2 to 12 with 1 pH unit increments. The weak background of NOA 65 was subtracted from all 4-MBA SER spectra (see Materials and Methods in the Supporting Information for details). The carboxylate stretching mode whose Raman peak is centered between 1400 and 1425 cm^{-1} increases in intensity and blue shifts (from 1400 to 1425 cm^{-1}) (Figure 4b) with increasing pH of the buffer. To control for overall intensity variations, we took the ratio between the integrated peak intensity of the carboxylate peak (A_{cbx}) as it shifts between 1400 and 1425 cm^{-1} and a mode at 1076 cm^{-1} (A_{1076}) that is not responsive to pH (A_{cbx}/A_{1076}). We then normalized this ratio by the maximum ratio (normalized A_{cbx}/A_{1076}). A plot of this normalized ratio versus pH gave the expected S-shaped calibration curve and revealed that the working pH region (or the linear region) is in the pH range of 5 to 9 (Figure 4c). We therefore determined a second calibration curve in this range but with smaller (0.2) pH increments to characterize the resolution of the sensor (Figure 4d). The resolution of the pH sensor is represented as twice the standard deviation at pH 7 (the middle of the curve), which in this case is 0.23 pH units.⁴⁸

We also compared data from our new microneedle sensing platform with a well-known SERS substrate, Au film-over-nanospheres (AuFON), with the LSPR appropriate for 785 nm excitation wavelength in water (Figure S2).⁴⁹ We found a small difference in apparent $\text{p}K_{\text{a}}$ value of 4-MBA on the two substrates and we attribute this small change to differences in the density of 4-MBA on the two surfaces. The $\text{p}K_{\text{a}}$ value of 4-MBA in solution (4.8) is known to differ from that on a monolayer (7.7) because of electrostatic destabilization of the carboxylates.⁵⁰ For the plasmonic microneedle array, we find the $\text{p}K_{\text{a}}$ value of 4-MBA is 7.13 (Figure S3). The $\text{p}K_{\text{a}}$ of 4-MBA on the AuFON is 7.38, which suggests that the density of 4-MBA is slightly lower on AuNR aggregates than on the AuFON. What this indicates is that the two platforms are comparable and function over similar pH ranges.

We next demonstrated that the sensor has the properties required for in situ sensing by measuring the pH inside a skin phantom agar gel. A SER spectrum of the plasmonic microneedle array was collected in air (Figure 5a, black trace) as a baseline before punching the microneedle arrays into an agar gel skin phantom. The plasmonic microneedle array was punctured into the agar gel skin phantom by pressing the gel onto the microneedle arrays by hand and SER spectra were collected through the microneedle array while it was in the skin phantom. The carboxylate peak at 1400 cm^{-1} (Figure 5a, black trace) increases in intensity and shifts to 1421 cm^{-1} when the microneedles are in the agar gel (Figure 5a, blue trace). The sensor was then taken out of the skin phantom and the SERS was measured again after rinsing it with water (Figure 5a, red trace). The resulting SER spectra before and after penetrating the agar gel were the same, showing the reversibility of the sensor. The mechanical robustness of the sensor was demonstrated by puncturing an agar gel skin phantom 10 times and observing that the SERS activity of the sensor was unchanged by the repeated puncturing (Figure 5b). To more robustly demonstrate the reversibility of the pH sensing, we collected SER spectra of the microneedle sensor after cycling the pH of the buffer solution between pH 2 to 12 and found that no signal loss occurred after six cycles (Figure 4e).

The stability of the sensor was then tested by incubating a microneedle array in phosphate-buffered saline for one month. As shown in Figure 5c, the SERS activity was unchanged by

the incubation as compared to a freshly prepared microneedle array. Additionally, the extinction spectrum of the microneedle sensor was taken every week during the month-long incubation to observe the presence of AuNR aggregates. As shown in Figure S6, the AuNR aggregates, represented by a broad peak around 780 nm, are still present on the microneedle array surface as of the final week.

To translate this study to ex vivo sensing, we applied the plasmonic microneedle array sensor to human skin (details in the Supporting Information). We determined the penetration depth of the microneedle array in skin using optical coherence tomography (OCT) and the sensor's ability to measure pH in situ by SERS. Compared with the flat region of the array as a control, the region with microneedles shows that microneedles can be placed in skin (Figure S7). Next, to change and measure the skin pH in the range between pH 6.6 and 7.6 (ISF pH) we incubated human skin in Dulbecco's phosphate-buffered saline (DPBS) with pH 6.6, 7.1, and 7.6 for 12 h while replacing the buffer solution every 2 h. As shown in Figure S8, the incubation was performed in a Petri dish while the tissue floated on a metal mesh boat so that the dermal side is in contact with the buffer solution and the top part (the stratum corneum) remains dry. After incubation, we dried the bottom of the skin with clean lab tissue paper and inserted a microneedle array in the skin. We then acquired SERS while the laser was focused through the polymer and onto the tips, as with the agar gel skin phantom. As controls, SER spectra of a microneedle array sensor was taken in the DPBS buffer solution in which the human skin was incubated. As shown in Figure 5d, compared to the controls, the A_{cbx}/A_{1076} corresponds well with the measurement from the skin. We then characterized the microneedle array sensor after puncturing human skin using UV-vis, SERS, and SEM. The SERS signal (Figure S9a) was stable and light extinction spectra (Figure S9b) showed AuNR aggregates present after puncturing human skin (with pH 6.6, 7.1, and 7.6). SEM images of microneedle arrays after insertion in human skin showed that the microneedles were still intact with AuNR aggregates present on the surface (Figure S10). Lastly, to test the mechanical stability of the sensor, the microneedle sensor was applied to human skin tissue repeatedly (10 punches). The SER spectra before and after 10 punches showed no loss of 4-MBA signal (Figure S11).

From the ex vivo study, we demonstrated the sensing ability and the mechanical robustness of the plasmonic microneedle arrays after insertion in skin. The next step is to investigate the toxicity and the lifetime of the sensor after a long period in skin. In order to do so, in vivo studies should be performed. Our studies have not characterized the long-term stability and performance of the sensor arrays in animal models, and we are not able to comment on the stability of the sensor to biofouling and any toxicity.

In conclusion, we report the design and fabrication of a plasmonic microneedle array SERS sensor and demonstrate its SERS activity to measure different pH levels in solutions, in an agar gel skin phantom, and in human skin. Measurement of pH with this sensor was both reversible and reproducible. The use of surface chemistries that provide reports of other molecular analytes, together with the mechanical and optical properties of the microneedle array, offer a strategy for sensing a variety of relevant analytes in interstitial fluids.

Supplementary Material

Refer to Web version on PubMed Central for supplementary material.

ACKNOWLEDGMENTS

The authors thank Ian Rubinoff and Pengxiao Hao with their support with OCT, Tirzah Abbott with her help with SEM, Eric J. Berns, Vitor Brasiliense, and Gyeongwon Kang for helpful discussions, and Shuangni Yang for her help with the experiments on human skin. We thank the Skin Tissue Engineering Core Facility in Feinberg school of Medicine at Northwestern University for providing us with human skin tissues. This work was supported by the Assistant Secretary of Defense for Health Affairs, through the Peer Reviewed Medical Research Program under Award No. W81XWH-16-1-0375. E.V.E. acknowledges support from the National Science Foundation Graduate Research Fellowship Program (DGE-1324585).

REFERENCES

- (1). Nie S; Emory SR Probing Single Molecules and Single Nanoparticles by Surface-Enhanced Raman Scattering. *Science* 1997, 275, 1102–1106. [PubMed: 9027306]
- (2). Camden JP; Dieringer JA; Wang Y; Masiello DJ; Marks LD; Schatz GC; Van Duyne RP Probing the Structure of Single-Molecule Surface-Enhanced Raman Scattering Hot Spots. *J. Am. Chem. Soc* 2008, 130 (38), 12616–12617. [PubMed: 18761451]
- (3). Willets KA; Van Duyne RP Localized Surface Plasmon Resonance Spectroscopy and Sensing. *Annu. Rev. Phys. Chem* 2007, 58, 267–297. [PubMed: 17067281]
- (4). Anker JN; Hall WP; Lyandres O; Shah NC; Zhao J; Van Duyne RP Biosensing with Plasmonic Nanosensors. *Nat. Mater* 2008, 7, 442–453. [PubMed: 18497851]
- (5). Heck KN; Janesko BG; Scuseria GE; Halas NJ; Wong MS Observing Metal-Catalyzed Chemical Reactions in Situ Using Surface-Enhanced Raman Spectroscopy on Pd-Au Nanoshells. *J. Am. Chem. Soc* 2008, 130, 16592–16600. [PubMed: 19554693]
- (6). Zaleski S; Clark KA; Smith MM; Eilert JY; Doty M; Van Duyne RP Identification and Quantification of Intravenous Therapy Drugs Using Normal Raman Spectroscopy and Electrochemical Surface-Enhanced Raman Spectroscopy. *Anal. Chem* 2017, 89 (4), 2497–2504. [PubMed: 28192951]
- (7). Roh JY; Matecki MK; Svoboda SA; Wustholz KL Identifying Pigment Mixtures in Art Using SERS: A Treatment Flowchart Approach. *Anal. Chem* 2016, 88 (4), 2028–2032. [PubMed: 26799174]
- (8). Ma K; Yuen JM; Shah NC; Walsh JT; Glucksberg MR; Van Duyne RP In Vivo, Transcutaneous Glucose Sensing Using Surface-Enhanced Spatially Offset Raman Spectroscopy: Multiple Rats, Improved Hypoglycemic Accuracy, Low Incident Power, and Continuous Monitoring for Greater than 17 Days. *Anal. Chem* 2011, 83 (23), 9146–9152. [PubMed: 22007689]
- (9). Yun YH; Eteshola E; Bhattacharya A; Dong Z; Shim JS; Conforti L; Kim D; Schulz MJ; Ahn CH; Watts N Tiny Medicine: Nanomaterial-Based Biosensors. *Sensors* 2009, 9 (11), 9275–9299. [PubMed: 22291565]
- (10). Walkey CD; Chan WCW Understanding and Controlling the Interaction of Nanomaterials with Proteins in a Physiological Environment. *Chem. Soc. Rev* 2012, 41 (7), 2780–2799. [PubMed: 22086677]
- (11). Lee WWY; McCoy CP; Donnelly RF; Bell SEJ Swellable Polymer Films Containing Au Nanoparticles for Point-of-Care Therapeutic Drug Monitoring Using Surface-Enhanced Raman Spectroscopy. *Anal. Chim. Acta* 2016, 912, 111–116. [PubMed: 26920779]
- (12). You Y-H; Biswas A; Nagaraja AT; Hwang J-H; Coté GL; McShane MJ Multidomain-Based Responsive Materials with Dual-Mode Optical Readouts. *ACS Appl. Mater. Interfaces* 2019, 11, 14286–14295. [PubMed: 30908908]
- (13). Yuen C; Liu Q Towards in Vivo Intradermal Surface Enhanced Raman Scattering (SERS) Measurements: Silver Coated Microneedle Based SERS Probe. *J. Biophotonics* 2014, 7 (9), 683–689. [PubMed: 23666825]

- (14). Yuen C; Liu Q Hollow Agarose Microneedle with Silver Coating for Intradermal Surface-Enhanced Raman Measurements: A Skin-Mimicking Phantom Study. *J. Biomed. Opt* 2015, 20 (6), 061102.
- (15). Kolluru C; Gupta R; Jiang Q; Williams M; Gholami Derami H; Cao S; Noel RK; Singamaneni S; Prausnitz MR Plasmonic Paper Microneedle Patch for On-Patch Detection of Molecules in Dermal Interstitial Fluid. *ACS Sensors* 2019, 4 (6), 1569–1576. [PubMed: 31070358]
- (16). Wang M; Hu L; Xu C Recent Advances in the Design of Polymeric Microneedles for Transdermal Drug Delivery and Biosensing. *Lab Chip* 2017, 17 (8), 1373–1387. [PubMed: 28352876]
- (17). Babity S; Roohnikan M; Brambilla D Advances in the Design of Transdermal Microneedles for Diagnostic and Monitoring Applications. *Small* 2018, 14 (49), 1803186.
- (18). Yang S; Wu F; Liu J; Fan G; Welsh W; Zhu H; Jin T Phase-Transition Microneedle Patches for Efficient and Accurate Transdermal Delivery of Insulin. *Adv. Funct. Mater* 2015, 25 (29), 4633–4641.
- (19). Donnelly RF; McCrudden MTC; Alkilani AZ; Larrañeta E; McAlister E; Courtenay AJ; Kearney MC; Raj Singh TR; McCarthy HO; Kett VL; Caffarel-Salvador E; Al-Zahrani S; Woolfson AD Hydrogel-Forming Microneedles Prepared from “Super Swelling” Polymers Combined with Lyophilised Wafers for Transdermal Drug Delivery. *PLoS One* 2014, 9 (10), No. e111547.
- (20). Arya J; Prausnitz MR Microneedle Patches for Vaccination in Developing Countries. *J. Controlled Release* 2016, 240, 135–141.
- (21). Chang H; Zheng M; Yu X; Than A; Seeni RZ; Kang R A Swellable Microneedle Patch to Rapidly Extract Skin Interstitial Fluid for Timely Metabolic Analysis. *Adv. Mater* 2017, 29, 1702243.
- (22). Li CG; Dangol M; Lee CY; Jang M; Jung H A Self-Powered One-Touch Blood Extraction System: A Novel Polymer-Capped Hollow Microneedle Integrated with a Pre-Vacuum Actuator. *Lab Chip* 2015, 15 (2), 382–390. [PubMed: 25352059]
- (23). Caffarel-Salvador E; Brady AJ; Eltayib E; Meng T; Alonso-Vicente A; Gonzalez-Vazquez P; Torrisi BM; Vicente-Perez EM; Mooney K; Jones DS; Bell SEJ; McCoy CP; McCarthy HO; McElnay JC; Donnelly RF Hydrogel-Forming Microneedle Arrays Allow Detection of Drugs and Glucose in Vivo: Potential for Use in Diagnosis and Therapeutic Drug Monitoring. *PLoS One* 2015, 10 (12), e0145644. [PubMed: 26717198]
- (24). Donnelly RF; Mooney K; Caffarel-Salvador E; Torrisi BM; Eltayib E; McElnay JC Microneedle-Mediated Minimally Invasive Patient Monitoring. *Ther. Drug Monit* 2014, 36 (1), 1.
- (25). Ribet F; Stemme G; Roxhed N Real-Time Intradermal Continuous Glucose Monitoring Using a Minimally Invasive Microneedle-Based System. *Biomed. Microdevices* 2018, 20 (4), 101. [PubMed: 30523421]
- (26). Miller PR; Skoog SA; Edwards TL; Wheeler DR; Xiao X; Brozik SM; Polsky R; Narayan RJ Hollow Microneedle-Based Sensor for Multiplexed Transdermal Electrochemical Sensing. *J. Visualized Exp* 2012, 3 (64), 1–6.
- (27). Bollella P; Sharma S; Cass AEG; Antiochia R Microneedle-Based Biosensor for Minimally-Invasive Lactate Detection. *Biosens. Bioelectron* 2019, 123, 152–159. [PubMed: 30177422]
- (28). Windmiller JR; Valdés-Ramírez G; Zhou N; Zhou M; Miller PR; Jin C; Brozik SM; Polsky R; Katz E; Narayan R; Wang J Bicomponent Microneedle Array Biosensor for Minimally-Invasive Glutamate Monitoring. *Electroanalysis* 2011, 23 (10), 2302–2309.
- (29). Mishra RK; Vinu Mohan AM; Soto F; Chrostowski R; Wang J A Microneedle Biosensor for Minimally-Invasive Transdermal Detection of Nerve Agents. *Analyst* 2017, 142 (6), 918–924. [PubMed: 28220163]
- (30). Mohan AMV; Windmiller JR; Mishra RK; Wang J Continuous Minimally-Invasive Alcohol Monitoring Using Microneedle Sensor Arrays. *Biosens. Bioelectron* 2017, 91, 574–579. [PubMed: 28088750]
- (31). Valdes TI; Moussy F In Vitro and In Vivo Degradation of Glucose Oxidase Enzyme Used for an Implantable Glucose Biosensor. *Diabetes Technol. Ther* 2000, 2 (3), 367–376. [PubMed: 11467339]

- (32). Harris JM; Reyes C; Lopez GP Common Causes of Glucose Oxidase Instability in In Vivo Biosensing. *J. Diabetes Sci. Technol* 2013, 7 (4), 1030–1038. [PubMed: 23911187]
- (33). Marunaka Y The Proposal of Molecular Mechanisms of Weak Organic Acids Intake-Induced Improvement of Insulin Resistance in Diabetes Mellitus via Elevation of Interstitial Fluid pH. *Int. J. Mol. Sci* 2018, 19 (10), 3244.
- (34). Marunaka Y Roles of Interstitial Fluid pH in Diabetes Mellitus: Glycolysis and Mitochondrial Function. *World J. Diabetes* 2015, 6 (1), 125–135. [PubMed: 25685283]
- (35). Baronzio G; Schwartz L; Kiselevsky M; Guais A; Sanders E; Milanese G; Baronzio M; Freitas I Tumor Interstitial Fluid as Modulator of Cancer Inflammation, Thrombosis, Immunity and Angiogenesis. *Anticancer Res* 2012, 32 (2), 405–414. [PubMed: 22287726]
- (36). Gromov P; Gromova I; Olsen CJ; Timmermans-Wielenga V; Talman ML; Serizawa RR; Moreira JMA Tumor Interstitial Fluid - A Treasure Trove of Cancer Biomarkers. *Biochim. Biophys. Acta, Proteins Proteomics* 2013, 1834 (11), 2259–2270.
- (37). Wagner M; Wiig H Tumor Interstitial Fluid Formation, Characterization, and Clinical Implications. *Front. Oncol* 2015, 5, 1–12. [PubMed: 25667919]
- (38). Jain RK; Shah SA; Finney PL Continuous Noninvasive Monitoring of Ph and Temperature in Rat Walker 256 Carcinoma during Normoglycemia and Hyperglycemia. *J. Natl. Cancer Inst* 1984, 73 (2), 429–436. [PubMed: 6589435]
- (39). Liliensiek SJ; Wood JA; Yong J; Auerbach R; Nealey PF; Murphy CJ Modulation of Human Vascular Endothelial Cell Behaviors by Nanotopographic Cues. *Biomaterials* 2010, 31 (20), 5418–5426. [PubMed: 20400175]
- (40). Meng J; Kong H; Han Z; Wang C; Zhu G; Xie S; Xu H Enhancement of Nanofibrous Scaffold of Multiwalled Carbon Nanotubes/Polyurethane Composite to the Fibroblasts Growth and Biosynthesis. *J. Biomed. Mater. Res, Part A* 2009, 88A (1), 105–116.
- (41). Sollier E; Murray C; Maoddi P; Di Carlo D Rapid Prototyping Polymers for Microfluidic Devices and High Pressure Injections. *Lab Chip* 2011, 11 (22), 3752–3765. [PubMed: 21979377]
- (42). Mohr JC; de Pablo JJ; Palecek SP 3-D Microwell Culture of Human Embryonic Stem Cells. *Biomaterials* 2006, 27 (36), 6032–6042. [PubMed: 16884768]
- (43). Sau TK; Murphy CJ Seeded High Yield Synthesis of Short Au Nanorods in Aqueous Solution. *Langmuir* 2004, 20 (15), 6414–6420. [PubMed: 15248731]
- (44). Nikoobakht B; El-Sayed MA Preparation and Growth Mechanism of Gold Nanorods (NRs) Using Seed-Mediated Growth Method. *Chem. Mater* 2003, 15 (10), 1957–1962.
- (45). Fernández-López C; Mateo-Mateo C; Álvarez-Puebla RA; Pérez-Juste J; Pastoriza-Santos I; Liz-Marzán LM Highly Controlled Silica Coating of PEG-Capped Metal Nanoparticles and Preparation of SERS-Encoded Particles. *Langmuir* 2009, 25 (24), 13894–13899. [PubMed: 19591480]
- (46). Rediguiéri CF; de Jesus Andreoli Pinto T; Bou-Chacra NA; Galante R; de Araújo GLB; do Nascimento Pedrosa T; Maria-Engler SS; De Bank PA Ozone Gas as a Benign Sterilization Treatment for PLGA Nanofibers Scaffolds. *Tissue Eng., Part C* 2016, 22 (4), 338–347.
- (47). Mathieson I; Bradley RH Improved Adhesion to Polymers by UV/Ozone Surface Oxidation. *Int. J. Adhes. Adhes* 1996, 16 (1), 29–31.
- (48). Wei H; Willner MR; Marr LC; Vikesland PJ Highly Stable SERS pH Nanoprobes Produced by Co-Solvent Controlled AuNP Aggregation. *Analyst* 2016, 141 (17), 5159–5169. [PubMed: 27143623]
- (49). Greenelch NG; Blaber MG; Henry A-I; Schatz GC; Van Duyne RP Immobilized Nanorod Assemblies: Fabrication and Understanding of Large Area Surface-Enhanced Raman Spectroscopy Substrates. *Anal. Chem* 2013, 85 (4), 2297–2303. [PubMed: 23343409]
- (50). Phan HT; Haes AJ Impacts of pH and Intermolecular Interactions on Surface-Enhanced Raman Scattering Chemical Enhancements. *J. Phys. Chem. C* 2018, 122 (26), 14846–14856.




Figure 1. Schematic representation of plasmonic microneedle array. (a) Array of polymer microneedles that are coated with gold nanorods (AuNRs) that are further functionalized with 4-MBA. This molecule is deprotonated at high pH and the intensity of the carboxylate vibrational mode in SERS can be correlated with the pH of the sensing environment. (b) The pH-sensitive molecule 4-MBA can be protonated at an acidic pH (pH 2) and deprotonated at a basic pH (pH 12).




Figure 2.

Schematic for the fabrication of the plasmonic microneedle array. A commercial PDMS mold (blue) is filled with the NOA prepolymer (orange) and cured with UV exposure (365 nm) to give the polymer microneedle array (gray). Following ozone treatment, the AuNRs (yellow) were deposited onto the surface of the array in the presence of the Raman reporter molecule 4-MBA resulting in situ AuNRs aggregation and formation of a self-assembled monolayer of 4-MBA on the AuNRs.




Figure 3.

Imaging and optical characterization of plasmonic microneedle array. (a) Optical image of the microneedle array (scale bars are 500 and 100 μm (inset)), (b) scanning electron micrograph of the microneedle array (scale bars are 200 and 100 μm (inset)), (c) scanning electron micrograph of AuNR aggregates on the side of a microneedle tip (1 μm scale bar) with contrast-adjusted scanning electron micrograph of an individual aggregate (inset, 100 nm scale bar), (d) light extinction spectra of microneedle (MN) array with (red trace) and without (black trace) AuNRs.




Figure 4. SERS activity of the plasmonic microneedle array. (a) SER spectra of 4-MBA on the plasmonic microneedle array in Britton-Robinson buffers (pH 2–12) with 1 pH unit increments. (b) Magnified spectra centered around the pH-sensitive 1400 cm^{-1} peak. For each pH level, seven spectra were collected from separate microneedle tips and averaged. (c) Calibration curve for the plasmonic microneedle array (MN, red trace) and for a standard SERS substrate (AuFON, black trace) over the entire pH 2–12 range. Error bars represent standard deviation. For MNs, seven spectra per pH increment were collected and averaged from three different samples ($n = 3$). For the AuFON, five spectra were collected from different spots across one sample per pH increment ($n = 5$). The parameters for the SER collection were $\lambda_{\text{ex}} = 785 \text{ nm}$, $t_{\text{acq}} = 1 \text{ min}$, $P_{\text{ex}} = 1 \text{ mW}$ for MNs, and $\lambda_{\text{ex}} = 785 \text{ nm}$, $t_{\text{acq}} =$

1 min, $P_{\text{ex}} = 270 \mu\text{W}$ for AuFON. (d) Calibration curve for the linear response range (pH 5–9) for the plasmonic microneedle array collected with 0.2 pH unit increment. (e) Cycling the pH (pH 2–12) of the plasmonic microneedle array environment demonstrating the reversibility of the sensor (six cycles).




Figure 5. Mechanical and thermal stability of plasmonic microneedle arrays. (a) SER spectra of 4-MBA on a plasmonic microneedle array before puncturing an agar gel skin phantom (black trace), while punctured (blue trace), and after puncturing (red trace). (b) SER spectra of 4-MBA on a plasmonic microneedle array before (black trace) and after (red trace) 10 punches into an agar gel skin phantom. (c) SER spectra of 4-MBA on a plasmonic microneedle array immediately after fabrication (black trace) and on a separate plasmonic microneedle array that was incubated for one month in 1× PBS (green trace). (d) Comparison plot of A_{cbx}/A_{1076} of MN array with AuNRs functionalized with 4-MBA inserted in human skin versus in DPBS solutions with pH 6.6, 7.1, and 7.6. Each point is an average of seven different microneedle tips, the error bars are the standard deviation at each pH, and the measurements

were done by focusing the laser through the polymer and on the tips. The parameters for the SERS data acquisitions were $\lambda_{\text{ex}} = 785 \text{ nm}$, 20 \times ELWD objective, $t_{\text{acq}} = 1 \text{ min}$, $P_{\text{ex}} = 1 \text{ mW}$.

Author Manuscript

Author Manuscript

Author Manuscript

Author Manuscript