

Trapped Embolic Protection Device: A Salvage Technique

Anna Luisa Kuhn¹, Ajit S. Puri², Katyucia De Macedo Rodrigues², Francesco Massari², Jasmeet Singh³

1. Department of Neurointerventional Radiology, University of Massachusetts, Worcester, USA 2. Division of Neuroimaging and Intervention, Department of Radiology, University of Massachusetts, Worcester, USA 3. Division of Neurointerventional Radiology, University of Massachusetts, Worcester, USA

Corresponding author: Jasmeet Singh, jasmeet.singh@umassmemorial.org

Abstract

After carotid artery stenting, retrieval of the embolic protection device can sometimes be difficult due to incomplete stent expansion, stent fracture, vasospasm, and vessel tortuosity. In this technical report, we describe a novel rescue technique used in a patient with diffuse calcific atherosclerosis of the left common and proximal left internal carotid arteries who underwent left internal carotid artery stenting with cerebral protection and in whom, due to an under-expanded proximal carotid stent strut in relation to a densely calcified plaque, we were initially unable to advance the retrieval device.

Categories: Neurology, Radiology, Neurosurgery

Keywords: carotid stent, carotid arterial disease, carotid stenting, internal carotid artery (ica), carotid angioplasty, carotid calcification, carotid atherosclerosis

Introduction

Carotid artery stenting is the preferred treatment for patients with symptomatic carotid stenosis and who are considered “high risk” for carotid endarterectomy [1]. The use of embolic protection devices (EPDs) is standard of care in carotid stenting and minimizes the risk for procedure-related strokes [2]. Retrieval of the EPD can sometimes be challenging, especially in the setting of incomplete stent expansion, stent fracture, vasospasm, and vessel tortuosity. Having a rescue plan in such scenarios is crucial.

Technical Report

A middle-aged patient presented to our clinic with a past medical history of multiple prior strokes, hypertension, hyperlipidemia, chronic kidney disease, and chronic obstructive pulmonary disease (COPD). The patient developed stroke symptoms while in rehabilitation for a fractured hip after a fall. On work-up, the patient was found to have extensive atherosclerotic changes throughout the cervical vasculature, including bilateral severe to critical internal carotid artery (ICA) stenoses, right vertebral artery occlusion and moderate stenosis of the left vertebral artery origin (Figure 1A and 1B). MRI of the brain revealed multiple small infarcts in both cerebral hemispheres (Figure 1C-E).

Received 06/22/2020

Review began 06/28/2020

Review ended 07/06/2020

Published 07/16/2020

© Copyright 2020

Kuhn et al. This is an open access article distributed under the terms of the Creative Commons Attribution License CC-BY 4.0., which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

How to cite this article

Kuhn A, Puri A S, De Macedo Rodrigues K, et al. (July 16, 2020) Trapped Embolic Protection Device: A Salvage Technique. Cureus 12(7): e9228. DOI 10.7759/cureus.9228




FIGURE 1: Bilateral vertebral artery disease and bihemispheric anterior circulation infarcts

A, B - CT angiography coronal and sagittal reconstructions: occlusion of the right vertebral artery at its origin with visualization of a vessel stump (A, arrow). Moderate stenosis at the origin of the left vertebral artery due to mixed calcified and non-calcified plaque (B, arrow). C, D, E - MRI DWI axial slices: magnetic resonance imaging of the brain reveals multiple small infarcts in both cerebral hemispheres (C-E).

The patient underwent right carotid endarterectomy but unfortunately developed hoarseness postoperatively due to vocal cord palsy. The patient also had a preexisting type two aortic dissection. For the treatment of her high-grade left internal carotid artery (ICA) stenosis, she opted for minimally invasive endovascular stenting. The procedure was performed under monitored anesthesia care (MAC) using dexmedetomidine (Precedex™) with continuous cardiovascular monitoring. A 65 cm long 6 French Super Arrow-Flex® sheath (Teleflex®, Morrisville, NC) was placed in the proximal descending aorta and a 6 French Envoy® guiding catheter (DePuy Synthes, Raynham, MA) was used to select the origin of the left common carotid artery (CCA). There was a severe diffuse atherosclerotic disease of the left CCA and ICA with extensive irregular atherosclerotic plaque formation and multiple plaque ulcers (Figure 2A). Two areas of critical, more than 90% stenosis, were noted in the proximal left CCA and left ICA origin. The Envoy catheter was successfully advanced beyond the origin of the left CCA. Through this catheter, a 5 mm Emboshield NAV™ embolic protection system (Abbott Vascular,

Santa Clara, CA) was carefully navigated into the high cervical left ICA and deployed uneventfully. A 4.5 mm x 20 mm Monorail™ over-the-wire balloon (Boston Scientific, Natick, MA) was advanced over the EPD wire to perform pre-stenting angioplasty of the CCA and proximal left cervical ICA (Figure 2B and 2C).



FIGURE 2: Left CCA and ICA anatomy, EPD placement and pre-stent balloon angioplasty

Severe calcific atherosclerotic disease of the left CCA and ICA (A). Successful placement of an embolic protection device in the distal cervical left ICA and pre-stent angioplasty (B and C).

CCA - common carotid artery; ICA - internal carotid artery; EPD - embolic protection device

An 8 x 36 mm Carotid Wallstent® (Boston Scientific) was then deployed in the left cervical ICA, spanning from the C3 to C7 level and extending into the left CCA (Figure 3A). Due to the previously described atherosclerotic plaque in the CCA, it was noted that the proximal portion of the stent did not appropriately open (Figure 3B, arrow). Attempts to advance the 4.5 mm x 20 mm Monorail over-the-wire balloon (Boston Scientific) to perform angioplasty at the level of the proximal stent and open the metal struts were unsuccessful. The balloon was removed, and a docking Traxcess® wire (Traxcess 14, Microvention® Terumo, Tustin, CA) was connected to the end of the monorail wire from the EPD. Serial angioplasty of the stent was then performed at the proximal opening of the stent using 1.5 mm x 9 mm (Figure 3C, arrow) and 3 mm x 15 mm Gateway over-the-wire balloons (Stryker® Neurovascular, Kalamazoo, MI) (Figure 3D).




FIGURE 3: Left ICA stenting and post-stent balloon angioplasty

Placement of an 8 x 36 mm Carotid Wallstent (Boston Scientific) spanning from the C3 to C7 level and extending into the left CCA (A). Incomplete opening of the proximal stent struts (B, arrow) and serial angioplasty of the proximal stent using a 1.5 mm x 9 mm (C, arrow) and 3 mm x 15 mm Gateway over-the-wire balloons (Stryker Neurovascular) (D).

ICA - internal carotid artery; CCA - common carotid artery

Despite successful angioplasty, the capture device for the EPD would not advance over the wire. We then advanced a 125 cm soft-tipped 5 French Sofia® intermediate/distal access catheter (MicroVention® Terumo) (Figure 4A, arrow) and used buddy wire technique with a 0.035-inch Terumo Glidewire® (Terumo Medical, Somerset, NJ) by the side of the embolic protection device wire (Figure 4B, arrow) to pass through the stent and successfully capture the embolic protection device. Once the EPD was removed, the deployed stent was re-accessed with a microwire.




FIGURE 4: Capture of the trapped EPD

As the capture device was still not able to enter the stent, a soft-tipped 5 French Sofia intermediate/distal access catheter (MicroVention Terumo) was advanced into the stent (A, arrow) using "buddy-wire" technique with a 0.035-inch Terumo Glidewire (Terumo Medical, Somerset, NJ) by the side of the embolic protection device wire (B, arrow).

EPD - embolic protection device

Two 8 x 29 mm overlapping Carotid Wallstents (Boston Scientific) were then placed across the severely diseased left CCA (Figure 5A-C) up to its origin in the aorta and angioplasty of the stented vessel segment was performed using a 5 mm x 20 mm Monorail balloon (Boston Scientific). The patient tolerated the procedure well and remained neurologically intact. After one night of in-hospital observation, the patient was discharged home the next day on a dual antiplatelet regimen consisting of ASA 81 mg and Plavix® 75 mg daily.



FIGURE 5: Left CCA stenting

After the recapture of the embolic protection device, placement of two additional overlapping 8 x 29 mm Carotid Wallstents (Boston Scientific) to cover the diseased left CCA (A-C). The final angiogram upon completion of the procedure is shown in image C.

CCA - common carotid artery

Discussion

There have been descriptions of rescue strategies to retrieve trapped embolic protection filters including external carotid compression, rotation of the head to the contralateral side to change carotid angulation, use of different catheters like the vertebral and JR-four catheters and carotid endarterectomy (CEA) [3-8]. CEA for EPD retrieval can be performed safely in an emergent setting [7, 8]. However, rescue strategies are mainly described in case reports and, therefore, the optimal management of a retained EPD remains a challenge. Calcified plaques and vessel tortuosity are independent risk factors associated with the difficult retrieval of embolic protection devices [9]. Both of these factors were present in our case. Also, since our guiding catheter position was very proximal, we anticipated some difficulty during device retrieval and, therefore, had a backup action plan in mind before initiating the procedure. Dauherty et al. encountered a similar case scenario in which they used a 5 French 125 cm vertebral catheter to recapture and remove a trapped Emboshield™ embolic protection device (Cordis Endovascular, Hialeah, FL) [4]. They had to advance the 125 cm 5 French vertebral catheter over the 180 cm Emboshield wire. Because only approximately 60 cm of the 180 cm wire was exposed outside the guiding sheath, they were unable to pin or hold the wire as the diagnostic catheter was advanced. Our technique offers more control as the attachment of the Traxcess docking wire (Microvention Terumo) essentially changes a Monorail system into an exchange microwire system. Use of this system allowed us to utilize the smaller over-the-wire Gateway balloon for angioplasty of the stent struts. This ultimately allowed us to pass the 5 French Sofia intermediate/distal access catheter and capture the EPD.

Conclusions

Having a "plan B" can make all the difference between a successful and unsuccessful procedural outcome. The navigability of a distal access catheter makes it a useful tool not only for intracranial interventions but also as a rescue tool to recapture a trapped EPD. Our rescue technique has not been described in the literature and could prove very useful in difficult situations during carotid artery stenting. Other helpful considerations include the use of

docking wires which provide extra length to EPD wires and the "buddy-wire" technique which can help provide stability and guidance for a catheter.

Additional Information

Disclosures

Human subjects: Consent was obtained by all participants in this study. University of Massachusetts Medical School IRB issued approval H00001860_10. **Animal subjects:** All authors have confirmed that this study did not involve animal subjects or tissue. **Conflicts of interest:** In compliance with the ICMJE uniform disclosure form, all authors declare the following: **Payment/services info:** All authors have declared that no financial support was received from any organization for the submitted work. **Financial relationships:** Ajit S. Puri declare(s) a grant and personal fees from Stryker. Ajit S. Puri declare(s) a grant and personal fees from Medtronic. Ajit S. Puri declare(s) personal fees from Cerenovus. Ajit S. Puri declare(s) personal fees from CereVasc . Ajit S. Puri declare(s) personal fees from Merit. Ajit S. Puri declare(s) stock/stock options from InNeuroCo. **Other relationships:** All authors have declared that there are no other relationships or activities that could appear to have influenced the submitted work.

References

1. Kumar PV, Lakshmi A, Shrivastava R, et al.: Protected carotid artery stenting in patients at high risk for carotid endarterectomy. *South Med J*. 2011, 104:257-263. [10.1097/SMJ.0b013e31820d8e39](https://doi.org/10.1097/SMJ.0b013e31820d8e39)
2. Zahn R, Mark B, Niedermaier N, et al.: Embolic protection devices for carotid artery stenting: better results than stenting without protection?. *Eur Heart J*. 2004, 25:1550-1558. [10.1016/j.ehj.2004.06.018](https://doi.org/10.1016/j.ehj.2004.06.018)
3. Nii K, Nakai K, Tsutsumi M, et al.: A manual carotid compression technique to overcome difficult filter protection device retrieval during carotid artery stenting. *J Stroke Cerebrovasc Dis*. 2015, 24:210-214. [10.1016/j.jstrokecerebrovasdis.2014.08.022](https://doi.org/10.1016/j.jstrokecerebrovasdis.2014.08.022)
4. Daugherty WP, White JB, Cloft HJ, Kallmes DF: Rescue retrieval of angioguard embolic capture system after failure of capture sheath retrieval in carotid angioplasty and stenting. *Am J Neuroradiol*. 2008, 29:1594-1595. [10.3174/ajnr.A1118](https://doi.org/10.3174/ajnr.A1118)
5. Ganim RP, Muench A, Giesler GM, Smalling RW: Difficult retrieval of the EPI filterwire with a 5 french FR4 coronary catheter following carotid stenting. *Catheter Cardiovasc Interv*. 2006, 67:309-311. [10.1002/ccd.20525](https://doi.org/10.1002/ccd.20525)
6. Page P, Niemann D, Son C, Li Y: Retained distal protection device during carotid artery stenting necessitating carotid endarterectomy: a complication and management considerations. *Surg Neurol Int*. 2018, 9:123. [10.4103/sni.sni_32_18](https://doi.org/10.4103/sni.sni_32_18)
7. Li T, Zha Y, Bo L, Wirthlin DJ, Zhang Q: Surgical management for retained distal embolic protection device and fractured guidewire after carotid artery stenting. *J Surg Case Rep*. 2016, 2016:105. [10.1093/jscr/rjw105](https://doi.org/10.1093/jscr/rjw105)
8. Shilling K, Uretsky BF, Hunter GC: Entrapment of a cerebral embolic protection device--a case report. *Vasc Endovascular Surg*. 2006, 40:229-233. [10.1177/153857440604000308](https://doi.org/10.1177/153857440604000308)
9. Lian X, Liu W, Li M, et al.: Risk factors and complications associated with difficult retrieval of embolic protection devices in carotid artery stenting. *Cardiovasc Intervent Radiol*. 2012, 35:43-48. [10.1007/s00270-011-0117-y](https://doi.org/10.1007/s00270-011-0117-y)