

ORIGINAL ARTICLE

Electrical stimulation therapy for the treatment of pressure ulcers in individuals with spinal cord injury: a systematic review and meta-analysis

Deena Lala¹, Sandi J Spaulding^{1,2}, Shauna M Burke³ & Pamela E Houghton^{1,4}

1 Health and Rehabilitation Sciences, Western University, London, ON Canada

2 School of Occupational Therapy, Western University, London, ON Canada

3 School of Health Studies, Western University, London, ON Canada

4 School of Physical Therapy, Western University, London, ON Canada

Key words

Electrical stimulation therapy;
Meta-analysis; Pressure ulcers; Spinal cord injuries; Systematic review

Correspondence to

P Houghton, PhD
School of Physical Therapy
Western University
Room 1458, Elborn College
1201 Western Road
London
Ontario N6G 1H1
Canada
E-mail: phoughto@uwo.ca

doi: 10.1111/iwj.12446

Lala D, Spaulding SJ, Burke SM, Houghton PE. Electrical stimulation therapy for the treatment of pressure ulcers in individuals with spinal cord injury: a systematic review and meta-analysis. *Int Wound J* 2016; 13:1214–1226

Abstract

To conduct a systematic review and meta-analysis on the effects of electrical stimulation therapy (EST) on healing pressure ulcers in individuals with spinal cord injury (SCI). CINAHL, The Cochrane Library, PubMed, SCOPUS, EMBASE, Nursing & Allied Health and Dissertation & Theses databases were searched for relevant English language articles from the date of inception to 31 January 2014. Separate searches were conducted in Google Scholar and academic journals specialised in wound care. Two reviewers independently assessed study eligibility. Studies were included if EST was used to treat pressure ulcers in individuals with SCI. A total of 599 articles were screened, and 15 studies met the inclusion criteria. A meta-analysis with five studies demonstrated that EST significantly decreased the ulcer size by 1.32%/day [95% confidence interval (CI): 0.58–2.05, $P < 0.001$] compared to standard wound care (SWC) or sham EST. Another meta-analysis conducted with four studies showed that EST increased the risk of wound healing by 1.55 times compared with standard wound care or sham EST (95% CI: 1.12 to 2.15, $P < 0.0001$). Because of the wide array of outcome measures across studies, a single meta-analysis could not be conducted. EST appears to be an effective adjunctive therapy to accelerate and increase pressure ulcer closure in individuals with SCI.

Introduction

Persons with spinal cord injury (SCI) experience a number of secondary health complications over their lifetime, with pressure ulcers being the most common. Despite the numerous prevention and management recommendations available (1,2), it has been estimated that 31–40% of individuals with SCI develop pressure ulcers in a given year (3–5) and up to 85% of them develop at least one pressure ulcer over the course of their lifetime (6). In addition to having a significant impact on the health care system, pressure ulcers are also associated with serious health implications for individuals with SCI including reductions in physical activity levels, social participation and overall quality of life (7).

Current recommendations for managing pressure ulcers involve standard wound care (SWC), which includes practices such as debridement, dressing, nutrition counselling and

Key Messages

- systematic review and meta-analysis on the effect of electrical stimulation therapy (EST) for pressure ulcer treatment
- fifteen studies were identified; of these, only eight were included in the meta-analysis
- the overall methodological quality of the studies included in the meta-analysis was identified as low. Heterogeneity between the studies was significant and may be attributed to lack of standardised method of measuring pressure ulcer size, different EST types and parameters and different treatment schedules
- EST was found to be an effective adjunctive therapy for treating pressure ulcers in individuals with spinal cord injury (SCI)

physical and occupational therapy (8). However, pressure ulcers may not always respond positively to SWC alone. Electrical stimulation therapy (EST) is an adjunctive approach to SWC that delivers low levels of electric current to the wound bed to increase the rate of wound healing (9). It has been suggested that EST accelerates wound healing by mimicking the natural electrical current of the skin when it is injured (9). EST has been shown to affect most phases of wound healing including inflammation, proliferation and remodelling (9). *In vitro* studies have demonstrated that EST can: induce fibroblastic activity by increasing DNA and protein synthesis (10), as well as calcium influx (11); promote migration and activation of macrophages (12); and promote myofibroblast transdifferentiation (13). Furthermore, *in vivo* studies have demonstrated that EST can: enhance angiogenesis by increasing capillary density, improve tissue oxygenation (14) through increased blood flow (15,16) and improve tissue tensile strength (17,18) by increasing collagen deposition (19). In addition, researchers have suggested that EST has both bacteriostatic and bactericidal effects on microbes that commonly infect wounds (20–22).

Earlier, systematic reviews (23,24) and meta-analyses (25–27) have been carried out to examine the healing potential of EST in various types of chronic wounds. The three meta-analyses (25–27), which each included data from a different group of publications, concluded that the application of EST as an adjunctive therapy for treating ulcers accelerates the rate of wound healing and promotes wound closure compared with SWC alone. Results from two systematic reviews (23,24) suggest there are quite a number of randomised controlled trials that have examined the effect of EST on healing of pressure ulcers. However, there have been no meta-analyses that combine the results and estimates of the pooled effect of EST on pressure ulcers. Furthermore, results from published meta-analyses involving a variety of wound types may not be transferable to the SCI population. Recent research has shown that the levels of various proteins including growth factors, cytokines and enzymes are quite different in pressure ulcers in people with SCI compared with those without SCI (28). Specifically, levels of pro-inflammatory cytokines such as IL-6 and IL-2, and cellular adhesion molecules (i.e. ICAM-1), are significantly higher in individuals with SCI (29,30), while concentrations of tissue inhibitor of metalloproteinases (TIMPs), matrix metalloproteinases (MMPs) and wound fluid proteins are significantly lower in individuals with SCI compared with those without SCI (28). Therefore, because of the differences in blood serum and wound biochemistry, it may be the case that the response to EST in pressure ulcer healing in individuals with SCI varies from that in the general population.

The purpose of this study was to conduct a systematic review and meta-analysis to determine the effectiveness of EST on the healing of pressure ulcers in individuals with SCI, specifically, in comparison with control groups. We also aimed to understand the patients' ability to adhere to the treatment schedule, the implications of EST on the quality of life and pain relief in this population and the economic impact on the health care system.

Table 1 Inclusion and exclusion criteria of the study

Inclusion criteria	Exclusion criteria
<ul style="list-style-type: none"> • EST designed to improve healing • EST using surface electrodes in or around the wound • All clinical trials • Open, full-thickness pressure ulcers • Men and women aged 18 years or older • Traumatic or non-traumatic SCI • Acute or chronic SCI • English articles 	<ul style="list-style-type: none"> • Electromagnetic fields • EST for nerve innervation [transcutaneous electrical nerve stimulation (TENS)] or muscle activation [functional electrical stimulation (FES)] • Use of indwelling electrodes • More than 50% of included ulcers caused by other aetiologies (e.g. diabetes mellitus, DVP) • Non-English language articles • Abstracts

EST, electrical stimulation therapy; SCI, spinal cord injury.

Methods

Types of studies

This systematic review included only studies where EST was applied to individuals with SCI and pressure ulcers. Clinical controlled trials comparing EST intervention (alone or in conjunction with SWC) with sham EST or SWC alone were examined for inclusion in the meta-analysis, while non-controlled trials (i.e. case studies) were described narratively. A summary of the inclusion and exclusion criteria used to select the studies included for the review is presented in Table 1.

Types of participants

Studies with men and women above the age of 18 years wherein at least 50% of participants had with acute or chronic SCI and the presence of pressure ulcers were considered for inclusion in the review. Pressure ulcers were defined as open, full-thickness skin lesions that were likely caused by pressure or other external forces, such as friction or shear.

Types of interventions

Studies that involved conductively coupled EST, which is applied using two or more surface electrodes placed on the wound bed or in the surrounding area, were included in this study. There were four distinct EST modalities examined:

- Low intensity direct current (LIDC) involves applying continuous, unidirectional flow of current of low intensity (<1 mA) for at least one second (9).
- Monophasic pulsed current (MPC) involves brief pulses of unidirectional flow of current followed by a finite off period (9). Common MPC waveforms include rectangular and twin-peaked [e.g. high-voltage pulsed current (HVPC)].
- Biphasic pulsed current (BPC) consists of brief pulses of bidirectional current that has either a symmetric or asymmetrical biphasic waveform. In symmetric BPC, the bidirectional pulsed current is equal and balanced, while

asymmetric BPC produces a bidirectional current that is unequal and may or may not be balanced. Balanced asymmetric BPC has no net positive or negative charge, while unbalanced asymmetric BPC creates a net positive or negative charge over time (9).

- Microcurrent (MC) includes MPC or BPC that provides current at a subsensory level.

All variations of frequency, amplitude and duration of EST were included. Control conditions in the studies included sham EST (i.e. power source disconnected, leads cut) or SWC. Studies where subjects were assigned to receive different EST protocols were also included by either totalling the number of healed ulcers or by taking the healing rate of the EST protocol with the largest sample of pressure ulcers.

Outcome measures

The primary outcome that was considered was wound healing, operationally defined as rate of change in ulcer size, absolute change in wound size and number of wounds completely healed post EST. Secondary outcomes including EST-related adverse events, quality of life, pain relief, economic analysis and EST compliance were collated and summarised narratively.

Data sources

Electronic searches were performed by one author using the following databases: CINAHL (Cumulative Index to Nursing and Applied Health Literature, 1981); The Cochrane Library (2009); Dissertation & Theses (1861); EMBASE (1947); Proquest – Nursing & Allied Health (1988); PubMed (1946) and SCOPUS (1966) from their date of availability through to 31 January 2014. The following keywords were used in each database: *spinal cord injury, spinal cord injuries, spinal cord, tetraplegia, quadriplegia, paraplegia, pressure ulcer, pressure sore, bedsore, decubitus ulcer, wound healing, electrical stimulation therapy, direct current, alternating current, galvanic current, monophasic current, biphasic current, pulsed current*. Articles not available via the university library, inter-library loan system or published in a language other than English were excluded. An experienced librarian was consulted to develop an appropriate search strategy for each database.

Separate searches (electronic and manual) were conducted in journals specialised in wounds including *Advances in Wound Care* (1988), *Advances in Skin and Wound Care* (2010), *International Wound Journal* (2004), *Ostomy Wound Management* (2001), *Wounds: A Compendium of Clinical Research and Practice* (1990) and *Wound Repair and Regeneration* (1993) from their date of availability through to 4 February 2014.

The reference lists of pertinent review articles and eligible studies were searched for additional relevant studies that were missed in the database search. Grey literature (i.e. articles not published through journals) was identified through Google Scholar (<http://scholar.google.ca>) and speaking with local experts in the field.

Study selection

Two reviewers screened the titles and abstracts (citations) independently. A paired consensus process was used to select relevant citations. Disagreements between reviewers were discussed until consensus was achieved.

Full articles were retrieved from all relevant citations and for those that were unclear or had missing abstracts from the database citation. Full-text articles were then reviewed and assessed for potential inclusion in the study. Paired consensus was repeated to confirm article eligibility. Any disagreement between the two reviewers was resolved through discussion.

Data extraction

One reviewer reviewed each article and extracted data using a standard data extraction sheet. The information extracted from each study included study design, participant characteristics (i.e. age, sex and SCI characteristics), pressure ulcer characteristics (i.e. stage, location and duration of ulcer), description of EST protocol used (i.e. type, waveform, pulse frequency, intensity, duration, polarity and electrode placement), treatment schedule (i.e. frequency and total treatment time) and primary and secondary outcomes. Crossover data were not examined. A second author checked the accuracy of all the data extracted.

Methodological quality assessment

The methodological quality of the clinical controlled trials was assessed using the PEDro (Physiotherapy Evidence Database) scale. The PEDro scale is based on a Delphi list developed by Verhagen *et al.* (31) and consists of 11 items. The first item on the PEDro scale (eligibility criteria) assesses external validity via a 'yes' or 'no' response, and the next 10 items assess the internal validity of a trial. Items 2 through 11 are scored out of 1, resulting in a total possible score of 10. The PEDro scale has demonstrated high construct validity (32) and good inter-rater reliability (33). One reviewer independently assessed the methodological quality of each study.

Data synthesis

All data were analysed using Comprehensive Meta-Analysis 2.0 (CMA; Biostat, NJ). Pooled analysis was conducted for trials that expressed healing as a rate [percent per day (%/day)] and number of pressure ulcers healed. No computations of raw data were performed for any of the trials. Dichotomous outcomes were expressed as risk ratio (RR) with 95% confidence interval (CI). RR was interpreted as follows: RR < 1, risk of healing is lower in EST group; RR = 1, risk of healing is same between EST and control groups; and RR > 1, risk of healing is greater in the EST group. Hedge's *g* was used to interpret the effect size of continuous variables. Hedge's *g* is a variation of Cohen's *d* that corrects for bias caused as a result of sample size (34). The criteria used for Cohen's *d* (35) were used to interpret Hedge's *g*: small, 0.2; moderate, 0.5; and large, 0.8. Treatment effect was significant if $P < 0.05$. Heterogeneity between the studies was determined using I^2 and P -value; I^2 exceeding 50% was used as the threshold to identify significant heterogeneity.




Figure 1 Study flow chart (PRISMA).

A fixed effects model was used when the threshold for heterogeneity was not reached, and a random effects model was used when the threshold for heterogeneity was exceeded (36). Funnel plots and Orwin's fail-safe N test were to be used to assess publication bias if more than 10 trials were included in the meta-analysis (37), although this was not possible because of the small number of trials in each pooled analysis.

Results

Results of the systematic review

A total of 599 citations including title and abstract were identified using electronic databases ($n = 498$) and secondary sources ($n = 101$), of which 507 citations were excluded. Full text was reviewed for 49 articles; 34 were subsequently excluded.

Reasons for exclusion were study participants were not individuals with SCI ($n = 13$), the articles were conference abstracts (i.e. full-text not available; $n = 8$), studies where <50% of participants had SCI ($n = 5$) or pressure ulcers ($n = 1$), there were no wound healing outcomes reported ($n = 5$) or EST was delivered using indwelling electrodes ($n = 2$). A PRISMA flow diagram is presented in Figure 1.

Inclusion criteria were met by 15 studies (Tables 2 and 3); six were randomised controlled clinical trials (RCTs) (38–43), three prospective controlled trials (44–46), four case studies (47–50) and two retrospective controlled trials (51,52). More than 500 persons with SCI were included in these reports, in which 214 patients were included in a large retrospective analysis (52). Six of the published reports were generated from a group of researchers from Slovenia. It appears that the same data are presented in one of the publications (46), and it is

Table 2 Summary of the clinical controlled studies included in the meta-analysis

Author, year, country	Study design; treatment; control	Number of patients; incomplete/complete (I/C); DOI	Number of pressure ulcers; duration of ulcers (DOU); Stage (NPUAP)	Electrode placement; EST waveform; stimulus parameter; treatment schedule	Results	
					Control	ES
Baker et al. (38), USA	RCT TRT ₁ : Asym BPC TRT ₂ : Sym BPC CON ₁ : MC CON ₂ : Sham MC	Asym BPC, n=20 Sym BPC, n=21 MC, n=20 Sham MC, n=19 I/C: 19/52 DOI: 1–420 months	Asym BPC, n=67 Sym BPC, n=58 MC, n=42 Sham MC, n=25 DOU: 2–1095 days Stage: n/a	Electrodes applied on opposite sides on the wound edge, active electrode more proximal Asym BPC: rectangular, 100 microseconds, 50 Hz, 7:7 on/off ratio Intensity: submotor Sym BPC: rectangular, 300 microseconds, 50 Hz, 7:7 on/off ratio Intensity: submotor MC: rectangular, 10 microseconds, 1 Hz, 4 mA Intensity: subsensory 30 minutes x 3/day x 5 days/weeks until healed	MC: 23.3 ± 4.8%/week, 18 healed Sham MC: 32.7 ± 7.0%/week, 6 healed Asym BPC: 36.4 ± 6.2%/week, 35 healed Sym BPC: 29.7 ± 5.1%/week, 33 healed	WHR significantly faster in Asym BPC versus MC and sham MC
Barczak et al. (44), Germany	PCT TRT: MPC CON: Sham MPC	MPC, n=10 Sham MCP, n=14 I/C: 6/18 DOI: n/a	MPC, n=16 Sham MPC, n=17 DOU: <1 to >6 months Stage: 3–5 according to Daniel	Active electrode applied directly over wound MPC: rectangular, 140 microseconds, 38 mA Intensity: n/a 60 minutes/day x 7 days/weeks x 4 weeks Active electrode applied directly over wound, dispersive applied on intact skin HVPC: 75 microseconds, 100 Hz, 500 µA Intensity: subsensory 60 minutes/day x 7 day/week x 2.9 weeks	1.1 ± 0.96%/day, 65% of initial WSA 2.0 ± 1.47%/day, 38% of initial wound size	WHR significantly faster in EST versus CON
Griffin et al. (39), USA	RCT TRT: HVPC CON: Sham HVPC	HVPC, n=8 Sham HVPC, n=9 I/C: 3/14 DOI: 3–1820 weeks	HVPC, n=8 Sham HVPC, n=9 DOU: 1–116 weeks Stage: II–IV	Active electrode applied directly over wound, dispersive applied on intact skin Intensity: subsensory 60 minutes/day x 7 day/week x 2.9 weeks	52% ↓ in WSA, 2 healed 80% ↓ in WSA, 3 healed	Significant ↓ in WSA in EST versus CON
Houghton et al. (40), Canada	RCT TRT: HVPC CON: SWC	HVCP n=16 SWC, n=18 I/C: n/a DOI: 1–51 years	HVCP, n=16 SWC, n=18 DOU: 0.3–20 years Stage: II–IV or unstageable	Active electrode applied directly over wound, dispersive applied on intact skin (n=11); electrodes applied on opposite sides of wound edge (n=2); electroconductive sock (n=3) HVPC: 50 microseconds, 10–100 Hz Intensity: submotor 240 minutes/day x 7 day/week x 13 weeks	36 ± 61% ↓ in WSA, 5 healed 70 ± 25% ↓ in WSA, 6 healed	Significant ↓ in WSA in EST versus CON
Jerčinič et al. (41), Slovenia	RCT TRT: Asym BPC CON: SWC	Asym BPC, n=42 SWC, n=31 I/C: n/a DOI: 32 ± 60 months	Asym BPC, n=61 SWC, n=48 DOU: 143.5 ± 229.2 days Stage: n/a	Electrodes applied on opposite sides on the wound edge Asym BPC: balanced, 250 microseconds, 40 Hz, 4:4 on/off ratio Intensity: motor 120 minutes/day x 5 days/week x 4 weeks	2.7 ± 3.6%/day 5.7 ± 7.1%/day	WHR significantly faster in EST versus CON

Table 2 Continued

Author, year, country	Study design; treatment; control	Number of patients; incomplete/complete I/C; DOI	Number of ulcers; duration of ulcers (DOU); Stage (NPUAP)	Electrode placement; EST waveform; stimulus parameter; treatment schedule	Results		Study conclusion
					Control	ES	
Karba et al. (45), Slovenia	PCT	LIDC+, n = 16	LIDC+, n = 16	LIDC+: Active electrode applied directly over wound, dispersive applied on intact skin, 0.6 mA	4.2 ± 1.1%/day	LIDC+: 7.4 ± 1.6%/day LIDC+/-: 4.8 ± 1.5%/day	WHR significantly faster in LIDC+ versus CON
	TRT ₁ : LIDC+	LIDC+/-, n = 18	LIDC+/-, n = 18	Intensity: subsensory			
	TRT ₂ : LIDC+/- CON: Sham LIDC+/-	Sham LIDC+/-, n = 16 I/C: n/a DOI: n/a	Sham LIDC+/-, n = 16 DOU: n/a Stage: III-IV	LIDC+/-: Electrodes applied on opposite sides of wound edge, 0.6 mA Intensity: subsensory 120 minutes/day x 7 days/week until healed			
Karba et al. (42), Slovenia	RCT	Aysm BPC, n = 6	Aysm BPC, n = 6	Electrodes applied on opposite sides of wound edge	-0.66 ± 1.16%/day, 0 healed	7.13 ± 1.46%/day, 6 healed	WHR significantly faster in EST versus CON
	TRT: Aysm BPC	Sham BPC, n = 6	Sham BPC, n = 6	Asym BPC: balanced, 250 microseconds, 4:4 on/off ratio			
	CON: Sham BPC	I/C: n/a DOI: n/a	DOU: n/a Stage: n/a	Intensity: motor 120 minutes/day x 7 days/week until healed			
Stefanovska et al. (51), Slovenia	Retrospective CT	Number of patients not specified	Asym BPC, n = 82	Electrodes applied on opposite sides of wound edge	2.21 ± 3.27%/day	Asym BPC: 5.43 ± 4.40%/day LIDC: 3.11 ± 3.83%/day	WHR significantly faster in Asym BPC versus SWC and LIDC
	TRT ₁ : Asym BPC	I/C: n/a	LIDC, n = 18	Asym BPC: balanced, 250 microseconds, 40Hz, 4:4 on/off ratio			
	TRT ₂ : LIDC	DOI: 31:0 ± 59.7 months	SWC, n = 50	Intensity: submotor			
	CON: SWC	DOI: 207.9 ± 482.14 days	DOU: 207.9 ± 482.14 days Stage: n/a	LIDC: 600 µA 120 minutes/day x 7 days/week x 4 weeks			

Study design: CS, case study or case studies; CT, controlled trial; PCT, prospective controlled trial; RCS, retrospective case series; RCT, randomised controlled trial.

Treatment: Asym BPC, asymmetrical biphasic pulsed current; Sym BPC, symmetrical BPC; HVPC, high voltage pulsed current; LIDC, low intensity direct current; MPC, monophasic pulsed current.

Results: WHR, wound healing rate; WSA, wound surface area.

n/a, Not available.

Table 3 Summary of the clinical studies not included in the meta-analysis


Author, year, country	Study design; treatment; control	Number of patients: incomplete/complete (I/C); DOI	Number of pressure ulcers; Duration of ulcers (DOU); Stage (NPUAP)	Electrode placement; EST waveform; stimulus parameter; treatment schedule	Results		Study conclusion
					Control	ES	
Adegoke and Badmos (43), Nigeria	RCT TRT: MPC CON: Sham MPC	MPC, n=3 Sham MPC, n=3 I/C: n/a DOI: n/a	MPC, n=3 Sham MPC, n=3 DOU: 10 ± 2.0 weeks Stage: IV	Active electrode applied directly over wound, dispersive applied on intact skin MPC: 30Hz, 2:1 on/off ratio Intensity: submotor 45 minutes/day x 3/week x 4 weeks	2.6 ± 0.0% ↓ in WSA	22.2 ± 1.4% ↓ in WSA	Significant ↓ in WSA in EST versus CON
Allen and Houghton (47), Canada	CS TRT: MPC	MPC, n=1 I/C: 0/1 DOI: 30 years	MPC, n=1 DOU: n/a Stage: III	Active electrode applied directly over wound, dispersive applied on intact skin MPC: rectangular, 140 microseconds, 128 Hz Intensity: n/a	n/a	19.85%/week	EST facilitates wound healing
Chalker (49), USA	CS TRT: BPC	BPC, n=1 I/C: n/a DOI: n/a	BPC, n=1 DOU: 4 weeks Stage: III	30 minutes/day x 7 days/week x 12 weeks Hand-held electrode applied directly over wound BPC: not specified Intensity: n/a	n/a	Ulcer healed	EST facilitates wound healing
Cukjati et al. (52), Slovenia	Retrospective CT TRT ₁ : BPC TRT ₂ : LIDC CON ₁ : Sham LIDC CON ₂ : SWC	214 patients (71.7% SCI) I/C: n/a DOI: 1–69 months	BPC, n=181 LIDC, n=42 Sham LIDC, n=23 SWC, n=54 DOU: 2–22 mos Stage: I–IV	60 minutes/day x 7 days/week until healed BPC: Electrodes applied on opposite sides on the wound edge, 250 microseconds, 40 Hz, 4:4 on/off ratio Intensity: motor LIDC: Electrodes applied on opposite sides on the wound edge or active electrode applied over wound, four dispersive electrodes applied on edge around wound, 0.6 mA Intensity: subsensory 120 minutes/day x 7 days/week x 60 weeks	Sham LIDC: 0.162 mm/day SWC: 0.145 mm/day	BPC: 0.190 mm/day LIDC: 0.168 mm/day	WHR significantly faster in BPC versus CON
Pollack et al. (48), USA	CS TRT: MPC	MPC, n=1 I/C: n/a DOI: 10 years	MPC, n=1 DOU: 23 months Stage: IV	Electrodes applied over bilateral gluteus, hamstring, and quadriceps muscles MPC: rectangular, 400 microseconds, 60 Hz Intensity: n/a	n/a	Ulcer healed	EST facilitates wound healing
Recio et al. (50), USA	RCS TRT: HVPC	HVPC, n=3 I/C: 1/3 DOI: n/a	HVPC, n=3 DOU: 11–14 months Stage: III–IV	0.37–2.23 minutes x 2/week x 28 weeks Active electrode applied directly over wound, dispersive applied on intact skin HVPC: 10 microseconds, 100 Hz Intensity: submotor	n/a	95.58–100% ↓ in WSA	EST facilitates wound healing
Trontelj et al. (46), Slovenia	PCT TRT: Asym BPC CON: SWC	Asym BPC, n=63 SWC, n=43 I/C: n/a DOI: n/a	Asym BPC, n=63 SWC, n=43 DOU: n/a Stage: n/a	60 minutes/day x 3–5/week until healed Electrodes applied on opposite sides on the wound edge Asym BPC: balanced, 1.25 milliseconds, 40 Hz, 4:4 on/off ratio Intensity: motor 120 minutes/day x 7 days/week until healed	2.60 ± 2.59%/day	4.89 ± 3.80%/day	WHR significantly faster in EST versus CON

Study design: CS, case study or case studies; CT, controlled trial; PCT, prospective case series; RCS, retrospective case series; RCT, randomised controlled trial. Treatment: Asym BPC, asymmetrical biphasic pulsed current; Sym BPC, symmetrical BPC; HVPC, high voltage pulsed current; LIDC, low intensity direct current; MPC, monophasic pulsed current. Results: WHR, wound healing rate; WSA, wound surface area. EST, electrical stimulation therapy; n/a, not available.

Table 4 Methodological quality scores of the clinical controlled trials using the Physiotherapy Evidence Database (PEDro) scale

Study, author, year	Items in PEDro scale											Score
	1 (Y/N)	2	3	4	5	6	7	8	9	10	11	
Adegoke and Badmos (43)	Y	1	1	1	0	0	0	1	0	0	1	5
Baker <i>et al.</i> (38)	N	1	0	0	1	0	0	1	0	1	1	5
Barczak <i>et al.</i> (44)	Y	0	0	1	1	1	0	0	0	1	1	5
Cukjati <i>et al.</i> (52)	Y	1	0	1	0	0	0	1	1	1	1	6
Griffin <i>et al.</i> (39)	Y	1	0	1	1	1	1	0	0	1	1	7
Houghton <i>et al.</i> (40)	Y	1	1	1	0	0	1	1	1	1	1	8
Jerčinović <i>et al.</i> (41)	Y	1	0	0	0	0	0	1	0	1	1	4
Karba <i>et al.</i> (45)	Y	0	0	0	1	1	0	0	0	1	1	4
Karba <i>et al.</i> (42)	N	1	0	0	0	0	0	1	1	1	1	5
Stefanovska <i>et al.</i> (51)	N	0	0	1	0	0	0	1	1	1	1	5
Trontelj <i>et al.</i> (46)	N	0	0	0	0	0	0	1	1	0	1	3
Average score												5.2

1, Eligibility; 2, random allocation; 3, concealed allocation; 4, baseline similarity; 5, blinding of subjects; 6, blinding of therapists; 7, blinding of all assessors; 8, measures of key outcomes from more than 85% of subjects; 9, intention-to-treat analysis; 10, between-group statistical comparison; 11, point measures of variability; Y, yes; N, no; 1, item present; 0, item absent.

**Figure 2** Forest plot illustrating the effect of electrical stimulation therapy (EST) on healing rate expressed as percent per day compared with standard wound care or sham EST.

not clear if there is any overlap in the data used within the two retrospective analyses (51,52). Across all these studies, regardless of study design, authors consistently concluded that EST accelerated wound healing rate and/or the incidence of complete wound closure.

Methodological quality of the controlled studies

The overall methodological quality of the controlled trials was low (33). The scores ranged from 3 to 8 with a mean PEDro score of 5.2 out of a total possible score of 10 (Table 4). Six studies stated that subjects were randomly allocated to groups (38–43), whereas three did not (44–46), and two studies collated data from a retrospective analysis. Only two studies (40,43) included a description of how allocation was concealed (51,52). Blinding of participants, therapists and assessors was performed in only one study (39), and partially in two others (44,45). None of the studies included an explicit statement of the use of intention-to-treat analysis; however, five studies did not report any participant dropout or loss to follow-up (40,42,46,51,52).

In general, the controlled trials had small sample sizes. Baker *et al.* (38) reported the largest sample; however, several recurrent or multiple ulcers were randomised to different groups.

Characteristics of the studies included in the meta-analysis

Of the 15 studies, eight had data available to calculate effect sizes and were included in the meta-analysis. One RCT (43) and one retrospective controlled trial (52) were excluded from the meta-analysis because the healing outcomes could not be pooled with the other trials. One additional controlled trial (48) was excluded from the meta-analysis because similar data were reported in previous publications (51,52). Details of the patients, treatment schedule, EST parameters including electrode placement and waveform and results are presented in Table 2. The meta-analyses included more than 302 participants; one trial did not report the number of participants involved (51). A total of 360 pressure ulcers were treated with EST, while 237 received either sham EST or SWC. Four studies used BPC (38,41,42,51), three used MPC (39,40,44), two used LIDC (45,51) and only one used MC (38). There was




Figure 3 Forest plot illustrating the effect of electrical stimulation therapy (EST) on healed events compared with standard wound care or sham EST.

considerable diversity in EST stimulus parameters (i.e. waveform, intensity, polarity), electrode placement (i.e. applied directly over the wound versus across the wound on intact skin) and treatment schedules. Treatment schedules for these studies ranged from 60 to 240 minutes/day, 5–7 days/week, for 2.9–13 weeks or until healed.

Multiple outcome measures were examined between and within the studies. Five studies expressed wound healing as a rate either in percent decrease per day (41,42,44,45,51) or percent decrease per week (38), while four studies calculated the total number of ulcers healed over the study period (38–40,42). Overall percentage change in wound surface area was also expressed in two studies (39,40).

Daily percent decrease in pressure ulcer size

Two RCTs (41,42), two prospective controlled trials (44,45) and one retrospective control trial (51) were used to evaluate the effects of EST on the rate of healing (i.e. daily percent decrease) of pressure ulcers in persons with SCI. A significant large overall effect size was found in favour of EST when examining the daily percent decrease in pressure ulcer size compared with SWC or sham EST (Hedge's $g = 1.32$, 95% CI: 0.58–2.05, $P < 0.001$, Figure 2). Significant heterogeneity was evident across these trials ($I^2 = 85.3\%$, $P < 0.0001$).

Two other studies including one retrospective control trial (52) and one prospective control trial (46) evaluated the effect of EST on the rate of healing (Table 3). Cukjati *et al.* (52) treated 300 wounds with either asymmetric BPC, LIDC, sham or SWC. After 60 weeks of 30–120 minutes/day, 7 days/week, those treated with BPC healed (0.190 mm/day) significantly faster than those treated with LIDC (0.168 mm/day), sham (0.162 mm/day) and conservative therapy (0.145 mm/day). Similarly, Trontelj *et al.* (46) found a significantly faster healing rate in pressure ulcers treated with asymmetric BPC ($4.89 \pm 3.80\%$ /day) compared with SWC ($2.60 \pm 2.59\%$ /day) when electrodes were applied on opposite sides of the wound edge. This study was excluded from the meta-analysis because the data were replicated in later publications (41,51).

Overall risk of healing

Four RCTs (38–40,42) evaluated the effects of EST on pressure ulcer healing (i.e. number of healed events). Healing of a

pressure ulcer with EST was found to be 1.55 (95% CI: 1.12–2.15, $P = 0.01$) times more likely than that with SWC or sham EST (Figure 3). The data were homogeneous ($I^2 = 0\%$, $P = 0.47$).

Other healing outcome measures

Three RCTs (39,40,43) expressed wound healing as an overall percent decrease in wound surface area (Tables 2 and 3); however, because of limited available data within the studies, a meta-analysis could not be conducted. Houghton *et al.* (40) found that pressure ulcers receiving HVPC with 20-minute cycles of 100 Hz, 10 Hz and no stimulation for 8 hours/day, 7 days/week for 3 months, had an overall percent decrease in wound surface area of $70 \pm 25\%$, which was statistically higher than that in the SWC group ($36 \pm 61\%$). Comparably, Adegoke and Badmos (43) found a noticeable percent decrease in wound surface area in the group provided with 'interrupted direct current' (22.2%) compared with the sham group (2.5%). In another RCT by Griffin *et al.* (39), eight males with SCI received HVPC stimulation at 100 Hz, 500 μ A for 60 minutes/day for 20 days, while nine males received sham stimulation. Although, all stage II pressure ulcers healed regardless of treatment type, those receiving EST showed a significant percent decrease in wound surface area of stage III and IV pressure ulcers compared with the sham group. The median percent decrease in wound surface area from pre-treatment was 80% in the HVPC group and 52% in the sham group.

Secondary outcome measures

Only one (40) of the 15 studies reported adverse events related to EST treatment, which were minor (i.e. skin irritation) and were quickly resolved.

None of the 15 studies reported on or measured the effect of EST in alleviating pain related to the pressure ulcer. Similarly, no studies assessed the impact of treating the pressure ulcer with EST on participants' quality of life.

One case study (47) assessed the economic costs associated with using EST for the treatment of pressure ulcers. Allen and Houghton (47) reported that the total cost to deliver EST in a community-based programme for 12 weeks was \$1477.46 (Canadian dollars).

Compliance with the use of EST was addressed in two (38,40) of the 15 trials. Houghton *et al.* (40) indicated that mean duration of EST was reported to be 3.0 ± 1.5 hours/day, which was approximately half of the recommended treatment time. Baker *et al.* (38) reported that 80% of the subjects were at least 'semi-compliant' (i.e. performed half the recommended treatment time).

Discussion

This systematic review and meta-analysis suggests that EST is an effective adjunctive therapy for treating pressure ulcers in individuals with SCI. The results of our meta-analysis have shown that EST decreases pressure ulcer size at a rate of 1.32%/day and increases pressure ulcer closure by 1.55 times more than with SWC alone or sham EST. These results are based on eight controlled clinical trials involving 597 pressure ulcers. The overall PEDro score for the clinical controlled trials was 5.2. This level of methodological quality is considered low.

High heterogeneity across some of the studies included in the meta-analysis was evident. This can be explained by the lack of standardised method of measuring pressure ulcer size, the considerable variations in type of EST and parameters used (i.e. waveform, frequency, intensity, polarity), the varying methods of delivering EST (i.e. the placement of electrodes) and the different treatment schedules. This was seen in studies with both large (38,41,51) and small sample sizes (39,40,42,44,45). The meta-analysis also included both randomised and non-randomised studies, which could have also contributed to the high heterogeneity. As a result of the heterogeneity between these studies, we used a more conservative model known as the random effects model. With this more conservative analysis, EST was still superior to SWC and sham EST in treating pressure ulcers in individuals with SCI.

The lower heterogeneity associated with the studies included in the meta-analysis of the number of healed ulcers is not surprising given the protocol by Baker *et al.* (38). They compared four treatment groups: asymmetrical BPC, symmetrical BPC, microcurrent (MC) and sham MC. It should be noted that as the authors of this study (38) originally created MC to be a 'stimulated control'; we decided to combine the healed events of the MC group with the sham group, and the healed events of the asymmetrical BPC group with the symmetrical BPC group. This is in line with multiple previous studies that have demonstrated that MC has little to no effect in decreasing the healing time of wounds (53–55). The study by Baker *et al.* (38) included a notable 192 pressure ulcers, which was a much larger sample size than other studies included in the meta-analysis. This resulted in the study having a large relative weight within the meta-analysis. Therefore, any negative effects associated with EST from smaller studies may be negated or masked. It should also be noted that several participants (42.5%) experienced recurrent or multiple pressure ulcers and each of these pressure ulcers was randomised to a particular group. This resulted in data from the same participant being recorded multiple times.

In addition, within this meta-analysis, two trials (39,40) reported the same numbers of pressure ulcers healed in both the EST and control groups. This was likely because of a

high number of stage II pressure ulcers assigned to the control groups. In both trials, all of the stage II ulcers healed completely, regardless of group assignment, suggesting that SWC alone may be sufficient to heal these ulcers in individuals with SCI. These results are consistent with recent best practice guidelines that recommend the use of EST combined with SWC to manage more severe, stage III and IV, pressure ulcers (2).

The results of this meta-analysis are comparable to previous meta-analyses; Gardner *et al.* (25), Barnes *et al.* (27) and Koel and Houghton (26) all demonstrated positive effects of EST on wound healing. However, our meta-analysis is unique in that we focused our findings on the effects of EST on pressure ulcers specifically in the SCI population. The previous meta-analyses included studies with different wound aetiologies beyond pressure ulcers and different patient populations. We included both randomised and non-randomised control trials and did not attempt to standardise the outcome measures as performed in the past (25). Our meta-analysis, which pooled studies expressing the daily percent decrease in pressure ulcer size, included four additional studies (one RCT, one retrospective controlled trial and two prospective controlled trials), which were not included in previous meta-analyses (27). Barnes *et al.* (27) included only one RCT in their meta-analysis that looked at the effects of EST on mean daily percentage change in ulcer size compared with SWC. Pressure ulcers were stratified by location to obtain the treatment effect of EST, which may have concealed the true strength of EST on healing pressure ulcers.

In the meta-analysis by Koel and Houghton (26), they found that unidirectional current flow was a more promising mode of EST delivery than bidirectional current flow. Baker *et al.* (38) found similar results showing that asymmetrical BPC is superior to symmetrical BPC. Although, both these currents flow in a bidirectional manner, the asymmetrical current has a net unidirectional flow. When we compared the results of unidirectional and bidirectional EST in these studies involving pressure ulcers and SCI, we actually found greater healing effects with bidirectional EST compared with unidirectional EST. Therefore, healing outcomes are likely dependent on many other variables, such as the different EST protocols in the included studies. Based on the aforementioned findings, the optimal stimulus parameters for EST remain to be determined.

Secondary outcome measures

There were no trials that measured pain relief and/or quality of life as an outcome measure in relation to EST treatment. Adverse events and compliance with EST were reported inconsistently; the majority of studies did not undertake a systematic approach to evaluating or recording adverse events and compliance, as collected in the methodological section. Only one RCT addressed device-related adverse events, including contact dermatitis associated with self-adhesive electrodes, while two trials (38,40) discovered that individuals with SCI were frequently unable to adhere to prescribed EST protocols.

Economic comparisons between SWC and EST were also scarce with only one case study suggesting that EST might be a cost-effective treatment (47). While there have been other studies (56,57) reporting similar findings in individuals with

SCI with pressure ulcers, these articles did not meet our search criteria. Given the cost-savings associated with using EST to improve healing rates in pressure ulcers, it is interesting that EST is not being used earlier during management. In Canada, for example, this therapy is either not provided or a last resort following other adjunctive therapies (58), while in the USA, EST is covered only if other adjunctive therapies were unsuccessful.

Methodological quality of the trials

The majority of the trials included in the meta-analysis had a quality score (out of 10) of 5 or less (38,41,42,44,45,51). Lack of description about the method of randomisation and allocation concealment, blinding and intention-to-treat analysis was the most common weakness in the trials. However, the low quality (33) across studies was not unexpected as the majority of trials were conducted prior to 2000, before most of the tools to assess methodological quality were developed. Unlike pharmacological trials, blinding of the participants and therapists to a treatment protocol is problematic because it produces visible muscle contractures and/or sensory stimulation. Only two studies used blinded assessor; however, it could be argued that wound size measures and determining that a wound has closed are objective outcomes free of assessor bias.

Study limitations

There are a number of limitations in this review. There were a relatively limited number of studies that met the inclusion criteria and the sample size of pressure ulcers and participants was small in each study. In addition, a wide array of healing outcome measures was used between the studies, which prevented the data from all studies from being pooled into one meta-analysis.

There was considerable variation in the EST intervention used including stimulus parameters (waveform, intensity, frequency and polarity) and treatment scheduling. This likely contributed to the significant heterogeneity between included studies when estimating the overall rate of pressure ulcer healing. However, from a clinical standpoint, these consistent positive results regardless of specific EST features could also suggest that a variety of EST paradigms, which ultimately deliver similar electrical charge to subcutaneous tissues, can stimulate physiological responses to healing. Furthermore, matching the EST treatment paradigm to the patient's preferences may be a more prudent approach rather than devising a single optimal EST treatment protocol for all to use.

Another limitation was the use of the PEDro scale for assessing methodological quality for both randomised and non-randomised controlled trials. Although the PEDro was designed specifically for the sake of comparing RCTs, we also used it to evaluate non-RCTs. In this meta-analysis, we believed that it was imperative that the same scale be used for accurate comparisons of quality across trials.

Publication bias was also a concern; it is well known that studies with insignificant results are less likely to be published. Unfortunately, we were unable to assess publication bias because of the small number of studies (i.e. <10 studies) (37). We attempted to minimise the bias by searching grey literature

and conducting a paired consensus process for study selection and data extraction.

Conclusion

In conclusion, this systematic review found a total of 15 published reports that evaluated the effect of EST on healing of pressure ulcers in people with SCI. Eight controlled clinical trials ($n=274$ patients) included in the meta-analysis found a significant overall effect favouring EST over SWC or sham EST. These results are generalisable to the majority of individuals with SCI as the controlled and uncontrolled studies were conducted in multiple countries and included a wide array of participant characteristics. Unfortunately, because of low methodological quality and high heterogeneity across some of the findings, the findings must be interpreted carefully.

This review also suggests that several types of EST and various treatment schedules can be applied to enhance pressure ulcer healing, allowing regulated health care providers to adapt to different EST programmes depending on a patient's needs. However, conclusions cannot be made regarding the optimal EST treatment protocol for healing pressure ulcers.

Researchers should attempt to investigate the ideal treatment protocol for treating pressure ulcers in individuals with SCI. In addition, future studies should address device-related adverse events, compliance rates and cost-effectiveness of EST compared with SWC, and the implications of EST for relieving pressure ulcer pain and improving the overall quality of life.

Acknowledgements

We would like to thank Marisa Surmacz for her assistance with the literature search. One of the authors of this meta-analysis was a principal investigator of a study included in the meta-analysis.

References

1. Consortium for spinal cord medicine. *Pressure ulcer prevention and treatment following spinal cord injury: a clinical practice guideline for health-care professionals*, 2nd edn. Washington, D.C: Paralyzed Veterans of America, 2014.
2. Houghton PE, Campbell KE, CPG Panel. *Canadian best practice guidelines for the prevention and management of pressure ulcers in people with spinal cord injury. A resource handbook for clinicians* [Internet]. URL http://onf.org/system/attachments/168/original/Pressure_Ulcers_Best_Practice_Guideline_Final_web4.pdf [accessed 20 February 2015].
3. Garber SL, Rintala DH, Hart KA, Fuhrer MJ. Pressure ulcer risk in spinal cord injury: predictors of ulcer status over 3 years. *Arch Phys Med Rehabil* 2000;**81**:465–71.
4. Saunders LL, Krause JS, Acuna J. Association of race, socioeconomic status, and health care access with pressure ulcers after spinal cord injury. *Arch Phys Med Rehabil* 2012;**93**:972–7.
5. Smith BM, Guihan M, LaVela SL, Garber SL. Factors predicting pressure ulcers in veterans with spinal cord injuries. *Am J Phys Med Rehabil* 2008;**87**:750–7.
6. Niazi ZB, Salzberg CA, Byrne DW, Viehbeck M. Recurrence of initial pressure ulcer in persons with spinal cord injuries. *Adv Wound Care* 1997;**10**:38–42.

7. Lala D, Dumont FS, Leblond J, Houghton PE, Noreau L. The impact of pressure ulcers on individuals living with a spinal cord injury. *Arch Phys Med Rehabil* 2014;**95**:2312–9.
8. Registered Nurses' Association of Ontario. *Risk assessment and prevention of pressure ulcers (Revised)*. Toronto: Registered Nurses' Association of Ontario, 2011.
9. Kloth LC. Electrical stimulation for wound healing: a review of evidence from in vitro studies, animal experiments, and clinical trials. *Int J Low Extrem Wounds* 2005;**4**:23–44.
10. Bourguignon GJ, Bourguignon LY. Electric stimulation of protein and DNA synthesis in human fibroblasts. *FASEB J* 1987;**1**:398–402.
11. Bourguignon GJ, Jy W, Bourguignon LY. Electric stimulation of human fibroblasts causes an increase in Ca²⁺ influx and the exposure of additional insulin receptors. *J Cell Physiol* 1989;**140**:379–85.
12. Cho MR, Thatte HS, Lee RC, Golan DE. Integrin-dependent human macrophage migration induced by oscillatory electrical stimulation. *Ann Biomed Eng* 2000;**28**:234–43.
13. Rouabhia M, Park H, Meng S, Derbali H, Zhang Z. Electrical stimulation promotes wound healing by enhancing dermal fibroblast activity and promoting myofibroblast transdifferentiation. *PLoS One* 2013;**8**:e71660.
14. Solis LR, Gyawali S, Seres P, Curtis CA, Chong SL, Thompson RB, Mushahwar VK. Effects of intermittent electrical stimulation on superficial pressure, tissue oxygenation, and discomfort levels for the prevention of deep tissue injury. *Ann Biomed Eng* 2011;**39**:649–63.
15. Petrofsky J, Hinds CM, Batt J, Prowse M, Suh HJ. The interrelationships between electrical stimulation, the environment surrounding the vascular endothelial cells of the skin, and the role of nitric oxide in mediating the blood flow response to electrical stimulation. *Med Sci Monit* 2007;**13**:CR391–7.
16. Ud-Din S, Perry D, Giddings P, Colthurst J, Zaman K, Cotton S, Whiteside S, Morris J, Bayat A. Electrical stimulation increases blood flow and haemoglobin levels in acute cutaneous wounds without affecting wound closure time: evidenced by non-invasive assessment of temporal biopsy wounds in human volunteers. *Exp Dermatol* 2012;**21**:758–64.
17. Stromberg BV. Effects of electrical currents on wound contraction. *Ann Plast Surg* 1988;**21**:121–3.
18. Brown M, Gogia PP. Effects of high voltage stimulation on cutaneous wound healing in rabbits. *Phys Ther* 1987;**67**:662–7.
19. Thawer HA, Houghton PE. Effects of electrical stimulation on the histological properties of wounds in diabetic mice. *Wound Repair Regen* 2001;**9**:107–15.
20. Rowley BA, McKenna JM, Chase GR, Wolcott LE. The influence of electrical current on an infecting microorganism in wounds. *Ann N Y Acad Sci* 1974;**238**:543–51.
21. Kincaid CB, Lavoie KH. Inhibition of bacterial growth in vitro following stimulation with high voltage, monophasic, pulsed current. *Phys Ther* 1989;**69**:651–5.
22. Szuminsky NJ, Albers AC, Unger P, Eddy JG. Effect of narrow, pulsed high voltages on bacterial viability. *Phys Ther* 1994;**74**:660–7.
23. Kawasaki L, Mushahwar VK, Ho C, Dukelow SP, Chan LLH, Chan KM. The mechanisms and evidence of efficacy of electrical stimulation for healing of pressure ulcer: a systematic review. *Wound Repair Regen* 2014;**22**:161–73.
24. Regan MA, Teasell RW, Wolfe DL, Keast D, Mortenson WB, Aubut J-AL. A systematic review of therapeutic interventions for pressure ulcers after spinal cord injury. *Arch Phys Med Rehabil* 2009;**90**:213–31.
25. Gardner SE, Frantz RA, Schmidt FL. Effect of electrical stimulation on chronic wound healing: a meta-analysis. *Wound Repair Regen* 1999;**7**:495–503.
26. Koel G, Houghton PE. Electrostimulation: current status, strength of evidence guidelines, and meta-analysis. *Adv Wound Care* 2014;**3**:118–26.
27. Barnes R, Shahin Y, Gohil R, Chetter I. Electrical stimulation vs. standard care for chronic ulcer healing: a systematic review and meta-analysis of randomised controlled trials. *Eur J Clin Invest* 2014;**44**:429–40.
28. Edsberg LE, Wyffels JT, Ogrin R, Craven C, Houghton P. A pilot study evaluating protein abundance in pressure ulcer fluid from people with and without spinal cord injury. *J Spinal Cord Med* 2014 [Epub ahead of print].
29. Segal JL, Gonzales E, Yousefi S, Jamshidipour L, Brunnemann SR. Circulating levels of IL-2R, ICAM-1, and IL-6 in spinal cord injuries. *Arch Phys Med Rehabil* 1997;**78**:44–7.
30. Davies AL, Hayes KC, Dekaban GA. Clinical correlates of elevated serum concentrations of cytokines and autoantibodies in patients with spinal cord injury. *Arch Phys Med Rehabil* 2007;**88**:1384–93.
31. Verhagen AP, de Vet HC, de Bie RA, Kessels AG, Boers M, Bouter LM, Knipschild PG. The Delphi list: a criteria list for quality assessment of randomized clinical trials for conducting systematic reviews developed by Delphi consensus. *J Clin Epidemiol* 1998;**51**:1235–41.
32. De Morton NA. The PEDro scale is a valid measure of the methodological quality of clinical trials: a demographic study. *Aust J Physiother* 2009;**55**:129–33.
33. Maher CG, Sherrington C, Herbert RD, Moseley AM, Elkins M. Reliability of the PEDro scale for rating quality of randomized controlled trials. *Phys Ther* 2003;**83**:713–21.
34. Borenstein M, Hedges LV, Higgins JPT, Rothstein H. *Chapter 4: effect sizes based on means. Introduction to meta-analysis*. Chichester: John Wiley & Sons, Ltd, 2009:21–32.
35. Cohen J. *Statistical power analysis for the behavioural sciences (Revised)*. New York: Academic Press, 1977.
36. Higgins JPT, Thompson SG, Deeks JJ, Altman DG. Measuring inconsistency in meta-analyses. *BMJ* 2003;**327**:557–60.
37. Sterne JAC, Sutton AJ, Ioannidis JPA, Terrin N, Jones DR, Lau J, Carpenter J, Rücker G, Harbord RM, Schmid CH. Recommendations for examining and interpreting funnel plot asymmetry in meta-analyses of randomised controlled trials. *BMJ* 2011;**343**:d4002.
38. Baker LL, Rubayi S, Villar F, Demuth SK. Effect of electrical stimulation waveform on healing of ulcers in human beings with spinal cord injury. *Wound Repair Regen* 1996;**4**:21–8.
39. Griffin JW, Tooms RE, Mendius RA, Clift JK, Vander Zwaag R, el-Zeky F. Efficacy of high voltage pulsed current for healing of pressure ulcers in patients with spinal cord injury. *Phys Ther* 1991;**71**:433–42; discussion 442–4.
40. Houghton PE, Campbell KE, Fraser CH, Harris C, Keast DH, Potter PJ, Woodbury GM. Electrical stimulation therapy increases rate of healing of pressure ulcers in community-dwelling people with spinal cord injury. *Arch Phys Med Rehabil* 2010;**91**:669–78.
41. Jerčinović A, Karba R, Vodovnik L, Stefanovska A, Kroselj P, Turk R. Low frequency pulsed current and pressure ulcer healing. *IEEE Trans Rehabil Eng* 1994;**2**:225–33.
42. Karba R, Benko H, Savrin R, Vodovnik L. Combination of occlusive dressings and electrical stimulation in pressure ulcer treatment. *Med Sci Res* 1995;**23**:671–3.
43. Adegoke BO, Badmos KA. Acceleration of pressure ulcer healing in spinal cord injured patients using interrupted direct current. *Afr J Med Med Sci* 2001;**30**:195–7.
44. Barczak M, Kluger P, Kluger J, Bäuerle J, Puhl W. Therapeutic effectiveness of electric stimulation in paraplegic patients with pressure sores. PhD thesis of M. Barczak submitted to Prof. Dr. W. Puhl at the Medical School of the University of Ulm, Germany, 2001.
45. Karba R, Semrov D, Vodovnik L, Benko H, Savrin R. DC electrical stimulation for chronic wound healing enhancement. Part 1. Clinical study and determination of electrical field distribution in the numerical wound model. *Bioelectrochem Bioenerg* 1997;**43**:265–70.
46. Trontelj K, Karaba R, Vodovnik L, Savrin R, Strukelj MP. Treatment of chronic wounds by low frequency pulsed electrical current. *J Tissue Viability* 1994;**4**:105–9.
47. Allen J, Houghton PE. A case study for electrical stimulation on a stage III pressure ulcer. *Wound Care Canada* 2004;**2**:34–6.

48. Pollack SF, Ragnarsson KT, Dijkers M, Dijkers M. The effect of electrically induced lower extremity ergometry on an ischial pressure ulcer: a case study. *J Spinal Cord Med* 2004;**27**:143–7.
49. Chalker M. Healing of decubitus ulcers of patients in neuro-kinesthetic program with the Electro-Acuscope 80. *Crit Care Update* 1983;**10**:50–2.
50. Recio AC, Felter CE, Schneider AC, McDonald JW. High-voltage electrical stimulation for the management of stage III and IV pressure ulcers among adults with spinal cord injury: demonstration of its utility for recalcitrant wounds below the level of injury. *J Spinal Cord Med* 2012;**35**:58–63.
51. Stefanovska A, Vodovnik L, Benko H, Turk R. Treatment of chronic wounds by means of electric and electromagnetic fields. Part 2. Value of FES parameters for pressure sore treatment. *Med Biol Eng Comput* 1993;**31**:213–20.
52. Cukjati D, Robnik-Sikonja M, Rebersek S, Kononenko I, Miklavcic D. Prognostic factors in the prediction of chronic wound healing by electrical stimulation. *Med Biol Eng Comput* 2001;**39**:542–50.
53. Malin EW, Galin CM, Laird KF, Huzar TF, Williams JF, Renz EM, Wolf SE, Cancio LC. Silver-coated nylon dressing plus active DC microcurrent for healing of autogenous skin donor sites. *Ann Plast Surg* 2013;**71**:481–4.
54. Ullah M. A study to detect the efficacy of micro-current electrical therapy on decubitus wound. *J Med Sci* 2007;**7**:1320–4.
55. Baker LL, Chambers R, DeMuth SK, Villar F. Effects of electrical stimulation on wound healing in patients with diabetic ulcers. *Diabetes Care* 1997;**20**:405–12.
56. Mittmann N, Chan BC, Craven BC, Isogai PK, Houghton PE. Evaluation of the cost-effectiveness of electrical stimulation therapy for pressure ulcers in spinal cord injury. *Arch Phys Med Rehabil* 2011;**92**:866–72.
57. Chan BC, Nanwa N, Mittmann N, Bryant D, Coyte PC, Houghton PE. The average cost of pressure ulcer management in a community dwelling spinal cord injury population. *Int Wound J* 2013;**10**:431–40.
58. Craven BC, Verrier M, Balioussis C, Wolfe DL, Hsieh JT, Noonan VK, Rasheed A, Cherban E. *Rehabilitation environmental scan atlas: capturing capacity in Canadian SCI rehabilitation*. Rick Hansen Institute: Vancouver, 2012.