

Quantification of Internal Carotid Artery Flow with Digital Subtraction Angiography: Validation of an Optical Flow Approach with Doppler Ultrasound

V. Mendes Pereira, R. Ouared, O. Brina, O. Bonnefous, J. Satwiaski, H. Aerts, D. Ruijters, F. van Nijnatten, F. Perren, P. Bijlenga, K. Schaller, and K.-O. Lovblad



ABSTRACT

BACKGROUND AND PURPOSE: Digital subtraction angiography is the reference standard technique to evaluate intracranial vascular anatomy and used on the endovascular treatment of vascular diseases. A dedicated optical flow-based algorithm was applied to DSA to measure arterial flow. The first quantification results of internal carotid artery flow validated with Doppler sonography are reported.

MATERIALS AND METHODS: We included 22 consecutive patients who underwent endovascular procedures. To assess the sensitivity of the algorithm to contrast agent-blood mixing dynamics, we acquired high-frame DSA series (60 images/s) with different injection rates: 1.5 mL/s ($n = 19$), 2.0 mL/s ($n = 18$), and 3.0 mL/s ($n = 13$). 3D rotational angiography was used to extract the centerline of the vessel and the arterial section necessary for volume flow calculation. Optical flow was used to measure flow velocities in straight parts of the ICAs; these data were further compared with Doppler sonography data. DSA mean flow rates were linearly regressed on Doppler sonography measurements, and regression slope coefficient bias from value 1 was analyzed within the 95% confidence interval.

RESULTS: DSA mean flow rates measured with the optical flow approach significantly matched Doppler sonography measurements (slope regression coefficient, $b = 0.83 \pm 0.19$, $P = .05$) for injection rate = 2.0 mL/s and circulating volumetric blood flow < 6 mL/s. For injection rate = 1.5 mL/s, volumetric blood flow < 3 mL/s correlated well with Doppler sonography ($b = 0.67 \pm 0.33$, $P = .05$). Injection rate = 3.0 mL/s failed to provide DSA–optical flow measurements correlating with Doppler sonography because of the lack of measurable pulsatility.

CONCLUSIONS: A new model-free optical flow technique was tested reliably on the ICA. DSA-based blood flow velocity measurements were essentially validated with Doppler sonography whenever the conditions of measurable pulsatility were achieved (injection rates = 1.5 and 2.0 mL/s).

ABBREVIATIONS: OF = optical flow; RMSE = relative root mean square errors; IR = injection rate; CA = contrast agent; USD = Doppler sonography; 3DRA = 3D rotational angiography

Even though digital subtraction angiography has traditionally been confined to standard anatomy assessment, quantification of blood flow based on DSA is becoming an important topic that could help neurointerventionists in making adequate periprocedural decisions. Some reports have described new techniques based on DSA that are able to assess flow or flow changes

during treatment of stented aneurysms.^{1,2} However, the development of clinically useful tools based on the integration of engineering, hemodynamic and physiologic knowledge still requires improved translation of biofluid mechanical information into clinical applications.³ X-ray video densitometry, based on the detection of the displacement of radiopaque contrast material through the vascular system, has been studied since the early 1960s and has been divided into 2 main classes: tracking and computational methods.⁴ Sarry et al⁵ estimated the flow by using an inverse advection model. Bogunović and Lončarić⁶ proposed the combination of DSA and 3D rotational angiography (3DRA), using an analysis of the time-attenuation curves. Rhode et al⁷ developed a model-based and weighted-optical field (OF) approach to improve already existing techniques and compared the results with simulation data, while Imbed et al^{8,9} developed a similar approach on the femoral artery and simulated angiographic data. Waechter et al¹⁰ developed a model-based approach to measure flow in the cerebral arteries.

Received October 8, 2012; accepted after revision March 28, 2013.

From the Interventional Neuroradiology Unit (V.M.P., R.O., O. Brina, K.-O.L.), Service of Neurosurgery (P.B., K.S.), and Service of Neurology (F.P.), University Hospital of Geneva, Geneva, Switzerland; and Medisys Research Laboratory (O. Bonnefous, J.S.) and Interventional X-Ray (H.A., D.R., F.v.N.), Philips Healthcare, Best, the Netherlands.

This work was supported by Swiss National Funds (SNF 32003B_141192 and SNF 320000_121565).

Please address correspondence to Vitor Mendes Pereira, MD, MSc, Interventional Neuroradiology, University Hospital of Geneva, Rue Gabrielle-Perret-Gentil, 4 Geneva, Switzerland 1211; e-mail: vitormpbr@hotmail.com

Indicates open access to non-subscribers at www.ajnr.org

Indicates article with supplemental on-line table

<http://dx.doi.org/10.3174/ajnr.A3662>

These approaches were all limited in case of fast flows, required longer straight-vessel segments, and were affected by low signal-to-noise ratios. A dedicated algorithmic scheme was developed to reduce instabilities due to temporal and spatial noise and to cope with fast flows, to address these issues.¹¹ Essentially, we used only the modulation by the cardiac cycle to extract flow-velocity values. The OF principle was then applied to the pulsating component of dye-concentration signal.¹¹ In this article, the first clinical results of the proposed algorithm applied to the ICA are reported. The clinical implementation and verification against Doppler ultrasound (USD) data in a consecutive cohort of patients are described, and the limits of the current technique are discussed.

MATERIALS AND METHODS

Clinical Data

The study protocol was approved by our institutional ethics committee (NEC 07–056). Between 2008 and 2009, all patients undergoing angiographic investigation or endovascular treatment for intracranial aneurysms in our department were considered for recruitment. Patients with intraventricular drainage, hydrocephalus, or hematoma were excluded because of a possible unstable intracranial pressure that could interfere with intracranial flow during the procedure and image acquisition. We included prospectively 22 consecutive patients with ruptured or unruptured intracranial aneurysms. All patients were evaluated under general anesthesia. A 5F 0.038-inch diagnostic catheter (Cook, Bloomington, Indiana) was placed selectively in the ICA (3 cm after the common carotid bifurcation) by using a femoral approach. The angiographic examinations were performed with a monoplane angiographic C-arm Allura FD20 system (Philips Healthcare, Best, the Netherlands). Every patient underwent 3DRA to acquire the 3D vascular anatomy with an injection rate (IR) = 3 mL/s during 6 seconds. An optimized projection, offering the best view of the vessel segment, was chosen for flow-rate measurement with the fewest overlapping vessels. High-speed DSA (60 frames/s with the 1024 × 1024 detector pixel matrix) was acquired in 10 seconds. The radiation exposure was equivalent to 2-frames-per-second sequences in terms of dose level. Two or 3 DSA acquisitions by using different contrast agent IRs were obtained for each patient. The contrast agent (CA) was iopamidol (Iopamiro 300; Bracco, Milan, Italy), and the injection duration was 3 seconds. CA was injected with an Imaxeon Avidia angiographic contrast injector (Medrad, Indianola, Pennsylvania). IRs were randomly selected before each acquisition among 3 protocols: 1.5 mL/s ($n = 19$), 2.0 mL/s ($n = 18$), and 3.0 mL/s ($n = 13$).

Optical Flow Principle

The OF-based algorithm used in this study to measure internal carotid flows was based on both 2D DSA and 3DRA image data.¹¹ The blood-contrast flow-velocity fields were measured from 2D DSA sequences. The injected contrast material was diluted in the blood stream while being transported through the vascular network. The pulsating contrast patterns generated by dye injection under the effect of the cardiac cycle were captured in DSA sequences. The main improvement provided by this approach consisted of extracting only the cardiac modulation from the signal, while discarding the low-frequency wash-in and washout compo-

nents. The OF approach is a mass-conserving technique, which was applied in an image-processing algorithm to extract velocity and flow curves. The second step consisted of projecting the DSA sequence on the arterial centerline extracted from the 3DRA data to generate the 2D “contrast wave map” or “flow map,” which defines the effective progression of the contrast wave along the arterial axis. Then, a 1D OF scheme was applied iteratively on the contrast wave map to extract the velocity profile. The iterative process was stopped when the quadratic error was no longer decreasing. The high-frame-rate acquisition was allowed to reduce constraints on the required arterial segment length and, therefore, to address more complex arterial geometries such as in cerebral vessels. For average velocities as high as 60 cm/s for instance, the required arterial segment length was approximately 20 mm to allow a proper warping, which was largely achieved at a frame rate of 60 Hz (2 times 10 mm). If $C(t, z)$ is the physiologically modulated contrast attenuation at instant time t , and z is the curvilinear axial position along the centerline, displacement velocity $V(t, z)$ satisfies the differential contrast wave equation:

$$1) \quad \frac{\partial C(t,z)}{\partial t} + \frac{\partial C(t,z)}{\partial z}V(t,z) = 0.$$

If $S(z)$ is the arterial cross-section at z , the volume flow curve is given by

$$2) \quad Q(t) = \frac{1}{L} \int_L S(z)V(t,z)dz,$$

where L is the curvilinear length of the arterial segment. $S(z)$ is measured from the segmented vasculature in the 3DRA dataset.

Figure 1 summarizes the whole velocity-extraction process,¹¹ starting from the DSA and 3DRA acquisitions (Fig 1A). The centerline of the artery was used to determine the curvilinear path represented by z . The contrast-level intensity along the curvilinear length z , namely the flow map or contrast wave map,^{10,12} is represented in Fig 1B. The velocity profile extracted from the flow map along with the volumetric flow is represented in Fig 1C.

To ensure good mixing between CA and blood, we selected the region of interest in the ICA at least 10 arterial diameters from the injection site.¹³ The region of measurement was reachable by sub-mandibular USD examination.

The presence of physiologic pulsatility in the contrast material motion pattern is effectively required for our OF algorithm¹¹ and was correlated with injection rate quotient (IQR) metric I_q ,

$$3) \quad I_q = \frac{Q_i}{Q_D},$$

where Q_i is the IR (milliliters per second) and Q_D is the time-average USD volume flow measure (milliliters per second). It has been shown, by comparing cardiac cycles measured from both electrocardiogram and time-intensity curves, that for $I_q \leq 1.2$, the physiologic pulsating character is conserved in a CA motion pattern.¹⁴

Doppler Measurement (USD)

The ICAs were examined by an experienced investigator with a 5- to 11-MHz color-coded duplex linear probe connected to an Aplio 770A scanner (Toshiba Medical Systems, Tokyo, Japan) by




FIG 1. Flowchart illustrating the treatment of DSA images to calculate the blood flow of patient 3. First, the 2D images are superimposed on the 3DRA to scale the distances on the 2D images (A). Next, the contrast values are distributed over a straight model (B) to extract a flow map from the contrast sequence. Last, the flow map is analyzed by using the optical flow principle to estimate flow velocity curves (C). The volume flow rate curve is estimated by using the section of the artery measured from the 3DRA volume.

using a standard technique.¹⁵ The sample volume, adjusted to the size of the artery, was placed in the ICA with a submandibular approach, to obtain velocity waveforms. Its placement was adjusted around the site of measurements to acquire the higher color-coded signal. The angle-corrected mean velocity, peak systolic velocity, and end-diastolic velocity were determined by tracing the maximum frequency envelope of the waveform. The reference flow curves, based on Womersley solutions, were extracted with a Matlab (MathWorks, Natick, Massachusetts) program from images of velocity profiles acquired with USD. The vessel diameter needed to estimate the volumetric flow rate was measured on the 3D reconstruction. For cases in which several USD measures were acquired, repeatability was estimated.

Statistics and Regression Analysis

The validation process was performed by using linear regression fit between Q_D (Doppler) and Q_x (x-ray). Q_D was either a single USD time-average flow rate measure ($n = 6$) or an average of several measures ($n = 16$). Q_x was the time-average measure provided by the OF method. A robust fitting technique was used to reduce the impact of outliers by iteratively reweighting least squares with a bisquare weighting function ($r = 4.6$)^{16,17}

$$4) \quad Q_x = b \cdot Q_D.$$

The slope b and offset a were the regression coefficients of the robust fit model. The analysis was made for every injection rate group separately (1.5, 2.0, and 3.0 mL/s) and for several Q_D thresholds ($Q_D \leq 3.0, \leq 4.0, \leq 5.0, \leq 6.0, \leq 7.0$ mL/s) to test the

volume flow limits of the current OF technique prototype. Ideally, the 95% CI of the regression slope estimate should contain the value 1. The y -intercept a coefficient was proportional to the CA injected volume flow and was not taken into account in the validation process because of the lack of knowledge of this factor.¹⁸ Nonparametric Wilcoxon rank sum U tests and the Kruskal-Wallis H test were used to test statistics for the continuous data: b -slope regression coefficient, injection rate quotient, and Doppler measurement precision. If the H test was significant among 3 groups, the U test was performed to find out which of the differences were statistically significant. The significance level was fixed to $\alpha = .05$.

RESULTS

The On-line Table summarizes the measurements from USD and DSA. The average Doppler flow rate of the whole patient sample was 3.2 ± 0.3 mL/s ($n = 22$). Of the 22 patients, 16 had at least 2 Doppler measures in the submandibular On-line ICA segment: 11 with relative root mean square errors (RMSE) $< 10\%$ and the rest with RMSE between 10% and 25%. This precision did not depend on the circulating mean flow rate ($P = .13$).

Impact of Volume Flow on X-Ray Flow Measurement

Q_x volume flow quantification was regressed on the reference USD Q_D for each injection rate (1.5 mL/s, 2.0 mL/s, and 3.0 mL/s) and for each Q_D threshold subsample (≤ 3 mL/s, ≤ 4 mL/s, ≤ 5 mL/s, ≤ 6 mL/s, ≤ 7 mL/s). Figure 2 illustrates the results for both $Q_D \leq 4$ mL/s (left) and $Q_D \leq 7$ mL/s (right) subsamples. For $Q_i = 1.5$ mL/s (Fig 2A, -B), the regression slopes were $b = 0.49$ (95% CI,




FIG 2. A and B, The regression fit at $Q_i = 1.5$ mL/s for $Q_D \leq 4$ mL/s (left) and $Q_D \leq 7$ mL/s (right). Q_x is represented on the y-axis, and Q_D on the x-axis. The data (blue star points), fit (red line), and 95% CI limits (magenta) are all represented. C and D, The regression fit at $Q_i = 2.0$ mL/s for $Q_D \leq 3$ mL/s (left) and $Q_D \leq 7$ mL/s (right). Q_x is represented on the y-axis and Q_D on the x-axis. The data (blue star points), fit (red line), and 95% CI limits (magenta) are all represented. E and F, The regression fit at $Q_i = 3.0$ mL/s for $Q_D \leq 3$ mL/s (left) and $Q_D \leq 7$ mL/s (right). Q_x is represented on the y-axis, and Q_D on the x-axis. The data (blue star points), fit (red line), and 95% CI limits (magenta) are all represented.

0.10–0.88; $R^2 = 0.72$) (Fig 2A) and $b = 0.42$ (95% CI, 0.21–0.63; $R^2 = 0.84$) (Fig 2B), respectively. For $Q_i = 2.0$ mL/s (Fig 2C, -D), the slopes were $b = 0.84$ (95% CI, 0.40–1.28; $R^2 = 0.76$) (Fig 2C)

and $b = 0.62$ (95% CI, 0.33–0.91; $R^2 = 0.81$) (Fig 2D), respectively. For $Q_i = 3.0$ mL/s (Fig 2E, -F), the slopes were $b = 0.35$ (95% CI, 0.03–0.67; $R^2 = 0.71$) (Fig 2E) and $b = 0.42$ (95% CI,




FIG 3. A, the boxplot for the fitted slope (y-axis) at 1.5, 2.0, and 3 mL/s (x-axis). Each boxplot represents the 5 slope values fitted for each Q_D subsample, with Q_D thresholds ≤ 3.0 , ≤ 4.0 , ≤ 5.0 , ≤ 6.0 , and ≤ 7.0 mL/s. B, results of the nonparametric ANOVA multiple comparison of the regression slope upper limit (x-axis) for each injection rate group (y-axis). These summarize the corresponding maximum slope values of 0.84, 1.21, and 0.72 with an SE of 0.12. The 95% CI error margins are also represented.




FIG 4. A, Slope versus Q_D for injection rates $Q_i = 1.5$ (blue), 2.0 (red), and 3.0 mL/s (green). The 95% CI limits have been superimposed to $Q_i = 2.0$ mL/s points. The graph shows that the slope bias calculated is not significantly different from 1 up to $Q_D = 6$ mL/s ($P > .05$). The $Q_i = 1.5$ mL/s group (blue points) follows the same evolution trend as $Q_i = 2.0$ mL/s, though with accentuated bias. A nonsignificant slope bias is likely for the Q_D range below 3 mL/s ($P = .06$). B, Injection rate quotients for $Q_i = 1.5$ mL/s (group 1), 2.0 (group 2), and 3.0 mL/s (group 3). The blue bars correspond to quotients < 1.2 , and the red bars, to quotients > 1.2 . The fact that the $Q_i = 1.5$ and 2.0 mL/s groups are reconstructed in similar ways with minimum slope bias, unlike the $Q_i = 3.0$ mL/s group, is partly due to differences in injection rate quotients ($P = .01$).

0.16–0.67; $R^2 = 0.80$) (Fig 2F), respectively. The slope bias from 1 significantly increased with the Q_D threshold for both $Q_i = 1.5$ and 2.0 mL/s groups. Unlike the 2 other IRs, the slope bias for $Q_i = 3.0$ mL/s followed a decreasing trend with increasing Q_D .

Sampling Effect on the Regression Slope Coefficient

The comparison of the regression slope coefficient among Q_i groups (Fig 3A) indicated that the only significant difference was between the $Q_i = 2.0$ mL/s and $Q_i = 3.0$ mL/s groups ($P = .0015$). The mean slopes, if sampling errors were ignored, were 0.48, 0.75, and 0.35 for $Q_i = 1.5$, 2.0, and 3.0 mL/s, respectively, with SE = 0.06. The corresponding slope upper values for the respective Q_i values were 0.84, 1.21, and 0.72, respectively, with a SE of 0.12 (Fig 3B). The $Q_i = 3.0$ mL/s group was significantly biased compared

with both $Q_i = 1.5$ mL/s and 2.0 mL/s ($P = .039$). For each CA injection rate Q_i , the sampling effect on slope driven by Q_D cutoffs at 3.0, 4.0, 5.0, 6.0, and 7.0 mL/s is plotted in Fig 4A. The error bars are superimposed for the group $Q_i = 2.0$ mL/s. For this group, the slope bias was not significant up to $Q_D = 6$ mL/s. Above $Q_D = 5$ mL/s, flow quantification was not sensitive to sampling. For $Q_i = 1.5$ mL/s, slope bias was significant ($P < .05$) for sampling cutoffs of > 3 mL/s; and for $Q_i = 3.0$ mL/s, the slope bias was effective for the whole sample ($P = .0004$).

Injection Rate Quotient

The ratio between the injection rate and the mean blood flow is illustrated in Fig 4B. The groups $Q_i = 1.5$ and 2.0 mL/s were not different ($P = .85$), while both of them were different from the




FIG 5. Comparison of contrast wave maps (top) and flow curves (bottom) for patient 13 for different IRs: IR = 1.5 mL/s (left), IR = 2.0 mL/s (middle), and IR = 3.0 mL/s (right). The red and blue curves in the bottom images correspond to USD and DSA measurements.

$Q_i = 3.0$ mL/s group ($P = .01$). In the first 2 groups, 90% of the data points had an injection rate quotient < 1.2 , implying that physiologic information and modulation were conserved in contrast-motion patterns.¹⁴ In the $Q_i = 3.0$ mL/s group, 45% of the points had an injection rate quotient > 1.2 . Therefore, the optical flow measures significantly matched the Doppler outcome for 2.0 mL/s injected cases when circulating blood flow ranged between 1 and 6 mL/s. With a 1.5 mL/s injection rate, flow measures matching was for circulating blood flow ranging from 1 to 3 mL/s. Figure 5 shows, for case 13, the contrast wave maps and flow curves corresponding to IR = 1.5, 2.0, and 3.0 mL/s. As one can notice, the contrast wave maps corresponding to IR = 1.5 and 2.0 mL/s show the best premises for velocity extraction (several pulses) compared with IR = 3.0 mL/s (1 detectable pulse).

DISCUSSION

Quantification of the Flow through DSA

The DSA-based approach represents a logical evolution of the criterion standard method for intracranial vascular anatomy assessment. Although DSA is an imaging technique used for endovascular treatments, the lack of tools for periprocedural quantitative flow measurements did not ease real-time therapeutic decision-making.^{1,2} The advantages for measuring blood velocity and flow changes during the procedure are numerous. For instance, one could evaluate the hemodynamic effect of an angioplasty in a stenotic segment for vasospasm or intracranial stenosis or the impact on the intra-aneurysmal hemodynamics of a flow-diverter stent to decide whether to place another one or coils.^{1,2} Other possible applications could be in acquiring patient-specific boundary conditions (flow curves) for computational flow dynamics. Different methods have been proposed to quantify both flow changes and characteristics from DSA sequences.^{3,7,10-12,19} However, most could not provide a reasonable level of confidence for clinical applications because of their inability to fully cope with both the complex advective and diffusive dispersion of the dye in blood stream and the complex pulsatile nature of the flow in the artery.³ Additionally, not all studies paid attention to CA injection protocols that might alter the physiologic modulation of CA motion waveform.^{13,20-22} This physiologic modulation is a key factor in dealing with the mixing problems related to CA transport in quantifying blood flow. The extraction of the modulated component with proper filtering is a necessary step that will condition the performance of such a functional algorithm. Con-

trary to the technique of Waechter et al,¹⁰ which also extracted the modulated component, our OF algorithm is model-free and can be applied to quantify flow in both the vessels and aneurysms.¹

In our study, we used 3 different IRs: 1.5, 2.0, and 3.0 mL/s. The Craya-Curtet number was 0.5–0.9 with a 5F catheter of 0.035-inch inner diameter, the tip of which was positioned at least a 10-artery diameter below the measurement zone.¹³ Measurable flow pulsatility was obtained for IR = 1.5 mL/s and 2 mL/s (injection rate quotient ≤ 1.2).^{13,14} Consequently, the OF technique performance was reasonably high, with slightly better results for 2 mL/s and for volume flows of < 6 mL/s. The performance with IR 1.5 mL/s was lower than expected because of the larger dilution of the dye with a deleterious effect on signal-to-noise ratio. The injection at 3 mL/s presented excessive reflux at the injection site, which altered the matching between dye and blood-flow dynamics in the region of interest.

General Trend of the OF Method

The general trend of our current OF algorithm was to measure flow with an underestimated slope bias that depends on the injection rate: 0.48, 0.75, and 0.35 for $Q_i = 1.5$ mL/s, 2.0 mL/s, and 3.0 mL/s, respectively, with RMSE = 0.06. Those values do not take into account sampling errors. The maximum slope value variations (95% CI) over the whole Q_D show that the slope bias for $Q_i = 1.5$ and 2.0 mL/s was not statistically significant (the maximum slope is, respectively, 0.84 and 1.21, with RMSE = 0.12, $P = .05$), unlike that for IR = 3.0 mL/s group. This general trend is consistent with the fact that the performance of the OF method depends first on the presence of the physiologic information on contrast wave maps and, consequently, on the injection rate quotient (≤ 1.2).

Thorough examinations of slope dependency with Q_D show that the slope bias for group IR = 2.0 mL/s was not statistically significant ($P > .05$) up to 6 mL/s, setting the limit on reliable velocity measurement to 60 ± 6 cm/s, taking into account the average inlet diameter of 4.8 mm and the relative RMSE = 0.12. For the IR = 1.5 mL/s group, the slope bias was small, within $Q_D \leq 3.0$ mL/s, and increased with Q_D . As Q_D increases, the corresponding CA signal-to-noise ratio is reduced, possibly because poor mixing occurs at the injection site due to high Craya-Curtet number.^{7,13} Consequently, intensity gradients in 2D images would have been overestimated in the denominator of the OF equation. More improvements to the OF technique are expected

in the near future to cope uniformly with different injection rates and flow rates.

Study Limitations

Doppler Measurements. The OF trend and limitations presented in the above section are discussed under the assumption that USD measures are infinitely accurate, which is, in fact, not correct. Biases due to geometric spectral broadening and to angulations between the beam direction and the vessel axis may lead to an overestimate of the USD velocity of approximately 30%.²³ The Doppler measurement repeatability inherent to the experienced investigator was reasonable (<23%). However, the mean-flow-rate repeatability of the measurements was <15% in 13 of 16 cases that had at least 2 USD measures, and for the other 3, it was $\leq 23\%$. The volumetric flow shapes were similar for repeated measures.

Acquisition Protocol. Sources of mismatch between OF and USD could have been due to the delay between the 2 acquisitions or to inaccurate estimates of the parent artery diameter from the 3D geometry. The diameter is influenced by the threshold selection and by the location of the measurement on the artery, which could deviate from the USD measurement zone. In this work, we focused on the most rectilinear part of the internal carotid artery to ensure the reliability of USD velocity measurements. In the submandibular part of ICA, the vessel narrowing is negligible within the OF measurement-segment length (2 cm). The range of flow rates in the ICA is quite exhaustive to validate the method. Nevertheless, it will be necessary to validate in other cerebral vessel locations where flow rates are smaller.

CONCLUSIONS

The DSA-based OF approach can be used to measure circulating flows in ICAs. The OF measure was validated with USD for flows up to 6 mL/s with CA injected at 2.0 mL/s and up to 3 mL/s with CA injected at 1.5 mL/s. The values obtained from the IR = 3.0 mL/s group were significantly biased because only partial dynamic information was propagated to the CA motion pattern. Improvements to the OF technique are expected in the near future to improve the accuracy and uniformity of velocity measurements over ranges of flow rates and injection rates. The capability of using this technique during the procedures and assessing fast flows in cerebral vessels would open new insights in intraoperative therapeutic decision-making.

APPENDIX

Statistical Analysis

The linear regression method we used applies generally to cases in which the independent variable (Q_D) is free from errors. In such situations, the robust fitting method consistently rejects the outliers and focuses on the bulk attenuation of data points. In our case, because 16 of 22 points were characterized by an estimated error on Q_D , errors-in-variables regression methods would have been used (eg, Deming regression) to avoid underestimating the regression coefficients. Some OF and Doppler measures were unique, and using Deming-like regression would have led to speculations on relative variances. Requiring that all data points have more than a single measure simultaneously in either OF or Dopp-

ler would have reduced the number of points by 50%. The robust regression fit is then a reasonable compromise to capture trends of the optical flow method.

The H test is generally performed among groups that are statistically independent. In our study, Q_D cutoff-based subsampling created the 3 IR (Q_i) groups. However, with such a low number of points (5) in each group, the lack of that statistical independence should not have dramatic consequences, given the level of precision and accuracy that we were dealing with.

USD

For several decades, the potential of Doppler instruments to measure volume flow rates at the same time as velocity waveforms has been discussed.²⁴ One article has reviewed all of the factors of inaccuracy and provided the orders of magnitude of the errors that were derived from the Doppler acquisition constraints and that were evaluated during validation tests.²³ Validation experimentation estimates the performance of the technique with in vivo and in vitro tests.²⁵ The authors found errors in the range of 5%–54% for in vivo experiments and errors <18% for in vitro flow-phantom experiments.

Disclosures: Vitor Mendes Pereira—*RELATED: Grant:* Swiss National Foundation.* Odile Bonnefous—*UNRELATED: Employment:* Philips Research Paris. Jean Stawiaski—*UNRELATED: Employment:* Philips Research. Fred van Nijnatten—*UNRELATED: Employment:* Philips Medical Systems Nederland BV (a Philips Healthcare Company). Hans Aerts—*UNRELATED: Employment:* Philips Healthcare. Daniel Ruijters—*UNRELATED: Employment:* Philips Healthcare. *Money paid to the institution.

REFERENCES

1. Pereira VM, Bonnefous O, Ouared R, et al. **A DSA-based method using contrast-motion estimation for the assessment of the intra-aneurysmal flow changes induced by flow-diverter stents.** *AJNR Am J Neuroradiol* 2013;34:808–15
2. Jou LD, Mawad ME. **Analysis of intra-aneurysmal flow for cerebral aneurysms with cerebral angiography.** *AJNR Am J Neuroradiol* 2012;33:1679–84
3. Lieber BB, Siebes M, Yamaguchi T. **Correlation of hemodynamic events with clinical and pathological observations.** *Ann Biomed Eng* 2005;33:1695–1703
4. Shpilfoygel SD, Close RA, Valentino DJ, et al. **X-ray videodensitometric methods for blood flow and velocity measurement: a critical review of literature.** *Med Phys* 2000;27:2008–23
5. Sarry L, Peng YJ, Boire JY. **Blood flow velocity estimation from x-ray densitometric data: an efficient numerical scheme for the inverse advection problem.** *Phys Med Biol* 2002;47:62
6. Bogunović H, Lončarić S. **Blood flow and velocity estimation based on vessel transit time by combining 2D and 3D X-ray angiography.** *Med Image Comput Assist Interv* 2006;9(pt 2):117–24
7. Rhode KS, Lambrou T, Hawkes DJ, et al. **Novel approaches to the measurement of arterial blood flow from dynamic digital x-ray images.** *IEEE Trans Med Imaging* 2005;24:500–13
8. Imbed B, Meunier J, Mongrain R, et al. **Optical flow assessment of parabolic velocity profile in cineangiography: a simulation study.** In: *Proceedings of the 17th International Conference of the Engineering in Medicine and Biology Society*, Montreal, Canada, September 20–25, 1995
9. Imbed B, Meunier J, Mongrain R, et al. **Blood flow assessment from optical flow in cineangiography.** In: *Proceedings of the 17th International Conference of the Engineering in Medicine and Biology Society*, Montreal, Canada, September 20–25, 1995
10. Waechter I, Bredno J, Barratt DC, et al. **Quantification of blood flow from rotational angiography.** *Med Image Comput Assist Interv* 2007;10(pt 1):634–41

11. Bonnefous O, Pereira VM, Ouared R, et al. **Quantification of arterial flow with digital subtracted angiography (DSA).** *Med Phys* 2012;39:6264–75
12. Waechter I, Bredno J, Hermans R, et al. **Model-based blood flow quantification from rotational angiography.** *Medical Image Anal* 2008;12:586–602
13. Lieber BB, Sadasivan C, Hao Q, et al. **The mixability of angiographic contrast with arterial blood.** *Med Phys* 2009;36:5064–78
14. Ouared R, Brina O, Bonnefous O, Groth A, Brunjs T, Babic D, Bijlenga P, Schaller K, Lovblad KO, Pereira VM. **Intracranial aneurysms: study of x-rays time-intensity-curves (TIC) pulsatility.** In: *Proceedings of 2nd International Conference on Mathematical and Computational Biomedical Engineering*, Washington DC. March 30-April 1, 2011
15. Forteza AM, Krejza J, Koch S, et al. *Ultrasound imaging of cerebrovascular disease.* In: Babikian VL, Wechsler L, Higashida RT [eds.]. **Imaging Cerebrovascular Disease.** Philadelphia: Butterworth-Heinemann; 2003, pp. 3–35.5
16. Holland PW, Welsch RE. **Robust regression using iteratively reweighted least-squares: theory and methods.** *Communications in Statistics: Theory and Methods* 1977;6:813–27
17. Street JO, Carroll RJ, Ruppert D. **A note on computing robust regression estimates via iteratively reweighted least squares.** *The American Statistician* 1988;42:152–54
18. Mulder G, Bogaerds AC, Rongen P, et al. **The influence of contrast agent injection on physiological flow in the circle of Willis.** *Med Eng Phys* 2011;33:195–203
19. Tenjin H, Asakura F, Nakahara Y, et al. **Evaluation of intra-aneurysmal blood velocity by time-density curve analysis and digital subtraction angiography.** *AJNR Am J Neuroradiol* 1998;19:1303–07
20. Sadasivan C, Cesar L, Seong J, et al. **An original flow diversion device for the treatment of intracranial aneurysms: evaluation in the rabbit elastase-induced model.** *Stroke* 2009;40:952–58
21. Sadasivan C, Cesar L, Seong J, et al. **Treatment of rabbit elastase-induced aneurysm models by flow diverters: development of quantifiable indexes of device performance using digital subtraction angiography.** *IEEE Trans Med Imaging* 2009;28:1117–25
22. Sadasivan C, Lieber BB, Cesar L, et al. **Angiographic assessment of the performance of flow diverters to treat cerebral aneurysms.** In: *Proceedings of the 28th Annual International Conference of the Engineering in Medicine and Biology Society*, New York, New York. August 31-September 3, 2006;3210–13
23. Hoskins PR. **Accuracy of maximum velocity estimates made using Doppler ultrasound systems.** *Br J Radiol* 1996;69:172–77
24. Burns PN. **Measuring volume flow with Doppler ultrasound: an old nut.** *Ultrasound Obstet Gynecol* 1992;2:238–41
25. Holland CK, Clancy MJ, Taylor KJ, et al. **Volumetric flow estimation in vivo and in vitro using pulsed-Doppler ultrasound.** *Ultrasound Med Biol* 1996;22:591–603