

Identification of Contaminating Fungal DNA Sequences in Zymolyase

DAGMAR RIMEK,^{1*} AMAR P. GARG,² WALTER H. HAAS,³ AND REINHARD KAPPE¹

Hygiene Institute¹ and Department of General Pediatrics, Children's Hospital,³ University of Heidelberg, Heidelberg, Germany, and Department of Botany, C. Charan Singh University, Meerut, India²

Received 27 August 1998/Returned for modification 22 October 1998/Accepted 11 December 1998

When different preparations of Zymolyase were included in the pretreatment protocol of a panfungal PCR assay using a primer system for the 18S rRNA gene, an amplification product occurred in negative controls. The amplified fragment showed 100.0% sequence identity to the *Saccharomyces sensu stricto* complex and *Kluyveromyces lodderae*. Lyticase, lysing enzymes, and proteinase K appeared to be free from fungal DNA.

Early diagnosis of invasive fungal infection is a major concern in modern medicine, because of the high morbidity and mortality in high-risk patients. Conventional methods, including microscopy, culture, and antigen detection, lack sufficient sensitivity and/or specificity. Therefore, sensitive PCR assays have been devised for detection of fungal DNA in blood and bronchoalveolar lavage fluid, often with panfungal primers (2, 16). The sample pretreatment often includes the use of enzymes for digestion of yeast cell walls in order to extract fungal DNA (1, 3, 4, 6, 9, 13–15, 17). We consistently amplified DNA from the enzyme Zymolyase by using a universal fungal primer system for the 18S rRNA gene (10). To determine the origin of the amplified DNA, we sequenced this fragment.

Universal precautions to prevent PCR assay contamination, including the use of separate rooms and plasticware supplies for PCR setup and products, aliquoted reagents, and aerosol-resistant pipette tips, have been observed (11).

Several different enzymes used for sample pretreatment for fungal PCR assays, including three different batches of Zymolyase-20T from two companies, lyticase, lysing enzymes (from *Trichoderma harzianum*), and proteinase K, were examined (Table 1). The lyophilized enzymes were dissolved in sterile, pyrogen- and DNA-free water (Aqua ad injectabilia; Braun, Melsungen, Germany) to result in a final concentration of 300 µg/ml. One hundred microliters of this solution were heated for 10 min at 95°C and then processed by a purification and concentration procedure for DNA by using the GeneClean II kit (Bio 101, La Jolla, Calif.) according to the recommendations of the manufacturer. The final DNA preparation was dissolved in 25 µl of water. Sterile water as a negative sample control was processed the same way each time a sample preparation was done. Control DNA was prepared from *Candida albicans* and *Aspergillus fumigatus* from the University of Heidelberg strain collection.

PCR amplifications were carried out in 100-µl volumes with 10 µl of sample added as described previously (10). Mixtures were subjected to an initial denaturation of 2 min at 95°C and 35 cycles of 95°C for 0.5 min, 53°C for 1 min, and 72°C for 1 min and a final extension at 72°C for 3 min and refrigeration. Two panfungal primer pairs complementary to fungal 18S rRNA sequences that are highly conserved were used: S1 and CUF1 (10), amplifying a 194-bp fragment (bp 86 to 279), including the variable region V2; and F5 (5'-AGTCTTAACC

ATAAACTATG) and F6 (5'-AGACAAATCACTCCACCA), amplifying a 295-bp fragment (bp 1012 to 1304), including the variable region V5.

Both primer pairs consistently yielded amplification products when the Zymolyase preparation was added to the master mix. The amplicons had exactly the sizes of the fragments which were amplified from fungal DNA (194 and 295 bp, respectively). No products occurred when water or the preparations of the other enzymes were added (Fig. 1 and Table 1). To identify the contaminating DNA, both PCR fragments were sequenced as described previously (5). Briefly, the resulting PCR products were purified by using the Qiaquick PCR purification kit (Qiagen, Hilden, Germany), according to the recommendations of the manufacturer. About 200 ng of purified PCR product was submitted to cycle sequencing with biotinylated dideoxynucleoside triphosphates (GATC-Biocyte sequencing kit; GATC, Konstanz, Germany) and Thermo sequenase (Amersham Buchler, Braunschweig, Germany). The primers S1, CUF1, F5, and F6 were chosen for sequencing. DNA fragments were transferred to a nylon membrane during gel electrophoresis with a GATC 1500 direct blotting electrophoresis sequencer and visualized as blue bands with streptavidin alkaline phosphatase and a nitroblue tetrazolium chloride-X-phosphate mix (Boehringer, Mannheim, Germany). The results of sequence analysis of the two amplified fragments of the contaminating fungal DNA from the three tested batches of Zymolyase were 100% identical among the three batches with both primer pairs.

DNA sequence similarity searches were performed by using the EMBL and GenBank sequence data banks. These searches revealed that both nucleotide sequences had 100% identity with 18S rRNA gene sequences of the following species: *Saccharomyces cerevisiae* (EMBL database accession no. Z75578), *Saccharomyces pastorianus* (accession no. X97805), *Saccharomyces paradoxus* (accession no. X97806), *Saccharomyces bayanus* (accession no. X97777), and *Kluyveromyces lodderae* (accession no. X83824). The four *Saccharomyces* species mentioned above display ≥99.9% sequence similarity in the 18S rRNA gene and form the *Saccharomyces sensu stricto* complex. *K. lodderae* is also a closely related yeast (8). Therefore, any of the five species could be the source of the contaminating fungal DNA in Zymolyase.

Thus, our finding raises concerns about possible false-positive results, if universal fungal primers are being used in sensitive PCR assays in combination with a Zymolyase pretreatment. Zymolyase is a registered trademark of the Kirin Brewery Co., Ltd., Tokyo, Japan. According to the distributors, Zymolyase-20T is manufactured by a submerged culture of *Arthro-bacter luteus* by Yeast Related Business, Development Dept.

* Corresponding author. Mailing address: Hygiene-Institut, Im Neuenheimer Feld 324, D-69120 Heidelberg, Germany. Phone: 49 6221 567816. Fax: 49 6221 564343. E-mail: dagmar_rimek@ukl.uni-heidelberg.de.

TABLE 1. Sensitive fungal PCR assays with fungi and commercial enzymes frequently used for pretreatment

Sample	Source	Amplification product result with primer pair:	
		S1-CUF1	F5-F6
<i>C. albicans</i>	DNA	Positive	Positive
<i>A. fumigatus</i>	DNA	Positive	Positive
Zymolyase-20T	Seikagaku, Tokyo, Japan	Positive ^a	Positive ^a
Zymolyase-20T	ICN, Aurora, Ohio	Positive ^a	Positive ^a
Lyticase	Sigma, St. Louis, Mo.	Negative	Negative
Lysing enzymes (<i>T. harzianum</i>)	Sigma, St. Louis, Mo.	Negative	Negative
Proteinase K	Amresco, Solon, Ohio	Negative	Negative
Proteinase K	Boehringer, Mannheim, Germany	Negative	Negative
Negative control	Sterile water	Negative	Negative

^a Identified by sequencing as DNA from a member of the *Saccharomyces sensu stricto* complex or *K. lodderae*.

(Kirin Brewery Co., Ltd.). Our data show that Zymolyase-20T is contaminated with DNA from the *Saccharomyces sensu stricto* complex or *K. lodderae*. This contamination is still present after partial purification of the Zymolyase by affinity chromatography, which was done by ICN Biomedicals, Inc.

Other enzymes, which are also widely used for pretreatment of samples for fungal PCR assays, including lysing enzymes from *T. harzianum* and lyticase (both from Sigma, Munich, Germany) and proteinase K (Amresco, Solon, Ohio, and Boehringer, Mannheim, Germany), did not yield amplification products when the panfungal primer pairs S1-CUF1 (Fig. 1) and F5-F6 (Table 1) were used; thus, they appeared to be largely free from fungal DNA.

The contamination of *Taq* polymerase with bacterial DNA,




FIG. 1. Ethidium bromide-stained polyacrylamide gel of PCR products obtained from enzymes used for pretreatment of samples for fungal PCR and controls by using primers S1 and CUF1. Lanes: L, 100-bp ladder; 1, 100 pg of *C. albicans* DNA; 2, 100 pg of *A. fumigatus* DNA; 3, Zymolyase (Seikagaku); 4, Zymolyase (ICN); 5, lysing enzymes (Sigma); 6, lyticase (Sigma); 7, proteinase K (Amresco); 8, proteinase K (Boehringer); 9, negative control.

which is of relevance for PCR with universal bacterial primers, has been described previously (7, 12). This is now the first documented report of fungal DNA being present in an enzyme used in the PCR assay process. We conclude that if Zymolyase pretreatment is used for fungal PCR, primers and probes must not cross-react with *Saccharomyces* DNA, or enzymes other than Zymolyase should be used, in order to avoid false-positive results.

This work was supported in part by the Deutscher Akademischer Austauschdienst (DAAD) (A. P. Garg) and by DFG grant Ha 1921/3-1/2.

We thank Matthias Maiwald for critical reading of the manuscript.

REFERENCES

- Burgener-Kairuz, P., J.-P. Zuber, P. Jaunin, T. G. Buchman, J. Bille, and M. Rossier. 1994. Rapid detection and identification of *Candida albicans* and *Torulopsis (Candida) glabrata* in clinical specimens by species-specific nested PCR amplification of a cytochrome P-450 lanosterol- α -demethylase (L1A1) gene fragment. *J. Clin. Microbiol.* **32**:1902-1907.
- Einsele, H., H. Hebart, G. Roller, J. Löffler, I. Rothenhöfer, C. A. Müller, R. A. Bowden, J.-A. van Burik, D. Engelhard, L. Kanz, and U. Schumacher. 1997. Detection and identification of fungal pathogens in blood by using molecular probes. *J. Clin. Microbiol.* **35**:1353-1360.
- Fujita, S.-I., B. A. Lasker, T. J. Lott, E. Reiss, and C. J. Morrison. 1995. Microtitration plate enzyme immunoassay to detect PCR-amplified DNA from *Candida* species in blood. *J. Clin. Microbiol.* **33**:962-967.
- Glee, P. M., P. J. Russell, J. A. Welsch, J. C. Pratt, and J. E. Cutler. 1987. Methods for DNA extraction from *Candida albicans*. *Anal. Biochem.* **164**:207-213.
- Haas, W. H., K. Schilke, J. Brand, B. Amthor, K. Weyer, P. B. Fourie, G. Bretzel, V. Sticht-Groh, and H. J. Bremer. 1997. Molecular analysis of *katG* gene mutations in strains of *Mycobacterium tuberculosis* complex from Africa. *Antimicrob. Agents Chemother.* **41**:1601-1603.
- Holmes, A. R., R. D. Cannon, M. G. Shepherd, and H. F. Jenkinson. 1994. Detection of *Candida albicans* and other yeasts in blood by PCR. *J. Clin. Microbiol.* **32**:228-231.
- Hughes, M. S., L.-A. Beck, and R. A. Skuce. 1994. Identification and elimination of DNA sequences in *Taq* DNA polymerase. *J. Clin. Microbiol.* **32**:2007-2008.
- James, S. A., J. Cai, I. N. Roberts, and M. D. Collins. 1997. A phylogenetic analysis of the genus *Saccharomyces* based on 18S rRNA gene sequences: description of *Saccharomyces kunashirensis* sp. nov. and *Saccharomyces martiniae* sp. nov. *Int. J. Syst. Bacteriol.* **47**:453-460.
- Kan, V. L. 1993. Polymerase chain reaction for the diagnosis of candidemia. *J. Infect. Dis.* **168**:779-783.
- Kappe, R., C. N. Okeke, C. Fauser, M. Maiwald, and H.-G. Sonntag. 1998. Molecular probes for the detection of pathogenic fungi in the presence of human tissue. *J. Med. Microbiol.* **47**:811-820.
- Kwok, S., and R. Higuchi. 1989. Avoiding false positives with PCR. *Nature* **339**:237-238.
- Maiwald, M., H.-J. Ditton, H.-G. Sonntag, and M. von Knebel-Doberitz. 1994. Characterization of contaminating DNA in *Taq* polymerase which occurs during amplification with a primer set for *Legionella* 5S ribosomal RNA. *Mol. Cell. Probes* **8**:11-14.
- Miyakawa, Y., T. Mabuchi, and Y. Fukazawa. 1993. New method for detection of *Candida albicans* in human blood by polymerase chain reaction. *J. Clin. Microbiol.* **31**:3344-3347.
- Müller, F.-M. C., K. E. Werner, M. Kasai, A. Francesconi, S. J. Chanock, and T. J. Walsh. 1998. Rapid extraction of genomic DNA from medically important yeasts and filamentous fungi by high-speed cell disruption. *J. Clin. Microbiol.* **36**:1625-1629.
- Rand, K. H., H. Houck, and M. Wolff. 1994. Detection of candidemia by polymerase chain reaction. *Mol. Cell. Probes* **8**:215-222.
- van Burik, J.-A., D. Myerson, R. W. Schreckhise, and R. A. Bowden. 1998. Panfungal PCR assay for detection of fungal infection in human blood specimens. *J. Clin. Microbiol.* **36**:1169-1175.
- van Deventer, A. J. M., W. H. F. Goessens, A. van Belkum, H. J. A. van Vliet, E. W. M. van Etten, and H. A. Verbrugh. 1995. Improved detection of *Candida albicans* by PCR in blood of neutropenic mice with systemic candidiasis. *J. Clin. Microbiol.* **33**:625-628.