




Since January 2020 Elsevier has created a COVID-19 resource centre with free information in English and Mandarin on the novel coronavirus COVID-19. The COVID-19 resource centre is hosted on Elsevier Connect, the company's public news and information website.

Elsevier hereby grants permission to make all its COVID-19-related research that is available on the COVID-19 resource centre - including this research content - immediately available in PubMed Central and other publicly funded repositories, such as the WHO COVID database with rights for unrestricted research re-use and analyses in any form or by any means with acknowledgement of the original source. These permissions are granted for free by Elsevier for as long as the COVID-19 resource centre remains active.



## Comparison of SARS-CoV-2 spike antibody quantitative titer reporting using the World Health Organization International Standard Units by four commercial assays

Ran Zhuo<sup>a,\*</sup>, Carmen Charlton<sup>a,b,c</sup>, Sabrina Plitt<sup>d,e</sup>, L. Alexa Thompson<sup>b</sup>, Sheila Braun<sup>f</sup>, Jacqueline Day<sup>g</sup>, Carla Osiowy<sup>g,h</sup>, Graham Tipples<sup>a,c,i</sup>, Jamil N Kanji<sup>a,b,f,j,k</sup>

<sup>a</sup> Public Health Laboratory, Alberta Precision Laboratories, Edmonton, Alberta, Canada

<sup>b</sup> Department of Laboratory Medicine and Pathology, University of Alberta, Edmonton, Alberta, Canada

<sup>c</sup> Li Ka Shing Institute of Virology, University of Alberta, Edmonton, Alberta, Canada

<sup>d</sup> School of Public Health, University of Alberta, Edmonton, Alberta, Canada

<sup>e</sup> Centre for Communicable Diseases and Infection Control, Public Health Agency of Canada, Ottawa, Ontario, Canada

<sup>f</sup> Public Health Laboratory, Alberta Precision Laboratories, Calgary, Alberta, Canada

<sup>g</sup> National Microbiology Laboratory, Public Health Agency of Canada, Winnipeg, MB, Canada

<sup>h</sup> University of Manitoba, Winnipeg, MB, Canada

<sup>i</sup> Department of Medical Microbiology and Immunology, University of Alberta, Edmonton, Alberta, Canada

<sup>j</sup> Section of Medical Microbiology, Department of Pathology and Laboratory Medicine, Cumming School of Medicine, University of Calgary, Calgary, Alberta, Canada

<sup>k</sup> Division of Infectious Diseases, Department of Medicine, University of Calgary, Calgary, Alberta, Canada

### ARTICLE INFO

#### Keywords:

COVID-19  
SARS-CoV-2  
Quantitative serological assays  
WHO international standard  
Antibodies  
Immunoglobulin

### ABSTRACT

The accurate measurement of serological response to SARS-CoV-2 vaccination is needed to correlate responses with effective protective immunity. The World Health Organization (WHO) has created an international standard to allow harmonization of immune response assessment to an arbitrary unit across different commercial assays; however, the accuracy of reporting of SARS-CoV-2 spike antibody titers in international standard units (BAU or IU/mL) from commercial assays is not well studied. Here, we report the performance comparison of four quantitative commercial assays testing for SARS-CoV-2 spike immunoglobins using the WHO's international standard. Sera, EDTA-plasma and heparinized plasma collected from individuals who are vaccine naïve or received BNT162b2 (Pfizer/BioNTech), mRNA-1273 (Moderna) or ChAdOx1-S (Oxford-AstraZeneca) were tested using Abbott Architect AdviseDx SARS-CoV-2 IgG II, DiaSorin LIAISON SARS-CoV-2 TrimericS IgG, Roche Elecsys Anti-SARS-CoV-2 S and GenScript cPass SARS-CoV-2 surrogate virus neutralization assays. The sensitivities ranged from 90% to 100%, and specificities from 88% to 100%. These four assays had excellent agreement (0.79–0.93) and correlation (0.87–0.97); however, Passing-Bablok regression analysis indicated that data generated by these assays were not comparable. Our data suggests that natural SARS-CoV-2 infection elicited a greater antibody response compared to vaccines, evident by a significantly higher neutralizing antibody titer in unvaccinated individuals who seroconverted.

### Abbreviations

BAU binding arbitrary unit  
CLIA chemiluminescent immunoassay  
ELISA enzyme-linked immunosorbent assays  
Ig immunoglobulin  
IS International Standard  
IU international unit

SARS-CoV 2 severe acute respiratory syndrome coronavirus 2  
WHO World Health Organization

### 1. Introduction

During the severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) pandemic, many serology assays were rapidly deployed into the market [1,2]. These serology assays utilize different chemistry [lateral

\* Corresponding author.

E-mail address: [rzhuo@ualberta.ca](mailto:rzhuo@ualberta.ca) (R. Zhuo).

<https://doi.org/10.1016/j.jcv.2022.105292>

Received 24 August 2022; Received in revised form 30 August 2022;

Available online 6 September 2022

1386-6532/© 2022 Elsevier B.V. All rights reserved.

flow immunoassays, enzyme-linked immunosorbent assays (ELISA), chemiluminescent immunoassay (CLIA)], target different antigens (spike glycoprotein, nucleocapsid) or immunoglobulin classes, and provide qualitative, semiquantitative or quantitative results [3–6].

Antibody tests can be used to confirm retrospective infection in cases of high clinical suspicion (when diagnostic testing may not have been performed, or was previously negative), for diagnosis of late-onset post-infection complications, such as the multisystem inflammatory syndrome in children [7,8]. Additionally, population-based seroprevalence can be used to evaluate the success of infection control measures at the community level [9,10], and estimate asymptomatic/non-reported symptomatic SARS-CoV-2 infection rates. Several SARS-CoV-2 vaccines based on mRNA or live adenovirus vector expressing the spike (S) glycoprotein have been approved and are widely available globally [11, 12]. As such, an increasing proportion of the world population has received at least one dose of vaccine [13,14]. Subsequently, serologic assays have become essential tools to measure vaccine induced humoral responses to the spike protein to better understand antibody longevity and effectiveness.

However, our current understanding of the effectiveness of vaccines to decrease transmission of new variants of concern, the duration of vaccine elicited immunity, and its correlation to protection against breakthrough infections are inadequate [15–18]. The challenge of comparing results between studies is the lack of harmonization of quantitative standards, as commercial assay manufacturers use their own arbitrary units.

The World Health Organization (WHO) Expert Committee on Biological Standards established an international standard (IS) for SARS-CoV-2 immunoglobulin based on pooled human plasma from convalescent patients with the aim to facilitate accurate calibration of serological assays for SARS-CoV-2 immunoglobulin detection [19,20]. The preparation is assigned the concentration of 1000 international unit per milliliter (IU/mL) for neutralizing antibody activity and is considered equivalent to 1000 binding antibody units per milliliter (BAU/mL), an arbitrary unit that can be used in comparison studies of binding antibody assays. Therefore, both units are considered numerically equivalent [19].

Recently, some commercial serological assays for SARS-CoV-2 have implemented conversion factors to allow antibody titers to be reported in international units; however, the accuracy of reporting in WHO IS units by these commercial assays is not well characterized.

Here, we evaluated serologic reporting of three commercial automated binding antibody immunoassays and one surrogate virus neutralization ELISA for anti-SARS-CoV-2 antibodies in WHO established IS units using clinically collected plasma or serum from vaccine recipients or vaccine naïve individuals. We describe the demographic characteristics of the donors and studied SARS-CoV-2 antibody levels associated with vaccination status.

## 2. Methods

### 2.1. Study participants and sample acquisition

Thirty SARS-CoV-2 antibody positive EDTA plasma, sera, and heparinized plasma samples, and 30 serologically negative samples of each type (based on previous Abbott Architect AdviseDx SARS-CoV-2 IgG II results) were used in this study. The WHO IS for Anti-SARS-CoV-2 Immunoglobulin was purchased from the National Institute for Biological Standards and Control (NIBSC, Hertfordshire, UK).

### 2.2. SARS-CoV-2 serologic testing

Abbott Architect AdviseDx SARS-CoV-2 IgG II Quant, DiaSorin TrimericS IgG, Roche Elecsys Anti-SARS-CoV-2 S, and GenScript cPass SARS-CoV-2 surrogate virus neutralization assays were used as per manufacturer recommendations to assess antibody titer [21–24].

### 2.3. Diagnostic sensitivity and specificity calculations

Assay diagnostic sensitivities and specificities were calculated using clinically defined anti-SARS-CoV-2 positive samples as the reference standard in SPSS v26 (IBM SPSS Statistics for Windows, Armonk, NY, USA). Clinically positive samples were defined as any sample determined as anti-SARS-CoV-2 positive by three of the four assays used in this study.

### 2.4. Statistical analysis

Data were summarized by percentages and frequencies for categorical variables, and medians and interquartile ranges (IQRs) for continuous variables. Categorical variables were compared using the Chi-square test or Fisher's exact test where appropriate; the continuous variables (age and antibody titers) were compared using Mann-Whitney U test or Kruskal-Wallis test with Dunn's multiple comparison correction. Assay agreements were analyzed with Cohen's Kappa and Spearman rank correlations. We defined kappa scores of 0.6 to 0.69 as moderate, 0.7 to 0.8 as good and above 0.8 as excellent agreement [25]. Assay results were compared with Passing-Bablok linear regression and Bland-Altman plots using STATA v16 (StataCorp LLC, College Station, TX, USA) [26]. All other statistical analyses were performed using SPSSv25 (IBM SPSS Statistics for Windows, Armonk, NY, USA) [27]. Graphs were generated in GraphPad Prism v9 (GraphPad Software, San Diego, CA, USA).

### 2.5. Ethics

Ethics approval was granted by the University of Alberta research ethics board (study reference numbers Pro00101916 and Pro00113764).

## 3. Results

Anti-SARS-CoV-2 spike seropositive individuals were older [Median age 46 vs 34, ( $P < 0.05$ )] and more likely to have received one or two doses of mRNA-1273 or BNT162b2 vaccines ( $P < 0.001$ ) (Table 1).

The Roche Elecsys assay reported the highest number of SARS-CoV-2 positives (10–13% more than Abbott RBD) across all sample types and the widest range of antibody titers. DiaSorin trimeric S assay reported the highest minimal titers, while Abbott RBD had the narrowest dynamic range (Table 2, Fig. 1). Eighty-eight percent (25/30) of EDTA plasma and 75% (24/32) of heparin plasma were interpreted as positive on both Abbott RBD and GenScript cPass assays, with 100% concordance for serum positivity.

Inter-assay agreement between the Abbott RBD, DiaSorin TrimericS and Roche Elecsys assays was excellent, with kappa scores ranging from 0.88 to 0.93; the agreement between GenScript cPass and the other three assays was slightly lower ( $k = 0.78 - 0.88$ ) (Table 3). The four assays correlated very well with Spearman's rho above 0.87; however, Passing-Bablok regression revealed the average measured values between any two of the four assays were significantly different (Table 3, Figure S1).

Direct testing of the WHO IS showed the Roche Elecsys was able to detect antibodies in up to 1000-fold dilutions and quantify titers as low as 1 BAU/mL (Table S1). Abbott RBD was able to identify IgG in dilutions corresponding to titers of 100 and 10 BAU/mL, while DiaSorin TrimericS and GeneScript cPass were able to detect and quantify titers as low as 100 BAU or IU/mL (Table S1).

All assays except GenScript cPass (89.8%) had sensitivities of 100% when compared to clinically defined positive samples (reference standard; Table 4). The specificity of Abbott RBD, DiaSorin TrimericS and Roche Elecsys ranged from 88% to 98% while GenScript cPass was 100% (Table 4).

Of the 90 samples originally chosen as positives, two (2.2%) tested positive on only two platforms and were re-classified as clinically

**Table 1**Donor demographic characteristics and SARS-CoV-2 vaccination status by serology status ( $n = 180$ ).

	SARS-CoV-2 Serology Status						P-value*
	Antibody negativen = 90			Antibody positiven = 90			
Age, Median (IQR)		34	(9, 54)	46	(9, 71)		0.5
Age Group, n (%)	under 18	29	32.2%	31	34.4%		0.003
	18–25	7	7.8%	0	0.0%		
	26–45	22	24.4%	14	15.6%		
	46–55	11	12.2%	9	10.0%		
	56–65	12	13.3%	8	8.9%		
	66–75	5	5.6%	14	15.6%		
	above 75	4	4.4%	14	15.6%		0.15
Sex, n (%)	Female	45	50.0%	35	38.9%		
	Male	15	16.7%	25	27.8%		
	Unknown	30	33.3%	30	33.3%		<0.001
Vaccination Status	Unvaccinated	73	81.1%	1	1.1%		
	1 Dose	9	10.0%	38	42.2%		
	2 Doses	2	2.2%	20	22.2%		
	Unknown	6	6.7%	31	34.4%		
Vaccine dose 1 Type (brand)	AstraZeneca	3	3.3%	5	5.6%		<0.001
	Moderna	1	1.1%	12	13.3%		
	Pfizer/BioNTech	7	7.8%	41	45.6%		
	Not Immunized	79	87.8%	32	35.6%		
Vaccine dose 2 type (brand)	Moderna	0	0.0%	8	8.9%		<0.001
	Pfizer/BioNTech	2	2.2%	23	25.6%		
	Not Immunized	88	97.8%	59	65.6%		

Serology status was defined based on the first Abbott Architect RBD results used for sample selection.

\* Mann-Whitney U test for Age and Chi-square test or Fisher's exact test for all categorical variables, unknown or not immunized groups not included in the analyses.

**Table 2**

Comparison of test measurements produced by Abbott RBD, DiaSorin Trimeric S, Roche Elecsys and GenScript cPass assays in EDTA plasma, heparin plasma and serum samples.

	Abbott RBD	DiaSorin TrimericS	Roche Elecsys	GenScript cPass	P-value
<b>EDTA Plasma, n = 60</b>					
Positive, n	30	31	34	25	
Titer, BAU/mL					0.30
Min	17.2	39	0.9	36.1	
Median, IQR	130.4 (47.2, 431.9)	291 (140, 1190)	138.6 (38.3, 1764.4)	259.2 (77, 453.6)	
Max	2405.9	33,100	54,001.7	21,021.3	
<b>Heparin Plasma, n = 60</b>					
Positive, n	32	29	32	24	
Titer, BAU/mL					0.10
Min	7.4	46	3.0	37.0	
Median, IQR	114.8 (27.3, 561.7)	419 (145, 1120)	97.6 (34, 640.5)	150.1 (69.6, 433.4)	
Max	2467.7	7060	10,293.1	2913.8	
<b>Serum, n = 60</b>					
Positive, n	30	30	33	30	
Titer, BAU/mL					<0.0001
Min	231.6	614	2.5	320.3	
Median, IQR	923 (465.8, 1670)	1750 (1170, 3000)	5917.7 (3010.3, 8154.3)	1279 (806.3, 2267.5)	
Max	6056.0	8770	33,954.5	9616.9	

P-values generated from comparing medians using the Kruskal-Wallis test, significant at  $< 0.05$ .

Min: minimum.

Max: maximum.

negative. Nine (10%) and 79 (87.8%) tested positive on three and four platforms, respectively. Two previous Abbott RBD negative heparin plasma samples were reclassified as positive by Abbott RBD (Table 2). Overall, SARS-CoV-2 antibody titers were found to be five- to ten-fold higher in serum compared to plasma sample types and significant

differences ( $P < 0.0001$ ) were observed (Table 1, Table S2–3).

Antibody titers in vaccine naïve participants, who were presumably infected by SARS-CoV-2 naturally, were higher than vaccinated participants who received one dose of either ChAdOx1-S or BNT162b2, quantified by all four assays (Fig. 2). Roche Elecsys reported a dosage dependent increase of antibody titer induced by BNT162b2 vaccine which was not found by the other three assays (Fig. 2). Neutralizing antibody levels were found to be significantly higher in seropositive vaccine naïve individuals than those who received ChAdOx1-S or BNT162b2 vaccines by GenScript cPass regardless of doses received (Fig. 2).

#### 4. Discussion

Overall, we observed excellent correlation ( $r = 0.87–0.97$ ) and agreement ( $k = 0.78–0.93$ ) between the three binding antibody assays and the neutralization antibody ELISA. These assays are suitable to be used clinically; however, data generated in standardized WHO IS units are not comparable, as Passing-Bablok regression revealed significant differences in proportional measurements between assays.

Abbott RBD and DiaSorin TrimericS had the best correlation, whereas Roche Elecsys and GenScript cPass results differed the most. The difference in assay chemistry, antigen specificity, and antibody class targeted (IgG vs. total Ig vs. pan-Ig neutralizing antibodies) may explain the small differences in diagnostic sensitivities and specificities, especially between GenScript cPass and Roche Elecsys S, because neutralizing antibodies are only a small proportion of the total antibodies measured by Elecsys S. It may also explain the observed differences in dynamic range (Roche Elecsys vs Abbott RBD) and minimal titer detected (DiaSorin TrimericS).

As the WHO IS is comprised of pooled plasma from eleven different SARS-CoV-2 convalescent patients [and therefore contains various isoforms of antibodies (IgA, IgM and IgG) targeting numerous epitopes on the spike protein, with some neutralizing ability [28], it is not surprising that assay platforms did not produce equivalent results given their different antibody class targets and epitope targets (RBD or trimeric spike protein). However, despite these differences, the numerical values of the positive dilutions produced by these assays are very similar to the expected WHO IS titer (Table S1).

To date, very few comparison studies exist that assess SARS-CoV-2




Fig. 1. SARS-CoV-2 antibody titer reported in binding antibody unit per mL by Architect RBD, DiaSorin TrimericS, Roche Elecsys and GenScript cPass in EDTA plasma, heparin plasma and serum. The dots depict titers of individual samples, and the lines represent median titer values and interquartile ranges.

**Table 3**  
Agreements and correlation of SARS-CoV-2 antibody titer in positive samples between Abbott RBD, DiaSorin Trimeric S, Roche Elecsys and GenScript cPass assays.

Assay pair	Cohen's k	Spearman's rho (95% CI)	Passing-Bablok average difference (95%CI)
RBD/ Trimeric S	0.93 (95% CI 0.88, 0.99) <sup>ε</sup>	0.97 (95%CI 0.96, 0.98)*	-74.95 (-10,101.94, -9952.04)**
RBD/Elecsys	0.92 (95% CI 0.87, 0.98) <sup>ε</sup>	0.87 (95%CI 0.81, 0.92)*	-2280.84 (-13,052.84, -8491.16)**
RBD/cPass	0.86 (95% CI 0.78, 0.93) <sup>ε</sup>	0.94 (95%CI 0.90, 0.96)*	573.27 (-12,507.97, -13,654.50)**
Trimeric S/ Elecsys	0.88 (95% CI 0.81, 0.95) <sup>ε</sup>	0.90 (95%CI 0.86, 0.93)*	-2205.89 (-11,716.78, -7304.99)**
Trimeric S/ cPass	0.88 (95% CI 0.81, 0.95) <sup>ε</sup>	0.94 (95%CI 0.91, 0.96)*	650.93 (-2231.08, -3532.94)**
Elecsys/ cPass	0.78 (95% CI 0.69, 0.87 s) <sup>δ</sup>	0.90 (95%CI 0.84, 0.93)*	3113.32 (-8083.30, -14,309.94)**

\* Correlation is significant at  $p < 0.01$ .  
 \*\* Average differences are significant at  $p < 0.05$ . Differences are measured in BAU/mL.  
<sup>δ</sup> Cohen's kappa score of 0.7–0.8 was defined as good.  
<sup>ε</sup> above 0.8 as excellent agreement.

**Table 4**  
Sensitivity and specificity of RBD, TrimericS, Elecsys and cPass calculated by using user-defined clinically positive reference standard.

N = 180	Compared to clinically positive Samples (3/4 assays positive)					
	Sensitivity	Specificity	TP	TN	FP	FN
Abbott Architect RBD	100.0%	95.7%	88	88	4	0
DiaSorin Trimeric S	100.0%	97.8%	88	90	2	0
Roche Elecsys	100.0%	88.0%	88	81	11	0
GenScript cPass	89.8%	100.0%	79	92	0	9

commercial assay performance against the WHO IS. Studies not using standardized unit BAU/mL found that numerical results of the Abbott RBD, DiaSorin Trimeric and Roche Elecsys S test kits are not interchangeable [29] and not surprisingly, the sensitivity (93.6–96%), correlation ( $r = 0.8–0.9$ ) and agreement ( $k = 0.6–0.8$ ) between the four assays was lower than we observed in this study (Table 2 and Table 3) [30, 31]. In line with our findings, Bradley et al., reported equivalent correlation between the Abbott RBD assays to Roche Elecsys S ( $r = 0.83$  vs 0.87); however, we observed an improved correlation with GenScript

cPass using standardized IU/mL ( $r = 0.86$  vs 0.94) rather than% neutralization [32]. In fact, 100% concordance of the Abbott RBD, Roche Elecsys S and GenScript cPass results was achieved in both Bradley et al., and our study, highlighting the importance of using the WHO IS when comparing methods (Table S1). Other groups described similar but slightly lower correlations (range 0.76–0.8) with Roche Elecsys S, Abbott RBD, and DiaSorin TrimericS to a virus neutralization test and correlation ranges from 0.5 to 0.9 between these three tests using the WHO IS units [33–35].

Average proportional differences of measured values between the four serological assays were revealed using Passing-Bablok regression analysis. Consistent with our findings, Perkmann et al., and Lukaszuk et al., found that Abbott RBD, DiaSorin TrimericS, Roche Elecsys had high proportional errors using Passing-Bablok regression, suggesting that results of these assays are not interchangeable [31,36]. An increased effort by the manufacturers to standardize assay output across platforms would be helpful to achieve cross-utility. Until then, comparison of results between assays should be interpreted with care.

We found a higher titer of neutralizing antibodies measured by the GenScript cPass in vaccine naïve seropositive individuals compared to those who received ChAdOx1-S or BNT162b2 regardless of dose (Fig. 2). The number of mRNA-1273 recipients were too small to draw any solid conclusions. This observation suggests that natural infection induces higher levels of antibody compared to vaccination. The ChAdOx1-S and the two mRNA vaccines were formulated to encode only the full-length spike protein [21,37,38], therefore individuals with natural infections likely generate polyclonal antibodies against multiple epitopes of the whole virus (including neutralizing antibodies) at a higher level. Moreover, the vaccine naïve participants belong to a young age group (Median age=9): age is another factor that may contribute to the increase in neutralizing antibody level observed, as children were shown to produce a robust neutralizing antibody response after SARS-CoV-2 infection [39]. In general, children with any upper respiratory infection may have higher viral loads than adults [40]. Recent research has suggested that natural infection generates longer antibody duration against future COVID-19 illnesses than two doses of vaccine in the previously unexposed individuals. Vaccinated individuals with prior infection have the highest level and longest duration of antibody response, which remains high after 12 months post vaccination [41,42]. However, more evidence is needed to gain a better understanding of the protective immunity induced by natural infection, or combinations of vaccine, and what risk factors may influence the immunity against breakthrough infections. Further standardization of serological assays would allow analysis using pooled data across studies to facilitate that.






Fig. 2. Anti-SARS-CoV-2 antibody titers in vaccinated and vaccine naïve participants quantified in BAU/mL by the four serologic assays under study.

Because the three types of samples used in this study were not paired (i.e. not collected from the same individuals), higher antibody titers observed in serum specimens cannot be considered an intrinsic characteristic of the specimen type, but rather reflects the characteristics of serological responses in the donors [people of all ages (plasma) vs. those under 25 (serum)]. In samples with higher levels of antibody, serum

reported a broader range of antibody level detected between the four assays compared to plasma specimens. There is likely a larger intrinsic error range at the upper limits of the quantifiable range, where the assays were saturated, which is where more of the serum specimen titers fall compared to plasma sample types (Fig. 1).

One of the limitations of this study is that we were unable to link

prior SARS-CoV-2 infection directly to vaccine naïve participants who tested positive for SARS-CoV-2 spike or neutralizing antibody, as nucleic acid amplification test (NAAT) results and anti-nucleocapsid serologic results were unavailable; however, the majority (24/32, 75%) of specimens were from children under 12 who were (at the time of specimen collection) ineligible for vaccination as per provincial immunization guidelines. This suggests the observed seroconversion was not due to vaccination but rather natural infection. Furthermore, the duration from prior SARS-CoV-2 infections to time of sample collection was unclear in unimmunized seropositive individuals. Therefore, the higher anti-SARS-CoV-2 immunoglobulin level observed in vaccine naïve participants (Fig 2) could be due to the difference in sample collection time and may not necessarily indicate that better protection is elicited by natural infection compared to vaccination. Lastly, the study is limited by a relatively small sample size and samples collected prior to September 2021 thus more recent samples are needed to evaluate assay performance on serologic responses to new variants (e.g. omicron) that harbor multiple mutations in the RBD.

In summary, we compared the quantitative reporting of SARS-CoV-2 spike immunoglobulin titers in WHO established international standard units by Abbott RBD, DiaSorin TrimericS, Roche Elecsys S and GenScript cPass in serum and plasma types. We found these assays were in excellent agreement and are suitable to be used clinically, but the data generated in standardized WHO IS units are not interchangeable. SARS-CoV-2 natural infections appeared to induce a greater level of neutralizing antibodies than vaccinations. Additional work to harmonize serologic assays for SARS-CoV-2 using the WHO international standard would be beneficial.

#### Declaration of Competing Interest

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

#### Acknowledgements

The authors thank Sean C. Taylor at GenScript USA inc. for providing the GenScript cPass SARS-CoV-2 neutralization antibody detection kits and the technologists at Public Health Laboratory Calgary, Alberta Precision Laboratories for specimen retrieval and generating the Abbott RBD and DiaSorin Trimeric S test results. This research did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.

#### Supplementary materials

Supplementary material associated with this article can be found, in the online version, at [doi:10.1016/j.jcv.2022.105292](https://doi.org/10.1016/j.jcv.2022.105292).

#### References

- R.M. West, A. Kobokovich, N. Connell, G.K. Gronvall, Antibody (Serology) tests for COVID-19: a case study, *mSphere* 6 (2021).
- L.A. Pinto, R.M. Shawar, B. O'Leary, T.J. Kemp, J. Cherry, N. Thornburg, et al., A trans-governmental collaboration to independently evaluate SARS-CoV-2 serology assays, *Microbiol Spectr.* 10 (2022), e0156421.
- Theel E.S. SARS-CoV-2 Serologic Testing, 4th edition. ed 2021.
- P. Van Caesele, Canadian Public Health Laboratory N, D. Bailey, Canadian society of clinical C, Forgie SE, association of medical M, et al. SARS-CoV-2 (COVID-19) serology: implications for clinical practice, laboratory medicine and public health, *CMAJ* 192 (2020). E973-E9.
- M. Lisboa Bastos, G. Tavaziva, S.K. Abidi, J.R. Campbell, L.P. Haraoui, J. C. Johnston, et al., Diagnostic accuracy of serological tests for covid-19: systematic review and meta-analysis, *BMJ* 370 (2020) m2516.
- A. Vengesai, H. Midzi, M. Kasambala, H. Mutandadzi, T.L. Mduleza-Jokonya, S. Rusakaniko, et al., A systematic and meta-analysis review on the diagnostic accuracy of antibodies in the serological diagnosis of COVID-19, *Syst. Rev.* 10 (2021) 155.
- L.R. Feldstein, E.B. Rose, S.M. Horwitz, J.P. Collins, M.M. Newhams, M.B.F. Son, et al., Multisystem inflammatory syndrome in U.S. children and adolescents, *N. Engl. J. Med.* 383 (2020) 334–346.
- D.S.Y. Ong, P.C. Fragkou, V.A. Schweitzer, R.F. Chemaly, C.D. Moschopoulos, C. Skevaki, et al., How to interpret and use COVID-19 serology and immunology tests, *Clin. Microbiol. Infect.* 27 (2021) 981–986.
- R. West, A. Kobokovich, N. Connell, G.K. Gronvall, COVID-19 antibody tests: a valuable public health tool with limited relevance to individuals, *Trends Microbiol.* 29 (2021) 214–223.
- A.M. Lerner, R.W. Eisinger, D.R. Lowy, L.R. Petersen, R. Humes, M. Hepburn, et al., The COVID-19 serology studies workshop: recommendations and challenges, *Immunity* 53 (2020) 1–5.
- E.E. Walsh, R.W. Frencik Jr., A.R. Falsey, N. Kitchin, J. Absalon, A. Gurtman, et al., Safety and immunogenicity of two RNA-based Covid-19 vaccine candidates, *N. Engl. J. Med.* 383 (2020) 2439–2450.
- M. Voysey, S.A.C. Clemens, S.A. Madhi, L.Y. Weckx, P.M. Folegatti, P.K. Aley, et al., Safety and efficacy of the ChAdOx1 nCoV-19 vaccine (AZD1222) against SARS-CoV-2: an interim analysis of four randomised controlled trials in Brazil, South Africa, and the UK, *Lancet* 397 (2021) 99–111.
- E. Mathieu, H. Ritchie, E. Ortiz-Ospina, M. Roser, J. Hasell, C. Appel, et al., A global database of COVID-19 vaccinations, *Nat. Hum. Behav.* 5 (2021) 947–953.
- E. Mathieu, H. Ritchie, E. Ortiz-Ospina, M. Roser, J. Hasell, C. Appel, et al., Author correction: a global database of COVID-19 vaccinations, *Nat. Hum. Behav.* 5 (2021) 956–959.
- R.W. Peeling, D.L. Heymann, Y.Y. Teo, P.J. Garcia, Diagnostics for COVID-19: moving from pandemic response to control, *Lancet* 399 (2022) 757–768.
- M. Ciotti, M. Ciccozzi, M. Pieri, S. Bernardini, The COVID-19 pandemic: viral variants and vaccine efficacy, *Crit. Rev. Clin. Lab. Sci.* 59 (2022) 66–75.
- Y. Galipeau, M. Greig, G. Liu, M. Driedger, M.A. Langlois, Humoral responses and serological assays in SARS-CoV-2 infections, *Front. Immunol.* 11 (2020), 610688.
- E. Lee, J.E. Oh, Humoral immunity against SARS-CoV-2 and the impact on COVID-19 pathogenesis, *Mol. Cells* 44 (2021) 392–400.
- P.A. Kristiansen, M. Page, V. Bernasconi, G. Mattiuzzo, P. Dull, K. Makar, et al., WHO international standard for anti-SARS-CoV-2 immunoglobulin, *Lancet* 397 (2021) 1347–1348.
- F. Zhu, T. Althaus, C.W. Tan, A. Costantini, W.N. Chia, N. Van Vinh Chau, et al., WHO international standard for SARS-CoV-2 antibodies to determine markers of protection, *Lancet Microbe* 3 (2022) e81-e2.
- L.R. Baden, H.M. El Sahly, B. Essink, K. Kotloff, S. Frey, R. Novak, et al., Efficacy and safety of the mRNA-1273 SARS-CoV-2 vaccine, *N. Engl. J. Med.* 384 (2021) 403–416.
- F. Bonelli, F.A. Blocki, T. Bunnell, E. Chu, O.A. De La, D.G. Grenache, et al., Evaluation of the automated LIAISON(RR) SARS-CoV-2 TrimericS IgG assay for the detection of circulating antibodies, *Clin. Chem. Lab Med.* 59 (2021) 1463–1467.
- E. Riestler, P. Findeisen, J.K. Hegel, M. Kabsch, C.M. Rank, et al., Performance evaluation of the Roche Elecsys Anti-SARS-CoV-2 S immunoassay, *J. Virol. Methods* 297 (2021), 114271.
- S.C. Taylor, B. Hurst, C.L. Charlton, A. Bailey, J.N. Kanji, M.K. McCarthy, et al., A new SARS-CoV-2 dual-purpose serology test: highly accurate infection tracing and neutralizing antibody response detection, *J. Clin. Microbiol.* (2021) 59.
- M.L. McHugh, Interrater reliability: the kappa statistic, *Biochem Med (Zagreb)* 22 (2012) 276–282.
- StataCorp. Stata statistical software release 15. 2017.
- Corp I. IBM SPSS statistics for windows, version 25.0. 2017.
- Mattiuzzo GB, E.M.; Hassall, M.; Routley, S.; Richardson, S.; Bernasconi, V.; Kristiansen, P.; Harvala, H.; Roberts, D.; Semple, M.G.; Turtle, L.C.W.; Openshaw P.J.M and Kenneth Baillie on behalf of the ISARIC4C Investigators, Lise Sofie Haug Nissen-Meyer, Arne Broch Brantsæter, Helen Baxendale, Eleanor Atkinson, Peter Rigsby, David Padley, Neil Almond, Nicola J. Rose, Mark page and the collaborative study participants. Establishment of the WHO international standard and reference panel for anti-SARS-CoV-2 antibody. In: WHO the Expanded Programme on Immunization of the Department of Immunization VaB, editor.z. p. 60.
- A. Di Meo, J.J. Miller, A. Fabros, D. Brinc, V. Hall, N. Pinzon, et al., Evaluation of three Anti-SARS-CoV-2 serologic immunoassays for post-vaccine response, *J. Appl. Lab Med.* 7 (2022) 57–65.
- K. Jung, S. Shin, M. Nam, Y.J. Hong, E.Y. Roh, K.U. Park, et al., Performance evaluation of three automated quantitative immunoassays and their correlation with a surrogate virus neutralization test in coronavirus disease 19 patients and pre-pandemic controls, *J. Clin. Lab Anal.* 35 (2021) e23921.
- T. Perkmann, N. Perkmann-Nagele, T. Koller, P. Mucher, A. Radakovic, R. Marculescu, et al., Anti-Spike Protein Assays to Determine SARS-CoV-2 antibody levels: a Head-to-Head comparison of five quantitative assays, *Microbiol. Spectr.* 9 (2021), e0024721.
- B.T. Bradley, A. Bryan, S.L. Fink, E.A. Goecker, P. Roychoudhury, M.L. Huang, et al., Anti-SARS-CoV-2 antibody levels measured by the AdviseDx SARS-CoV-2 assay are concordant with previously available serologic assays but are not fully predictive of sterilizing immunity, *J. Clin. Microbiol.* 59 (2021), e0098921.
- A.E. Egger, C. Irsara, B. Holzer, C. Winkler, R. Bellmann-Weiler, G. Weiss, et al., Borderline and weakly positive antibody levels against the S-protein of SARS-CoV-2 exhibit limited agreement with virus neutralization titres, *J. Clin. Virol. Plus* 2 (2022), 100058.
- K. Saker, V. Escuret, V. Pitiot, A. Massardier-Pilonchery, S. Paul, B. Mokdad, et al., Evaluation of commercial Anti-SARS-CoV-2 antibody assays and comparison of standardized titers in vaccinated health care workers, *J. Clin. Microbiol.* 60 (2022), e0174621.

- [35] M. Infantino, M. Pieri, M. Nuccetelli, V. Grossi, B. Lari, F. Tomassetti, et al., The WHO international standard for COVID-19 serological tests: towards harmonization of anti-spike assays, *Int. Immunopharmacol.* 100 (2021), 108095.
- [36] K. Lukaszuk, J. Kiewisz, K. Rozanska, M. Dabrowska, A. Podolak, G. Jakiel, et al., Usefulness of IVD kits for the assessment of SARS-CoV-2 antibodies to evaluate the humoral response to vaccination, *Vaccines (Basel)* (2021) 9.
- [37] J. Sadoff, M. Le Gars, G. Shukarev, D. Heerwegh, C. Truyers, A.M. de Groot, et al., Interim results of a phase 1-2a trial of Ad26.COV2.S Covid-19 vaccine, *N. Engl. J. Med.* 384 (2021) 1824–1835.
- [38] F.P. Polack, S.J. Thomas, N. Kitchin, J. Absalon, A. Gurtman, S. Lockhart, et al., Safety and efficacy of the BNT162b2 mRNA Covid-19 vaccine, *N Engl J Med* 383 (2020) 2603–2615.
- [39] P. Liu, J. Cai, R. Jia, S. Xia, X. Wang, L. Cao, et al., Dynamic surveillance of SARS-CoV-2 shedding and neutralizing antibody in children with COVID-19, *Emerg Microbes Infect.* 9 (2020) 1254–1258.
- [40] C.L. Charlton, E. Babady, C.C. Ginocchio, T.F. Hatchette, R.C. Jerris, Y. Li, et al., Practical Guidance for Clinical Microbiology Laboratories: viruses Causing Acute Respiratory Tract Infections, *Clin. Microbiol. Rev.* 32 (2019).
- [41] V. Hall, S. Foulkes, F. Insalata, P. Kirwan, A. Saei, A. Atti, et al., Protection against SARS-CoV-2 after Covid-19 vaccination and previous infection, *N. Engl. J. Med.* 386 (2022) 1207–1220.
- [42] G. Milne, T. Hames, C. Scotton, N. Gent, A. Johnsen, R.M. Anderson, et al., Does infection with or vaccination against SARS-CoV-2 lead to lasting immunity? *Lancet Respir. Med.* 9 (2021) 1450–1466.