

Multimedia Appendix 15. Group-averaged macro F1 scores: subgroup analysis of the apnea-hypopnea index and demographic characteristics in Clionic Lifecare Clinic.

	AHI			Sleep Efficiency			Body Mass Index			Gender		
	≤ 15	> 15	P	≤ 85%	> 85%	P	≤ 25	> 25	P	Male	Female	P
Airable												
SleepRoutine (38)	0.54±0. 14(30)	0.68±0. 07(8)	0.012	0.57±0. 19(9)	0.57±0. 12(29)	0.9499	0.57±0. 13(28)	0.59±0. 16(10)	0.6541	0.63±0. 1(12)	0.54±0. 14(26)	0.0587
SleepScore (12)	0.25±0. 11(10)	0.37±0. 13(2)	0.2657	0.21±0. 02(3)	0.29±0. 14(9)	0.4221	0.31±0. 16(6)	0.23±0. 04(6)	0.2963	0.24±0. 0(2)	0.28±0. 14(10)	0.7527
Pillow (37)	0.22±0. 1(30)	0.22±0. 06(7)	0.9687	0.23±0. 13(9)	0.22±0. 08(28)	0.9326	0.23±0. 09(27)	0.2±0.1 1(10)	0.3346	0.23±0. 07(12)	0.22±0. 11(25)	0.6755
Nearable												
Withings Sleep Tracking Mat (38)	0.42±0. 15(30)	0.38±0. 14(8)	0.5694	0.34±0. 16(9)	0.43±0. 14(29)	0.0952	0.42±0. 15(28)	0.38±0. 16(10)	0.5591	0.46±0. 11(12)	0.39±0. 16(26)	0.1826
Google Nest Hub 2 (19)	0.3±0.1 1(16)	0.28±0. 03(3)	0.7758	0.31±0. 13(6)	0.29±0. 09(13)	0.7409	0.32±0. 08(14)	0.24±0. 13(5)	0.1788	0.38±0. 06(2)	0.29±0. 11(17)	0.2574
Amazon Halo Rise (16)	0.55±0. 07(13)	0.63±0. 02(3)	0.0835	0.65±0. 0(1)	0.56±0. 07(15)	-	0.57±0. 07(12)	0.54±0. 07(4)	0.5254	0.56±0. 08(8)	0.57±0. 06(8)	0.6740
Wearable												
Google Pixel Watch (18)	0.55±0. 18(15)	0.59±0. 09(3)	0.7245	0.47±0. 25(6)	0.59±0. 07(12)	0.1535	0.57±0. 08(13)	0.5±0.2 9(5)	0.4219	0.6±0.1 1(2)	0.55±0. 17(16)	0.6642
Galaxy Watch 5 (19)	0.51±0. 11(14)	0.52±0. 09(5)	0.8354	0.44±0. 11(3)	0.52±0. 1(16)	0.2216	0.48±0. 11(14)	0.6±0.0 5(5)	0.0304	0.49±0. 1(10)	0.53±0. 11(9)	0.4871
Fitbit Sense 2 (17)	0.56±0. 17(14)	0.63±0. 09(3)	0.5236	0.51±0. 24(6)	0.61±0. 07(11)	0.2346	0.59±0. 07(13)	0.49±0. 29(4)	0.3002	0.66±0. 06(2)	0.56±0. 17(15)	0.4262
Apple Watch 8 (18)	0.38±0. 19(14)	0.57±0. 2(4)	0.1072	0.15±0. 06(2)	0.45±0. 19(16)	0.0490	0.43±0. 21(13)	0.4±0.2 1(5)	0.8432	0.49±0. 21(9)	0.35±0. 17(9)	0.1471
Oura Ring 3 (31)	0.48±0. 13(25)	0.52±0. 11(6)	0.4699	0.44±0. 2(9)	0.5±0.0 6(22)	0.1852	0.49±0. 09(23)	0.48±0. 19(8)	0.9234	0.47±0. 09(9)	0.49±0. 13(22)	0.6314

The number in the parenthesis indicates the number of participants tested with each device. Mean ± standard deviation of macro-averaged F1 scores across participants. The bold values represent statistical significance ($p < 0.05$). Abbreviations: CLC, Clionic Life Center; AHI, Apnea-Hypopnea Index.