

SUPPLEMENTARY INFORMATION

Evaluation of Unity 1.5 T MR-Linac plan quality in patients with prostate cancer

Shohei Tanaka¹, Noriyuki Kadoya¹, Miyu Ishizawa², Yoshiyuki Katsuta¹, Kazuhiro Arai¹, Haruna Takahashi¹, Yushan Xiao¹, Noriyoshi Takahashi¹, Kiyokazu Sato³, Ken Takeda², Keiichi Jingu¹

¹Department of Radiation Oncology, Tohoku University Graduate School of Medicine, Sendai, Japan

²Department of Radiological Technology, School of Health Sciences, Faculty of Medicine, Tohoku University, Sendai, Japan

³Radiation Technology, Tohoku University Hospital, Sendai, Japan




FIGURE S1 Box plots for monitor units (MU), optimization time, and delivery time achieved using various segment numbers with 7 beams.

TABLE S1 Gantry angle settings for each number of beams

Number of beams	Gantry angle (°)
5	0, 72, 144, 216, 288
7	0, 51, 102, 153, 204, 255, 306
9	0, 50, 75, 100, 150, 210, 260, 285, 310
16	5, 25, 45, 65, 85, 105, 136, 157, 203, 224, 255, 275, 295, 315, 335, 355
20	6, 24, 42, 60, 78, 96, 114, 136, 153, 170, 190, 207, 224, 246, 263, 280, 297, 314, 331, 348

TABLE S2 Optimization parameters for low-risk and moderate-risk patients

		Manu al	Weig ht	Minimum volume (%)	Reference dose (Gy)	Power law exponent	Shrink margin (cm)	Isoconstr aint
Structure name	Cost function							
	Target penalty	✓	3	99				36.5
	Quadratic							
		✓	1		37			1
PTV–Rectum	overdose							
	Quadratic							
		✓	2		38			0.02
	overdose							
PTV/Rectum	Target penalty	✓	1	99.9				34.5
	Quadratic							
		✓	0.2		35.1		0	0.2
Rectum	overdose							
	Serial	✓	1			12	0.3	24

	Serial	✓	5		10	0.6	14
	Quadratic						
	overdose	✓	0.5	36.8			0.04
Bladder	Quadratic						
	overdose	✓	0.5	32.63		0	0.4
	Serial	✓	2		10	0.6	17
	Quadratic						
Left femoral head	overdose	✓	0.3	14		0	0.1
Right femoral head	Quadratic						
	overdose	✓	0.3	14		0	0.1
Urethra PRV	Quadratic						
	overdose	✓	0.3	37.5			0.02

	Quadratic				
	✓	0.5	36.25	0.1	0.02
	overdose				
Body					
	Quadratic				
	✓	0.3	18.17	1.5	0.5
	overdose				

PRV, planning organ at risk volume; PTV, planning target volume; PTV/Rectum, overlap between PTV and rectum.

TABLE S3 Optimization parameters for high-risk patients

Structure name	Cost function	Man ual	Weig ht	Minimum volume (%)	Reference dose (Gy)	Power law exponent	Shrink margin (cm)	Isoconstr aint
Urethra PRV	Quadratic overdose	✓	1		39.2			0.02
CTV	Target penalty	✓	1.5	99.5				41
PTV–Rectum–Ureth ra PRV	Quadratic overdose	✓	2	95				36.3
PTV/Rectum	Target penalty	✓	3		42			0.02
Rectum	Target penalty	✓	1	99				34.4
PTV/Rectum	Target penalty	✓	1	99.9				34.5
Rectum	Quadratic	✓	0.5		36		0	0.1

		overdose				
	Serial	✓	1		12	0.3
	Serial	✓	5		10	0.6
<hr/>						
	Quadratic					
		✓	0.3	39	0.1	0.05
	overdose					
Bladder	Quadratic					
		✓	0.5	36	0.1	0.1
	overdose					
	Serial	✓	2		10	0.6
<hr/>						
	Quadratic					
Left femoral head		✓	0.3	14	0	0.1
	overdose					
	Quadratic					
Right femoral head		✓	0.3	14	0	0.1
	overdose					
<hr/>						

	Quadratic				
Prostate + 1.5 cm	✓	0.5	36.25	0.1	0.1
	overdose				

	Quadratic				
Body	✓	0.3	18.75	1.5	0.3
	overdose				

CTV, clinical target volume; PRV, planning organ at risk volume; PTV, planning target volume; PTV/Rectum, overlap between PTV and rectum.

TABLE S4 Average dose parameters for the target and organ at risk (OAR) using various segment numbers with 7 beams for low- and high-risk patients

The average dose parameters for target and OAR										
Low-risk patients						High-risk patients				
Dosimetric parameters			p value			Dosimetric parameters			p value	
60 segments	100 segments	150 segments	60 segments vs. 100 segments	100 segments vs. 150 segments		60 segments	100 segments	150 segments	60 segments vs. 100 segments	100 segments vs. 150 segments
PTV-Rectum										
D98% (Gy)	35.87	35.88	35.90	0.26	0.01	D98% (Gy)	35.33	35.39	35.37	< 0.01
Dmax (Gy)	38.60	38.35	38.32	< 0.01	0.14	Dmax (Gy)	42.77	42.52	42.54	< 0.01
PTV/Rectum										
D98% (Gy)	34.82	34.44	34.63	0.99	0.18	D98% (Gy)	34.83	34.67	34.54	0.72
Rectum										
V36Gy (cc)	0.04	0.03	0.02	0.02	0.32	V36Gy (%)	0.06	0.05	0.05	0.01
V32.6Gy (%)	1.25	1.15	1.11	< 0.01	0.06	V32.6Gy (%)	0.60	0.54	0.53	0.01
V29Gy (%)	2.93	2.84	2.78	0.15	0.14	V29Gy (%)	1.85	1.77	1.74	0.12
V25.3Gy (%)	4.97	4.93	4.85	0.21	0.48	V25.3Gy (%)	4.58	4.48	4.47	0.16
V21.7Gy (%)	7.65	7.69	7.55	0.37	0.12	V21.7Gy (%)	10.48	10.41	10.46	0.40
V18.1Gy (%)	12.18	12.37	12.20	0.79	0.19	V18.1Gy (%)	24.40	24.59	24.80	0.78
Bladder										
V37Gy (cc)	1.80	1.21	1.22	< 0.01	0.52	V37Gy (%)	5.17	4.97	5.01	0.03
V18.1Gy (%)	37.40	36.86	36.72	0.01	0.13	V18.1Gy (%)	36.82	36.73	36.80	0.26
Urethra PRV										
V38Gy (cc)	0.00	0.00	0.00	0.25	0.50	Dmax (Gy)	39.79	39.58	39.63	< 0.01
Left Femoral Head										
Dmax (Gy)	17.87	17.44	17.41	0.03	0.48	Dmax (Gy)	15.41	15.27	15.51	0.31
V14.5Gy (%)	3.61	2.66	2.43	0.02	0.34	V14.5Gy (%)	0.64	0.31	0.33	0.11
Right Femoral Head										
Dmax (Gy)	16.97	16.70	16.74	0.04	0.64	Dmax (Gy)	15.76	15.61	15.61	0.07
V14.5Gy (%)	1.98	1.66	1.63	0.11	0.39	V14.5Gy (%)	0.77	0.56	0.53	0.17
CTV-Urethra PRV										
			D95%			39.39	39.30	39.35	0.99	0.06

CTV, clinical target volume; Dmax, maximal dose; D95%, dose administered to 95% of volume; D98%, dose administered to 98% of volume; OAR, organ at risk; PRV, planning organ at risk volume; PTV, planning target volume; PTV–Rectum–Urethra, PTV minus rectum and urethra structure; PTV/Rectum, common areas of PTV and rectum.