

## Human CXCL10/IP-10 DuoSet ELISA, 15 Plate

Catalog Number: DY266 Expiration: 21 Jun 2026

Lot: P294754

## **Certificate of Analysis**

## **SPECIFICATIONS**

REAGENT	PART NUMBER	# OF VIALS	AMOUNT PER VIAL	WORKING CONCENTRATION	LOT#
Capture	840420	3	120 µg	2.00 μg/mL	ADN0621041
Detection	840421	3	1.20 µg	20.0 ng/mL	AAP1521041
Standard	840422	3	55.0 ng	31.2-2000 pg/mL	1513403
Streptavidin-HRP	893975	3	N/A	40-fold dilution	P289330

## **PREPARATION & STORAGE**

Store unopened kit at 2-8 °C. Do not use past kit expiration date.

REAGENT	PREPARATION	STORAGE OF OPENED/RECONSTITUTED MATERIAL	
Capture	Reconstitute with 0.5 mL of PBS	Store at 2-8 °C for up to 8 weeks or aliquot and store at -20 °C to -70 °C in a manual defrost freezer for up to 12 weeks.*	
Detection	Reconstitute with 1.0 mL of Reagent Diluent		
Standard	Reconstitute with 0.5 mL of Reagent Diluent	Store reconstituted standard at 2-8 °C for up to 8 weeks or aliquot and store at -70 °C for up to 12 weeks.*	
Streptavidin-HRP	Dilute with Reagent Diluent	Store undiluted at 2-8 °C for up to 12 weeks. DO NOT FREEZE.*	

<sup>\*</sup>Provided this is within the expiration date of the kit.

It is hereby certified that the above product has been tested for proper performance and function under our established Quality Control Testing criteria. It is authorized by our Quality Assurance program to be released for sale.

John Beauchamp

Manager, Quality Assurance/Regulatory Affairs This C of A was updated on 7/30/2021, Rev. 0

For Research Use or Further Manufacturing Purposes Only





