



**Figure S1** The signal-to-noise ratio was not different between the dNAV-CMRA and the sNAV-CMRA (dNAV-CMRA: 310; sNAV-CMRA: 313) in the phantom. The purple circles indicate region of interest. dNAV-CMRA, diaphragmatic-navigated coronary magnetic resonance angiography; sNAV-CMRA, self-navigated coronary magnetic resonance angiography.