

SUPPLEMENTARY INFORMATION

Direct Mass Spectrometric Analysis of Brominated Flame Retardants in Synthetic Polymers

Krista Grönlund¹, Ville H. Nissinen¹, Ilkka Rytöluoto², Milad Mosallaei², Joonas Mikkonen², Kirsi Korpijärvi³, Paavo Auvinen¹, Mika Suvanto¹, Jarkko J. Saarinen¹ and Janne Jänis^{1}*

¹Department of Chemistry, University of Eastern Finland, Yliopistonkatu 7, 80130 Joensuu, Finland

²VTT Technical Research Centre of Finland Ltd., Visiokatu 4, 33101 Tampere, Finland

³VTT Technical Research Centre of Finland Ltd., Koivurannantie 1, 40400 Jyväskylä, Finland

*Corresponding author: janne.janis@uef.fi

Table S1. Model compound formulations.

Sample	Br target concentration	Experimental Br concentration (XRF)	Ratio	Target formulation		
	[ppm]	[ppm]	Exp. vs target	Polymer [mg]	BFR [mg]	Sb ₂ O ₃ [mg]
ABS-0	0	0 ± 2	—	10000.00	0.00	0.00
ABS-HBCD-1000	1,000	675 ± 10	0.71	9979.92	13.38	6.69
ABS-HBCD-2500	2,500	1,982 ± 20	0.81	9949.81	33.46	16.73
ABS-HBCD-5000	5,000	4,395 ± 39	0.89	9899.61	66.92	33.46
ABS-HBCD-10000	10,000	8,610 ± 73	0.86	9799.23	133.85	66.92
ABS-TBBPA-1000	1,000	982 ± 12	0.98	9974.48	17.02	8.51
ABS-TBBPA-2500	2,500	2,251 ± 22	0.89	9936.19	42.54	21.27
ABS-TBBPA-5000	5,000	4,484 ± 39	0.90	9872.38	85.08	42.54
ABS-TBBPA-10000	10,000	8,914 ± 77	0.89	9744.76	170.16	85.08
ABS-decaBDE-1000	1,000	916 ± 12	0.92	9981.99	12.00	6.00
ABS-decaBDE-2500	2,500	2,109 ± 21	0.84	9954.99	30.01	15.00
ABS-decaBDE-5000	5,000	4,311 ± 38	0.86	9909.97	60.02	30.01
ABS-decaBDE-10000	10,000	8,750 ± 75	0.87	9819.94	120.04	60.02
HIPS-0	0	9 ± 2	—	10000.00	0.00	0.00
HIPS-HBCD-1000	1,000	656 ± 10	0.66	9979.92	13.38	6.69
HIPS-HBCD-2500	2,500	1,688 ± 18	0.68	9949.81	33.46	16.73
HIPS-HBCD-5000	5,000	3,860 ± 34	0.78	9899.61	66.92	33.46
HIPS-HBCD-10000	10,000	8,282 ± 71	0.83	9799.23	133.85	66.92
HIPS-TBBPA-1000	1,000	877 ± 11	0.88	9974.48	17.02	8.51
HIPS-TBBPA-2500	2,500	2,033 ± 20	0.81	9936.19	42.54	21.27
HIPS-TBBPA-5000	5,000	4,334 ± 38	0.87	9872.38	85.08	42.54
HIPS-TBBPA-10000	10,000	8,609 ± 75	0.86	9744.76	170.16	85.08
HIPS-decaBDE-1000	1,000	793 ± 11	0.79	9981.99	12.00	6.00
HIPS-decaBDE-2500	2,500	2,043 ± 21	0.82	9954.99	30.01	15.00
HIPS-decaBDE-5000	5,000	4,148 ± 37	0.83	9909.97	60.02	30.01
HIPS-decaBDE-10000	10,000	8,773 ± 74	0.88	9819.94	120.04	60.02

Table S2. Utilized parameters in DIP-APCI-MS analysis.

Ion source		Collision cell	
End plate offset	500 V	Collision energy	10.0 eV
Capillary	4000 V	Collision RF	600.0 V _{pp}
Corona	4000 nA	Transfer time	140.0 µs
Nebulizer	2.0 Bar	Pre pulse storage	12.0 µs
Dry gas	5.0 L/min	Flight tube	
Dry temperature	200 °C	Flight tube	9900.0 V
Vaporizer temperature	150-450 °C	Reflector	
Transfer		Decelerator	882.0 V
Deflection 1 Delta	40.0 V	Reflector	2500.0 V
Funnel 1 RF	300.0 V _{pp}	Detector	
Funnel 2 RF	400.0 V _{pp}	TOF	2268.9 V
Multipole RF	250.0 V _{pp}	Corrector	
Quadrupole		Fill	64.7 V
Ion energy	5.0 eV	Extract	-611.1 V
Low mass	100.00 m/z	Lens	-6454.3 V

Table S3. Utilized calibration lists for HIPS and ABS samples.

HIPS		ABS	
Ion formula	m/z	Ion formula	m/z
C ₁₃ H ₉	165.069877	C ₁₁ H ₁₂ N	158.096426
C ₁₄ H ₁₀	178.077702	C ₁₄ H ₁₅ N ₂	211.122975
C ₁₆ H ₁₅	207.116827	C ₁₉ H ₂₀ N	262.159026
C ₂₄ H ₂₄	312.187252	C ₂₂ H ₂₃ N ₂	315.185575
C ₂₅ H ₂₄	324.187252	C ₂₅ H ₂₆ N ₃	368.212124
C ₃₂ H ₃₁	415.242027	C ₃₀ H ₃₁ N ₂	419.248175
C ₃₃ H ₃₂	428.249852	C ₃₃ H ₃₄ N ₃	472.274725
C ₄₁ H ₄₀	532.312453	C ₃₆ H ₃₇ N ₄	525.301274
C ₄₉ H ₄₈	636.375053	C ₄₁ H ₄₂ N ₃	576.337325
C ₅₇ H ₅₆	740.437653	C ₄₄ H ₄₅ N ₄	629.363874
C ₆₅ H ₆₄	844.500253	C ₄₇ H ₄₈ N ₅	682.390423
C ₇₃ H ₇₂	948.562854	C ₅₂ H ₅₃ N ₄	733.426474
		C ₅₅ H ₅₆ N ₅	786.453023
		C ₅₈ H ₅₉ N ₆	839.479572
		C ₆₃ H ₆₄ N ₅	890.515624

Table S4. Summary of selected signals observed in DIP-MS analysis of ABS-decaBDE sample (Br concentration of 1,000 ppm).

Experimental m/z	Theoretical m/z	Error (ppm)	Observed ion	Ion formula	Proposed compound**	Observation time frame (min)***
881.259062*	881.257073	2.26	M ⁺ •	C ₁₂ HBr ₉ O	nonaBDE	5.5–7.0
959.170006*	959.167571	2.54	M ⁺ •	C ₁₂ Br ₁₀ O	decaBDE	5.5–7.0
270.234020	270.234202	-0.67	M ⁺ •	C ₂₀ H ₃₀	B ₅	7.0–8.5
324.281936	324.281153	2.41	M ⁺ •	C ₂₄ H ₃₆	B ₆	7.0–8.5
378.328252	378.328103	0.39	M ⁺ •	C ₂₈ H ₄₂	B ₇	7.0–8.5
432.375130	432.375053	0.18	M ⁺ •	C ₃₂ H ₄₈	B ₈	7.0–8.5
486.421991	486.422003	-0.02	M ⁺ •	C ₃₆ H ₅₄	B ₉	7.0–8.5
540.469117	540.468953	0.30	M ⁺ •	C ₄₀ H ₆₀	B ₁₀	7.0–8.5
156.080700	156.080776	-0.49	[M + H] ⁺	C ₁₁ H ₁₀ N	AS	8.5–10.0
210.115029	210.115150	-0.58	M ⁺ •	C ₁₄ H ₁₄ N ₂	A ₂ S	8.5–10.0
261.151354	261.151201	0.59	M ⁺ •	C ₁₉ H ₁₉ N	AS ₂	8.5–10.0
314.178076	314.177750	1.04	M ⁺ •	C ₂₂ H ₂₂ N ₂	A ₂ S ₂	8.5–10.0
365.213930	365.213801	0.35	M ⁺ •	C ₂₇ H ₂₇ N	AS ₃	8.5–10.0
418.240289	418.240350	-0.15	M ⁺ •	C ₃₀ H ₃₀ N ₂	A ₂ S ₃	8.5–10.0
471.266989	471.266899	0.19	M ⁺ •	C ₃₃ H ₃₃ N ₃	A ₃ S ₃	8.5–10.0
522.302608	522.302951	-0.66	M ⁺ •	C ₃₈ H ₃₈ N ₂	A ₂ S ₄	8.5–10.0

* Instead of monoisotopic peak, the most abundant signal is reported in the table.

** The observed polymer fragments were marked with abbreviation A_aB_bS_s, in which a, b, and s corresponds to the number of acrylonitrile, butadiene, and styrene monomers, respectively.

*** The observation time frame refers to temperature ramp program utilized in the DIP-MS analyses: time frames 5.5–7.0 min, 7.0–8.5 min and 8.5–10.0 min correspond to approximately 350, 400 and 450 °C, respectively.

Table S5. Summary of signals observed in HIPS-decaBDE sample (Br concentration of 1,000 ppm) during the DIP-MS analysis.

Experimental m/z	Theoretical m/z	Error (ppm)	Observed ion	Ion formula	Proposed compound**	Observation time frame(min)***
530.469526	530.469347	0.34	M ⁺ •	C ₃₅ H ₆₂ O ₃	Irganox 1076	5.5–7.0
662.446238	662.445849	0.59	M ⁺ •	C ₄₂ H ₆₃ O ₄ P	Irgafos 168 oxide	5.5–7.0
881.257641*	881.257073	0.64	M ⁺ •	C ₁₂ HBr ₉ O	nonaBDE	5.5–7.0
959.167551*	959.167571	-0.02	M ⁺ •	C ₁₂ Br ₁₀ O	decaBDE	5.5–7.0
270.233973	270.234202	-0.85	M ⁺ •	C ₂₀ H ₃₀	B ₅	7.0–8.5
324.281062	324.281153	-0.28	M ⁺ •	C ₂₄ H ₃₆	B ₆	7.0–8.5
378.328110	378.328103	0.02	M ⁺ •	C ₂₈ H ₄₂	B ₇	7.0–8.5
432.374988	432.375053	-0.15	M ⁺ •	C ₃₂ H ₄₈	B ₈	7.0–8.5
486.421913	486.422003	-0.19	M ⁺ •	C ₃₆ H ₅₄	B ₉	7.0–8.5
207.116834	207.116827	0.03	[M + H] ⁺	C ₁₆ H ₁₅	S ₂	8.5–10.0
312.187281	312.187252	0.09	M ⁺ •	C ₂₄ H ₂₄	S ₃	8.5–10.0
415.241766	415.242027	-0.63	[M + H] ⁺	C ₃₂ H ₃₁	S ₄	8.5–10.0
520.312282	520.312453	-0.33	M ⁺ •	C ₄₀ H ₄₀	S ₅	8.5–10.0

* Instead of monoisotopic peak, the most abundant signal is reported in the table.

** The observed polymer fragments were marked with abbreviations B_b and S_s, in which b and s corresponds to the number of butadiene and styrene monomers, respectively.

*** The observation time frame refers to temperature ramp program utilized in the DIP-MS analyses: time frames 5.5–7.0 min, 7.0–8.5 min and 8.5–10.0 min correspond to approximately 350, 400 and 450 °C, respectively.




Figure S1. Averaged mass spectrum (time frame of 5.5–7.0 min) from DIP-MS measurement of HIPS sample without any BFRs.




Figure S2. Total ion chromatograms from DIP-MS analysis of decaBDE (top), TBBPA (middle) and HBCD (bottom). The utilized APCI vaporizer temperature program is presented as a red line in the upmost chromatogram.




Figure S3. Total ion chromatograms from DIP-MS analysis of HIPS-decaBDE and ABS-decaBDE samples having Br concentrations of 1,000 ppm.




Figure S4. Averaged mass spectrum (time frame of 5.5–7.0 min) from DIP-MS measurement of ABS-TBBPA sample (Br concentration 1,000 ppm).




Figure S5. Averaged mass spectrum (time frame of 5.5–7.0 min) from DIP-MS measurement of ABS-HBCD sample (Br concentration 1,000 ppm).




Figure S6. Averaged mass spectrum (time frame of 5.5–7.0 min) from DIP-MS measurement of HIPS-TBBPA sample (Br concentration 1,000 ppm).




Figure S7. Averaged mass spectrum (time frame of 5.5–7.0 min) from DIP-MS measurement of HIPS-HBCD sample (Br concentration of 1,000 ppm).




Figure S8. Normalized intensity of m/z 959.168 (decaBDE) versus m/z 207.117 (styrene dimer) compared to the mass fraction of decaBDE based on XRF data. The MS data was collected using a small sample size (<0.1 mg) and the error bars represent standard deviation of the mean between five parallel measurements.