

# Development and External Evaluation of a Self-Learning Auto-Segmentation Model for Colorectal Cancer Liver Metastases Assessment (COALA)

## ELECTRONIC SUPPLEMENTARY MATERIAL

**Table S1:** Reconstruction and acquisition parameters used to obtain the CTs.

<b>Parameter</b>	<b>INTERNAL dataset N = 373</b>	<b>EXTERNAL dataset N = 50</b>
Pitch/table speed	0.6 – 1.48 / 19.2 – 182.8 mm	0.6 – 1.08 / 23 – 86.3 mm
X-Ray tube current	157 – 581 mA	90 – 553 mA
kVp	90 – 120 kVp	100 – 120 kVp
Tube rotation speed	0.4 – 0.75 ms	0.5 – 0.8 ms
Slice thickness	3.0 – 5.0 mm	3.0 mm
Width X height	512 – 1024 X 512 – 1024	512 X 512