

Table 3. MR scanning parameters

	176 subjects		34 subjects	
	T1-weighted MPRAGE	BOLD	T1-weighted MPRAGE	BOLD
Field strength	1.5T	1.5T	1.5T	1.5T
TR	9.7 msec	2.5 sec	9.7 msec	3 sec
TE	4 msec	50 msec	4 msec	50 msec
Flip angle	12	90	10	90
Voxel size	1.0 × 1.0 × 1.25 mm	3.75 × 3.75 mm	1.0 × 1.0 × 1.25 mm	3.75 × 3.75mm