

**Table 1**

Determination Results of PFOS by the CPC-ECR colorimetric method and HPLC-MS

(n=3)

C <sub>LO</sub> . PFOS (mg/l)	CPC-ECR (mg/l)	HPLC-MS (µg/l)
100.0	18.30±0.121	17.77±0.003
150.0	59.42±0.022	57.77±0.003
200.1	88.79±0.034	87.50±0.004