

Table S5 Summary of AEs related to the use of the device CZ-901

Disease	Case	Adverse event	Incidence
<i>Malignant</i>			
Bone	7	Tumor rupture 1	14.29% (1/7)
	7	1	14.29
<i>Benign</i>			
Uterine fibroid	328	Burn 3 Severe/prolonged abdomen pain 9	3.66% (12/328)
Bone	8		
	336	12	3.57%
Total	343	13	3.79%