

Table S6 Summary of AEs related to the use of the device 2000

Disease	Case	Adverse event	Incidence
<i>Malignant</i>			
Prostate	34	Hematuria 11	32.35%
	34	11	(11/34) 32.35%
<i>Benign</i>			
Uterine adenomyoma	30		
Hypersplenism	30	Skin burn 1	3.33%
	60	1	(1/30) 1.67%
Total	94	12	12.77%