

Table S7 Summary of AEs related to the use of the device HY2900

Disease	Case	Adverse event	Incidence
Uterine fibroid/adenomyoma	31	Burn 4	12.90% (4/31)
Total	31	4	12.90%