

SUPPLEMENTARY TABLE S3. COMPARISON OF OVERNIGHT SOF-SENSOR WITH PARADIGM PUMP AND ENLITE CONTINUOUS GLUCOSE MONITORING WITH VEO PUMP POINT ACCURACY BASED ON REFERENCE SOURCE

	<i>Number of pairs</i>		<i>Difference^a</i>		<i>RAD^b</i>		<i>ISO^c</i>	
	<i>Sof-Sensor</i>	<i>Enlite</i>	<i>Sof-Sensor</i>	<i>Enlite</i>	<i>Sof-Sensor</i>	<i>Enlite</i>	<i>Sof-Sensor</i>	<i>Enlite</i>
Reference source								
Central lab/YSI	1,171	312	-3 (-18, +10)	-10 (-27, +8)	9% (5%, 19%)	15% (7%, 23%)	78%	72%
GlucoScout	2,342	1,103	+2 (-14, +16)	-13 (-26, 0)	13% (6%, 22%)	14% (6%, 23%)	71%	70%
HemoCue	114	251	-16 (-33,+4)	-31 (-50, -15)	19% (9%, 27%)	24% (15%, 40%)	54%	39%
Calibration of studies								
Capillary	866	661	-7 (-28, +11)	-24 (-43, -3)	13% (6%, 23%)	20% (11%, 32%)	68%	54%
Venous	2,761	1,005	+1 (-12, +15)	-11 (-23, +1)	12% (6%, 21%)	13% (6%, 21%)	75%	73%

^aDifference=continuous glucose monitoring value minus reference value. Data are median values (25th, 75th percentiles).

^bRelative absolute difference (RAD)=absolute difference/reference. Data are median values (25th, 75th percentiles).

^cInternational Organization for Standardization (ISO) criteria are continuous glucose monitoring measurements within ± 15 mg/dL for reference glucose values ≤ 75 mg/dL and within $\pm 20\%$ for reference glucose values >75 mg/dL.