

Supplementary Table S5. Time Course Studied for JC-10 Mitochondria Potential Assay

Time (min)	Average \pm SD (max)	Average \pm SD (min)	Z'-Factor	Assay window
Variation of granularity output with CCCP incubation time (0.8 mM, $n = 12$)				
60	6,946 \pm 166.2	18.0 \pm 16.1	0.92	386
90	4,922 \pm 22.6	18.0 \pm 10.0	0.98	273
120	2,008 \pm 70.0	0.7 \pm 0.6	0.89	3,011