

Additional file 2

Reproducibility of ALBIA-HMGCR

Serum number	Anti-HMGCR level (AU/mL)	Intra-assay variation (within plate) %CV (n)	Inter-assay variation (assay to assay) %CV (n)	Inter-assay variation (batch to batch) %CV (n)
# 1	455 (high)	5 (25)	9 (30)	2.8 (2)
# 2	223 (medium)	5 (25)	11 (30)	4.5 (2)
# 3	29 (low)	13 (25)	10 (30)	5.7 (2)