

SUPPLEMENTARY MATERIAL

**A blinded international study on the reliability of genetic testing
for GGGGCC-repeat expansions in *C9orf72* reveals marked
differences in results among 14 laboratories**

Chizuru Akimoto, Alexander E Volk *et al.*

Supplementary table 1. The results of 78 samples.

RP-PCR results		
All laboratories same results	50 samples (64.1%)	
1 laboratory different result	20 samples (25.6%)	13 samples (16.7%): 1 laboratory Q result 5 samples (6.41%): 1 laboratory FP result 2 samples (2.56%): 1 laboratory FN result
2 laboratories different results	5 samples (6.41%)	4 samples (5.12%): 2 laboratories Q results 1 sample (1.28%): 1 laboratory FN and 1 laboratory Q results
6 laboratories different results	2 samples (2.56%)	1 sample (1.28%): 1 laboratory FN and 5 laboratories Q results 1 sample (1.28%): 6 laboratories Q results
7 laboratories different results	1 sample (1.28%)	1 sample (1.28%): 1 laboratory FN and 6 laboratories Q results
Amplicon-length analysis results		
All laboratories same results	72 samples (92.3%)	
1 laboratory different result	2 samples (2.56%)	1 sample (1.28%): 1 laboratory different peak number 1 sample (1.28%): 1 laboratory Q result
2 laboratories different results	1 sample (1.28%)	1 sample (1.28%): 1 laboratory different peak number and 1 laboratory Q result
3 laboratories different results	1 sample (1.28%)	1 sample (1.28%): 1 laboratory different peak number and 2 laboratories Q results
5 laboratories different results	2 samples (2.56%)	2 samples (2.56%): 4 laboratories different peak number and 1 laboratory Q result
RP-PCR + amplicon-length analysis results		
All laboratories same results	53 samples (67.9%)	
1 laboratory different result	19 samples (24.4%)	16 samples (20.5%): 1 laboratory Q result 2 samples (2.56%): 1 laboratory FP result 1 sample (1.28%): 1 laboratory FN result
2 laboratories different results	3 samples (3.85%)	2 samples (2.56%): 2 laboratories Q results 1 sample (1.28%): 1 laboratory FP and 1 laboratory Q results
5 laboratories Different results	2 samples (2.56%)	1 sample (1.28%): 1 laboratory FN and 4 laboratories Q results 1 sample (1.28%): 5 laboratories Q results
6 laboratories Different results	1 sample (1.28%)	1 sample (1.28%): 1 laboratory FN and 5 laboratories Q results

FP: False positive, FN: False negative, Q: Undecided or unclassified sample.

Supplementary table 2. The GGGGCC-repeat numbers on 78 samples from 32 families.

Sample	Allele 1	Allele 2	Sample	Allele 1	Allele 2
1.1	7	Expanded	16.1	2	Expanded
1.2	6	Expanded	16.3	2	Expanded
2.1	2	Expanded	17.1	2	6
2.2	2	Expanded	17.2	2	5
3.1	5	12	18.1	5	Expanded
3.2	2	Expanded	18.2	2	Expanded
4.1	4	Expanded	19.1	2	2
4.2	4	Expanded	19.2	2	5
4.3	8	Expanded	20.1	5	13
4.4	4	Expanded	20.2	5	6
5.1	2	2	21.1	5	Expanded
5.2	2	2	21.2	5	Expanded
6.1	5	5	22.1	11	Expanded
6.2	2	Expanded	22.2	2	Expanded
6.3	2	Expanded	23.1	8	Expanded
6.4	2	Expanded	23.2	8	Expanded
6.5	23-24	Expanded	24.1	2	Expanded
6.6	8	Expanded	24.2	2	Expanded
6.7	8	Expanded	25.1	2	2
6.8	2	2	25.2	2	8
7.1	2	6	26.1	2	Expanded
7.2	2	5	26.2	2	Expanded
8.1	2	6*	26.3	23-25	Expanded
8.2	2	6*	27.1	2	Expanded
9.1	2	10	27.2	8	Expanded
9.2	2	8	27.3	5	Expanded
10.1	5	Expanded	27.4	5	Expanded
10.2	2	Expanded	28.1	2	7
11.1	2	Expanded	28.2	5	Expanded
11.2	2	Expanded	29.1	2	5
12.1	2	5	29.2	2	6
12.2	2	8	30.1	2	6
12.3	2	5	30.2	2	5
13.1	2	Expanded	31.1	7	Expanded
13.2	2	Expanded	31.2	2	2
14.1	2	2	31.3	2	Expanded
14.2	2	8	32.1	6*	Expanded
15.1	2	6	32.2	2	Expanded
15.2	6	10	32.3	28-32	Expanded

*: mutant allele.

Supplementary table 3A. Total methods of 14 laboratories.

Laboratory	A	B	C	D	E	F	G	H	I	J	K	L	M	N
RP-PCR method														
Primers*	P1, 2, 3	P7, 8, 9	P1, 2, 3	P4, 13	P4, 5, 6	P10, 11, 12	P4, 5, 6	P1, 2, 3	P4, 5, 6	P4, 5, 6	No	P1, 2, 3	P4, 5, 6	P1, 2, 3
PCR reaction*	PR1	PR2	PR1	PR3	PR4	PR5	PR6	PR1	PR7	PR8	No	PR1	PR6	PR9
Primer 1 dose (μM)	1.4	0.2	1.4	0.2	0.4	0.33	1.0	1.4	1.4	2.5	No	1.4	1.0	1.4
Primer 2 dose (μM)	0.7	0.2	0.7	0.2	0.2	0.33	1.0	0.7	0.7	2.5	No	0.7	1.0	0.7
Primer 3 dose (μM)	1.4	0.2	1.4	-	0.4	0.33	1.0	1.4	1.4	0.25	No	1.4	1.0	1.4
Deaza-GTP (mM)	0.18	0.2	0.18	0.2	0.0375	0	0.25	0.18	0.18	0	No	0.18	0.25	0.16
DNA dose (ng/μl)	7.1	5.0	2.14	2.0	20	10.0	12.5	7.1	12.0	3.3	No	7.1	5.0	3.3
PCR protocol*	PP1	PP2	PP3	PP4	PP5	PP6	PP7	PP8	PP7	PP9	No	PP1	PP7	PP1
Extension time (min)	3	8-18	3	3	40 sec	2-10	8-16	3	8-16	8-16	No	3	8-16	3
Total cycle	40	45	40	32	65	35	35	43	35	34	No	40	35	40
Analysis machine*	M1	M2	M2	M1	M1	M1	M1	M1	M3	M1	No	M1	M1	M2
PCR product conc.	10%	10%	25%	20%	0.47%	16.2%	3.3%	0.94%	8.8%	10%	No	21%	4.8%	13.3%
Analysis software*	A1	A2	A1	A1	A1	A5	A2	A2	A2	A2	No	A1	A1	A3
Sensitivity	100%	100%	100%	100%	91.3%	100%	97.8%	93.5%	93.5%	93.5%	71.7%	95.7%	89.1%	93.5%
Specificity	100%	100%	100%	100%	96.8%	87.5%	100%	100%	100%	93.8%	96.9%	100%	100%	87.5%
Unclassified	0%	0%	0%	0%	5.1%	2.6%	0%	3.8%	3.8%	3.8%	14.1%	2.6%	5.1%	9.0%
Amplicon length analysis method														
Primers*	P14, 16	P14, 16	P14, 16	P15, 16	P14, 16	P17, 18	P14, 16	P14, 16	P15, 16	P14, 16	No	N.D.	N.D.	N.D.
PCR reaction*	PR10	PR11	PR12	PR13	PR10	PR14	PR15	PR16	PR13	PR17	No	N.D.	N.D.	N.D.
DNA dose (ng/μl)	2.2	5.0	2.4	4.2	2.0	3.3	3.35	7.1	2.0	3.3	No	N.D.	N.D.	N.D.
PCR protocol*	PP10	PP11	PP12	PP13	PP14	PP15	PP16	PP17	PP18	PP19	No	N.D.	N.D.	N.D.
Extension time (min)	1.5	1	2	2	1.5	1	1	3	2	8-16	No	N.D.	N.D.	N.D.
Total cycle	34	38	30	35	35	33	36	40	35	34	No	N.D.	N.D.	N.D.
Analysis machine*	M1	M2	M2	4% gel	M1	M1	M1	M1	3% gel	M1	No	N.D.	N.D.	N.D.
PCR product conc.	10%	0.5%	25%		0.47%	16.2%	0.17%	0.47%		10%	No	N.D.	N.D.	N.D.
Analysis software*	A1	A2	A1		A1	A5	A2	A2		A2	No	N.D.	N.D.	N.D.

*: Listed in supplementary table 3B and 3C, N.D.: not done. No: no method disclosed.

Supplementary table 3B. List of products and equipment.

Primers for RP-PCR	
P1	FAM-AGTCGCTAGAGGCGAAAGC
P2	tacgcatcccagtttgagacGGGGCCGGGGCCGGGGCCGGGG
P3	tacgcatcccagtttgagacg
P4	FAM-tgtaaaacgacggccagtCAAGGAGGGAAACAACCGCAGCC
P5	caggaaacgctatgaccGGGCCCGCCCCGACCACGCCCGGCCCGGCCCGG
P6	caggaaacgctatgacc
P7	FAM-CAAGGAGGGAAACAACCGCAGCC
P8	ggataacaattcacacaggGGGCCCGCCCCGACCACGCCCGGCCCGGCCCGG
P9	ggataacaattcacacagg
P10	FAM-AGTACTCGCTGAGGGTGAAC
P11	cgtacgcatcccagtttgagaGCCCCGGCCCCGGCCCCGG
P12	cgtacgcatcccagtttgaga
P13	cacgacgttgtaaacgaCCCCGGCCCCGGCCCCGG

Primers for amplicon-length analysis	
P14	FAM-CAAGGAGGGAAACAACCGCAGCC
P15	CAAGGAGGGAAACAACCGCAGCC
P16	GCAGGCACCGCAACCGCAG
P17	FAM-CAGGTGTGGGTTTAGGAGGT
P18	CCAGCTTCGGTCAGAGAAAT
Size standard	
S1	GeneScan 500XL ROX and 500ROX Size Standard (Applied Biosystems, USA)
S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
S3	GeneScan 400HD Rox Size Standard (Applied Biosystems, USA)
S4	GeneScan 600LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	
M1	3730 and 3730xl DNA Analyzer (Applied Biosystems, USA)
M2	3130 and 3130xl Genetic Analyzer (Applied Biosystems, USA)
M3	3500 Genetic Analyzer (Applied Biosystems, USA)
Analysis software	

A1	Peak Scanner Software v1.0 or v2.0 (Applied Biosystems, USA)
A2	GeneMapper Software v4.0 or v4.1 (Applied Biosystems, USA)
A3	Geotyper v4 (Applied Biosystems, USA)
A4	Agarose gel
A5	In-house developed Tracl genotyping software

Supplementary table 3C-1. Concentration of ingredients in RP-PCR.

PCR reaction number	PR1	PR2	PR3	PR4	PR5	PR6	PR7	PR8	PR9
Laboratory	ACHL	B	D	E	F	GM	I	J	N
Qiagen buffer *1	-	-	-	1x	-	1x	-	-	-
FastStart PCR Master *2	1x	-	-	-	-	-	-	-	-
Taq DNA polymerase (units)	-	-	1.8	0.75	-	2.5	1.0	-	-
Other buffer and polymerase	-	*4	*5	-	*6	-	*7	*8	*9
Primer 1 (μ M)	1.4	0.2	0.2	0.4	0.33	1.0	1.4	2.5	1.4
Primer 2 (GGGGCC repeat) (μ M)	0.7	0.2	0.2	0.2	0.33	1.0	0.7	2.5	0.7
Primer 3 (μ M)	1.4	0.2	-	0.4	0.03	1.0	1.4	0.25	1.4
dCTP, dATP, dTTP (each) (mM)	-	0.2	0.2	0.05	-	0.25	0.18	-	-
7-deaza-2-deoxy GTP (mM)	0.18	0.2	0.2	0.0375	-	0.25	0.18	-	0.16
dGTP (mM)	-	-	-	0.0125	-	-	-	-	-
dNTPs (each) (mM)	-	-	-	-	0.5	-	-	-	-
DMSO	7%	5%	5%	-	-	5%	-	-	7%
Betaine (M)	-	-	1.0	-	2.0	1.0	-	1.75	0.83
Q solution *3	1x	-	-	1x	-	-	-	-	-
MgCl ₂ (mM)	0.89	2.0	-	-	-	-	-	-	0.83

*1: *Taq* DNA polymerase (QIAGEN, Germany)

*2: FastStart PCR Master mix (Roche, Switzerland)

*3: Q solution (QIAGEN, Germany)

*4: *Taq* Robust (Kapa Biosystems, USA), Kapa Enhancer 1x.

*5: RG25 buffer (160mM NH₄SO₄, 670mM Tris-HCl (pH 8.8), 1% Tween, 25mM MgCl₂)

*6: Expand Long Template PCR system (Roche, Switzerland), Buffer 2 1x, Titanium *Taq* polymerase 2x.

*7: AccuPrime GC-rich DNA polymerase (Invitrogen, USA), Buffer A 1x.

*8: Extensor Mastermix (Thermo Scientific, USA)

*9: True Allele PCR Premix (Applied Biosystems, USA), PCR master mix 1x.

Supplementary table 3C-2. Concentration of ingredients in amplicon-length analysis PCR.

PCR reaction number	PR10	PR11	PR12	PR13	PR14	PR15	PR16	PR17
Used laboratory	AE	B	C	DI	F	G	H	J
Primer 4 (μM)	0.4	0.2	0.4	0.2	0.3	1.0	0.5	0.1
Primer 5 (μM)	0.4	0.2	0.4	0.2	0.3	1.0	0.5	0.1
dCTP, dATP, dTTP (each) (mM)	-	0.2	-	-	-	0.25	-	-
7-deaza-2-deoxy GTP (mM)	-	0.2	0.2	-	-	0.25	0.18	-
dNTP (each) (mM)	-	-	0.25	-	0.3	-	-	-
DMSO (%)	-	5.0	4.0	-	-	5.0	7.1	1.0
Betaine (M)	-	-	1.0	-	1.0	1.0	-	-
Q solution *1	-	-	-	-	-	-	1x	-
MgCl ₂ (mM)	-	2.0	-	-	-	-	0.89	-
Kit buffer	1x	1x	1x	1x	1x	1x	1x	1x
Kit polymerase (U)	-	0.2	1.25	0.4	0.3	2.5	-	-
Kit and polymerase	*2	*3	*4	*5	*6	*7	*8	*9

*1: Q solution (QIAGEN, Germany)

*2: One *Taq* 2X Master Mix with GC Buffer (New England BioLabs, UK)

*3: KAPA Enhncer and *Taq* Robust (Kapa Biosystems, USA)

*4: AmpliTaq and GeneAmp PCR buffer (Roche, Switzerland)

*5: AccuPrime GC-Rich DNA Polymerase (Invitrogen, USA)

*6: Kapa HiFi Hotstart PCR kit, GC buffer (Kapa Biosystems, USA)

*7: *Taq* DNA Polymerase (QIAGEN, Germany)

*8: FastStart PCR Master mix (Roche, Switzerland)

*9: Megamix Blue (Microzone, UK)

Supplementary table 3C-3. PCR protocols for RP-PCR (PP1-9).

PP1, Laboratory A, L, N			PP2, Laboratory B			PP3, Laboratory C		
Temp	Time	Cycle	Temp	Time	Cycle	Temp	Time	Cycle
95°C	15 min	1	98°C	10 min	1	95°C	8 min	1
94°C	1 min	2	98°C	35 sec	13	95°C	1 min	8
70°C	1 min		65-58°C	2 min	Every cycle	70-56°C	1 min	Every cycle
72°C	3 min		72°C	8 min	-1°C	72°C	3 min	-2°C
94°C	1 min	3	97°C	35 sec	32	95°C	1 min	32
68°C	1 min		58°C	2 min	Every cycle	56°C	1 min	
72°C	3 min		72°C	8-18 min	+20 sec	72°C	3 min	
94°C	1 min	4	72°C	10 min	1	72°C	10 min	1
66°C	1 min		4°C	hold	1	4°C	hold	1
72°C	3 min							
94°C	1 min	5						
64°C	1 min							
72°C	3 min							
94°C	1 min	6	PP4, Laboratory D			PP5, Laboratory E		
62°C	1 min		Temp	Time	Cycle	Temp	Time	Cycle
72°C	3 min		95°C	5 min	1	95°C	5 min	1
94°C	1 min	7	95°C	30 sec	32	95°C	30 sec	49
60°C	1 min		68°C	3 min		70-53.2°C	30 sec	Every cycle
72°C	3 min		72°C	7 min		68°C	40 sec	-0.35°C
94°C	1 min	8	12°C	hold	1	95°C	30 sec	16
58°C	1 min					58°C	30 sec	
72°C	3 min					68°C	40 sec	
94°C	1 min	5				68°C	10 min	1
56°C	1 min					12°C	hold	1
72°C	3 min							
72°C	10 min	1						
4°C	hold	1						
PP6, Laboratory F			PP7, Laboratory G, I, M			PP8, Laboratory H		
Temp	Time	Cycle	Temp	Time	Cycle	Temp	Time	Cycle
98°C	10 min	1	98°C	10 min	1	95°C	5 min	1
97°C	35 sec	10	97°C	35 sec	10	95°C	30 sec	8
53°C	2 min		64°C	2 min		70-56°C	30 sec	Every cycle
68°C	2 min		68°C	8 min		72°C	3 min	-2°C
97°C	35 sec	25	97°C	35 sec	25	95°C	30 sec	35
53°C	2 min	Every cycle	64°C	2 min	Every cycle	56°C	30 sec	
68°C	2-10 min	+20 sec	68°C	8-16 min	+20 sec	72°C	3 min	
68°C	10 min	1	16°C	hold	1	72°C	20 min	1
4°C	hold	1				4°C	hold	1

PCR protocols for RP-PCR (continued).

PP9, Laboratory J

Temp	Time	Cycle
95°C	5 min	1
95°C	30 sec	8
70-56°C	30 sec	Every cycle
72°C	8 min	-2°C
95°C	30 sec	26
56°C	30 sec	Every cycle
72°C	8-16 min	+20 sec
72°C	7 min	1
4°C	hold	1

Supplementary table 3C-4. PCR protocols for amplicon-length analysis (PP10-19).

PP10, Laboratory A			PP11, Laboratory B			PP12, Laboratory C		
Temp	Time	Cycle	Temp	Time	Cycle	Temp	Time	Cycle
98°C	5 min	1	97°C	5 min	1	95°C	5 min	1
97°C	30 sec	10	97°C	30 sec		95°C	1 min	
65-56°C	30 sec	Every cycle	60°C	30 sec	10	56°C	1 min	30
68°C	1.5 min	-1°C	72°C	1 min		72°C	2 min	
97°C	30 sec		97°C	10 sec		72°C	10 min	1
55°C	30 sec	24	60°C	15 sec	28	4°C	hold	1
68°C	1.5 min		72°C	30 sec				
68°C	10 min	1	72°C	10 min	1			
12°C	hold	1	4°C	hold	1			

PCR protocols for Amplicon length analysis (continued).

PP13, Laboratory D			PP14, Laboratory E			PP15, Laboratory F		
Temp	Time	Cycle	Temp	Time	Cycle	Temp	Time	Cycle
P			98°C	5 min	1	95°C	5 min	1
95°C	3 min	1	97°C	30 sec	11	98°C	20 sec	
95°C	30 sec		65-55°C	30 sec	Every cycle	59°C	15 sec	33
68°C	30 sec	35	68°C	1.5 min	-1°C	72°C	1 min	
72°C	2 min		97°C	30 sec		72°C	5 min	1
72°C	10 min	1	55°C	30 sec	24	4°C	hold	1
12°C	hold	1	68°C	1.5 min				
			68°C	10 min	1			
			12°C	hold	1			
PP16, Laboratory G			PP17, Laboratory H			PP18, Laboratory I		
Temp	Time	Cycle	Temp	Time	Cycle	Temp	Time	Cycle
94°C	4 min	1	95°C	4 min	1	95°C	3 min	1
94°C	30 sec		95°C	30 sec	14	95°C	30 sec	
62°C	30 sec	36	70-57°C	30 sec	Every cycle	63°C	30 sec	35
72°C	1 min		72°C	3 min	-1°C	72°C	2 min	
72°C	10 min	1	95°C	30 sec		72°C	10 min	1
			56°C	30 sec	26	16°C	hold	1
			72°C	3 min				
			72°C	10 min	1			
			4°C	hold	1			
PP19, Laboratory J								
Temp	Time	Cycle						
95°C	5 min	1						
95°C	30 sec	8						
70-56°C	30 sec	Every cycle						
72°C	8 min	-2°C						
95°C	30 sec	26						
56°C	30 sec	Every cycle						
72°C	8-16 min	+20 sec						
72°C	7 min	1						
4°C	hold	1						

Supplementary table 3D. Methods of each laboratory.

Laboratory A**RP-PCR**

List

Primer 1	P1	FAM-AGTCGCTAGAGGCGAAAGC
Primer 2 (with GGGGCC repeat)	P2	tacgcatcccagtttgagacgGGGGCCGGGGCCGGGGCCGGGG
Primer 3	P3	tacgcatcccagtttgagacg
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730x/DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR1)

Products	Final conc.
FastStart PCR Master	1x
Primer 1	1.4 µM
Primer 2	0.7 µM
Primer 3	1.4 µM
7-deaza-2-deoxy GTP	0.18 mM
DMSO	7 %
Q solution	1x
MgCl ₂	0.89 mM
DNA	7.1 ng/µl
Total	14.0 µl

FastStart PCR Master (Roche, Switzerland)
Q solution (QIAGEN, Germany).

PCR protocol (PP1)

Temp	Time	Cycle
95°C	15 min	1
94°C	1 min	2-8 *
70-58°C	1 min	
72°C	3 min	
94°C	1 min	5
56°C	1 min	
72°C	3 min	
72°C	10 min	1
4°C	hold	1

* Every 2°C decrease add 1 cycle.

Capillary electrophoresis

PCR product	1.0 µl
Size standard	0.5 µl
Formamide	8.5 µl
Total	10.0 µl

95°C 5min and on ice.
Analyse with DNA analyzer.

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730x/ DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR10)

Products	Final conc.
Primer 4	0.4 µM
Primer 5	0.4 µM
Kit buffer	1x
DNA	2.2 ng/µl
Total	15 µl

Kit buffer: One *Taq* 2x Master Mix with GC Buffer (New England BioLabs, UK).

PCR protocol (PP10)

Temp	Time	Cycle
98°C	5 min	1
97°C	30 sec	10
65-56°C	30 sec	Every cycle -1°C
68°C	1.5 min	
97°C	30 sec	24
55°C	30 sec	
68°C	1.5 min	
68°C	10 min	1
12°C	hold	1

Capillary electrophoresis

PCR product	1.0 µl
Size standard	0.5 µl
Formamide	8.5 µl
Total	10 µl

95°C 5min and on ice.
Analyse with DNA analyzer.

Supple table 3D. Methods of each laboratory

Laboratory B**RP-PCR**

List

Primer 1	P7	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P8	ggataacaatttcacacaggGGGCCCCGCCCGACCACGCCCCCGGC CCCGGCCCCCGG
Primer 3	P9	ggataacaatttcacacagg
Size standard	S2	GeneScan 500 LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M2	3130xl Genetic Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper Software v4.0 (Applied Biosystems, USA)

PCR reaction (PR2)		PCR protocol (PP2)			Capillary electrophoresis
Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H ₂ O=1:1
KAPA Enhancer	1x	98°C	10 min	1	PCR product 2.0 µl Size standard 0.1 µl Formamide 7.9 µl
Primer 1	0.2 µM	98°C	35 sec	13	
Primer 2	0.2 µM	65-58°C	2 min	Every cycle	Total 10.0 µl
Primer 3	0.2 µM	72°C	8 min	-1°C	
dCTP, dATP, dTTP (each)	0.2 mM	97°C	35 sec	32	95°C 3min and analyse with genetic analyzer.
7-deaza-2-deoxy GTP	0.2 mM	58°C	2 min	Every cycle	
DMSO	5%	72°C	8-18 min	+20 sec	
MgCl ₂	2.0 mM	72°C	10 min	1	
Taq Robust	0.2 U	72°C	10 min	1	
DNA	5 ng/µl	72°C	10 min	1	
Total	20 µl	4°C	hold	1	

KAPA Enhancer and *Taq* Robust (Kapa Biosystems, USA). Ramping up and down: 0.5°C/sec

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M2	3130xl Genetic Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper Software v4.0 (Applied Biosystems, USA)

PCR reaction (PR11)		PCR protocol (PP11)			Capillary electrophoresis
Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H ₂ O=1:39
KAPA Enhancer	1x	97°C	5 min	1	Diluted PCR product 2.0 µl Size standard 0.1 µl Formamide 7.9 µl
Primer 4	0.2 µM	97°C	30 sec	10	
Primer 5	0.2 µM	60°C	30 sec	10	Total 10.0 µl
dCTP, dATP, dTTP (each)	0.2 mM	72°C	1 min	28	
7-deaza-2-deoxy GTP	0.2 mM	97°C	10 sec	28	95°C 3min and analyse with genetic analyzer.
DMSO	5%	60°C	15 sec	28	
MgCl ₂	2.0 mM	72°C	30 sec	1	
Taq Robust	0.2 U	72°C	10 min	1	
DNA	5 ng/µl	72°C	10 min	1	
Total	20 µl				

KAPA Enhancer and *Taq* Robust (Kapa Biosystems, USA)

Laboratory C**RP-PCR**

List

Primer 1	P1	FAM-AGTCGCTAGAGGCGAAAGC
Primer 2 (with GGGGCC repeat)	P2	tacgcatcccagtttgagacgGGGGCCGGGGCCGGGGCCGGGG
Primer 3	P3	tacgcatcccagtttgagacg
Size standard	S2	GeneScan 500 LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M2	3130xl genetic analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR1)

Products	Final conc.
FastStart PCR Master	1x
Primer 1	1.4 μ M
Primer 2	0.7 μ M
Primer 3	1.4 μ M
7-deaza-2-deoxy GTP	0.18 mM
DMSO	7 %
Q solution	1x
MgCl ₂	0.89 mM
DNA	2.14 ng/ μ l
Total	28 μl

FastStart PCR Master (Roche, Switzerland)
Q solution (QIAGEN, Germany)

PCR protocol (PP3)

Temp	Time	Cycle
95°C	8 min	1
95°C	1 min	8
70-56°C	1 min	Every cycle
72°C	3 min	-2°C
95°C	1 min	
56°C	1 min	32
72°C	3 min	
72°C	10 min	1
4°C	hold	1

Capillary electrophoresis

PCR product	2.5 μ l
Size standard	0.25 μ l
Formamide	7.25 μ l
Total	10.0 μl
95°C 2min and 4°C. Analyse with genetic analyzer.	

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M2	3130xl genetic analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR12)

Products	Final conc.
Primer 4	0.4 μ M
Primer 5	0.4 μ M
7-deaza-2-deoxy GTP	0.2 mM
dNTPs (each)	0.25 mM
DMSO	4 %
Betaine	1.0 M
GeneAmp PCR buffer	1x
AmpliAmp DNA polymerase	1.25 U
DNA	2.4 ng/ μ l
Total	25.0 μl

AmpliAmp and GeneAmp PCR buffer (Roche, Switzerland)

PCR protocol (PP12)

Temp	Time	Cycle
95°C	5 min	1
95°C	1 min	
56°C	1 min	30
72°C	2 min	
72°C	10 min	1
4°C	hold	1

Capillary electrophoresis

PCR product	2.5 μ l
Size standard	0.25 μ l
Formamide	7.25 μ l
Total	10.0 μl
95°C 2min and 4°C. Analyse with genetic analyzer.	

Laboratory D**RP-PCR**

List

Primer 1	P4	FAM-tgtaaacgacggccagtCAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P13	cacgacgttgtaaacgaCCCCGGCCCCGGCCCCGG
Primer 3	-	none
Size standard	S2	GeneScan 500LIZ Size Standard (Life Technologies, USA)
Analysis machine	M1	3730xL DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v2.0 (Applied Biosystems, USA)

PCR reaction (PR3)

Products	Final conc.
RG25 buffer	1x
Primer 1	0.2 µM
Primer 2	0.2 µM
dCTP, dATP, dTTP (each)	0.2 mM
7-deaza-2-deoxy GTP	0.2 mM
DMSO	5%
Betaine	1.0 M
Taq DNA polymerase	1.8 U
DNA	2.0 ng/µl
Total	20.0 µl

RG25 buffer (160mM NH₄SO₄, 670mM Tris-HCl (pH 8.8), 1% Tween, 25mM MgCl₂)

PCR protocol (PP4)

Temp	Time	Cycle
95°C	5 min	1
95°C	30 sec	32
68°C	3 min	
72°C	7 min	1
12°C	hold	1

Capillary electrophoresis

PCR product	2.0 µl
Size standard	0.5 µl
Formamide	7.5 µl
Total	10.0 µl
95°C 5min and analyse with DNA analyzer.	

Amplicon length analysis

List

Primer 4	P15	CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Analysis system	A4	4 % agarose gel

PCR reaction (PR13)

Products	Final conc.
Primer 4	0.2 µM
Primer 5	0.2 µM
AccuPrime buffer A	1x
AccuPrime Taq	0.4 U
DNA	4.2 ng/µl
Total	30.0 µl

AccuPrime GC-Rich DNA Polymerase (Invitrogen, USA)

PCR protocol (PP13)

Temp	Time	Cycle
95°C	3 min	1
95°C	30 sec	
68°C	30 sec	35
72°C	2 min	
72°C	10 min	1
12°C	hold	1

Laboratory E**RP-PCR**

List

Primer 1	P4	FAM-tgtaaacgacggccagtCAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P5	caggaacagctatgaccGGGCCCCGCCCCGACCACGCCCCGCCCCG GCCCCG
Primer 3	P6	caggaacagctatgacc
Size standard	S1	GeneScan 500ROX Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR4)**PCR protocol (PP5)****Capillary electrophoresis**

Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H ₂ O=1.5:100
Qiagen buffer	1x	95°C	5 min	1	Diluted PCR product 5.0 µl Size standard 0.07 µl H ₂ O 10.0 µl
Primer 1	0.4 µM	95°C	30 sec	49	
Primer 2	0.2 µM	70-53.2°C	30 sec	Every cycle -0.35°C	Total 15.7 µl
Primer 3	0.4 µM	68°C	40 sec		
ATP, CTP, TTP (each)	0.05 mM				95°C 5min and analyse with DNA analyzer.
GTP	0.0125 mM	95°C	30 sec		
7-deaza-2-deoxy GTP	0.0375 mM	58°C	30 sec	16	
Q solution	1x	68°C	40 sec		
Taq DNA polymerase	0.75 U				
DNA	20 ng/µl	68°C	10 min	1	
Total	25.0 µl	12°C	hold	1	

Taq DNA polymerase (QIAGEN,
Germany)

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S1	GeneScan 500ROX Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR10)**PCR protocol (PP14)****Capillary electrophoresis**

Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H ₂ O=1.5:100
Primer 4	0.4 µM	98°C	5 min	1	Diluted PCR product 5.0 µl Size standard 0.07 µl H ₂ O 10.0 µl
Primer 5	0.4 µM	97°C	30 sec	11	
Kit buffer	1x	65-55°C	30 sec	Every cycle -1°C	Total 15.7 µl
DNA	2.0 ng/µl	68°C	1.5 min		
Total	25.0 µl	97°C	30 sec		95°C 5min and analyse with DNA analyzer.
Kit buffer: One Taq 2x Master Mix with GC Buffer (New England BioLabs, UK)		55°C	30 sec	24	
		68°C	1.5 min		
		68°C	10 min	1	
		12°C	hold	1	

Supple table 3D. Methods of each laboratory

Laboratory F**RP-PCR**

List

Primer 1	P10	FAM-AGTACTCGCTGAGGGTGAAC
Primer 2 (with GGGGCC repeat)	P11	cgtacgcatcccagtttgagaGCCCCGCCCCGCCCCGG
Primer 3	P12	cgtacgcatcccagtttgaga
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A5	In-house developed Tracl genotyping software

PCR reaction (PR5)

Products	Final conc.
Buffer 2 (kit)	1x
Primer 1	0.33 μ M
Primer 2	0.33 μ M
Primer 3	0.03 μ M
dNTP (each)	0.5 mM
Betaine	2.0 M
Polymeras (TiTaq 50x)	2x
DNA	10 ng/ μ l
Total	20.0 μl

Expand Long Template PCR system
(Roche, Switzerland)**PCR protocol (PP6)**

Temp	Time	Cycle
98°C	10 min	1
97°C	35 sec	10
53°C	2 min	
68°C	2 min	
97°C	35 sec	25
53°C	2 min	Every cycle
68°C	2-10 min	+20 sec
68°C	10 min	1
4°C	hold	1

Capillary electrophoresis

PCR product	2.0 μ l
Size standard	0.35 μ l
Formamide	10.0 μ l
Total	12.35 μl
95°C 1.5min and analyse with DNA analyzer.	

Amplicon length analysis

List

Primer 4	P17	FAM-CAGGTGTGGGTTTAGGAGGT
Primer 5	P18	CCAGCTTCGGTCAGAGAAAT
Size standard	S4	GeneScan 600LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A5	In-house developed Tracl genotyping software

PCR reaction (PR14)

Product	Final conc.
Primer 4	0.3 μ M
Primer 5	0.3 μ M
dNTPs (each)	0.3 mM
Betaine	1.0 M
Kapa HiFi GC buffer	1x
Polymerase (kit)	0.3U
DNA	3.3 ng/ μ l
Total	15.0 μl

Kapa HiFi Hotstart PCR kit (Kapa Biosystems, USA)

PCR protocol (PP15)

Temp	Time	Cycle
95°C	5 min	1
98°C	20 sec	33
59°C	15 sec	
72°C	1 min	
72°C	5 min	1
4°C	hold	1

Ramping rate: 50%

Capillary electrophoresis

PCR product	2.0 μ l
Size standard	0.35 μ l
Formamide	10.0 μ l
Total	12.35 μl
95°C 1.5min and analyse with DNA analyzer.	

Laboratory G**RP-PCR**

List

Primer 1	P4	FAM-tgtaaacgacggccagtCAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P5	caggaaacagctatgaccGGGCCC GCCCGACCGCCCCGCCCCG GCCCCG
Primer 3	P6	caggaaacagctatgacc
Size standard	S3	GeneScan 400HD Rox Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.0 (Applied Biosystems, USA)

PCR reaction (PR6)		PCR protocol (PP7)			Capillary electrophoresis	
Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H2O=1:5	
Qiagen buffer	1x	98°C	10 min	1		
Primer 1	1.0 µM	97°C	35 sec	10	Diluted PCR product	2.0 µl
Primer 2	1.0 µM	64°C	2 min		Size standard	0.12 µl
Primer 3	1.0 µM	68°C	8 min		Formamide	7.88 µl
dCTP, dATP, dTTP (each)	0.25 mM				Total	10.0 µl
7-deaza-2-deoxy GTP	0.25 mM	97°C	35 sec	25	95°C 2min and analyse with DNA analyzer.	
DMSO	5%	64°C	2 min	Every cycle		
Betaine	1.0 M	68°C	8-16 min	+20 sec		
Taq DNA polymerase	2.5 U					
DNA *1	12.5 ng/µl	Reduce ramping rate up and down to 0.5°C/sec.				
Total	20.0 µl					

Taq DNA polymerase, buffer (QIAGEN, Germany)

*1: Initial DNA concentration ~250 ng/µl.

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S3	GeneScan 400HD Rox Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.0 (Applied Biosystems, USA)

PCR reaction (PR15)		PCR protocol (PP16)			Capillary electrophoresis	
Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H2O=1:180	
Primer 4	1.0 µM	94°C	4 min	1		
Primer 5	1.0 µM	94°C	30 sec	36	Diluted PCR product	3.0 µl
dCTP, dATP, dTTP (each)	0.25 mM	62°C	30 sec		Size standard	0.10 µl
7-deaza-2-deoxy GTP	0.25 mM	72°C	1 min		HiDi Formamide	6.90 µl
DMSO	5%				Total	10.0 µl
Betaine	1.0 M	72°C	10 min	1	95°C 2min and analyse with DNA analyzer.	
Qiagen buffer	1x					
Taq DNA polymerase	2.5 U					
DNA *2	3.35 ng/µl					
Total	20.0 µl					

Taq DNA polymerase, buffer (QIAGEN, Germany)

*2: Initial DNA concentration ~67 ng/µl.

Supple table 3D. Methods of each laboratory

Laboratory H**RP-PCR**

List

Primer 1	P1	FAM-AGTCGCTAGAGGCGAAAGC
Primer 2 (with GGGGCC repeat)	P2	tacgcatcccagtttgagacgGGGGCCGGGGCCGGGGCCGGGG
Primer 3	P3	tacgcatcccagtttgagacg
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.0 (Applied Biosystems, USA)

PCR reaction (PR1)**PCR protocol (PP8)****Capillary electrophoresis**

Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H ₂ O=1:19
FastStart PCR Master	1x	95°C	5 min	1	
Primer 1	1.4 µM	95°C	30 sec	8	Diluted PCR product 2.0 µl
Primer 2	0.7 µM	70-56°C	30 sec	Every cycle	Formamide mix 8.6 µl
Primer 3	1.4 µM	72°C	3 min	-2°C	Total 10.6 µl
7-deaza-2-deoxy GTP	0.18 mM				
DMSO	7 %	95°C	30 sec		Formamide mix=8µl Size standard + 850µl Formamide
Q solution	1x	56°C	30 sec	35	
MgCl ₂	0.89	72°C	3 min		
DNA	7.1 ng/µl				No denaturation. Analyze with DNA analyzer.
Total	28.1 µl	72°C	20 min	1	
		4°C	hold	1	

FastStart PCR Master (Roche,
Switzerland)

Q solution (QIAGEN, Germany)

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730 DNA Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.0 (Applied Biosystems, USA)

PCR reaction (PR16)**PCR protocol (PP17)****Capillary electrophoresis**

Products	Final conc.	Temp	Time	Cycle	Dilute PCR product PCR product:H ₂ O=1:39
Primer 4	0.5 µM	95°C	4 min	1	
Primer 5	0.5 µM	95°C	30 sec	14	Diluted PCR product 2.0 µl
7-deaza-2-deoxy GTP	0.18 mM	70-57°C	30 sec	Every cycle	Formamide mix 8.6 µl
DMSO	7.1 %	72°C	3 min	-1°C	Total 10.6 µl
Q solution	1x				
MgCl ₂	0.89 mM	95°C	30 sec		Formamide mix=8µl Size standard + 850µl Formamide
FastStart PCR master	1x	56°C	30 sec	26	
DNA	7.1 ng/µl	72°C	3 min		
Total	28.0 µl	72°C	10 min	1	No denaturation. Analyze with DNA analyzer.
		4°C	hold	1	

FastStart PCR Master (Roche,
Switzerland)

Q solution (QIAGEN, Germany)

Laboratory I**RP-PCR**

List

Primer 1	P4	FAM-tgtaaacacgacggccagtCAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P5	caggaaacagctatgaccGGGCCCCGCCCGACACGCCCCGGCCCCG GCCCCGG
Primer 3	P6	caggaaacagctatgacc
Size standard	S1	GeneScan 500ROX Size Standard (Applied Biosystems, USA)
Analysis machine	M3	3500 Genetic Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.0 (Applied Biosystems, USA)

PCR reaction (PR7)**PCR protocol (PP7)****Capillary electrophoresis**

Products	Final conc.	Temp	Time	Cycle	PCR product	1.5 µl	
AccuPrime buffer A	1x	98°C	10 min	1	Size standard	0.5 µl	
Primer 1	1.4 µM	97°C	35 sec	10	Formamide	15.0 µl	
Primer 2	0.7 µM	64°C	2 min		25	Total	17.0 µl
Primer 3	1.4 µM	68°C	8 min			Every cycle +20 sec	95°C 5min and analyse with genetic analyzer.
dCTP, dATP, dTTP (each)	0.18 mM	97°C	35 sec	1			
7-deaza-2-deoxy GTP	0.18 mM	64°C	2 min				
Taq DNA polymerase	1.0 U	68°C	8-16 min				
DNA	12 ng/µl						
Total	25.0 µl	16°C	hold				

AccuPrime GC-rich DNA polymerase,
buffer A (Invitrogen, USA)

Ramp rate adjusted to 64% for the
denaturation and extension steps.
Ramp rate adjusted to 25% for the
annealing step.

Amplicon length analysis

List

Primer 4	P15	CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Analysis system	A4	3% agarose gel

PCR reaction (PR13)**PCR protocol (PP18)**

Products	Final conc.	Temp	Time	Cycle
Primer 4	0.2 µM	95°C	3 min	1
Primer 5	0.2 µM	95°C	30 sec	35
Kit buffer	1x	63°C	30 sec	
Kit polymerase	0.4 U	72°C	2 min	
DNA	2.0 ng/µl			1
Total	25.0 µl	72°C	10 min	
Kit: AccuPrime GC-rich DNA polymerase (Invitrogen, USA)		16°C	hold	1

Supple table 3D. Methods of each laboratory

Laboratory J**RP-PCR**

List

Primer 1	P4	FAM-tgtaaacgacggccagtCAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P5	caggaaacagctatgaccGGGCCCCGCCCGACCACGCCCGGCCCGGCCCCGG
Primer 3	P6	caggaaacagctatgacc
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730x/DNA Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.1 (Applied Biosystems, USA)

PCR reaction (PR8)		PCR protocol (PP9)			Capillary electrophoresis	
Products	Final conc.	Temp	Time	Cycle	PCR product	1.0 µl
Extensor Mastermix	1x	95°C	5 min	1	Size standard	0.5 µl
Primer 1	2.5 µM	95°C	30 sec	8	Formamide	8.5 µl
Primer 2	2.5 µM	70-56°C	30 sec	Every cycle	Total	10.0 µl
Primer 3	0.25 µM	72°C	8 min	-2°C	95°C 2min and on ice.	
Betaine	1.75 M	95°C	30 sec	26	Analyse with DNA analyzer.	
DNA	3.3 ng/µl	56°C	30 sec	Every cycle		
Total	15 µl	72°C	8-16 min	+20 sec		
Extensor Mastermix (Thermo Scientific, USA)		72°C	7 min	1		

Amplicon length analysis

List

Primer 4	P14	FAM-CAAGGAGGGAAACAACCGCAGCC
Primer 5	P16	GCAGGCACCGCAACCGCAG
Size standard	S2	GeneScan 500LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730x/ DNA Analyzer (Applied Biosystems, USA)
Analysis software	A2	GeneMapper v4.1 (Applied Biosystems, USA)

PCR reaction (PR17)		PCR protocol (PP19)			Capillary electrophoresis	
Products	Final conc.	Temp	Time	Cycle	PCR product	1.0 µl
Primer 4	0.1 µM	95°C	5 min	1	Size standard	0.5 µl
Primer 5	0.1 µM	95°C	30 sec	8	Formamide	8.5 µl
DMSO	1%	70-56°C	30 sec	Every cycle	Total	10.0 µl
Megamix Blue	1x	72°C	8 min	-2°C	95°C 2min and on ice.	
DNA	3.3 ng/µl	95°C	30 sec	26	Analyse with DNA analyzer.	
Total	15 µl	56°C	30 sec	Every cycle		
Megamix Blue (Microzone, UK)		72°C	8-16 min	+20 sec		
		72°C	7 min	1		
		4°C	hold	1		

Laboratory L**RP-PCR**

List

Primer 1	P1	FAM-AGTCGCTAGAGGCGAAAGC
Primer 2 (with GGGGCC repeat)	P2	tacgcatcccagtttgagacgGGGGCCGGGGCCGGGGCCGGGG
Primer 3	P3	tacgcatcccagtttgagacg
Size standard	S1	GeneScan 500XL ROX Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730x/DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR1)

Products	Final conc.
FastStart PCR Master	1x
Primer 1	1.4 µM
Primer 2	0.7 µM
Primer 3	1.4 µM
7-deaza-2-deoxy GTP	0.18 mM
DMSO	7 %
Q solution	1x
MgCl ₂	0.89 mM
DNA	100 ng
Total	28.0 µl

FastStart PCR Master (Roche, Switzerland)
Q solution (QIAGEN, Germany)

PCR protocol (PP1)

Temp	Time	Cycle
95°C	15 min	1
94°C	1 min	2-8 *
70-58°C	1 min	
72°C	3 min	
94°C	1 min	5
56°C	1 min	
72°C	3 min	
72°C	10 min	1

* Every 2°C decrease add 1 cycle.

Capillary electrophoresis

PCR product	2.0 µl
Size standard	0.5 µl
HiDi Formamide	7.0 µl
Total	9.5 µl

95°C 3min and analyse with DNA analyzer.

Amplicon length analysis...not performed.

Laboratory M**RP-PCR**

List

Primer 1	P4	FAM-tgtaaacgacgcccagtCAAGGAGGGAAACAACCGCAGCC
Primer 2 (with GGGGCC repeat)	P5	caggaaacagctatgaccGGGCCCCGCCCCGACCACGCCCCGGCCCCG GCCCCGG
Primer 3	P6	caggaaacagctatgacc
Size standard	S2	GeneScan 500 LIZ Size Standard (Applied Biosystems, USA)
Analysis machine	M1	3730x/DNA Analyzer (Applied Biosystems, USA)
Analysis software	A1	Peak Scanner Software v1.0 (Applied Biosystems, USA)

PCR reaction (PR6)		PCR protocol (PP7)			Capillary electrophoresis	
Products	Final conc.	Temp	Time	Cycle	PCR product	
Qiagen buffer	1x	98°C	10 min	1	PCR product	1.0 µl
Primer 1	1.0 µM	97°C	35 sec	10	Size standard	0.1 µl
Primer 2	1.0 µM	64°C	2 min		Formamide	19.9 µl
Primer 3	1.0 µM	68°C	8 min		Total	21.0 µl
dCTP, dATP, dTTP (each)	0.25 mM	97°C	35 sec	25	95°C 5min and analyse with DNA analyzer.	
7-deaza-2-deoxy GTP	0.25 mM	64°C	2 min	Every cycle		
DMSO	5%	68°C	8-16 min	+20 sec		
Betaine	1.0 M	Ramping up and down: 0.5°C/sec				
Taq DNA polymerase	2.5 U					
DNA	5.0 ng/µl					
Total	20.0 µl					

Taq DNA polymerase, buffer (QIAGEN, Germany)

Amplicon length analysis...not performed.

Supple table 3D. Methods of each laboratory

Laboratory N**RP-PCR**

List

Primer 1	P1	FAM-AGTCGCTAGAGGCGAAAGC
Primer 2 (with GGGGCC repeat)	P2	tacgcatcccagtttgagacgGGGGCCGGGGCCGGGGCCGGGG
Primer 3	P3	tacgcatcccagtttgagacg
Size standard	S1	GeneScan 500ROX Size Standard (Applied Biosystems, USA)
Analysis machine	M2	3130 Genetic Analyzer (Applied Biosystems, USA)
Analysis software	A3	Genotyper v4 (Applied Biosystems, USA)

PCR reaction (PR9)		PCR protocol (PP1)			Capillary electrophoresis	
Products	Final conc.	Temp	Time	Cycle	PCR product	2.0 µl
True Allele PCR premix	1x	95°C	15 min	1	Size standard	0.5 µl
Primer 1	1.4 µM	94°C	1 min	2-8 *	Formamide	12.5 µl
Primer 2	0.7 µM	70-58°C	1 min		Total	15.0 µl
Primer 3	1.4 µM	72°C	3 min		95°C 2min and analyse with DNA analyzer.	
7-deaza-2-deoxy GTP	0.16 mM	94°C	1 min	5		
DMSO	7%	56°C	1 min			
Betaine	0.83 M	72°C	3 min	1		
MgCl ₂	0.83 mM	72°C	10 min			
DNA	3.3ng/µl	4°C	hold	1		
Total	30.0 µl					
True Allele PCR premix (Applied Biosystems, USA)		* Every 2°C decrease add 1 cycle. Ramp speed: 1°C/sec				

Amplicon length analysis...not performed.

**Supplementary table 4. The definitions of an expansion of the GGGGCC-repeat in
C9orf72 as used in 13 laboratories.**

GGGGCC-repeat number (4 laboratories)

- > 24 repeats
- ≥ 30 repeats
- > 30 repeats (2 laboratories)

RP-PCR shape (2 laboratories)

Clear saw-tooth sloping pattern

GGGGCC-repeat number + RP-PCR shape (1 laboratory)

>30 repeats + Stutter profile in RP-PCR

GGGGCC-repeat number + RP-PCR shape + Amplicon-length analysis result (1 laboratory)

Saw-tooth pattern over 460 bp (30 repeats) in RP-PCR + One amplicon in amplicon-length analysis

RP-PCR shape + Amplicon-length analysis result (1 laboratory)

Saw-tooth pattern in RP-PCR + One amplicon in amplicon-length analysis

RP-PCR shape + Amplicon-length analysis result + Southern blot result (4 laboratories)


- Saw-tooth pattern with concave shape in RP-PCR + One amplicon in amplicon-length analysis + Expanded band in Southern blot
 - One amplicon in amplicon-length analysis + Saw-tooth profile in RP-PCR + Expanded band in Southern blot
 - One amplicon in amplicon-length analysis + Long stutter in RP-PCR + Expanded band in Southern blot (2 laboratories)
-

Supplementary table 5. Summary of the Southern blot methods of three laboratories.


	Laboratory I	Laboratory II	Laboratory III
Labelling	Radioisotope	Radioisotope	DIG system
Dose of gDNA	10 µg	10 µg	6-10 µg
Probe primer-F	TGACACACCAAGCGTCA TCT	TATTAAGGTTCGCACAC GCTA	AGAACAGGACAAGTTGCC
Probe primer-R	CACTGTGGGAGCAGTGT CAT	GACGGCTGACACACCAA G	AACACACACCTCCTAAAC C
Probe length	1000bp	210 bp	241 bp
Restriction enzyme	Xba I and Hind III	Xba I and Hind III	Xba I

Supplementary Figure 1a,b,c,d

A



Amplicon-length analysis



129bp: Peak of wild type allele with 2 repeats.
131bp: Peak of mutant allele with 6 repeats.

B


Sequence

WT CGCGCTAGGGGCGGGGCGGGGCGTGGTCGGGGCGGGCCCGGGGCGGGCCCGGGG


Mut. CGCGCTAGGGGCGGGGCGGGGCGGGGCGGGGCGGGGCGGGGCGGGGCGGGCCCGGGG

1 2 deletion 15bp

1 2 3 4 5 6 insertion 17bp



Supplementary figure 1a.b.c.d (cont.)



Supplementary figure 1. Detailed analyses of the results of the three samples with mutations.


(A) The Rp-PCR and amplicon length analysis of sample 8.1. In RP-PCR there are four peaks before the wild-type allele peak at 290 bp. In amplicon-length analysis there are two peaks at 129bp (wild-type allele) and 131bp (the mutant allele), respectively and the two peaks are close together.

As a consequence, four laboratories interpreted this sample as having one amplicon.

- (B) The sequence of sample 8.1. The wild-type allele has two GGGGCC-repeats and the mutant allele has six GGGGCC-repeats with a complex 15 deletion/17bp insertion mutation. Sample 8.1 and Sample 8.2 both carries this mutation.
- (C) The RP-PCR and amplicon length analysis results of sample 32.1. In RP-PCR there are four peaks before the 290bp peak (mutant allele) and a saw-tooth pattern peaks (expanded allele). In amplicon-length analysis there is one peak at 131bp (mutant allele) with 6 repeats.
- (D) The sequence of 32.1 illustrating the mutant allele which has six repeats with a complex 15bp deletion/17bp insertion mutation.




Supplementary Figure 2

Laboratory I




Supplementary Figure 2 (cont.)

Laboratory II



Supplementary Figure II (cont.)

Laboratory III



Supplementary Figure 2. The Southern blot results in three laboratories.

Laboratory III ran out of DNA from one individual (sample 31.1). This laboratory therefore only

performed SB on 77 samples.

PC: Positive control, NC: negative control, +: sample with large repeat expansion, -: sample without large repeat expansion.