

Table S4. Unadjusted outcomes at follow-up (using paced AV delay).

	Nominal (n=1596)	AV Adjusted (n=2666)	P-value (vs. Nominal)	VV Adjusted (n=583)	P-value (vs. Nominal)	AV and VV adjusted (n=873)	P-value (vs. Nominal)
Follow-up, days	948.60±326.91	984.09±323.14	0.0009	938.98±341.18	0.5576	955.82±331.95	0.6133
PCT activity	6.79±4.14	7.59±4.51	<.0001	6.95±4.49	0.5092	7.27±4.10	0.0279
Change from baseline	-0.06±1.34	0.01±1.56	0.2002	-0.05±1.60	0.8760	-0.01±1.46	0.4799
HRV SDANN	72.56±20.71	73.40±19.16	0.4978	74.84±22.65	0.2459	72.55±21.71	0.9932
Change from baseline	0.61±9.54	0.98±10.21	0.5543	0.47±9.86	0.8896	0.58±10.14	0.9700
HRV footprint	28.81±9.27	28.90±8.84	0.8631	29.26±10.38	0.6116	28.83±9.71	0.9751
Change from baseline	0.10±3.68	0.21±3.89	0.6623	0.61±4.46	0.2015	-0.11±4.71	0.5107
HRV mean HR	77.35±8.22	76.95±8.06	0.4361	76.16±10.25	0.1498	77.01±9.44	0.6071
Change from baseline	-0.29±3.52	-0.52±3.59	0.3510	-0.51±3.92	0.5714	0.03±5.05	0.2985
AF burden > 5%	29.9%	17.9%	<.0001	27.8%	0.3759	19.3%	<.0001
Change from baseline	4.00%±22.0%	4.00%±17.0%	0.2460	2.00%±18.0%	0.3681	4.00%±18.0%	0.9394
LV pacing < 95%	17.1%	17.8%	0.5753	18.0%	0.6214	19.4%	0.1626
Change from baseline	-1.00%±14.0%	0.00%±14.0%	0.8868	-1.00%±12.0%	0.0214	-2.00%±15.0%	0.1597
ATP therapy	28.4%	28.7%	0.8827	30.4%	0.3832	33.1%	0.0158
Shock during follow-up	12.8%	13.6%	0.4948	13.0%	0.9061	16.5%	0.0128
Death	19.7%	18.3%	0.2600	20.9%	0.5393	19.1%	0.7158

Numbers are % or mean ± standard deviation.

AV: atrioventricular timing; VV: ventricular-to-ventricular timing; ATP: anti-tachycardia pacing; AF: atrial fibrillation; LV: left-ventricular; HRV: heart rate variability; SDANN: standard deviation of the average NN intervals; HR heart rate.