

Additional file 2: Table S2A: Intra-assay variability of antigen-specific IFN γ responses ranging between 1-10 SFC/well^a

Donor	Antigen	Replicate 1	Replicate 2	Replicate 3	Mean	SD	%CV
S12	BZLF1	2	12	8	7	5.03	68.63
S23	BZLF1	11	6	3	7	4.04	60.62
S28	BZLF1	8	9	5	7	2.08	28.39
S11	EBNA3A	4	3	2	3	1.00	33.33
S15	EBNA3A	4	8	7	6	2.08	32.87
S24	EBNA3A	3	8	4	5	2.65	52.92
S30	EBNA3A	10	8	4	7	3.06	41.66
Mean						2.85	46.28

Additional file 2: Table S2B: Intra-assay variability of antigen-specific IL2 responses ranging between 1-10 SFC/well^a

Donor	Antigen	Replicate 1	Replicate 2	Replicate 3	Mean	SD	%CV
S17	EBNA3A	2	9	6	6	3.51	61.97
S24	EBNA3A	10	6	9	8	2.08	24.98
Mean						2.80	43.48

Additional file 2: Table S2C: Intra-assay variability of antigen-specific IFN γ +IL2 responses ranging between 1-10 SFC/well^a

Donor	Antigen	Replicate 1	Replicate 2	Replicate 3	Mean	SD	%CV
S12	BZLF1	2	1	6	3	2.65	88.19
S13	BZLF1	2	7	5	5	2.52	53.93
S24	BZLF1	1	7	1	3	3.46	115.47
S11	EBNA3A	7	2	3	4	2.65	66.14
S14	EBNA3A	7	7	3	6	2.31	40.75
S15	EBNA3A	1	5	3	3	2.00	66.67
S24	EBNA3A	7	2	4	4	2.52	58.08
Mean						2.59	69.74

^a Values represent the number of detected antigen-specific IFN γ (table S2A), IL2 (table S2B), and IFN γ +IL2 (table S2C) SFC/2x10⁵ PBMC (stimulated with 1 μ g/ml BZLF1 or EBNA3A peptide pools) (data is not background subtracted) in the IFN γ /IL2 FluoroSpot assay; SD = standard deviation. CV = coefficient of variation.