

	A	B	C	D	E	F	G	H	I
1		Protocol Program Sheet	Program 1: Rinsing out cryoprotectant and incubating NaBH followed by rinsing						
2		Creation Date:							
3		Protocol number of steps:	13	Total 18 minutes in MW					
4		Protocol Description:	NABH						
5		Load cooler set point:							
6		Vacuum set point:							
7		Vacuum cycle Vacuum time:							
8		Vacuum cycle Vent time:							
9		Stirrer Speed:							
10	Step #	Description	User Prompt (on/off)	Time (Hr:min:sec)	Power (Watts)	Temp (-C)	Load Cooler (off/auto/on)	Vacuum/Bubbler Pump (off/bub/vac cycle/vac on/vap)	SteadyTemp pump (on/off)
11	1	Rinse PBS 1	On	0: 1: 0	150	35	Off	Off	Off
12	2	Rinse PBS 2	On	0: 1: 0	150	35	Off	Off	Off
13	3	Rinse PBS 3	On	0: 1: 0	150	35	Off	Off	Off
14	4	Rinse PBS 4	On	0: 1: 0	150	35	Off	Off	Off
15	5	Rinse PBS 5	On	0: 1: 0	150	35	Off	Off	Off
16	6	NaBH on	On	0: 3: 0	150	35	Off	Off	Off
17	7	NaBH off	Off	0: 2: 0	0	35	Off	Off	Off
18	8	NaBH on	On	0: 3: 0	150	35	Off	Off	Off
19	9	Rinse PBS 1	On	0: 1: 0	150	35	Off	Off	Off
20	10	Rinse PBS 2	On	0: 1: 0	150	35	Off	Off	Off
21	11	Rinse PBS 3	On	0: 1: 0	150	35	Off	Off	Off
22	12	Rinse PBS 4	On	0: 1: 0	150	35	Off	Off	Off
23	13	Rinse PBS 5	On	0: 1: 0	150	35	Off	Off	Off

	J
1	
2	
3	
4	
5	
6	
7	
8	
9	
10	SteadyTemp temp (°C)
11	30
12	30
13	30
14	30
15	30
16	30
17	30
18	30
19	30
20	30
21	30
22	30
23	30