

	A	B	C	D	E	F	G	H	I
1		<b>Protocol Program Sheet</b>	Program 5: Tyramide Signal Amplification with Biotinyl Tyramine						
2		Creation Date:							
3		Protocol number of steps:	7	Total 11 minutes in MW					
4		Protocol Description:	TSA						
5		Load cooler set point:							
6		Vacuum set point:							
7		Vacuum cycle Vacuum time:							
8		Vacuum cycle Vent time:							
9		Stirrer Speed:							
10	<b>Step #</b>	<b>Description</b>	<b>User Prompt (on/off)</b>	<b>Time (Hr:min:sec)</b>	<b>Power (Watts)</b>	<b>Temp (-C)</b>	<b>Load Cooler (off/auto/on)</b>	<b>Vacuum/Bubbler Pump (off/bub/vac cycle/vac on/vap)</b>	<b>SteadyTemp pump (on/off)</b>
11	1	BT on	On	0: 3: 0	150	35	Off	Off	Off
12	2	BT off	Off	0: 1: 0	0	35	Off	Off	Off
13	3	BT on	On	0: 3: 0	150	35	Off	Off	Off
14	4	Rinse 1	On	0: 1: 0	150	35	Off	Off	Off
15	5	Rinse 2	On	0: 1: 0	150	35	Off	Off	Off
16	6	Rinse 3	On	0: 1: 0	150	35	Off	Off	Off
17	7	Rinse 4	On	0: 1: 0	150	35	Off	Off	Off

	J
1	
2	
3	
4	
5	
6	
7	
8	
9	
10	SteadyTemp temp (°C)
11	30
12	30
13	30
14	30
15	30
16	30
17	30